



Hammer, Steven J and Good, Daniel W and Scanlan, Paul and Palacio-Torralba, Javier and Phipps, Simon and Stewart, Grant D and Shu, Will and Chen, Yuhang and McNeill, S Alan and Reuben, Robert L (2017) Quantitative mechanical assessment of the whole prostate gland ex vivo using dynamic instrumented palpation. Proceedings of the Institution of Mechanical Engineers, Part H: Journal of Engineering in Medicine, 231 (12). pp. 1081-1100. ISSN 0954-4119 , <http://dx.doi.org/10.1177/0954411917734257>

This version is available at <https://strathprints.strath.ac.uk/62577/>

Strathprints is designed to allow users to access the research output of the University of Strathclyde. Unless otherwise explicitly stated on the manuscript, Copyright © and Moral Rights for the papers on this site are retained by the individual authors and/or other copyright owners. Please check the manuscript for details of any other licences that may have been applied. You may not engage in further distribution of the material for any profitmaking activities or any commercial gain. You may freely distribute both the url (<https://strathprints.strath.ac.uk/>) and the content of this paper for research or private study, educational, or not-for-profit purposes without prior permission or charge.

Any correspondence concerning this service should be sent to the Strathprints administrator: strathprints@strath.ac.uk

Quantitative mechanical assessment of the whole prostate gland *ex vivo* using dynamic instrumented palpation

¹SJ Hammer, ^{2,3}D W Good, ¹P Scanlan, ¹J Palacio-Torralba ^{2,3}S Phipps, ³G D Stewart, ¹W Shu, ¹Y Chen, ^{2,3}S A McNeill, ¹RL Reuben

¹Institute for Mechanical, Process and Energy Engineering, School of Engineering and Physical Sciences, Heriot Watt University, Edinburgh, UK.

²Dept. of Urology, Western General Hospital, Edinburgh UK

³Edinburgh Urological Cancer Group, Institute of Genetics and Molecular Medicine, University of Edinburgh, UK.

Abstract

An instrumented palpation sensor, designed for measuring the dynamic modulus of tissue *in vivo*, has been developed and trialled on *ex vivo* whole prostate glands. The sensor consists of a flexible membrane sensor/actuator with an embedded strain gauge and is actuated using a dynamically varying airflow at frequencies of 1 and 5Hz. The device was calibrated using an indentation stiffness measurement rig and gelatine samples with a range of static modulus similar to that reported in the literature for prostate tissue. The glands were removed from patients with diagnosed prostate cancer scheduled for radical prostatectomy and the stiffness was measured within 30 minutes of surgical removal. Each prostate was later examined histologically in a column immediately below each indentation point and graded into one of four groups; normal, BPH, cancerous, and mixed cancer and BPH. Eleven prostates were assessed using multiple point probing and the complex modulus at 1Hz and 5Hz was calculated on a point-by-point basis. The device yielded values of quasi-static modulus of 15 ± 0.5 kPa and dynamic modulus of 20 ± 0.5 kPa for whole prostates and a sensitivity of up to 80% with slightly lower specificity was achieved on diagnosis of prostate cancer using a combination of mechanical measures. This assessment did not take into account some obvious factors such as edge effects, overlap and clinical significance of the cancer, all of which would improve performance. The device, as currently configured, is immediately deployable *in vivo*. A number of improvements are also identified which could improve the sensitivity and specificity in future embodiments of the probe.

Keywords: prostate, palpation, dynamic stiffness, tissue assessment, trans-rectal sensor

1. Introduction

Prostate cancer (PCa) is the most common type of cancer affecting men (1). Early diagnosis of PCa is critical in the management of its treatment and particular recent attention has focused on distinguishing between indolent and aggressive cancers, which may involve monitoring using several

1
2
3 modes over a period of months or years. Current methods of diagnosis of PCa include; digital rectal
4 examination (DRE), a prostate-specific antigen (PSA) blood test (2), prostate biopsy, trans-rectal
5 ultrasonography (TRUS) (3), trans-rectal sonoelastography (TRSE) (4) and MRI (5). DRE is currently
6 used widely for diagnosis and screening of PCa and the current work is aimed at enhancing the
7 reproducibility and quantification of the finding of a DRE by instrumenting it (iDRE). In conventional
8 DRE, a lubricated and gloved finger is inserted into the rectum and the posterior surface of the
9 prostate is palpated to identify any compliance anomalies which may indicate the presence of PCa
10 (which essentially results in lack of differentiation of the cells, thus altering the mechanical
11 properties). The prostate can also be affected by benign prostatic hyperplasia (BPH) (a proliferation
12 of the glandular component of the prostate leading to enlargement and internal pressure), a
13 condition which is also known to affect its mechanical properties. Because of its subjectivity, it is
14 difficult to use DRE to provide any monitoring, which would require comparison of assessments
15 made over a period of time, perhaps even by different practitioners. However, its ease of
16 deployment and relatively low invasiveness make it an ideal candidate for regular monitoring with
17 an immediate result.
18
19
20
21

22 Elevated levels of PSA in a blood test may indicate the presence of PCa but suffers from a high rate
23 of false positives, which may lead to overdiagnosis and overtreatment (6). In a prostate biopsy,
24 several core samples are removed from the prostate using a hollow needle. The microstructure of
25 the core tissue is examined microscopically and any disease present is evaluated using the Gleason
26 severity grade (7, 8). The disadvantage of biopsy is that, even when guided using ultrasound or other
27 imaging technique, the needles may miss the affected tissue entirely, especially when any cancerous
28 lesions present are small. Prostate biopsy is invasive and can lead to complications and there is some
29 motivation to provide better pre-biopsy information than is currently available. TRUS and TRSE
30 involve ultrasound waves being directed into the prostate *via* a trans-rectal transducer/sensor giving
31 a map related to the stiffness of the prostate. In sonoelastography, an external compression is
32 additionally applied to the prostate through the rectal wall using a fluid-filled cover attached to the
33 probe, with images being captured in the compressed and uncompressed states to produce a
34 relative stiffness map (9, 10). Shear-wave sonoelastography is a developing technique which uses a
35 compressive acoustic pulse and beam correlation techniques to track the motion of shear waves
36 through tissue (11). Several commercially available clinical ultrasound scanners are available with
37 these ultrasonic techniques although their chief disadvantage is that some types of prostate disease
38 cannot be reliably distinguished due to stiffness artefacts (12). Another image-based stiffness
39 measurement technique is Magnetic Resonance Elastography (MRE) which measures wave field
40 motion through tissue that is induced by an external actuator (13). MRE studies of the prostate have
41 been performed *in vivo* with actuation either *via* the urethra (14,15), the rectum (16) or the pubic
42 bone (17) and *ex vivo* studies (18) on whole prostate glands have also been carried out.
43
44
45
46
47
48

49 Despite this range of available diagnostic methods and devices, there is an acknowledged clinical
50 need (19) for complementary methods, at least at the early stages. A relatively easily deployable
51 probe providing reliable, reproducible and diagnostically significant information would make a
52 significant difference to the treatment and staging of prostate cancer. The main access point for
53 mechanical assessment of prostate cancer is *via* the rectum and some devices have already been
54 reported for *in vivo* trans-rectal use, one example (20) using a balloon inserted trans-rectally and
55 inflated, pressing against the prostate. The shape and texture of the posterior surface was recorded
56 at two different pressures using a camera, the change in volume giving a stiffness map. A number of
57
58
59
60

1
2
3 others have already developed point-probe devices and have measured the prostate stiffness *ex vivo*
4 on prostate slices (21–24) and on prostate chips removed during trans-urethral prostate resection
5 (25, 26). Although there is some broad agreement on the absolute values of elastic properties
6 between these studies, it is also clear that there is significant within-patient and between-patient
7 variability and that the test conditions (pre-strain, strain rate and specimen size) all have a significant
8 effect on the values measured (and reported) (27). Nevertheless, there is a strong body of evidence
9 that prostate condition can be assessed mechanically and that, at the very least, contrast between
10 diseased and normal areas can be obtained for a given patient.
11
12

13
14 In order to tackle patient variability and to use values of elastic modulus for supporting numerical
15 (e.g. FE) modelling, it is necessary to determine the relationship between the tissue structure and its
16 mechanical properties. The tissue structure can be quantified in a number of different ways, perhaps
17 the most useful, in the case of the prostate, being in terms of the histological elements (the
18 glandular acinae with their associated epithelial lining and surrounding stromal tissue) or by the
19 distribution and grade of cancer foci. Earlier attempts at this based on point-probing of prostate
20 slices (27) have indicated that pattern recognition techniques are required even with the
21 heterogeneity exhibited in two-dimensions.
22
23

24
25 Mechanical measurements on whole prostates are relatively rare. Åstrand *et al.* (28) have used a
26 tactile sensor equipped with a PZT element to demonstrate contrast between cancerous and non-
27 cancerous areas in a single human prostate sample. The method uses changes in resonance
28 frequency of the piezoelectric element as the sensor is pressed against the surface and yields a
29 “stiffness parameter” as opposed to a value of elastic modulus. Although the range of resonance
30 frequency is not mentioned in their paper, it is likely to be in the range of a few tens of kHz, making
31 the stiffness more akin to an ultrasonic (wave propagation) measurement than a mechanical
32 palpation. Most relevant to the current work, Kim *et al.* (29) have mechanically assessed whole
33 prostates post-laparoscopic prostatectomy using a simulation of the *in vivo* environment, addressing
34 such issues as the support conditions for the gland, an issue also explored by Palacio-Torrallba *et al.*
35 (30) in an FE model.
36
37

38
39 The work presented here is ultimately aimed at making an *in vivo* assessment, devices for which
40 need to be compact and also need to avoid causing harm to the patient or the operator of the device
41 (31). The device whose measurements are reported here is designed for trans-rectal deployment
42 using dynamic instrumented palpation (32) to generate complex modulus measurements over a
43 range of frequencies. In this paper, results of DIP measurements on *ex vivo* whole prostates are
44 reported to demonstrate the viability of the device for distinguishing tissue quality.
45
46
47

48 2. Methods

49 2.1. Patient cohort

50
51 Patients with a diagnosis of localised prostate cancer were invited to participate in the study for the
52 reason that a contrast could be expected between the two lobes. Diagnosis had been confirmed in
53 all patients through the use of biopsy, DRE, PSA testing, multi-parametric MRI and TRUS and the
54 patients were scheduled for radical prostatectomy. Ethical approval was obtained from the local
55 NHS Medical Ethics Committee [Ref: REC (12/SS/1048)] before the study began. The eleven patients
56
57
58
59
60

1
2
3 whose data is presented gave their informed consent to the study. The patients underwent radical
4 prostatectomy (removal of the whole prostate and seminal vesicles) *via* laparoscopic surgery. The
5 excised whole prostate of each patient was taken from the operating theatre to the laboratory for
6 testing, which was done within 1 hour of removal and took less than 2 hours to complete.
7

8 9 2.2. Dynamic instrumented palpation (DIP)

10 Dynamic instrumented palpation (DIP) involves the measurement of the dynamic stiffness of a
11 material by applying an oscillating displacement and measuring the resulting force or *vice versa* (32).
12 The dynamic response is often important in distinguishing healthy and diseased tissue, the principle
13 being that it measures time-dependent properties and hence reveals viscous components as well as
14 elastic ones (33). The essential experimental variables in DIP are the frequency of oscillation, the
15 amplitude of the applied force or displacement and the mean value of force or displacement around
16 which oscillation is applied. DIP has been used before to measure the stiffness of prostate chips from
17 trans-urethral prostate resection operations (25, 26) and on slices from excised whole prostates (27)
18 using an electromagnetic actuator to deliver an oscillating displacement, a method that is not, of
19 course, deployable *in vivo*.
20
21
22

23 The prototype *in vivo* device used here consists of a flexible silicone membrane that is inflated by
24 pressurised air and has been described elsewhere (34). The device is made up of several laser-cut
25 acrylic layers with engraved channels to allow the entrance of water or air into the membrane
26 chamber. The strain in the membrane from the pressurised fluid is measured using a strain gauge.
27 The membrane consists of two layers of 50 μm thick silicone fixed above the membrane chamber
28 using transfer adhesive. The strain gauge is sandwiched between the two layers of membrane to
29 ensure its adhesion to the membrane under dynamic load conditions, and to prevent condensation
30 build up inside the protective sheath used during examinations from causing a short circuit. The
31 membrane is dynamically actuated using a closed air system using an actuator with another driving
32 membrane inside. The driving membrane is actuated using a controlled release of compressed air
33 which causes it to expand, compressing the air inside the closed system and causing the membrane
34 mounted to the device to expand. When the compressed air is exhausted, the driving membrane
35 relaxes and the pressure in the closed system is reduced, leading to the deflation of the device
36 membrane. An image of the probe end of the device used in testing is shown in **Error! Reference
37 source not found..** As part of the development of the device the membrane mechanical response
38 was characterised and its time dependent behaviour not found to be significant at the frequencies of
39 interest [35]. The calibration tests, which were carried out before and after use of the device, did not
40 indicate a significant variation in membrane response with repeated cycling.
41
42
43
44
45
46

47 The pressure behind the membrane in the closed system is measured with a proprietary transducer.
48 In the *ex vivo* tests, the device is applied to the prostate surface using an XYZ stage which also serves
49 to position the device on the prostates. The applied contact force is measured using a proprietary
50 force sensitive resistor (FSR).
51

52 Data from all three sensors were captured as synchronous time series using DAQ hardware mounted
53 in a custom built instrument case, which was connected to a PC running National Instruments
54 LabVIEW. The DAQ hardware was also used to control the solenoid valve, which, in turn, controlled
55 the input of pressurised air into the actuator. A typical actuation waveform and resulting strain
56 gauge waveform are shown in **Error! Reference source not found..**
57
58
59
60

Typically, data were captured for each probe point at a sampling rate of 1000Hz over around 20 seconds and the signals averaged to produce a mean value, a mean amplitude and a phase at the fundamental actuation frequency for each of the strain and pressure. A relative stiffness map for each prostate was calculated based on three characteristics of the strain gauge and flow pressure sensor signals: the mean ratio (MR), amplitude ratio (AR) and the phase difference (PD).

Mean ratio was calculated using

$$MR = \frac{\bar{P}_{flow}}{\bar{\epsilon}}, \quad (1)$$

where \bar{P}_{flow} is the mean pressure measured by the flow sensor, and $\bar{\epsilon}$ is the mean strain, related to the displacement of the membrane into the tissue. The mean ratio thus gives a measure of the quasi-static stiffness of the prostate tissue at that point.

Amplitude ratio was calculated using

$$AR = \frac{\hat{P}_{flow}}{\hat{\epsilon}}, \quad (2)$$

where \hat{P}_{flow} is the amplitude of the pressure measured by the flow sensor and is related to the oscillating force, and $\hat{\epsilon}$ is the amplitude of the strain, so that AR gives a relative measure of the magnitude of the dynamic stiffness, with the phase difference completing the description of the dynamic response. The contact force (and hence depth of indentation before the membrane is oscillated) does not figure in the calculation of the stiffness map, although it needs to be recorded since both the dynamic and static stiffness can, especially in soft materials, vary significantly with mean stress or strain.

2.3. Calibration

In the interest of designing a probe that is deployable with minimum patient discomfort, it was necessary to adopt a “soft” design for the sensor and actuator. Such an approach requires careful calibration to ensure that the static and dynamic properties of the actuator can be de-convolved from the tissue response and also to enable a conversion of stiffness (force per unit displacement) to modulus (stress/strain).

Accordingly, the DIP device was run back-to-back with a calibrated loading frame (Biomomentum Mach-1, www.biomomentum.com) specifically designed for viscoelastic specimens (**Error! Reference source not found.**), measuring the dynamic and quasi-static stiffness of a range of gelatine samples. The gelatine samples were made with 3g, 4g, 5g, 6g, 7g or 8g of gelatine powder, dissolved in 40ml of boiling water and left overnight in a refrigerator at 4°C to solidify. On the day of testing, the gel samples were removed from the sample container and applied to the XYZ measurement stage (for DIP device measurement) or the measurement stage found on the Mach-1.

The gelatine samples were chosen to have a range of stiffnesses and applied forces similar to those measured on the prostates. For each indentation, the 6mm diameter hemispherical probe tip of the

Mach-1 was advanced to the corresponding depth below the surface of the gelatine sample. Each dynamic actuation used a 0.5mm amplitude sinusoidal displacement waveform. This amplitude is 0.2mm less than the mean height of the 6mm diameter membrane used on the DIP device when it is dynamically inflated (35), but allows the Mach-1 to operate at its maximum actuation frequency of 5Hz. A reduced range of indentation depths was used for some samples to avoid damage to the gel sample. Actuation frequencies of 1 and 5Hz were used on all samples and an extended range was used for the membrane to assess its dynamic performance. The stiffness measurements (in gf/mm) from the Mach-1 were converted into values of elastic modulus, E , using the Hertz-Sneddon equation

$$E = \frac{3}{4} \cdot \frac{(1 - \nu^2) \cdot f_z}{\sqrt{R} \cdot \sqrt{\delta_z^3}}, \quad (3)$$

where f_z is the reaction force (N), R is the radius of the indenter (mm), δ_z is the indentation depth (mm) and ν is Poisson's ratio (taken to be 0.499) (36, 37).

Error! Reference source not found. shows that the static modulus increases with gel strength and does not vary significantly with frequency, as expected. The static modulus also increases a little with pre-indentation, suggesting a slightly stiffening stress-strain curve, apart from the initial stage. The dynamic modulus (**Error! Reference source not found.**) is generally higher than the static modulus and increases more markedly with indentation depth. The change with frequency is again small. The values obtained for the phase difference did not vary in any systematic way with depth, frequency or gel strength and the mean was close to zero in all measurements, although with a scatter of around 1 radian (Figure 6), indicating that the calibration material has no measurable phase response at the frequencies used.

The DIP device was pressed onto the gel samples using indentations of between 0 and 5 mm at 1 mm intervals covering the range used on the *ex vivo* prostate specimens. For each gelatine sample, the stiffness measured using the DIP device was plotted against the elastic modulus calculated from the Mach-1 results. A linear fit was applied, yielding calibration equations which could be used to determine the dynamic modulus ($E_{dynamic}$) (Figure 7) and the quasi-static modulus (E_{static}) (Figure 8) from the measured force and displacement in the DIP device. For the dynamic modulus (Figure 7), the calibration is quite reliable at 5Hz, but less so at 1Hz, where higher gel strengths and deeper pre-indentation levels (i.e. higher reaction forces) lead to outliers. For the static modulus (Figure 8) the correlation is similarly poor for both frequencies and there is a distinct stratification in terms of gel strength, probably again reflecting non-linear behaviour in the membrane under higher forces. The phase does not require calibration between the two methods of measurements since it is related to time. However, it does reveal the viscous behaviour of the DIP probe against which any phase difference measured on the prostate samples can be set. As with the Mach 1, the phase measured with the DIP device showed no systematic relationship to gel strength or depth of penetration. Plotted against frequency (with the effects of depth and gel strength embedded in the scatter), Figure 9, reveals a tendency for the phase to increase with frequency. The values of phase are considerably larger for the DIP device than they are for the Mach 1. It might be noted that the negative values recorded at 15Hz are likely to be due to the 2π ambiguity in the processing algorithm. Overall, the calibrations show that amplitude ratio is likely to be more reliable way of

1
2
3 assessing prostate material using the DIP probe than is a static modulus, since it behaves more
4 reliably under small changes in displacement.
5
6
7
8
9

10 2.4. Measurement protocol

11 Point measurement on excised prostates was carried out using a manual XYZ positioning stage
12 (**Error! Reference source not found.**10) with a working volume of 100 x 100 x 100 mm, large enough
13 to allow probing of the majority of prostates encountered. The DIP device was covered with a sterile
14 sheath and fixed to the stage using a screw clamp, while the prostate rested on the flat surface of
15 the stage with its anterior surface downwards. The stiffness of the prostate was measured at up to
16 35 locations, spaced around 6mm apart, on the posterior surface, the position being recorded using
17 the scales on the XYZ stage. At each probe point, the device was advanced by 3mm **in the Z-direction**
18 after touching the surface and oscillated at 1Hz and 5Hz for around 10 cycles. The process was
19 repeated for depths of 5mm and 8mm. The pressure and strain were used, as before, to determine
20 mean ratio, amplitude ratio and phase and the corresponding values converted to moduli using the
21 calibration equations described in the previous section.
22
23
24
25

26 **It should be emphasised that this protocol was somewhat idealised in order to provide a comparison**
27 **with the calibration measurements and to register the probe points with the histological analysis.**
28 **Such a protocol could not be reproduced easily *in vivo* and the implications of support conditions,**
29 **attitude of the probe relative to the gland, number of probe points and test duration are discussed**
30 **in Section 4.**
31
32
33
34
35

36 2.5. Histological analysis

37 After testing, each prostate was subjected to histological examination to determine the type and
38 distribution of tissue throughout the gland. The positions of the surgical clips seen in Figure 10 were
39 noted on the pathology report after processing according to International Society of Urological
40 Pathology guidelines. The H&E-stained transverse sections corresponding to the rows where
41 stiffness measurements had been made were imaged using a microscope and classified using image
42 analysis software (Image Pro Premier, Media Cybernetics Inc., <http://www.mediacy.com/>). Four
43 main tissue classifications were used for each location: 'Normal' (no visible disease of any kind
44 present), 'BPH' (benign prostatic hyperplasia nodules were present), 'Cancer' (a prostate cancer
45 tumour was identified) and 'Mixed' (both prostate cancer and BPH were identified).
46
47
48

49 Figure 11 shows a typical probing sequence on the posterior surface along with an example of the
50 corresponding transverse histological section. The probe points were set apart by the nominal
51 diameter of the probe and so the entire posterior surface was covered without significant overlap.
52 Each probe point corresponded to a column of diameter 6mm and the depth of the prostate and the
53 assessment referred to above applied to the entire column. Probing was carried out at frequencies
54 of 1Hz and 5Hz and at depths of 3mm (Depth A), 5mm (Depth B) and 8mm (Depth C). As can be seen
55 for the example section shown in Figure 11, the depth of the prostate can vary with position as can
56
57
58
59
60

1
2
3 the depth of different histological features, so different depths may reveal different aspects of the
4 mechanical behaviour more clearly.
5
6

7 8 **3. Results**

9 10 **3.1 Data summary from 11 prostates**

11 Eleven prostate glands were measured *ex vivo* using a single build of the DIP sensor, whose
12 calibration was described earlier. A total of 359 points were assessed under six conditions (two
13 frequencies and three depths), yielding a mechanical feature vector of between 6 and 18 elements,
14 depending on whether the depth is stratified. The histological classification (4 target) of the
15 positions measured on these samples is shown Table 1, which also gives a simpler classification in
16 terms of presence or absence of cancer (2 target).
17
18

19 Figure 12 shows boxplots of quasi-static modulus, dynamic modulus and phase difference, classified
20 by histology for the shallowest indentation (Depth A). It is clear that variations of the expected type
21 are present, e.g. that cancer is stiffer than normal or BPH tissue and that it exhibits smaller phase
22 differences. Table 2 shows the values of modulus and phase difference for the 11 patients, classified
23 in two ways, normal tissue vs. cancer (excluding the “mixed” group), and benign (i.e. normal plus
24 BPH) vs. malignant (i.e. cancer plus mixed). Besides showing the mean and standard deviation for
25 each mechanical feature, Table 2 also shows the result of a *t*-test between the two classifications, *P*-
26 values less than 0.05 normally being taken to indicate that the differences between the mean values
27 are statistically significant.
28
29
30

31 From Figure 12 and Table 2, it is clear that, using one of the three depths and the limited range of
32 frequency studied, only some of the mechanical indicators are able to distinguish cancer from
33 normal tissue over a range of patients and also between benign and malignant conditions. However,
34 there is no obvious explanation as to why some indicators are better than others, requiring a more
35 detailed analysis of patient-specific, location-specific and device-specific effects.
36
37
38

39 **3.2 In-patient variation (effect of heterogeneity)**

40 A typical example of the variation in quasi-static modulus, dynamic modulus and phase difference
41 for a single patient is shown alongside the corresponding histology slices in Figure 13. Even with this
42 simplified presentation, it is possible to discern patterns in the mechanical indicators that reveal the
43 presence of the area of cancer, highlighted in dark colour.
44
45

46 Because of the complexity of the mechanical information (six features at three depths) and the
47 heterogeneity of the column, even using the simplified classification, a pattern-recognition tool was
48 used to assess in-patient sensitivity (i.e. proportion of points with disease which return a positive
49 result from the mechanical indicators) and specificity (i.e. proportion of points without disease
50 which return a negative result from the mechanical indicators). A proprietary algorithm (Matlab)
51 consisting of a two-layer feed-forward Artificial Neural Network (ANN) was used, incorporating
52 sigmoid neurons in the hidden layer and “softmax” neurons in the output layer. The mechanical data
53 at each point were input as a feature vector and the target consisted of a two- or four-element
54 vector with zeroes in all but the correct classification (Table 1).
55
56
57
58
59
60

1
2
3 In each ANN run, 10 iterations of training (using scaled conjugate back-propagation) were used, each
4 time with the data set being divided randomly into three parts, 70% for training, 15% for validation
5 and 15% for testing. This was chosen to give a reasonable compromise between training and stability
6 and allowed comparison between the various choices of feature vector and data set described
7 below.
8

9
10 In total, 35 probe points were assessed on this prostate, yielding 35 different target vectors of two
11 elements or four elements (as described above). With this number, it was not effective to use a 4-
12 target output vector as the performance in validation and test could vacillate depending upon
13 whether or not some of the classifications were actually represented in the validation or test set.
14 Even with the 2-element target, the performance of the network over all categories was found to be
15 a more reliable indicator of the presence (or absence) of a pattern than relying on the test
16 performance alone.
17

18
19 Figure 1 shows the ANN Confusion Matrix for the base case, where the six-element feature vectors
20 consisting of dynamic modulus, static modulus and phase angle at each of the frequencies were
21 considered as independent measurements for each of the depths, i.e. 3×35 (105) input vectors.
22 Looking first at the "All" confusion matrix, it can be seen that, over training, validation and test 40 of
23 the 59 PCa targets were correctly classified (i.e. 68% sensitivity) and 29 of the 46 benign targets were
24 correctly classified (i.e. 63% specificity). In the following sensitivity analysis, the overall ANN
25 performance (square at bottom right of each 3×3 array) is used for comparison, the percentage
26 correct for all runs and for the test set being recorded in Table 3. Comparing Figure 14 with Table 3
27 highlights the difficulty with the relatively small number of data sets, where the performance of the
28 validation set and the test set are rather different, the test set indicating that the success rate is
29 about the same as guessing and the validation set somewhat better than the overall.
30
31

32
33 It might be noted that Figure 13 only shows 28 of the 35 measurement points on this particular
34 patient. This is because one histological slice was a little fragmented and spread over a number of
35 slides, making it difficult to present in such an intuitive way as the four other sections (although the
36 histology could be measured fairly reliably). If the base case is re-run on the 28 measurement points
37 (i.e. $3 \times 28 = 84$ input vectors), the overall performance (74% sensitivity and 73% specificity) and the
38 test performance are both somewhat better. Looking at Figure 13, the good performance on the 28
39 points is remarkable, given the evident heterogeneity in depth and coverage of the test column by
40 the cancer-affected areas, not to mention the variation in column height and support conditions.
41
42

43
44 Returning to the base case (35 points), the remaining rows of Table 3 show the effect of selecting
45 which mechanical indicators go into the feature vector. Considering both the "All" and "Test" sets, it
46 seems that performance is generally made worse (or, at least, not any better) by the selection of
47 particular elements of the 6-element vector. The exception to this is the static modulus whose
48 presence in the feature vector seems to reduce sensitivity and specificity.
49

50
51 Although these results are encouraging, several other factors have not been taken into
52 consideration, which, if they had, would almost certainly improve sensitivity and specificity. First of
53 all, there is a pronounced edge effect (most evident at points 7, 8, 21 and 22 in Figure 14) where the
54 modulus is low, most likely due to the probe engaging on a downward-sloping surface but also due
55 to the reduced thickness of the tissue and poorer support on the anterior surface. Also, because the
56 probed volume extends beyond the radius of the membrane, results in one column are likely to be
57
58
59
60

1
2
3 affected by the adjacent columns, so that a cancer-free column adjacent to a column containing
4 cancer may well give a different mechanical result to a cancer-free column with adjacent cancer-free
5 columns. Finally, no shading has been applied for the clinical significance of the cancer such that
6 points 6 and 23, say, would both be classified as "cancer".
7
8
9

10 3.3 Between-patient variation (patient specificity)

11 Across the cohort, some prostates only contained two types of tissue classification while others
12 contained all four. Some prostates contained small multi-focal tumours while others contained a
13 single large tumour. Thus, it might be expected that applying similar pattern-recognition techniques
14 as applied to Patient 21 across the whole cohort would allow an assessment of the extent to which
15 the mechanical features were indicative of fundamental differences in tissue structure associated
16 with malignancy or other disease conditions. Figure 15 shows the confusion matrix corresponding to
17 Figure 14 (cancer vs. benign using the six-element feature vector), this time for all 359 points listed
18 in Table 1, measured at each of the three depths. For the "All" confusion matrix, the sensitivity is
19 60% and the specificity 54%, both being rather worse than for the Patient 21 data (see lines 1 and 2,
20 Table 4).
21
22
23

24 Because of the larger number of samples, it is possible to concatenate the feature vector to contain
25 the measurements at all three depths, which is more realistic for an actual *in vivo* intervention.
26 Figure 16 shows the relevant confusion matrices trained onto the two targets (cancer and benign)
27 and the four targets listed in Table 1. For the "All" confusion matrix (Row 3, Table 4), the sensitivity
28 and specificity are a bit closer to that for the Patient 21 data. As for Patient 21, removing the static
29 modulus improves diagnostic performance significantly (Row 4, Table 4) although not to as high a
30 level as when considering the nearest equivalent for Patient 21 only (Row 5, Table 4). For the four
31 targets, it is possible to assess sensitivity to any disease (including BPH) and sensitivity to the
32 presence of cancer separately and, as can be seen in Table 4. Here, both sensitivities and, indeed the
33 specificity, are a relatively poor 52%, although the numbers of samples are rather small, even in the
34 "all" matrix. If, on the other hand, all points with a BPH or "Mixed" classification are removed, the
35 sensitivity again improves although the specificity is not quite as good.
36
37
38
39
40
41
42
43
44

45 4. Discussion

46 4.1. DIP device performance

47 The DIP device described here was able to measure the complex modulus of excised whole prostates
48 in an efficient way. The device was deployed at around 35 positions on each prostate gland giving
49 complete coverage of the posterior surface (that accessible trans-rectally) at a spatial resolution
50 limited only by the membrane diameter of 6mm. The soft actuator made minimal marking on the
51 prostate gland, leaving specimens that could still be successfully used for histological analysis. The
52 speed of the measurement protocol also meant that prostates could be completely assessed within
53 an hour to reduce drying out of the tissue, thus minimising potential changes in the mechanical
54 behaviour due to fluid loss over time. The speed of measurement per point and the relative softness
55
56
57
58
59
60

1
2
3 of the probe means that the device configuration is suitable for trans-rectal deployment to measure
4 the stiffness of the prostate gland *in vivo*.
5

6 From the calibration studies and from the ANN analysis, it is clear that the quasi-static modulus was
7 the least discriminating of the three mechanical features and its exclusion from the feature vector
8 either improved sensitivity and specificity or, at least, made no difference. It is thought that this is
9 due to medium time (over a few tens of seconds) “fade” in the actuator associated with small leaks,
10 and that a few changes to the design and data processing would help to improve the quality of this
11 measurement while still having the probe deployable *in vivo*. The two frequencies chosen (1Hz and
12 5Hz) appeared to give complementary information in the feature vector and, although the device is
13 capable of delivering higher frequencies, these were not used in the current study to maintain the
14 calibration. There is a potential ambiguity in phase angle as the deployment of the device did not
15 make it convenient to establish a time zero. Nevertheless, the phase features contain useful
16 diagnostic information, although the absolute value of the phase is not known. A wider range of
17 frequencies that allow a pattern of phase vs frequency to be established will add to the diagnostic
18 power of the device.
19
20
21
22

23 4.2. Stiffness measurements compared to other published values

24 The advantage of the current probe is that it returns values that can be traced to measurement
25 standards and so allows the use of data from other investigators to be used, provided that the
26 measurement conditions are clearly known. The use of relatively low amplitude dynamic probing
27 allows information on time constants to be obtained efficiently and the option of a variety of probed
28 depths allows potential non-linearity in the stress-strain response to be assessed as well as allowing
29 the probed volume to be varied. A final, and immensely powerful, advantage of making calibrated
30 measurements *in vivo* is that, in conjunction with detailed mechanical modelling and/or simulation,
31 it allows the development of structure-property relationships whose use can enhance iDRE and can,
32 furthermore, lead to other applications, for example in surgery and in different types of soft tissue
33 (37, 38).
34
35
36

37 The literature on prostate stiffness measurement reports a variety of methods and a variety of
38 different units, some of them device-specific. More importantly, if they are even acknowledged, the
39 effects of specimen and probe size, non-linearity of stress-strain response and temporal (or
40 viscoelastic) behaviour are rarely examined systematically. Table 5 summarises the reported values
41 of elastic or viscoelastic properties of prostate tissue, focussing on low frequency or low strain rate
42 palpation measurements as opposed to those which involve ultrasonic or high frequency wave
43 propagation (of which only one example, 39, is given). The variation in absolute values of modulus
44 (from around 10kPa to a few hundred kPa) is perhaps unsurprising, given the range in specimen size,
45 strain rate and strain level used. One common conclusion amongst investigators, however, is that,
46 within the method they have used, tissue affected by cancer is stiffer than “normal” tissue, although
47 there is considerable disagreement about the magnitude and significance of the difference in the
48 face of other variations in their respective datasets.
49
50
51
52

53 The investigators working with pieces of excised tissue (40) and (25, 26, 27 and 41, all of whom used
54 essentially the same test protocol) tend to report rather higher values of modulus, which could be
55 explained by the expression of water from the samples leading to a more compacted material. In
56 particular, Krouskop *et al* (40) reported a very large increase in elastic modulus in PCa specimens
57
58
59
60

1
2
3 when tested with 4% pre-strain as opposed to 2% pre-strain. In common with (25, 26, 27 and 41)
4 and in contrast with (40), Zhang *et al* (42) observed viscoelastic (or, at least time-dependent)
5 behaviour in their study on 17 prostate cores from the anterior regions of the prostates from 8
6 patients. Although the Kelvin-Voigt viscoelastic model they used is unsuitable for describing the
7 stress-strain behaviour of materials, it is nevertheless helpful in describing the frequency (or strain
8 rate) behaviour of the measurements. The values of dynamic modulus cited in Table 4 are at 150Hz,
9 specifically used because of the authors' interest in applying the results to crawling wave
10 elastography. Using the values they published for the K-V parameters, it is possible to calculate the
11 mean values of dynamic modulus to be 3.61kPa for normal and 8.65kPa for PCa at 1Hz and 5.10kPa
12 for normal and 12.41kPa for PCa at 5Hz, indicating that the effect of frequency over this range is
13 almost as much as the effect of tissue composition. The K-V model implies that there will be no
14 change in phase difference with frequency. It should be emphasised here that the foregoing is based
15 on extrapolation of a model, rather than direct observation. Furthermore, Zhang *et al* fitted the K-V
16 model to relaxation curves over several hundred seconds, and it is well-known (e.g. Palacio-Torralba
17 *et al*, 47) that it is very difficult to fit a single time constant (as demanded by the K-V model) to
18 relaxation curves, the tendency being to obtain a principal time constant approximately equal to the
19 length of observation. The results of the very detailed measurements of Muruyama *et al* (43) make
20 plain the need for a multi-scale approach in understanding the structure-property relationship in
21 such a heterogeneous and complex structure as the prostate gland.
22
23

24
25
26
27 The advantage of working with prostate pieces is that the local tissue condition can be assessed
28 carefully, but, as discussed above, this can give problems in obtaining a realistic (from the point of
29 view of clinical application) measure of mechanical properties. Of the three systematic studies on
30 whole prostates (including the current one) only one (45) reports unambiguously a statistically
31 significant difference between the moduli (however measured) associated with normal and
32 cancerous tissue, the remaining two suggesting the specificity (i.e. the variability within "normal"
33 tissue) to be a greater issue than sensitivity, a general finding echoed in all studies. All three of the
34 whole prostate studies used a similar maximum strain, although only the current study uses an
35 oscillating probe with a relatively small range of strain (for the dynamic modulus). Ahn and Kim (45)
36 used a strain rate that is similar to the current study for the quasi-static modulus, which corresponds
37 to holding of the mean indentation over several oscillations (typically a few seconds), whereas the
38 oscillatory strain rates (approximately amplitude divided by one quarter of a period) are up to two
39 orders of magnitude higher. The strain rates used by Carson *et al* (44) are around an order of
40 magnitude lower than those used by Ahn and Kim, making them the slowest of the three. Two of the
41 three studies (present one and 46) on whole prostates have used the Hertzian contact model to
42 determine modulus from the force and indentation depth, which requires an assumption of a value
43 for Poisson's ratio, although, in later work, Kim *et al* (46) have refined their analysis using an FE
44 model, leading to a slightly reduced value for normal tissue and a more substantial increase for PCa.
45 At face value, then, the current study (quasi-static modulus) is in agreement with the measurement
46 of Ahn and Kim for normal tissue, although in stark disagreement with their value for PCa. The
47 measurements of Carson *et al* yielded a substantially higher elastic modulus, although there is no
48 direct comparator in the other two studies due to the much lower strain rate used. Similarly, the
49 dynamic modulus and PD from the current study have no direct comparator in other whole prostate
50 studies.
51
52
53
54
55
56
57
58
59
60

1
2
3 The key question in whole prostate studies is, however, how best to distinguish between healthy
4 and diseased prostate during an examination, whether using sonoelastography or purely mechanical
5 palpation, or a combination of both. Ahn and Kim measured 21 points on the posterior surfaces of
6 35 prostates using a smaller probe than the one used in the current work, but also with a relatively
7 small depth of indentation. Instead of carrying out a point-by-point statistical analysis, they mapped
8 the pathology of each prostate, dividing it into those regions classified as cancerous and normal and
9 then compared measured and estimated properties from the map (although this process is not
10 entirely clear). Carson *et al*, (44) assessed a total of 32 prostates, 26 of which were from radical
11 prostatectomies. They measured stiffness with a 12 mm diameter steel ball-end probe, recording
12 the force as the probe was advanced into the surface by 30% of its thickness (typically 9mm) and
13 retracted again over a period of 6-10min, using the rate of change of force with position immediately
14 on unloading (so-called Oliver-Pharr stiffness) to determine the elastic modulus which they simply
15 calculated from the force divided by hemispherical surface area. They probed five points on each
16 prostate and carried out some additional measurements on slices 6-10mm thick for a closer
17 correlation between histology and measurement. For the whole prostates, they reported a median
18 Young's modulus of 43.2kPa with a SD of 59.8. The corresponding values for prostates with palpable
19 abnormalities was 46.5 ± 22.2 and 31 ± 63.1 from which they were able to conclude only that prostates
20 with cancer were stiffer. Although they do not specifically measure an elastic constant, the
21 preliminary study on a whole excised prostate gland carried out by Astrand *et al* [48] needs also to
22 be mentioned here. Using the resonance sensor referred to earlier (22), they have reported a
23 statistically significant difference in dynamic stiffness between areas considered to contain cancer
24 and other areas of the prostate.

31 4.3 Implications of findings for *in vivo* deployment

32 The probe, as described here, is immediately deployable *in vivo* by mounting it on a surgeon's index
33 finger under a glove for trans-rectal insertion, essentially delivering it as in conventional DRE. A
34 typical DRE on a non-anaesthetised patient will last for the order of a few minutes, so it is important
35 to reduce the number of measurements and the number of points to the minimum necessary to
36 establish the condition. On anaesthetised patients, this is less of a problem, although interventions
37 should still be kept as short as possible.

40 Since the probe does not currently have a location device and its manoeuvrability *in situ* is limited,
41 the attitude of the probe and the exact position are only known with respect to anatomical
42 references such as the bladder neck and the anal sphincter which the practitioner can sense directly.
43 Taking a small number of points (6-8) against the references of top, bottom, left and right ensures at
44 least that there is coverage of the palpable region accessible through the rectum.

47 Finally, the applied force and the constraint are, to an extent, patient- and operator-specific so these
48 could also affect the measurements and hence the capacity of the probe to distinguish tissue
49 condition. Modelling work based on patient MRI scans by Palacio-Torralba *et al* [30], has suggested
50 that the local anatomy of the patient as well as the intrabladder pressure can have a significant
51 effect on the apparent stiffness. The effects of the overlying glove and the rectal wall, along with the
52 operator's subjective assessment of light, medium and heavy contact pressure will also serve to blur
53 the measurements somewhat. On the positive side, however, it is well-known that, despite patient-
54 specific variations, prostate cancer is manually palpable using conventional DRE with reasonable
55 inter-operator agreement, so the prospect of a meaningful measurement is good.

1
2
3 The main value of making a measurement *in vivo* with the existing probe is to answer the most
4 important question, which is whether or not the stiffness of the prostate *in vivo* is significantly
5 different to that measured with the same method *ex vivo* and also whether or not disease is more or
6 less distinguishable *in vivo* than *in vitro*. To this end, some preliminary measurements *in vivo* using
7 the existing probe is warranted.
8
9

10 11 12 13 14 15 **3. Conclusions**

16 The advantage of the current probe is that it returns values that can be traced to measurement
17 standards and so allows the use of data from other investigators in its detailed design and
18 development of measurement protocols, provided that the measurement conditions are clearly
19 known. The use of relatively low amplitude dynamic probing allows information on time constants to
20 be obtained efficiently and the option of a variety of probed depths allows potential non-linearity in
21 the stress-strain response to be assessed as well as allowing the probed volume to be varied. A final,
22 and immensely powerful, advantage of making calibrated measurements *in vivo* is that, in
23 conjunction with detailed mechanical modelling and/or simulation, it allows the development of
24 structure-property relationships whose used can enhance iDRE and can, furthermore, lead to wider
25 applications, for example in surgery and in different types of soft tissue
26
27

28 This work reports the first measurements of the dynamic mechanical properties of whole excised
29 prostates. The dynamic modulus is around 20kPa at 1Hz and around 14kPa at 5Hz. The method also
30 allows a measurement of the quasi-static modulus from the mean stress, effectively applied over the
31 few seconds' duration of a single point measurement, and values of around 14kPa were measured.
32 Although not directly comparable in terms of allowed relaxation times, the values determined are in
33 broad accord with the findings of the two other reported systematic studies of whole excised
34 prostates.
35
36

37 Whilst there was no single mechanical measure which yielded a generalised statistically significant
38 difference between diseased and normal tissue, combinations of the mechanical measures were
39 found to give sensitivity and specificity of up to 78% and 75%, respectively, applied to a single
40 patient and 73% and 62% when applied to all 11 patients. This performance could have been
41 improved by taking into account such factors such as edge effects, overlap and clinical significance of
42 the cancer.
43
44

45 In its current state of development, the device is deployable *in vivo* and was designed as such. This
46 necessitated a number of compromises in the degree of control of load and displacement, which
47 ultimately means that each embodiment has its own dynamic properties. A few modifications will
48 suffice to improve the sensitivity of the device. Changes in the measurement protocol would also
49 allow a wider range of time constants to be assessed and improve the precision of the assessment of
50 the quasi-static modulus.
51
52
53
54
55
56
57
58
59
60

1
2
3 Future areas of research include the deployment of the probe *in vivo*, the use of a wider panel of
4 diagnostic indicators, miniaturisation and modelling of the tissue histological structure to allow a
5 more robust structure-property relationship to be devised.
6
7

8 Acknowledgements

9
10 The authors acknowledge the support of the Engineering and Physical Sciences Research Council
11 under grant nos. (EP/I019472/1, EP/I020101/1 and EP/K036939/1).
12


13 References

- 14 1. Jemal A, Bray F, Center MM, Ferlay J, Ward E, Forman D. Global cancer statistics. *CA Cancer J*
15 *Clin.* 2011 Mar 1;61(2):69–90.
- 16 2. Gambert SR. Screening for prostate cancer. *Int Urol Nephrol.* 2001 Jun 1;33(2):249–57.
- 17 3. Goddi A, Sacchi A, Magistretti G, Almolla J. Transrectal real-time elastography of the prostate:
18 Normal patterns. *J Ultrasound.* 2011 Diciembre;14(4):220–32.
- 19 4. Pallwein L, Mitterberger M, Struve P, Pinggera G, Horninger W, Bartsch G, et al. Real-time
20 elastography for detecting prostate cancer: preliminary experience. *BJU Int.* 2007
21 Jul;100(1):42–6.
- 22 5. Kundra V. Prostate Cancer Imaging. *Semin Roentgenol.* 2006 Abril;41(2):139–49.
- 23 6. Auvinen A, Moss SM, Tammela TLJ, Taari K, Roobol MJ, Schröder FH, et al. Absolute Effect of
24 Prostate Cancer Screening: Balance of Benefits and Harms by Center within the European
25 Randomized Study of Prostate Cancer Screening. *Clin Cancer Res Off J Am Assoc Cancer Res.*
26 2016 Jan 1;22(1):243–9.
- 27 7. Gleason DF. Histologic grading of prostate cancer: A perspective. *Hum Pathol.* 1992
28 Mar;23(3):273–9.
- 29 8. Epstein JI. An Update of the Gleason Grading System. *J Urol.* 2010 Feb;183(2):433–40.
- 30 9. Garra BS. Imaging and Estimation of Tissue Elasticity by Ultrasound. *Ultrasound Q Dec 2007.*
31 2007;23(4):255–68.
- 32 10. Ophir J, Céspedes I, Ponnekanti H, Yazdi Y, Li X. Elastography: A quantitative method for
33 imaging the elasticity of biological tissues. *Ultrason Imaging.* 1991 Apr;13(2):111–34.
- 34 11. Hoskins PR. Principles of ultrasound elastography. *Ultrasound.* 2012 Jan 2;20(1):8–15.
- 35 12. Pallwein L, Mitterberger M, Pinggera G, Aigner F, Pedross F, Gradl J, et al. Sonoelastography of
36 the prostate: Comparison with systematic biopsy findings in 492 patients. *Eur J Radiol.* 2008
37 Feb;65(2):304–10.
- 38 13. V. Litwiller D, K. Mariappan Y, L. Ehman R. Magnetic Resonance Elastography. *Curr Med*
39 *Imaging Rev.* 2012 Feb 1;8(1):46–55.
- 40 14. Arani A, Plewes D, Chopra R. Transurethral prostate magnetic resonance elastography:
41 Prospective imaging requirements. *Magn Reson Med.* 2011 Feb;65(2):340–9.
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

15. Chopra R, Arani A, Huang Y, Musquera M, Wachsmuth J, Bronskill M, et al. In vivo MR elastography of the prostate gland using a transurethral actuator. *Magn Reson Med*. 2009 Jul 1;62(3):665–71.
16. Arani A, Plewes D, Krieger A, Chopra R. The feasibility of endorectal MR elastography for prostate cancer localization. *Magn Reson Med*. 2011 Dec;66(6):1649–57.
17. Li S, Chen M, Wang W, Zhao W, Wang J, Zhao X, et al. A Feasibility Study of MR Elastography in the Diagnosis of Prostate Cancer at 3.0T. *Acta Radiol*. 2011 Jan 4;52(3):354–8.
18. Sahebjavaher RS, Nir G, Gagnon LO, Ischia J, Jones EC, Chang SD, et al. MR elastography and diffusion-weighted imaging of ex vivo prostate cancer: quantitative comparison to histopathology. *NMR Biomed*. 2015 Enero;28(1):89–100.
19. Good D W *et al.*, *Elasticity as a biomarker for prostate cancer: a systematic review*. *Br J Urol International*, 2014, 113, pp. 523-534.
20. Panteliou SD, Tzortzis V, Anagnostopoulos GT, Sunaric MM, Sarris J, Hatzimouratidis K, et al. Development of a New Optical Device and Its Feasibility in Prostate Cancer Detection. *Urol Int*. 2012;89(3):290–5.
21. Jalkanen V, Andersson BM, Bergh A, Ljungberg B, Lindahl OA. Resonance sensor measurements of stiffness variations in prostate tissue in vitro—a weighted tissue proportion model. *Physiol Meas*. 2006 Nov 10;27(12):1373.
22. Jalkanen V, Andersson BM, Bergh A, Ljungberg B, Lindahl OA. Prostate tissue stiffness as measured with a resonance sensor system: a study on silicone and human prostate tissue in vitro. *Med Biol Eng Comput*. 2006 Jul;44(7):593–603.
23. Jalkanen V, Andersson BM, Bergh A, Ljungberg B, Lindahl OA. Explanatory models for a tactile resonance sensor system—elastic and density-related variations of prostate tissue in vitro. *Physiol Meas*. 2008 Jul 1;29(7):729–45.
24. Jalkanen V, Andersson BM, Bergh A, Ljungberg B, Lindahl OA. Indentation loading response of a resonance sensor—discriminating prostate cancer and normal tissue. *J Med Eng Technol*. 2013 Oct;37(7):416–23.
25. Yang THJ, Leung SKW, Phipps S, Reuben RL, McNeill SA, Habib FK, et al. In-vitro dynamic micro-probing and the mechanical properties of human prostate tissues. *Technol Health Care Off J Eur Soc Eng Med*. 2006;14(4–5):281–96.
26. Phipps S, Yang THJ, Habib FK, Reuben RL, McNeill SA. Measurement of tissue mechanical characteristics to distinguish between benign and malignant prostatic disease. *Urology*. 2005 Aug;66(2):447–50.
27. Yang, T H J et al. Dynamic instrumented palpation (DIP) - a new method for soft tissue quality assessment: application to prostate disease diagnosis, submitted to Proceedings IMechE, Part H, *Journal of Engineering in Medicine*, 2016
28. Åstrand A P *et al.* *Initial measurements on whole human prostate ex vivo with a tactile resonance sensor in order to detect prostate cancer*, IFMBE Proceedings, 2015, 48, pp. 120-123.

- 1
2
3 29. Kim Y, Ahn B, Na Y, Shin T, Rha K, Kim J. Digital rectal examination in a simulated environment
4 using sweeping palpation and mechanical localization. *Int J Precis Eng Manuf.* 2014 Jan
5 9;15(1):169–75.
6
7 30. Palacio-Torralba J et al. Patient specific modelling of palpation-based prostate cancer diagnosis:
8 effects of pelvic cavity anatomy and intrabladder pressure, *Int. J. Numer. Methods in Biomed. Eng.*,
9 2016, 32(1), pp. 1-13
10
11 31. BS EN 60601-1:2006 +A11:2011 Medical electrical equipment — Part 1: General requirements
12 for basic safety and essential performance [Internet]. The British Standards Institution; 2012.
13 Available from: [https://bsol.bsigroup.com/en/Bsol-Item-Detail-](https://bsol.bsigroup.com/en/Bsol-Item-Detail-Page/?pid=00000000030256124)
14 [Page/?pid=00000000030256124](https://bsol.bsigroup.com/en/Bsol-Item-Detail-Page/?pid=00000000030256124)
15
16
17 32. Yang T H J *et al.*, Dynamic instrumented palpation (DIP) - a new method for soft tissue quality
18 assessment; application to engineered mechanical phantom materials. *Biomedical Physics and*
19 *Engineering Express*, 2017, 3(2), 025103
20
21 33. McNeill SA, Reuben RL. Apparatus for mapping biological tissue quality. US2006051734 (A1),
22 2006.
23
24 33. Fung Y-C. *Biomechanics: Mechanical Properties of Living Tissues*. Springer; 1993. 590 p.
25
26 34. Scanlan, P., Hammer, S. J., Good, D., Shu, W., Reuben, R. L., Phipps, S., Stewart, G. A. & McNeill,
27 S. A. (2015) Development of a novel actuator for the dynamic palpation of soft tissue for use in
28 the assessment of prostate tissue quality. *Sensors and Actuators A (Physical)*, 232, pp 310-318
29
30 35. Scanlan P, Hammer SJ, Shu W, Reuben RL. A Scalable, Minimal Contact Device for the
31 Characterization of Elastomer Membrane Deformation. *Procedia Eng.* 2014;87:508–11.
32
33 36. Weisenhorn AL, Khorsandi M, Kasas S, Gotzos V, Butt H-J. Deformation and height anomaly of
34 soft surfaces studied with an AFM. *Nanotechnology.* 1993;4(2):106.
35
36 37. Posnansky O. On the influence of microscopic architecture elements to the global viscoelastic
37 properties of soft biological tissue. *Physica D.* 2014, 289:1–11.
38
39 38. Palacio-Torralba, J et al. Histology-based homogenization analysis of soft tissue: application to
40 prostate cancer. *J. Roy. Soc. Interface* 2017 14, 20170088; DOI: 10.1098/rsif.2017.0088.
41
42 39. Parker KJ, Huang SR, Lerner RM, Lee, F. J, Rubens D, Roach D. Elastic and ultrasonic properties
43 of the prostate. In: *Ultrasonics Symposium, 1993 Proceedings, IEEE 1993.* 1993. p. 1035–8
44 vol.2.
45 40. Krouskop TA, Wheeler TM, Kallel F, Garra BS, Hall T. Elastic Moduli of Breast and Prostate
46 Tissues under Compression. *Ultrason Imaging.* 1998 Oct 1;20(4):260–74.
47
48 41. Phipps S et al., Measurement of the mechanical characteristics of benign prostatic tissue: a novel
49 method for assessing benign prostatic disease, *Urology*, 2005, 65(3), pp. 1024-1028.
50
51 42. Zhang M *et al.* *Quantitative characterization of viscoelastic properties of human prostate*
52 *correlated with histology*, *Ultrasound in Med. and Biol.*, 2008, 34(7), pp. 1033-1042.
53
54
55
56
57
58
59
60

- 1
2
3 43. Murayama Y, Omata S, Yajima T, Peng Q, Shishido K, Peehl DM, et al. High Resolution Regional
4 Elasticity Mapping of the Human Prostate. In: Engineering in Medicine and Biology Society,
5 2007 EMBS 2007 29th Annual International Conference of the IEEE. 2007. p. 5802–5.
6
7 44. Carson WC, Gerling GJ, Krupski TL, Kowalik CG, Harper JC, Moskaluk CA. Material
8 characterization of ex vivo prostate tissue via spherical indentation in the clinic. Med Eng Phys.
9 2011 April;33(3):302–9.
10
11 45. Ahn B and Kim J. Measurement and characterization of soft tissue behavior with surface
12 deformation and force response under large deformations. Med Image Anal. 2010
13 April;14(2):138–48
14
15 46. Kim Y, Ahn B, Lee JW, Rha KH and Kim J. Local property characterisation of prostate glands
16 using inhomogeneous modeling based on tumor volume and location analysis, Med Biol Eng
17 Comput, 2013, 51, 197-205.
18
19 47. Palacio-Torralba, J., Hammer, S., Good, D.W., Alan McNeill, S., Stewart, G.D., Reuben, R.L., Chen,
20 Y., 2015a. Quantitative diagnostics of soft tissue through viscoelastic characterization using time-
21 based instrumented palpation. J. Mech. Behav. Biomed. Mater. 41, 149-160
22
23 48. Åstrand A P *et al.* *Initial measurements on whole human prostate ex vivo with a tactile resonance*
24 *sensor in order to detect prostate cancer*, IFMBE Proceedings, 2015, 48, pp. 120-123
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60



16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60

Figure 1: The main components of the probe end of the prototype in vivo DIP device. The strain sensor and the force sensor on the rear of the device are visible through the transparent membrane and body materials.




Figure 1: An actuation waveform (top) used to actuate the membrane on the DIP device. The strain waveform (bottom) shows the response of the membrane when applied to the prostate gland and actuated. The figure shows 1Hz (left) and 5Hz (right) portions of the actuation waveform.

For Peer Review




Figure 1: The Biomomentum Mach-1 indentation materials tester. Left – a view of the whole tester. Right – a closeup of the 6mm diameter hemispherical indenter being applied to a 3g gelatine sample.




Figure 1: Static elastic moduli of gelatine samples used in calibration as measured by Mach-1 probe.




Figure 1: Dynamic elastic moduli of gelatine samples used in calibration as measured by Mach-1 probe.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60




Figure 6: Phase difference of gelatine samples used in calibration as measured by Mach-1 probe at 1Hz (left), 4Hz (middle) and 5Hz (right).

Peer Review




Figure 7: Calibration results for amplitude ratio measured by DIP device and dynamic modulus from the Mach-1.




Figure 8: Calibration results for mean ratio measured by DIP device and quasi-static modulus from the Mach-1.

Peer Review




Figure 9: Calibration results for phase difference measured by DIP device (bottom) and from the Mach-1 (top).




Figure 10: Locating the DIP device on the ex vivo prostate. The device was presented to the posterior surface of the prostate under a sheath using an XYZ table (left). The position of each row measured on the prostate was marked using thread (right) as a guide to the positioning of the DIP device. Surgical clips were inserted to mark the position of each row of measurements to allow a histological section to be prepared at the correct axial position.




Figure 11: Typical probing sequence on posterior surface. Left, plan view showing points pitched at 6mm apart. Right, transverse histological section at A-A with areas of PCa outlined, showing typical column of assessment (point 11). Number of probe points depends on size of gland.

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49
50
51
52
53
54
55
56
57
58
59
60




Figure 12: Boxplots showing quasi-static modulus, dynamic modulus and phase difference for 11 prostates assessed with a mean depth of 3mm.

For Peer Review




Figure 13: Histological sections and point probe measurements (Patient 21)



Figure 14: Confusion matrix for P21 using six-element features and all depths. Target Class 1 is PCa and Target Class 2 is benign.



Figure 15: Confusion matrix for all patients using six-element features and all depths. Target Class 1 is PCa and Target Class 2 is benign.



Figure 16: Confusion matrix for all patients using six-element features and all depths. Target Class 1 is PCa and Target Class 2 is benign for the 2-target matrix and, for the 4-target matrix, Class 1 is cancer, Class 2 is mixed, Class 3 is BPH and Class 4 is normal.

Table 1: The number of measurement points recorded at 1Hz and 5Hz actuation frequencies and three depths for each tissue type identified in the 11 prostates measured.

Tissue histological classification	N_{4T}	Simplified classification	N_{2T}
Normal	83	Benign	149
BPH	66		
Cancer	149	Cancer	210
Cancer and BPH	61		
Total			359

Table 2: Statistical analysis of quasi-static modulus, dynamic modulus and phase difference for 11 prostates assessed with a mean depth of 3mm (Figure 12).

	Normal (n=83)	Cancer (n=149)	P-value	Benign (n=149)	Malignant (n=210)	P-value
E_{static} 1Hz (kPa)	15.1±0.1	15.1±0.2	0.7	15.1±0.2	15.1±0.2	0.7
$E_{dynamic}$ 1Hz (kPa)	20.7±0.5	20.8±0.6	0.07	20.7±0.7	20.8±0.7	0.5
PD 1Hz (rad)	-0.57±0.03	-0.58±0.03	0.003	- 0.57±0.04	- 0.59±0.03	0.002
E_{static} 5Hz (kPa)	14.2±0.2	14.4±0.2	5×10^{-7}	14.3±0.2	14.4±0.2	0.0004
$E_{dynamic}$ 5Hz (kPa)	14.0±0.6	14.3±0.8	0.005	14.1±1.0	14.3±0.8	0.13
PD 5Hz (rad)	1.19±0.15	1.14±0.14	0.04	1.18±0.15	1.14±0.14	0.007

Table 3: Sensitivity analysis of ANN performance for P21 (Figure 13).

Points	Feature vector	Percentage correct classifications	
		Test set	All
All (1-35)	All mech. features at each depth 3×35×6	50	65.7
1 to 28	All mech. features at each depth 3×28×6	53.8	73.8
All	Modulus only at each depth 3×35×4	56.3	50.5
All	Dyn Modulus only at each depth	68.8	64.8

	3×35×2		
All	Dyn Modulus and phase only at each depth 3×35×4	56.3	69.5
All	5Hz only at each depth 3×35×3	68.8	57.1

Table 4: Sensitivity and specificity for all 11 patients using "All" confusion matrix.

Target	Feature vector	Percentage correct classifications	
		Sensitivity	Specificity
Cancer and Benign	All mech. features at each depth 3×35×6 (P21 only)	68	63
Cancer and Benign	All mech. features at each depth 3×359×6	60	54
Cancer and Benign	All mech. features at all depths 359×18	65	56
Cancer and Benign	All mech. features (ex. E_{static}) at all depths 359×12	67	63
Cancer and Benign	All mech. features (ex. E_{static}) at each depth (P21 only) 3×35×4	78	76
Cancer, Mixed, BPH and Normal	All mech. features at all depths 359×18	52 (any disease) 52 (cancer)	52
Cancer and Normal	All mech. features at all depths 232×18	73	62
Cancer and Normal	All mech. features (ex. E_{static}) at all depths 232×12	73	62

Table 5: Summary of published values of prostate elastic modulus

Authors	Sample type	Test method	Notes	Modulus value(s)
Parker et al. (39)	Whole prostates	Ultrasonic resonance (3-7MHz)	Normal tissue only, linear elastic modulus	2.15 kPa
Krouskop et al. (40)	"Small pieces"	Sinusoidal indentation (4.83mm dia.), 0.1, 1, 4Hz, 2% and 4% pre-strain	Normal, BPH and PCa. No phase difference observed so only modulus reported.	40-85kPa normal, 30-55 BPH, 75-270 PCa
Phipps et al. (41)	TURP chips	Sinusoidal indentation (2mm dia.), 1-50Hz, 20% mean strain	Dynamic modulus	20-160kPa
Phipps et al. (26)	TURP chips	Sinusoidal indentation (2mm dia.), 1-50Hz, 20% mean strain	BPH vs PCa, Dynamic modulus and tan(PD)	100, 0.22 (BPH), 110, 0.15 (PCa)
Yang et al. (25)	Transverse slices, 6mm thick	Sinusoidal indentation (2mm dia.), 5-30Hz, 20% mean strain	Dynamic modulus and tan(PD), variation with frequency. No cancer present in samples.	100-450kPa, 0-0.3
Zhang et al (42)	Cores 8mm diameter and 7mm long	Stress relaxation at 5% compressive strain.	Results over 700s fitted to viscoelastic model and extrapolated dynamic modulus and tan(PD) to 150Hz. Normal vs PCa.	10-22kPa, 0.3-0.4 (normal, 24-56, 0.3-0.4 (PCa)
Murayama et al (43)	Sections 300 micron thick, obtained from cores 7mm dia and 10mm long	Microscale sensor (10 micron dia.), resonant at 89kHz	Modulus (E) calculated from frequency shift. 42k points probed on "normal" core from one patient and 55k points from patient with PCa used to give elasticity distribution	Mode around 10kPa for normal and PCa. Wider distribution for PCa
Carson et al. (44)	Whole prostates and some 5-10mm thick slices thereof	Load-unload indentation to around 30% strain over several minutes.	Young's modulus calculated from Oliver-Pharr stiffness. Focus on prostates with palpable abnormalities.	46.5±22.2 palpable and 31±63.1 non palpable
Ahn and Kim (45)	Whole prostates	2mm diameter hemispherical indenter	Indentation from 0 to 3mm at 1mm/s	15.2 ± 5.8 kPa (normal), 28.8 ± 11.2 kPa (cancer)

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49

Yang et al (27)	Transverse slices, 6mm thick	Sinusoidal indentation (2mm dia.), 5-30Hz, 20% mean strain	Quasi-static modulus. Dynamic modulus and tan(PD), variation with frequency. Nodular vs stromal histological components in normal or BPH prostates.	292kPa, 141-420kPa, 0.28-0.41 (nodular), 325kPa, 115-236kPa, 0.31-0.45 (stromal)
Yang et al (27)	TURP chips	Sinusoidal indentation (2mm dia.), 5Hz, 20% mean strain	Quasi-static modulus. Dynamic modulus and tan(PD). Patients with BPH, some with PCa.	21kPa, 83kPa, 0.22 (BPH), 40kPa, 111kPa, 0.19 (PCa)
This work	Whole prostates	Sinusoidal indentation (4mm dia.), 1Hz, 5Hz, light, medium, heavy mean strain	Quasi-static modulus. Dynamic modulus and tan(PD). Probe points classified as Normal, BPH, mixed and PCa.	Typically 15±0.5kPa, 20±0.5kPa, -0.6±0.03 rad (1Hz), 1.2± 0.1 rad (5Hz)

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49

For Peer Review

1
2
3
4
5
6
7
8
9
10
11
12
13
14
15
16
17
18
19
20
21
22
23
24
25
26
27
28
29
30
31
32
33
34
35
36
37
38
39
40
41
42
43
44
45
46
47
48
49

For Peer Review