

ORIGINAL SCIENTIFIC PAPER

COMPARISON OF TWO METHODS FOR DESTRUCTION OF BIOLOGICAL MATERIAL FOR DETERMINATION OF SELENIUM

MARIJAN MATEK AND MAJABLANUŠA

Institute for Medical Research and Occupational Health, Zagreb, Croatia

Received 3 November 1998

The purpose of the study was to compare two methods for destruction of biological material for selenium (Se) analysis: wet digestion by conductive heating in programmed digestion block and digestion in microwave oven. In both methods samples were prepared in a closed system using nitric acid. Selenium was analysed by electrothermal atomic absorption spectrometry.

The results have shown that both methods are convenient for complete mineralisation and are accurate in determining selenium in a variety of foodstuffs. Microwave digestion, however, has the advantage of speed and simplicity over the conventional heating procedure.

Key words:
dry ashing, fish, meat, microwave digestion, wet digestion

Preparation of samples is a critical stage in the analysis of trace elements. The well-established standard methods usually involve wet digestion by conductive heating (hot plate, digestion block) or dry ashing (muffle furnace). Mineral acids, oxidants, and ashing aids are used in various sample preparation techniques (1-3). Two types of the wet digestion system are currently in use: closed-vessel and open-vessel system (1, 2). Ever since the first application of microwave energy as a heat source for wet digestion in 1975, (1) the method has increasingly been used for preparation of samples (1, 4, 5).

Some advantages of microwave digestion over conventional sample preparation techniques for volatile elements have already been described – short sample prepara-

tion time, less acid required for digestion, pressure/temperature control, automation, and flexibility (1).

The purpose of our study was to compare and verify two methods for preparation of samples used in determination of selenium in food items: classical conductive heating in a programmed digestion block and digestion in a microwave oven. Electrothermal atomic absorption spectrometry (ET AAS) was used to quantify selenium.

METHODS

Equipment

For decomposition of samples we used a commercial microwave oven (CEM, USA, Model MDS-2000 with 950 W power) and a conductive heater with programmed digestion block (Tecator, Sweden, DS-40 and Autostep 1012 Controller). Digestion vessels (100 ml) for the microwave digestion consisted of a Teflon-PFA inner liner with a poly (ether imide) outer casing that can withstand a pressure of 1380 kPa (200 psi) and a temperature of 200 °C. For conductive heating in digestion block, we used borosilicate glass tubes (100 ml) with Teflon caps closed with special metal holders.

Varian AA300, Se Super lamp, pyrolytic coated partitioned tubes, and standard addition and deuterium background correction were used to measure selenium.

Reagents

We used nitric acid and selenium standard solution (1000±2 mg Se/L) of suprapur quality produced by Merck (Germany). The ascorbic acid and PdCl₂ as a matrix modifier were produced by Sigma (Switzerland) and Fluka (Germany), respectively. The conductivity of deionised water used in the experiment was 0.06 µS/cm.

Digestion procedures

For microwave digestion, 3 g of each fresh food item (hake, fried chicken, and a farm rainbow trout) or 0.2 g of lyophilised certified reference material (Community Bureau of Reference, Belgium – BCR, Pig kidney 186) were mixed with 3–5 ml of concentrated, suprapur HNO₃, and heated in the microwave oven. Tables 1 and 2 show detailed experimental conditions.

Table 1 *Experimental conditions for microwave oven digestion of fresh food items**

Parameter	Stage		
	I	II	III
Power (%)	70	70	70
Pressure (psi)	50	100	150
Running time (min)	10	10	5
Fan speed (%)	100	100	100

*3 g of fresh food was mixed with 5 ml of concentrated HNO₃.

Table 2 *Experimental conditions for microwave oven digestion of lyophilised certified reference material (BCR, Pig kidney 186)**

Parameter	Stage			
	I	II	III	IV
Power (%)	90	90	90	90
Pressure (psi)	20	40	80	135
Running time (min)	15	10	10	35
Fan speed (%)	100	100	100	100

*0.2 g of certified reference material was mixed with 3 ml of concentrated HNO₃.

For conventional heating in digestion block, 1 g of each of the above fresh food items or 0.5 g of lyophilised standard reference material were mixed with 1 ml of concentrated suprapur HNO₃. The samples were left to stand at room temperature in open tubes overnight. The tubes were then closed and heated in the block according to the temperature programme given in Table 3.

After digestion, both series of samples were adjusted to 10 ml with deionised water and analysed for Se by ET AAS.

Table 3 *Experimental conditions for conventional heating of fresh food items in digestion block or lyophilised certified reference material (BCR, Pig kidney 186) in a closed tube system**

Parameter	Stage		
	I	II	III
Temperature (°C)	50	60	80
Time increase (min)	10	10	10
Time at T (min)**	30	30	300

*1 g of fresh food item or 0.5 g of certified reference material was mixed with 1 ml concentrated HNO₃.

**Time at T indicates the duration of digestion at preset temperature

Electrothermal atomic absorption spectrometry

The ET AAS method was modified for Se measurements in food, as previously described for serum (6, 7). Before the measurement, Se (VI) was reduced to Se (IV) by mixing ascorbic acid and a matrix modifier (PdCl₂) with the sample. Standard addition method and deuterium background correction were applied for Se analysis in all samples of biological materials. The accuracy of both destruction techniques was tested by analysing pig kidney samples as certified reference standard.

RESULTS AND DISCUSSION

Conditions for microwave decomposition of lyophilised material differed from those for fresh food because of different mixture volumes taken in the decomposition tube

(Table 1 and 2). Higher power and longer decomposition time were needed for the lyophilised standard. At the end of decomposition procedure, all biological material was dissolved and the solutions were transparent. Decomposition by conventional heating did not require different temperature programmes. Measurements of the certified reference material showed that the recovery after destruction in the digestion block and in the microwave oven was 95.8% and 104%, respectively (Table 4). Such recovery is considered very good for both methods. The relative standard deviations, obtained by measuring ten replicates, were between 5.0% and 17.7%, depending on the concentration of Se in the tested biological material (Table 5). The precision was lower for Se concentrations in food below 0.1 $\mu\text{g/g}$.

Table 4 Accuracy of the two methods for wet digestion

Certified standard	Certified value ($\mu\text{g/g}$)	Determined value ($\mu\text{g/g}$)*	Recovery (%)	Digestion method
Pig kidney 186	10.3 \pm 0.5	10.6 \pm 0.4 9.9 \pm 0.2	104 95.8	Microwave oven Digestion block

* $\bar{X}_{10} \pm \text{SD}$

Table 5 Repeatability of Se measurements in biological material decomposed by the two methods

Biological material	$\bar{X}_{10} \pm \text{SD}$ ($\mu\text{g Se/g}$)	RSD (%)	Digestion method
Fried chicken	0.33 \pm 0.02	5.8	Microwave oven
Hake	0.25 \pm 0.02	6.0	Microwave oven
Hake	0.22 \pm 0.01	5.0	Digestion block
Farm rainbow trout	0.08 \pm 0.01	17.7	Digestion block

RSD – relative standard deviation

CONCLUSION

To conclude, both digestion methods can be used for a complete mineralisation and are accurate in determining selenium in a variety of food items. However, our study shows that microwave digestion has the advantage of speed and simplicity over the conventional heating digestion procedure.

REFERENCES

1. Gawalko EJ, Nowicki TW, Babb J, et al. Comparison of closed-vessel and focused open-vessel microwave dissolution for determination of cadmium, copper, lead, and selenium in

- wheat, wheat products, corn bran, and rice flour by transverse-heated graphite furnace atomic absorption spectrometry. *J AOAC Int* 1997;80:379–87.
2. *World Health Organisation International Programme on Chemical Safety (WHO IPCS) Selenium (Environmental Health Criteria 58)*. Geneva: WHO, 1987.
 3. *Foster LH, Sumar S*. Methods of analysis used for the determination of selenium in milk and infant formula. A review. *Food Chem* 1995;53:453–66.
 4. *Saraswati R, Vetter TW, Waters RL*. Comparison of reflux and microwave oven digestion for determination of arsenic and selenium in sludge reference material using flow injection hydride generation and atomic absorption spectrometry. *Analyst* 1995;120:95–9.
 5. *Bryce DW, Izquierdo A, Luque de Castro MD*. Use of focused microwaves for expeditive shortening of sample pre-treatment: digestion and reduction procedures prior to selenium speciation as selenium (IV) or selenium (VI). *Analyst* 1995;120:2171–4.
 6. *Teague-Nishimura JE, Tominaga T, Katsura T, et al*. Direct experimental evidence for in situ graphite and palladium selenide formations with improvement on the sensitivity of selenium in graphite furnace atomic absorption spectrometry. *Anal Chem* 1987;59:1647–51.
 7. *Pohl B, Schneider T*. Determination of normal levels of selenium in blood serum by GFAAS. *Varian Optical Spectroscopy Instruments*. AA Instrument at Work. Darmstadt: Varian GmbH, 1991;AA-103:1–6.

Sažetak

USPOREDBA DVIJU METODA RAZARANJA BIOLOŠKOGA MATERIJALA ZA MJERENJE SELENIJA

Priprema uzoraka kritična je faza određivanja elemenata u tragovima. Klasične metode razaranja biološkog materijala jesu mokro (digestijski blok) ili suho (mufolna peć) spaljivanje. Primjena mikrovalova u istu svrhu naglo raste tek osamdesetih godina.

U ovom radu uspoređene su dvije metode mokrog razaranja biološkog materijala za kasniju analizu selenija (Se): razaranje u aluminijskom bloku i razaranje mikrovalovima. Postupak kod obje metode proveden je u zatvorenim posudama koncentriranom HNO_3 suprapur čistoće. Obje metode validirane su određivanjem Se u standardnom referentnom materijalu istom metodom. Izmjereni rezultati uspoređeni su s referentnim vrijednostima. Točnost određena mjerenjem Se u referentnom materijalu spaljenom u bloku za digestiju iznosi 95,8%, odnosno 104% za spaljivanje u mikrovalnom uređaju. Relativna standardna devijacija (RSD), određena mjerenjem Se u 10 paralelno priređenih uzoraka, kretala se od 2,3 do 17,7% i od 3,8 do 6,0% ovisno o koncentraciji Se u ispitivanom materijalu. Može se zaključiti da su obje metode primjenjive za potpuno razaranje biološkog materijala za naknadnu analizu Se. Dakako, razaranje mikrovalovima ima prednost jer je sam postupak jednostavniji i brži.

Ključne riječi:

meso, mokro razaranje, razaranje mikrovalovima, riba, suho razaranje

Requests for reprints:

Marijana Matek, M.Sc.
Institute for Medical Research and Occupational Health
Ksaverska cesta 2, P.O. Box 291, HR-10001 Zagreb, Croatia
E-mail: mmatek@imi.hr