

# BIODEVICES 2008

International Conference on  
Biomedical Electronics and Devices

## Proceedings

Volume I

Funchal, Madeira - Portugal · 28 - 31 January, 2008

TECHNICAL CO-SPONSORSHIP BY



CO-ORGANIZED BY



IN COOPERATION WITH



# BIODEVICES 2008

Proceedings of the  
First International Conference on  
Biomedical Electronics and Devices

Volume 1

Funchal, Madeira - Portugal

January 28 – 31, 2008

Co-organized by  
**INSTICC – Institute for Systems and Technologies of Information,  
Control and Communication**  
and  
**UMA – University of Madeira**

Technical Co-sponsorship by  
**IEEE EMB – Engineering in Medicine and Biology Society**

In Cooperation with  
**AAAI – Association for the Advancement of Artificial Intelligence**

# BRAIN COMPUTER INTERFACE

## *Comparison of Neural Networks Classifiers*

José Luis Martínez Pérez and Antonio Barrientos Cruz

*Grupo de Robótica y Cibernética, Universidad Politécnica de Madrid, C/José Gutiérrez Abascal 2, Madrid, España*  
*jlmartinez@etsii.upm.es, antonio.barrientos@upm.es*

**Keywords:** Electroencephalography, Brain Computer Interface, Spectral Analysis, Biomedical Signal Detection, Pattern recognition.

**Abstract:** Brain Computer Interface is an emerging technology that allows new output paths to communicate the user's intentions without use of normal output ways, such as muscles or nerves (Wolpaw, J. R.; et al., 2002). In order to obtain its objective BCI devices shall make use of classifier which translate the inputs provided by user's brain signal to commands for external devices. The primary uses of this technology will benefit persons with some kind blocking disease as for example: ALS, brainstem stroke, severe cerebral palsy (Donchin et al., 2000). This report describes three different classifiers based on three different types of neural networks: Radial Basis Functions *RBF*, Probabilistic Neural Networks *PNN*, and Multi-Layer Perceptrons *MLP*. The report compares the results produced by them in order to obtain conclusions to apply to an on-line BCI device, it also describes the experimental procedure followed in the experiments. As result of the tests carried out on five healthy volunteers an estimation of the success rate for each type of classifier, the type and architecture of the classifier, and filtering windows are established.

## 1 INTRODUCTION

Brain Computer Interface technology (Wolpaw, J.R.; et al., 2000), BCI, is aimed to communicate human beings with external computerised devices using the electroencephalographic signal as primary source of commands (Birbaumer, N; et al., 2000); in the first international meeting for BCI technology in 1999 it was established that BCI “*must not depend on the brain's normal output pathways of peripheral nerves and muscles*”

In order to control an external device using thoughts it is necessary to associate some mental patterns to device commands, so an algorithm that detects, acquires, filters and classifies the human electroencephalographic signal is required (Kostov, A.; Polak, M., 2000) (Pfurtscheller et al., 2000).

This article compares results coming from three different classifiers based on neural networks: Radial Basis Function, Probabilistic Neural Networks, and Multi Layer Perceptron.

In the experiments considered for this article a low number of electrodes has been used to capture the endogenous electroencephalographic subject's signal. In order to facilitate the use of this technology it is important to make it easy to use, the number of electrodes employed in these devices is a global key feature, as the fewer of electrodes used, the higher the comfort (Wolpaw, 2007).

Because the main changes in brain activity are associated to changes in the power amplitude of band frequencies, spectrograms based on FFT are used to obtain initial feature vectors (Obermaier et al., 2001) (Proakis and Manolakis, 1997). Principal Component Analysis (PCA) is used to combine these initial features in order to reduce the dimensionality of the input space. To minimise the leakage effect seven different types of preprocess windows has been considered: rectangular, triangular, Blackman's, Hamming's, Hanning's, Kaiser's and Tukey's (Harris, 1978). The existence of statistical evidence in the feature population associated to different brain activities

has been previously shown (Peña Sánchez de Rivera, 1986) (Martinez, J.L.; et al., 2006).

The results provided by each classifier are compared using the confusion matrix (Duda et al., 2001).

This article is composed of the following sections: Section 2 briefly describes the methodology.

Section 3 describes the algorithmics used in the experiments.

Section 4 and 5 present and analyse the results.

Section 6 is devoted to conclusions.

## 2 EXPERIMENTAL PROCEDURE

The tests described below were carry out on five male healthy subjects, one of them has been trained before, but the other four were novice in the use of the system.

In order to facilitate the mental concentration on the proposed activities, the experiments were carried on in a room with low level of noise and under controlled environmental conditions, all electronic equipments external to the experiment around subject were switched off to avoid electromagnetic artifacts. The experiments were carry out between 10:00 a.m. and 14:00 p.m. The subjects were sat-down in front of the acquisition system monitor, at 50 cm from the screen, their hands were in a visible position, the supervisor of the experiment controlled the correct development of it.

### 2.1 Methodology

The experimental process is shown on figure 1.

*Test of system devices.* Checks the correct level of battery, and the correct state of the electrodes.

*System assembly.* Device connections: superficial electrodes (Grass Au-Cu), battery, bio-amplifier (g.BSamp by g.tec), acquisition signal card (PCI-MIO-16/E-4 by National Instrument), computer.

*System test.* Verifies the correct operation of the whole system. To minimise noise from the electrical network the Notch filter (50Hz) of the bio-amplifier is switched on.

*Subject preparation for the experiment.* Application of electrodes on subject's head. It is verified that electrode impedance was lower than 4 KOhms.

*System initialisation and setup.* Verification of data register. The temporal signal evolution is monitored, in the spectrogram should appear a very low 50 Hz component.

*Experiment setup.* The supervisor of the experiment sets-up the number of replications,  $N_{rep} = 10$ , and the quantity of different mental activities. The



Figure 1: Diagram of the experiment realization.

duration of each trial is  $t = 7s$ , the acquisition frequency is  $f_s = 384Hz$ . The system suggests to the subject to think about the proposed mental activity. A short relax is allowed at the end of each trial; between replications the relax time is  $t = 7s$ .

### 2.2 Position of Electrodes

Electrodes were placed in the central zone of the skull, next to C3 and C4 (Penny, W. D.; et al., 2000), two pair of electrodes were placed in front of and behind of Rolandic sulcus, this zone is one with the highest discriminant power, it takes signal from motor and sensory areas of the brain (Birbaumer, N; et al., 2000). Reference electrode was placed on the right mastoid, two more electrode are placed near to the corner of the eyes to register blinking.




Figure 2: Electrode placement.

### 2.3 Description of Cerebral Activities

The supervisor of the experiment asks the subject to figure out the following mental activities, these activities will be the cerebral patterns to differentiate among them (Neuper, C.; et al., 2001).

*Activity A. Mathematical task.* Recursive subtraction of a prime number, i.e. 7, from a big quantity, i.e. 3.000.000.

*Activity B. Movement task.* This task is subdivided in:

*B-1 Movement imagination.* The subject imagines moving their limbs or hands, but without the materialisation of the movement.

*B-2 Movement realization.* The subject is able to move their hands.

*Activity C. Relax.* The subject is relaxed.

### 3 ALGORITHM

This section describes the procedure applied to recorded signal just before its classification.




Figure 3: Algorithm.

#### 3.1 Window Analysis Generator

In this block the registered signal is chopped in packages of samples, similar to the bundles of samples obtained from an acquisition card in an on-line BCI application. The number of samples in each package is a compromise between the goodness of the classification and the amount of time taken by this classification. An algorithm with very good classification and low number of mistakes will take a very big package, so the time between classifications will be also very big, it will do the algorithm useless for a real on-line BCI system, neither a very fast algorithm with small packages of samples but with a high number of mistakes will be useful.

In this work we have considered packages of 128 samples, the sample frequency is  $F_s = 384Hz$ , and the classification latency is  $t = 1/3s$ .

The duration of each activity is 7s, so there will be 21 classifications obtained from each register, no overlap between windows have been considered.

#### 3.2 Standardisation

To compare the signal of different sessions is necessary to standardise the samples, avoiding for exam-

ple that variations in the impedance of the electrodes changes the classification result.

The standardisation of each analysis window consists in the subtraction of the average value and the division by the standard deviation.

$$\mu = \frac{\sum_{i=1}^N x_i}{N}; \quad \sigma^2 = \frac{(x - \mu)^2}{N}$$

$$x' = \frac{x - \mu}{\sigma}$$

#### 3.3 Windowing

In this block different kind of windows are convoluted with the standardise signal.

The frequency leakage effect occurs when signals with low frequency components are chopped or convoluted with windows with sharp edges, in this cases in the spectrogram appears high frequency components (Harris, F.J., 1978).

The following types of windows have been considered:

- Rectangular.
- Triangular.
- Blackman's.
- Hamming's.
- Hanning's.
- Kaiser's.
- Tukey's.

#### 3.4 FFT

The cerebral activity becomes apparent mainly through the frequency components of the electroencephalographic signal. Different kind of mental activities have different frequency components (Harris, F.J., 1978). For this reason it is necessary to transform the sampled time domain signal to frequency domain, so a Fast Fourier Transform is applied to each block of  $2^7$  sampled data.

Having in mind that the sample frequency is 384Hz, the frequency resolution is:

$$\Delta f = \frac{384Hz}{128} = 3Hz.$$

In this application the useful information is in the amplitude of the frequency components, so the phases are discarded, we focus our attention on the spectrograms of each of the analysis windows.

Considering the properties of the Fourier Transform and having in mind that the signal in the time domain only have real components, in the Nyquist

frequency is produced the reflection effect, so the signal information is in the first half of the components (Harris, F.J., 1978).

### 3.5 Feature Selection

A vector of features is extracted from each signal analysis window. This vector is made up as the mean of the amplitudes of the frequency bands. Because the frequency of normal human brain is under 40-50Hz, only frequencies between 6 and 38Hz have been considered.

Table 1: Feature vector.

FFT index.	Frequency.	Denomination.
1	0 - 2	Not considered
2	3 - 5	Not considered
3	6 - 8	$\theta$ .
4	9 - 11	$\alpha_1$ .
5	12 - 14	$\alpha_2$ .
6 - 7	15 - 20	$\beta_1$ .
8 - 10	21 - 29	$\beta_2$ .
11 - 13	30 - 38	$\beta_3$ .
14 - 64	39 - 192	Not considered

### 3.6 Classifiers

Three different types of classifiers have been considered, each one of them based on different types of neural networks (Ripley, 2000) (Bishop, 1995):

- Multi-Layer Perceptrons (MLP).
- Radial Basis Functions (RBF).
- Probabilistic Neural Networks (PNN).

Each classifier applies the following procedure to the vector of features extracted previously:

1. Determination of the learning (50%), test (25%) and validation (25%) sets of data.
2. Attainment of the normalisation matrix for the learning data set.
3. Application of Principal Component Analysis to the learning data set in order to reduce the dimensionality of the data input space.
4. Learning of the input data set by the neural network.
5. Application of the neural network to the test data set, if the error test is below the goal error the learning process is stopped, in other case the network is trained again.

6. Application of the neural network to the validation data set in order to estimate the performance error.
7. Application of the neural net to the whole data set and result registration.
8. Attainment of the confusion matrices for each experiment.

#### 3.6.1 Multi-Layer Perceptron Classifier

The setup parameters used in this classifier are:

- Learning algorithm: Levenberg-Marquardt (Backpropagation).
- Number of hidden unit neurons: 60.
- Number of output neurons: 3.
- Goal error =  $1e^{-5}$ .
- Epochs = 400.
- Max. fail = 5.
- Mem. reduc. = 1.
- Min. grad. =  $1e^{-10}$ .
- $\mu = 1e^{-3}$ .
- $\mu_{dec} = 0.1$ .
- $\mu_{inc} = 10$ .
- $\mu_{max} = 1e^{-5}$ .

#### 3.6.2 Radial Basis Function Classifier

The setup parameters used in this classifier are:

- Number of hidden neurons: The learning algorithm used by this type of neural networks determine the number of neurons in the hidden layer through an iterative process, it starts with a reduced number of hidden neurons and it is increased meanwhile the goal error is not achieved or a maximum number of neurons is reached.
- Spread constant : 0.25 (Determine the zone of influence of each neuron).
- Number of output neurons : 3.

#### 3.6.3 Probabilistic Neural Network Classifier

The setup parameters used in this classifier are:

- Number of hidden neurons: The learning algorithm used as much hidden neurons as pairs of input vector - target vector were in the learning data set.
- Spread constant : 0.25 (Determine the zone of influence of each neuron).
- Number of output neurons : 3.

## 4 RESULTS

The figures in the appendix summarised on vertical axis the percentage of correct classifications obtained from the confusion matrices applied to each one of the three classifiers. It shall be noted that the scale has been broken in order to appreciate the scattering results. On the horizontal axis appears the different types of filtering windows taken into account.

For each filtering window appears a bar with the results of each classifier: maximum, minimum and median percentage values.

It is also shown the results obtained when the classifier use two different types of architectures, one with only one neural network that gathers all vectors of features for each electroencephalographic channel, and other that employs two neural networks, one for each electroencephalographic channel.

## 5 DISCUSSION

From the analysis of the results the following considerations are extracted:

- Classifiers based on Probabilistic Neural Networks or Radial Basis Functions perform better than ones based on Multi Layer Perceptrons.
- Result stability. For all test the procedure was replicated three times, both PNN and RBF classifiers produced the same confusion matrices, instead of MLP classifiers which produced different confusion matrices for each replica.
- Comparison between PNN and RBF classifiers showed higher maximum percentages of correct classifications for PNN but also a higher variability.
- Classifiers based on only one neural network that considers at the same time features obtained from both electroencephalographic channels not always perform better than classifiers based on two neural networks, one for each channel.
- Considering the different types of filtering windows, the best results are obtained for Kaiser's, rectangular and Tukey's windows.

## 6 CONCLUSIONS

This report demonstrates that it is possible to discriminate mental activity from the electroencephalographic signal, it also compares three different types of neural networks classifiers applied to an off-line

prototype of BCI device that use FFT in order to estimate the power spectrum of the recorded signal when volunteers carried out specific mental tasks.

Both classifiers based on Probabilistic Neural Networks and Radial Basis Functions produced better and more stable results than the classifier based on Multi Layer Perceptrons. It is possible due to the vector feature distributions associate to each mental activity and to the interpolation capability of PNN and RBF, this capability is higher in PNN and RBF than in MLP neural networks.

It is hoped that On-line BCI devices based on classifiers that make use of neural networks like RBF or PNN will perform better than other based on MLP or equivalents.

In order to improve the success rate of classifications the use of filtering windows has been proved to be a good technique. In the same manner a classifier with a multiple network architecture followed by a block that weighs the network outputs could produce better results than classifiers based on only one neural network.

## REFERENCES

- Birbaumer, N; et al. (2000). The thought translation device (TTD) for completely paralyzed patients. *IEEE Transactions on Rehabilitation Engineering.*, 8(2):190–193.
- Bishop, C. (1995). *Neural Networks for Pattern Recognition Analysis*. Oxford University Press, London, 1st edition.
- Donchin, E., Spencer, K. M., and Wijesinghe, R. (2000). The mental prosthesis: assessing the speed of a p300-based brain-computer interface. *Rehabilitation Engineering, IEEE Transactions on [see also IEEE Trans.on Neural Systems and Rehabilitation]*, 8(2):174–179.
- Duda, R. O., Hart, P. E., and Strok, D. G. (2001). *Pattern classification*. John Wiley & sons, New York etc. Richard O. Duda, Peter E. Hart, David G. Strok.
- Harris, F. J. (1978). On the use of windows for harmonic analysis with the discrete fourier transform.
- Harris, F.J. (1978). On the Use of Windows for Harmonic Analysis with the Discrete Fourier Transform. *Proceedings of the IEEE*, 66(1):51–83.
- Kostov, A.; Polak, M. (2000). Parallel man-machine training in development of EEG-based cursor control. *IEEE Transactions on Rehabilitation Engineering.*, 8(2):203–205.
- Martinez, J.L.; et al. (2006). The windowing Effect in Cerebral Pattern Classification. An Application to BCI Technology. *IASTED Biomedical Engineering BioMED 2006*, pages 1186–1191.

Neuper, C.; et al. (2001). Motor Imagery and Direct Brain-Computer Communication. *Proceedings of the IEEE*, 89(7):1123–1134.

Obermaier, B., Neuper, C., Guger, C., and Pfurtscheller, G. (2001). Information transfer rate in a five-classes brain-computer interface. *IEEE Transactions on Neural Systems and Rehabilitation Engineering*, 9(3):283–288. Importante.

Peña Sánchez de Rivera, D. (1986). *Estadística : modelos y métodos*, volume 109-110. Alianza, Madrid. Daniel Pea Sanchez de Rivera; 2 v. 23 cm; 1. Fundamentos – 2. Modelos lineales y series temporales.

Penny, W. D.; et al. (2000). EEG-based communication: A pattern recognition approach. *IEEE Transactions on Rehabilitation Engineering*, 8(2):214–215.

Pfurtscheller et al. (2000). Current trends in Graz brain-computer interface (BCI) research. *IEEE Transactions on Rehabilitation Engineering*, 8(2):216–219.

Proakis, J. G. and Manolakis, D. G. (1997). *Tratamiento digital de seales : [principios, algoritmos y aplicaciones]*. Prentice-Hall, Madrid.

Ripley, B. (2000). *Pattern Recognition and Neural Networks*. Cambridge University Press, London, 2nd edition.

Wolpaw, J. R. (2007). Brain-computer interfaces as new brain output pathways. *The Journal of Physiology*.

Wolpaw, J. R.; et al. (2002). Brain-Computer interface for communication and control. *Clinical Neurophysiology*, 113:767–791.

Wolpaw, J.R.; et al. (2000). Brain-Computer Interface Technology: A Review of the First International Meeting. *IEEE Transactions on Rehabilitation Engineering*, 8(2):164–171.

## APPENDIX




Figure 4: Channel 1. Correct classification.




Figure 5: Channel 2. Correct classification.




Figure 6: Channel 1 and 2. Correct classification.






Figure 7: Channel 1. Correct classification.




Figure 10: Channel 1. Correct classification.




Figure 8: Channel 2. Correct classification.




Figure 11: Channel 2. Correct classification.




Figure 9: Channel 1 and 2. Correct classification.




Figure 12: Channel 1 and 2. Correct classification.




Figure 13: Channel 1. Correct classification.




Figure 16: Channel 1. Correct classification.




Figure 14: Channel 2. Correct classification.




Figure 17: Channel 2. Correct classification.




Figure 15: Channel 1 and 2. Correct classification.




Figure 18: Channel 1 and 2. Correct classification.