

SYKE Proficiency Test 5/2010

Gross and net calorific values in fuels

**Mirja Leivuori, Minna Rantanen,
Kaija Korhonen-Ylönen and Markku Ilmakunnas**

**REPORTS OF FINNISH ENVIRONMENT
INSTITUTE 4 | 2011**

SYKE Proficiency Test 5/2010

Gross and net calorific in fuels

**Mirja Leivuori, Minna Rantanen,
Kaija Korhonen-Ylönen and Markku Ilmakunnas**



Helsinki 2011

Finnish Environment Institute



REPORTS OF FINNISH ENVIRONMENT INSTITUTE 4 | 2011
Finnish Environment Institute SYKE

The organizer of the intercomparison test:
Finnish Environment Institute SYKE, Laboratories
Hakuninmaantie 6, 00430 Helsinki
phone +358 20 610 123, fax +358 9 495 913
Publication is available only in the internet :
www.environment.fi/publications

ISBN 978-952-11-3846-1 (PDF)
ISSN 1796-1726 (online)

CONTENT

ALKUSANAT/PREFACE	5
1 INTRODUCTION	6
2 ORGANIZING THE PROFICIENCY TEST	6
2.1 Responsibilities	6
2.2 Participants	7
2.3 Samples and delivery	7
2.4 Homogeneity studies	7
2.5 Feedback from the proficiency test	8
2.6 Analytical methods	8
2.6.1 Gross and net calorific value	8
2.6.2 Measurement of carbon, hydrogen, nitrogen, sulphur, moisture and ash	8
2.7 Processing of the data	9
2.7.1 Testing of normality of data, outliers and replicate results	9
2.7.2 Assigned values and uncertainties	9
2.7.3 Standard deviation for proficiency assessment and z score	10
3 RESULTS AND CONCLUSIONS	11
3.1 Results	11
3.2. Analytical methods and status to the results	12
3.2.1 Ash, moisture and elemental measurements	13
3.2.2 Gross and net calorific value	13
3.3 Uncertainties of the results	14
3.4 Estimation of emission factor	15
4. EVALUATION OF PERFORMANCE	15
5 SUMMARY	16
6 YHTEENVETO	17
7 REFERENCES	17

APPENDICES

Appendix 1	Participants in the proficiency test 5/2010	20
Appendix 2	Preparation of the samples	22
Appendix 3	Testing of the samples	23
Appendix 4.1	Comments sent by the participants	25
Appendix 4.2	Comments to the participants	26
Appendix 5.1	Analytical methods	27
Appendix 5.2	Significant differences in the results reported using different standard methods	33
Appendix 5.3	Results grouped according to the methods	34
Appendix 6	Evaluation of the assigned values and their uncertainties	40
Appendix 7	Terms in the result tables	41
Appendix 8	Results of each participant	42
Appendix 9	Summary of the z scores	53
Appendix 10	Measurement uncertainties reported by the laboratories	55
Appendix 11	z and zeta scores of laboratories	62
 DOCUMENTATION PAGE		68
KUVAILULEHTI		69
PRESENTATIONBLAD		70

ALKUSANAT

Suomen ympäristökeskus (SYKE) on toiminut ympäristöalan kansallisena vertailulaboratoriona vuodesta 2001 lähtien. Toiminta perustuu ympäristöministeriön määräykseen, mikä on annettu ympäristönsuojelulain (86/2000) nojalla. Vertailulaboratorion tarjoamista palveluista yksi tärkeimmistä on pätevyyskokeiden ja muiden vertailumittausten järjestäminen. Vertailumittausten järjestäminen täyttää kansainvälisen ohjeiden ISO/IEC Guide 43-1 ja ILAC-G13 asettamat vaatimukset. SYKE:n laboratoriot on FINAS-akkreditointipalvelun akkreditoima testauslaboratorio T003 ja vertailumittausten järjestäjä PT01 (Proftest SYKE, www.finias.fi).

Tämä pätevyyskoe on toteutettu SYKE:n vertailulaboratorion pätevyysalueella ja se antaa tietoa osallistujien pätevyyden lisäksi tulosten vertailukelpoisuudesta myös yleisemmällä tasolla.

Pätevyyskokeen onnistumisen edellytys on järjestäjän ja osallistujien välinen luottamuksellinen yhteistyö.

Parhaat kiitokset yhteistyöstä kaikille osallistuille!

PREFACE

Finnish Environment Institute (SYKE) has served as the National Reference Laboratory in the environmental sector designated by the Ministry of the Environment under the section 24 of the Environment Protection Act (86/2000) since 2001. The duties of the reference laboratory service include providing proficiency tests and other interlaboratory comparisons for analytical laboratories and other producers of environmental information. The proficiency testing service is a part of the SYKE laboratory management system based on the standard EN ISO/IEC 17025. The SYKE proficiency testing service also conforms to the requirements of ISO/IEC GUIDE 43-1 and ILAC G-13. The SYKE laboratories have been accredited by the Finnish Accreditation service as the testing laboratory T003 and as the proficiency testing provider PT01 (Proftest SYKE, www.finias.fi).

This proficiency test has been carried out under the scope of the SYKE reference laboratory and it provides information about performance of the participants as well as comparability of the results at more general level.

The success of the proficiency test requires confidential co-operation between the provider and participants.

Thank you for your co-operation!

Helsingissä 4. Helmikuuta 2011 / Helsinki 4 February 2011



Marja Luotola

Laboratorionjohtaja / Chief of laboratory

1 INTRODUCTION

The Finnish Environment Institute (Proftest SYKE) carried out the proficiency test for the analysis of the gross and the net calorific value as well as for content of ash, carbon, hydrogen, nitrogen, sulphur and analytical moisture content in fuels in September 2010. The samples were prepared from peat (B1) and coal (K1). Additionally, the participants were asked to estimate/calculate the emission factor for both samples.

The test was carried out in accordance with the international guidelines, ISO/IEC Guide 43-1 [1], ISO/IEC 17043 [2], ISO 13528 [3] and IUPAC Recommendations [4]. The Proftest SYKE has been accredited by the Finnish Accreditation Service as a proficiency testing provider (PT01, www.finias.fi). Proftest SYKE is the accredited proficiency test provider on the field of the present test.

The proficiency test performed as the joint work with the working group (WG) of the European co-operation for Accreditation (EA) for Interlaboratory Comparisons (EA WG ILC in Testing) and the data will be confidentially handled in the work of this WG.

2 ORGANIZING THE PROFICIENCY TEST

2.1 Responsibilities

Organizing laboratory:

Finnish Environment Institute (SYKE), Laboratories
Hakuninmaantie 6, 00430 Helsinki
tel. +358 20 610 123, fax +358 9 448 320

Subcontractors:

The peat sample B1 prepared by Enas in Jyväskylä (Finland) and the coal sample K1 was prepared by Helsinki Energia (Finland). Both samples were homogenized and divided to the sub samples at the laboratory of Water Protection Association of the Kokemäenjoki River in Tampere (Finland, accredited testing laboratory T064 by the Finnish Accreditation Service, www.finias.fi).

The samples were tested at Mibrag mbH in Zeitz (Germany, the accredited testing laboratory DGA-PL-1161.00 by the Deutsche Gesellschaft für Akkreditierung mbH, www.dakks.de). The calorific value of the peat sample was, additionally, tested at the laboratory of Ramboll Analytics in Vantaa (Finland, the accredited testing laboratory T039 by the Finnish Accreditation Service, www.finias.fi).

The responsibilities in organizing the proficiency test were as follows:

Mirja Leivuori, coordinator
Kaija Korhonen-Ylönen, substitute of coordinator
Keijo Tervonen, technical assistance
Sari Lanteri, technical assistance
Markku Ilmakunnas, technical assistance and layout of the report.

The analytical expert was:

Minna Rantanen, Ramboll Analytics

2.2 Participants

In this proficiency test (PT) totally 58 laboratories participated, from which 16 were from Finland and 41 from other European countries and one laboratory from Asia (Appendix 1). The sample testing laboratories Mibrag mbH has the code 60 and Ramboll Analytics has the code 59 in the result tables.

2.3 Samples and delivery

The preparation of the samples is presented more detailed in Appendix 2.

The sample B1 was the peat sample from the Finnish marshland. The material was air dried and grounded by the mill with 500 µm sieve before homogenization and sample dividing.

The coal sample (K1) was prepared from a Russian crushed coal. The material was air dried and grounded to particle size < 200 µm before homogenization and sample dividing.

The samples were delivered 14 September 2010. They were requested to be analyzed and reported before 7 October 2010.

The samples and the requested measurands were as follows:

Sample	Sample type	Measurements
B1	Peat	Gross and net calorific value, C, S, N, H, moisture content of the analysis sample (M_{ad}), ash content
K1	Coal	Gross and net calorific value, C, S, N, H, moisture content of the analysis sample (M_{ad}), ash content

In the covering letter with the samples was noted that the moisture content of the analysis had to be measured as the first measurement after storing samples closed one day on the measuring laboratory. The samples were asked to homogenate before measurements and to store in dry place at room temperature.

Additionally, the participants had the possibility to estimate/calculate the emission factor (as received) for the both samples. For this estimation/calculation the organizer of this PT reported the total moisture contents as received (M_{ar}) for peat sample B1 44.3% and for coal sample K1 11.2% in the covering letter of the samples.

2.4 Homogeneity studies

Homogeneity of the samples B1 and K1 was tested by analyzing the gross calorific value and the ash content as replicate determinations from fifteen subsamples. The homogeneity of the gross calorific value of the peat sample B1 was tested from eight additional samples by other laboratory due to the somewhat high variability of the first results. After this additional testing the peat sample was considered also as homogenous (Appendix 3). Moreover carbon, nitrogen and hydrogen from fifteen and sulphur from eight subsamples were measured. According to the all homogeneity test results the both samples B1 and K1 were considered homogenous.

Particle size distribution was also tested from one sub sample of peat (B1) and coal (K1). The results show that the samples were appropriate for measurement of calorific value (Appendix 3).

2.5 Feedback from the proficiency test

Appendix 4.1 contains the comments sent by the participants. The comments were mainly relating to the data input protocols in the laboratories. The provider gives some comments to the participants considering mainly the reporting of the results in Appendix 4.2.

2.6 Analytical methods

2.6.1 Gross and net calorific value

The analytical methods based on different standard methods were used for the measurements in the PT. The used analytical methods of the participants are shown in more detail in Appendix 5.1.

Mostly, the standard methods or the CEN/technical specification were used for measurement of calorific value (CEN/TS 14918 [5], ISO 1928 [6], DIN 51900 [7], ASTM D 5865-07 [8]). A few laboratories were used some national standards (e.g. lab 29, 44). The participants used mainly the sample amount 0.5–1 g for measurement of the calorific value. Generally, the analyses were carried out from air dried samples (Appendix 5.1).

The measurements of calorific value were mainly done by IKA, LECO and PARR equipments. The volume of water added into a reaction bomb varied mainly from 1 to 10 ml depending on the type of measuring equipment (Appendix 5.1). In the calibration used benzoic acid from 8-10 different producers. Mainly, the calibration standard was used without correction to the value given in the certificate.

In the calculation of gross calorific value ($q\text{-V,gr,d}$) various correction methods were used. Basically, fuse wire, ignition, acid, cotton, moisture, nitrogen and sulphur corrections were used. However, the participants used several combinations of them (Appendix 5.1). In the calculation of net calorific value ($q\text{-p,net,d}$) different combinations of correction factors were used as well. Mainly, the measured hydrogen content with or without nitrogen and oxygen corrections was used. However, in some cases also calculated hydrogen content was used for corrections.

2.6.2 Measurement of carbon, hydrogen, nitrogen, sulphur, moisture and ash

In the PT several standard methods or technical specifications were used mainly for measurement of different parameters as follows:

- C, H and N: CEN/TS 15104 [9], ISO/TS 12902 [10], ASTM D 5373 [11]
- S: ASTM D 4239 [12], CEN/TS 15289 [13], ISO 334 [14]
- Analytical moisture content: CEN/TS 14774 [15], ISO 589 [16], DIN 51718 [17], ASTM D 5142 [18]
- Ash content: CEN/TS 14775 [19], ISO 1171 [20], DIN 51719 [21], ASTM D 5142 [18]

However, in some cases other international standards or national standards were used. For example, sulphur was measured using standards: ASTMD 3177 (lab 28), ISO-351, ISO 19579 (lab 51) or DIN51900-1+DIN10304-1 (lab 46). For moisture measurements were used also standard: ASTM D 3173 (lab 28), ISO 11722 (lab 35, 47, 51), ISO 5066-2 or ASTM D 3302 (lab 57). Additionally, one laboratory used infrared moisture analysis (21).

Carbon, hydrogen and nitrogen were also measured using standards: ASTM D 3178 (lab 37), ASTM D5291-10 (lab 46) or ISO 1028 (lab 47). Also some national standards (e.g. lab 29, 37, 38, 4, 42, 44, 48, 49) or in-house methods (EDXRF, lab 57) were used.

Carbon, hydrogen and nitrogen were measured by using different equipments (e.g. VARIOMAX, LECO, ELTRA, ELTRA CHS, Appendix 5.1). Different elemental analyzers (e.g. ELTRA, LECO, Appendix 5.1) were also used for measurements of sulphur. Sulphur was measured also by using O₂-combustion and IC-measurement and some other techniques.

Ash content was measured also using some other standards, e.g. ASTMD 3174 (lab 28, 35, 42) and some national standards (e.g. lab 29, 37, 38, 41) or in-house methods were reported as well. Ash content was determined gravimetrically by heating mainly at the temperature 550 °C (Sample B1) or 815 °C (Sample K1). Also some other temperatures were used for ash content measurements (Appendix 5.1).

2.7 Processing of the data

2.7.1 Testing of normality of data, outliers and replicate results

Before the statistical treatment, the data was tested according to the Kolmogorov-Smirnov normality test and the outliers were rejected according to the Hampel test for calculation of the mean value (H in the results sheets). Also before the robust calculation some extreme outliers were rejected in case that the results deviated from the robust mean more than 50 %. The replicate results were tested using the Cochran-test (C in the result sheets). If the result was reported < DL (detection limit), it has not been included in calculation of the results (H in the results sheets).

2.7.2 Assigned values and uncertainties

The robust mean was used as the assigned value for each measurement of the sample B1 and K1 (Appendix 6). In the calculation of the robust mean outliers were not normally rejected, but they were iterated before the final calculation of the robust mean. However, in this proficiency test some extreme results (at most 1-7 results/measurement) had to be rejected because of rather strict requirements for reproducibility given in the standards for analysis described in the covering letter of the samples. Especially in estimation of the assigned value of gross and net calorific value, the base for extreme value was either the anomalous calorific value or anomalous value in the measured moisture or/and element value used in the calculation or the reported results were not according to the given instructions. In addition, a few laboratories reported the anomalous values in measurement of the gross and net calorific value, which may indicate systematic errors in measurement. Also the mean value (after using the Hampel outlier test) and the median value of the data were calculated, which were quite near with the assigned values (Table 1). Also the results of homogeneity testing of the samples were used as background information in estimation of the assigned values. Additionally, the calculated assigned values of the calorific values were compared with the results obtained in the kernel density plots [4].

When using the robust mean of the participant results as the assigned value, the uncertainties of the assigned values for calorific values varied from 0.18 % to 0.28 %. For the other measurements the uncertainty varied from 0.34 % to 5.6 % (Appendix 6).

The participants also calculated emission factors (EF) according to the given total moisture contents as received (M_{ar}) for the samples and the results were evaluated as well. According to the evaluation of results at least laboratories 34 and 43 were not calculated the emission factor for peat as requested. The case was the same for the emission factor of coal reported by the laboratory 43. These results were excluded from the assigned value calculation. For these laboratories, the performance evaluation of emission factor was not satisfactory, thus weakening the reliability of the performance evaluation of EF as a whole.

After reporting the preliminary results in October 2010 no corrections had done to the assigned values.

2.7.3 Standard deviation for proficiency assessment and z score

For the total standard deviation for proficiency assessment used in the calculation of the z score was used the target value for reproducibility recommended in the international standards or technical specifications for measurement of calorific values and other determinants [5, 6, 9, 10, 12, 19, 20, 21].

The reproducibility recommended in the standards was mainly fulfilled for the gross calorific values ($\pm 300 \text{ J g}^{-1}$). For the net calorific value the reproducibility was increased to $\pm 368 \text{ J g}^{-1}$ for the sample B1 and to $\pm 374 \text{ J g}^{-1}$ for the sample K1. The reason was partly the variability of the results and the missing of clear reproducibility information for the net calorific value in the standard methods [e.g. 5-8]. In calculation of net calorific value there are more uncertainty sources than in calculation of gross calorific value. Particularly, on the final results of net calorific value uncertainty and errors of other measurements (i.e. moisture, S, H) can affect. For some other measured parameters (i.e. C, H, N, S and ash) the total standard deviation for proficiency assessment had to be increased from the reproducibility presented in the standard methods. However, noticeable was that for H and ash the reproducibility of results were better than in the previous PT and the total standard deviation was able to lower in this PT when compared to the previous PT [22].

The results of analysis moisture (M) have not been evaluated because of rather great variation of the results, but the assigned values for both sample types are presented.

The performance evaluation was carried out by using z scores (Appendix 7).

In the performance evaluation z scores were interpreted as follows:

$ z \leq 2$	satisfactory results
$2 < z < 3$	questionable results
$ z \geq 3$	unsatisfactory results

The performance evaluation of participants using calculated z scores are presented in Appendix 8.

The reliability of the assigned value was tested according to the criterion:

$$u/s_p \leq 0.3 \quad , \text{ where}$$

u is the standard uncertainty of the assigned value (the uncertainty of the assigned value (U) divided by 2) and

s_p the standard deviation for proficiency assessment (total standard deviation divided by 2).

The test criterion for the reliability of the assigned value was fulfilled in every case, which indicated that the assigned values were very reliable.

The reliability of the target value for the total deviation and the reliability of the corresponding z score were estimated by comparing the deviation for proficiency assessment (s_p) with the robust standard deviation of the reported results (s_{rob}). The criterion $s_{rob} < 1.2 * s_p$ was fulfilled in every case. Due to this the evaluation of performance is reliable for this proficiency test.

The performance was not evaluated for moisture contents partly due to high variability between the results. The evaluation of emission factor (EF) is only indicative due to the noticed calculation errors for the both sample types.

After reporting the preliminary results in October 2010 no corrections had done to the standard deviations for the proficiency assessment.

3 RESULTS AND CONCLUSIONS

3.1 Results

The summary of the results is presented in Table 1. The reported results and their uncertainties grouped by the measurement methods are presented graphically in Appendix 5.3. Explanations to terms used in the result tables are presented in Appendix 7. The results and the performance of each laboratory are presented in Appendix 8. The summary of z scores is shown in Appendix 9. The measurement uncertainties reported by the laboratories grouped according to the evaluation procedure is reported in Appendix 10. The comparison of z and zeta scores (Appendix 7) is shown in Appendix 11.

Table 1. Summary of the result in the proficiency test 5/2010.

Analyte	Sample	Unit	Ass. val.	Mean	Mean rob.	Md	SD rob	SD rob, %	Num. of labs	2*Targ SD%	Accepted z-val%
Ash,d	B1	w%	7,7	7.69	7.70	7.70	0.18	2,3	30	6	90
	K1	w%	10,9	10.88	10.90	10.91	0.10	0,9	48	2,5	92
C,d	B1	w%	53,4	53.39	53.43	53.66	0.79	1,5	22	3	86
	K1	w%	73,3	73.21	73.25	73.26	0.92	1,3	38	2,5	87
EF	B1	t CO ₂ /TJ	106	105.85	105.93	105.90	1.15	1,1	14	4	64
	K1	t CO ₂ /TJ	93,8	93.83	93.84	93.96	1.13	1,2	23	4	91
H,d	B1	w%	5,63	5.63	5.63	5.69	0.16	2,8	20	7	90
	K1	w%	4,6	4.60	4.60	4.57	0.18	3,8	32	7	91
Mad,d	B1	w%	2,88	2.84	2.88	2.87	0.49	17	30		
	K1	w%	3,52	3.50	3.52	3.54	0.22	6,3	49		
N,d	B1	w%	1,62	1.62	1.62	1.59	0.075	4,6	20	15	85
	K1	w%	2,14	2.14	2.14	2.17	0.15	7,2	30	15	93
q-p,net,d	B1	J/g	20460	20495.19	20460.12	20456.00	90.31	0,4	27	1,8	74
	K1	J/g	28810	28782.99	28806.49	28810.00	187.57	0,7	42	1,3	71
q-V,gr,d	B1	J/g	21680	21696.23	21680.50	21683.00	99.36	0,5	31	1,4	77
	K1	J/g	29820	29805.91	29818.50	29810.00	134.64	0,5	49	1	71
S,d	B1	w%	0,25	0.25	0.25	0.25	0.015	6,1	25	15	92
	K1	w%	0,31	0.31	0.31	0.31	0.018	5,9	45	15	84

Ass. Val.	the assigned value
Mean	the mean value
Mean rob	robust mean
Md	the median value
SD %	the standard deviation as percent
SD rob	the robust standard deviation
SD rob %	the robust standard deviation as percents
Num of Labs	the number of participants
2*Targ. SD%	the total standard deviation for proficiency assessment at 95 % confidence level (2^*s_p)
Accepted z-val%	the satisfactory z scores: the results (%), where $ z \leq 2$.

The robust standard deviation of results was lower than 2 % for 50 % of the results and it was lower than 6 % for 78 % of the results (Table 1). For nitrogen (N) in the sample K1 and for sulphur (S) in the sample B1 the robust standard deviation was 7.2 and 6.1 %, respectively. In measurement of moisture the robust standard deviation was 6.3 % in the coal sample and more than duplicated (17 %) in measurement of the peat sample. The standard deviations of the results in this PT were nearly in the same range than in the previous respective PT SYKE 5/2009 [22], where the deviations varied from 0.5 % to 11 %.

In this PT the participants were requested to report the replicate results for all measurements. The results of the replicate determinations based on the ANOVA statistical handling are presented in Table 2. The international standards or technical specifications relating to measurements in fuels recommend the targets for the repeatability.

In particular, in measurement of the calorific values, the requirement for the repeatability is ± 120 J/g. In this PT the requirements for the repeatability in measurement of the gross calorific value are 0.55 % for the sample B1 and 0.40 % for the sample K1 and in measurement of the net calorific value 0.59 % and 0.42 %, respectively. In each case, the obtained repeatability in measurement of the gross calorific value and the net calorific value was lower than the repeatability requirement (Table 2, the column s_w %). The repeatability was mainly acceptable only for carbon C in the elemental measurements (Table 2, the column s_w %).

The summary of the robustness of the methods, the ratio s_b/s_w , is presented in Table 3. The ratio s_b/s_w should not be exceeded 3 for robust methods. However, in Table 3 is seen that in many cases the robustness exceed the value 3. For the gross calorific value, the ratio s_b/s_w , was 2.9 % (the sample B1) and 4.6 % (the sample K1), for the net calorific values 5.3 % and 6.0 %, respectively. The high ratio s_b/s_w for the net calorific value was at least partly resulting from variable procedures used in calculation of the results.

Table 2. Summary of repeatability on the basis of duplicate determinations (ANOVA statistics).

Analyte	Sample	Unit	Ass. val.	Mean	Md	sw	sb	st	sw %	sb %	st %	2*Targ SD %	Num of labs	Accepted z-val %
Ash,d	B1	w%	7,7	7,696	7,7	0,0992	0,1463	0,1768	1,3	1,9	2,3	6	30	90
	K1	w%	10,9	10,88	10,91	0,04347	0,1307	0,1377	0,4	1,2	1,3	2,5	48	92
C,d	B1	w%	53,4	53,41	53,61	0,1358	0,7927	0,8043	0,25	1,5	1,5	3	22	86
	K1	w%	73,3	73,22	73,25	0,1307	0,9912	0,9998	0,18	1,4	1,4	2,5	38	84
EF	B1	t CO ₂ /TJ	106	105,8	106,1	0,3619	2,152	2,182	0,34	2	2,1	4	14	64
	K1	t CO ₂ /TJ	93,8	93,78	93,91	0,1764	1,609	1,619	0,19	1,7	1,7	4	23	91
H,d	B1	w%	5,63	5,629	5,675	0,03594	0,1464	0,1508	0,64	2,6	2,7	7	20	90
	K1	w%	4,6	4,596	4,57	0,03232	0,2082	0,2107	0,7	4,5	4,6	7	32	88
Mad,d	B1	w%	2,88	2,852	2,87	0,03728	0,49	0,4914	1,3	17	17		30	
	K1	w%	3,52	3,507	3,53	0,0369	0,3003	0,3025	1,1	8,6	8,6		49	
N,d	B1	w%	1,62	1,621	1,601	0,02274	0,0727	0,07617	1,4	4,5	4,7	15	20	85
	K1	w%	2,14	2,137	2,17	0,02053	0,1437	0,1451	0,96	6,7	6,8	15	30	93
q-p,net,d	B1	J/g	20460	20490	20460	30,68	164,4	167,3	0,15	0,8	0,82	1,8	27	74
	K1	J/g	28810	28790	28820	35,54	211,9	214,8	0,12	0,74	0,75	1,3	42	71
q-V,gr,d	B1	J/g	21680	21690	21700	54,08	153,6	162,8	0,25	0,71	0,75	1,4	31	77
	K1	J/g	29820	29810	29820	35,6	162	165,8	0,12	0,54	0,56	1	49	71
S,d	B1	w%	0,25	0,2504	0,248	0,008775	0,01508	0,01745	3,5	6	7	15	25	92
	K1	w%	0,31	0,3068	0,31	0,006946	0,02101	0,02213	2,3	6,8	7,2	15	45	84

Ass. val. - assigned value, Md - median, sw - repeatability standard error, sb - standard error between laboratories, st - reproducibility standard error

Table 3. The robustness (s_b/s_w) of the replicate results in the PT5/2010.

Analyte	Sample	s_b/s_w	Analyte	Sample	s_b/s_w
Ash	B1	1.5	S	B1	1.7
	K1	3.0		K1	3.0
C	B1	5.8	q-p,net,d	B1	5.3
	K1	7.6		K1	6.0
H	B1	4.1	q-V,gr,d	B1	2.9
	K1	6.4		K1	4.6
M _{ad}	B1	13	EF	B1	5.9
	K1	8.1		K1	9.2
N	B1	3.2			
	K1	7.0			

3.2 Analytical methods and status to the results

In Appendix 5.1 is summarized the used analytical methods in the proficiency test. The difference between the average concentrations of elements measured by different sample preparation methods was tested using the t-test. The results of the t-test are shown in Appendix 5.2. In Appendix 5.3 is presented the results of participated laboratories grouped based on used different standard methods.

3.2.1 Ash, moisture and elemental measurements

In measurement of the ash content only a few laboratories reported too high or too low values, and different techniques have not clearly affected the results (Appendix 5.3). In the ash content, no statistically significant difference between the results was noticed.

The analysis moisture (M_{ad}) was measured using different standard methods (Appendices 5.1 and 5.3). There was the statistically significant difference between the moisture results in the sample K1 measured by the standards ISO 589 and DIN 51718 and between DIN 51718 and ASTM D 5142 (Appendix 5.2). The differences between the results can be varied from the amounts of used sample amount for analyses, the drying time at an oven and drying atmosphere at an oven or the humidity content of the ambient. The methods ISO 589 and DIN 51718 are based on gravimetric methods and ASTM D 5142 is instrumental procedure. It was difficult to estimate the correct values for analysis moisture at proficiency tests as a result above mentioned reasons. The correct measurement of moisture is important, because it plays a significant role in the calculation of the calorific values.

In measurement of carbon, nitrogen, hydrogen and sulphur different standard methods were used (Appendices 5.1 and 5.3). Only for the results of sulphur in the sample K1 there was the statistically significant difference between the standards ISO 334 and ASTM D 4239 (Appendix 5.3). The difference can be due to the various measuring procedures. ASTM D 4239 is an instrumental procedure using high temperature tube furnace combustion and ISO 334 is Eschka method based on mainly gravimetric technique. ASTM D 4239 is a rapid method and ISO 334 is time consuming containing several analysis steps.

3.2.2 Gross and net calorific value

In the Appendices 5.1 and 5.3 are shown the results of the gross calorific values for the samples B1 and K1 with the reported information of the used standard method and the more detail information of the used methods. Basically, there was no clear difference between the gross calorific values obtained using the different standard methods and no statistically significant difference between the results were found. For the laboratories 26 (the sample K1 and B1) and 47 (the sample K1) the deviation of the net calorific value is evident due to the errors in the data reporting (Appendices 4.1, 8). In many cases, the anomalous result was explained by the errors in the data reporting and/or errors in the measurement. There was some variance between results, but mainly the gross calorific values and net calorific values were between the acceptable variances. The main mistakes might be due to erroneous data handling and calculations.

The reasons for deviated results are fairly complicated to obtain in this PT, because there might be several errors effecting on the final results at the same time.

There are several factors, which have to be taken into account measurement of calorific value:

- Sample should be mixed well before analyses are carried out.
- Analytical moisture and calorific value should be measured at a same time (within a few hours), if the gross calorific value is analyzed from air dried sample. Analytical moisture has a great effect for calculation the gross calorific value as a dry weight basis. The porous fuel material adsorbs moisture very easily and the changes in the moisture content of the laboratory air can cause inaccuracies to the calorific value reported as a dry weight basis.
- In measurement of calorific value from the dried sample, moisture can absorb into a sample very easily from a laboratory environment. The sample should be analyzed as soon as possible after drying or a dried sample should keep in a small, effectively closed small container until the calorific value analysis is carried out.

- The coal samples can be analyzed in a power form. Usually biomass and peat samples shall be tested in a pellet form or in power form in a closed combustion bag or capsule. In the future, it could be valuable to ask participants report more information from the way of sample handling in the analyzing step as this can be an important issue in the verifying of test results.
- If the sample contains a high amount of sulphur and nitrogen, the correction for sulphur and total acids can affect to a great extent, especially in the case of coal samples.
- The laboratory has to take into account the calibration conditions, whether benzoic acid has been weighed in air or in vacuum (on the basis of a certificate). Further, in the measurement of the sample the conditions should be similar as during the calibration process (e.g. a pressure, an amount of calorimeter water, a correction for total acids).
- The mass of the sample (g) has to be adequate to meet the valid temperature rise and the linear calibration range.
- Stability of the calorimeter has to be checked before sample measurements with the certified benzoic acid.
- The calculation of gross and net calorific value should be based on the formulas of the international standards. If in the calculation any literature values for the parameters needed are used, those should be reported with the calorific values. To get more accurate results the measured parameters for the correction parameters are recommended to use.

3.3 Uncertainties of the results

Several approaches were used for estimating of measurement uncertainty (Appendix 10). The approach based on X-chart (Meth 2), existing IQC and validation data (Meth 3) or CRM data (Meth 4) were most common. Generally, the approach for estimating measurement uncertainty has not made a definite impact on the uncertainty estimates

From 6 to 30 laboratories reported the expanded measurement uncertainties with their results (Table 4, Appendix 11). The estimated uncertainties varied greatly, e.g. for coal (the sample K1) from 0.02 to 30 %. For the calorific value the uncertainty variation was also very large. Typically, about a half of the reported calorific value uncertainties were higher than the requirements for repeatability presented in the standard methods [5, 6].

Particularly, very low uncertainties (around 0.01 %) can be considered as questionable. Possibly, some uncertainties have been wrongly reported, not as percent as the provider of this PT had requested. In many other cases, the reported measurement uncertainties did not meet the requirements presented in the standard methods for the repeatability of the method. On the other hand, almost for each measurement also extremely high measurement uncertainties have been reported (Appendix 11).

As indicative information also the calculated zeta scores are reported in Appendix 11. These have been informed to the participants with the preliminary results as well. In the calculation of zeta score both the uncertainty of assigned value and the measuring uncertainties reported by the laboratories has been taken account (Appendix 7). As there were reported very variable and some cases high measuring uncertainties, no performance evaluation based on the zeta score was performed (see Appendix 4.2). However, as an example there can be seen some cases where the reported high uncertainty of laboratory has really affected the zeta score and improved the performance of the participant comparing with the z score (i.e. lab 20 for C in the sample B1, Appendix 11). On the other hand, very low measuring uncertainties have resulted in poor performance in the use of the zeta score (i.e. lab 51 for q-V,gr,d in the sample B1) and a satisfactory performance based on z score (Appendix 11). The participants can evaluate their performances based on these scores. In the case that the participant provided data of good quality, there is not a large deviation between the z and zeta scores. Thus this information could be useful when laboratories re-evaluate the measurement

uncertainties [23].

Based on the reported measuring uncertainties it is evident that harmonization in the estimating of uncertainties should be continued.

3.4 Estimation of emission factor

Additionally, the laboratories were asked to estimate the emission factors for the samples distributed in the PT by taking into account their own net calorific values and the total moisture values as received 44.3 % for the peat sample, B1, and 11.2 % for the coal sample, K1, which informed in the covering letter of the samples. In total, 14 laboratories reported the emission factor in measurement of the peat sample, and 23 laboratories reported it for the coal sample (Table 1, Appendix 8).

According to the evaluation of results at least laboratories 34 and 43 had not calculated the emission factor for peat sample B1 as correctly (based on the total moisture as received). The case was the same for the emission factor of coal sample K1 reported by the laboratory 43. These results were excluded from the assigned value calculation. For these laboratories the performance evaluation of emission factor is not satisfactory, and thus weakening the reliability of the performance evaluation of EF as a whole.

The participants were asked to calculate EF-values using the equation presented in the EC directive 2007/589/EC [24]. Mainly the participants informed that the calculation of EF-value was based on the EC directive 2007/589/EC (Appendix 5.1). Some national guides of the equation for the calculation of EF-value are available (e.g. in Finland). In Finland the Energy Market Authority has made the guideline for the calculation of emission factor for fossile fuels. (<http://www.energiemarkkinavirasto.fi/files/Paastokerroin11112008.pdf>). This is presented in the Appendix 5.1. The one aim has been to harmonized the used equation for the calculation of EF values within the Finnish accredited laboratories.

The emission factors are used in the European emission trading of the energy. This PT showed that the common procedure for calculation of EF-values within the different EU countries is urgently needed.

4 EVALUATION OF PERFORMANCE

The evaluation of the participants was based on z scores, which were calculated using the estimated target values for the total deviation. The calculated z scores are presented with the results of each participant (Appendix 8) and the summary of z scores is presented in Appendix 9. Additionally, zeta scores have been reported with the preliminary results for those laboratories, which have reported the measuring uncertainties. These sores are only informative and no performance evaluation based on those was done (Appendix 11).

The total number of laboratories participating in this PT was 58. The robust standard deviation of the results was mostly lower than 7.5 %, while for the calorific values it was lower than 1 %.

The criteria for performance had been mainly set according to the target value for reproducibility recommended in the international standards or technical specifications for measurement of the calorific values and other determinants. The reproducibility required in the standards was fulfilled for the gross calorific values. For the net calorific value reproducibility was somewhat increased from the value for the gross caloric value partly due to the variability of the results and the missing clear reproducibility information for the net caloric value in the standard methods. For some other measured parameters (i.e. C, H, N, S and ash) total standard deviation for proficiency assessment

had to be increased from the reproducibility of standards (Table 1). The reliability of the performance evaluation of EF was weak due to the errors in the calculation of the emission factor (EF).

Peat

Accepting the deviations of 1.4–15 % from the assigned values for the peat sample (B1) 82 % of results were satisfactory (Table 1). In the measurement of ash, H and S at least 90 % of the results were satisfactory. In the measurement of gross and net calorific values 77 % and 74 % of results, respectively, were satisfactory when accepting the deviations 1.4 % and 1.8 % from the assigned values. In this PT the number of satisfactory results of the calorific values for the peat sample was in the same range than in the previous PT 5/2009 [22]. There were more difficulties in the estimation of EF, where less than 64 % of results were satisfactory.

Coal

Accepting the deviations of 1–15 % from the assigned values for the coal sample (K1) 85 % of results were satisfactory. For the emission factor, EF, the performance evaluation was somewhat non-reliable probably due to the erroneous calculation. In the measurement of ash, H and N over 90 % of the results were satisfactory. In the measurement of gross and net calorific values 71 % of results were satisfactory, when accepting the deviations 1 and 1.3 % from the assigned values. In this PT the number of satisfactory result of the gross calorific value was higher than in the previous PT/2009 (65 %), while the satisfactory net calorific value was in the same range than in the previous test [22].

This PT showed again that the common procedure for calculation of EF-values is not available at this moment. However, the performance for EF in the peat sample was in the same range than in the previous PT5/2009. Thus, it is urgently needed harmonized equation for the calculation of EF-values within the EU countries.

In total, 84 % from the results were satisfactory when the deviations of 1–15 % from the assigned values were accepted. About 78 % of the participants used the accredited methods and 87 % of their results were satisfactory. SYKE arranged a similar proficiency test in 2009 and then 83 % of the results were satisfactory [22].

5 SUMMARY

Proftest SYKE carried out the proficiency test for measurement the gross and the net calorific value, the content of ash, carbon, nitrogen, hydrogen, moisture and sulfur in fuels in September 2010. One peat sample and one coal sample were delivered to the laboratories for the analysis of each measurement. In total, 58 laboratories participated in the proficiency test. The test performed as the joint work with the working group (WG) of the European co-operation for Accreditation (EA) for Interlaboratory Comparisons (EA WG ILC in Testing) and the data will be confidentially handled in the work of this WG.

The robust means of the reported results by the participants were used as the assigned values for measurements. The uncertainties of the calculated assigned values were mainly less than 0.3 % for calorific values and at maximum 5.6 % for the other measurements.

The evaluation of performance was based on the z score which was calculated using the standard deviation for proficiency assessment at 95 % confidence level. The evaluation of performance was not done for the measurement of moisture. In total, 84 % of the participating laboratories reported the satisfactory results when the deviations of 1–15 % from the assigned values were accepted.

About 78 % of the participants used the accredited methods and 87 % of their results were satisfactory. In measurement of the gross calorific value from the peat sample 77 % of the results were satisfactory and respectively in measurement of the coal sample 71 % from the results were satisfactory. In measurement of the net calorific value from the peat sample 74 % of the results were satisfactory and respectively in measurement of the coal sample 71 % from the results were satisfactory.

This PT showed that the common procedure for calculation of EF-values is not available at this moment. The emission factors are used in the European emission trading of the energy. Thus, it is urgently needed harmonized equation for the calculation of EF-values within the EU countries.

6 YHTEENVETO

Suomen ympäristökeskus (Proftest SYKE) järjesti syyskuussa 2010 pätevyyskokeen kalorimetrisen ja tehollisen lämpöarvon sekä tuhkan, vedyn, typen, rikin ja kosteuden määrittämiseksi turpeesta ja kivihiilestä. Pätevyyskoe järjestettiin yhteistyössä Euroopan akkreditoointielinten yhteistyöjärjestön (EA) testauksen vertailumittaustyöryhmän kanssa (EA WG ILC). Pätevyyskokeen tuloksia käytetään jatkossa luottamuksellisesti akkreditoinnin toiminnan seurantaan kansainvälisellä tasolla.

Pätevyyskokeeseen osallistui yhteensä 58 laboratoriota. Laboratorioiden pätevyyden arviointi tehtiin z-arvon avulla ja sen laskemisessa käytetyn kokonaishajonnan tavoitearvot olivat määrityksestä riippuen välillä 1–15 %. Mittaussuureen vertailuarvona käytettiin osallistujien ilmoittamien tulosten robustia keskiarvoa. Tavoitearvon epävarmuus oli lämpöarvon määritykssä alhaisempi kuin 0.3 % ja muiden määritysten osalta korkeintaan 5.6 %. Tulosten arviointia ei tehty kosteuspitoisuuden määrityksselle, koska tulosten hajonta oli suuri. Arviointi on jonkin verran epävarma hiilen päästökertoimelle, koska kaikki laboratoriot eivät olleet laskeneet arvoa tulokosteutta kohti.

Koko tulosaineistossa hyväksyttäviä tuloksia oli 84 %, kun vertailuarvosta sallittiin 1–15 %.n poikkeama. Noin 78 % osallistujista käytti akkreditoituja määritysmenetelmiä ja näistä tuloksista oli hyväksyttäviä 87 %. Kalorimetrisen lämpöarvon tuloksista oli hyväksyttäviä 77 % (turve) ja 71 % (kivihiili). Tehollisen lämpöarvon tuloksille vastaavat hyväksyttävien tulosten osuudet olivat 74 % (turve) ja 71 % (kivihiili).

Pätevyyskokeessa havaittiin, että selvää laskentakaavaa päästökertoimelle ei ole kuvattuna direktiivissä 2007/589/EC [24]. Esimerkiksi Suomessa on tehty kansallinen ohjeistus kiinteiden fossiilisten polttoaineiden päästökertoimen laskentaan. Yhtenäinen ohjeistus päästökertoimen laskennalle eri EU-maissa todettiin puuttuvan, mistä johtuen laskentatapa vaihtelee. Päästökerrointa käytetään Euroopan laajuisessa energian päästökaupassa. Täten yhtenäisen, dokumentoidun, laskentakaavan käyttöönotto EU-laajuisesti on erityisen tärkeää.

7 REFERENCES

1. Proficiency Testing by Interlaboratory Comparison – Part 1: Development and Operation of Proficiency Testing Schemes. 1996. ISO/IEC Guide 43-1.
2. ISO/IEC 17043, 2010. Conformity assessment – General requirements for proficiency testing.
3. ISO 13528, 2005. Statistical methods for use in proficiency testing by interlaboratory comparisons.

4. Thompson, M., Ellison, S.L.R., Wood. R., 2005. The International Harmonized Protocol for the Proficiency Testing of Analytical Chemistry laboratories (IUPAC Technical report. Draft). International Union of Pure and Applied Chemistry. Analytical, Applied and Clinical Chemistry Division. Interdivisional Working Party for Harmonization of Quality Assurance Schemes for Analytical Laboratories.
5. CEN/TS 14918, 2005. Solid Biofuels. Method for the determination of calorific value.
6. ISO 1928, 1995. Solid mineral fuels- Determination of gross calorific value by a bomb calorimetric method, and calculation of net calorific value.
7. DIN 51900, 2000. Testing of solid and liquid fuels - Determination of gross calorific value by the bomb calorimeter and calculation of net calorific value.
8. ASTM D 5865-07, 2007. Standard Test Method for Gross Calorific Value of Coal and Coke.
9. CEN/TS 15104, 2005. Solid biofuels. Determination of total content of carbon, hydrogen and nitrogen. Instrumental methods.
10. ISO/TS 12902, 2001. Solid mineral fuels - Determination of total carbon, hydrogen and nitrogen - Instrumental methods
11. ASTM D 5373, 2008. Standard Test Methods for Instrumental Determination of Carbon, Hydrogen, and Nitrogen in Laboratory Samples of Coal and Coke.
12. ASTM D 4239, 2008. Standard Test Methods for Sulfur in the Analysis Sample of Coal and Coke Using High - Temperature Combustion and Infrared Absorption.
13. CEN/TS 15289, 2006. Solid biofuels - Determination of total content of sulphur and chlorine.
14. ISO 334, 1992. Solid mineral fuels - Determination of total sulfur - Eschka method.
15. CEN/TS 14774-3, 2005. Solid biofuels. Methods for the determination of moisture content. Oven dry method. Part 3: Moisture in general analysis sample.
16. ISO 589. 2003. Hard coal - Determination of total moisture.
17. DIN 51718. 2002. Determining the moisture content of solid fuels.
18. ASTM D 5142. 2009. Standard Test Methods for Proximate Analysis of the Analysis Sample of Coal and Coke by Instrumental Procedures.
19. CEN/TS 14775, 2005. Solid biofuels. Method for the determination of ash content
20. ISO 1171, 1997. Solid mineral fuels. Determination of ash.
21. DIN 51719. 1997. Determination of ash in solid mineral fuels.
22. Leivuori, M., Mäkinen, I., Rantanen, M., Salonen, M. Korhonen, K. and Ilmakunnas, M. 2009. SYKE Proficiency Test 5/2009. Gross and net calorific value in fuels. Reports of Finnish Environment Institute 2/2010. (<http://www.environment.fi/download.asp?contentid=115724&lan=en>)
23. Mäkinen, I. 2009. Use of uncertainty estimates as reported by participants in proficiency testing for the evaluation of their results: pros and cons. Accred Qual Assur (2009) 14:35-41.

24. 2007/589/EC (18.7.2007). Commission Decision of 18 July 2007 establishing guidelines for the monitoring and reporting of greenhouse gas emissions pursuant to Directive 2003/87/EC of the European Parliament and of the Council.

PARTICIPANTS IN THE PROFICIENCY TEST 5/2010

Air pollution national research and development institute for industrial ecology, Bucarest, Romania
Ambitec Laboratorio Medioambiental, Madrid, Spain
Analisi & Controlli, Genova, Italy
BELAB AB, Norrköping, Sweden
Cementa Reasearch, Östersund, Sweden
Central Laboratory Bayer Antwerpen NV, Belgium
Čez, a.s., Elektrárna Počerady, AZL, Czech Republic
Co "Akmenés CEMENTAS" cement testing laboratory, Naujoji Akmene, Lithuania
CTG, Bergamo, Italy
Edenderry Power Operations Ltd, Edenderry, Republic of Ireland
Eesti Energia Narva Elektrijaamad AS Chemical lab. BEJ, Narva, Estonia
Eesti Energia Narva Elektrijaamad AS Eesti elektrijaama keemialabor, Narva, Estonia
Ekokem Oy Ab, Riihimäki, Finland
Elektrocieplownia EC Nowa Sp. z.o.o., Dąbrowa Górnica, Poland
Enas Oy, Jyväskylä, Finland
Eurofest - control EAD, Sofia, Bulgaria
Eurofins Environment Sweden AB, Lidköping, Sweden
Finnsementti Oy, Parainen, Finland
Geological analysis laboratory, Bucharest, Romania
Helsingin Energia, Power Plant Chemistry, Helsinki, Finland
HEP-Proizvodnja d.o.o. CKTL, Zagreb, Croatia
HERKON d.o.o., Mostar, Bosnia and Herzegovina
Hjortens Laboratorium AB, Östersund, Sweden
Inspectorate Estonia AS, Viimsi Vald, Estonia
Inspekt RGH d.o.o. - Testing laboratory Kakanj, Sarajevo, Bosnia and Herzegovina
Intertek Belgium NV, Antwerpen, Belgium
ISD Dunaferr Ztr. Coal Chemistry Material Testing Department, Dunaújváros, Hungary
JP Elektroprivreda d.d. Sarajevo, Z.D. RMU Kakanj d.o.o. Kakanj, E.J. Stručne službe-Služba Laboratorijska, Kakanj, Bosnia and Herzegovina
KCL Kymen Laboratorio Oy, Kuusankoski, Finland
Kraftwerk Mehrum GmbH, Hohenhameln, Germany
Kristiinan voimalaitos, Kristiinankaupunki, Finland
Kuopion Energia Oy, Kuopio, Finland
Laboratory for hard mineral fuels and mineral raw materials, Belgrade, Serbia
Laboratory of Fuel Technology and Lubricants, Athens, Greece
Lafarge Cementos S.A.U (Lab. Técnico Comercial), Madrid, Spain
LMA Laboratorium voor Multi-element Analyses, Sleeuwijk, Netherlands
METLA, Kannus, Finland
Mibrag, Zeitz, Germany
N.V. Inspectorate, Ghent, Belgium
Nab Labs Oy, Rauma, Finland
Pegop-Energia Eléctrica, Pego, Portugal
Ramboll Analytics, Vantaa, Finland
Rautaruukki Oyj Ruukki metals/production, Raahe, Finland
Stora Enso Oyj, Research Centre Imatra, R&D Services, Trouble shooting and Microscopy, Imatra, Finland
SGS, Immingham, England
SOCOR, Dechy, France
SP Technical Research Institute of Sweden, Borås, Sweden
Stazione sperimentale per i combustibili, Milano, Italy

Teknologiakeskus Ketek Oy, Kokkola, Finland

The Foundation of Agr. Tech. Commercialization and Transfer, Suin-ro Gwonseon Suwon-si, Korea

Tubitak-Bursa test and analysis Laboratory, Bursa, Turkey

TÜV NORD CZECH Laboratoře a zkušebny, Brno, Czech Republic

UPt Endesa, AS Pontes, A Corlina, Spain

Vaskiluodon Voima Oy, Seinäjoen voimalaitos, Finland

Vaskiluodon Voima Oy, Vaasa, Finland

Vattenfall A/S Amagerværket, København S, Denmark

VTT Expert Services Oy, Espoo, Finland

Vyzkumny ustav pro hnedé uhli, Most, Czech Republic

PREPARATION OF THE SAMPLES

Sample B1, peat

The sample B1 was prepared from the sample material taken from the Finnish marshal. The peat was dried at room temperature and grounded by a mill with 500 µm sieve at the laboratory of Enas. The dried and sieved sample was mixed by a mechanized sample mixer and distributed in sub samples of 50 g using a rotary sample divider equipped with a vibratory sample feeder at the laboratory of Water Protection Association of the Kokemäenjoki River. The particle size distribution of peat was measured by the laboratory of Enas using laser diffraction (Malvern).

Sample K1, steam coal fuel

The sample K1 was a Russian crushed coal. The coal was dried at room temperature and grounded to particle size < 200 µm at the Helsinki Energy. The dried and sieved sample was mixed by a mechanized sample mixer and distributed in sub samples of 50 g using a rotary sample divider equipped with a vibratory sample feeder at the laboratory the laboratory of Water Protection Association of the Kokemäenjoki River. The particle size distribution of coal was measured by the Helsinki Energia, Power Plant Chemistry using laser diffraction (Malvern).

TESTING OF THE SAMPLES

Homogeneity

Homogeneity was tested from duplicate measurements of calorific value and ash content in fifteen samples, which were homogenised before sampling (Table 1). The gross calorific value in the peat sample B1 was additionally tested from eight samples by other laboratory (2nd test results in Table 1) due to the some variability of the test results of the first analysing laboratory (1st test results in Table 1). In the 2nd test the sampling variation (s_{sam}) was much smaller than the variation obtained in the 1st test. In addition, carbon, hydrogen and nitrogen content from fifteen and sulphur content from eight samples were tested. The analytical variation s_{an} and the sampling variation s_{sam} was calculated using one-way variance analysis. For this proficiency test, the analytical results were statistically handled according to the IUPAC guidelines for the treatment of homogeneity testing data and the total standard deviation for proficiency assessment [4].

Table 1. Results from the homogeneity testing of the peat B1 and coal K1 samples.

Measurements	Value	$s_p\%$	s_p	s_{an}	s_{an}/s_p	Is $s_a/s_p < 0.5$?	s_{sam}	s_{sam}^2	c	Is $s_{sam}^2 < c$?
Peat (B1)										
Gross calorific value, J/g, (the 1 st test)	21701	0.69	150	43.5	0.29	yes	71.4	5110	5255	yes
Gross calorific value, J/g, (the 2 nd test)	21649	0.69	150	29.3	0.20	yes	9.22	85	5926	yes
Ash, w-%	7.75	1.9	0.15	0.056	0.37	yes	0.063	0.004	0.006	yes
Coal (K1)										
Gross calorific value, J/g	29801	0.50	150	32.9	0.22	yes	45.2	2042	4190	yes
Ash, w-%	10.9	1.4	0.15	0.014	0.093	yes	0.012	0.0001	0.004	yes

where,

s_p = standard deviation for proficiency assessment, (total standard deviation divided by 2)

$s_p\%$ = standard deviation for proficiency assessment as percent, (total standard deviation divided by 2)

s_{an} = analytical deviation, mean standard deviation of results in a sub sample

s_{sam} = sampling deviation, standard deviation of results between sub samples

$$c = F1 \cdot s_{all}^2 + F2 \cdot s_a^2$$

where:

$$s_{all}^2 = (0.3 \cdot s_i)^2$$

$F1 = 2.01$ when the number of sub samples is 8, $F2 = 1.25$ when the number of sub samples is 8

$F1 = 1.79$ when the number of sub samples is 12, $F2 = 0.86$ when the number of sub samples is 12

$F1 = 1.69$ when the number of sub samples is 15, $F2 = 0.71$ when the number of sub samples is 15


Conclusion: In each case, the criteria were fulfilled. Additionally, the results of the other tested parameters support the homogeneity of samples. The samples could be regarded as homogenous.

TESTING OF THE SAMPLES


Particle size

To test the particle size of samples one sample of each sample type was tested using laser diffraction (Malvern).

In Figure 1 is showing the distribution of particle size for the samples B1 and K1. For peat sample B1 the mean size of particles was 39.7 μm and 99.8 % of the particles were smaller than 550 μm . For coal sample K1 the mean size of particles was 60.5 μm and 100 % of the particles were smaller than 212 μm . For the both sample material the requirement of particle size given in the international standards was fulfilled [5, 6].



a) The particle size distribution of peat B1.



b) The particle size distribution of coal K1.

Figure 1. The particle size distribution of the fuel samples.

COMMENTS SENT BY THE PARTICIPANTS

Lab	Comment to the samples / PT	Action/Proftest
13, 16	The samples were delivered crosswise between the two local offices of the participants.	The local offices changed their samples themselves. The provider will be more carefully in the packing of samples.
23	The participants ordered both testing samples, but only the sample B1 (peat) delivered.	The sample B1 was unfortunately unpacked and the participant did not want the sample later. The missed sample was not charged. The provider will be more carefully in the packing of samples.
33, 39	The samples were delayed.	The local, national, post services had difficulties to deliver the packets to the participants.
42	The samples were delayed due to the incomplete address.	The samples were posted to the address which was informed on the registration form to the provider.

Lab	Comment to the results	Action/Proftest
26	<p>The results of q-V,gr,d and q-p,net,d for the samples were reported erroneously.</p> <p>The laboratory informed new results for the samples after receiving the preliminary results:</p> <p>B1: q-V,gr,d 21436 and q-p,net,d 20429 J/g,</p> <p>K1: q-V,gr,d 29753 and q-p,net,d 28746 J/g</p>	<p>The results were not corrected into the final data. They were outliers in the statistical treatment, and so they have not affected the performance evaluation.</p> <p>If the results should have been reported rightly they should have been satisfactory.</p> <p>The participant can re-calculate z scores according to the guide for participating laboratories in Proftest proficiency testing schemes (www.environment.fi/syke/proftest).</p>
33, 44	The results of calorific values were reported in MJ/g instead of the requested J/g.	The units were corrected by the provider. However, in the future proficiency tests the unit corrections will be not done.
46	<p>After sending the preliminary results the participants informed that they were reported M_{ad} in the sample K1 incorrectly.</p> <p>The right results were:</p> <p>1.82 % and 1.74 %</p>	The results were handled as outliers in the statistical treatment.
47	<p>The laboratory has not reported the results in dry weight basis (as analysed).</p> <p>The laboratory informed new results for the samples after receiving the preliminary results:</p> <p>K1: q-V,gr,d 29920 and q-p,net,d 28970 J/g, Ash% 10.9, S% 0.32</p>	<p>The results were handled as outliers in the statistical treatment. If the results should have been reported rightly they should have been satisfactory.</p> <p>The participant can re-calculate z scores according to the guide for participating laboratories in Proftest proficiency testing schemes (www.environment.fi/syke/proftest).</p>

COMMENTS TO THE PARTICIPANTS

Laboratory	Comments on results
28	The laboratory provided results several times; new results and corrections to the older ones. The provider recommended that the results will be provided at once and only minor corrections later, if needed.
30, 38, 58	Laboratories reported only one result, though replicate results were requested. These results were not included in the calculation of assigned values.
16, 31, 28	Laboratories reported four parallel results, though replicate results were requested. The two first results were included in the final database.
6, 25, 26, 36, 37, 51, 52, 54	The laboratories reported the measurement uncertainties as \pm J/g instead of UC %. The values were corrected by the provider. The provider did not assure that all erroneously reported values have been corrected. No comments to the uncertainty corrections were given from the participants after the preliminary results. The provider strongly recommended that the participants are more carefully to report the uncertainty correctly. The performance evaluation using zeta scores, which based on the measurement uncertainties, is unsure due to the errors in the result reporting by the participants.
-	After receiving the preliminary results several participants informed their missing data in the z and zeta score comparison (Appendix 6 in the preliminary results). These participants were not reported their measurement uncertainties with the results. As the zeta score is based on the uncertainties the data of non-reported participants were not reported in this Annex. The provider will more clearly describe the zeta score reporting in the covering letter of preliminary results.
-	Fewer errors than the previous test were found in the wrongly reported units, which is a good feedback. In the future wrong unit will be not corrected by the provider, unless a total amount of results is too low for the statistical calculations.

ANALYTICAL METHODS

Analyte	Code	Method
q-V,gr,d q-p,net,d	1	CEN/TS 14918
	2	ISO 1928
	3	DIN 51900
	4	ASTM D 5865
	5	Other
Ash	1	CEN/TS 14775
	2	ISO 1171
	3	DIN 51719
	4	ASTM D 5142
	5	Other
C, H, N	1	CEN/TS 15104
	2	ISO/TS 12902
	3	ASTM D 5373
	4	Other
S	1	CEN/TS 15289
	2	ISO 334
	3	ASTM D 4239
	4	Other
M_{ad}	1	CEN/TS 14774
	2	ISO 589
	3	DIN 51718
	4	ASTM D 5142
	5	Other

ANALYTICAL METHODS**Measurements:**

Measurement of gross calorific value	Sample B1	Sample K1
Date of analysis in 2010	20.9, 22.9, 21.9, 23.9, 28.9, 30.9, 1.10, 5.10, 6.10, 7.10, 29.9-4.10, 4-7.10, 9-10.10, 13.10	29.9/4.10, 27.9, 28.9, 22.9, 24.9, 26.9, 27.9, 29.9, 30.9, 1.10, 4-5.10, 4.10, 5.10, 6.10 29.9-4.10, 4-7.10, 5-7.10, 29-30.9, 30.9-1.10, 13.10
Sample amount	Mainly 0.5-1.2 g used < 0.5 g used: 0.3 g (lab 20), 0.2 g (lab 30)	Mainly 0.5-1 g used < 0.5 g used: 0.3 g (lab 20), 0.3-0.4 g (lab 46)
Drying of sample	Air dried: lab 7, 8, 9, 12, 15, 16, 21, 22, 24, 26, 31, 37, 57 At 105 °C dried: lab 4, 14, 27, 34, 38 Air dried & at 105 °C dried: lab 58 No drying: lab 5, 20, 29, 30, 40, 43, 60	Air dried: lab 2, 3, 4, 11, 12, 13, 15, 21, 22, 23, 24, 25, 28, 31, 36, 37, 42, 44, 45, 47, 48, 52, 57 At 105 °C dried: lab 13, 29, 34, 38, 39, 46, 58 Air dried & at 105 °C dried: lab 33, 49, 52 No drying: lab 5, 6, 20, 32, 40, 43, 49, 54, 56, 60
Equipment (manufacturer, volume of bomb and volume of added water)	14 labs: IKA (model C2000, C5000, C5003 or C7000) with different bomb volumes (200 – 260 ml) and with the added water of 1 or 5 ml 3 labs: PARR (model 1261, 1281, 6300 or 6400) with bomb volume 250 ml (or not given) and with the added water of 0 ml (lab 14, 16) 8 labs: LECO (model AC 300 or 350) with bomb volume 260 - 400 ml and with the added water of 1 ml or 5 ml	17 labs: IKA (model 2000, 5000, 5003 or 7000) with different bomb volumes (200 – 260 ml) and with the added water of 1 or 5 ml 10 labs: LECO (model AC 300, 350 or 500) with bomb volume 300 - 400 ml (2000 ml lab 6) and with added water of 1 – 10 ml 10 labs: PARR (model 1261, 1281, 6200, 6300 or 6400) with bomb volume 250-500 ml (or not given) and with the added water of 1 ml Sanio Gallenkamp with the bomb volume 250 ml and added water 1 ml (lab 3), 1 lab: Julius Peters with the bomb volume 301 ml and with the added water 10 ml 1 lab: Lagent MS with the bomb volume 300 ml and with the added water 1 ml 1 lab: Precizia KL 10 with the bomb volume 350 ml and with the added water 5 ml

ANALYTICAL METHODS

Measurements:

Measurement of gross calorific value	Sample B1	Sample K1
Calibration	<p>Benzoeacid</p> <p>4 labs: PARR, 26454, 26434.9 or 26432.1 J/g</p> <p>6 labs: BAS-BCS, 26439.7 J/g</p> <p>3 labs: IKA, 26456, 26460-26470 J/g</p> <p>6 labs: ALPHA, 26564, 26454 or 26457 J/g</p> <p>1 lab: NIST</p> <p>2 lab: Fluka, 26470 J/g</p> <p>4 lab: LECO, 26451, 26440, 28190 J/g</p> <p>1 lab: Federal state unitary enterprise, 26454 J/g</p> <p>1 lab: Sigma-Aldrich</p> <p>7 labs: as weighed</p> <p>9 labs: in air</p> <p>2 labs: in vacuum</p>	<p>Benzoeacid</p> <p>10 labs: PARR, 26454 or 26432.1 J/g</p> <p>8 labs: IKA, 26460, 26456 or 26556 J/g</p> <p>8 labs: BAS-BCS, 26439.7 J/g</p> <p>5 labs: ALPHA, 26434, 26454, 26464 or 26457 J/g</p> <p>3 labs: NIST, 26434 J/g</p> <p>4 lab: LECO, 26451, 26440 J/g</p> <p>1 lab: Fluka, 26470 J/g</p> <p>1 lab: Sigma-Aldrich</p> <p>1 lab: Riedel, 26453 J/g</p> <p>1 lab: BHD, 26434.9 J/g</p> <p>1 lab: Federal state unitary enterprise, 26454 J/g</p> <p>9 labs: as weighed</p> <p>19 labs: in air (lab 3, 11, 13, 15, 24, 28, 33, 34, 37, 38, 40, 41, 42, 45, 46, 48, 58, 60)</p> <p>7 labs: in vacuum</p>
	<p>Correction of the certified value¹</p> <p>17 labs: no</p> <p>9 labs: yes (lab 4, 5, 8, 16, 21, 22, 58, 59)¹</p>	<p>Correction of the certified value¹</p> <p>32 labs: no</p> <p>12 labs: yes (lab 4, 5, 12, 21, 22, 36, 48, 49, 58)¹</p>

¹Correction the value given in the certificate

lab 5: corrected value 26449 J/g (5 ml water)

lab 8: mean has been calculated from the control sample charts (x-charts)

lab 1, 12, 59: volume, temperature, pressure

lab 16: corrected value 26442 J/g

lab 21: It has been used as an input reference value for the calorimeter

lab 36: wire and N in air

lab 4, 22, 48, 49, 58: with factor that is calculated according to the certificate

ANALYTICAL METHODS**Calculations:**

Gross calorific value	Sample B1	Sample K1
Correction taken into account in calculations	wire, ignition, acid, moisture (lab 60) wire, ignition, S, acid (lab 4) moisture (lab 24) wire, moisture (lab 9) analysis moisture (lab 7) wire, S, acid (lab 30, 57) wire, S, acid, moisture (lab 11, 16, 20, 22, 26, 27, 59) wire, S and N (lab 15) N, S, moisture (lab 31) cotton, wire, moisture, (lab 8) wire, S, moisture (lab 2, 29) wire, moisture, N, S (lab 5) not in detail (lab 34, 37, 43) fixed acid correction (lab 14) wire, cotton, S, combustion aid (dodecane) (lab 40)	wire, ignition, acid, moisture (lab 60) wire, ignition, S, acid, moisture (lab 4, 5, 41) wire, ignition, S, moisture, acid (lab 2) not in detail (lab 11, 34, 37, 43) S, N, moisture (lab 24, 31) wire, acid (lab 22) wire, S (lab 48) wire (lab 53) wire, S, acid (lab 33, 46, 47, 52, 56, 57) wire, N (lab 36) wire, S, acid, moisture (lab 12, 20, 22, 28, 32, 54) correction 156 J/g (lab 38) wire, S, N (lab 15) wire, acid, moisture (lab 13) wire, N, S, moisture (lab 5, 29) wire, S, moisture (lab 21, 25, 42) wire, cotton, moisture (lab 3, 44) wire, cotton, S, combustion aid (dodecane) (lab 40) wire, paper, S, acid (lab 45) moisture (lab 49) wire, moisture, paper (lab 50)
Net calorific value	Sample B1	Sample K1
Correction taken into account in calculations	N+O, H/measured (lab 1, 4, 40) H/fixed (lab 7, 8 (5.60%)) H/measured (lab 22, 30) N+O (40 lab 24), H (6, lab 24) N+O, H/calculated (lab 27) [N+O]=100 % -[ash-C-H-S] (lab 21) not in detail (lab 34, 37, 43) S, H measured (lab 9) N+O, H (lab 29, 60) only H (lab 12, 3, 59) N, H fixed (1.22 MJ/kg) N-5.6% (fixed)+N%=35% (lab 16) S, N, acid, wire, moisture (lab 26) moisture, acid, ash, H, N/measured (lab 57)	DIN (212 J/% for H, 24.4 J/% for H ₂ O (lab 2) N+O, H/measured (lab 1, 4, 22, 40) N+O, H/calculated (lab 41) not in detail (lab 5, 6, 11, 34, 37, 43 and 38) H/measured (lab 3, 50, 12) H/calculated (lab 25, 46, 53) H/calculated from volatiles (lab 13) N+O]=100 % -[ash-C-H-S] (lab 21) only H (lab 12, 6:lab 24, lab 31) N+O, H (lab 29, 60) %H ₂ O=8.94*H(%), heat of vaporization of water (25°C,p) 43985 J/mol (lab 32) N (59 kJ/mol), S (55.45 J/mol), H (lab 36) Fixed H, ash, moisture (lab 42) H, N (lab 45) H, O, moisture (lab 48) H 5.8%, O 32%, N 1.5% use in calculation ISO1928:1995, annex E3.2 (lab 49) N+O, H ISO 1925 (lab 52) H, O (lab 56) moisture, acid, ash, H, N/measured (lab 57)

ANALYTICAL METHODS

Measurements:

Measurement of C, H and N	Sample B1	Sample K1
Sample amount	1 – 350 mg (depending on an equipment), 1 g (N), 796 mg	1 – 250 mg (depending on an equipment), 1 g (N) or > 1g, 919 mg
Equipment (manufacturer)	3 labs: VarioMAX CHN 2 labs: Vario EL 6 labs: LECO CHN or CS labs: ELTRA CHS-500 1 lab: CE Instruments 1 lab: EURO EA 1 lab: Skalar Primacs SLC (only C) 1 lab: Tru Spe CHN 1 lab: Jena multi EA 2000 (only C) 1 lab: Dionex ICSS2000 (only N) 1 lab: Martinek analyzer (C, H) 1 lab: Pramas-Vagner (N) 1 lab: Dennstedt oven (C, N)	3 labs: VarioMAX CHN 2 labs: Vario EL 10 labs: LECO CHN, CHN/S, TC, CR or CS 3 labs: ELTRA CHS 4 lab: Vario MACRO 1 lab. Bichi, KjellFix 1 lab: EURO EA 1 lab: CE Instruments 1 lab: Skalar Primacs SLC (only C) 1 lab: Martinek analyzer (C, H) 1 lab: Pramas-Vagner (N) 1 lab: Dennstedt oven (C, N) 1 lab: TruSpec CHN (N) 1 lab: Thermo Finning Flash 1112 1 lab: Elemental analyzer Flash EA 1112, Thermo C, H calculated (lab 47)
Measurement of S	Sample B1	Sample K1
Sample amount	Mainly 0.1-0.5 g (depending on an equipment), 6 g (lab 57) Eschka: 1 g (lab 31)	Mainly 0.1-0.4 g (depending on an equipment), 1 mg (lab 3), 1 g (lab 24, 3), 70 mg (lab 52), 6 g (lab 57) Eschka method: 1 g (lab 49)
Equipment (manufacturer)	2 labs: ELTRA CS 8 labs: LECO SC or S 1 lab: IKA AOD 1 lab: EURO EA 1 lab: LECO CHNS 1 lab: Oxidation + ion chromatography (lab 21) 1 lab: Iris Interip thermo elementar (lab 9) 1 lab: Jena multi EA 2000 1 lab: Mitsubishi TS-100V 1 lab: Esche mixture 1 lab: EDXRF (lab 57)	6 labs: ELTRA CS or CHS 14 labs: LECO, SC, S or CHN/S 1 labs: Vario EL 1 lab: EURO EA 1 lab: LECO CHNS 1 lab: Oxidation + ion chromatography (lab 21) 1 lab: Muffle furnace F6000, Vamstead/Thermolyne Corporation 1 lab: S.I.E. 1 lab: Esche mixture 1 lab: Parr 1261 + ion chromatography 1 lab: Elemental analyzer Flash EA 1112 + FDP 1112, Thermo 1 lab: EDXRF (lab 57) 1 lab: carbolite tube furnace

ANALYTICAL METHODS

Measurements:

Measurement of ash content	Sample B1	Sample K1
Sample amount	0.6-4 g (mainly 1 g)	0.15-5 g (mainly 1 g)
Measurement and temperature	Measurement 19 labs: Gravimetric 5 labs: TGA (lab 15, 21, 37) 1 lab: Tgl (lab 29) Temperature 14 labs: 550 °C 1 lab: 525 °C (lab 9) 7 labs: 815 °C (lab 8, 14, 16, 38, 57, 60) 1 lab: 925 °C (lab 21) 1 lab: 950 °C (lab 26) 1 lab: 725 °C (lab 43) 1 lab: 740 °C (lab 58, no method)	Measurement 35 labs: Gravimetric 2 labs: TGI (lab 3, 29) 5 lab: TGA (lab 15, 21, 36, 45) 1 lab. MAC-400 (lab 37) Temperature 30 labs: 815 °C 4 lab: 750 °C (lab 6, 28, 32, 54) 2 lab: 550 °C (lab 20, 49) 1 lab: 925 °C (lab 21) 1 lab: 960 °C (lab 36) 1 lab: 820 °C (lab 42) 1 lab: 725 °C (lab 43) 1 lab: 800 °C (lab 46) 1 lab: 740 °C (lab 58, no method)
Measurement of moisture	Sample B1	Sample K1
	17 an air (lab 4, 5, 7, 9, 12, 15, 16, 20, 21, 22, 24, 31, 34, 38, 57, 58) 7 lab an N ₂ atmosphere (26, 27, 29, 37, 40, 43, 60)	25 an air (lab 3, 4, 13, 15, 20, 21, 22, 23, 24, 28, 31, 33, 34, 38, 39, 41, 42, 44, 45, 46, 48, 49, 50, 52, 53, 57, 58) 14 lab an N ₂ atmosphere (lab 2, 4, 11, 12, 25, 29, 32, 36, 37, 40, 43, 47, 54, 56, 60)

Calculations:

Emission factor (EF) ¹	Sample B1	Sample K1
Equation according to the decision 2007/589/EC	9 labs: According to EC decision ¹ 1 lab: national (lab 40) 1 lab (lab 21): based on the equation: $1000 \times 3.664 \times (C/100) \times (1 - (50.8/100)/Q(p_{net}, 50.8\%))$	15 labs: According to EC decision 1 labs: national (lab 40) 1 lab (lab 29): based on the equation: $1000 \times 3.664 \times (C/100) \times (1 - (8.27/100)/Q(p_{net}, 8.27\%))$ 1 lab: EU. D. 2003/87/CE 1 lab: $F = C(tC/t_{com}) \times 3.664 / Pci$ (lab 32)

¹In the sample letter the provider gave a possibility to the participants to calculate the EF-value using the procedure presented in the EC directive and using the total moisture content as received presented in the letter. Later has been obtained, that in the EC directives has not been given the detailed equation for calculation of EF-values. However, a written description has been given. Due to this some national guides for the equation of EF-value calculation has been produced.

In Finland the Energy Market Authority has made the guideline for the calculation of emission factor for fossile fuels as follows (<http://www.energiemarkkinavirasto.fi/files/Paastokerroin11112008.pdf>):

$$EF = 1000 \times 3.664 \times (C/100) \times (1 - M_{ar}/100)/Q_{net,ar}, \text{ where}$$

EF emission factor, g CO₂/MJ

C carbon content as dry, %

M_{ar} total moisture as received, %

Q_{net,ar} net calorific value as received, MJ/kg

SIGNIFICANT DIFFERENCES IN THE RESULTS REPORTED USING DIFFERENT STANDARD METHODS

In the statistical comparison of the methods has included the data, in which the number of the results was ≥ 3 .

Analyte	Sample	Method	X	sd	n	Significant difference
S	K1	2. ISO 334	0.25	0.06	6	X: meth 2-3
		3. ASTM D 4239	0.31	0.01	15	
M_{ad}	K1	2. ISO 589	3.43	0.19	9	X: meth 2-3, 3-4
		3. DIN 51718	3.76	0.18	4	
		4. ASTM D 5142	3.63	0.13	7	


where, X: the mean value

sd: the standard deviation

n: the number of the result


LIITE 5.3. RESULTS GROUPED ACCORDING TO THE METHODS

APPENDIX 5.3. Method code - see appendix 5.1




Analyytti (Analyte) C,d

Näyte (Sample) K1




Analyytti (Analyte) H,d


Näyte (Sample) B1





Analyytti (Analyte) H,d

Näyte (Sample) K1










Analytti (Analyte) S,d

Näyte (Sample) K1



EVALUATION OF THE ASSIGNED VALUES AND THEIR UNCERTAINTIES

Analyte	Sample	Unit	Assigned value	Estimation of assigned value	Uncertainty ($U = 2*u$) ¹⁾ , %	u/s_p ²⁾
Ash	B1	w%	7.7	Robust mean	1.2	0.2
	K1	w%	10.9	Robust mean	0.34	0.1
C	B1	w%	53.4	Robust mean	0.89	0.3
	K1	w%	73.3	Robust mean	0.52	0.2
H	B1	w%	5.6	Robust mean	1.7	0.3
	K1	w%	4.6	Robust mean	1.8	0.3
EF	B1	tCO ₂ /TJ	106	Robust mean	2.5	0.2
	K1	tCO ₂ /TJ	93.8	Robust mean	2.3	0.2
M	B1	w%	7.68	Robust mean	3.9	-
	K1	w%	3.17	Robust mean	5.6	-
N	B1	w%	1.62	Robust mean	2.8	0.2
	K1	w%	2.14	Robust mean	3.5	0.2
q-p,net,d	B1	J/g	20460	Robust mean	0.25	0.1
	K1	J/g	28810	Robust mean	0.28	0.2
q-V,gr,d	B1	J/g	21680	Robust mean	0.23	0.2
	K1	J/g	29820	Robust mean	0.18	0.2
S	B1	w%	0.25	Robust mean	3.3	0.2
	K1	w%	0.31	Robust mean	2.3	0.2

The expanded uncertainty of the assigned value¹⁾ was estimated according to the equation [3]:

$$U\% = \frac{100 \times \left(\frac{2 \times 1.25 \times s_{rob}}{\sqrt{n}} \right)}{AV}$$

where,

U% = the expanded uncertainty of the assigned value

n = the number of the results

s_{rob} = the robust standard deviation

AV = the assigned value

To test the reliability of uncertainty of assigned value the ratio, u/s_p ²⁾, was calculated [4], where:

s_p = the total standard deviation for proficiency assessment divided by 2

u = the standard uncertainty of the assigned value

If $u/s_p \leq 0.3$ the assigned value is reliable and the z scores are qualified.

TERMS IN THE RESULT TABLES

Sample	The code of the sample
z-Graphics	z score - the graphical presentation
z score	calculated as follows: $z = (x_i - X)/s_p$, where x_i = the result of the individual laboratory X = the assigned value s_p = the target value of the standard deviation for proficiency assessment.
zeta score	$\text{zeta} = (x_i - X)/\sqrt{u_{lab}^2 + u_c^2}$, u_{lab} = the standard uncertainty of the participant's result u_c = the standard uncertainty of the assigned value
Outl test OK	yes - the result passed the outlier test H = Hampel test (a test for the mean value)
Assigned value	the reference value
Assigned value 2^*U_c	the expanded uncertainty of the assigned value
2^*Targ SD \%,	the target value of total standard deviation for proficiency assessment (s_p) at
Targ 2SD%	95 % confidence level
Lab's result	the result reported by the participant (the mean value of the replicates)
Md.	Median
Mean	Mean
Robust mean	Robust mean
SD	Standard deviation
SD%	Standard deviation, %
SD %rob	Robust standard deviation, %
Passed	The results passed the outlier test
Missing	i.e. < DL
Num of labs	the total number of the participants
Nr	the number of the result in the diagram (<i>Appendix 11</i>)

Summary on the z scores

S – satisfactory ($-2 \leq z \leq 2$)

Q – questionable ($2 < z < 3$), positive error, the result deviates more than $2 * s_p$ from the assigned value

q – questionable ($-3 > z < -2$), negative error, the result deviates more than $2 * s_p$ from the assigned value

U – unsatisfactory ($z \geq 3$), positive error, the result deviates more than $3 * s_p$ from the assigned value

u – unsatisfactory ($z \leq -3$), negative error, the result deviates more than $3 * s_p$ from the assigned value

Robust analysis

The items of data is sorted into increasing order, $x_1, x_2, x_i, \dots, x_p$.

Initial values for x^* and s^* are calculated as:

$$X^* = \text{median of } x_i \quad (i = 1, 2, \dots, p)$$

$$s^* = 1.483 \text{ median of } |x_i - x^*| \quad (i = 1, 2, \dots, p)$$

For each x_i ($i = 1, 2, \dots, p$) is calculated:

$$x_i^* = x^* - \varphi \quad \text{if } x_i < x^* - \varphi$$

$$x_i^* = x^* + \varphi \quad \text{if } x_i > x^* + \varphi$$

$$x_i^* = x_i \quad \text{otherwise}$$

The new values of x^* and s^* are calculated from:

$$x^* = \sum x_i^* / p$$

$$s^* = 1.134 \sqrt{\sum (x_i^* - x^*)^2 / (p-1)}$$

The robust estimates x^* and s^* can be derived by an iterative calculation, i.e. by updating the values of x^* and s^* several times, until the process converges.

Ref: Statistical methods for use in proficiency testing by inter laboratory comparisons, Annex C [3].

LIITE 8. RESULTS OF EACH PARTICIPANT

APPENDIX 8.

Analyte	Unit	Sample	z-Graphics							Z-value	Outl. test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
Laboratory 1																						
Ash,d	w%	B1		-0.649	yes	7,7	6	7,85	7,7	7,696	0,1756	2,3	27	3	0	30						
	w%	K1		-0,514	yes	10,9	2,5	10,83	10,91	10,88	0,1371	1,3	47	1	0	48						
C,d	w%	B1		1,124	yes	53,4	3	54,3	53,61	53,41	0,7942	1,5	20	2	0	22						
	w%	K1		-0,218	yes	73,3	2,5	73,1	73,25	73,22	0,9928	1,4	36	2	0	38						
H,d	w%	B1		0,355	yes	5,63	7	5,7	5,675	5,629	0,1487	2,6	18	2	0	20						
	w%	K1		-1,242	yes	4,6	7	4,4	4,57	4,596	0,209	4,5	31	1	0	32						
Mad,d	w%	B1			yes	2,88		3,8	2,87	2,852	0,4869	17,0	28	2	0	30						
	w%	K1			yes	3,52		3,5	3,53	3,507	0,301	8,6	48	1	0	49						
N,d	w%	B1		0,658	yes	1,62	15	1,7	1,601	1,621	0,07511	4,6	17	3	0	20						
	w%	K1		0,374	yes	2,14	15	2,2	2,17	2,137	0,1439	6,7	28	2	0	30						
q-p,net,d	J/g	B1		0,831	yes	20460	1,8	20610	20460	20490	165,3	0,8	21	6	0	27						
	J/g	K1		-0,190	yes	28810	1,3	28770	28820	28790	213,3	0,7	34	8	0	42						
q-V,gr,d	J/g	B1		1,064	yes	21680	1,4	21840	21700	21690	161,4	0,7	26	5	0	31						
	J/g	K1		-0,644	yes	29820	1	29720	29820	29810	164,9	0,6	41	8	0	49						
S,d	w%	B1		0,533	yes	0,25	15	0,26	0,248	0,2504	0,0173	6,9	23	2	0	25						
	w%	K1		0,430	yes	0,31	15	0,32	0,31	0,3068	0,022	7,2	41	4	0	45						
Laboratory 2																						
Ash,d	w%	K1		0,220	yes	10,9	2,5	10,93	10,91	10,88	0,1371	1,3	47	1	0	48						
C,d	w%	K1		0,005	yes	73,3	2,5	73,31	73,25	73,22	0,9928	1,4	36	2	0	38						
H,d	w%	K1		-0,683	yes	4,6	7	4,49	4,57	4,596	0,209	4,5	31	1	0	32						
Mad,d	w%	K1			yes	3,52		4,025	3,53	3,507	0,301	8,6	48	1	0	49						
q-p,net,d	J/g	K1		0,491	yes	28810	1,3	28900	28820	28790	213,3	0,7	34	8	0	42						
q-V,gr,d	J/g	K1		0,292	yes	29820	1	29860	29820	29810	164,9	0,6	41	8	0	49						
S,d	w%	K1		-0,430	yes	0,31	15	0,3	0,31	0,3068	0,022	7,2	41	4	0	45						
Laboratory 3																						
Ash,d	w%	K1		0,991	yes	10,9	2,5	11,04	10,91	10,88	0,1371	1,3	47	1	0	48						
Mad,d	w%	K1			yes	3,52		3,93	3,53	3,507	0,301	8,6	48	1	0	49						
q-p,net,d	J/g	K1		1,602	yes	28810	1,3	29110	28820	28790	213,3	0,7	34	8	0	42						
q-V,gr,d	J/g	K1		2,354	yes	29820	1	30170	29820	29810	164,9	0,6	41	8	0	49						
S,d	w%	K1		-0,237	yes	0,31	15	0,3045	0,31	0,3068	0,022	7,2	41	4	0	45						
Laboratory 4																						
Ash,d	w%	B1		1,277	yes	7,7	6	7,995	7,7	7,696	0,1756	2,3	27	3	0	30						
	w%	K1		0,514	yes	10,9	2,5	10,97	10,91	10,88	0,1371	1,3	47	1	0	48						
C,d	w%	B1		-0,568	yes	53,4	3	52,95	53,61	53,41	0,7942	1,5	20	2	0	22						
	w%	K1		-2,079	yes	73,3	2,5	71,4	73,25	73,22	0,9928	1,4	36	2	0	38						
EF	t	B1		-0,542	yes	106	4	104,8	106,1	105,9	2,125	2	10	4	0	14						
	t	K1		-0,453	yes	93,8	4	92,95	93,91	93,78	1,6	1,7	22	1	0	23						
H,d	w%	B1		0,414	yes	5,63	7	5,712	5,675	5,629	0,1487	2,6	18	2	0	20						
	w%	K1		0,888	yes	4,6	7	4,743	4,57	4,596	0,209	4,5	31	1	0	32						
Mad,d	w%	B1			yes	2,88		2,235	2,87	2,852	0,4869	17,0	28	2	0	30						
	w%	K1			yes	3,52		2,735	3,53	3,507	0,301	8,6	48	1	0	49						
N,d	w%	B1		-0,280	yes	1,62	15	1,586	1,601	1,621	0,07511	4,6	17	3	0	20						
	w%	K1		-0,819	yes	2,14	15	2,008	2,17	2,137	0,1439	6,7	28	2	0	30						
q-p,net,d	J/g	B1		-0,179	yes	20460	1,8	20430	20460	20490	165,3	0,8	21	6	0	27						
	J/g	K1		-1,853	yes	28810	1,3	28460	28820	28790	213,3	0,7	34	8	0	42						
q-V,gr,d	J/g	B1		-0,112	yes	21680	1,4	21660	21700	21690	161,4	0,7	26	5	0	31						
	J/g	K1		-2,341	yes	29820	1	29470	29820	29810	164,9	0,6	41	8	0	49						
S,d	w%	B1		1,280	yes	0,25	15	0,274	0,248	0,2504	0,0173	6,9	23	2	0	25						
	w%	K1		0,129	yes	0,31	15	0,313	0,31	0,3068	0,022	7,2	41	4	0	45						
Laboratory 5																						
Ash,d	w%	B1		0,649	yes	7,7	6	7,85	7,7	7,696	0,1756	2,3	27	3	0	30						
	w%	K1	<img alt="z-Graphics plot for Ash,d K1 showing a horizontal line at 0, a vertical dashed line at 0,64																			

Analyte	Unit	Sample	z-Graphics							Z-value	Outl test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
			-3	-2	-1	0	+1	+2	+3													
Laboratory 6																						
Ash,d	w%	K1								0,330	yes	10,9	2,5	10,95	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1								-0,224	yes	73,3	2,5	73,09	73,25	73,22	0,9928	1,4	36	2	0	38
H,d	w%	K1								0,932	yes	4,6	7	4,75	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1									yes	3,52		3,795	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1								0,561	yes	2,14	15	2,23	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	K1								-0,494	yes	28810	1,3	28720	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1								-0,010	yes	29820	1	29820	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1								0,645	yes	0,31	15	0,325	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 7																						
Ash,d	w%	B1								0,411	yes	7,7	6	7,795	7,7	7,696	0,1756	2,3	27	3	0	30
Mad,d	w%	B1									yes	2,88		2,57	2,87	2,852	0,4869	17,0	28	2	0	30
q-p,net,d	J/g	B1								-0,274	yes	20460	1,8	20410	20460	20490	165,3	0,8	21	6	0	27
q-V,gr,d	J/g	B1								-0,254	yes	21680	1,4	21640	21700	21690	161,4	0,7	26	5	0	31
Laboratory 8																						
Ash,d	w%	B1								-0,433	yes	7,7	6	7,6	7,7	7,696	0,1756	2,3	27	3	0	30
Mad,d	w%	B1									yes	2,88		2,31	2,87	2,852	0,4869	17,0	28	2	0	30
q-p,net,d	J/g	B1								-0,445	yes	20460	1,8	20380	20460	20490	165,3	0,8	21	6	0	27
q-V,gr,d	J/g	B1								-0,540	yes	21680	1,4	21600	21700	21690	161,4	0,7	26	5	0	31
Laboratory 9																						
Ash,d	w%	B1								-0,390	yes	7,7	6	7,61	7,7	7,696	0,1756	2,3	27	3	0	30
C,d	w%	B1								-0,843	yes	53,4	3	52,72	53,61	53,41	0,7942	1,5	20	2	0	22
H,d	w%	B1								0,553	yes	5,63	7	5,739	5,675	5,629	0,1487	2,6	18	2	0	20
Mad,d	w%	B1									yes	2,88		2,045	2,87	2,852	0,4869	17,0	28	2	0	30
N,d	w%	B1								0,156	yes	1,62	15	1,639	1,601	1,621	0,07511	4,6	17	3	0	20
q-p,net,d	J/g	B1								-0,668	yes	20460	1,8	20340	20460	20490	165,3	0,8	21	6	0	27
q-V,gr,d	J/g	B1								-0,659	yes	21680	1,4	21580	21700	21690	161,4	0,7	26	5	0	31
S,d	w%	B1								-0,080	yes	0,25	15	0,2485	0,248	0,2504	0,0173	6,9	23	2	0	25
Laboratory 10																						
Ash,d	w%	K1								0,514	yes	10,9	2,5	10,97	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1								2,210	yes	73,3	2,5	75,33	73,25	73,22	0,9928	1,4	36	2	0	38
Mad,d	w%	K1									yes	3,52		3,575	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	K1								0,441	yes	28810	1,3	28890	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1								0,466	yes	29820	1	29890	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1								0,430	yes	0,31	15	0,32	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 11																						
Ash,d	w%	K1								0,147	yes	10,9	2,5	10,92	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1								-0,971	yes	73,3	2,5	72,41	73,25	73,22	0,9928	1,4	36	2	0	38
H,d	w%	K1								2,950	yes	4,6	7	5,075	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1									yes	3,52		3,59	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1								-1,340	yes	2,14	15	1,925	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	K1								-0,171	yes	28810	1,3	28780	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1								-0,335	yes	29820	1	29770	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1								0,000	yes	0,31	15	0,31	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 12																						
Ash,d	w%	B1								1,017	yes	7,7	6	7,935	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1								0,917	yes	10,9	2,5	11,03	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1								0,568	yes	53,4	3	53,86	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1								0,742	yes	73,3	2,5	73,98	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	B1								0,283	yes	106	4	106,6	106,1	105,9	2,125	2	10	4	0	14
	t	K1								0,426	yes	93,8	4	94,6	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	B1								0,099	yes	5,63	7	5,649	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1								-0,867	yes	4,6	7	4,46	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1									yes	2,88		3,525	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1									yes	3,52		3,774	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1								-0,963	yes	1,62	15	1,503	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1								-0,779	yes	2,14	15	2,015	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1								-0,065	yes	20460	1,8	20450	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1								0,766	yes	28810	1,3	28950	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1								0,721	yes	29820	1	29930	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1								0,587	yes	0,25	15	0,261	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1								0,150	yes	0,31	15	0,3135	0,31	0,3068	0,022	7,2	41	4	0	45

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

Analyte	Unit	Sample	z-Graphics							Z-value	Outl. test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
Laboratory 13																						
Ash,d	w%	K1	-	-	-	-	-	-	-	-2,349	yes	10,9	2,5	10,58	10,91	10,88	0,1371	1,3	47	1	0	48
Mad,d	w%	K1	-	-	-	-	-	-	-	-	yes	3,52	-	3,575	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	K1	-	-	-	-	-	-	-	0,214	yes	28810	1,3	28850	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1	-	-	-	-	-	-	-	0,268	yes	29820	1	29860	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1	-	-	-	-	-	-	-	-1,699	yes	0,31	15	0,2705	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 14																						
Ash,d	w%	B1	-	-	-	-	-	-	-	-0,173	yes	7,7	6	7,66	7,7	7,696	0,1756	2,3	27	3	0	30
Mad,d	w%	B1	-	-	-	-	-	-	-	-	yes	2,88	-	2,265	2,87	2,852	0,4869	17,0	28	2	0	30
q-p,net,d	J/g	B1	-	-	-	-	-	-	-	0,234	yes	20460	1,8	20500	20460	20490	165,3	0,8	21	6	0	27
q-V,gr,d	J/g	B1	-	-	-	-	-	-	-	0,283	yes	21680	1,4	21720	21700	21690	161,4	0,7	26	5	0	31
Laboratory 15																						
Ash,d	w%	B1	-	-	-	-	-	-	-	-1,342	yes	7,7	6	7,39	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1	-	-	-	-	-	-	-	-0,220	yes	10,9	2,5	10,87	10,91	10,88	0,1371	1,3	47	1	0	48
Mad,d	w%	B1	-	-	-	-	-	-	-	C	2,88	-	2,655	2,87	2,852	0,4869	17,0	28	2	0	30	
	w%	K1	-	-	-	-	-	-	-	yes	3,52	-	3,33	3,53	3,507	0,301	8,6	48	1	0	49	
q-V,gr,d	J/g	B1	-	-	-	-	-	-	-	-5,288	H	21680	1,4	20880	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1	-	-	-	-	-	-	-	-6,190	H	29820	1	28900	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1	-	-	-	-	-	-	-	-1,413	yes	0,25	15	0,2235	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1	-	-	-	-	-	-	-	-1,161	yes	0,31	15	0,283	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 16																						
Ash,d	w%	B1	-	-	-	-	-	-	-	0,368	yes	7,7	6	7,785	7,7	7,696	0,1756	2,3	27	3	0	30
Mad,d	w%	B1	-	-	-	-	-	-	-	-	yes	2,88	-	3,4	2,87	2,852	0,4869	17,0	28	2	0	30
q-p,net,d	J/g	B1	-	-	-	-	-	-	-	-0,432	yes	20460	1,8	20380	20460	20490	165,3	0,8	21	6	0	27
q-V,gr,d	J/g	B1	-	-	-	-	-	-	-	-0,550	yes	21680	1,4	21600	21700	21690	161,4	0,7	26	5	0	31
Laboratory 17																						
Ash,d	w%	K1	-	-	-	-	-	-	-	-2,165	yes	10,9	2,5	10,61	10,91	10,88	0,1371	1,3	47	1	0	48
Mad,d	w%	K1	-	-	-	-	-	-	-	-	yes	3,52	-	3,635	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	K1	-	-	-	-	-	-	-	-1,768	yes	28810	1,3	28480	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1	-	-	-	-	-	-	-	-0,369	yes	29820	1	29770	29820	29810	164,9	0,6	41	8	0	49
Laboratory 18																						
Ash,d	w%	K1	-	-	-	-	-	-	-	0,183	yes	10,9	2,5	10,93	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1	-	-	-	-	-	-	-	-0,126	yes	73,3	2,5	73,19	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	K1	-	-	-	-	-	-	-	0,187	yes	93,8	4	94,15	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	K1	-	-	-	-	-	-	-	-0,472	yes	4,6	7	4,524	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1	-	-	-	-	-	-	-	-	yes	3,52	-	3,463	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1	-	-	-	-	-	-	-	-0,327	yes	2,14	15	2,088	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	K1	-	-	-	-	-	-	-	-0,134	yes	28810	1,3	28790	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1	-	-	-	-	-	-	-	-0,181	yes	29820	1	29790	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1	-	-	-	-	-	-	-	-0,602	yes	0,31	15	0,296	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 19																						
C,d	w%	K1	-	-	-	-	-	-	-	0,164	yes	73,3	2,5	73,45	73,25	73,22	0,9928	1,4	36	2	0	38
H,d	w%	K1	-	-	-	-	-	-	-	1,118	yes	4,6	7	4,78	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1	-	-	-	-	-	-	-	-	yes	3,52	-	3,415	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1	-	-	-	-	-	-	-	1,651	yes	2,14	15	2,405	2,17	2,137	0,1439	6,7	28	2	0	30
S,d	w%	K1	-	-	-	-	-	-	-	0,237	yes	0,31	15	0,3155	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 20																						
Ash,d	w%	B1	-	-	-	-	-	-	-	3,355	H	7,7	6	8,475	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1	-	-	-	-	-	-	-	-0,734	yes	10,9	2,5	10,8	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1	-	-	-	-	-	-	-	6,055	H	53,4	3	58,25	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1	-	-	-	-	-	-	-	7,040	H	73,3	2,5	79,75	73,25	73,22	0,9928	1,4	36	2	0	38
Mad,d	w%	B1	-	-	-	-	-	-	-	-	yes	2,88	-	3,1	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1	-	-	-	-	-	-	-	-	yes	3,52	-	3,52	3,53	3,507	0,301	8,6	48	1	0	49
q-V,gr,d	J/g	B1	-	-	-	-	-	-	-	-2,405	yes	21680	1,4	21320	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1	-	-	-	-	-	-	-	-0,469	yes	29820	1	29750	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1	-	-	-	-	-	-	-	-0,613	yes	0,25	15	0,2385	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1	-	-	-	-	-	-	-	-0,172	yes	0,31	15	0,306	0,31	0,3068	0,022	7,2	41	4	0	45

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

Analyte	Unit	Sample	z-Graphics							Z-value	Outl test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
			-3	-2	-1	0	+1	+2	+3													
Laboratory 21																						
Ash,d	w%	B1								-1,104	yes	7,7	6	7,445	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1								-0,587	yes	10,9	2,5	10,82	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1								0,531	yes	53,4	3	53,83	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1								-0,513	yes	73,3	2,5	72,83	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	B1								0,189	yes	106	4	106,4	106,1	105,9	2,125	2	10	4	0	14
	t	K1								-0,773	yes	93,8	4	92,35	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	B1								-0,140	yes	5,63	7	5,603	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1								-0,158	yes	4,6	7	4,575	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1									yes	2,88		3,165	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1									yes	3,52		4,34	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1								0,593	yes	1,62	15	1,692	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1								0,156	yes	2,14	15	2,165	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1								0,128	yes	20460	1,8	20480	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1								2,096	yes	28810	1,3	29200	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1								0,122	yes	21680	1,4	21700	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1								2,431	yes	29820	1	30180	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1								0,347	yes	0,25	15	0,2565	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1								0,065	yes	0,31	15	0,3115	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 22																						
Ash,d	w%	B1								0,065	yes	7,7	6	7,715	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1								0,294	yes	10,9	2,5	10,94	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1								0,337	yes	53,4	3	53,67	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1								-0,115	yes	73,3	2,5	73,19	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	B1								-0,071	yes	106	4	105,8	106,1	105,9	2,125	2	10	4	0	14
	t	K1								-0,400	yes	93,8	4	93,05	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	B1								-0,926	yes	5,63	7	5,447	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1								-0,534	yes	4,6	7	4,514	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1									yes	2,88		3,2	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1									yes	3,52		3,105	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1								-0,502	yes	1,62	15	1,559	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1								-0,209	yes	2,14	15	2,107	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1								0,307	yes	20460	1,8	20520	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1								1,709	yes	28810	1,3	29130	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1								0,171	yes	21680	1,4	21710	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1								1,858	yes	29820	1	30100	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1								1,573	yes	0,25	15	0,2795	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1								1,161	yes	0,31	15	0,337	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 23																						
q-V,gr,d	J/g	K1								0,154	yes	29820	1	29840	29820	29810	164,9	0,6	41	8	0	49
Laboratory 24																						
Ash,d	w%	B1								-0,238	yes	7,7	6	7,645	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1								0,073	yes	10,9	2,5	10,91	10,91	10,88	0,1371	1,3	47	1	0	48
Mad,d	w%	B1									yes	2,88		2,92	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1									yes	3,52		3,44	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	B1								-0,163	yes	20460	1,8	20430	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1								-1,335	yes	28810	1,3	28560	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1								0,366	yes	21680	1,4	21740	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1								0,335	yes	29820	1	29870	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1								-3,355	H	0,31	15	0,232	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 25																						
Ash,d	w%	K1								-0,367	yes	10,9	2,5	10,85	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1								1,261	yes	73,3	2,5	74,45	73,25	73,22	0,9928	1,4	36	2	0	38
Mad,d	w%	K1									yes	3,52		3,65	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	K1								0,355	yes	28810	1,3	28880	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1								0,892	yes	29820	1	29950	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1								-1,054	yes	0,31	15	0,2855	0,31	0,3068	0,022	7,2	41	4	0	45

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

Analyte	Unit	Sample	z-Graphics							Z-value	Outl test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
Laboratory 26																						
Ash,d	w%	B1			-0,390	yes	7,7	6	7,79	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			0,514	yes	10,9	2,5	10,97	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	B1			-0,125	yes	53,4	3	53,5	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			0,682	yes	73,3	2,5	73,92	73,25	73,22	0,9928	1,4	36	2	0	38					
H,d	w%	B1			0,634	yes	5,63	7	5,755	5,675	5,629	0,1487	2,6	18	2	0	20					
	w%	K1			0,714	yes	4,6	7	4,715	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	B1				yes	2,88		2,785	2,87	2,852	0,4869	17,0	28	2	0	30					
	w%	K1				yes	3,52		3,58	3,53	3,507	0,301	8,6	48	1	0	49					
N,d	w%	B1			-0,288	yes	1,62	15	1,585	1,601	1,621	0,07511	4,6	17	3	0	20					
	w%	K1			-0,187	yes	2,14	15	2,11	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	B1			5,276	H	20460	1,8	21430	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			5,142	H	28810	1,3	29770	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			-5,598	H	21680	1,4	20830	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			-7,495	H	29820	1	28700	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			0,640	yes	0,25	15	0,262	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			0,538	yes	0,31	15	0,3225	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 27																						
Ash,d	w%	B1			-0,693	yes	7,7	6	7,54	7,7	7,696	0,1756	2,3	27	3	0	30					
C,d	w%	B1			0,112	yes	53,4	3	53,49	53,61	53,41	0,7942	1,5	20	2	0	22					
H,d	w%	B1			0,964	yes	5,63	7	5,82	5,675	5,629	0,1487	2,6	18	2	0	20					
Mad,d	w%	B1				yes	2,88		3,39	2,87	2,852	0,4869	17,0	28	2	0	30					
N,d	w%	B1			-7,753	H	1,62	15	0,678	1,601	1,621	0,07511	4,6	17	3	0	20					
q-p,net,d	J/g	B1			-8,1380	H	20460	1,8	5476	20460	20490	165,3	0,8	21	6	0	27					
q-V,gr,d	J/g	B1			0,814	yes	21680	1,4	21800	21700	21690	161,4	0,7	26	5	0	31					
S,d	w%	B1			-0,827	yes	0,25	15	0,2345	0,248	0,2504	0,0173	6,9	23	2	0	25					
Laboratory 28																						
Ash,d	w%	K1			0,147	yes	10,9	2,5	10,92	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	K1			-0,065	C	73,3	2,5	73,24	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	K1			0,067	yes	93,8	4	93,925	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	K1			-0,404	yes	4,6	7	4,535	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	K1				yes	3,52		3,355	3,53	3,507	0,301	8,6	48	1	0	49					
N,d	w%	K1			13,270	H	2,14	15	4,27	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	K1			-1,722	yes	28810	1,3	28490	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	K1			-2,384	yes	29820	1	29460	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	K1			-5,441	H	0,31	15	0,1835	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 29																						
Ash,d	w%	B1			-1,147	yes	7,7	6	7,435	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			0,037	yes	10,9	2,5	10,91	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	B1			-0,911	yes	53,4	3	52,67	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			0,584	yes	73,3	2,5	73,84	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	B1			0,024	yes	106	4	106,1	106,1	105,9	2,125	2	10	4	0	14					
	t	K1			0,426	yes	93,8	4	94,6	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	B1			-1,632	yes	5,63	7	5,309	5,675	5,629	0,1487	2,6	18	2	0	20					
	w%	K1			-1,149	yes	4,6	7	4,415	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	B1				C	2,88		3,475	2,87	2,852	0,4869	17,0	28	2	0	30					
	w%	K1				yes	3,52		3,675	3,53	3,507	0,301	8,6	48	1	0	49					
N,d	w%	B1			-0,152	yes	1,62	15	1,602	1,601	1,621	0,07511	4,6	17	3	0	20					
	w%	K1			0,059	yes	2,14	15	2,149	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	B1			1,773	yes	20460	1,8	20790	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			0,542	yes	28810	1,3	28910	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			1,509	yes	21680	1,4	21910	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			0,178	yes	29820	1	29850	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			-0,560	yes	0,25	15	0,2395	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			0,237	yes	0,31	15	0,3155	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 30																						
Ash,d	w%	B1			-0,519	yes	7,7	6	7,58	7,7	7,696	0,1756	2,3	27	3	0	30					
C,d	w%	B1			-5,718	H	53,4	3	48,82	53,61	53,41	0,7942	1,5	20	2	0	22					
H,d	w%	B1			-28,510	H	5,63	7	0,012	5,675	5,629	0,1487	2,6	18	2	0	20					
Mad,d	w%	B1				yes	2,88		1,43	2,87	2,852	0,4869	17,0	28	2	0	30					
N,d	w%	B1			35,230	H	1,62	15	5,90	1,601	1,621	0,07511	4,6	17	3	0	20					
q-p,net,d	J/g	B1			-6,712	H	20460	1,8	19224	20460	20490	165,3	0,8	21	6	0	27					
q-V,gr,d	J/g	B1			-6,128	H	21680	1,4	20750	21700	21690	161,4	0,7	26	5	0	31					
S,d	w%	B1			60,270	H	0,25	15	1,38	0,248	0,2504	0,0173	6,9	23	2	0	25					

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

Analyte	Unit	Sample	z-Graphics							Z-value	Outl. test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
			-3	-2	-1	0	+1	+2	+3													
Laboratory 31																						
Ash,d	w%	B1								-0,171	yes	7,7	6	7,661	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1								0,451	yes	10,9	2,5	10,96	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1								-1,292	yes	53,4	3	52,37	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1								-1,899	yes	73,3	2,5	71,56	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	B1								-2,288	yes	106	4	101,2	106,1	105,9	2,125	2	10	4	0	14
	t	K1								-1,972	yes	93,8	4	90,1	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	B1								-0,944	yes	5,63	7	5,444	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1								-0,453	C	4,6	7	4,527	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1									yes	2,88		3,26	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1									yes	3,52		3,36	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1								-0,428	yes	1,62	15	1,568	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1								-1,903	yes	2,14	15	1,835	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1								2,599	yes	20460	1,8	20940	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1								3,300	H	28810	1,3	29430	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1								2,982	yes	21680	1,4	22130	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1								4,031	H	29820	1	30420	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1								0,400	yes	0,25	15	0,2575	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1								0,882	yes	0,31	15	0,3305	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 32																						
Ash,d	w%	K1								1,798	yes	10,9	2,5	11,14	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1								0,235	yes	73,3	2,5	73,52	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	K1								-0,091	yes	93,8	4	93,63	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	K1								0,994	yes	4,6	7	4,76	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1									yes	3,52		3,49	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1								-0,903	yes	2,14	15	1,995	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	K1								-0,227	yes	28810	1,3	28770	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1								-0,094	yes	29820	1	29810	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1								-0,430	yes	0,31	15	0,3	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 33																						
Ash,d	w%	K1								0,330	yes	10,9	2,5	10,95	10,91	10,88	0,1371	1,3	47	1	0	48
Mad,d	w%	K1									yes	3,52		3,52	3,53	3,507	0,301	8,6	48	1	0	49
q-V,gr,d	J/g	K1								-0,959	yes	29820	1	29680	29820	29810	164,9	0,6	41	8	0	49
Laboratory 34																						
Ash,d	w%	B1								1,082	yes	7,7	6	7,95	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1								-0,404	yes	10,9	2,5	10,84	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1								-1,261	yes	53,4	3	52,39	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1								-0,928	yes	73,3	2,5	72,45	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	B1								-5,755	H	106	4	93,8	106,1	105,9	2,125	2	10	4	0	14
	t	K1								-0,666	yes	93,8	4	92,55	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	B1								0,855	yes	5,63	7	5,799	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1								0,848	yes	4,6	7	4,736	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1									yes	2,88		2,84	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1									yes	3,52		3,32	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1								-0,259	yes	1,62	15	1,589	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1								-0,586	yes	2,14	15	2,046	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1								0,521	yes	20460	1,8	20560	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1								-0,123	yes	28810	1,3	28790	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1								0,053	yes	21680	1,4	21690	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1								0,080	yes	29820	1	29830	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1								-0,160	yes	0,25	15	0,247	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1								-1,075	yes	0,31	15	0,285	0,31	0,3068	0,022	7,2	41	4	0	45

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

SYKE - Interlaboratory comparison test 5/2010

Analyte	Unit	Sample	z-Graphics							Z-value	Outl test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
Laboratory 35																						
Ash,d	w%	B1			-0,216	yes	7,7	6	7,75	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			-0,367	yes	10,9	2,5	10,85	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	B1			0,437	yes	53,4	3	53,75	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			0,546	yes	73,3	2,5	73,8	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	B1			9,835	H	106	4	126,8	106,1	105,9	2,125	2	10	4	0	14					
	t	K1			8,769	H	93,8	4	110,3	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	B1			-0,736	yes	5,63	7	5,485	5,675	5,629	0,1487	2,6	18	2	0	20					
	w%	K1			-0,621	yes	4,6	7	4,5	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	B1					yes	2,88		2,9	2,87	2,852	0,4869	17,0	28	2	0	30				
	w%	K1					yes	3,52		3,74	3,53	3,507	0,301	8,6	48	1	0	49				
N,d	w%	B1			0,165	yes	1,62	15	1,64	1,601	1,621	0,07511	4,6	17	3	0	20					
	w%	K1			0,218	yes	2,14	15	2,175	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	B1			-26,820	H	20460	1,8	15520	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			-22,950	H	28810	1,3	24510	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			0,007	yes	21680	1,4	21680	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			-2,059	yes	29820	1	29510	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			0,000	yes	0,25	15	0,25	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			0,430	yes	0,31	15	0,32	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 36																						
Ash,d	w%	K1			-0,954	yes	10,9	2,5	10,77	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	K1			0,819	yes	73,3	2,5	74,05	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	K1			0,240	yes	93,8	4	94,25	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	K1			-0,280	yes	4,6	7	4,555	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	K1					yes	3,52		3,535	3,53	3,507	0,301	8,6	48	1	0	49				
N,d	w%	K1			0,685	yes	2,14	15	2,25	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	K1			0,401	yes	28810	1,3	28890	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	K1			0,805	yes	29820	1	29940	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	K1			-0,860	yes	0,31	15	0,29	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 37																						
Ash,d	w%	B1			-0,065	yes	7,7	6	7,685	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			-1,908	yes	10,9	2,5	10,64	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	B1			-0,175	yes	53,4	3	53,26	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			0,224	yes	73,3	2,5	73,5	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	B1			-0,519	yes	106	4	104,9	106,1	105,9	2,125	2	10	4	0	14					
	t	K1			0,000	yes	93,8	4	93,8	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	B1			-0,355	yes	5,63	7	5,56	5,675	5,629	0,1487	2,6	18	2	0	20					
	w%	K1			0,807	yes	4,6	7	4,73	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	B1					yes	2,88		2,875	2,87	2,852	0,4869	17,0	28	2	0	30				
	w%	K1					yes	3,52		3,57	3,53	3,507	0,301	8,6	48	1	0	49				
N,d	w%	B1			0,535	yes	1,62	15	1,685	1,601	1,621	0,07511	4,6	17	3	0	20					
	w%	K1			0,436	yes	2,14	15	2,21	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	B1			0,410	yes	20460	1,8	20540	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			1,129	yes	28810	1,3	29020	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			0,399	yes	21680	1,4	21740	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			1,422	yes	29820	1	30030	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			-0,240	yes	0,25	15	0,2455	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			-0,323	yes	0,31	15	0,3025	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 38																						
Ash,d	w%	B1			-2,814	H	7,7	6	7,05	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			-1,761	yes	10,9	2,5	10,66	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	B1			1,699	yes	53,4	3	54,761	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			-0,045	yes	73,3	2,5	73,259	73,25	73,22	0,9928	1,4	36	2	0	38					
Mad,d	w%	B1					yes	2,88		2,75	2,87	2,852	0,4869	17,0	28	2	0	30				
	w%	K1					yes	3,52		3,57	3,53	3,507	0,301	8,6	48	1	0	49				
q-p,net,d	J/g	B1			-4,670	H	20460	1,8	19600	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			-5,447	H	28810	1,3	27790	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			9,093	H	21680	1,4	23060	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			25,420	H	29820	1	33610	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			-1,149	yes	0,25	15	0,22845	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			-0,877	yes	0,31	15	0,28962	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 39																						
Ash,d	w%	K1			0,367	yes	10,9	2,5	10,95	10,91	10,88	0,1371	1,3	47	1	0	48					
Mad,d	w%	K1					yes	3,52		3,325	3,53	3,507	0,301	8,6	48	1	0	49				
q-V,gr,d	J/g	K1			-0,654	yes	29820	1	29720	29820	29810	164,9	0,6	41	8	0	49					

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

Analyte	Unit	Sample	z-Graphics			Z-value	Outl test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl failed	Missing	Num of labs
Laboratory 40																		
Ash,d	w%	B1		-0.628	yes	7,7	6	7,555	7,7	7,696	0,1756	2,3	27	3	0	30		
	w%	K1		1,284	yes	10,9	2,5	11,07	10,91	10,88	0,1371	1,3	47	1	0	48		
C,d	w%	B1		0,424	yes	53,4	3	53,74	53,61	53,41	0,7942	1,5	20	2	0	22		
	w%	K1		0,191	yes	73,3	2,5	73,47	73,25	73,22	0,9928	1,4	36	2	0	38		
EF	t	B1		0,165	yes	106	4	106,3	106,1	105,9	2,125	2	10	4	0	14		
	t	K1		0,187	yes	93,8	4	94,15	93,91	93,78	1,6	1,7	22	1	0	23		
H,d	w%	B1		-0,771	yes	5,63	7	5,478	5,675	5,629	0,1487	2,6	18	2	0	20		
	w%	K1		-1,981	yes	4,6	7	4,281	4,57	4,596	0,209	4,5	31	1	0	32		
Mad,d	w%	B1			yes	2,88		2,98	2,87	2,852	0,4869	17,0	28	2	0	30		
	w%	K1			yes	3,52		3,8	3,53	3,507	0,301	8,6	48	1	0	49		
N,d	w%	B1		0,514	yes	1,62	15	1,683	1,601	1,621	0,07511	4,6	17	3	0	20		
	w%	K1		0,760	yes	2,14	15	2,262	2,17	2,137	0,1439	6,7	28	2	0	30		
q-p,net,d	J/g	B1		0,024	yes	20460	1,8	20460	20460	20490	165,3	0,8	21	6	0	27		
	J/g	K1		0,507	yes	28810	1,3	28910	28820	28790	213,3	0,7	34	8	0	42		
q-V,gr,d	J/g	B1		-0,184	yes	21680	1,4	21650	21700	21690	161,4	0,7	26	5	0	31		
	J/g	K1		0,003	yes	29820	1	29820	29820	29810	164,9	0,6	41	8	0	49		
S,d	w%	B1		-0,720	yes	0,25	15	0,2365	0,248	0,2504	0,0173	6,9	23	2	0	25		
	w%	K1		-1,140	yes	0,31	15	0,2835	0,31	0,3068	0,022	7,2	41	4	0	45		
Laboratory 41																		
Ash,d	w%	K1		0,037	yes	10,9	2,5	10,9	10,91	10,88	0,1371	1,3	47	1	0	48		
C,d	w%	K1		0,769	yes	73,3	2,5	74	73,25	73,22	0,9928	1,4	36	2	0	38		
EF	t	K1		0,773	yes	93,8	4	95,25	93,91	93,78	1,6	1,7	22	1	0	23		
Mad,d	w%	K1			yes	3,52		3,42	3,53	3,507	0,301	8,6	48	1	0	49		
q-p,net,d	J/g	K1		-0,184	yes	28810	1,3	28780	28820	28790	213,3	0,7	34	8	0	42		
q-V,gr,d	J/g	K1		-0,107	yes	29820	1	29800	29820	29810	164,9	0,6	41	8	0	49		
S,d	w%	K1		-0,538	yes	0,31	15	0,2975	0,31	0,3068	0,022	7,2	41	4	0	45		
Laboratory 42																		
Ash,d	w%	K1		-11,050	H	10,9	2,5	9,395	10,91	10,88	0,1371	1,3	47	1	0	48		
Mad,d	w%	K1			yes	3,52		3,145	3,53	3,507	0,301	8,6	48	1	0	49		
q-p,net,d	J/g	K1		-24,820	H	28810	1,3	24160	28820	28790	213,3	0,7	34	8	0	42		
q-V,gr,d	J/g	K1		-29,190	H	29820	1	25470	29820	29810	164,9	0,6	41	8	0	49		
S,d	w%	K1		-2,151	yes	0,31	15	0,26	0,31	0,3068	0,022	7,2	41	4	0	45		
Laboratory 43																		
Ash,d	w%	B1		-0,649	yes	7,7	6	7,55	7,7	7,696	0,1756	2,3	27	3	0	30		
	w%	K1		0,294	yes	10,9	2,5	10,94	10,91	10,88	0,1371	1,3	47	1	0	48		
C,d	w%	B1		1,373	yes	53,4	3	54,5	53,61	53,41	0,7942	1,5	20	2	0	22		
	w%	K1		-1,626	yes	73,3	2,5	71,81	73,25	73,22	0,9928	1,4	36	2	0	38		
EF	t	B1		-4,021	H	106	4	97,47	106,1	105,9	2,125	2	10	4	0	14		
	t	K1		-1,364	yes	93,8	4	91,24	93,91	93,78	1,6	1,7	22	1	0	23		
H,d	w%	B1		0,837	yes	5,63	7	5,795	5,675	5,629	0,1487	2,6	18	2	0	20		
	w%	K1		-0,031	yes	4,6	7	4,595	4,57	4,596	0,209	4,5	31	1	0	32		
Mad,d	w%	B1			yes	2,88		2,86	2,87	2,852	0,4869	17,0	28	2	0	30		
	w%	K1			yes	3,52		3,695	3,53	3,507	0,301	8,6	48	1	0	49		
N,d	w%	B1		-0,905	yes	1,62	15	1,51	1,601	1,621	0,07511	4,6	17	3	0	20		
	w%	K1		-1,464	yes	2,14	15	1,905	2,17	2,137	0,1439	6,7	28	2	0	30		
q-p,net,d	J/g	B1		0,144	yes	20460	1,8	20490	20460	20490	165,3	0,8	21	6	0	27		
	J/g	K1		0,141	yes	28810	1,3	28840	28820	28790	213,3	0,7	34	8	0	42		
q-V,gr,d	J/g	B1		0,366	yes	21680	1,4	21740	21700	21690	161,4	0,7	26	5	0	31		
	J/g	K1		0,047	yes	29820	1	29830	29820	29810	164,9	0,6	41	8	0	49		
S,d	w%	B1		-0,213	yes	0,25	15	0,246	0,248	0,2504	0,0173	6,9	23	2	0	25		
	w%	K1		-0,645	yes	0,31	15	0,295	0,31	0,3068	0,022	7,2	41	4	0	45		
Laboratory 44																		
Ash,d	w%	K1		-0,110	yes	10,9	2,5	10,89	10,91	10,88	0,1371	1,3	47	1	0	48		
Mad,d	w%	K1			yes	3,52		3,615	3,53	3,507	0,301	8,6	48	1	0	49		
q-V,gr,d	J/g	K1		-0,218	yes	29820	1	29790	29820	29810	164,9	0,6	41	8	0	49		
Laboratory 45																		
Ash,d	w%	K1		0,367	yes	10,9	2,5	10,95	10,91	10,88	0,1371	1,3	47	1	0	48		
C,d	w%	K1		-1,506	yes	73,3	2,5	71,92	73,25	73,22	0,9928	1,4	36	2	0	38		
EF	t	K1		-0,053	yes	93,8	4	93,7	93,91	93,78	1,6	1,7	22	1	0	23		
H,d	w%	K1		-3,975	yes	4,6	7	3,96	4,57	4,596	0,209	4,5	31	1	0	32		
Mad,d	w%	K1			yes	3,52		3,705	3,53	3,507	0,301	8,6	48	1	0	49		
q-p,net,d	J/g	K1		-20,350	H	28810	1,3	25000	28820	28790	213,3	0,7	34	8	0	42		
q-V,gr,d	J/g	K1		-0,372	yes	29820	1	29760	29820	29810	164,9	0,6	41	8	0	49		
S,d	w%	K1		-0,086	yes	0,31	15	0,308	0,31	0,3068	0,022	7,2	41	4	0	45		

Analyte	Unit	Sample	z-Graphics							Z-value	Outl. test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs	
			-3	-2	-1	0	+1	+2	+3														
Laboratory 46																							
Ash,d	w%	K1								0,220	yes	10,9	2,5	10,93	10,91	10,88	0,1371	1,3	47	1	0	48	
C,d	w%	K1								0,611	yes	73,3	2,5	73,86	73,25	73,22	0,9928	1,4	36	2	0	38	
EF	t	K1								0,906	yes	93,8	4	95,5	93,91	93,78	1,6	1,7	22	1	0	23	
H,d	w%	K1								-0,615	yes	4,6	7	4,501	4,57	4,596	0,209	4,5	31	1	0	32	
Mad,d	w%	K1								H	3,52			98,22	3,53	3,507	0,301	8,6	48	1	0	49	
N,d	w%	K1								1,153	yes	2,14	15	2,325	2,17	2,137	0,1439	6,7	28	2	0	30	
q-p,net,d	J/g	K1								-2,480	yes	28810	1,3	28350	28820	28790	213,3	0,7	34	8	0	42	
q-V,gr,d	J/g	K1								-3,484	H	29820	1	29300	29820	29810	164,9	0,6	41	8	0	49	
S,d	w%	K1								-0,602	yes	0,31	15	0,296	0,31	0,3068	0,022	7,2	41	4	0	45	
Laboratory 47																							
Ash,d	w%	K1								-2,936	yes	10,9	2,5	10,5	10,91	10,88	0,1371	1,3	47	1	0	48	
Mad,d	w%	K1									yes	3,52			3,75	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	K1								-5,393	H	28810	1,3	27800	28820	28790	213,3	0,7	34	8	0	42	
q-V,gr,d	J/g	K1								-6,841	H	29820	1	28800	29820	29810	164,9	0,6	41	8	0	49	
S,d	w%	K1								-0,215	yes	0,31	15	0,305	0,31	0,3068	0,022	7,2	41	4	0	45	
Laboratory 48																							
Ash,d	w%	K1								1,211	yes	10,9	2,5	11,06	10,91	10,88	0,1371	1,3	47	1	0	48	
C,d	w%	K1								0,884	yes	73,3	2,5	74,11	73,25	73,22	0,9928	1,4	36	2	0	38	
H,d	w%	K1								2,022	yes	4,6	7	4,925	4,57	4,596	0,209	4,5	31	1	0	32	
Mad,d	w%	K1									yes	3,52			3,5	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1								0,822	yes	2,14	15	2,272	2,17	2,137	0,1439	6,7	28	2	0	30	
q-p,net,d	J/g	K1								0,539	yes	28810	1,3	28910	28820	28790	213,3	0,7	34	8	0	42	
q-V,gr,d	J/g	K1								0,956	yes	29820	1	29960	29820	29810	164,9	0,6	41	8	0	49	
S,d	w%	K1								1,914	yes	0,31	15	0,3545	0,31	0,3068	0,022	7,2	41	4	0	45	
Laboratory 49																							
Ash,d	w%	B1								-0,022	yes	7,7	6	7,695	7,7	7,696	0,1756	2,3	27	3	0	30	
Mad,d	w%	B1									yes	2,88			2,145	2,87	2,852	0,4869	17,0	28	2	0	30
q-p,net,d	J/g	B1								-1,784	yes	20460	1,8	20130	20460	20490	165,3	0,8	21	6	0	27	
q-V,gr,d	J/g	B1								-1,918	yes	21680	1,4	21390	21700	21690	161,4	0,7	26	5	0	31	
S,d	w%	B1								-1,200	yes	0,25	15	0,2275	0,248	0,2504	0,0173	6,9	23	2	0	25	
Laboratory 50																							
Ash,d	w%	K1								1,101	yes	10,9	2,5	11,05	10,91	10,88	0,1371	1,3	47	1	0	48	
C,d	w%	K1								-1,266	yes	73,3	2,5	72,14	73,25	73,22	0,9928	1,4	36	2	0	38	
EF	t	K1								-0,533	yes	93,8	4	92,8	93,91	93,78	1,6	1,7	22	1	0	23	
H,d	w%	K1								1,708	yes	4,6	7	4,875	4,57	4,596	0,209	4,5	31	1	0	32	
Mad,d	w%	K1									yes	3,52			3,245	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1								0,249	yes	2,14	15	2,18	2,17	2,137	0,1439	6,7	28	2	0	30	
q-p,net,d	J/g	K1								-0,027	yes	28810	1,3	28810	28820	28790	213,3	0,7	34	8	0	42	
q-V,gr,d	J/g	K1								0,329	yes	29820	1	29870	29820	29810	164,9	0,6	41	8	0	49	
S,d	w%	K1								-8,172	H	0,31	15	0,12	0,31	0,3068	0,022	7,2	41	4	0	45	
Laboratory 51																							
q-V,gr,d	J/g	B1								0,689	yes	21680	1,4	21780	21700	21690	161,4	0,7	26	5	0	31	
Laboratory 52																							
Ash,d	w%	K1								-1,688	yes	10,9	2,5	10,67	10,91	10,88	0,1371	1,3	47	1	0	48	
C,d	w%	K1								-0,169	yes	73,3	2,5	73,15	73,25	73,22	0,9928	1,4	36	2	0	38	
EF	t	K1								0,320	yes	93,8	4	94,4	93,91	93,78	1,6	1,7	22	1	0	23	
H,d	w%	K1								0,491	yes	4,6	7	4,679	4,57	4,596	0,209	4,5	31	1	0	32	
Mad,d	w%	K1									yes	3,52			3,36	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1								0,495	yes	2,14	15	2,22	2,17	2,137	0,1439	6,7	28	2	0	30	
q-p,net,d	J/g	K1								-0,675	yes	28810	1,3	28680	28820	28790	213,3	0,7	34	8	0	42	
q-V,gr,d	J/g	K1								-0,902	yes	29820	1	29690	29820	29810	164,9	0,6	41	8	0	49	
S,d	w%	K1								1,312	yes	0,31	15	0,3405	0,31	0,3068	0,022	7,2	41	4	0	45	
Laboratory 53																							
Ash,d	w%	K1								-0,991	yes	10,9	2,5	10,77	10,91	10,88	0,1371	1,3	47	1	0	48	
C,d	w%	K1								-2,794	yes	73,3	2,5	70,74	73,25	73,22	0,9928	1,4	36	2	0	38	
Mad,d	w%	K1									yes	3,52			3,465	3,53	3,507	0,301	8,6	48	1	0	49
q-p,net,d	J/g	K1								0,323	yes	28810	1,3	28870	28820	28790	213,3	0,7	34	8	0	42	
q-V,gr,d	J/g	K1								-0,091	yes	29820	1	29810	29820	29810	164,9	0,6	41	8	0	49	
S,d	w%	K1								-2,237	yes	0,31	15	0,258	0,31	0,3068	0,022	7,2	41	4	0	45	

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

Analyte	Unit	Sample	z-Graphics			Z-value	Outl test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
			-3	-2	-1	0	+1	+2	+3									
Laboratory 54																		
Ash,d	w%	K1				0,000	yes	10,9	2,5	10,9	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1				-0,398	yes	73,3	2,5	72,94	73,25	73,22	0,9928	1,4	36	2	0	38
H,d	w%	K1				0,839	yes	4,6	7	4,735	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1					yes	3,52		3,8	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1				0,343	yes	2,14	15	2,195	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	K1				-0,481	yes	28810	1,3	28720	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1				-0,013	yes	29820	1	29820	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1				0,323	yes	0,31	15	0,3175	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 55																		
Ash,d	w%	B1				0,909	yes	7,7	6	7,91	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1				0,073	yes	10,9	2,5	10,91	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1				-1,167	yes	53,4	3	52,47	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1				0,610	yes	73,3	2,5	73,86	73,25	73,22	0,9928	1,4	36	2	0	38
H,d	w%	B1				0,507	yes	5,63	7	5,73	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1				-0,425	yes	4,6	7	4,532	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1					yes	2,88		2,495	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1					yes	3,52		3,125	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1				-0,288	yes	1,62	15	1,585	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1				-1,464	yes	2,14	15	1,905	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1				-0,369	yes	20460	1,8	20390	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1				-2,787	yes	28810	1,3	28290	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1				-1,031	yes	21680	1,4	21520	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1				-0,825	yes	29820	1	29700	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1				-0,267	yes	0,25	15	0,245	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1				-0,258	yes	0,31	15	0,304	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 56																		
Ash,d	w%	K1				0,330	yes	10,9	2,5	10,95	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	K1				-0,060	yes	73,3	2,5	73,25	73,25	73,22	0,9928	1,4	36	2	0	38
H,d	w%	K1				-0,230	yes	4,6	7	4,563	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	K1					yes	3,52		3,45	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	K1				0,489	yes	2,14	15	2,219	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	K1				0,513	yes	28810	1,3	28910	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	K1				0,382	yes	29820	1	29880	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	K1				-0,086	yes	0,31	15	0,308	0,31	0,3068	0,022	7,2	41	4	0	45
Laboratory 57																		
Ash,d	w%	B1				0,325	yes	7,7	6	7,775	7,7	7,696	0,1756	2,3	27	3	0	30
	w%	K1				-0,697	yes	10,9	2,5	10,8	10,91	10,88	0,1371	1,3	47	1	0	48
C,d	w%	B1				1,380	yes	53,4	3	54,5	53,61	53,41	0,7942	1,5	20	2	0	22
	w%	K1				2,079	yes	73,3	2,5	75,2	73,25	73,22	0,9928	1,4	36	2	0	38
EF	t	B1				1,863	yes	106	4	110	106,1	105,9	2,125	2	10	4	0	14
	t	K1				2,186	yes	93,8	4	97,9	93,91	93,78	1,6	1,7	22	1	0	23
H,d	w%	B1				3,636	H	5,63	7	6,347	5,675	5,629	0,1487	2,6	18	2	0	20
	w%	K1				-1,307	yes	4,6	7	4,389	4,57	4,596	0,209	4,5	31	1	0	32
Mad,d	w%	B1					yes	2,88		2,72	2,87	2,852	0,4869	17,0	28	2	0	30
	w%	K1					yes	3,52		3,405	3,53	3,507	0,301	8,6	48	1	0	49
N,d	w%	B1				1,440	yes	1,62	15	1,795	1,601	1,621	0,07511	4,6	17	3	0	20
	w%	K1				1,480	yes	2,14	15	2,377	2,17	2,137	0,1439	6,7	28	2	0	30
q-p,net,d	J/g	B1				7,342	H	20460	1,8	21810	20460	20490	165,3	0,8	21	6	0	27
	J/g	K1				16,460	H	28810	1,3	31890	28820	28790	213,3	0,7	34	8	0	42
q-V,gr,d	J/g	B1				11,150	H	21680	1,4	23370	21700	21690	161,4	0,7	26	5	0	31
	J/g	K1				21,960	H	29820	1	33090	29820	29810	164,9	0,6	41	8	0	49
S,d	w%	B1				1,893	yes	0,25	15	0,2855	0,248	0,2504	0,0173	6,9	23	2	0	25
	w%	K1				4,710	H	0,31	15	0,4195	0,31	0,3068	0,022	7,2	41	4	0	45

Outlier test failed: C - Cochran, G1 - Grubbs(1-outlier algorithm), G2 - Grubbs(2-outliers algorithm), H - Hampel, M - manual

SYKE - Interlaboratory comparison test 5/2010

Analyte	Unit	Sample	z-Graphics							Z-value	Outl. test OK	Assigned value	2* Targ SD%	Lab's result	Md.	Mean	SD	SD%	Passed	Outl. failed	Missing	Num of labs
Laboratory 58																						
Ash,d	w%	B1			-3,506	H	7,7	6	6,89	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			-1,835		yes	10,9	2,5	10,65	10,91	10,88	0,1371	1,3	47	1	0	48				
C,d	w%	B1			-2,647	yes	53,4	3	51,28	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			-0,524	yes	73,3	2,5	72,82	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	B1			-9,528	H	106	4	85,8	106,1	105,9	2,125	2	10	4	0	14					
	t	K1			1,066	yes	93,8	4	95,8	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	B1			0,376	yes	5,63	7	5,704	5,675	5,629	0,1487	2,6	18	2	0	20					
	w%	K1			0,534	yes	4,6	7	4,686	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	B1				yes	2,88		3,52	2,87	2,852	0,4869	17,0	28	2	0	30					
	w%	K1				yes	3,52		2,44	3,53	3,507	0,301	8,6	48	1	0	49					
N,d	w%	B1			-6,469	H	1,62	15	0,834	1,601	1,621	0,07511	4,6	17	3	0	20					
	w%	K1			-10,430	H	2,14	15	0,466	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	B1			1,320	yes	20460	1,8	20703	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			-2,237	yes	28810	1,3	28391	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			1,535	yes	21680	1,4	21913	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			-2,918	yes	29820	1	29385	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			8,640	H	0,25	15	0,412	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			2,925	yes	0,31	15	0,378	0,31	0,3068	0,022	7,2	41	4	0	45					
Laboratory 59																						
q-V,gr,d	J/g	B1			0,007		yes	21680	1,4	21680	21700	21690	161,4	0,7	26	5	0	31				
Laboratory 60																						
Ash,d	w%	B1			-0,541	yes	7,7	6	7,575	7,7	7,696	0,1756	2,3	27	3	0	30					
	w%	K1			0,661	yes	10,9	2,5	10,99	10,91	10,88	0,1371	1,3	47	1	0	48					
C,d	w%	B1			0,549	yes	53,4	3	53,84	53,61	53,41	0,7942	1,5	20	2	0	22					
	w%	K1			-0,497	yes	73,3	2,5	72,84	73,25	73,22	0,9928	1,4	36	2	0	38					
EF	t	B1			0,189	yes	106	4	106,4	106,1	105,9	2,125	2	10	4	0	14					
	t	K1			-0,133	yes	93,8	4	93,55	93,91	93,78	1,6	1,7	22	1	0	23					
H,d	w%	B1			0,036	yes	5,63	7	5,637	5,675	5,629	0,1487	2,6	18	2	0	20					
	w%	K1			-0,345	yes	4,6	7	4,544	4,57	4,596	0,209	4,5	31	1	0	32					
Mad,d	w%	B1				yes	2,88		2,785	2,87	2,852	0,4869	17,0	28	2	0	30					
	w%	K1				yes	3,52		3,68	3,53	3,507	0,301	8,6	48	1	0	49					
N,d	w%	B1			0,177	yes	1,62	15	1,642	1,601	1,621	0,07511	4,6	17	3	0	20					
	w%	K1			-0,436	yes	2,14	15	2,07	2,17	2,137	0,1439	6,7	28	2	0	30					
q-p,net,d	J/g	B1			0,103	yes	20460	1,8	20480	20460	20490	165,3	0,8	21	6	0	27					
	J/g	K1			0,214	yes	28810	1,3	28850	28820	28790	213,3	0,7	34	8	0	42					
q-V,gr,d	J/g	B1			0,132	yes	21680	1,4	21700	21700	21690	161,4	0,7	26	5	0	31					
	J/g	K1			0,013	yes	29820	1	29820	29820	29810	164,9	0,6	41	8	0	49					
S,d	w%	B1			0,800	yes	0,25	15	0,265	0,248	0,2504	0,0173	6,9	23	2	0	25					
	w%	K1			0,645	yes	0,31	15	0,325	0,31	0,3068	0,022	7,2	41	4	0	45					

LIITE 9. SUMMARY OF TEH z SCORES

APPENDIX 9.

Analyte	Sample\Lab	1	2	3	4	5	6	7	8	9	10	11	12	13	14	15	16	17	18	19	20	21	22	23	
Ash,d	B1	S	.	S	S	.	S	S	S	.	S	.	S	S	S	.	.	U	S	S	.	.			
	K1	S	S	S	S	S	S	.	.	S	S	S	q	.	S	.	q	S	.	S	S	S	.		
C,d	B1	S	.	S	.	.	.	S	.	S	.	S	U	S	S	.	.		
	K1	S	S	.	q	.	S	.	.	Q	S	S	S	S	U	S	S	.		
EF	B1	.	.	.	S	S	S	.	.	S	S	
	K1	.	.	.	S	S	S	.	.	S	S	.		
H,d	B1	S	.	S	.	S	.	S	.	S	.	S	S	S	.	S	S	.	
	K1	S	S	.	S	.	S	.	S	.	Q	S	S	S	.	S	S	.	S		
Mad,d	B1	
	K1	
N,d	B1	S	.	S	.	S	.	S	.	S	.	S	.	S	.	S	.	S	S	.	S	S	.		
	K1	S	.	S	.	S	.	S	.	S	S	S	.	S	.	S	.	S	S	.	S	S	.		
q-p,net,d	B1	S	.	S	.	S	.	S	S	S	.	S	.	S	.	S	.	S	.	S	S	.	S		
	K1	S	S	S	S	S	.	S	.	S	S	S	.	S	.	S	.	S	S	.	Q	S	.		
q-V,gr,d	B1	S	.	S	S	.	S	S	S	.	S	.	S	.	S	u	S	.	.	q	S	S	.		
	K1	S	S	Q	q	S	S	.	.	S	S	S	S	.	S	u	.	S	S	.	S	Q	S	S	
S,d	B1	S	.	S	S	.	S	S	.	S	.	S	.	S	.	S	.	S	.	S	S	S	.		
	K1	S	S	S	S	S	S	.	.	S	S	S	S	.	S	.	S	S	S	S	S	S	.		
% Accredited		100	100	75	88	100	100	100	100	100	80	86	100	75	100	67	100	67	100	100	50	88	100	100	
	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	
Analyte	Sample\Lab	24	25	26	27	28	29	30	31	32	33	34	35	36	37	38	39	40	41	42	43	44	45	46	
Ash,d	B1	S	.	S	S	.	S	S	S	.	S	S	.	S	q	.	S	.	.	S	.	.	.		
	K1	S	S	S	.	S	S	.	S	S	S	S	S	S	S	S	S	S	u	S	S	S	S		
C,d	B1	.	.	S	S	.	S	u	S	.	S	S	.	S	S	.	S	.	S	.	S	.	.		
	K1	.	S	S	.	S	S	.	S	S	.	S	S	S	S	.	S	S	.	S	.	S	S		
EF	B1	S	.	q	.	.	u	U	.	S	.	.	S	.	.	u	.	.	.		
	K1	S	S	.	S	S	.	S	U	S	S	.	S	S	.	S	.	S	S		
H,d	B1	.	.	S	S	.	S	u	S	.	S	S	.	S	S	.	S	.	S	.	S	.	.		
	K1	.	.	S	.	S	S	.	S	S	.	S	S	S	S	.	S	.	S	.	u	S	.		
Mad,d	B1	
	K1	
N,d	B1	.	.	S	u	.	S	U	S	.	S	S	.	S	S	.	S	.	S	.	S	.	.		
	K1	.	.	S	.	U	S	.	S	S	.	S	S	S	S	.	S	.	S	.	S	.	S		
q-p,net,d	B1	S	.	U	u	.	S	u	Q	.	S	u	.	S	u	.	S	.	S	.	S	.	.		
	K1	S	S	U	.	S	S	.	U	S	.	S	u	S	S	u	.	S	S	u	S	.	u		
q-V,gr,d	B1	S	.	u	S	.	S	u	Q	.	S	S	.	S	U	.	S	.	S	.	S	.	.		
	K1	S	S	u	.	q	S	.	U	S	S	q	S	S	U	S	S	S	u	S	S	S	u		
S,d	B1	.	.	S	S	.	S	U	S	.	S	S	.	S	S	.	S	S	.	S	.	S	.		
	K1	u	S	S	.	u	S	.	S	S	.	S	S	S	S	.	S	S	q	S	.	S	S		
% Accredited		86	100	71	71	62	100	14	69	100	100	94	69	100	100	50	100	100	100	0	94	100	71	75	
	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	
Analyte	Sample\Lab	47	48	49	50	51	52	53	54	55	56	57	58	59	60	%									
Ash,d	B1	.	.	S	S	.	S	u	.	S	90										
	K1	q	S	.	S	.	S	S	S	S	S	S	S	.	S	92									
C,d	B1	S	.	S	q	.	S	86										
	K1	.	S	.	S	.	S	q	S	S	S	Q	S	.	S	87									
EF	B1	S	u	.	S	64											
	K1	.	.	.	S	.	S	.	.	Q	S	.	S	91											
H,d	B1	S	.	U	S	.	S	90										
	K1	.	Q	.	S	.	S	.	S	S	S	S	S	.	S	91									
Mad,d	B1		
	K1		
N,d	B1	S	.	S	u	.	S	85										
	K1	.	S	.	S	.	S	.	S	S	S	S	u	.	S	93									
q-p,net,d	B1	.	.	S	.	.	.	S	.	U	S	.	S	74											
	K1	u	S	.	S	.	S	S	S	q	S	U	q	.	S	71									
q-V,gr,d	B1	.	.	S	.	S	.	S	.	S	U	S	S	S	77										
	K1	u	S	.	S	.	S	S	S	S	U	q	.	S	71										
S,d	B1	.	.	S	.	.	.	S	.	S	U	.	S	92											
	K1	S	S	.	u	.	S	q	S	S	S	U	Q	.	S	84									
% Accredited		25	86	100	88	100	100	60	100	93	100	50	44	100	100										
	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	yes	

Analyte	Sample\Lab	47	48	49	50	51	52	53	54	55	56	57	58	59	60	%
---------	------------	----	----	----	----	----	----	----	----	----	----	----	----	----	----	---

S - satisfactory ($-2 \leq z \leq 2$), Q - questionable ($2 < z < 3$), q - questionable ($-3 < z < -2$),
U - unsatisfactory ($z \geq 3$), u - unsatisfactory ($z \leq -3$)

%* - percentage of satisfactory results

Totally satisfactory, % In all: 84 In accredited: 87 In non-accredited: 70

MEASUREMENT UNCERTAINTIES REPORTED BY THE LABORATORIES

For evaluation of the measurement uncertainty the participants have used the procedures as follows:




In the figures the procedures have been presented using the same code number.




1. using the variation of the results in X chart (for the artificial samples)
2. using the variation of the results in X chart and the variation of the replicates (r%- or R- chart for real samples)
3. using the data obtained in method validation and IQC, see e.g. NORDTEST TR 537¹⁾
4. using the data obtained in the analysis of CRM (besides IQC data). see e.g. NORDTEST TR 537¹⁾
5. using the IQC data and the results obtained in proficiency tests. see e.g. NORDTEST TR 537¹⁾
6. using the "modelling approach" (GUM Guide or EURACHEM Guide Quantifying Uncertainty in Analytical Measurements²⁾
7. other procedure
8. no uncertainty estimation

IQC= internal quality control

¹⁾ <http://www.nordicinnovation.net>


²⁾ <http://www.eurachem.org>

LIITE 10.
APPENDIX 10.Analytti (Analyte) **Ash,d** Näyte (Sample) B1Analytti (Analyte) **Ash,d** Näyte (Sample) K1Analytti (Analyte) **C,d** Näyte (Sample) B1




Analyytti (Analyte) H,d

Näyte (Sample) B1




Analyytti (Analyte) H,d

Näyte (Sample) K1




Analyytti (Analyte) Mad,d

Näyte (Sample) B1




Analyytti (Analyte) **M_{ad,d}**


Näyte (Sample) K1

Analyytti (Analyte) **N,d**


Näyte (Sample) B1

Analyytti (Analyte) **N,d**


Näyte (Sample) K1




Analyytti (Analyte) q-p,net,d Näyte (Sample) B1




Analyytti (Analyte) q-p,net,d Näyte (Sample) K1




Analyytti (Analyte) q-V,gr,d Näyte (Sample) B1






z AND zeta SCORES OF LABORATORIES

<u>Analyte</u>	<u>Ash,d</u>	<u>Sample</u>	<u>B1</u>
Assigned value	Assigned 2*UC	Targ 2SD, %	
7,7	0,09	6	


z&zeta-Score diagram

Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	7,935	8	1,02	0,73
2	22	7,715	10	0,06	0,04
3	20	8,475	15	3,35	1,22
4	4	7,995	3	1,28	2,30
5	1	7,85	10	0,65	0,38
6	60	7,575	1	-0,54	-2,13
7	49	7,695	8	-0,02	-0,02
8	43	7,55	1,6	-0,65	-1,99
9	37	7,685	8	-0,06	-0,05
10	34	7,95	1,78	1,08	2,98
11	35	7,75	0,39	0,22	1,05
12	57	7,775	0,637	0,32	1,46
13	31	7,6605	0,23848	-0,17	-0,86
14	24	7,645	2,9	-0,24	-0,46

<u>Analyte</u>	<u>Ash,d</u>	<u>Sample</u>	<u>K1</u>
Assigned value	Assigned 2*UC	Targ 2SD, %	
10,9	0,04	2,5	

z&zeta-Score diagram

<u>Analyte</u>	<u>C,d</u>	<u>Sample</u>	<u>B1</u>
Assigned value	Assigned 2*UC	Targ 2SD, %	
53,4	0,48	3	


z&zeta-Score diagram

Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	53,855	2	0,57	0,77
2	22	53,67	2	0,34	0,46
3	20	58,25	30	6,05	0,55
4	4	52,945	1	-0,57	-1,27
5	1	54,3	5	1,12	0,65
6	60	53,84	0,2	0,55	1,79
7	43	54,5	1,1	1,37	2,86
8	37	53,26	5	-0,17	-0,10
9	34	52,39	0,62	-1,26	-3,49
10	28	53,5	1,09	0,12	0,26
11	35	53,75	0,53	0,44	1,25
12	57	54,505	0,246	1,38	4,43
13	31	52,365	0,2364	-1,29	-4,18

Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	11,025	6	0,92	0,38
2	22	10,94	3	0,29	0,24
3	18	10,925	5	0,18	0,09
4	20	10,8	15	-0,73	-0,12
5	4	10,97	2	0,51	0,63
6	1	10,83	7	-0,51	-0,18
7	48	11,065	2,4	1,21	1,23
8	10	10,97	1,82	0,51	0,69
9	2	10,93	1,64	0,22	0,33
10	60	10,99	1	0,66	1,54
11	43	10,94	11,2	0,29	0,07
12	54	10,9	0,3	0,00	0,00
13	37	10,64	8	-1,91	-0,61
14	34	10,845	0,31	-0,40	-2,11
15	35	10,85	0,27	-0,37	-2,02
16	36	10,77	2,3	-0,95	-1,04
17	42	9,395	1,6	-11,05	-19,35
18	53	10,765	0,51	-0,99	-3,97
19	28	10,92	0,13	0,15	0,94
20	50	11,05	0,04	1,10	7,45
21	33	10,945	0,25	0,33	1,86
22	25	10,85	0,53	-0,37	-1,43
23	57	10,805	0,72	-0,70	-2,17
24	39	10,95	0,45	0,37	1,58
25	45	10,95	1,2	0,37	0,73
26	31	10,9615	0,23848	0,45	2,57
27	52	10,67	0,26	-1,69	-9,45
28	24	10,91	1,6	0,07	0,11
29	41	10,905	3,62	0,04	0,03
30	6	10,945	0,3	0,33	1,74

Analyte C,d		Sample K1
Assigned value	Assigned 2*UC	Targ 2SD, %
73,3	0,38	2,5


z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	73,98	2	0,74	0,89
2	22	73,195	3	-0,11	-0,09
3	18	73,185	1	-0,13	-0,28
4	20	79,75	30	7,04	0,54
5	19	73,45	2	0,16	0,20
6	4	71,395	2	-2,08	-2,58
7	1	73,1	5	-0,22	-0,11
8	48	74,11	0,5	0,88	3,05
9	10	75,325	0,66	2,21	6,47
10	2	73,305	2,22	0,01	0,01
11	60	72,845	0,2	-0,50	-2,24
12	29	73,835	1,5	0,58	0,91
13	43	71,81	0,8	-1,63	-4,33
14	54	72,935	0,6	-0,40	-1,26
15	37	73,505	5	0,22	0,11
16	34	72,45	0,23	-0,93	-4,10
17	26	73,925	1,09	0,68	1,40
18	35	73,8	1,85	0,55	0,71
19	36	74,05	0,2	0,82	3,68
20	53	70,74	2,81	-2,79	-2,53
21	28	73,24	1,8	-0,07	-0,09
22	50	72,14	0,04	-1,27	-6,09
23	25	74,455	0,63	1,26	3,83
24	57	75,205	0,028	2,08	10,01
25	45	71,92	0,5	-1,51	-5,28
26	31	71,56	0,2364	-1,90	-8,37
27	52	73,145	0,32	-0,17	-0,69
28	46	73,86	5	0,61	0,30
29	41	74,005	0,89	0,77	1,85
30	6	73,095	0,6	-0,22	-0,71

Analyte EF		Sample B1
Assigned value	Assigned 2*UC	Targ 2SD, %
106	0,91	4


z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	106,6	10	0,28	0,11
2	22	105,85	6	-0,07	-0,05
3	4	104,85	3	-0,54	-0,70
4	43	97,475	2,5	-4,02	-6,55
5	35	126,85	3,16	9,83	10,14
6	57	109,95	0,579	1,86	7,11

Analyte EF		Sample K1
Assigned value	Assigned 2*UC	Targ 2SD, %
93,8	0,63	4

z&zeta-Score diagram




Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	94,6	2	0,43	0,80
2	22	93,05	6	-0,40	-0,27
3	18	94,15	2	0,19	0,35
4	4	92,95	5	-0,45	-0,36
5	43	91,242	1,9	-1,36	-2,77
6	35	110,25	2,76	8,77	10,59
7	36	94,25	0,63	0,24	1,04
8	57	97,9	0	2,19	13,02
9	41	95,25	4,9	0,77	0,62

Analyte H,d

Assigned value	Assigned 2*UC	Targ 2SD, %
5,63	0,1	7

z&zeta-Score diagram




Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	5,6495	16	0,10	0,04
2	22	5,4475	10	-0,93	-0,66
3	4	5,7115	3	0,41	0,82
4	1	5,7	10	0,36	0,24
5	60	5,637	0,6	0,04	0,13
6	43	5,795	2,7	0,84	1,78
7	37	5,56	5	-0,36	-0,47
8	34	5,7985	1,52	0,86	2,53
9	26	5,755	0,36	0,63	2,45
10	35	5,485	0,11	-0,74	-2,89
11	57	6,3465	0,546	3,64	13,54
12	31	5,444	1,9212	-0,94	-2,57

Analyte H,d

Assigned value	Assigned 2*UC	Targ 2SD, %
4,6	0,09	7

z&zeta-Score diagram




Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	4,4605	5	-0,87	-1,16
2	22	4,514	15	-0,53	-0,25
3	18	4,524	5	-0,47	-0,62
4	19	4,78	5	1,12	1,41
5	4	4,743	3	0,89	1,70
6	1	4,4	10	-1,24	-0,89
7	48	4,9255	0,36	2,02	7,10
8	2	4,49	6,44	-0,68	-0,73
9	60	4,5445	0,6	-0,34	-1,18
10	29	4,415	7	-1,15	-1,15
11	43	4,595	3,4	-0,03	-0,06
12	54	4,735	0,2	0,84	2,98
13	37	4,73	5	0,81	1,03
14	34	4,7365	2,52	0,85	1,83
15	26	4,715	0,36	0,71	2,51
16	35	4,5	0,29	-0,62	-2,20
17	36	4,555	0,5	-0,28	-0,97
18	28	4,535	0,2	-0,40	-1,44
19	50	4,875	0,04	1,71	6,11
20	57	4,3895	0,693	-1,31	-4,43
21	45	3,96	0,3	-3,98	-14,10
22	31	4,527	1,9212	-0,45	-1,17
23	52	4,679	0,06	0,49	1,75
24	46	4,501	11	-0,61	-0,39
25	6	4,75	0,2	0,93	3,31

Analyte N,d

Assigned value	Assigned 2*UC	Targ 2SD, %
1,62	0,05	15

z&zeta-Score diagram




Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	1,503	8	-0,96	-1,80
2	22	1,559	8	-0,50	-0,91
3	4	1,586	12	-0,28	-0,35
4	1	1,7	10	0,66	0,90
5	60	1,6415	2,8	0,18	0,63
6	43	1,51	6,4	-0,91	-2,02
7	37	1,685	15	0,53	0,50
8	26	1,585	0,22	-0,29	-1,40
9	35	1,64	0,24	0,16	0,80
10	57	1,795	0,394	1,44	6,93
11	31	1,568	5,8478	-0,43	-1,00

Analyte N,d

Assigned value	Assigned 2*UC	Sample K1
2,14	0,08	Targ 2SD, % 15

z&zeta-Score diagram




Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	2,015	3	-0,78	-2,49
2	22	2,1065	8	-0,21	-0,36
3	18	2,0875	8	-0,33	-0,57
4	19	2,405	5	1,65	3,67
5	4	2,0085	13	-0,82	-0,96
6	1	2,2	10	0,37	0,51
7	48	2,272	0,36	0,82	3,28
8	60	2,07	2,8	-0,44	-1,42
9	29	2,1495	6	0,06	0,13
10	43	1,905	5,1	-1,46	-3,73
11	54	2,195	0,3	0,34	1,37
12	37	2,21	15	0,44	0,41
13	26	2,11	0,22	-0,19	-0,75
14	35	2,175	0,22	0,22	0,87
15	36	2,25	0,3	0,69	2,74
16	28	4,27	0,1	13,27	53,17
17	50	2,18	0,04	0,25	1,00
18	57	2,3775	0,149	1,48	5,93
19	31	1,8345	5,8478	-1,90	-4,57
20	52	2,2195	0,04	0,50	1,99
21	46	2,325	19	1,15	0,82
22	6	2,23	0,3	0,56	2,24

Analyte q-p.net,d

Assigned value	Assigned 2*UC	Sample B1
20460	55	Targ 2SD, % 1,8

z&zeta-Score diagram




Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	20448	2	-0,07	-0,06
2	22	20516,5	2	0,31	0,27
3	16	20380,5	4	-0,43	-0,19
4	4	20427	1	-0,18	-0,31
5	1	20613	5	0,83	0,30
6	7	20409,5	0,18	-0,27	-1,53
7	49	20131,5	5	-1,78	-0,65
8	43	20486,5	0,6	0,14	0,39
9	37	20535,5	5	0,41	0,15
10	26	21431,5	0,4	5,28	19,08
11	35	15521	1,5	-26,82	-41,29
12	57	21812	0,156	7,34	41,81
13	31	20938,5	0,40233	2,60	9,51
14	24	20430	0,4	-0,16	-0,61

Analyte q-p.net,d

Assigned value	Assigned 2*UC	Sample K1
28810	82	Targ 2SD, % 1,3


z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	28953,5	1,3	0,77	0,75
2	22	29130	2	1,71	1,09
3	18	28785	1,5	-0,13	-0,11
4	4	28463	1	-1,85	-2,34
5	1	28774,5	5	-0,19	-0,05
6	48	28911	0,82	0,54	0,81
7	10	28892,5	2,2	0,44	0,26
8	29	28911,5	2,5	0,54	0,28
9	43	28836,5	0,4	0,14	0,37
10	37	29021,5	5	1,13	0,29
11	26	29773	0,4	5,14	13,32
12	35	24513	1,5	-22,95	-22,81
13	36	28885	1,7	0,40	0,30
14	42	24162	1	-24,82	-36,43
15	53	28870,5	1,5	0,32	0,27
16	28	28487,5	3,4	-1,72	-0,66
17	50	28805	20	-0,03	0,00
18	25	28876,5	0,4	0,36	0,94
19	57	31893	0,146	16,46	65,39
20	45	24999	0,7	-20,35	-39,44
21	31	29428	0,40233	3,30	8,58
22	52	28683,5	0,62	-0,68	-1,29
23	24	28560	0,5	-1,34	-3,04
24	46	28345,5	2	-2,48	-1,62

Analyte q-V,gr,d		Sample B1
Assigned value	Assigned 2*UC	Targ 2SD, %
21680	51	1,4


z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	22	21706	2	0,17	0,12
2	20	21315	30	-2,41	-0,11
3	16	21596,5	4	-0,55	-0,19
4	4	21663	1	-0,11	-0,15
5	1	21841,5	5	1,06	0,30
6	7	21641,5	0,18	-0,25	-1,20
7	60	21700	0,2	0,13	0,60
8	49	21389	5	-1,92	-0,54
9	43	21735,5	0,6	0,37	0,79
10	37	21740,5	5	0,40	0,11
11	34	21688	0,02	0,05	0,31
12	26	20830,5	0,4	-5,60	-17,39
13	35	21681	1,4	0,01	0,01
14	51	21784,5	0,09	0,69	3,83
15	57	23371,5	0,076	11,15	62,64
16	31	22132,5	0,40233	2,98	8,82
17	24	21735,5	0,5	0,37	0,92
18	59	21681	1,5	0,01	0,01

Analyte q-V,gr,d		Sample K1
Assigned value	Assigned 2*UC	Targ 2SD, %
29820	52	1


z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	29927,5	1,3	0,72	0,55
2	22	30097	2	1,86	0,92
3	18	29793	1,5	-0,18	-0,12
4	20	29750	30	-0,47	-0,02
5	4	29471	1	-2,34	-2,33
6	1	29724	5	-0,64	-0,13
7	48	29962,5	0,82	0,96	1,13
8	10	29889,5	2,2	0,47	0,21
9	2	29863,5	0,23	0,29	1,01
10	60	29822	0,2	0,01	0,05
11	29	29846,5	2,5	0,18	0,07
12	43	29827	0,4	0,05	0,11
13	54	29818	0,8	-0,01	-0,02
14	37	30032	5	1,42	0,28
15	34	29832	0,007	0,08	0,46
16	26	28702,5	0,4	-7,49	-17,73
17	35	29513	1,95	-2,06	-1,06
18	36	29940	1,7	0,80	0,47
19	42	25468,5	1	-29,19	-33,48
20	53	29806,5	1,5	-0,09	-0,06
21	28	29464,5	3,6	-2,38	-0,67
22	50	29869	20	0,33	0,02
23	33	29677	0,1	-0,96	-4,78
24	25	29953	0,4	0,89	2,04
25	57	33094	0,171	21,96	85,20
26	39	29722,5	0,14	-0,65	-2,93
27	45	29764,5	0,7	-0,37	-0,52
28	31	30421	0,40233	4,03	9,04
29	23	29843	0,013	0,15	0,88
30	52	29685,5	0,6	-0,90	-1,45
31	24	29870	1,2	0,34	0,28
32	46	29300,5	2	-3,48	-1,77
33	41	29804	1,46	-0,11	-0,07
34	6	29818,5	0,8	-0,01	-0,01

<u>Analyte S,d</u>	<u>Sample B1</u>	
Assigned value	Assigned 2*UC	Targ 2SD, %
0,25	0,008	15


z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	0,261	10	0,59	0,81
2	22	0,2795	10	1,57	2,03
3	21	0,2565	18	0,35	0,28
4	20	0,2385	35	-0,61	-0,27
5	4	0,274	6	1,28	2,63
6	1	0,26	5	0,53	1,31
7	60	0,265	1	0,80	3,56
8	49	0,2275	16	-1,20	-1,21
9	43	0,246	7,3	-0,21	-0,41
10	37	0,2455	0,08	-0,24	-1,12
11	34	0,247	0,78	-0,16	-0,73
12	26	0,262	0,41	0,64	2,97
13	35	0,25	0,04	0,00	0,00
14	57	0,2855	3,715	1,89	5,34
15	31	0,2575	4,83762	0,40	1,01

<u>Analyte S,d</u>	<u>Sample K1</u>	
Assigned value	Assigned 2*UC	Targ 2SD, %
0,31	0,007	15

z&zeta-Score diagram



Nr	Lab	Mean	Reported UC,%	z	zeta
1	12	0,3135	5	0,15	0,41
2	22	0,337	10	1,16	1,57
3	21	0,3115	18	0,06	0,05
4	18	0,296	8	-0,60	-1,13
5	20	0,306	35	-0,17	-0,07
6	19	0,3155	6	0,24	0,55
7	4	0,313	4	0,13	0,42
8	1	0,32	5	0,43	1,15
9	48	0,3545	2,8	1,91	7,33
10	10	0,32	4,38	0,43	1,28
11	2	0,3	6,47	-0,43	-0,97
12	60	0,325	1	0,65	3,89
13	29	0,3155	16	0,24	0,22
14	43	0,295	7,3	-0,65	-1,32
15	54	0,3175	0,1	0,32	2,14
16	37	0,3025	0,08	-0,32	-2,14
17	34	0,285	1,29	-1,08	-6,32
18	26	0,3225	0,41	0,54	3,51
19	35	0,32	0,009	0,43	2,86
20	36	0,29	0,09	-0,86	-5,71
21	42	0,26	11,5	-2,15	-3,26
22	53	0,258	3,27	-2,24	-9,49
23	28	0,1835	0,12	-5,44	-36,12
24	50	0,12	0,04	-8,17	-54,28
25	25	0,2855	0,08	-1,05	-7,00
26	57	0,4195	3,203	4,71	14,45
27	45	0,308	0,06	-0,09	-0,57
28	31	0,3305	4,83762	0,88	2,35
29	52	0,3405	0,03	1,31	8,71
30	24	0,232	7,5	-3,35	-8,32
31	46	0,296	4	-0,60	-2,04
32	41	0,2975	16,205	-0,54	-0,51
33	6	0,325	0,1	0,65	4,28

Documentation page

Publisher	Finnish Environment Institute (SYKE)	Date February 2011
Author(s)	Mirja Leivuori, Minna Rantanen, Kaija Korhonen-Ylönen ja Markku Ilmakunnas	
Title of publication	Proficiency test 5/2010 Gross and net calorific value in fuels	
Parts of publication/ other project publications	The publication is available on the internet: www.ymparisto.fi/julkaisut	
Abstract	<p>The Finnish Environment Institute (Proftest SYKE) carried out the proficiency test for measurement the gross and the net calorific value, the content of ash, carbon, nitrogen, hydrogen, moisture and sulfur in fuels in September 2010. One peat sample and one coal sample were delivered to the laboratories for the analysis of each measurement. In total, 58 laboratories participated in the proficiency test. The test performed in collaboration with the working group (WG) of the European co-operation for Accreditation (EA) for Interlaboratory Comparisons (EA WG ILC in Testing).</p> <p>The robust means of the reported results by the participants were used as the assigned values for measurements. The evaluation of performance was based on the z score which was calculated using the standard deviation for proficiency assessment at 95 % confidence level. The total standard deviation for performance assessment was mainly set on the basis of the reproducibility requirements presented the standard methods. The evaluation of performance was not done for the measurement of moisture and for the emission factor of the coal sample. In total, 84 % of the participating laboratories reported the satisfactory results. About 78 % of the participants used accredited methods and 87 % of their results were satisfactory. In measurement of the gross calorific value from the peat sample 77 % of the results were satisfactory and respectively in measurement of the coal sample 71 % from the results were satisfactory. In measurement of the net calorific value from the peat sample 74 % of the results were satisfactory and respectively in measurement of the coal sample 71 % from the results were satisfactory.</p>	
Keywords	Proficiency test, interlaboratory comparison, Proftest, SYKE, coal, peat, measurement of calorific value, emission factor, measurement of ash, moisture, carbon, sulphur, nitrogen and hydrogen, environmental laboratories	
Publication series and number	Suomen ympäristökeskuksen raportteja 4/2011	
Theme of publication		
Project name and number, if any		
Financier/ commissioner		
Project organization		
	ISSN 1796- 1726 (online)	ISBN 978-952-11-3846-1 (PDF)
	No. of pages 70	Language English
	Restrictions Public	Price
For sale at/ distributor	Finnish Environment Institute, Customer service E-mail: neuvonta.syke@ymparisto.fi Tel. + 358 20 610 183 Telefax. + 358 20 490 2190	
Financier of publication	Finnish Environment Institute, P.O.Box 140, FIN-00251 Helsinki, Finland	
Printing place and year	Helsinki 2011	
Other information		

Kuvailulehti

Julkaisija	Suomen ympäristökeskus (SYKE)		Julkaisuaika Helmikuu 2011
Tekijä(t)	Mirja Leivuori, Minna Rantanen, Kaija Korhonen-Ylönen ja Markku Ilmakunnas		
Julkaisun nimi	Laboratorioiden välinen pätevyyskoe 5/2010 Kalorimetrisen ja tehollinen lämpöarvo polttoaineista		
Julkaisun osat/ muut saman projektin tuottamat julkaisut	Julkaisu on saatavana myös internetistä: http://www.ymparisto.fi/julkaisut		
Tiivistelmä	<p>Suomen ympäristökeskus (Proftest SYKE) järjesti syyskuussa 2011 pätevyyskokeen kalorimetrisen ja tehollisen lämpöarvon sekä tuhkan, vedyn, typen, rikin ja kosteuden määrittämiseksi turpeesta ja kivihiilestä. Pätevyyskoe järjestettiin yhteistyössä Euroopan akkreditointielinten yhteistyöjärjestön (EA) testauksen vertailumittaustyöryhmän kanssa (EA WG ILC in testing).</p> <p>Pätevyyskokeeseen osallistui yhteensä 58 laboratoriota. Laboratorioiden pätevyyden arvionti tehtiin z-aron avulla ja sen laskemisessa käytetyn kokonaishajonnan tavoitearvot olivat välillä 1-15 %. Mittaussuureen vertailuarvona käytettiin osallistujien ilmoittamien tulosten robustia keskiarvoa. Tavoitearvon epävarmuus oli lämpöarvon määritysissä alhaisempi kuin 0.3 % ja muiden testisuureiden osalta korkeintaan 5.6 %. Tulosten arvointia ei tehty kosteuspitoisuuden määrityselle, koska osallistujien välinen hajonta oli suuri. Arvointia ei myöskään tehty päästökertoimelle hiilen osalta, koska kaikki laboratoriot eivät olleet laskeneet arvoa tulokosteutta kohti.</p> <p>Koko tulosaineistossa hyväksyttäviä tuloksia oli 84 %, kun vertailuarvosta sallittiin 1-15 %.n poikkeama. Noin 78 % osallistujista käytti akkreditoituja määritysmenetelmiä ja näistä tuloksista oli hyväksyttäviä 87 %. Kalorimetrisen lämpöarvon tuloksista oli tydyttäviä 77 % (turve) ja 71 % (kivihiili). Tehollisen lämpöarvon tuloksille vastaavat tydyttävien tulosten osuudet olivat 74 % (turve) ja 71 % (kivihiili).</p>		
Asiasanat	Pätevyyskoe, vertailumittaus, Proftest, SYKE, kalorimetrisen lämpöarvo, tehollinen lämpöarvo, päästökerroin, tuhkan, kosteuden, hiilen, rikin, typen ja vedyn määritys, turve, hiili, ympäristölaboratoriot		
Julkaisusarjan nimi ja numero	Suomen ympäristökeskuksen moniste 4/2011		
Julkaisun teema			
Projektihankkeen nimi ja projektinumero			
Rahoittaja/ toimeksiantaja			
Projektiryhmään kuuluvat organisaatiot			
	ISSN 1796- 1726(verkkoj.)	ISBN 978-952-11-3846-1 (PDF)	
	Sivuja 70	Kieli englanti	
	Luottamuksellisuus Julkinen	Hinta	
Julkaisun myynti/ jakaja	Suomen ympäristökeskus, Asiakaspalvelu E-mail: neuvonta.syke@ymparisto.fi Puh. 020 610 183 Telefax. 020 490 2190		
Julkaisun kustantaja	Suomen ympäristökeskus, PL 140, 00251 Helsinki		
Painopaikka ja -aika	Helsinki 2011		
Muut tiedot			

Presentationsblad

Utgivare	Finlands Miljöcentral (SYKE)	Datum Februari 2011
Författare	Mirja Leivuori, Minna Rantanen, Kaija Korhonen-Ylönen och Markku Ilmakunnas	
Publikationens titel	Provningsjämförelse 5/2010 Kalorimetriskt och effektivt värmevärde i bränsle	
Publikationens delar/ andra publikationer inom samma projekt	Publikationen finns tillgänglig på internet www.ymparisto.fi/julkaisut	
Sammandrag	<p>Under September 2010 genomförde Finlands Miljöcentral (Proftest SYKE) en provningsjämförelse, som omfattade bestämningen av kalorimetriskt och effektivt värmevärde, svavel, väte, kol, nitrogen, aska och fuktighet i torv och stenkol. Tillsammans 58 laboratorier deltog i jämförelsen. Testet utfördes in samarbete med arbetsgruppen för arbetsgruppen (WG) för Europeiska samarbetet för ackreditering (EA) för Provningsjämförelser (EA WG ILC i Testing).</p> <p>Som referensvärde av analytens koncentration användes mest det robust medelvärdet av deltagarnas resultat. Resultaten värderades med hjälp av z-värden. I jämförelsen var 84 % av alla resultaten acceptabelt, när total deviation på 1–15 % från referensvärdet accepterades.</p>	
Nyckelord	provningsjämförelse, Proftest, SYKE, kalorimetriskt och effektivt värmevärde, utsläppskoefficient, svavel, väte, kol, nitrogen, aska och fuktighet, stenkol, torv, miljölaboratorier	
Publikationsserie och nummer	Suomen ympäristökeskuksen raportteja 4/2011	
Publikationens tema		
Projektets namn och nummer		
Finansiär/ uppdagsgivare		
Organisationer i projektgruppen		
	ISSN 1796- 1726 (online)	ISBN 978-952-11-3846-1 (PDF)
	Sidantal 70	Språk Engelska
	Offentlighet Offentlig	Pris
Beställningar/ distribution	Finlands miljöcentral, informationsstjänsten neuvonta.syke@ymparisto.fi Tfn 020 610 183 Fax 09 5490 2190	
Förläggare	Finlands Miljöcentral, PB 140, 00251 Helsingfors	
Tryckeri/ tryckningsort och –år	Helsingfors 2011	
Övriga uppgifter		



ISBN 978-952-11-3846-1 (PDF)

ISSN 1796-1726 (online)