

Spectrophotometric determination of trace Arsenic (III) ion based on complex formation with galloyanine

Abstract

In this study, a simple, selective and sensitive method, for spectrophotometric determination of As(III) with galloyanine as the sensitive reagent was developed. The wavelength of an analytical measurement, for the determination of As (III), using galloyanine was at 630nm with an optimum response at pH 2. The RSD for the reproducibility of 100 ppm As(III) was 2.3%. The LOD was 0.04 ppm with linear dynamic range in As(III) concentration of 0.2 - 1.5 ppm. The developed method has been validated against Atomic Absorption Spectrophotometry (AAS). The interference study of several metal ions was carried out and it revealed that that Mn (II) ion was interfered the most.

Keyword: Galloyanine; Arsenic determination; Metal toxicity