# Development of Microwave Moisture Sensors and Low Intensity Ultrasonic Sensors for Agriculture, Industrial and Scientific Applications\*

Kaida bin Khalid, Sidek Abd Aziz, Wan Daud Wan Yusof and Jumiah Hassan

Faculty of Science and Environmental Studies Universiti Putra Malaysia 43400 UPM, Serdang, Selangor Malaysia

E-mail of Corresponding Author: kaida@fsas.upm.edu.my

Key words: moisture sensor, microwave, ultrasonic, hevea latex, and oil palm fruit.

### Introduction

Low intensity microwave has been extensively used for biological sensing, in particular, monitoring of moisture content (MC) which is an important parameter in determining quality of agriculture produce, degree of ripeness and completeness of industrial processes. In the 6th Malaysian Plan we have successfully developed Microwave Reflection Type Latexometer<sup>1</sup> and moisture meter for hevea latex and various lossy liquids and microstrip sensor for oil palm fruit<sup>2,3</sup>. Further studies on the development of microwave sensors for oil palm fruit, application of latexometer for industrial application, development of ultrasonic sensor for characterization food emulsion and microwave and ultrasonic characterization of rubber wood are main focus in the present work.

# **Materials and Methods**

New design, analysis and performance of a conductor-backed coplanar waveguide (CBCPW) moisture sensor4 and other microwave sensors<sup>5</sup> are conducted in this project. The CBCPW sensor consists of three components the coupling system representing the transition between the coaxial line and the CBCPW, the two-layer structures of the CBCPW and the sensing area with its semi-infinite three-layer CPCPW structure. A relationship between the scattering parameter  $S_{21}$  and the moisture content of the oil palm fruit has been developed. The reflection and transmission phenomena in the sensor structure can be represented by a signal flow graph and can be simplified by using non-touching loop rules. The calculation of S21 is based on the quasitransverse electromagnetic mode approximation. Oil palm fruits in various stages of maturity were selected from the Tenera variety. The fresh mesocarp was pressed firmly into the sensing

area to ensure a good contact. The magnitude of attenuation was measured using automatic network analyzer at 2.2 GHz and temperature at 26°C.

The second part of the project deals with the development of MRT-latexometer for latex-based industries<sup>7</sup>. Basically the sensor consists of microwave transmitter, non-lossy protective cover, receiver, detector, signal conditioning and display unit. The sensor is dipped in the latex and the reflected wave detected by the meter can be calibrated with the total solid content (TSC) of the latex. This sensor covers a wide range of TSC, ranging from 0% to 60%. For a trial run this sensor is installed at the Rubber Glove Factory RRIM Sg. Buloh, Kuala Lumpur.

### **Results and Discussion**

The performance of the small-gap and large gap CBCPW sensors on oil palm mesocarp is presented. Reasonably close agreement of about 0.5 dB is found between the experimental and theoretical data. Both attenuation and sensitivity of the sensors are affected significantly by the thickness of the protective layer. With a sensing area length of about 1.6cm (i.e. small gap sensor) the sensitivity for s/h = 0.13(ratio of protective cover with respect to the thickness of substrate) is about 0.14 dB/MC while for s/h=0.22 the sensitivity drop to 0.11 dB/MC. This phenomenon is due to the decreasing interaction between the e.m field and the sample as the thickness of the protective layer increases. It is also found that the large-gap sensor has a better sensitivity than the small-gap sensor. For example at s/h=0.18, the sensitivity of the large-gap sensor is about 0.17dB/MC while the small -gap sensor is 0.11 dB/MC. This effect is possibly due to higher field density in the large-gap sensor.

MRT-Latexometer is found to be suitable for the process control in the rubber glove factory especially to prepare the latex with correct TSC for dipping process.

### Conclusions

Close agreement has been found between theoretical analysis and experimental results of attenuation and sensitivity of CBCPW sensor with respect to the thickness of the protective layer and the size of the gap between the conducting strip and the upper ground plane. Based on this analysis, optimum geometrical sensor parameters can be obtained in order to give highest sensitivity at a particular dynamic range. Furthermore, the strong relationship between attenuation and moisture content for this kind of sensor offers the possibility of the development of a compact microwave instrument for assessing the quality of oil palm fruits that could reach the factory.

MRT-Latexometer developed for latexbased factory is a future instrument for a rapid, accurate and non-destructive method for determining the TSC or moisture content of latex.

# Benefits from the study

CBCPW moisture sensor has a great potential in determining the quality of oil palm fruit, which reach the processing plant. Latex-based industries can use MRT-Latexometer for process control and determination of TSC of hevea latex.

## Literature cited in the text

Khalid, KB. 1994. A Portable Microwave Moisture Meter for Lossy Liquids. Proc. of the 1994 Asia- Pacific Microwave Conf. "Makuhari Messe, Tokyo, 477-481.

Khalid, KB. and Abbas ZB, 1996.

Development of Microstrip Sensor for Oil
Palm Fruits. in Microwave Aquametry
:Electromagnetic Wave Interaction with

- Water-Containing Materials edited by A. Kraszewski, TAB-IEEE Press Book Series, 239-248.
- Khalid K.and Abbas. Z. 1992. A
  Microstrip Sensor for Determination of
  Harvesting Time for Oil Palm Fruits
  (Tenera: Elacis Guineensis). J.
  Microwave Power and EM Energy, 27
  (1): 3-10.
- Khalid KB. and Hua.TL. 1998.

  Development of Conductor-Backed

  Coplanar Waveguide Moisture Sensor for

  Oil Palm Fruit. Meas. Sci. Technol,

  9(8): 1191-1195.
- Khalid KB and Teh Seng Guan. 1998.

  Determination of Moisture Content in various Liquids using Microstrip Disk Antenna. International Science Conference, 7-9 May 1998, UPM, Malaysia.

## Project Publications in Refereed Journals

- Hassan, J., Khalid, K.B. and Daud, W.M. 1997. Microwave Dielectric Properties of Hevea Latex at Temperatures -30°C to 50°C. Pertanika J. Sci & Technol. 5(2): 1-11.
- Khalid, K.B., Zakaria, Z. and Daud, W.M. 1996. Variation of Dielectric Properties of

- Oil Palm Mesocarp with Moisture Content and Fruit Maturity at Microwave Frequencies. *Elaeis*. 8(2): 83-91.
- Kabir, M.F., Sidek, H.A.A., Daud, W.M. and Khalid, K.B. 1997. Effect of Moisture Content and Grain Angle on the Ultrasonic Properties of Rubber Wood. Holzforschu.ng 51: 263-267
- Kabir, M.F., Sidek, H.A.A., Daud, M.A. and Khalid, K. 1997. Detection of knot and split of rubber wood by non-destructive ultrasonic method. *Journal of Tropical Forest Products*. 3(1): 88-96.
- Kabir, M.F., Khalid, K., Daud, W.M. and Sidek, H.A.A. 1997. Dielectric properties of rubber wood at microwave frequencies measured with an open ended coaxial line. Wood and Fiber Science. 29(4).
- Kabir, M.F., Daud, W.M., Khalid, K. and Sidek, H.A.A. 1997. Viscoelastic properties of rubber wood at various moisture content, grain direction, frequency and temperature. Holz als Rohund Werkstoff (European Journal of Wood and Wood Industries). 55(5).
- Talib, Z.A., Daud, W.M., Khalid, K.B.,
  Mat- Yunus, W.M. and Mansor, A.R.
  1998. Dielectric Study of Electrolyzed
  Rubber. IEEE Trans. On Dielectrics and Electrical Insulation. 5(2): 195-198.

- Kabir, M.F., Daud, W.M., Khalid, K. and
  Sidek, H.A.A. 1998. Dielectric and
  Ultrasonic Properties of Rubber
  Wood:Effect of Moisture Content, Grain
  Direction and Frequency. J. Solid St. Sci.
  Technol. Lett. 5(1): 43-52.
- Sidek, H.A.A., Chow, S.P., Shaari, A.H. and Senin, H.B. 1996. Ultrasonic Studies of Palm Oils and Other Vegetables Oils.
  ELAIS (International Journal of Palm Oils). 8(1): 37-44.

# Project Publications in Conference Proceedings

Khalid, K.B. and Teh Seng Guan. 1998.
Determination of Moisture Content in various Liquids using Microstrip Disk Antenna. International Science Conference: 7-9 May 1998, UPM, Malaysia.

## **Graduate Research**

- Jumiah bte Hassan. 1999. Dielectric Physics [Ph.D]. Universiti Putra Malaysia.
- Mohamad F Kabir. 1998. Dielectric Physics [Ph.D]. Universiti Putra Malaysia.
- Zanariah bt Zakaria. 1998. Dielectric Physics [M. Sc.]. Universiti Putra Malaysia.

<sup>\*</sup>An update of the abstract published in OPM Research Report 1993.