

## Implementation of the Z-scan technique using NI cRIO 9074 system

R. Gutiérrez Fuentes<sup>β,\*</sup>, M.A. Camacho López

*Laboratorio de Fotomedicina, Biofotónica y Espectroscopia Láser de Pulsos Ultracortos, Facultad de Medicina, Universidad Autónoma del Estado de México, Toluca, Edo. Mex., 50120, México.*

<sup>β</sup> *Universidad Politécnica del Valle de Toluca  
Santiaguillo Tlalcilcali, Edo. Mex., 50904, México.*

J.L. Jiménez Pérez

*Unidad Profesional Interdisciplinaria en Ingeniería y Tecnologías Avanzadas del I.P.N.  
Gustavo A. Madero, Cd. Mx., 07340, México.*

Z.N. Correa Pacheco

*Centro de Desarrollo de Productos Bióticos del Instituto Politécnico Nacional  
Yautepec, Morelos, 62730, México.*

(Received: November 6th, 2015; Accepted: November 22nd, 2016)

This paper presents the implementation of Z-scan technique using NI cRIO 9074 system to characterize different types of nanofluids and films. This technique was implemented in the *Universidad Autónoma del Estado de México*. For this experiment a chassis NI cRIO9074 of National Instruments, a linear translation stage NRT150E and a stepper motor controller BSC203 were used, both of Thorlabs. Three steps were followed for this implementation first, the connection between NI cRIO9074 and BSC203 controller. Second, the program on LabVIEW was developed and finally, all optic part of the z-scan technique was implemented. This implementation can be used to characterize relatively thin (< 5mm) optical materials. The system testing was done with gold nanoparticles. The results showed the nonlinear optical properties of these samples.

*Keywords:* instrumentation; nonlinear refractive index; Z-scan; nanoparticles.

### Introduction

In recent years, the studies for new nonlinear optical materials have been increasing interest, due to the numerous applications, such as characterization of biological nano-markers, chemical and biological sensors, energy transport, thermal materials, and photothermal therapy, in general in nanoscience and nanotechnology. For this, the Z-scan technique has been widely used, due to versatility of applications in optical study of nonlinear index of refraction,  $n_2$ , and the nonlinear absorption coefficient,  $\beta$  [1, 2]. On the other hand, the cRIO9074 system is a reconfigurable embedded control and data acquisition system. This controller offers powerful stand-alone embedded execution for deterministic LabVIEW Real-Time applications. Additionally, cRIO9074 is programmed with LabVIEW. It is a graphical programming language that uses icons instead of lines of text to create applications. In contrast to text-based programming languages, where instructions determine program execution, LabVIEW uses dataflow programming, where the flow of data determines execution. In LabVIEW, the programmer builds a user interface with a set of tools and objects. The front panel is user interface, and the block diagram contains code of functions to control the front panel objects, using graphical representations. Then LabVIEW can be used in a variety of embedded control and monitoring applications, like to Z-scan [3, 4]. To characterize different types of nanofluids the Z-scan technique was implemented using NI cRIO 9074 system. This technique was implemented in the *Facultad de Medicina* of the *Universidad*

*Autónoma del Estado de México*. In this work, the steps for the implementation of the Z-scan technique are described. Finally, the system was tested with Au nanofluids. The results showed the nonlinear optical properties of these samples.


### NI cRIO9074 system and BSC203 driver communication

The BSC203 is a driver of THORLABS for linear translation stages and it has different protocols of communication like to USB, RS232, Ethernet and TTL signals (trigger in/out). TTL communication between cRIO9074 and BSC203 was used. For this communication, the cRIO9074 used NI9401 module was also used. This module is an 8-channel bidirectional digital, which received and send TTL signals to BSC203. In the Figure 1(a) this connection is shown, where the DB25 and DB15 connectors correspond to NI9401 and BSC203 respectively.

### Construction of program on LabVIEW

The Z-scan program contains “while loop” and “case structure” instructions. The program has five parts. The first is the introduction of parameters for Z-scan such as scanning length, scan resolution and variation tolerance laser (Figure 1(b)). The second part is the start trigger, it has three conditions before of data acquisition, one is the trigger output of BSC203 and the other is the reference sensor signal ( $D_R$ ). When the reference voltage is higher than “BAJO” and lower than “ALTO”, the data acquisition begins. The

\* [ruben\\_tol@yahoo.com.mx](mailto:ruben_tol@yahoo.com.mx)



**Figure 1.** (a) NI cRIO9074 system and BSC203 driver communication. For 9401: pin 1 = GND, pin 14 = digital input and pin 20 = digital output; and for BSC203: pin 9 = GND, pin 4 = Trigger input and pin 12 = Trigger output [5, 6]; (b) Introduction of parameters; (c) Z-scan Technique Control Panel.

module NI9234 provides this signal reference. These conditions are shown in Figure 2(a).

The next part is three-sensor data acquisition. The first sensor is the reference phototransistor ( $D_R$ ); the second is the closed aperture ( $D_C$ ) and third is open aperture ( $D_O$ ). These signals are the RMS values (Figure 2(b)).

The fourth part of program is the advance of linear translation stage. This advance depends on the scan resolution. In the case of Au nanofluid, it is 0.5mm. Finally, a table received the experimental data (Figure 2(c)). These data are processed and fits with the equation 2 in Origin to determine the nonlinear refraction index.

Then in general, the program in LabVIEW after the RUN, it calculates the number of measures, then it waits for the BSC203 trigger, takes the data, writes in the table and sends a trigger for next measure, and so on until the number of measures is completed. The control panel is shown in Figure 1(c).

**Experimental set-up**

The implementation of all optic parts of the z-scan technique begins with the characterization of diode laser. In this case  $\omega_0 = 30.1 \mu\text{m}$  (this laser beam waist at the focus length was calculated with  $\omega = 450 \mu\text{m}$ ) and the Rayleigh length was found to be satisfied the basic criteria of the Z-scan experiment,  $I_0 = 1.75 \text{ KW/cm}^2$  is the on-axis irradiance at focus ( $z = 0$ ),  $\lambda = 532 \text{ nm}$  [6]. A microcontroller circuit

modulated the laser to 10 Hz. The lens ( $L_1$ ) is 8 cm focal length. Other important part is the alignment; ensuring that the laser is perpendicular to the lens, sample and detectors. Also, the Linear Translation Stage should to be parallel to Z axis. The Figure 3 shows the experimental set-up.

To calculate the nonlinear refractive index ( $n_2$ ) were used the next equations [2,4].

$$\Delta\Phi_0 = kn_2I_0L_{eff} \tag{1}$$


Where  $L_{eff} = (1 - \exp(-\alpha_0L))/\alpha_0$  is the effective length of nonlinear medium,  $k = 2\pi/\lambda$  is the wave vector,  $\alpha_0$  is the linear absorption coefficient of the samples (L denotes the sample thick-ness),  $\Delta\Phi_0$  is induced phase shift, proportional at the transmittance variation between peak and valley positions describe for Eq. 2.

$$T(z, \Delta\Phi) = 1 + 4\Delta\Phi_0 \left( \frac{z}{z_0} \right) / \left( \left( \frac{z}{z_0} \right)^2 + 1 \right) \left( \left( \frac{z}{z_0} \right) + 9 \right) \tag{2}$$

Where  $z$  is the position,  $z_0$  is the Rayleigh length, finally the induced phase shift is determinates by Eq. 3.

$$\Delta T_{p-v} = 0.406(1 - S)^{0.25} |\Delta\Phi_0| \tag{3}$$

Here  $S$  is the linear transmittance of the aperture. The sample was scanned along a Z-axis by our system.



**Figure 2.** (a) Conditions to begin the data acquisition; (b) Data acquisition; (c) Data table in block diagram.




Figure 3. Z-scan experimental set-up.

**Results and Discussion**

The synthesized sample was composed of Au nanoparticles of 30 nm in average diameter and the tested sample consists of an Au nanofluid with 0.033 mmol in 3 ml of water (initial concentration). After characterization, it had a SPR to 520 nm. The nanofluid was placed in a quartz cuvette of 1mm of optical thickness. All the experiments were performed at room temperature. In addition, was diluted two ml of nanofluid with 0.5 ml of water, until get four samples, the last was 50 % nanofluid and 50 % of water to observed the signal of Z-scan respect to concentration. Figure 4 (a - e) shows the closed aperture Z-scan curves obtained for Au nanofluids.

The circle symbols represent the experimental data and the solid lines is the theoretical fit. The experimental data show symmetry curves. The peak followed by valley illustrates a self-focusing effect for a negative change in refraction. The solid lines show the theoretical fitting using a well-known normalized transmittance. In Table 1 these experiments are summarizing.

It is important to specify that there are many works on the synthesis of nanoparticles with different experimental conditions such as the size of the nanoparticle, the type of concentration, solvent stabilizer and solvent type [7-15]. Therefore, these differences cause different effects on colloid nonlinearities. Esmacil Shahriari *et al.* [11] found nonlinear refractive indices using nanoparticles of 30.5 nm particle size and concentration of  $2.3 \times 10^{-3}$  M in water. These results are very similar to those obtained in this work. In both papers, reduction reactions (redox) where a salt of gold is reduce for nanoparticles synthesis are used. These results are very similar to those obtained in this work.

Finally, Marcano *et al.* [14], in laboratory practice made inexpensive experiments for observing thermal lens and Z-scan for a sample of iodine in ethanol, in which there is no automatic control system for the experimental data with greater accuracy. Therefore, in this work, automation system by controlling the linear actuator and synchronization with

Table 1. Data summary for Z-scan experiments

Au nanofluid / water (mL)	$\Delta T_{p-v}$	$\Delta \Phi_0$	$\alpha$ (cm <sup>-1</sup> )	$n_2$ (cm <sup>2</sup> /W) $\times 10^{-7}$
2/0	0.55	1.61	0.65	-8.02 ± 0.02
2/0.5	0.528	1.55	0.38	-7.60 ± 0.03
2/1	0.403	1.18	0.30	-5.78 ± 0.03
2/1.5	0.315	0.92	0.17	-4.49 ± 0.04
2/2	0.274	0.80	0.22	-3.91 ± 0.03




Figure 4. Au nanofluid (a) Initial concentration (2 ml); (b) - (e) Successive dilutions with 0.5 ml, 1.0 ml, 1.5 ml and 2.0 ml water.

the laser pulse with a more versatile and fast application was proposed.

## Conclusion

The implementation of the Z-scan technique using NI cRIO 9074 system is functional and the results show coherence with other author's obtained values of refractive index coefficient nonlinear of similar order [11].

The implementation of this system with NI cRIO 9074 system offers flexible implementation and is relatively simple, compared to other techniques in optic laboratories [14]. In addition, it can be seen that the measurements are accurate and reliable.

## Acknowledgements

The authors are thankful to CONACYT, UAEMex, COFAA and SIP - IPN for financial support.

## References

- [1]. M. Sheik-Bahae, A.A. Said, E. W. Van Stryland, *Opt. Lett.* **14**, 955 (1989).
- [2]. M. Sheik-Bahae, A.A. Said, T.H. Wei, D.J. Hagan, E.W. Van Stryland, *IEEE J. Quant. Electron.* **26**, 760 (1990).
- [3]. M.F. Jiménez-Salinas, A.A. Rodríguez-Rosales, R. Ortega-Martínez, C.J. Román-Moreno, O.G. Morales-Saavedra, N. Qureshi, *Proc. SPIE* **8011**, 80112B (2011).
- [4]. J.L. Jimenez-Pérez, R. Gutiérrez-Fuentes., J.F. Ramirez-Sanchez, O. Vidal, D. Téllez-Sánchez, Z. Correa-Pacheco, A. Cruz-Orea, J. García, *Adv. Nanopart.* **2**, 223 (2013).
- [5]. Thorlabs Inc., *BSC202 and BSC203 Benchtop Stepper Motor Controller User Guide* (rev. Dec. 2016).
- [6]. C.L. Clark, *LabVIEW Digital Signal Processing* (McGraw-Hill, 2005).
- [7]. M. Trejo-Durán, D. Cornejo-Monroy, E. Alvarado-Méndez, A. Olivares-Vargas, V.M. Castaño, *J. Eur. Opt. Soc., Rapid Publ.* **9**, 14030 (2014).
- [8]. H. Nadjari, Z.A. Azad, *Opt. Laser Technol.* **44**, 1629 (2012).
- [9]. M.H. Majles Ara, Z. Dehghani, R. Sahraei, A. Daneshfar, Z. Javadi, F. Divsar, *J. Quant. Spectrosc. Radiat. Transfer* **113**, 366 (2012).
- [10]. M.A. Camacho López, R. Gutiérrez Fuentes, O.A. García Garduño, Y. León Marroquín, I. Hernández Ponce, D.A. Bobadilla Díaz, A. Pliego Carrillo, *VI congreso Nacional de Tecnología Aplicada a Ciencias de la Salud*, (Puebla, Pue. 2015).
- [11]. E. Shahriari, W. Mahomood, M. Yunnus, K. Naghavi, Z.A. Talib, *Opt. Commun.* **283**, 1929 (2010).
- [12]. E. Shahriari, W. Mahommod, M. Yunus, K. Naghavi, E. Saion, *Am. J. Eng. Appl. Sci.* **3**, 260 (2010).
- [13]. E. Shahriari, W. Mahmood, M. Yunus, E. Saion, *Braz. J. Phys.* **40**, 256 (2010).
- [14]. L. Rodríguez, J. Ramírez, A. Marcano, *Rev. Mex. Fis.* **E51**, 1 (2005).
- [15]. J.L. Jiménez Pérez, G. López Gamboa, Z.N. Correa Pacheco, J.F. Sánchez Ramírez, M. Sanchez Rivera, M. Salazar Villanueva, *Int. J. Eng. Tech. Res.* **3**, 162 (2015).