



Laccase-Based Biosensor for Determination of Acetaminophen

Diogo D. CORDEIRO & Éric de S. GIL*

*Faculdade de Farmácia – UFG – Av. Universitária com 1ª Avenida s/n,
Setor Universitário - Goiânia - Goiás – Brasil*

SUMMARY. The use of biosensors based on laccase is an interesting alternative for monitoring phenolic compounds in pharmaceutical and environmental analysis. The detection and determination of acetaminophen in pharmaceutical samples, using a biosensor with carbon paste modified with laccase is described. The method showed good linearity ($r = 0.9954$) and low detection and quantification limits ($2.4 \times 10^{-6} \text{ mol.L}^{-1}$ and $7.9 \times 10^{-6} \text{ mol.L}^{-1}$, respectively). Furthermore, the proposed methodology using an enzyme electrode showed appropriated reproducibility and selectivity.

KEY WORDS: Acetaminophen, Biosensor, Laccase.

* Author to whom correspondence should be addressed. *E-mail:* ericsgil@gmail.com