

Tinning, Peter W. and Scrimgeour, Ross and McConnell, Gail (2018) Standing wave microscopy of red blood cell membrane morphology with high temporal resolution. In: ELMI 2018, 2018-06-05 - 2018-06-08.

This version is available at https://strathprints.strath.ac.uk/64471/

Strathprints is designed to allow users to access the research output of the University of Strathclyde. Unless otherwise explicitly stated on the manuscript, Copyright © and Moral Rights for the papers on this site are retained by the individual authors and/or other copyright owners. Please check the manuscript for details of any other licences that may have been applied. You may not engage in further distribution of the material for any profitmaking activities or any commercial gain. You may freely distribute both the url (<u>https://strathprints.strath.ac.uk/</u>) and the content of this paper for research or private study, educational, or not-for-profit purposes without prior permission or charge.

Any correspondence concerning this service should be sent to the Strathprints administrator: strathprints@strath.ac.uk

The Strathprints institutional repository (https://strathprints.strath.ac.uk) is a digital archive of University of Strathclyde research outputs. It has been developed to disseminate open access research outputs, expose data about those outputs, and enable the management and persistent access to Strathclyde's intellectual output.

Standing wave microscopy of red blood

cell membrane morphology with high temporal resolution

Peter W. Tinning¹, Ross Scrimgeour² and Gail McConnell¹

1 Department of Physics, University of Strathclyde, Glasgow, United Kingdom.

2 Strathclyde Institute of Pharmacy and Biomedical Sciences, University of Strathclyde, Glasgow, United Kingdom.

Introduction

Cool

A)

D)

2 µm

2 µm

Simply Better Control

Widefield fluorescence microscopy is an integral tool for life science imaging though the achievable resolutions are limited by the diffraction nature of light. One technique to increase the axial resolution is known as standing wave microscopy [1]. The standing wave can be generated by placing a mirror at the specimen plane which causes interference between the incoming and reflected excitation illumination. The axial resolution is reduced to $\lambda/4n$ as only fluorophores which are in the location of the full width at the half maximum of the antinodes are excited [2] resulting in periodic bands of fluorescence. Previous investigations using this technique imaged red blood cells upon a mirror using a confocal laser scanning microscope meaning that the temporal resolution was limited [2] and fast membrane movements could not be captured. We present here the first demonstration of widefield standing wave microscopy upon red blood cells with high temporal resolution and an axial resolution better than 100 nm.

Red blood cells were isolated and labelled with the fluorescent membrane dye Dil (10 μ M). They were then plated onto a broadband reflector under a coverslip

Methods

- \pm 1.5 nm and using specimen plane powers of 1.71 \pm 0.01 mW.
- A camera binning n = 2 was used and an exposure time of 33 ms with a recording duration of 33 seconds to obtain movies 1000 frames in





Fig. 2: Frames taken from the video-rate standing wave movie of the bottom half of a red blood cell. The frames presented as A-F are frames 1, 200, 400, 600, 800 and 999. Using this technique we are able to observe rapid membrane deformations in real time with an axial resolution of 91.4 nm and a lateral resolution of 304 nm.

Fig. 3: 2D computational reconstruction of the video-rate standing wave movie. (Fig. 2). Extracting the antinodal planes allows for **quantitative analysis to be carried out** on the data as well as easier observation of antinodal plane movements.

2 µm

Fig. 4: 3D reconstruction of frame one of the standing wave red blood cell in which the axial aspect ratio has been exaggerated to aid visualisation. A total of 835 reconstructed frames were obtained from the raw 1000 frame movie which allows the **observation of how the 3D** membrane topography of the red blood cell changes in real time.

Photobleaching Rates Compared To Widefield Microscopy



Fig. 5: Average normalised intensity obtained from A) standing wave imaging and B) widefield epifluorescence imaging of red blood cells (n = 10). It can be seen from these

Phototoxicity Compared To Widefield Microscopy

Fig. 6: Standing wave images of a red blood cell. The frames presented as A-F are at time points 15, 360, 720, 1080, 1440 and 1800 s. The decay of blood cell overtime the red 2 µm resembles the process of photohemolysis observed under focussed low power laser illumination [4]. It is hypothesised that this process is due to the interaction between reactive oxygen species and the membrane proteins spectrin and band 3 [3].

Fig. 7: Widefield images of a red blood cell at the same timepoints as Fig. 6. It can be seen when compared to Fig. 6 that the red blood cell appears to decay in a similar manner and rate. This demonstrates that our standing wave technique does not cause any greater amount of photo-toxicity to the specimen.

graphs that;

1) Standing wave imaging causes no significantly greater amount of photobleaching than standard widefield imaging.

2) Standing wave imaging obtains significantly brighter images due to multiple in focus antinodal planes



B)

Conclusions and Future Work

- This is the first demonstration of widefield standing wave imaging of red blood cells at a video rate with an axial resolution of under 100 nm.
- By applying this technique we have been able to observe rapid membrane deformations in real time by simply placing our specimens upon a mirror.
- We have also shown that using this technique does not induce any greater • amount of phototoxicity or bleaching to the specimen than widefield microscopy.
- This work has now been published in Biomedical Optics Express (2018).
- Future work involves applying this technique in other cell types in order to study morphological changes (blebbing, cell division or migration) and membrane potential imaging in excitable cells.
- We also aim to use multiple wavelengths (excitation or emission) in order to reduce the nodal plane contributions in the images and determine the morphology of unknown specimens by the antinodal plane ordering. References: [1] B. Bailey et. al., Nature, 1993. [2] R. Amor et. al., Sci. Rep. 2014. [3] F. Wong et. al., J. Microscopy, 2007.