

## UNIVERSITY OF LEADING THE WAY WESTMINSTER<sup>#</sup>

## WestminsterResearch

http://www.westminster.ac.uk/research/westminsterresearch

### Editorial

Introducing EbolaCheck: potential for point-of-need infectious disease diagnosis

**Sterghios Moschos** 

This is an Author's Accepted Manuscript of an article published by Informa Healthcare in Expert Review of Molecular Diagnostics on 31 August 2015, 15 (10). pp. 1237-1240. © Informa 2015. Available online at:

http://dx.doi.org/10.1586/14737159.2015.1084228

The WestminsterResearch online digital archive at the University of Westminster aims to make the research output of the University available to a wider audience. Copyright and Moral Rights remain with the authors and/or copyright owners.

Users are permitted to download and/or print one copy for non-commercial private study or research. Further distribution and any use of material from within this archive for profit-making enterprises or for commercial gain is strictly forbidden.

Whilst further distribution of specific materials from within this archive is forbidden, you may freely distribute the URL of WestminsterResearch: (<u>http://westminsterresearch.wmin.ac.uk/</u>).

In case of abuse or copyright appearing without permission e-mail <u>repository@westminster.ac.uk</u>

# Introducing EbolaCheck: potential for point-of-need infectious disease diagnosis

Sterghios Athanasios Moschos

Westminster Genomic Services, Department of Biomedical Sciences, Faculty of

Science and Technology, University of Westminster, London, UK.

Tel: +44 (0)207 911 5000

s.moschos@westminster.ac.uk

#### Abstract

The 2013-5 Ebolavirus disease (EVD) humanitarian crisis has spurred the development of laboratory-free, point of care (POC) nucleic acid testing (NAT) solutions. EbolaCheck is an international consortium of public health, academic and biotechnology industry stakeholders aiming to deliver clinical molecular diagnostic (MDx) standard of care (SOC) testing suitable for the West African milieu within 12 months. In this article the current status of the EbolaCheck platform is discussed in the context of the current regulatory framework. Future goals to achieve differential diagnosis of hemorrhagic fever disease from <5 microliters of whole blood samples (WBS) or mucosal biofluids, in a single tube process, in under 40 minutes and with minimal operator training are presented.

Keywords: Ebola, molecular diagnostics, nucleic acid testing, point of care, portable, non-profit.

#### Background: Clinical diagnosis of Ebolavirus disease

Ebolavirus disease (EVD) is a haemorrhagic fever disease (HFD) caused by members of the *filoviridae* family of RNA viruses. The filamentous Ebolavirus virion (~90 x 1000 nm) houses a 7 gene, ~19 kb genome packed in a nucleoprotein (NP) sheath. Transmission is mediated via the Ebolavirus transmembrane glycoprotein (GP) primarily via macrophage/monocytes. The glycoprotein also features immunomodulation, immune evasion and endothelial barrier disruption roles.[1] The monocytic tropism of Ebolavirus mediates proinflammatory responses during replication that amplify infectivity and pathology, collectively resulting in the internal haemorrhage and organ failure characteristic of the later stages of disease. [2,3]

Diagnosis is extremely difficult [2,4] as symptoms mimic other HFDs, flu, or gastrointestinal infections, which do not preclude Ebolavirus co-infection. [4,5] Transmission risk increases in line with symptom severity, mirroring viraemia;[6] pre-symptomatic patients are not considered contagious and may remain asymptomatic for up to 21 days. [3] Confirmation of Ebolavirus as the causal disease agent requires clinical molecular diagnostic (MDx) laboratory solutions. To date, USD\$100, <8 hr long, transcription polymerase chain reaction (RT-PCR) nucleic acid tests (NAT) on RNA extracts from 3.5 ml of whole blood sample (WBS) are the method of choice. [7]

However, at the height of the EVD outbreak lack of capacity in West Africa required sample shipment overseas resulting in 3-5 day turnaround times and post-mortem diagnosis. [1,8] The need for a true point-of-need NAT was acute, yet no *in vitro* diagnostic (IVD) test had received regulatory clearance. 'Homebrew' assays were based on Trombley *et al.* (the 'Trombley' assays; United States Army Medical Research Institute for Infectious Disease; USAMRIID) [9] or Panning *et al.*. [10] These eventually received USFDA emergency use authorisation (EUA; EZ1 assay) or were made commercially available under the self-certification CE marking principles (Altona RealStar<sup>®</sup> Filovirus Screen), [11] respectively.

#### Molecular diagnostics for infectious diseases at the point of need

Following 9/11 and the subsequent airborne viral disease pandemics, efforts were made to develop decentralised, point of care (POC) NAT's [12]. The resulting solutions, however, were not designed with resource-limited settings in mind, [13] despite the ASSURED criteria espoused by the WHO. [14] Thus, the need for a safe, cheap, simple, robust, portable and battery operated solution remained, presenting an attractive development opportunity for emerging NAT technologies. However, clinical development costs, [13] convoluted intellectual property landscapes and industry doubts over outbreak duration and return on investment potential, presented substantial obstacles. Poignantly, despite corporate social responsibility opportunities, to date, all of the major diagnostics manufacturers that engaged in the Ebola response offered primer-probe kits for existing lab-based platforms, or developed 'cassette' kits for existing, closer-topatient systems. Importantly, these cassette systems maintain for-profit pricing structures for low and middle-income countries, even following receipt of philanthropic donations in support of their development. The monetisation/investment barrier remains cornerstone to both regulator and non-government support organisation efforts. [15]

Yet despite large industry indifference several academic groups and start up/spinout companies sought to address the POC clinical diagnostic need. However, they faced scepticism from some regulatory bodies regarding manufacturing capacity, quality assurance, commercial launch/support and distribution capability. [16] Thus, little consideration was given to post proof-ofprinciple, non-profit production and distribution opportunities similar to regulator-certified, generic pharmaceuticals supply chain models. Under normal circumstances this would appear appropriate considering the high risk to individual and public health on account of false positive or false negative misdiagnosis (WHO category 4 IVD classification). However, on August 2014 WHO declared the West African Ebola outbreak as a public health emergency of international concern (PHEIC). Interestingly, this motivated the FDA to enable EUA approvals; in contrast, the WHO demanded engagement through the full pre-qualification process. This diverged significantly from the documented successes with other WHO-listed, FIND Diagnostics-vetted, but academic-lead efforts to address neglected disease diagnostics need. The net result was limited performance validation facilitation (access to stored patient samples managed by the WHO) for innovations aiming to address the humanitarian need in the affected countries at the point of care, in lieu for questionable support to preferred lab-based platforms.

#### EbolaCheck: the team

The EbolaCheck consortium was formed in response to the August 2014 call of the Research for Health in Humanitarian Crises (r2hc) programme, managed by Enhancing Learning and Research for Humanitarian Assistance (ELRHA; www.elrha.org). The Research for Health in Humanitarian Crises (R2HC) programme aims to improve health outcomes by strengthening the evidence base for public health interventions in humanitarian crises (visit www.elrha.org/work/r2hc for more information). The goal of the joint effort between University of Westminster, BioGene Ltd., Public Health England (PHE), USAMRIID, and the Kwame Nkrumah University of Science and Technology (KNUST) funded through R2HC is to deliver by November 2015 a novel point of need NAT solution for simple, rapid and safe patient triage for EVD anywhere in West Africa.

#### **EbolaCheck: Key principles**

EbolaCheck can be divided into four sub-systems: the NAT instrument, the EVD assay, the WBS reaction formulation and the reaction consumable. Together, they aim to replace the clinical MDx standard of care (SOC) with a rapid, point-of-need, sample-to-answer format.

#### Low cost suitable for West Africa

A simple, patent-protected, energy and engineering-efficient method enables rapid (<2 min), single-tube access to pathogen & host nucleic acids in biofluids with no need for microfluidics. Direct compatibility with standard, cryoprotectable RT-PCR biochemistries further reduces overall cost. Crucially, EbolaCheck will be available to support the on-going, WHO-declared, EVD humanitarian crisis in Africa at cost only.

#### Clinical standard of care reliability

The Trombley assay sets for Ebolavirus Zaire GP and NP [9] were migrated to EbolaCheck (Trombley+) to i) minimise delays, ii) avoid complex licensing negotiations and iii) on account of emerging field performance evaluation data. Multiplexed use of the Trombley+ assay sets also discriminate vaccinated from infected patients; NP is not found in the two most advanced EVD clinical vaccine candidates [17,18], a problem in on-going vaccination clinical trials pursued by other r2hc funded programmes (Gilbert S., personal communication). USAMRIID have demonstrated performance across 5 logs of viral RNA genome equivalents (GE) with 100% analytical specificity against 65 other pathogens and analytical sensitivities of 0.001 (NP) and 0.0001 (GP) plaque-forming units (PFU) per reaction. [9] The roughly 4,000 GE/PFU ratio observed under biosafety level 4 (BSL4) experimentation [19], suggests a lower limit of detection (LLOD) of 10 GE/reaction, or 10<sup>4</sup> GE/ml of WBS. Given typical time-to-presentation in autumn 2014 was >3 days post symptom onset, a LLOD goal of  $10^4$  GE/ml WBS was set for the Trombley+ assays on EbolaCheck. Present performance data on surrogate pseudoviral templates indicate 9 logs of quantitative linear dynamic range with a lower limit of quantification of 66 GE/reaction and LLOD of 6 GE/reaction, i.e. in line with our performance targets.

#### Simple, sample-to-result standard operating procedure.

The plethora of reports on 'simple' medical device misuse by end-users in the developed world underscore the importance of ensuring device reliability, particularly with category 4 IVD devices operated under significant duress, in environmentally challenging conditions. [13] The EbolaCheck standard operating procedure (SOP) consists of:

1) reagent unpacking and automated rehydration,

2) 5 microliter WBS collection by fingertip lancet puncture and MicroSafe<sup>®</sup> capillary collection,

3) Sample ejection into the rehydrated consumable,

4) lock and loading onto the EbolaCheck instrument, and

5) run initiation by touchscreen input.

Availability and status of the 8 random-access testing stations is visually identified on the front-facing touchscreen. Patient status is simply reported as positive, negative or problematic, with the latter indicating a need to repeat the test due to a failure. Full run kinetics, analytics and diagnostics can be accessed on-screen or over a WiFi connection.

#### Safety

The 5 microliter WBS requirement of EbolaCheck presents a significant risk reduction to both HCW and HFD patients compared to the closed system, 3.5 ml Vacutainer<sup>®</sup> Eclipse<sup>™</sup> needle and Vacutainer<sup>®</sup> sample SOC protocol. Thermal cycling is expected to destroy EVD [20]; used, sealed consumables are nonetheless discarded as BSL4 clinical waste. The instrument is fully compatible with chlorine dioxide surface sterilisation [20] and designed against ingress of liquids or internal condensation [13]. Secure WiFi interface permits remote system checks, maintenance and full reaction data off-boarding. The random access stations also self-diagnose errors and automatically shut down to prevent misdiagnosis.

Tests with full personal protective equipment suggest the EbolaCheck HCW SOP takes under 2 minutes to complete by minimally trained individuals, with time to results in <40 minutes; real time reaction progression monitoring suggests high viraemia positive results could be called in as little as 20 minutes.

#### Portability

Field experience from in-country PHE response teams advised against easily removed, small-form designs, highlighting the need for higher throughput. The ruggedized, 8-well form maintains power supply independence through either mains and/or car battery/alternator power sources. Furthermore, energy consumption modelling indicates solar power supply to be achievable. Design for safety also achieves durability and reliable operation in savannah, coastal and jungle conditions, without corrosion or performance deterioration: simulated environment tests indicate the instrument can complete runs at temperatures as high as 50°C with 98% humidity, and as low as -20°C.

#### **Development timeline**

Prototype design, engineering and assay development was initiated in November 2014. Internal assay standards containing the Trombley assay targets were developed in MS2 phage icosahedron (Armored RNA®) [21] (commercially available) and lipid bilayer enveloped HIV pseudovirus [22] (open access) formats. Although 26nm and 80-100nm in size respectively, these represent a vast cadre of viral pathogens. Thus, BSL4 study requirements have been reduced to confirmatory studies using live Ebolavirus, and yielded data supporting EbolaCheck platform utility against other viral pathogens. BSL4 studies are thus

limited to performance evaluation testing against the clinical SOC NAT Trombley assay on culture preparations of Ebolavirus and fresh WBS derived from nonhuman primate models of Ebolavirus infection. In-country testing with fresh or stored patient samples is not expected on account of continued outbreak decline and current WHO priorities to established technologies. However, at least 3 instruments will be tested in West Africa using mock sample preparations to confirm system operation, portability and reliability in urban, rural and remote environments.

#### **Future directions**

Our early data support multiplexed detection and quantification potential of 3-4 NAT targets in WBS on EbolaCheck. As positive [23] and detrimental [5] coinfections are common amongst EVD patients, expansion of multiplexing is necessary, but unlikely to exceed concomitant amplification capability need beyond 5 targets. Field data also indicate mucosal biofluids such as semen [24], ocular fluid [25] and breast milk [26] might be viral depots in convalescence. Interestingly, culturally acceptable alternatives such as saliva [27] and gingivalcrevicular fluid [28], might also be of use for HFD diagnosis. Thus, demonstrating EbolaCheck compatibility with these mucosal biofluids will expand point of need monitoring and surveillance capability and introduce the opportunity for needlefree testing. Early feasibility studies indicate this may enable differential HFD diagnosis with minimal cost of goods increase.

#### **Concluding remarks**

Of the 9 EVD NATs that have received to date USFDA EUA, 3 involve complex cartridge/microfluidic systems. Only the 90 minute Cepheid Xpert<sup>®</sup> Ebola assay (May 2015) is reasonably priced for the West African milieu at ~US\$20 per test, despite charitable backing. With a comparable assay cadre and LLOD to EbolaCheck, it features a 3 log, non-quantitative dynamic range in highly diluted WBS, requires sample pre-processing, multiple mechanical steps and a separate personal computer and barcode scanner. Despite >10,000 instruments placed worldwide this WHO-selected platform costs US\$17,000-17,500 to eligible countries. Thus, per-unit scaled production costs are comparable to the current manufacturing cost of EbolaCheck prototypes and the Trombley+ EVD assays. The EbolaCheck consortium has demonstrated that humanitarian crises can motivate efforts to the significant potential benefit of those in need as well as leverage development opportunities for appropriately positioned technologies from socially responsible industry with commercial interests in the West. The EbolaCheck consortium is presently seeking charitable support towards scale-up production and delivery of the first differential HFD diagnosis solution, to be provided at cost for any future WHO-declared humanitarian crises.

#### References:

- Cenciarelli O, Pietropaoli S, Malizia A *et al.* Ebola virus disease 2013-2014 outbreak in west Africa: an analysis of the epidemic spread and response. *International journal of microbiology*, 2015, 769121 (2015).
- 2. Kalra S, Kelkar D, Galwankar SC *et al.* The emergence of ebola as a global health security threat: from 'lessons learned' to coordinated multilateral

containment efforts. *Journal of global infectious diseases*, 6(4), 164-177 (2014).

- 3. Ghazanfar H, Orooj F, Abdullah MA, Ghazanfar A. Ebola, the killer virus. *Infectious diseases of poverty*, 4, 15 (2015).
- 4. Boisen ML, Schieffelin JS, Goba A *et al.* Multiple circulating infections can mimic the early stages of viral hemorrhagic fevers and possible human exposure to filoviruses in Sierra Leone prior to the 2014 outbreak. *Viral immunology*, 28(1), 19-31 (2015).
- Gire SK, Goba A, Andersen KG *et al.* Genomic surveillance elucidates Ebola virus origin and transmission during the 2014 outbreak. *Science*, 345(6202), 1369-1372 (2014).
- 6. Towner JS, Rollin PE, Bausch DG *et al.* Rapid diagnosis of Ebola hemorrhagic fever by reverse transcription-PCR in an outbreak setting and assessment of patient viral load as a predictor of outcome. *Journal of virology*, 78(8), 4330-4341 (2004).
- Spengler JR, McElroy AK, Harmon JR, Stroher U, Nichol ST, Spiropoulou CF. Relationship Between Ebola Virus Real-Time Quantitative Polymerase Chain Reaction-Based Threshold Cycle Value and Virus Isolation From Human Plasma. *The Journal of infectious diseases*, (2015).
- 8. Tambo E, Ugwu EC, Ngogang JY. Need of surveillance response systems to combat Ebola outbreaks and other emerging infectious diseases in African countries. *Infectious diseases of poverty*, 3, 29 (2014).
- 9. Trombley AR, Wachter L, Garrison J *et al.* Comprehensive panel of realtime TaqMan polymerase chain reaction assays for detection and absolute quantification of filoviruses, arenaviruses, and New World hantaviruses.

The American journal of tropical medicine and hygiene, 82(5), 954-960 (2010).

- 10. Panning M, Laue T, Olschlager S *et al.* Diagnostic reverse-transcription polymerase chain reaction kit for filoviruses based on the strain collections of all European biosafety level 4 laboratories. *The Journal of infectious diseases*, 196 Suppl 2, S199-204 (2007).
- 11. Moschos SA. Genomic biomarkers for patient selection and stratification: the cancer paradigm. *Bioanalysis*, 4(20), 2499-2511 (2012).
- 12. Holland CA, Kiechle FL. Point-of-care molecular diagnostic systems--past, present and future. *Current opinion in microbiology*, 8(5), 504-509 (2005).
- 13. Peeling RW, McNerney R. Emerging technologies in point-of-care molecular diagnostics for resource-limited settings. *Expert review of molecular diagnostics*, 14(5), 525-534 (2014).
- 14. Peeling RW, Holmes KK, Mabey D, Ronald A. Rapid tests for sexually transmitted infections (STIs): the way forward. *Sexually transmitted infections*, 82 Suppl 5, v1-6 (2006).
- 15. Chua AC, Cunningham J, Moussy F, Perkins MD, Formenty P. The Case for Improved Diagnostic Tools to Control Ebola Virus Disease in West Africa and How to Get There. *PLoS Negl Trop Dis*, 9(6), e0003734 (2015).
- 16. Perkins MD, Kessel M. What Ebola tells us about outbreak diagnostic readiness. *Nature biotechnology*, 33(5), 464-469 (2015).
- 17. Sarwar UN, Costner P, Enama ME *et al.* Safety and immunogenicity of DNA vaccines encoding Ebolavirus and Marburgvirus wild-type glycoproteins in a phase I clinical trial. *The Journal of infectious diseases*, 211(4), 549-557 (2015).

- 18. Regules JA, Beigel JH, Paolino KM *et al.* A Recombinant Vesicular Stomatitis Virus Ebola Vaccine - Preliminary Report. *The New England journal of medicine*, (2015).
- 19. Rossi CA, Kearney BJ, Olschner SP *et al.* Evaluation of ViroCyt(R) Virus Counter for rapid filovirus quantitation. *Viruses*, 7(3), 857-872 (2015).
- 20. Cook BW, Cutts TA, Nikiforuk AM *et al.* Evaluating environmental persistence and disinfection of the Ebola virus Makona variant. *Viruses*, 7(4), 1975-1986 (2015).
- 21. Pasloske BL, Walkerpeach CR, Obermoeller RD, Winkler M, DuBois DB. Armored RNA technology for production of ribonuclease-resistant viral RNA controls and standards. *Journal of clinical microbiology*, 36(12), 3590-3594 (1998).
- 22. Mather ST, Wright E, Scott SD, Temperton NJ. Lyophilisation of influenza, rabies and Marburg lentiviral pseudotype viruses for the development and distribution of a neutralisation-assay based diagnostic kit. *Journal of virological methods*, (2014).
- 23. Lauck M, Bailey AL, Andersen KG, Goldberg TL, Sabeti PC, O'Connor DH.
  GB virus C coinfections in west African Ebola patients. *Journal of virology*, 89(4), 2425-2429 (2015).
- 24. Christie A, Davies-Wayne GJ, Cordier-Lasalle T *et al.* Possible sexual transmission of ebola virus liberia, 2015. *MMWR. Morbidity and mortality weekly report*, 64(17), 479-481 (2015).
- 25. Varkey JB, Shantha JG, Crozier I *et al.* Persistence of Ebola Virus in Ocular Fluid during Convalescence. *The New England journal of medicine*, (2015).

- 26. Moreau M, Spencer C, Gozalbes JG *et al.* Lactating mothers infected with Ebola virus: EBOV RT-PCR of blood only may be insufficient. *Euro surveillance : bulletin Europeen sur les maladies transmissibles = European communicable disease bulletin*, 20(3) (2015).
- 27. Formenty P, Leroy EM, Epelboin A *et al.* Detection of Ebola virus in oral fluid specimens during outbreaks of Ebola virus hemorrhagic fever in the Republic of Congo. *Clinical infectious diseases : an official publication of the Infectious Diseases Society of America*, 42(11), 1521-1526 (2006).
- Corstjens PL, Abrams WR, Malamud D. Detecting viruses by using salivary diagnostics. *Journal of the American Dental Association*, 143(10 Suppl), 12S-18S (2012).

**Acknowledgements**: This work was funded by grant no. 13335 of the Research for Health in Humanitarian Crises (R2HC) Programme managed by ELRHA, funded equally by the Wellcome Trust and DFID.