

MID-INFRARED FREQUENCY COMB SPECTROSCOPY USING A VIRTUALLY IMAGED PHASED ARRAY

ADAM J. FLEISHER, *Chemical Sciences Division, National Institute of Standards and Technology, Gaithersburg, MD, USA.*

Here we present a new mid-infrared frequency comb system for rapid spectral acquisition using a virtually imaged phased array (VIPA) spectrometer.^a A difference-frequency generation comb, tuneable from 4.4 μm to 4.7 μm , was used to interrogate a single-pass absorption cell containing either N₂O or CO dilute in either N₂ or air. Precision molecular spectroscopy capabilities at timescales of less than 1 ms will be presented, and progress toward cavity-enhanced and time-resolved comb spectroscopies^b will be discussed.

^aL. Nugent-Glandorf et al., *Opt. Lett.* **37**, 3285 (2012)

^bA.J. Fleisher et. al., *J. Phys. Chem. Lett.* **5**, 2241 (2014)