BMJ Paediatrics Open

Improved oxygen systems in district hospitals in Lao PDR: a prospective field trial of the impact on outcomes for childhood pneumonia and equipment sustainability

Amy Zigrida Gray,^{1,2,3} Melinda Morpeth,^{1,2} Trevor Duke,^{1,2,3} David Peel,^{1,4} Christian Winter,⁵ Manivanh Satvady,⁶ Kongkham Sisouk,⁷ Bouasengnignom Prasithideth,⁷ Khamsay Detleuxay⁷

ABSTRACT

To cite: Gray AZ, Morpeth M, Duke T, *et al.* Improved oxygen systems in district hospitals in Lao PDR: a prospective field trial of the impact on outcomes for childhood pneumonia and equipment sustainability. *BMJ Paediatrics Open* 2017;1:e000083. doi:10.1136/ bmjpo-2017-000083

Received 16 May 2017 Revised 13 July 2017 Accepted 20 July 2017



¹Department of Paediatrics, Centre for International Child Health. The University of Melbourne, Parkville, Victoria, Australia ²The Royal Children's Hospital, Melbourne Australia. Parkville. Victoria, Australia ³Murdoch Childrens Research Institute, Parkville, Victoria, Australia ⁴Ashdown Consultants, Hartfield, UK ⁵WHO Country Office, Vientiane, Lao PDR ⁶Department of Health Care. Ministry of Health, Vientiane, Lao PDR ⁷Mahosot Hospital, Ministry of Health, Vientiane, Lao PDR

Correspondence to

Dr Amy Zigrida Gray; amy.gray@ rch.org.au

Background Hypoxaemia is a common and potentially fatal complication of many childhood, newborn and maternal conditions but often not well recognised or managed in settings where resources are limited. Oxygen itself is often inaccessible due to cost or logistics. This paper describes implementation of oxygen systems in Lao district hospitals, clinical outcomes after 24 months and equipment outcomes after 40 months postimplementation.

Methods A prospective field trial was conducted in 20 district hospitals, including 10 intervention hospitals that received oxygen concentrators and 10 control hospitals. Equipment outcomes were evaluated at baseline, 12, 24 and 40 months. Clinical outcomes of children under 5 years of age with pneumonia were evaluated using a before-and-after controlled study design with information retrospectively collected from medical records.

Results Fourteen (37%), 7 (18%) and 12 (34%) of 38 concentrators required repair at 12, 24 and 40 months, respectively. The proportion of children discharged well increased in intervention (90% (641/712) to 95.2% (658/691)) and control hospitals (87.1% (621/713) to 92.1% (588/606)). In intervention hospitals, case fatality rates for childhood pneumonia fell from 2.7% (19/712) preintervention to 0.80% (6/691) postintervention with no change in control hospitals (1.7% (12/713) preintervention and 2.3% (14/606) postintervention).

Conclusion Medium-term sustainability of oxygen concentrators in hospitals accompanied by reduced case fatality for childhood pneumonia has been demonstrated in Lao PDR. Significant local engineering capacity to address multiple causes of equipment malfunction was critical. The ongoing requirements and fragile structures within the health system remain major risks to long-term sustainability.

INTRODUCTION

Hypoxaemia, or insufficient oxygen levels in the blood, is a potentially fatal complication of many common causes of mortality in children, newborns and adults.¹ It is estimated

What is already known on this topic?

- Hypoxaemia is a potentially fatal complication of many common childhood, neonatal and adult health problems.
- Hypoxaemia is often not well recognised or treated in many low-resource settings, and oxygen is often inaccessible.
- Systematic use of pulse oximetry for detecting hypoxaemia and availability of reliable oxygen sources can reduce death rates from pneumonia in children.

What this study hopes to add?

- Represents one of the most comprehensive evaluations of implementation of oxygen concentrator systems in district hospitals.
- Improved outcomes for childhood pneumonia in hospitals that received oxygen concentrator systems that included equipment, engineering and clinical training.
- The potential maintenance and engineering burden of concentrators on already fragile health systems.

that at least 13% of children with WHO-defined pneumonia requiring hospitalisation have hypoxaemia, with wide geographical variation.^{2 3} Despite its importance, hypoxaemia is often not well recognised or managed in settings where resources are limited.^{4 5} Evidence from previous studies suggests systematic use of pulse oximetry for detecting hypoxaemia and availability of reliable oxygen sources can reduce death rates from pneumonia in children by up to 35%.⁶

In low and middle-income countries, oxygen can represent a significant hospital

expense. In countries with a user-pays health system oxygen may be available, but the cost to the patient can limit use or precipitate early discharge. High costs of oxygen partly reflect logistical difficulties in transporting oxygen, which also lead to intermittent availability of oxygen cylinders. High costs to patients and lack of access to oxygen potentially undermine community confidence in hospitals and are a major impediment to universal access to healthcare.

Oxygen concentrators—small machines that remove nitrogen from atmospheric air to generate 85%–95% oxygen—are one potential solution.⁷ They have been found to be feasible for delivery of oxygen in low-income countries, although there is variability in performance.^{8–10} For concentrator systems to be sustainable, they require a continuous and reliable power supply, a system for monitoring, maintenance and repair, as well as clinical staff trained in their use.⁷

The Lao Oxygen Therapy Pilot Project was established in recognition of the complex issues, and multiple barriers, relating to oxygen supply in Lao PDR. The aims of the project were to implement an affordable oxygen system in district hospitals and to evaluate its feasibility, sustainability and impact on patient care and outcomes. This paper describes the implementation approach, clinical outcomes at 24 months and equipment outcomes at 40 months postimplementation.

METHODS

Hospital selection

Intervention hospitals were selected based on criteria including availability of reliable electricity, absent or limited oxygen supply, remote geographical location and anticipated admission numbers. Two northern and three southern provinces were included for different climates and disease patterns. Randomisation was not feasible given the selection criteria; however, control hospitals were recruited and selected from the same provinces based on similarity to intervention hospitals. One provincial hospital was included as a control as it was equivalent to the larger district hospitals.

Implementation of oxygen systems

The project consisted of four phases; a situational assessment, establishment of the oxygen team and partners, systematic implementation of the oxygen systems and evaluation of clinical use of oxygen, equipment use and maintenance requirements, clinical outcomes and patient costs. (figure 1) The Lao Oxygen team included staff from the Lao Ministry of Health, local adult and paediatric intensive care physicians, provincial paediatricians, international paediatricians, an international biomedical engineer and staff from the Lao Country Office of the WHO. An agreement was made that oxygen delivered from project concentrators would be free to patients, removing any financial impediments to access.

Implementation of oxygen systems in selected hospitals occurred from September to November 2011. All 10 intervention hospitals were visited by the installation team to check existing resources, install the oxygen system and provide biomedical engineering training on its use and maintenance. This was followed closely by the clinical training team, consisting of clinicians from the Lao Oxygen team.

The clinical training team delivered Lao-specific training materials based on WHO guidelines for oxygen¹¹ and hospital care of children¹² which included videos, guidelines and case-based teaching sessions. On completion of training, each participant underwent an observational assessment based on recognition and treatment of hypoxia, as well as oximeter use. Control hospitals were given pulse oximeters but did not receive any training on their use.

Equipment

The VisionAire Oxygen Concentrator (AirSep, Buffalo, New York, USA) was chosen because of several unique features that were considered to be appropriate for Laos.⁷ Unlike other concentrators, the VisionAire has no external filter that typically require regular maintenance and replacement. Removing this maintenance step was felt to be advantageous. The concentrator was specified to deliver 5L per minute of oxygen in the concentration range of 82%–95% at 21°C degrees and operate at temperatures between 5°C and 35°C degrees with relative humidity up to 95%. Oxygen analysers (Hudson RCI, http://hudsonrci.com/) were provided to monitor the performance of concentrators.

A mains-powered, table-top oximeter was chosen from a major German manufacturer (Bitmos). It was selected to minimise risk of equipment loss and because the software rejects movement signals facilitating accurate use in children. Oximeter sensors with a guaranteed life of 1 year were sourced from Carril Instruments, Barcelona.

Evaluation design

A prospective field trial was conducted in 20 district hospitals, including 10 intervention hospitals that received oxygen concentrator systems and 10 control hospitals. This included a before-and-after controlled study of the outcomes of children less than 5 years of age with an admission diagnosis of pneumonia.

Equipment outcomes

Prospective evaluation of equipment outcomes was undertaken with data drawn from a number of sources including the field notes from meetings and the installation of oxygen systems, hospital reports and site visits to both control and intervention hospitals at baseline, 3, 12 and 24 months. Hospital equipment checklists and surveys were completed at both intervention and control hospitals at each visit. Data collected at these times included routine hospital admission data, hours of oxygen use as read from the concentrator display and the availability and function of equipment. An additional visit to intervention hospitals was conducted 40 months after implementation, more than 1 year after the formal

Situati	Time point	Intervention Control			
Situational assessment		A			
Hospital recruitment (n=20)		B	B		
Delivery of intervention and baseline evaluation		C D E F G I			
12 mo follow	nth post intervention up	F G H	F G H		
equipr		D			
24 mo follow	nth post intervention up	F G H	F G H		
40 mo follow	nth post intervention up	G			
Key A B	availability and use in La Hospital recruitment. Se proposed by the oxygen current availability of a s or limited oxygen supply	team, provincial and dist sufficiently reliable electric or difficulties in oxygen d	2011). pitals was based on criteria rict hospital staff including ity supply, current absence		
D	Control hospitals were s comparison. Delivery of oxygen syste installation team delivere equipment use, mainten Delivery of pulse oximet oximeters (Bitmos). Con Clinical training on oxyg workshops to all availab	elected from the same or em equipment and engine ed hands-on training to loc ance and repair. ers. Intervention hospitals trol hospitals received ha	ers with respiratory illness. neighbouring provinces for ering training. The cal hospital staff on received table-top ndheld oximeters (Lifebox). ver two-days via interactive material from the WHO		

Figure 1 Graphical representation of study design and time points.

project ended, to understand the sustainability beyond the project timeframe.

Clinical outcomes

6

Clinical data were collected in both intervention and control hospitals before and after the implementation. Admission books were retrospectively reviewed at each hospital preintervention (2011) and postintervention (2014) to identify children under 5 years of age who were admitted with pneumonia or a lower respiratory tract infection. Data were collected by a paediatric resident or paediatrician familiar with WHO classification of pneumonia and trained to identify signs or symptoms of pneumonia from records. Patients with diagnoses of 'pneumonia', 'cough', 'respiratory distress', 'bronchitis', 'respiratory infection' and 'bronchiolitis' were included in the initial identification of patients but excluded later if (1) the medical record had been lost or (2) the clinical description in the medical record was inconsistent with the WHO-defined pneumonia.

Preintervention we aimed to identify 100 cases of pneumonia/lower respiratory tract infection from each hospital. If 100 cases were not found, data collection continued back to a maximum of 5 years (2007). Postintervention data were collected for cases admitted between January 2012 and December 2013.

A standardised data abstraction form was used to collect patient demographics, diagnosis, pneumonia severity, documentation of SpO_2 , oxygen therapy and patient outcome. Clinical signs and severity of pneumonia documented by the admitting doctor were reviewed by data collectors and investigators (MM/AZG) and if clinical signs indicated a different severity, then pneumonia severity was reclassified based on available data.

The primary outcome measure was the proportion of patients under 5 years of age with pneumonia who stayed in hospital to complete a course of treatment and were discharged well, with no signs of ongoing illness documented in the medical record. This was based on our hypothesis that improving access to oxygen would facilitate treatment completion and reduce early discharge.

Secondary outcomes included: the proportion of children under 5 years of age with pneumonia who were discharged unwell with ongoing signs of illness; the proportion of children who died; and the proportion of children in whom peripheral oxygen saturation (SpO_2) and oxygen therapy was documented.

Statistical analysis

The sample size calculation for clinical outcome data was based on estimates from our baseline assessment that 5% of children with pneumonia are discharged while still unwell, or die in hospital, and an expectation that the oxygen intervention might halve this proportion. A sample size of 906 cases was required in both the intervention and control arms. For other outcome measures, such as the proportion of patients monitored, this estimated sample size was likely to be conservative due to higher prevalence of the measured outcome or larger expected effect size.

Data were entered and stored securely in EpiData databases. Analysis was done using SPSS software (version 16.0). Descriptive analysis of quantitative data was performed with categorical variables described in terms of numbers and percentages and numerical variables described according to their median and range. Tests of proportions were performed on categorical variables preintervention and postintervention using χ^2 .

Ethics

Ethical clearance was granted by the Lao National Ethics Committee for Health Research and the Human Research Ethics Committee of The University of Melbourne in 2011. Written consent was obtained from each participating hospital. Table 1Characteristics of study hospitals in 10intervention and 10 control hospitals in the Lao OxygenProject at baseline

	Hospital cohort			
	Control Median (range)	Intervention Median (range)		
Inpatient beds	16 (10–35)	19 (10–30)		
Admissions in 2010				
Total	610 (442–4573)	1236 (270–6194)		
Adult admissions	383 (350–2042)	1354 (221–3284)		
Paediatric admissions	225 (59–2531)	189 (49–2910)		
Cost per oxygen cylinder (US\$)	23 (14–64)	29 (14–64)		
Total staff numbers				
Doctors	2 (1–27)	2 (0–6)		
Medical assistants	3 (0–20)	2 (0–20)		
Nurses	9 (3–44)	9 (4–24)		

RESULTS

Equipment outcomes

Control and intervention cohorts had similar bed and staff numbers as well as oxygen costs at baseline, while annual admission numbers varied (table 1). The median costs of provision of oxygen to a patient was approximately 200 000 Kip (US\$25) per cylinder.

In the first year following installation 37% (14/38) of concentrators experienced a fault, with a higher proportion in southern compared with northern provinces. (table 2) Seven concentrators (18%) needed sieve beds replaced, and four concentrators (11%) required a new compressor, with other repairs being relatively minor. Following repair, 92% (35/38) of the concentrators were functional. The limiting factor was the quantity of spare parts rather than local engineering capacity. Given the equipment repair requirements at this time a request was made to the manufacturer to replace 10 concentrators with an updated model with improved airflow and compressors and to replenish spare parts. Nine were distributed to hospitals with concentrators requiring major repairs, or those in which problems were anticipated, with one kept in reserve in the central equipment stores.

At 24 months, 18% (7/38) of concentrators required repair including one new model machine, and all but one were repaired. Forty months postimplementation, one-third of concentrators (12/32) required repair, of which 10 (83%) were original machines (table 2). It is useful to understand the cumulative failure rate of concentrators over time; however, due to concentrators being replaced or removed from operation over time, this was difficult. Instead cumulative failure rates can be expressed as a proportion of the number of 'concentratory years of operation' (ie, number of failed concentrators/ total number of concentrators in field × number of years).

 Table 2
 Concentrator repair and maintenance at 12, 24 and 40 months postimplementation of oxygen systems in 10 intervention hospitals and hours of concentrator use in each time interval

		Concentrator repair and maintenance needs						
	Machines supplied, n	12months			24 months		40 months	
Hospital		Number needing repair, n(%)	Working after repair n (%)	Machines replaced n(%)	Number needing repair, n(%)	Working after repair n (%)	Number needing repair, n(%)	Working after repair n (%)
Northern prov	vinces							
H1P2	2	1 (50)	2	0	1	2	1	1
H2P2	3	0 (0)	3	2	0	3	0	3
H1P3	4	2 (50)	4	0	1	4	2	2
H2P3	4	1 (25)	4	0	1	3	1	3
Subtotal	13	4 (31)	13 (100)	2 (15)	3 (23)	12 (92)	4 (31)	9 (69)
Southern prov	vinces							
H1P1	4	2 (50)	3	2	1	4	2	2
H2P1	3	1 (33)	3	1	2	3	0	3
H1P4	4	4 (100)	3	2	1	4	2	2
H2P4	6	2 (33)	6	1	0	6	2	6
H1P5	4	1 (25)	3	1	0	4	2	3
H2P5	4	0 (0)	4	0	0	4	0	4
Subtotal	25	10 (40)	22 (88)	7 (28)	4 (16)	25 (100)	8 (32)	20 (80)
Total	38	14 (37)	35 (92)	9 (24)	7 (18)	37 (97)	12 (32)	29 (76)
Hours of use		13465 (data available from 37/38 machines)			17050 (38/38 machines)		17839 (35/38 machines)	

For the original models, the proportion of machines failing was 0.37, 0.30 and 0.28 at 12, 24 and 40 months, respectively, compared with 0.11 and 0.14 for new model concentrators at 24 and 40 months, respectively.

Concentrators were used for 13465 hours and 17050 hours in the first and second years of the project respectively, with an overall total of around 30000 hours. Accurate measurement of hours past 24 months was difficult due to movement of equipment between hospitals and concentrators taken out of service. There were no adverse incidents, such as a fire, from the use of concentrators during the study period.

All oximeters and sensors were functional 1 year after implementation. After 2 years, 9 of the 10 oximeters in intervention hospitals were still functioning and the majority of the oximeter sensors. By 40 months, 70% of oximeters remained functional, while the proportion of functional sensors varied widely between hospitals from 100% (H2P3) to none (H1P2).

Clinical outcomes

A total of 2912 standardised data abstraction forms were collected from intervention and control hospitals, of which 2777 remained after excluding children greater than 5 years of age and 2722 after excluding other diagnoses. Table 3 outlines the characteristics of childhood pneumonia cases in the preimplementation and postimplementation era from control and intervention hospitals. Patient numbers, age, length of stay and pneumonia severity were similar between hospital cohorts both preintervention and postintervention. There was variation in the proportion of patients contributed by different study hospitals preintervention and postintervention. There was a reduction in the proportion of severe/very severe pneumonia cases in the postintervention era, but the proportions were comparable between intervention and control hospital cohorts.

Preintervention, between group analysis demonstrated a small difference between control and intervention hospitals in the proportion of patients discharged well, with no difference in the proportion discharged with ongoing signs of illness, or who died in hospital, both overall and in the most severe pneumonia cases (table 4). Postintervention, the proportion of all patients with pneumonia discharged home well improved in both the control and intervention hospitals. In patients with severe or very severe pneumonia, the proportional increase in patients discharged well was greater in the intervention compared with control hospital (table 4).

There was a corresponding reduction in the proportion of pneumonia patients discharged home unwell with documented signs of ongoing illness in both control (8.1% (58/713) preintervention and 3.3% (20/606) postintervention, RR 0.41 (95% CI 0.25 to 0.67)) and intervention hospitals (6.2% (44/712) preintervention and 1.9% (13/691) postintervention, RR 0.30 (95% CI 0.16 to 0.56)) (table 4). In the preintervention era, approximately 70% of these patients had documentation of discharge against medical advice, and this fell to 25% of cases postintervention accounting for most of change in patient numbers.

In intervention hospitals, case fatality rates for childhood pneumonia fell from 2.7% (19/712)

Table 3 Admission numbers and characteristics of children under 5 years of age with pneumonia whose medical records were reviewed in preimplementation and postimplementation of oxygen systems in Lao district hospitals

	Contro	ol hospitals n=1319	Intervention hospitals n=1403		
	Pre*		Pre*		
Patient numbers, n	713	606	712	691	
Patient numbers by province n (9	%)				
Champasack	62 (8.7)	83 (13.7)	161 (22.6)	41 (5.9)	
Luang Namtha	146 (20.5)	112 (18.5)	69 (9.7)	78 (11.3)	
Oudomxay	170 (23.8)	81 (13.4)	139 (19.5)	211 (30.5)	
Savannakhet	235 (33.0)	250 (41.3)	144 (20.2)	216 (31.3)	
Sekong	100 (14.0)	80 (13.2)	199 (27.9)	145 (21.0)	
Age in months, median (IQR)	12 (6–24)	15 (7–24)	12 (6–24)	12 (7–25)	
Length of stay, median (IQR)	2 (2–4)	3 (2–4)	2 (1–4)	2 (1–3)	
Pneumonia severity n (%)					
Non-severe pneumonia	372 (52.2)	396 (65.3)	412 (57.9)	423 (62.2)	
Severe/very severe pneumonia	341 (47.8)	210 (34.7)	300 (42.1)	261 (37.8)	

*Refers to the preintervention era with medical records collected between 2007 and 2011. Data collection in hospitals in 2011 stopped once oxygen systems had been installed.

†Refers to the postintervention era with medical records collected between 2012 and 2013.

preintervention to 0.80% (6/691) postintervention (RR 0.33, 95% CI 0.13 to 0.81) with no change in control hospitals (1.7% (12/713) and 2.3% (14/606), RR 1.37, 95% CI 0.64 to 2.95) (table 4). The same trend was seen for the most severe pneumonia cases (table 4).

Improved processes of care were seen in both intervention and control hospitals including increased

documentation of SpO₉ in medical records, with a larger proportional increase in intervention hospitals (figure 2a). In both cohorts, SpO₉ was more frequently documented in the more severe cases. Less improvement was seen in the proportion of children having oxygen therapy and/or oxygen flow rates documented, with minimal difference between hospitals (figure 2b).

	All pneumonia cases			Severe or very severe pneumonia cases			
	Preintervention	Postinterve	ntion	Preintervention	Postinterve	ntion	
Admissions	Ν	N		Ν	N		
Control	713	606		341	210		
Intervention	712	691		300	261		
Well at discharge	n (%)	n (%)	Pre-post comparison*	n (%)	n (%)	Pre-post comparison*	
Control, n (%)	621 (87.1)	588 (92.1)	p=0.003	261 (76.5)	179 (85.2)	p=0.013	
Intervention, n (%)	641 (90.0)	658 (95.2)	p<0.001	243 (81.0)	244 (93.5)	p=0.28	
Between group analysis†	p<0.001			p=0.17			
Died							
Control	12 (1.7)	14 (2.3)	p=0.41	12 (3.5)	14 (6.7)	p=0.09	
Intervention	19 (2.7)	6 (0.8)	p=0.011	19 (6.3)	6 (2.3)	p=0.021	
Between group analysis†	p=0.20			p=0.097			
Discharged unwell/with sig	gns of illness						
Control	58 (8.1)	20 (3.3)	p<0.001	48 (14.1)	9 (4.3)	p<0.001	
Intervention	44 (6.2)	13 (1.9)	p<0.001	31 (10.3)	6 (2.3)	p<0.001	
Between group analysis†	p=0.15			p=0.15			

*Using χ^2 test of proportions to compare preintervention and postintervention data within the hospital cohort. \pm Using χ^2 test of proportions for preintervention (baseline) data between control and intervention groups.

6




Figure 2 Percentage of cases of pneumonia in children less than 5 years of age in control and intervention hospitals in which (A) SpO₂ and (B) oxygen therapy were documented in the medical records, preintervention and postintervention.

DISCUSSION

In 2010, the Lao Ministry of Health identified the lack of affordable oxygen therapy as a major health system challenge and responded with a pilot project to deliver affordable oxygen concentrator systems to district hospitals. The project built on lessons learnt in other settings and an increasing knowledge basis of the types of concentrators and equipment that may best suit different environments.^{9 10} There is evidence of its impact on care and outcomes, with a reduction in case fatality rates for childhood pneumonia in intervention but not control hospitals. However, there are also ongoing challenges, and the engineering investment and health system support required to achieve these outcomes must be recognised.

The project was embedded in the Ministry of Health and implementation decisions were made by a team supported by international partners. Lao clinicians and engineers received hands-on training aimed at building local. The local engineering capacity that was developed and sustained was evident by their ability to address substantial maintenance and repair requirements for concentrators. Despite this, many oxygen concentrators failed, and the impact of the maintenance requirements on feasibility and sustainability of concentrator systems should not be underestimated. Failure rates of the newer model replacement concentrators appeared to be less. Concentrator faults were reported either through the normal health system communication channels or through project evaluation activities. The majority of the equipment issues were only identified through the latter. In the absence of an 'oxygen project', there remains a major risk that equipment falters and is abandoned.

Identifying reasons for equipment failure is essential. The high proportion of the original model concentrators requiring repair likely reflects poor suitability to the environmental conditions. The majority of major faults occurred in southern provinces where temperatures of 40°C degrees are combined with humidity and dust. Laboratory testing of concentrators that occurred after the project began support our observations that the original machines do not function optimally in these conditions. Investigations demonstrated that the AirSep VisionAire was less reliable at high temperature and simultaneous high humidity than the earlier model AirSep Elite.¹⁰ This was attributed to an inadequate internal cooling system in the VisionAire that leads to higher running temperatures and in turn failure of the compressor and sieve beds as we saw in Laos. Despite modifications, the manufacturer no longer recommends the VisionAire for use in tropical conditions.

Environmental conditions will not be the sole explanation for equipment failures. Hospital differences are apparent, even in geographically similar locations. The marked discrepancies in functioning oximeter sensors in some hospitals compared with others suggests local factors such as attention to maintenance and loss are also important. Further exploration of high-performing and low-performing case examples may provide insight into hospital-level influences on the effectiveness of oxygen concentrator systems and medical equipment in general.

The total of $30\,000$ hours over the first 2 years of the project represented only 5% of the potential 665 760 hours the machines were available for use during this time. This in part will be due to relatively low admission numbers of Lao district hospitals. It is may also reflect underuse of oxygen in cases in which it may have been indicated, or equipment downtime from faults.

Despite these problems, this study provides evidence of improved oxygen use and clinical outcomes for children under 5 years of age with pneumonia. Reduction in the proportion of children with pneumonia who were discharged unwell before completing treatment was seen in both intervention and control hospitals, with a corresponding increase in the proportion discharged well. In control hospitals, this may reflect a number of factors including a Hawthorn effect from participation in the study, with increased awareness of the importance of oxygen therapy. It may also reflect access to oximetry to determine completion of treatment, or other secular trends such as improved financial circumstances of the community. Finally, case severity impacts on outcomes and in both the control and intervention groups, there was a reduced proportion of severe and very severe pneumonia in the postintervention era. However, these

changes were comparable between control and intervention hospitals—they do not explain the reduction in the proportion of patients discharged unwell, which was observed among the most severe pneumonia cases, nor do they explain the reduction in case fatality observed only in intervention hospitals. Therefore, although there are many confounding factors to consider, the reduction in pneumonia case fatality documented in intervention hospitals seems likely to be explained, at least in part, by access to improved oxygen systems as seen in previous studies of oxygen implementation.⁶

Improvements in documentation of SpO_2 and oxygen delivery were also seen with a greater degree of improvement in the intervention group. For pulse oximetry, this may reflect improved availability of oximeters to control hospitals, compared with more intense training in their use at intervention hospitals. Perceived relevance of documenting SpO_2 is also likely to be a factor as suggested by higher levels of documentation in more severe cases. Gaps remain between actual and ideal practice. A range of other factors in individual hospitals are likely to contribute to the lack uptake of recommended practices, including the lack of a specific location on monitoring charts to document oximetry and oxygen, and established hospital practices for patient monitoring which require a broader cultural change.

There are limitations in the study design. The inclusion of a control group of hospitals improves on previous assessments of oxygen systems. However, the inability to allocate oxygen systems using an experimental design still limits interpretations of outcomes. Improved clinical outcomes and mortality rates may have been due to confounding factors, as discussed previously, but we are unaware of other interventions that may have affected hospitals differentially. The accuracy and detail of medical records limited our ability to accurately determine pneumonia severity or patient outcomes in some instances. Medical records represented those we could locate; we cannot be sure how many children were admitted with pneumonia in the 20 hospitals in each of the two time periods. Variation between hospitals is evident and, similar to equipment outcomes, changes in clinical practice may be influenced by hospital characteristics that we were not able to explore in detail. For some hospitals, the project provided the first reliable source of hospital oxygen, while for others, it only removed the financial barrier.

Delivery of an affordable, feasible and sustainable oxygen system to prevent hypoxaemia and reduce mortality is a critical intervention in resource-limited settings. It requires a multifaceted approach including robust oxygen equipment, well supported local clinical and engineering capacity and ongoing support for the clinical use of oxygen. However, oxygen concentrator systems can create significant ongoing maintenance demands, which already fragile health systems in low resource settings must be able to support to achieve long-term sustainability, unless more robust equipment solutions can be found. There would be clear benefit from a more detailed understanding of contextual factors influencing equipment longevity and clinical outcomes in order for future oxygen programme to have maximal impact.

Contributors AZG contributed to study design, developed the clinical training materials, assisted with the implementation of oxygen systems, supervised data collection and training of data collectors, analysed the data and was responsible for the first draft of the manuscript. MM contributed to supervision of data collection in the postimplementation era, entered, cleaned and analysed data and contributed to both the first draft and subsequent revisions of the manuscript. TD contributed to study design, planning of implementation and contributed to interpretation of results and revision of the manuscript. DP contributed to study design, provided technical expertise into the engineering aspects of the programme, supervised implementation of equipment and contributed to revision of the manuscript. CW contributed to study design, equipment procurement and implementation of oxygen systems as well as revision of the manuscript. MS contributed to study design, in particular hospital selection and liaison, implementation of oxygen systems, monitoring of equipment in the field and reviewed the manuscript. KS contributed to study design, clinical training in oxygen use, data collection from medical records and supervised other data collectors and reviewed the manuscript. BP contributed to study design, clinical training in oxygen use, data collection from medical records and supervised other data collectors and reviewed the manuscript. KD contributed to study design, clinical training in oxygen use and supervision of data collectors and reviewed the manuscript.

Funding This work was supported by the Japanese Embassy Lao PDR through the WHO Office, who procured the equipment for the project. Additional funding for evaluation came via AusAID funding to the Centre for International Child Health, The University of Melbourne. We thank the manufacturer Airsep for responding generously to our request for replacement machines and spare parts in the second year of the project. The funding sources had no role in the design of the study or the decision to publish this article.

Competing interests None declared.

Ethics approval Human Research Ethics Committee of The University of Melbourne and Lao National Ethics Committee for Health Research.

Provenance and peer review Not commissioned; externally peer reviewed.

Data sharing statement On request we are happy to share the data pertaining to this study. Requests should be made by email to the corresponding author.

Open Access This is an Open Access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: http://creativecommons.org/ licenses/by-nc/4.0/

© Article author(s) (or their employer(s) unless otherwise stated in the text of the article) 2017. All rights reserved. No commercial use is permitted unless otherwise expressly granted.

REFERENCES

- Subhi R, Adamson M, Campbell H, et al. The prevalence of hypoxaemia among ill children in developing countries: a systematic review. Lancet Infect Dis 2009;9:219–27.
- Black RE, Cousens S, Johnson HL, et al. Global, regional, and national causes of child mortality in 2008: a systematic analysis. Lancet 2010;375:1969–87.
- Wandi F, Peel D, Duke T. Hypoxaemia among children in rural hospitals in Papua New Guinea: epidemiology and resource availability--a study to support a national oxygen programme. *Ann Trop Paediatr* 2006;26:277–84.
- Usen S, Weber M, Mulholland K, et al. Clinical predictors of hypoxaemia in Gambian children with acute lower respiratory tract infection: prospective cohort study. BMJ 1999;318:86–91.
- Weber MW, Usen S, Palmer A, *et al.* Predictors of hypoxaemia in hospital admissions with acute lower respiratory tract infection in a developing country. *Arch Dis Child* 1997;76:310–4.
- Duke T, Wandi F, Jonathan M, et al. Improved oxygen systems for childhood pneumonia: a multihospital effectiveness study in Papua New Guinea. Lancet 2008;372:1328–33.
- Duke T, Peel D, Graham S, *et al.* Oxygen concentrators: a practical guide for clinicians and technicians in developing countries. *Ann Trop Paediatr* 2010;30:87–101.

6

Open Access

- La Vincente SF, Peel D, Carai S, *et al.* The functioning of oxygen concentrators in resource-limited settings: a situation assessment in two countries. *Int J Tuberc Lung Dis* 2011;15:693–9.
 Mardi E, indext and E, indext and E. Statistical Statistics and Statistis and Statistics and Statistics and Statistics and Statistics
- Matai S, Peel D, Wandi F, et al. Implementing an oxygen programme in hospitals in Papua New Guinea. *Ann Trop Paediatr* 2008;28:71–8.
 Peel D, Neighbour R, Eltringham RJ. Evaluation of oxygen
- Peel D, Neighbour R, Eltringham RJ. Evaluation of oxygen concentrators for use in countries with limited resources. *Anaesthesia* 2013;68:706–12.
- Duke TD, ed. The clinical use of oxygen in hospitals with limited resources; guidelines for health-care workers, hospital engineers and managers.(internet). Geneva: WHO, 2011. http://video.rch.org.au/ cich/The_Clinical_Use_of_Oxygen_November_2011.pdf (accessed 29 March 2016).
- 12. World Health Organisation (WHO). WHO pocketbook of Hospital Care for Children. 1st ed. Geneva: WHO Press, 2005.

University Library



A gateway to Melbourne's research publications

Minerva Access is the Institutional Repository of The University of Melbourne

Author/s:

Gray, AZ; Morpeth, M; Duke, T; Peel, D; Winter, C; Satvady, M; Sisouk, K; Prasithideth, B; Detleuxay, K

Title:

Improved oxygen systems in district hospitals in Lao PDR: a prospective field trial of the impact on outcomes for childhood pneumonia and equipment sustainability

Date:

2017-08-01

Citation:

Gray, A. Z., Morpeth, M., Duke, T., Peel, D., Winter, C., Satvady, M., Sisouk, K., Prasithideth, B. & Detleuxay, K. (2017). Improved oxygen systems in district hospitals in Lao PDR: a prospective field trial of the impact on outcomes for childhood pneumonia and equipment sustainability. BMJ PAEDIATRICS OPEN, 1 (1), https://doi.org/10.1136/bmjpo-2017-000083.

Persistent Link: http://hdl.handle.net/11343/208019

File Description: Published version License: CC BY-NC