



## Developing an automated functional cardiotoxicity drug screening platform

Christensen, Rie Kjær; Skafte-Pedersen, Peder; Wilson, Sandra; Larsen, Niels Bent

*Publication date:*  
2018

*Document Version*  
Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

*Citation (APA):*  
Christensen, R. K., Skafte-Pedersen, P., Wilson, S., & Larsen, N. B. (2018). Developing an automated functional cardiotoxicity drug screening platform. Poster session presented at Organ-on-a-Chip & Tissue-on-a-Chip Europe 2018, Rotterdam, Netherlands.

---

### General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

# Developing an automated functional cardiotoxicity drug screening platform


Rie Kjær Christensen<sup>1,2</sup>, Peder Skafte-Pedersen<sup>1</sup>, Sandra Wilson<sup>1</sup>, Niels Bent Larsen<sup>2</sup>

## INTRODUCTION

Cardiotoxicity is a major challenge in drug development. Current cardiotoxicity testing relies heavily on ion channel measurements using patch clamp analysis on heart cells (cardiomyocytes). An automated functional cardiotoxicity drug testing method that extends beyond ion channel analysis has not yet been developed, but is required, as many cardiotoxic events do not correlate with ion channel behavior.

The key function in evaluating cardiac toxicity is cardiac mini-tissue contraction. The project aim is to develop an automated analysis platform with contracting mini-tissues in a multi-well format connected to a read-out system. A micro-molded platform manufactured using state-of-the-art 3D microprinting has shown proof of concept with model cells (mouse fibroblasts and myoblasts) cultured for up to 4 weeks.


## PROCESS



**Fig. 1: Manufacturing process of the drug screening platform**

(a) Induced pluripotent stem cells (iPSC) are grown in 2D and differentiated into cardiac progenitor cells in 0-15 days. (b) Cardiac progenitor cells are seeded onto the drug screening platform. (c) A stereolithographic method is used to produce the platform for cardiomyocyte cell culturing providing a fast and easy manufacturing process, [1],[2]. (d) After seeding cardiac progenitor cells onto the platform mini-tissues are formed and culturing is controlled to mature the mini-tissues into mature cardiomyocyte mini-tissues in day 15-35. (e) Drug screening platform with mature cardiomyocyte mini-tissues ready for drug screening protocols.

## FUNCTIONALITY



**Fig. 2: Drug screening analysis platform**

(a) An automated drug screening analysis platform to use for evaluating cardiotoxicity in a multiple construct configuration. By culturing in mechanically relevant conditions [2] it is possible to obtain a mature state of the cardiomyocyte that optimize the peak force generation. (b) The mini-tissues function independently (c) and will contract simultaneously based on an induced stimulation.

## BACKGROUND FOR DEVELOPMENT

A lot of effort is going into drug development which is both time- and cost-intensive and robust methods to test cardiotoxic effects are still missing. In recent years development of a contractile mini-tissue in different shapes, configurations, sizes and dimensions have been investigated [3], [4], [5], [6], [7] as the contractile force is a good indicator of cardiomyocyte mini-tissue function. There is a tight correlation between culturing methods to reach maturity of the cardiomyocytes, stiffness, and contractile behavior, all of which has been considered in the presented drug screening platform.

**Contact:** Rie K. Christensen, rie.christensen@sophion.com.

<sup>1</sup>Sophion Bioscience A/S, Baltorpvej 154 2750 Ballerup, Denmark.

<sup>2</sup>Dept. of Micro- and Nanotechnology, DTU Nanotech, Technical University of Denmark, 2800 Lyngby, Denmark

## REFERENCES

- [1] Billiet, T.; Vandenhaute, M.; Schelfhout, J.; Van Vlierberghe, S.; Dubrue, P. *Biomaterials* 2012, 33 (26), 6020–6041 DOI: 10.1016/j.biomaterials.2012.04.050.
- [2] R. Zhang and N. B. Larsen, *Lab Chip*, 2017, 17, 4273–4282.
- [3] A. W. Feinberg, P. W. Alford, H. Jin, C. M. Ripplinger, A. A. Werdich, S. P. Sheehy, A. Grosberg and K. K. Parker, *Biomaterials*, 2012, 33, 5732–41.
- [4] W. R. Legant, A. Pathak, M. T. Yang, V. S. Deshpande, R. M. McMeeking and C. S. Chen, *Proc. Natl. Acad. Sci. U. S. A.*, 2009, 106, 10097–102.
- [5] N. Thavandiran, N. Dubois, A. Mikryukov, S. Massé, B. Beca, C. A. Simmons, V. S. Deshpande, J. P. McGarry, C. S. Chen, K. Nanthakumar, G. M. Keller, M. Radisic and P. W. Zandstra, *Proc. Natl. Acad. Sci. U. S. A.*, 2013, 110, E4698-707.
- [6] A. Eder, I. Vollert, A. Hansen and T. Eschenhagen, *Adv. Drug Deliv. Rev.*, 2016, 96, 214–224.
- [7] V. Y. Sidorov, P. C. Samson, T. N. Sidorova, J. M. Davidson, C. C. Lim and J. P. Wikswo, *Acta Biomater.*, 2017, 48, 68–78.