

Novel penalised likelihood reconstruction of PET in the assessment of histologically verified small pulmonary nodules

Eugene J. Teoh^{1,2} · Daniel R. McGowan^{2,3} · Kevin M. Bradley¹ · Elizabeth Belcher⁴ · Edward Black⁴ · Fergus V. Gleeson^{1,2}

Received: 23 December 2014 / Revised: 20 April 2015 / Accepted: 28 April 2015 / Published online: 20 May 2015
© The Author(s) 2015. This article is published with open access at Springerlink.com

Abstract

Objectives Investigate the effect of a novel Bayesian penalised likelihood (BPL) reconstruction algorithm on analysis of pulmonary nodules examined with 18F-FDG PET/CT, and to determine its effect on small, sub-10-mm nodules.

Methods 18F-FDG PET/CTs performed for nodule evaluation in 104 patients (121 nodules) were retrospectively reconstructed using the new algorithm, and compared to time-of-flight ordered subset expectation maximisation (OSEM) reconstruction. Nodule and background parameters were analysed semi-quantitatively and visually.

Results BPL compared to OSEM resulted in statistically significant increases in nodule SUV_{max} (mean 5.3 to 8.1, $p < 0.00001$), signal-to-background (mean 3.6 to 5.3, $p < 0.00001$) and signal-to-noise (mean 24 to 41, $p < 0.00001$). Mean percentage increase in SUV_{max} ($\% \Delta \text{SUV}_{\text{max}}$) was significantly higher in nodules ≤ 10 mm ($n = 31$, mean 73 %) compared to > 10 mm ($n = 90$, mean 42 %) ($p = 0.025$). Increase in signal-to-noise was higher in nodules ≤ 10 mm (224 %, mean 12 to 27) compared to > 10 mm (165 %, mean 28 to 46). When applying optimum

SUV_{max} thresholds for detecting malignancy, the sensitivity and accuracy increased using BPL, with the greatest improvements in nodules ≤ 10 mm.

Conclusion BPL results in a significant increase in signal-to-background and signal-to-noise compared to OSEM. When semi-quantitative analyses to diagnose malignancy are applied, higher SUV_{max} thresholds may be warranted owing to the SUV_{max} increase compared to OSEM.

Key Points

- Novel Bayesian penalised likelihood PET reconstruction was applied for lung nodule evaluation.
- This was compared to current standard of care OSEM reconstruction.
- The novel reconstruction generated significant increases in lung nodule signal-to-background and signal-to-noise.
- These increases were highest in small, sub-10-mm pulmonary nodules.
- Higher SUV_{max} thresholds may be warranted when using semi-quantitative analyses to diagnose malignancy.

Keywords Solitary pulmonary nodule · Positron-emission tomography · Image reconstruction · Signal-to-noise ratio · Image quality enhancement

✉ Eugene J. Teoh
eugene.teoh@oncology.ox.ac.uk

¹ Department of Radiology, Churchill Hospital, Oxford University Hospitals NHS Trust, Old Road, Headington, Oxford OX3 7LE, UK

² Department of Oncology, University of Oxford, Oxford OX3 7DQ, UK

³ Radiation Physics and Protection, Churchill Hospital, Oxford University Hospitals NHS Trust, Old Road, Oxford OX3 7LE, UK

⁴ Department of Thoracic Surgery, John Radcliffe Hospital, Oxford University Hospitals NHS Trust, Headley Way, Oxford OX3 7DU, UK

Introduction

18F-FDG PET/CT is now commonly performed in the assessment of potentially malignant solitary pulmonary nodules [1, 2], but its use is limited in the evaluation of small pulmonary nodules [3, 4], with a paucity of studies evaluating its value in small, sub-10-mm pulmonary nodules [1, 4].

The use of semi-quantitative measurements in the evaluation of small pulmonary nodules is also uncertain, with no agreed standardised uptake value (SUV_{max}) threshold for

differentiating benign from malignant aetiologies [1, 5–7]. Signal acquisition and technical factors such as the image reconstruction methodologies used have a greater impact on the accuracy of SUV measurements in small nodules compared to large nodules [8, 9].

There are two main types of PET reconstruction available: analytical and iterative. Generally iterative methods are more commonly used because of their improved signal-to-noise ratios compared to analytical algorithms, such as filtered back projection (FBP) [8–10]. Iterative methods also allow the modelling of various system factors that can be included in the reconstruction, such as the point spread function (PSF) [11–13]. The main iterative algorithm used is ordered subset expectation maximisation (OSEM) [14] which aims to find the most likely image from the raw data through repeated iterations of possibilities. Each iteration gives an image with a greater likelihood of describing the measured data. The main disadvantage of iterative algorithms is that it is not possible to run them to full convergence as the image noise increases with each iteration, becoming unacceptable before full convergence is reached [8, 10]. In general, OSEM is stopped after a predetermined number of iterations resulting in an underconverged image. The failure to accurately attribute the signal results in an underestimation of SUVs, and has the greatest effect in small FDG-avid foci such as sub-10-mm nodules.

Recently a new iterative PET reconstruction algorithm, Bayesian penalised likelihood (BPL), has been developed by GE Healthcare, called Q.Clear (GE Healthcare, Milwaukee, USA) and includes PSF modelling [15]. Whilst penalised likelihood algorithms were first reported in 1987 [16], and their advantage over FBP was described in 1996 [17], their clinical use has so far been very limited. The BPL considered here includes a relative difference penalty, first introduced in 2002 [18]. The main advantage this has over other possible penalties is that it is a function of the difference between neighbouring voxels as well as a function of their sum [19]. This penalty function acts as a noise suppression term, which allows an increased number of iterations without the noise usually seen in OSEM [15]. The strength of this penalty function is controlled by the penalisation factor (beta), which is the only user-input variable to the algorithm. Modified block sequential regularized expectation maximization (BSREM) is used as an optimiser for this BPL algorithm, which, as a result of the penalty function, allows effective convergence to be achieved in images, potentially providing a more accurate SUV [19, 20].

The aim of this study was to investigate the effect of the iterative reconstruction technique using BPL on the analysis of nodules scanned with PET/CT, and to determine its effect on small, sub-10-mm nodules.

Materials and methods

Patient selection

All patients who underwent 18F-FDG PET/CT at our institution between November 2010 and December 2013 for the evaluation of pulmonary nodules, with subsequent histological diagnosis, were retrospectively identified. Nodules were defined as having a long-axis diameter of ≤ 30 mm on lung windows. Informed consent is not required for retrospective reviews of this nature in our institution.

18F-FDG PET/CT imaging protocol

PET/CT examinations were performed on a 3D mode time of flight (ToF) GE Discovery 690 PET/CT system (GE Healthcare, Milwaukee, USA). The patients fasted for at least 6 h prior to their scan. Their blood glucose was measured prior to intravenous injection, with 4 MBq/kg of 18F-FDG. Imaging commenced 90 min post-injection (93 ± 7 min) and covered the skull base to upper thighs. The PET/CT images were acquired under normal tidal respiration for 4 min per bed position. The CT was performed using a pitch of 0.984, 120 kV, auto mA with a noise index of 25.

PET reconstructions

PET images were reconstructed using two different algorithms both of which used the CT for attenuation correction and the same normalisation correction factors. The standard of care PET reconstruction algorithm used is ToF OSEM (VPFX, GE Healthcare, Milwaukee, USA). This was used with two iterations, 24 subsets and 6.4 mm Gaussian filter in our institution. The sinograms generated at the time of scanning were retrospectively processed using the new ToF BPL reconstruction algorithm (Q.Clear, GE Healthcare, Milwaukee, USA) using a penalisation factor (beta) of 400, the only user-input variable for this algorithm.

Imaging analysis

Semi-quantitative analysis

The pre-existing PET images (reconstructed using OSEM) and new PET images reconstructed using BPL were fused with the CT component of the original study for analysis. The SUV_{max} of each nodule was recorded using a standard volume of interest (VOI) tool. Background SUVs were measured in the right lobe of liver and descending aorta at the level of the carina, with 3.0-cm and 1.0-cm diameter spherical VOIs, respectively. SUV_{max} , SUV_{mean} and standard deviation within the VOI were recorded for both reference organs. Signal-to-background ratio (SBR) for each nodule was

calculated as nodule SUV_{max} divided by descending aorta SUV_{mean} . Nodule signal-to-noise ratio (SNR) was defined as nodule SUV_{max} divided by liver SUV_{sd} using the standard deviation on a liver reference VOI (SUV_{sd}) as the measure of noise. To identify possible change in image noise the background SNR was calculated as liver SUV_{mean} divided by liver SUV_{sd} .

Visual analysis

Visual analysis of the OSEM and BPL PET/CT images was performed by a radiology resident with 4 years of radiology (including 1 year of PET/CT) experience. Nodules were scored according to degree of FDG uptake (above-background, at-background or below-background). The reference organ for background uptake was the descending aorta. The scorer reviewed the cases in a randomised and unpaired order, blinded to the clinical details.

Statistical analysis

Statistical analyses were performed using IBM SPSS Statistics 20.0 (IBM Corporation, New York, USA). P values ≤ 0.05 were considered as statistically significant.

Background analysis

This was performed across the entire cohort. Differences in background SUV_{mean} , SUV_{max} and SNR across the entire cohort were analysed using paired t tests.

Nodule analysis

Nodules were classified according to three different categories for analysis:

1. Size: ≤ 10 mm or > 10 mm in long-axis diameter on lung windows
2. FDG uptake: FDG-positive (above-background) or FDG-negative (at/below-background). Background uptake was patient-specific and set as the descending aorta SUV_{mean} on the OSEM algorithm.
 - a. A sub-analysis of FDG-positive nodules was performed according to size
3. Aetiology

For the first two categories, differences in nodule SUV_{max} , SBR and SNR were analysed using Wilcoxon rank-sum tests. The percentage difference in nodule SUV_{max} ($\% \Delta SUV_{max}$) was also calculated. Differences in $\% \Delta SUV_{max}$ within each category were analysed using the Mann–Whitney U test.

Across the entire cohort, and within the FDG uptake classification, correlation between $\% \Delta SUV_{max}$ and nodule size was analysed using Spearman's rank correlation coefficient.

Within the aetiological classification, differences in $\% \Delta SUV_{max}$ and nodule size were analysed according to three categories: primary lung cancer, metastases and benign aetiologies, using the Kruskal–Wallis test.

Diagnostic performance

The performance of both algorithms to detect malignant nodules (primary lung cancer and metastases) was assessed using both semi-quantitative and visual criteria. For semi-quantitative criteria, receiver operating characteristic (ROC) curves were plotted, and area under the curve (AUC) values calculated. The areas under both ROC curves were compared using the method described by Hanley and McNeil [21]. The optimal SUV threshold for the diagnosis of malignancy was defined as the point on the curve closest to the upper left corner of the ROC space. Sensitivity, specificity and accuracy for malignancy detection were calculated for these thresholds. For visual criteria, nodules scored as above-background were designated malignant and nodules at-background or below-background were designated benign. Sensitivity, specificity and accuracy for malignancy detection were then calculated.

Results

Clinical characteristics

One hundred and four patients (45 male, 59 female, mean age 68 years, range 23–89 years) met the inclusion criteria. A total of 121 nodules were included for analysis, comprising various malignant ($n=106$) and benign ($n=15$) aetiologies. Malignant nodules comprised of non-small cell lung cancer (NSCLC) ($n=64$), metastases of extra-thoracic malignancy ($n=32$) and carcinoid ($n=10$). Benign aetiologies included hamartoma, sarcoidosis and mycobacterial granulomata, benign alveolar adenoma and inflammatory scar. The mean nodule diameter was 15 mm (range 3–28 mm) with 31 nodules ≤ 10 mm and 90 nodules > 10 mm.

Background analysis

Differences in all the background SUV parameters between OSEM and BPL were statistically significant (Table 1). The largest difference was in liver SUV_{mean} (mean difference 0.17, 95 % CI 0.14–0.21). The average background SNR on OSEM was 10.2 (range 6.9–15.0), increasing to 12.3 on BPL (range 8.0–17.9, $p < 0.0001$) (Table 1).

Table 1 Background analysis

Parameter	OSEM Mean (range)	BPL Mean (range)	Paired <i>t</i> test (BPL vs OSEM)	
			Mean difference (95 % CI)	<i>p</i> value
Liver SUV _{mean}	2.3 (1.4–3.2)	2.5 (1.3–3.7)	0.17 (0.14–0.21)	<0.001
Liver SUV _{max}	3.1 (1.6–4.5)	3.2 (1.6–4.7)	0.07 (0.02–0.12)	0.006
Liver SUV _{peak}	2.6 (1.4–3.6)	2.7 (1.4–4.1)	0.16 (0.12–0.20)	<0.001
Liver SUV _{sd}	0.23 (0.09–0.38)	0.21 (0.09–0.32)	–0.027 (0.020–0.034)	<0.001
D. aorta SUV _{mean}	1.5 (0.8–2.5)	1.6 (0.8–2.4)	0.07 (0.05–0.10)	<0.001
D. aorta SUV _{max}	1.8 (1.0–3.1)	1.8 (0.9–2.7)	0.04 (0.01–0.07)	0.016
SNR	10.2 (6.9–15.0)	12.3 (8.0–17.9)	2.1 (1.7–2.4)	<0.0001

Nodule analysis: entire cohort

There was a statistically significant difference in overall nodule SUV_{max} (mean difference 2.8, $p < 0.00001$), SNR (mean 17.0 on OSEM almost doubling to 41.1 on BPL, $p < 0.0001$) and SBR (mean difference 1.7, $p < 0.00001$), with mean increase $\% \Delta \text{SUV}_{\text{max}}$ of 50.0 % (median 40.5 %, range –16.7 to +301 %). There was a statistically significant inverse correlation between $\% \Delta \text{SUV}_{\text{max}}$ and nodule size ($p = 0.0028$). Results of this analysis are described in Table 2.

Nodule analysis: size

Consistent with analysis of the entire cohort, there were statistically significant differences between BPL and OSEM in nodule SUV_{max}, SBR and SNR in both ≤ 10 mm and > 10 mm categories. Results of this analysis are detailed in Table 2.

Nodule analysis: FDG uptake

The same trends in change to SUV_{max}, SBR and SNR were observed in FDG-positive (above-background) nodules (all $p < 0.00001$). There was also a stronger correlation between $\% \Delta \text{SUV}_{\text{max}}$ and nodule size in this group compared to the entire cohort ($p < 0.00001$ vs 0.0028). However in FDG-negative (at/below-background) nodules, statistical significance of SUV_{max} and SBR change was relatively weaker ($p = 0.031$ and 0.0076, respectively). There was no statistically significant change to SBR or $\% \Delta \text{SUV}_{\text{max}}$ in FDG-negative nodules ($p = 0.196$ and 0.522, respectively). Results of this analysis are summarised in Table 2 and Fig. 1.

In view of the stronger correlation between $\% \Delta \text{SUV}_{\text{max}}$ and nodule size in FDG-positive nodules compared to the entire cohort, a sub-analysis was performed according to size. The difference in $\% \Delta \text{SUV}_{\text{max}}$ between FDG-positive nodules ≤ 10 mm and > 10 mm was accentuated in this subgroup (92.1 % vs 45.3 %, $p = 0.00013$).

Nodule analysis: aetiology

When categorised into three groups, NSCLC, metastases and benign, there was an observed difference in $\% \Delta \text{SUV}_{\text{max}}$ between the groups (mean 65.2 % in metastases, 46.6 % in NSCLC, 35.0 % in benign), but this was not statistically significant. The highest $\% \Delta \text{SUV}_{\text{max}}$ in metastases could not be accounted for by differences in nodule size, as there was no significant difference in nodule size among the three groups (Table 3). The OSEM SUV_{max}, BPL SUV_{max} and $\% \Delta \text{SUV}_{\text{max}}$ of the three groups are summarised in Table 3.

Visual analysis of FDG uptake

On visual analysis of FDG uptake on OSEM compared to BPL (Table 4), scores were concordant in 113 nodules (93 %). Seven nodules had a higher score on BPL (six background increased to above-background, of which four were malignant and two benign; one below-background increased to background–malignant), and one nodule had a lower score on BPL (above-background decreased to below-background–malignant).

Four of the seven nodules that scored higher on BPL using visual analysis were ≤ 10 mm. By semi-quantitative criteria, three of these seven nodules were FDG-negative on OSEM, one remaining FDG-negative on BPL by semi-quantitative criteria. The latter nodule was scored as below-background on OSEM, and background on BPL. The two benign nodules in this group were FDG-positive by semi-quantitative criteria on both algorithms. These results are summarised in Table 4.

Diagnostic performance

ROC curves were plotted to evaluate the usefulness of BPL and OSEM to detect malignant nodules on the basis of semi-quantitative criteria (Fig. 2). The AUC values were 0.719 ($p = 0.009$) and 0.709 ($p = 0.006$), respectively, with no statistically significant difference between the two algorithms ($p = 0.98$).

Table 2 Summary of SUV_{max} , SNR, SBR and percentage difference in SUV_{max} across the entire cohort, classified according to size (≤ 10 mm and >10 mm) and FDG uptake (FDG-positive and FDG-negative)

	All ($n=121$)		≤ 10 mm ($n=31$)		>10 mm ($n=90$)		FDG-positive ($n=105$)		FDG-negative ($n=16$)	
	OSEM	BPL	OSEM	BPL	OSEM	BPL	OSEM	BPL	OSEM	BPL
SUV_{max}										
Mean	5.3	8.1	2.8	5.3	6.2	9.0	5.9	9.1	1.2	1.3
Median	4.1	5.8	2.4	4.0	5.1	7.7	4.6	7.6	1.2	1.2
Range	0.5–19.4	0.5–30.7	0.5–5.7	0.5–19.3	1.0–19.4	1.0–30.7	1.4–19.4	1.3–30.7	0.5–1.7	0.5–2.0
Wilcoxon p value	<0.00001		<0.00001		<0.00001		<0.00001		0.031	
SNR										
Mean	23.8	41.1	12.2	27.3	27.7	45.8	26.6	46.4	4.9	6.2
Median	17.0	29.8	9.5	20.4	20.5	35.0	19.9	35.7	4.7	5.6
Range	1.3–104.2	1.9–155.1	1.3–35.6	1.9–101	4.1–104	4.2–155	5.3–104	5.5–155	1.3–8.1	1.9–11.2
Wilcoxon p value	<0.00001		0.00019		<0.00001		<0.00001		0.0076	
SBR										
Mean	3.6	5.3	1.9	3.4	4.1	6.0	4.0	6.0	0.7	0.8
Median	2.6	3.8	1.7	2.5	3.2	4.6	2.9	4.9	0.8	0.8
Range	0.3–14.3	0.3–26.2	0.3–5.2	0.3–13.1	0.6–14.3	0.6–26.2	1.0–14.3	0.9–26.2	0.3–1.0	0.3–1.3
Wilcoxon p value	<0.00001		0.00005		<0.00001		<0.00001		0.196	
%ΔSUV_{max}										
Mean	50.0 %		73.3 %		42.0 %		56.0 %		10.9 %	
Median	40.5 %		61.6 %		37.9 %		43.3 %		7.2 %	
Range	-16.7 to 301 %		-13.7 to 301 %		-16.7 to 156 %		-12.7 to 301 %		-16.7 to 49.5 %	
Spearman p value (correlation vs size)	0.0028		-		-		<0.00001		0.522	
Mann–Whitney U p value	-		0.025		-		<0.00001		-	

The optimum SUV_{max} threshold for detection of malignancy was 3.5 and 4.4 for OSEM and BPL, respectively. The sensitivities, specificities and accuracies at these thresholds as an entire cohort and categorised according to size are summarised in Table 5. Across these groups, there was an increase in accuracy and a divergence in sensitivity and

specificity between OSEM and BPL, with sensitivity tending to increase and specificity tending to decrease with BPL. The greatest increases in accuracy (48 % to 58 %) and sensitivity (44 % to 56 %) were demonstrated in nodules ≤ 10 mm.

On the basis of visual criteria, sensitivity and accuracy was consistently higher across the cohort compared to semi-


**Fig. 1** Scatter plot of % ΔSUV_{max} (relative change of SUV_{max} from OSEM to BPL) against nodule size according to FDG-positivity

Table 3 SUV_{max}, %ΔSUV_{max} and size of NSCLC, metastases and benign nodules

	NSCLC (n=74)		Metastases (n=32)		Benign (n=15)	
	OSEM	BPL	OSEM	BPL	OSEM	BPL
SUV _{max}						
Mean	6.5	9.5	3.7	6.7	2.9	4.0
Median	5.5	8.8	2.9	4.7	2.5	3.8
Range	1.0–19.4	1.1–30.7	0.5–12.1	0.5–26.1	1.0–6.3	1.0–9.1
%ΔSUV _{max}						
Mean	46.6 %		65.1 %		35.0 %	
Median	39.8 %		48.6 %		39.7 %	
Range	–16.7 to 190 %		–13.7 to 301 %		0–85 %	
Kruskall–Wallis <i>p</i> value	0.393					
Size (mm)						
Mean	15		17		14	
Median	13.5		16.5		13	
Range	3–28		5–28		8–24	
Kruskall–Wallis <i>p</i> value	0.131					

quantitative criteria, although specificity was lower (Table 5). The differences in accuracy and sensitivity between OSEM and BPL were relatively smaller, if not unchanged (Table 5). The largest change in sensitivity was by 3 % (84 % to 87 % in the entire cohort), and 2 % in accuracy (84 % to 82 % in nodules >10 mm).

Discussion

Overall there was an almost twofold increase in nodule SNR across the cohort. Significant increments in SUV_{max} and SBR in FDG-positive nodules (SUV_{max}>blood pool SUV_{mean}) were also observed with an average SUV_{max} increase of 3.2 (56 %) and SBR increment of 2.0, while changes in

background SUV parameters were minimal (mean difference ≤0.17, Table 1). These changes are best illustrated by two example cases in Figs. 3 and 4. There was no significant increase in SUV_{max} and SBR in FDG-negative nodules. There was an inverse correlation between nodule size and %ΔSUV_{max}: for nodules ≤10 mm %ΔSUV_{max} was 92.1 %, compared to nodules >10 mm (%ΔSUV_{max} 45.3 %).

This overall observation is considered to be due to BPL running to ‘effective convergence’ [19, 20], compared to the limited convergence of OSEM (reported here using two iterations). The greater differences in SUV_{max} in nodules ≤10 mm would also be due to BPL including PSF modelling in the reconstruction process, giving higher values for small lesions compared to reconstructions that do not include PSF modelling, with SUV_{max} approaching phantom truth [11, 13].

Table 4 Results of visual analysis of FDG uptake compared to semi-quantitative criteria

	OSEM	BPL
Visual		
FDG-positive (above-background)	99	104
FDG-negative (at/below-background)	22	17
Semi-quantitative		
FDG-positive (above-background)	105	106
FDG-negative (at/below-background)	16	15
	Visual	Semi-quantitative
Concordant	113 (93 %)	116 (96 %)
Increased score /	6	3
became FDG-positive	(4 malignant, 2 benign)	(2 malignant, 1 benign)
Decreased score /	1	2
became FDG-negative	(malignant)	(malignant)




Fig. 2 ROC curves for evaluation of pulmonary nodules on OSEM and BPL based on SUV_{max} as a single determinant of malignant involvement

Interestingly, the greatest increases were seen in small malignant nodules. While comparison between OSEM with PSF modelling (SharpIR on GE systems) and BPL might seem more appropriate, the former reconstruction has not been adopted as standard of care in our institution. This is due to the increased intervoxel covariance seen with the PSF modelling (Fig. 2 in [22]), which causes images to appear very heterogeneous, for example in the liver.

Table 5 Diagnostic performance of OSEM and BPL in detecting malignant nodules on the basis of semi-quantitative analysis using optimum SUV_{max} threshold (3.5 and 4.4, respectively) and visual analysis

	Semi-quantitative		Visual	
	OSEM	BPL	OSEM	BPL
All (<i>n</i> =121)				
Sensitivity	60.4 %	67.0 %	84.0 %	86.8 %
Specificity	73.3 %	66.7 %	33.3 %	20.0 %
Accuracy	62.0 %	67.0 %	77.7 %	78.5 %
≤10 mm (<i>n</i> =31)				
Sensitivity	44.4 %	55.6 %	100.0 %	100.0 %
Specificity	75.0 %	75.0 %	69.2 %	60.0 %
Accuracy	48.4 %	58.1 %	87.1 %	87.1 %
>10 mm (<i>n</i> =90)				
Sensitivity	65.8 %	70.9 %	89.9 %	89.9 %
Specificity	72.7 %	63.6 %	45.5 %	27.3 %
Accuracy	66.7 %	70.0 %	84.4 %	82.2 %





Fig. 3 Coronal PET and PET/CT images demonstrating an 11-mm left upper lobe adenocarcinoma which had an SUV_{max} of 2.1 on OSEM, and 4.6 on BPL. Liver and descending aorta SUV_{mean} differed by 0.1 between the two reconstructions. SNR increased by more than twofold from 15 to 35. All PET images are displayed on SUV scale 0–6

The effect of BPL on nodule evaluation and diagnostic performance was assessed using semi-quantitative and visual criteria. Improvements in sensitivity and accuracy were observed with BPL when semi-quantitative criteria were applied. This was accompanied by decreases in specificity which were expected, considering there was no significant difference in $\% \Delta SUV_{max}$ between benign and malignant nodules and AUC values using either algorithm. Remarkably, sensitivity and accuracy were consistently higher using visual criteria compared to semi-quantitative criteria (e.g. accuracy 62–67 % on semi-quantitative criteria compared to 78–79 % on visual criteria), and the differences of these parameters between OSEM and BPL were less varied using visual criteria. The specificities were generally lower for BPL, although this may be accounted for by the use of FDG uptake as a single determinant of malignancy. It is anticipated that the incorporation of other factors such as clinical history, risk, prior imaging and CT appearance would improve specificity alongside the other diagnostic performance parameters.

The two most used methods to analyse 18F-FDG PET data for distinguishing benign from malignant lung

Fig. 4 Axial PET and PET/CT images of an 8-mm right upper lobe metastasis from colorectal adenocarcinoma which had an SUV_{max} of 1.8 on OSEM, and 4.4 on BPL. Liver and descending aorta SUV_{mean} differed by up to 0.28. Signal-to-noise ratio increased by more than twofold from 8 to 20. All PET images are displayed on SUV scale 0–6



nodules are either by subjectively assessing FDG avidity in comparison to background or using semi-quantitative analyses, SUV, most commonly SUV_{max} , with variable thresholds used as a cut-off to separate benign from malignant disease. In their widely cited meta-analysis, Gould et al. demonstrated that semi-quantitative analysis did not improve the accuracy of 18F-FDG PET in the diagnosis of pulmonary nodules in comparison to subjective visual analysis [1], and our observations are generally consistent with this finding. Other groups have also demonstrated visual analysis of 18F-FDG PET to be useful [7, 23]. Potentially, when either a semi-quantitative method or subjective visual analysis is used to differentiate benign from malignant disease, a higher threshold to report a nodule as benign may be appropriate.

Factors which influence SUV, and in turn affect semi-quantitative analysis, include blood glucose levels, radiotracer uptake times and respiratory movement [24]; the last of these being of particular relevance to lung nodules. Our study suggests that BPL does not improve the differentiation of benign from malignant nodules when compared to conventional OSEM analysis. However, this may be because we did not include enough benign nodules to enable a separation of nodules with borderline FDG avidity to show that BPL increases the detectable signal from those that are malignant, and those that are not. This seems unlikely to be the case, because non-malignant FDG-avid nodules are just as likely to have their visibility and SUVs increased as are malignant nodules. As such, using BPL instead of OSEM improves the sensitivity of PET/CT in nodule characterisation but does not appear to improve its specificity.

The value of BPL is most likely to be its ability to provide better visibility and more accurate quantitative data. This has been shown in phantom studies [15, 25] and our results demonstrate this advancement specifically

in small nodules. This should enable PET/CT to be used in the assessment of nodules currently not thought suitable for imaging, potentially 6- to 9-mm nodules, when a low SUV or visibility is known to be an inaccurate assessment of their true metabolic activity. In contrast to the clinical setting of the solitary pulmonary nodule in patients without a history of known malignancy, BPL may be of particularly added value in assessing small nodules in the context of known malignancy. The highest $\% \Delta SUV_{max}$ was observed in metastatic nodules in our cohort (65.1 %, Table 3).

Alongside the small numbers of benign nodules analysed, this study found relatively low AUC values derived compared to larger group data [1], owing to the stand-alone use of SUV_{max} as a single determinant of malignant involvement. Although, this supports the published observations that semi-quantitative analysis does not improve the accuracy of FDG-PET in lung nodule assessment.

Conclusion

The use of BPL, an iterative reconstruction technique using a Bayesian penalised likelihood reconstruction algorithm, results in a significant increase in signal-to-noise and signal-to-background measures in comparison to conventional OSEM reconstruction. While it does not improve the overall accuracy of 18F-FDG PET/CT for differentiating benign from malignant nodules, it appears to provide a more accurate report on the metabolic activity of the nodules. When a semi-quantitative analysis is applied in the assessment of pulmonary nodules whose signal is reconstructed using BPL, a higher SUV_{max} threshold may be warranted owing to the general increase in SUV_{max} values in comparison to conventional OSEM.

Acknowledgments The authors thank Alice Asseraf and Hugo Arques at GE Healthcare for their assistance in the processing of sinograms using BPL.

The scientific guarantor of this publication is Prof. Fergus Gleeson. The authors of this manuscript declare no relationships with any companies whose products or services may be related to the subject matter of the article. Daniel McGowan is funded by a National Institute for Health Research (NIHR) Healthcare Scientist Chief Scientific Officer Doctoral Award. The research was supported by the NIHR Oxford Biomedical Research Centre Programme. This paper presents independent research funded by the NIHR. The views expressed are those of the authors and not necessarily those of the NHS, the NIHR or the Department of Health.

No complex statistical methods were necessary for this paper. Institutional review board approval was not required because informed consent is not required for retrospective reviews of this nature in our institution. Written informed consent was not required for this study because informed consent is not required for retrospective reviews of this nature in our institution. Methodology: retrospective, diagnostic or prognostic study, performed at one institution.

Open Access This article is distributed under the terms of the Creative Commons Attribution-NonCommercial 4.0 International License (<http://creativecommons.org/licenses/by-nc/4.0/>), which permits any noncommercial use, distribution, and reproduction in any medium, provided you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made.

References

- Gould MK, Maclean CC, Kuschner WG, Rydzak CE, Owens DK (2001) Accuracy of positron emission tomography for diagnosis of pulmonary nodules and mass lesions: a meta-analysis. *JAMA* 285: 914–924
- Fletcher JW, Kymes SM, Gould M et al (2008) A comparison of the diagnostic accuracy of 18F-FDG PET and CT in the characterization of solitary pulmonary nodules. *J Nucl Med* 49:179–185
- van Tinteren H, Hoekstra OS, Smit EF et al (2002) Effectiveness of positron emission tomography in the preoperative assessment of patients with suspected non-small-cell lung cancer: the PLUS multicentre randomised trial. *Lancet* 359:1388–1393
- Maffione AM, Grassetto G, Rampin L et al (2014) Molecular imaging of pulmonary nodules. *AJR Am J Roentgenol* 202:W217–W223
- Khalaf M, Abdel-Nabi H, Baker J, Shao Y, Lamonica D, Gona J (2008) Relation between nodule size and 18F-FDG-PET SUV for malignant and benign pulmonary nodules. *J Hematol Oncol* 1:13-8722-1-13
- Kim SK, Allen-Auerbach M, Goldin J et al (2007) Accuracy of PET/CT in characterization of solitary pulmonary lesions. *J Nucl Med* 48:214–220
- Nomori H, Watanabe K, Ohtsuka T, Naruke T, Suemasu K, Uno K (2005) Visual and semiquantitative analyses for F-18 fluorodeoxyglucose PET scanning in pulmonary nodules 1 cm to 3 cm in size. *Ann Thorac Surg* 79:984–988
- Adams MC, Turkington TG, Wilson JM, Wong TZ (2010) A systematic review of the factors affecting accuracy of SUV measurements. *AJR Am J Roentgenol* 195:310–320
- Kinahan PE, Fletcher JW (2010) Positron emission tomography-computed tomography standardized uptake values in clinical practice and assessing response to therapy. *Semin Ultrasound CT MR* 31:496–505
- Tong S, Alessio AM, Kinahan PE (2010) Image reconstruction for PET/CT scanners: past achievements and future challenges. *Imaging Med* 2:529–545
- Alessio AM, Stearns CW, Tong S et al (2010) Application and evaluation of a measured spatially variant system model for PET image reconstruction. *IEEE Trans Med Imaging* 29:938–949
- Lasnon C, Hicks RJ, Beaugreard JM et al (2012) Impact of point spread function reconstruction on thoracic lymph node staging with 18F-FDG PET/CT in non-small cell lung cancer. *Clin Nucl Med* 37: 971–976
- Bettinardi V, Presotto L, Rapisarda E, Picchio M, Gianolli L, Gilardi MC (2011) Physical performance of the new hybrid PETCT discovery-690. *Med Phys* 38:5394–5411
- Hudson HM, Larkin RS (1994) Accelerated image reconstruction using ordered subsets of projection data. *IEEE Trans Med Imaging* 13:601–609
- Ross S (2013) Q.Clear (GE Healthcare White Paper). http://www3.gehealthcare.co.uk/~media/documents/us-global/products/petct/whitepaper/q%20clear/ge-healthcare-white-paper_qclear.pdf. Accessed 16 Apr 2015
- Geman S, McClure DE (1987) Statistical methods for tomographic image reconstruction. *Bull Int Stat Inst* 52:5–21
- Mumcuoglu EU, Leahy RM, Cherry SR (1996) Bayesian reconstruction of PET images: methodology and performance analysis. *Phys Med Biol* 41:1777–1807
- Nuyts J, Beque D, Dupont P, Mortelmans L (2002) A concave prior penalizing relative differences for maximum-a-posteriori reconstruction in emission tomography. *IEEE Trans Nucl Sci* 49:56–60
- Asma E, Ahn S, Ross SG, Chen A, Manjeshwar RM (2012) Accurate and consistent lesion quantitation with clinically acceptable penalized likelihood images. Nuclear Science Symposium and Medical Imaging Conference (NSS/MIC), 2012 IEEE:4062–4066
- Ahn S, Fessler JA (2003) Globally convergent image reconstruction for emission tomography using relaxed ordered subsets algorithms. *IEEE Trans Med Imaging* 22:613–626
- Hanley JA, McNeil BJ (1983) A method of comparing the areas under receiver operating characteristic curves derived from the same cases. *Radiology* 148:839–843
- Rahmim A, Qi J, Sossi V (2013) Resolution modeling in PET imaging: theory, practice, benefits, and pitfalls. *Med Phys* 40: 064301
- Herder GJ, Golding RP, Hoekstra OS et al (2004) The performance of (18)F-fluorodeoxyglucose positron emission tomography in small solitary pulmonary nodules. *Eur J Nucl Med Mol Imaging* 31:1231–1236
- Kwee TC, Cheng G, Lam MG, Basu S, Alavi A (2013) SUVmax of 2.5 should not be embraced as a magic threshold for separating benign from malignant lesions. *Eur J Nucl Med Mol Imaging* 40: 1475–1477
- McGowan DR, Teoh EJ, Franklin J, Bradley KM, Gleeson FV, Fenwick JD (2014) Improvement using a new Bayesian penalised likelihood reconstruction (QClear) on contrast recovery and background variability in the NEMA image quality phantom and patients. *Eur J Nucl Med Mol Imaging* 41:S172