

Technical University of Denmark



A bioimpedance-based sensing system for monitoring cellular dynamics in a 3D culture environment

Canali, Chiara; Larsen, Layla Bashir; Dufva, Martin; Wolff, Anders; Grøttem Martinsen, Ørjan ; Heiskanen, Arto; Emnéus, Jenny

Published in:

Proceedings of the 15th International Conference on Biomedical Engineering

Publication date:

2013

Document Version

Publisher's PDF, also known as Version of record

[Link back to DTU Orbit](#)

Citation (APA):

Canali, C., Muhammad, H. B., Dufva, M., Wolff, A., Grøttem Martinsen, Ø., Heiskanen, A., & Emnéus, J. (2013). A bioimpedance-based sensing system for monitoring cellular dynamics in a 3D culture environment. In Proceedings of the 15th International Conference on Biomedical Engineering [320]

DTU Library

Technical Information Center of Denmark

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

ABSTRACT OF CONTRIBUTION ICBME 2013

ID: 320

Oral Presentation

Topics: D1) Biochips, D6) Lab-on-Chip

Keywords: Lab-On-A-Chip, 3D-culture, bioimpedance, scaffold, stem cells

A bioimpedance-based sensing system for monitoring cellular dynamics in a 3D culture environment

Chiara Canali¹, Haseena Bashir Muhammad¹, Martin Dufva¹, Anders Wolff¹, Ørjan Grøttem Martinsen^{2,3}, Arto Heiskanen¹, Jenny Emnéus¹

¹Department of Micro- and Nano-technology, Technical University of Denmark, Kgs Lyngby, Denmark; ²Department of Physics, University of Oslo, Oslo, Norway; ³Department of Biomedical and Clinical Engineering, Rikshospitalet, Oslo University Hospital, Oslo Norway

chca@nanotech.dtu.dk

There is currently an increasing interest in developing sensitive analytical methods for real-time monitoring of the entire process of tissue engineering, starting from a bare 3D polymer scaffold, to cell attachment, growth and differentiation, up to the formation of vascularised organ-on-a-chip systems.

Bioimpedance has shown to be a powerful tool for the study and modeling of biological systems both in vivo and in vitro, establishing a physical correlation between bioelectrical measurements and tissue growth characteristics.

In this work, a bioimpedance-based 3D culture Lab-On-A-Chip (LOC) system is presented which is designed and optimized to investigate cellular dynamics occurring under in vitro tissue growth conditions. The novelty of this device is to enable label-free monitoring of real cell dynamics in a more physiological microenvironment mimicking the complex network of interactions taking place among cells and their matrix.

Since the important electrical characteristics of an electrode/tissue system are primarily determined by the electrode configuration, simplified finite element models were used to optimize electrode density and orientation within the 3D chip to enhance measurement sensitivity. Based on the simulation results, a prototype chip was developed with two opposite couples of vertical rectangular plate electrodes in a three-terminal configuration where current-carrying (CC) and pick-up (PU) electrodes could be switched between different positions to extend a positive sensitivity field to the whole bulk scaffold volume.

To demonstrate proof-of-concept, the bioimpedance-based LOC system was evaluated by growing different concentrations of mesenchymal stem cells embedded within a 3D gelatin scaffold. Preliminary results indicate that the presented device has a high potential for real-time monitoring of the entire process of tissue engineering without affecting cell viability.

Since different combinations of CC and PU electrodes can be exploited, this system will pave the way towards electrical impedance tomography applications enabling imaging of the changes occurring in 3D cell culture environments.