

Johnston, Lauren and Flamini, Vittoria and Kazakidi, Asimina (2018) Blood flow in an intracranial aneurysmal artery with a dual-layer stent. In: BioMedEng18 Proceedings. BioMedEng, London, p. 382. ISBN 978-1-9996465-0-9,

This version is available at https://strathprints.strath.ac.uk/65495/

Strathprints is designed to allow users to access the research output of the University of Strathclyde. Unless otherwise explicitly stated on the manuscript, Copyright © and Moral Rights for the papers on this site are retained by the individual authors and/or other copyright owners. Please check the manuscript for details of any other licences that may have been applied. You may not engage in further distribution of the material for any profitmaking activities or any commercial gain. You may freely distribute both the url (<u>https://strathprints.strath.ac.uk/</u>) and the content of this paper for research or private study, educational, or not-for-profit purposes without prior permission or charge.

Any correspondence concerning this service should be sent to the Strathprints administrator: strathprints@strath.ac.uk

The Strathprints institutional repository (https://strathprints.strath.ac.uk) is a digital archive of University of Strathclyde research outputs. It has been developed to disseminate open access research outputs, expose data about those outputs, and enable the management and persistent access to Strathclyde's intellectual output.

BioMedEng18

Blood flow in an intracranial aneurysmal artery with a dual-layer stent

L. Johnston¹, V. Flamini², A. Kazakidi¹

¹ Department of Biomedical Engineering, University of Strathclyde, Glasgow, UK ² Tandon School of Engineering, New York University, New York, USA

Intracranial aneurysms are focal dilatations of arteries in the cerebral vasculature that require surgical intervention once detected. Dual-layer flow-diverting stents are the latest innovation intended for the endovascular treatment of intracranial aneurysms. A fine-mesh device, comprised of a low porosity inner stent and a high porosity outer stent, is implanted within the parent artery bearing the aneurysm. The inner layer is designed to alter the haemodynamics of both the aneurysm and parent vessel, and is therefore the active flow-diverting element, while the outer layer ensures the patency of nearby perforating arteries^{1,2}. The FRED (Flow-Redirection Endoluminal Device) stent is the first device of its kind, with clinical approval in European and international markets. Several studies investigating the efficacy of the FRED stent have found its design to be safe and effective in the treatment of difficult-to-treat or otherwise untreatable intracranial aneurysms^{2,3}. However, clinical radiological post-procedural data is limited. As of January 2018, only 370 patients across ten studies reported the efficacy of the FRED device, with primarily saccular and small aneurysms being treated. This work is the first detailed computational fluid dynamic analysis of blood flow in a section of an intracranial aneurysmal artery after treatment with a dual-layer stent. In particular, two-dimensional computational models implemented a virtual bench test to investigate local haemodynamic changes in part of the treated vessel. Preliminary results consider steady flow conditions, based on a parabolic inlet profile, for the solution of the incompressible Newtonian Navier-Stokes equations. Computational fluid dynamic analysis focussed on the variance of flow parameters, including velocity, pressure, vorticity, and wall shear stress, for three dual-layer stent models: a baseline model with the same number of total wires as the FRED stent on both layers; an alternative model with half the number of wires in the outer layer; and a second alternative model with half the number of wires in the inner layer⁴. The most favourable flow behaviour in the aneurysmal artery was observed for the design in which the outer layer has half the number of wires than the inner layer. For all designs, the radial clearance between the two layers was seen to reduce the velocity of the surrounding blood flow to 98% of the blood flowing at the vessel centre. This research provides an aid to quantitatively estimate the flow of blood in a dual-layer treated vessel and will facilitate the optimisation of such devices.

References

- 1. Diaz O. et al., Journal of NeuroInterventional Surgery 2014; 6(8): 614 617.
- 2. Kocer N. et al., Journal of Neurosurgery 2014; 120(5): 1158 1171.
- 3. Mahboobani N.R. et al., Neurointervention 2017; 12(1): 11 19.
- 4. Alherz A.I., Tanweer O., Flamini V., Journal of Biomechanics 2016; 49(1): 2420 2427

Acknowledgments

This work is supported in part from the University of Strathclyde Research Studentship Scheme (SRSS) Student Excellence Awards (SEA) Project No 1619, and the European Union's Horizon 2020 research and innovation programme under the Marie Skłodowska-Curie grant agreement No 749185.

