

Spectrochimica acta, Part B: Atomic spectroscopy 1998 vol.53 B N6-8, pages 1003-1014

---

## Multidimensional integration of absorbances: An approach to absolute analyte detection

Gilmutdinov A., Harnly J.

Kazan Federal University, 420008, Kremlevskaya 18, Kazan, Russia

---

### Abstract

The problem of absolute analyte detection is considered in this paper. It is shown that integration in absorbance, not in intensity, is a pre-requisite for absolute detection in atomic absorption spectrometry. A design for an atomic absorption spectrometer of the future is described which measures absorbance resolved in three key areas: wavelength, space and time. Intensity must be measured with sufficient temporal, spatial and spectral resolution to guarantee the accuracy of the computed absorbance. Technically, such measurements can be made simultaneously with a continuum source, a high resolution echelle spectrometer and a two dimensional solid-state array detector. All computed absorbances are fully background and stray light corrected. With such measurements, and a proper optical configuration, absolute analyte detection can become a reality and the possibility of absolute analysis becomes more obtainable.

---