

Applied Magnetic Resonance 2003 vol.25 N2, pages 249-259

A Q-Band Pulsed ENDOR Spectrometer for the Study of Transition Metal Ion Complexes in Solids

Hoentsch J., Rosentzweig Y., Heinhold D., Köhler K., Gutjahr M., Pöppl A., Völkel G., Böttcher R.
Kazan Federal University, 420008, Kremlevskaya 18, Kazan, Russia

Abstract

We describe the design of a pulsed electron nuclear double resonance (ENDOR) spectrometer operating at Q-band frequencies (35 GHz) for studies of transition metal ion complexes in the temperature range between 4.2 and 297 K. Specific features of the spectrometer are a microwave IMPATT generator, a homebuilt cavity, and a commercial Bruker magnet. Standard Davies and Mims ENDOR sequences have been implemented. The performance of the spectrometer is demonstrated for a broad radio frequency range by ^1H , ^{14}N , ^{31}P , ^{133}Cs , and ^{207}Pb pulsed ENDOR experiments of Cu^{2+} , Cr^{5+} , and V^{4+} transition metal ion complexes in both single crystals and disordered materials.
