


Probing Shallow trapped electrons of TiO₂ with UHV-IRRAS

Hikmet Sezen, Carsten Natzeck, Alexei Nefedov, Christof Wöll


1 Scope

Probing trapped shallow state electrons delivering from both atomic hydrogen and UV treatments on powder and single crystal TiO₂ samples by a novel method of the infrared reflection absorption spectroscopy in an ultrahigh vacuum environment (UHV-IRRAS).^[1-2] However, now we have a more furnished UHV system at KIT, which is called **THEO**!



2 Instrumental

THEO





Performance of the system: CO absorption on r-TiO₂^[2]

Single Crystals:

E-beam sample holder:
 ΔT : 1500K-100K (LHe 30K)


Powders:

Transmission sample holder:
 ΔT : 1200K-100K (LHe 30K)




3 Experiment

Sample Preparation: >



4 Results



5 References

- 1.) M. Xu, Y. Gao, E. M. Moreno, M. Kunst, M. Muhler, Y. Wang, H. Idriss, C. Wöll, Phys. Rev. Lett. 106, 138302 (2011).
- 2.) M. Xu, H. Noei, K. Fink, M. Muhler, Y. Wang, C. Wöll, Angew. Chem. Int. Ed. 51, 4731 (2012)
- 3.) C. Di Valentin, G. Pacchioni, A. Selloni, J. of Phys. Chem. C 113, 20543 (2009)
- 4.) D. A. Panayotov and J. T. Yates, Chem. Phys. Lett. 436, 204 (2007).
- 5.) D. A. Panayotov, S. P. Burrows, J. R. Morris, J. Phys. Chem. C 116, 4535 (2012).