



Institutional Repository - Research Portal

Dépôt Institutionnel - Portail de la Recherche

researchportal.unamur.be

RESEARCH OUTPUTS / RÉSULTATS DE RECHERCHE

Oriented transfer of proteins for biosensor applications

Volcke, Cedric; Ghandiraman, Ram Prasad; Basabe-Desmots, Lourdes; Riaz, Asif; Cafolla, Attilio A.; McCraith, Brian D.

Published in:
European Cells and Materials

Publication date:
2007

[Link to publication](#)

Citation for pulished version (HARVARD):

Volcke, C, Ghandiraman, RP, Basabe-Desmots, L, Riaz, A, Cafolla, AA & McCraith, BD 2007, 'Oriented transfer of proteins for biosensor applications' European Cells and Materials, vol. 14, no. Suppl. 3, pp. 42-42.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- You may not further distribute the material or use it for any profit-making activity or commercial gain
- You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Oriented transfer of proteins for biosensor applications

C. Volcke^{1,2}, R. P. Ghandiraman¹, L. Basabe-Desmonts¹, A. Riaz¹, A. Cafolla¹ and B. McCraith¹

¹ Biomedical Diagnostics Institute, Dublin City University, Dublin, Ireland. ² Laboratoire Lasers et Spectroscopies, University of Namur, Namur, Belgium.

INTRODUCTION: The orientation of antibodies upon their immobilisation on different surfaces is a crucial factor affecting their recognition properties. This phenomenon is strongly dependent on surface properties such as composition, reactivity, wettability, roughness etc. [1-2]. Some polymeric substrates, such as Zeonor, are used in bio-medical or -technological devices, mainly because of their interesting optical properties, low production time and cost. However, their surface modification/functionalisation is more difficult to perform as compared to other widely studied metallic substrates (such as gold or silicon).

Our objective is to overcome these problems by performing a two-step experiment. The first step uses the advantages of well-known modified metallic/semiconducting substrates (in particular organosilane-modified silicon) for the orientation of antibodies [3-4]. The second step consists of transferring the oriented antibodies onto a second substrate, which may have unfavourable characteristics. This transfer process is based on the template-imprinted nanostructures technique developed by Ratner et al. [5]. In this poster, we focus in particular on the transfer process of proteins from one substrate to the other.

METHODS: The transfer process may be described briefly as follows: a selected protein (fibrinogen, avidin or antibody) is microcontact printed onto a SiO₂ surface, a sugar layer is then spin coated on top of the patterned surface. This is followed by the deposition of a polymer-like layer using a 13.56 MHz RF capacitively coupled plasma in a PECVD reactor. This upper polymer-like layer is then glued to a second substrate. Finally the substrates are separated. The microprinted pattern of proteins is therefore transferred to the second substrate. During these experiments, the wettability, film thickness, film morphology were characterised using contact angle measurements (CA), spectroscopic ellipsometry (SE) and atomic force microscopy (AFM), respectively. Fluorescence microscopy was also used to determine the presence of tagged-proteins on the transferred substrate.

RESULTS. We first demonstrate (using AFM) the transfer of microprinted patterns of non oriented

proteins (fibrinogen – Fig. 1 left). We then show that, when using Cy5-labelled anti-human IgG, such patterns are composed of proteins (as revealed by fluorescence microscopy – Figure 1 right). Finally the function of the transferred proteins is investigated using antibodies, which are recognised by antigens after transfer.





Fig. 1: AFM image (50μm) of fibrinogen microprinted lines after transfer process (left). Fluorescence image of microcontact printed lines of Cy5-labelled anti-human IgG after transfer process (right).

CONCLUSIONS: In this poster, we have demonstrated the transfer of protein from one substrate to another without modifying the protein and its properties. Particularly, the recognition between antigens and antibodies was also proved after transfer of antibodies.

REFERENCES: ¹ A. Sethuraman, M. Han, R.S. Kane et al (2004) *Langmuir* **20**:7779-88. ² L.-C. Xu and C.A. Siedlecki (2007) *Biomaterials* **28**:3273-83. ³ S. Chen, L. Liu, J. Zhou et al (2003) *Langmuir* **19**: 2859-64. ⁴ H. Wang, D.G. Castner, B.D. Ratner et al (2003) *Langmuir* **20**:1877-87. ⁵ H. Shi, W.-B. Tsai, M.D. Garrison et al (1999) *Nature* **398**:593-597.

ACKNOWLEDGEMENTS: C.V. is a postdoctoral researcher of the Belgian National Fund for Scientific Research (F.N.R.S.). Financial support of the Science Foundation Ireland (S.F.I.) is acknowledged.