

Serveur Académique Lausannois SERVAL serval.unil.ch

Author Manuscript

Faculty of Biology and Medicine Publication

This paper has been peer-reviewed but does not include the final publisher proof-corrections or journal pagination.

Published in final edited form as:

Title: Automation in the clinical microbiology laboratory.

Authors: Burnham CA, Dunne WM Jr, Greub G, Novak SM, Patel R

Journal: Clinical chemistry

Year: 2013 Dec

Volume: 59

Issue: 12

Pages: 1696-702

DOI: 10.1373/clinchem.2012.201038

In the absence of a copyright statement, users should assume that standard copyright protection applies, unless the article contains an explicit statement to the contrary. In case of doubt, contact the journal publisher to verify the copyright status of an article.

1 **Q & A: Automation in the Clinical Microbiology Laboratory**

2

3 **Moderator:** Carey-Ann D. Burnham^{1*}

4

Experts: W. Michael Dunne, Jr.², Gilbert Greub³, Susan M. Novak⁴, and Robin Patel⁵

5

6 ¹Department of Pathology & Immunology and Pediatrics, Washington University School of Medicine, St.

7 Louis, MO, ²bioMerieux, Inc., Durham, NC, ³Institute of Microbiology, University of Lausanne and

8 University Hospital Center, Lausanne, Switzerland, ⁴Southern California Permanente Medical Group—

9 Regional Reference Laboratories, ⁵Division of Clinical Microbiology, Mayo Clinic, Rochester, MN

10

11

12 *Address correspondence to this author at:

13 Department of Pathology & Immunology

14 Washington University School of Medicine

15 St. Louis, MO 63110

16 Telephone: (314) 362-1547

17 E-mail: cburnham@path.wustl.edu

18

19 The clinical microbiology laboratory has historically been considered “low-tech”, especially when
20 compared to the clinical chemistry laboratory. However, systems are emerging for the clinical
21 microbiology laboratory with the potential to automate almost all areas of testing, including inoculation
22 of primary culture plates, detection of growth on culture media, identification of microorganisms,
23 susceptibility testing, and extraction and detection of nucleic acids in clinical specimens. As a result, the
24 workflow in the microbiology laboratory is changing at a rapid pace and microbiologists have the
25 challenge of selecting the most appropriate, clinically useful, and cost-effective automation for their
26 laboratories. We have asked four experts in this field, from clinical microbiology laboratories in the
27 United States and Europe, as well as from industry, to comment on the feasibility and impact of
28 automation in the clinical microbiology laboratory.

29

30 ***Are you currently using or do you anticipate using an automation platform in***
31 ***your microbiology laboratory? If yes, which sections of your laboratory are***
32 ***automated?***

33 **Robin Patel:** Mayo Clinic’s clinical microbiology laboratory has been performing testing since 1911.
34 Although select tests today resemble those performed a century ago, we have many examples of
35 automated, state-of-the-art tests. These include blood cultures, infectious diseases serologic platforms
36 and nucleic acid and proteomic diagnostics, to name a few. For over two decades, microbiology
37 laboratories have been using automated blood culture instruments that “sense” microbial growth in
38 blood culture bottles and “flag” positive bottles for immediate attention by laboratory technologists.
39 Prior to the availability of such systems (in the not-so-distant past), laboratory technologists manually

40 evaluated *each* blood culture bottle on *multiple* occasions. Technologists today could not fathom
41 returning to the manual approach used a mere three decades ago. As with many chemistry tests, a
42 myriad of infectious diseases serologic tests are performed on automated platforms. Nucleic acid
43 diagnostics, which have been used in our laboratory for over two decades, are evolving to automated,
44 “black-box”-type formats. We use a core nucleic acid extraction facility preparatory to downstream
45 molecular microbiology testing. And, several of our nucleic acid amplification platforms are fully
46 automated. Finally, matrix assisted laser desorption ionization time of flight mass spectrometry has
47 revolutionized identification of bacteria and fungi in our laboratory. The success of this automated
48 platform lies in advances in mass spectrometry and bioinformatics. With its implementation, we have
49 been able to not only to reduce turn-around-times but also the associated cost of identification of
50 organisms.

51 **Susan M. Novak:** I direct a high volume laboratory servicing 14 hospitals and >100 clinics in the
52 southwestern United States where automation is a necessity. Our integrated health care system
53 services members in a broad geographic area relying on a centralized reference laboratory for
54 processing and work up of approximately 4000 microbiology specimens per day. Our microbiology
55 laboratory has been automated since 2002 with semi-automated and automated plating
56 instrumentation. Greater than 80% of the microbiology specimens arrive at the central laboratory for
57 processing and plating. Based on the high specimen volume the decision was made years ago to
58 capitalize on the economies of scale and allow centralized plating. This improved the plating quality and
59 mitigated ergonomic injuries by automating repetitive manual functions. Specimen processing delay in
60 the medical centers was another reason our health care system chose to centralize and automate
61 microbiology plating. Our laboratory is anticipated to expand over the next several years. This
62 expansion and increased testing volume will continue to put us in the position to incorporate additional

63 automation into the laboratory setting to continue to meet the goal of cost effective quality laboratory
64 testing.

65 **Gilbert Greub:** Given the current shortage of financial resources and the concomitant increase in
66 activity of our clinical diagnostic microbiology laboratory of about 4 to 12% per year, we decided to
67 move towards a fully automated bacteriology laboratory. We also considered automation a priority for
68 our serological and molecular diagnostic laboratories. Given the ease to automatically handle samples
69 received in a liquid format, automation of serological laboratories is relatively straightforward. The same
70 is true regarding molecular diagnostic laboratories, as soon as the first steps of DNA extraction have
71 been done. Thus, in our molecular diagnostic laboratory, we now use automation for most DNA
72 extraction and PCRs. Automation of PCR was associated with (i) reduced technician workload, (ii)
73 reduced rate of PCR contamination (< 1%), and (iii) reduced time to results (< 24h).

74 In our bacteriology laboratory, since fall 2011, we are using an automated inoculation system
75 which was selected among all systems available at that time based on various criteria including number
76 and diversity of samples daily received in our laboratory. This system considerably reduces the
77 technician workload by automatically inoculating and spreading each sample, labeling and sorting
78 inoculated agar plates and preparing smears for Gram staining. For microbial identification we
79 currently use two stand-alone automated systems. In 2015, we will move to a fully automated
80 laboratory, by adding the missing pieces to the puzzle, i.e. smart incubators, high quality digital imaging,
81 automated colony picking system and all required transport belts in-between.

82 ***Which features or criteria are/would be most important to you when selecting a***
83 ***laboratory automation system?***

84 **Susan M. Novak:** Based on our experiences there are several equally important features that need to
85 be considered when selecting laboratory automation. Specimen volume, mix and timing of specimen
86 arrival in the laboratory will drive the size of instrumentation, throughput requirements and type of
87 automation. Throughput is a very important metric that must be assessed carefully. The laboratory
88 automation capacity must parallel and support the overall specimen volume of each individual
89 laboratory. In addition to throughput, reliability (mean time to failure) of the automation is extremely
90 important. If adjustments to staffing are based on the implementation of automation it is extremely
91 important to have robust instrumentation that has little down time and negative impact on the day to
92 day work flow. Excessive down time impacts staffing (i.e. overtime) in addition to turn-around-time and
93 even the morale of the staff. Implementation of automation in the laboratory is a complex decision and
94 the negative effect of a poor decision cannot be stressed enough. It is important that vendors realize
95 that they must become partners in helping laboratories work through change management with the
96 staff as automation is integrated into the laboratory setting. Employee “buy in” is critical to the success
97 of a new test or piece of automation. Other criteria are important and it is advisable for laboratories
98 considering automation to get input from colleagues that have prior experience.

99 **Robin Patel:** Several issues bear consideration. Cost is paramount – will the system result in increased
100 or decreased operating costs? Increased costs are associated with instrumentation, reagents and/or
101 disposables. Automated systems should be green; in other words, reagents and disposables should be
102 minimal. Decreased costs would primarily be expected to be personnel-related. Laboratory safety is an
103 important consideration as there may be hazards to laboratory personnel associated with automation,
104 either as a result of exposure to hazardous microorganisms or dangerous equipment. As bizarre as the
105 latter may seem, we have had instruments harm laboratory personnel! Space must also be considered –
106 is there space to accommodate automation (and how much remodeling will be needed)? Automation

107 can “solve” problems in the laboratory. Our technologists and laboratory assistants suffered ergonomic
108 injuries (e.g., as a result of repeat pipetting), which are avoided with automation. Quality control *may* be
109 improved, due to avoidance of human error. The possibility of nucleic acid contamination of molecular
110 diagnostic assays deserves careful consideration, even in the era of automation. Systems must interface
111 with electronic medical records for test ordering and result reporting. Volume throughput must be taken
112 into consideration. Can one system/instrument accommodate the laboratory’s volumes? Are modules
113 available so that the system can evolve with changing test volumes? Specimen collection containers may
114 need to be standardized to integrate with automated systems. This seemingly trivial task can be
115 monumental in a large healthcare setting. Instrument evaluation may not always be feasible with full-
116 laboratory automation, impacting the decision-making process (i.e., should the laboratory adopt a
117 specific platform). Laboratories will need to address strategies to validate automation. Finally, as
118 laboratories become more reliant on automation, it will be challenging to deal with system failure (due
119 to instrument breakdown or lack of availability of disposables). Back-up plans will need to be carefully
120 considered. We have experienced multiple situations where we have moved all of a certain type of
121 testing to an automated platform only to experience an inability to utilize the platform due to
122 instrument failure or lack of availability of required disposables (which are often offered from sole
123 source suppliers). In such situations, the laboratory may be stranded in terms of testing. Lack of
124 availability of disposables for commonly used systems may have a substantial impact on healthcare
125 across the nation and around the world.

126

127 ***How would automation impact your laboratory operation and workflow? Will***
128 ***microbiology laboratories make a shift towards “24 hour microbiology”?***

129 **Robin Patel:** In our experience, automation has dramatic affects in these areas. For each system
130 adopted, laboratories need to reevaluate operations, workflow, space and staffing assignments. Our
131 laboratory operates 24/7, so increased automation won't change our routine operating hours! Over the
132 next decade, however, I expect to see more automation, including robotic incubation of culture plates,
133 and automated reading of culture plates, as well as automated identification and antimicrobial
134 susceptibility testing of colonies growing on plates (i.e., using "colony pickers"). Automation of tedious
135 tasks will allow reassignment of staff to high value work areas requiring increased use of their
136 intellectual skills. The microbiology technologist of the future will likely apply some of their expertise via
137 computer interfaces (rather than handling culture plates one-by-one).

138 **Wm. Michael Dunne, Jr.:** If a tree falls in the forest and no one is there to hear it, does it make a sound?
139 It follows that the sooner clinically actionable patient results were provided to clinicians, the sooner
140 appropriate therapy would be administered. This, in turn, could potentially reduce overall length of
141 hospitalization with concomitant reduction of patient care costs, risk for nosocomial infection, and
142 improved outcomes. Time to results (TTR) can be reduced in a number of ways including the use of
143 rapid diagnostic platforms (e.g., real-time PCR) or the performance of testing on a more frequent basis.
144 Automation in microbiology holds the promise of specimen processing on a real-time basis that, when
145 coupled to rapid diagnostic modalities could significantly reduce TTR and, by inference, generate
146 positive patient outcomes. Automation, however, doesn't mean the system runs by itself. Additional
147 shifts of operation require personnel support (at least until artificial intelligence algorithms permit
148 accurate interpretation and processing of cultures in the absence of humans). If pertinent patient
149 results are generated and posted on the laboratory information system in the wee hours but not acted
150 upon by medical personnel, then the opportunity is lost and the additional expense of 24 hour operation
151 is not realized in terms of patient outcome. I recall a situation where the laboratory adopted RT-PCR

152 screening for MRSA. Because of the cost of controls, the test was performed once daily in the
153 afternoon. Upon further evaluation, the results were not reviewed by infection prevention staff until
154 the following morning. If the samples had been plated on MRSA screening media when they were
155 received, the results were available sooner in many cases with similar performance characteristics and
156 an isolate for subsequent testing. This does not necessarily apply to reference laboratory testing where
157 results are billed directly to clients and the additional workload of off-shift testing generates real
158 income. In this case, the decision to offer 24-hour testing is found in the balance sheets where
159 additional revenues are offset by labor expenditures and operational costs. In summary, the answer to
160 this question should be based on at least two options - improved patient outcome and/or improved
161 laboratory income. It would be nice if both were realized.

162

163 **Gilbert Greub:** A 24/24 7/7 availability of some immunochromatographic tests, some direct microscopic
164 examination such as blood smear for *Plasmodium* spp. and of selected PCRs is warranted given the
165 clinical impact and the added value of short time to results. However, given the painfulness of night
166 shift, the associated social cost and the relatively low clinical impact of 24/24 7/7 shift for most
167 microbiology processes, a 24 hour microbiology is not a goal of our core bacteriology laboratory.
168 Nevertheless, automation will clearly modify our workflow and organization. Since agar plates may be
169 automatically checked by digital imaging at regular intervals and sterile plates automatically discarded,
170 time to results is significantly improved, especially if the technician working time is extended for
171 instance from 6 a.m. to 10 p.m. Moreover, moving towards paperless telebacteriology will be a major
172 opportunity to improve the working atmosphere by reading agar plates in quiet offices located next but
173 outside the core laboratory, with an expected gain in quality and productivity.

174

175 ***Is a full-laboratory automation system appropriate for all types of laboratories?***

176 ***Is there a minimum daily specimen volume required to support an automated***
177 ***platform?***

178 **Wm. Michael Dunne, Jr:** The simple answer is “it depends.” So what does it depend upon? 1.
179 Scalability and modularity. Can an automated microbiology system be appropriately scaled to
180 accommodate the specimen volume and sample types of individual microbiology laboratories with
181 associated cost adjustments and the potential to increase throughput at a later date? For example, I
182 recall an instrument developed in the early 2000’s that would screen 300 urine samples per hour and
183 generate rudimentary identification and susceptibility data for the samples if positive. The problem with
184 this business model was that very few laboratories had volumes to justify purchase of the instrument
185 and a scaled down version wasn’t available. The system was also very large and was difficult to get into
186 the laboratory proper without demolition activities. So if a one-size-fits-all option is developed by
187 manufacturers, it might limit the market to very large volume laboratories. Modularity would also allow
188 laboratories of various sizes to implement specific units of the system. 2. Availability of skilled labor. As
189 the pipeline of individuals completing training in clinical laboratory science programs at all degree levels
190 decreases, the competition for their services among laboratories will likely increase with the associated
191 cost of competitive salaries and benefits (supply and demand). The reduced availability of a skilled
192 workforce even in smaller laboratories would likely push the equation in favor of automated systems
193 such that fewer individuals could handle the workload while devoting more time to interpretation and
194 other complex processes. 3. Medical economics. Labor, I believe, still remains the single most costly
195 component of any laboratory budget line item so it would be desirable to stabilize and minimize the
196 effects of rising labor costs and shortages. Reimbursement should also be considered in the equation

197 and I can't help but believe that reimbursement rates from third-party payers and government agencies
198 won't be going up any time soon. 4. Specimen complexity. Clearly a laboratory that processes several
199 hundred urine samples per day would be more amenable to automation than one that processes a mix
200 of complex sample types. To get back to the question, the answer obviously lies somewhere in between
201 the office lab that processes 5 throat samples and 10 urine cultures per day and a large university-based
202 or reference laboratory that evaluates hundreds to thousands of mixed sample type cultures daily.

203

204 **Gilbert Greub:** Below a given threshold, the cost of the automated system and its maintenance will
205 outweigh the benefit in term of hands on time and quality. The minimum daily specimen volume
206 required to support an automated platform will be different for each platform and cost-effectiveness
207 studies are warranted to define such cut-off. However, these cost analyses are yet completely lacking
208 for automated microbiology systems. For example, when considering the different inoculation systems
209 currently on the market, it readily appear that they target mid-size laboratories since they may inoculate
210 as many as 180 to 270 agar plates per hour. Such throughput of these inoculation systems is also
211 optimal for larger laboratories that generally use several inoculation systems in parallel to ease
212 workflow organization and to provide the necessary back-up during technical failures. Automated
213 systems are optimal when the workflow is continuous, i.e. the efficiency decreases when samples are
214 processed by batch.

215 **Susan M. Novak:** Because all laboratories are not created equal based on depth and scope of work
216 performed, I do not believe that full laboratory automation systems will be appropriate for all
217 laboratories. Incorporation of full laboratory automation will be highly dependent on specimen volume
218 and the need to become more efficient in the workplace and reduce FTE's associated with manual tasks
219 that can be automated. I believe the medium to large size laboratories will be able to justify full

220 laboratory automation more readily because the purchasing of equipment can be tied to a “return on
221 investment” relative to the amount of staff needed in the laboratory to perform certain tasks. Smaller
222 volume laboratories will perhaps take a modular approach with automation, choosing modules or pieces
223 of automation that support the laboratory. In speaking with various vendors of plating instrumentation
224 platforms, the cut off for a plating instrument appears to be approximately 200 specimens per day. But
225 some vendors do believe smaller labs could still use this automation. That said, given that full
226 laboratory automation is new to microbiology, each lab must perform an analysis to determine what
227 automation, if any, would be suitable.

228

229 ***Do you anticipate that adoption of an automated microbiology platform would***
230 ***restrict you to using products from one vendor?***

231 **Susan M. Novak:** This is a very interesting question and one many laboratories that are considering
232 automation in microbiology are most likely grappling with. Today there are several pieces of
233 automation that have been fixtures in the clinical microbiology laboratory for years (i.e.
234 identification/susceptibility and blood culture systems). In some laboratory settings, such as ours,
235 specimen plating instruments have been incorporated to meet pre-analytical needs. Because these
236 instruments are unrelated to one another there has never been the need for connectivity. The need for
237 integration all changes with the onset of full laboratory automation. For example, if plates are
238 transferred via a conveyor below to a smart incubator with a digital camera and a colony is chosen for
239 identification or susceptibility testing, the plate must be sent to a piece of equipment that can pick the
240 colony and inoculate needed materials. In this scenario, connectivity is a prerequisite to full laboratory
241 automation. How will that occur if each piece of automation is from a different vendor? Just because a

242 laboratory chooses pre-analytical plating instruments or a digital incubator should not mean that the
243 identification/susceptibility instrument would need to be changed or that the laboratory will have the
244 money to do so. The question remains will the software on these systems be “open” and can various
245 instruments from varied manufacturers talk to one another? If this is not the case then this could place
246 a huge burden on the laboratory. Switching out several instruments at once would impact the capital
247 expenditure and also impact the number of validations that need to be performed when integrating new
248 equipment into the laboratory setting.

249 **Gilbert Greub:** Adoption of an automated microbiology platform may restrict the use of products
250 proposed by another vendor. Some companies try to protect their market by proposing automated
251 system with no or very low compatibility with systems from other manufacturers. Fortunately, the
252 increasing awareness of clinical microbiologists of the importance of such technical compatibility is
253 progressively pushing the industry to propose flexible solutions. Moreover, some middleware options
254 are available that may solve some incompatibilities between different automated systems and between
255 an automated system and the laboratory information system (LIS).

256

257 ***In an era of laboratory automation, do you think that “routine” microbiology***
258 ***will continue to be done by specialized microbiology technologists, or do you***
259 ***think that generalists and/or chemistry technologists will be involved in this***
260 ***testing?***

261 **Gilbert Greub:** Microbiology laboratories will still need technologists trained in microbiology at several
262 steps, including microscopy and agar plate interpretation, since these activities will still represent a

263 pivotal step governing decision on downstream steps. With automation, the proportion of repetitive
264 tasks requiring limited knowledge will decrease whereas specialized tasks will increase, as will the need
265 for specific skills in information technology (IT).

266 **Wm. Michael Dunne, Jr.:** Yes (another hedge!). On one end of the spectrum, less-complicated
267 processes i.e., specimen accessioning, loading of samples into automated plating systems, reading and
268 interpretation of screening media, pulling, subculturing and preparing gram-stains of positive blood
269 cultures, setting up samples for nucleic acid extraction, antimicrobial susceptibility testing (AST),
270 identification (including MALDI-TOF) and interpretation of certain culture types (e.g., urine) can be
271 readily accomplished by well-trained laboratory technicians and generalists. I am a firm believer,
272 however, that the correct interpretation of more complicated specimen cultures (respiratory, GU, GI,
273 nephrostomy urine, CSF, blood, tissue, abscess material, implants, etc.) in the context of relevant patient
274 information requires more focused training and experience. This is especially true for the interpretation
275 of antimicrobial susceptibility testing results. Heck, I've been a clinical microbiologist for more than 30
276 years and still run across results that I can't explain. That is not to say that generalists and/or clinical
277 chemistry technologists can't gain this level of experience over time, but it requires guidance and
278 mentoring. Automation in microbiology doesn't negate the need for interpretive skills. On the contrary,
279 automated systems should enhance the efficacy of microbiology specialists by allowing them to devote
280 their attention toward problem solving and interpretation of complicated sample results or processes.
281 Further, automation in microbiology has not yet supplanted the need to explain test results to clinicians
282 in a clear and knowledgeable manner or to suggest further supplementary testing. So in my estimation,
283 it would behoove the fully automated clinical microbiology laboratory to establish a good mix of
284 microbiology specialists for training, hierarchical result interpretation, and physician interface with
285 generalists capable of maintaining the accelerated workflow.

286 **Susan M. Novak:** Clinical laboratory scientist pools are diminishing due to many technologists retiring
287 and fewer replacements in the workplace due to the elimination of training programs. I believe that the
288 microbiology laboratory will continue to evolve and change relative to how it appears today. Based on
289 the development of more sophisticated automation and software tools there will be a place for the less
290 specialized laboratorian in microbiology. With the advancement of digital microbiology and integrated
291 software programs perhaps a less skilled laboratorian can review cultures that once required a licensed
292 technologist. That said, microbiology will still need to report clinically relevant information to the
293 provider, since not all bacteria growing from a culture is relevant to the patient condition or illness. The
294 specialized microbiologist will still have a place in the laboratory but it is anticipated that the overall mix
295 will change. Advances in molecular testing have already resulted in the introduction of moderately
296 complex instrumentation that is easy to run and can be run by a generalist or chemistry technologist
297 today lacking microbiology expertise.

298

299 ***What are the challenges in bringing full-laboratory automation in microbiology***
300 ***to market?***

301 **Wm. Michael Dunne, Jr.:** From my point of view (and, again, factor in 30+ years as a clinical
302 microbiologist), the primary hurdle to overcome toward total laboratory automation in microbiology is
303 perception. Unlike our colleagues in clinical chemistry and hematology who have handled much higher
304 levels of automation for many years, clinical microbiology has traditionally been a very tactile, hands-
305 on, interpretive practice that is equal parts art and science. Converting an analog discipline into a digital
306 one will require a paradigm shift. The process of evaluating microbial growth by holding plates, touching
307 colonies with loops and the associated odors (accidentally acquired of course) would be replaced by

308 high resolution images displayed on a flat screen monitor and robotic manipulation of colonies using
309 robotics. Factors that are driving microbiology in the digital direction include laboratory consolidation
310 and associated workload, an aging skilled workforce with more folks reaching full-time shell-collecting
311 age than those entering the profession, and paper-thin margins on reimbursement for services. It is my
312 feeling that the conversion to automation is happening at a more rapid pace in Europe, Asia and South
313 America than it is in the US at the moment but that might change shortly. To accommodate this trend, a
314 number of modules have been or will be developed which replicate activities that are currently being
315 performed manually. These include specimen processing, incubation, culture evaluation, identification,
316 antimicrobial susceptibility testing, and result reporting. The modules supporting each of these
317 processes must be able to function independently as well as in a coordinated fashion to deliver the
318 desired results whether laboratories select individual components or purchase the whole shooting
319 match. With respect to the latter, laboratories and manufacturers will need to form "partnerships" to
320 gain the level of trust, support and service necessary to keep the systems operating 24 hours/day.
321 Simple functions that we currently take for granted in daily operations become complex problems for
322 engineers in the design process such as handling unusual specimen types, movement of plates and
323 imaging within an incubator without contamination, processing anaerobic cultures, or even something
324 as simple as preparing a purity plate for AST. I could go on but imagine the process of setting up disk
325 diffusion plates by automation and you get the point.

326 Another issue that will certainly generate a fair degree of angst for system providers will be the
327 design of clinical trials to gain regulatory clearance for automated microbiology systems. What would
328 be the reference standard, manually processed cultures? What if the culture evaluation module of an
329 automated system provides additional sensitivity in terms of positive growth detection such that
330 specificity appears decreased relative to a manual methods? Should modules be evaluated individually

331 or as a complete system? These issues will have to be settled with regulatory agencies prior to the onset
332 of clinical trials.

333 To conclude, expanding automation in clinical microbiology is a certainty that will necessitate a
334 change in skill sets and expertise among microbiologists or laboratorians in the future. Individuals
335 coming out of clinical laboratory science programs at any degree level will need to accommodate these
336 changes. Other major issues that will need to be addressed with the automation of the microbiology
337 laboratory include overall capital outlay, minimizing contamination within the system (moulds,
338 mycobacteria, etc.), instrument maintenance and downtime, and in-house validation processes.

339 **Robin Patel:** Challenges include the varied nature of individual laboratories (sizes, physical layouts, test
340 menus, etc.), cost, and ever-evolving biology (including emerging antimicrobial resistance and
341 description of novel infectious agents). Although, chemistry laboratories have enjoyed the benefit of
342 full-laboratory automation for many years, microbiology laboratories have lagged behind because of
343 their multiple specimen types with varying viscosities and container types and sizes, and because of the
344 wide diversity of organisms detected. This variability has hindered the availability of full-laboratory
345 automation for clinical microbiology laboratories. Technology is rapidly changing and improving,
346 however, such that automation is increasingly used in clinical microbiology laboratories, and full-
347 laboratory automation is becoming possible. However, challenges remain. Platforms are not yet fully
348 developed. Laboratories that are early adopters may face availability of newer, “better” options in the
349 near future, and be “stuck” with “antiquated” platforms (analogous to the ever-changing mobile phone
350 or computer industry). There may be a perception on the part of laboratory technologists that they will
351 be “replaced by machines”. My personal perspective is that what will be mainly replaced will be the
352 uninteresting aspects of the job of the laboratory technologist. I find it hard to imagine a scenario where
353 laboratories would not need the expertise of experienced, well-trained microbiologists to oversee
354 microbiology testing.

355 ***What is the impact of automation in clinical microbiology on patient care?***

356 **Robin Patel:** This is an area of tremendous potential. Automation may provide expedited patient
357 results, facilitating prompt diagnosis and treatment and avoidance of unnecessary additional testing and
358 associated costs. (Although it would be intuited that automation would decrease turnaround time for
359 patient results, depending on the system design, turnaround time may be increased.) Automation may
360 increase the availability of results after-hours, necessitating new systems for acting on such results (i.e.,
361 at times when healthcare staff is not traditionally available). Automation may avoid falsely positive or
362 negative results associated with human error. Automated reading of culture plates may facilitate
363 recognition of colony growth not appreciated with the naked eye, enabling enhanced sensitivity and
364 earlier detection. Finally, technologic advances will likely bring testing closer to the patient; I predict that
365 some automated microbiology tests will move out of the laboratory, in a variety of formats, including
366 automated patient self-testing (for certain organisms).

367 **Gilbert Greub:** During the last ten years, several innovations in microbiology such as MALDI-TOF MS and
368 point of care molecular assays have had major impact on patient care, since these new diagnostic
369 approaches are significantly reducing the time to results. Automation in molecular diagnosis has also
370 contributed to a significant decrease in turnaround time. Automation in clinical microbiology will also
371 have some impact on patient care by improving traceability, reproducibility and quality. This is especially
372 true for automated inoculation system that helps much (i) in term of traceability and (ii) in getting high-
373 quality spreading of the inoculum over the agar plate, with a higher number of isolated colonies that
374 may be used for downstream identification and characterization. Thus, although the precise impact of
375 automation on patient care has not yet been precisely assessed, I am convinced that in addition to the
376 benefit of automation on the laboratory workflow and in avoiding repetitive uninteresting tasks, there is
377 also some benefit of automation on patient care.

378 **Author Acknowledgments, Disclosures or Potential Conflicts of Interest:**

379 **C.A. Burnham:** Has received research support from Cepheid, Accelerate Technology Corporation, and
380 bioMerieux.

381 **W.M. Dunne, Jr:** Employee of bioMerieux. The author would like to thank Dr. Kirk Doing for reviewing
382 these responses and suggesting more intelligent options.

383 **G. Greub:** No conflict of interest. Many thanks to A Croxatto, C Durussel and G. Prod'hom for helpful
384 discussion on automation.

385 **S. M. Novak:**

386 **R. Patel:** I would like to acknowledge the helpful suggestions of Dr. Patricia J. Simner (Clinical
387 Microbiology Fellow, Mayo Clinic), and Mayo Clinic's Microbiology Operations Managers, Mss. Emily A.
388 Vetter and Mary F. Jones.

389

390