

# NETWORK SELECTION MECHANISM FOR TELECARDIOLOGY APPLICATION IN HIGH SPEED ENVIRONMENT

YEW HOE TUNG

A thesis submitted in fulfilment of the  
requirements for the award of the degree of  
Doctor of Philosophy (Biomedical Engineering)

Faculty of Biosciences and Medical Engineering  
Universiti Teknologi Malaysia

JULY 2017

Specially dedicated to *my beloved parents and family*.

## **ACKNOWLEDGEMENT**

First of all, I would like to thank and express my deepest appreciation to my supervisors, Prof. Dr. Eko Supriyanto, Dr. Hau Yuan Wen and Dr. Muhammad Haikal Satria for their valuable guidance, support and personal kindness throughout my Ph.D study. Without their advice, this research would be very difficult to complete successfully.

I am deeply indebted to Ministry of Education Malaysia and Universiti Malaysia Sabah for the financial support. I would also like to thank the administration staff of Faculty of Biosciences and Medical Engineering for their support during my Ph.D study at Universiti Teknologi Malaysia.

Lastly, I would like to express my sincerest thanks to my beloved family for their continuous support and encouragement during my Ph.D study.

## ABSTRACT

The existing network selection schemes biased either to cost or Quality of Service (QoS) are not efficient enough for telecardiology application in high traveling speed environment. Selection of the candidate network that is fulfilling the telecardiology service requirements as well as user preference is a challenging issue. This is because the preference of telecardiology user might change based on the patient health condition. This research proposed a novel Telecardiology-based Handover Decision Making (THODM) mechanism that consists of three closely integrated algorithms: Adaptive Service Adjustment (ASA), Dwelling Time Prediction (DTP) and Patient Health Condition-based Network Evaluation (PHCNE). The ASA algorithm guarantees the quality of telecardiology service when none of the available networks fulfils the service requirements. The DTP algorithm minimizes the probability of handover failure and unnecessary handover to Wireless Local Area Network (WLAN), while optimizing the connection time with WLAN in high traveling speed environment. The PHCNE algorithm evaluates the quality of available networks and selects the best network based on the telecardiology services requirement and the patient health condition. Simulation results show that the proposed THODM mechanism reduced the number of handover failures and unnecessary handovers up to 80.0% and 97.7%, respectively, compared with existing works. The cost of THODM mechanism is 20% and 85.3% lower than the Speed Threshold-based Handover (STHO) and Bandwidth-based Handover (BWHO) schemes, respectively. In terms of throughput, the proposed scheme is up to 75% higher than the STHO scheme and 370% greater than the BWHO scheme. For telecardiology application in high traveling speed environment, the proposed THODM mechanism has better performance than the existing network selection schemes.

## ABSTRAK

Skim pemilihan rangkaian yang sedia ada mementingkan kos atau Kualiti Perkhidmatan (QoS) didapati tidak cukup berkesan untuk aplikasi telekardiologi dalam persekitaran kelajuan yang tinggi. Pemilihan rangkaian yang memenuhi keperluan perkhidmatan telekardiologi serta keutamaan pengguna adalah satu isu yang mencabar kerana keutamaan pengguna telekardiologi akan berubah berdasarkan keadaan kesihatan pesakit. Kajian ini mencadangkan mekanisme *Telecardiology-based Handover Decision Making* (THODM) yang terdiri daripada tiga algoritma: *Adaptive Service Adjustment* (ASA), *Dwelling Time Prediction* (DTP) and *Patient Health Condition-based Network Evaluation* (PHCNE). Algoritma ASA menjamin kualiti perkhidmatan telekardiologi apabila tiada rangkaian yang dapat memenuhi keperluan perkhidmatan telekardiologi. Algoritma DTP meminimumkan kebarangkalian kegagalan penyerahan dan penyerahan yang tidak perlu kepada Rangkaian Kawasan Tempatan Tanpa Wayar (WLAN) semasa mengoptimumkan masa sambungan dengan WLAN dalam persekitaran kelajuan yang tinggi. Algoritma PHCNE menilai kualiti rangkaian yang tersedia dan memilih rangkaian yang terbaik berdasarkan keperluan perkhidmatan telekardiologi dan keadaan kesihatan pesakit. Keputusan simulasi menunjukkan bahawa mekanisme THODM mengurangkan bilangan kegagalan penyerahan dan penyerahan yang tidak perlu sebanyak 80.0% dan 97.7% berbanding dengan kerja-kerja yang sedia ada. Kos bagi mekanisme THODM adalah 20% dan 85.3% lebih rendah daripada skim *Speed Threshold-based Handover* (STHO) dan *Bandwidth-based Handover* (BWHO). Dari segi pemprosesan, THODM adalah 75% dan 370% lebih tinggi daripada skim STHO dan BWHO. Aplikasi telekardiologi dalam persekitaran kelajuan yang tinggi, mekanisme THODM mempunyai prestasi yang lebih baik daripada skim pemilihan rangkaian yang sedia ada.

## TABLE OF CONTENTS

CHAPTER	TITLE	PAGE
	DECLARATION	ii
	DEDICATION	iii
	ACKNOWLEDGEMENT	iv
	ABSTRACT	v
	ABSTRAK	vi
	TABLE OF CONTENTS	vii
	LIST OF TABLES	xi
	LIST OF FIGURES	xii
	LIST OF ABBREVIATIONS	xv
	LIST OF SYMBOLS	xviii
	LIST OF APPENDICES	xxii
<b>1</b>	<b>INTRODUCTION</b>	<b>1</b>
	1.1 Research Motivation	1
	1.2 Problem Background	2
	1.3 Problem Statement	6
	1.4 Research Objectives	7
	1.5 Research Contributions	8
	1.6 Research Scopes	10
	1.7 Thesis Organization	11
<b>2</b>	<b>LITERATURE REVIEW</b>	<b>12</b>
	2.1. Introduction	12

2.2	Wireless Technologies	12
2.3	Network Coverage in Malaysia (Developing Country)	14
2.4	Wireless Telecardiology Systems	15
2.5	Handover in Heterogeneous Wireless Networks	18
	2.5.1 Background of Handover	18
	2.5.2 Handover Process	20
	2.5.3 Handover Criteria	22
	2.5.4 Handover Performance	24
2.6	Media Independent Handover (IEEE802.21)	25
2.7	Handover Decision Strategies	27
	2.7.1 RSS Based Handover Algorithm	28
	2.7.2 Function Based Handover Algorithm	33
	2.7.3 Multiple Criteria Based Handover Algorithm	41
	2.7.4 Intelligent Based Handover Algorithm	49
2.8	Comparison of Handover Decision Strategies	52
2.9	Summary	54
<b>3</b>	<b>RESEARCH METHODOLOGY</b>	<b>57</b>
3.1	Introduction	57
3.2	Research Framework	57
	3.2.1 Problem Investigation	59
	3.2.2 Design and Implementation	60
	3.2.2.1 Self-inspection Phase	61
	3.2.2.2 Pre-handover filtering Phase	62
	3.2.2.3 Network Selection Phase	65
	3.2.3 Performance Evaluation	65
	3.2.3.1 Experiment Setup for Dwelling Time Prediction Algorithm	66
	3.2.3.2 Experiment Setup for Patient Health Condition based Network Evaluation Algorithm	68

	3.2.3.3 Experiment Setup for Adaptive Service Adjustment Algorithm	69
	3.2.3.4 Simulation Parameters	70
	3.2.3.5 Results and Analysis	70
3.3	Summary	71
<b>4</b>	<b>ALGORITHMS DESIGN</b>	<b>72</b>
4.1	Introduction	72
4.2	Proposed Framework of Telecardiology System in Heterogeneous Wireless Network	72
4.3	Overview of THODM Mechanism	74
4.4	Self-inspection Phase	76
4.5	Pre-handover Filtering Phase	79
4.6	Network Selection Phase	88
4.7	Summary	90
<b>5</b>	<b>RESULT AND DISCUSSION</b>	<b>91</b>
5.1	Introduction	91
5.2	Performance of Adaptive Service Adjustment Algorithm	92
5.3	Performance of Dwelling Time Prediction Algorithm	95
	5.3.1 Root Mean Square Error	96
	5.3.2 Probability of Unnecessary Handover And Handover Failure	101
5.4	Performance of THODM Mechanism	108
	5.4.1 Number of Unnecessary Handovers	110
	5.4.2 Throughput	113
	5.4.3 Cost	114
5.5	Summary	116
<b>6</b>	<b>CONCLUSION</b>	<b>118</b>
6.1	Summary	118



6.2	Achievements	119
6.2.1	Adaptive Service Adjustment Algorithm	119
6.2.2	Dwelling Time Prediction Algorithm	120
6.2.3	Patient Health Condition Based Network Evaluation Algorithm	120
6.3	Future Work Directions	121
<b>REFERENCES</b>		<b>122</b>
Appendix A		134-135

## LIST OF TABLES

<b>TABLE NO.</b>	<b>TITLE</b>	<b>PAGE</b>
2.1	Comparison of wireless technologies	14
2.2	Percentage of wireless network coverage over population by three major network service providers in Malaysia [55-57]	14
2.3	Comparison of vertical and horizontal handover [48]	19
2.4	Summary of RSS based handover schemes	33
2.5	Comparative summary of function based handover algorithms	40
2.6	Saaty 1-9 fundamental scale [95]	46
2.7	Random Index [95]	48
2.8	A comparative summary of vertical handover schemes	56
3.1	Summary of research questions and solutions	60
3.2	Simulation parameters applied in the proposed algorithms	70
4.1	Inputs of THODM mechanism	74
4.2	Data rate required by different types of telecardiology services [38, 108]	75
5.1	Simulation parameters	93
5.2	Summary of performance comparison of Yan et al.[79], Hussain et al.[50] and the proposed DTP schemes	108
5.3	Grouping of relevant schemes	109

## LIST OF FIGURES

<b>FIGURE NO.</b>	<b>TITLE</b>	<b>PAGE</b>
2.1	A simple taxonomy of wireless technologies based on range, data rate and mobility support [54]	13
2.2	Maxis network coverage in Malaysia [58]	15
2.3	Architecture of Wireless Telecardiology System	16
2.4	Horizontal and vertical handover in heterogeneous wireless networks	19
2.5	Handover Management Concept	21
2.6	Media Independent Handover Function (MIHF)	27
2.7	Handover from WLAN to cellular network [76]	30
2.8	Scenario of prediction process in Hussain et al. [50] and Yan et al. [79]	32
2.9	Handover scheme proposed by Hong et al. [48]	34
2.10	Network environment in Park et al. [83]	37
2.11	Vertical handover algorithm in Dong and Maode [84]	38
2.12	TOPSIS analysis of distance to ideal and anti-ideal solution [92]	41
2.13	AHP system [95]	46
2.14	Fuzzy based handover scheme proposed by Kaleem et al. [49]	50
2.15	Architecture of VHM proposed by Nasser et al. [104]	52
3.1	Research methodology framework	58
3.2	Concept of proposed THODM mechanism	61
3.3	Block diagram of the proposed DTP algorithm	63
3.4	Dwelling time prediction scenario	64

3.5	Scenario of MT traveling trajectory in the micro cell coverage (OPNET Modelling)	67
3.6	Heterogeneous wireless network scenario	69
4.1	Proposed Telecardiology Framework	73
4.2	Block diagram of THODM mechanism	75
4.3	Flow chart of the self-evaluation process	77
4.4	Flow chart of pre-handover filtering process	80
4.5	MT traveling trajectory in WLAN cell coverage	81
4.6	Measured RSS value: (a) original RSS value and (b) averaged RSS value	84
4.7	Relationship between (a) the $d$ value and traveling distance $l$ and (b) $d_{th}$ , acceleration and velocity	87
4.8	Flow chart of network selection phase	89
5.1	Network selected by MT	94
5.2	Comparison of $DR_{REQ}$ and data rate achieved by BWHO based handover scheme ( $DR_{BWHO}$ )	94
5.3	Comparison of $DR_{REQ}$ and data rate achieved by BWHO+ASA scheme ( $DR_{BWHO+ASA}$ )	95
5.4	RSS curve generated using OPNET Modeler	96
5.5	Estimated and actual $R$ and $r$ radius values based on the predefined RSS thresholds	99
5.6	Comparison between estimated traveling distance and actual traveling distance, $D$ , within the WLAN coverage	99
5.7	Comparison between estimated traveling time and actual traveling time, $T_{WLAN}$ , within the WLAN coverage	100
5.8	RMSE of the estimated traveling distance based on the RSS monitoring time interval of 100 ms, 50 ms, and the proposed method that is dynamically adjusted according to the traveling speed of MT	101
5.9	Total number of handovers to the WLAN based on: (a) handover failure threshold and (b) unnecessary handover threshold	104

5.10	Number of (a) handover failures and (b) unnecessary handovers	105
5.11(a)	Ratio of handover failures to total number of handovers	106
5.11(b)	Ratio of unnecessary handovers to the total number of handovers	107
5.12	Number of handovers that occurred in BWHO, THODM and ideal solution	112
5.13	Number of unnecessary handovers that occurred in THODM and BWHO schemes	112
5.14	Average throughput achieved by MT in single loop at the speed of 11.11 m/s to 38.88 m/s	115
5.15	Percentage of throughput gain of THODM over STHO and BWHO schemes	115
5.16	Average cost per Mb achieved by THODM, BWHO and STHO mechanisms	116

## LIST OF ABBREVIATIONS

2G	-	Second Generation Wireless Telephone Technology
3G	-	Third Generation Wireless Telephone Technology
4G	-	Fourth Generation Wireless Telephone Technology
AHP	-	Analytic Hierarchy Process
ANN	-	Artificial Neural Networks
AP	-	Access Point
ASA	-	Adaptive Service Adjustment
BS	-	Base Station
C	-	Cost per Mb
CI	-	Consistency Ratio
CNN	-	Current Connect Network
CNs	-	Corresponding Nodes
CR	-	Consistency Ratio
CVD	-	Cardiovascular Disease
DR	-	Data Rate
DTP	-	Dwelling Time Prediction
e-health	-	Electronic health
ECG	-	Electrocardiogram
EP	-	Ending Point
IEEE	-	Institute of Electrical and Electronics Engineers
FAHP	-	Fuzzy Analytic Hierarchy Process
FIS	-	Fuzzy Inference System
GPRS	-	General Packet Radio Service
GSM	-	Global System for Mobile Communications
HA	-	Home Agent
HMIPv6	-	Hierarchical Mobile IP version 6

LTE	-	Long Term Evolution
m-health	-	Mobile health
MAC	-	Medium Access Control
MADM	-	Multiple Attributes Decision Making
Mb	-	Megabit
MICS	-	Media Independent Command Service
MIES	-	Media Independent Event Service
MIH	-	Media Independent Handover
MIHF	-	Media Independent Handover Function
MIIS	-	Media Independent Information Service
MoH	-	Ministry of Health
MT	-	Mobile Terminal
NHO	-	Number of handover
NUHO	-	Number of unnecessary handover
PCI	-	Percutaneous Coronary Intervention
PD	-	Packet Delay
PHC	-	Patient Health Condition
PHCNE	-	Patient Health Condition based Network Evaluation
QNL	-	Qualify Network List
QoS	-	Quality of Service
RI	-	Random Consistency Index
RMSE	-	Root Mean Square Error
RSS	-	Received Signal Strength
SAW	-	Simple Additive Weighting
SINR	-	Signal to Interference Plus Noise Ratio
SNR	-	Signal to Noise Ratio
SP	-	Starting Point
STHO	-	Speed Threshold Based Handover Scheme
THODM	-	Telecardiology based Handover Decision Making
TOPSIS	-	Technique of Order Preference by Similarity to Ideal Solution
$T_{WLAN\_1}$	-	Traveling time within the WLAN_1 coverage
$T_{WLAN\_2}$	-	Traveling time within the WLAN_2 coverage
$T_{WLAN\_3}$	-	Traveling time within the WLAN_3 coverage

$T_{\text{WLAN}_4}$	-	Traveling time within the $\text{WLAN}_4$ coverage
UMTS	-	Universal Mobile Telecommunications System
VHOM	-	Vertical Handover Manager
WBAN	-	Wireless Body Area Network
WiMAX	-	Worldwide Interoperability for Microwave Access
WLAN	-	Wireless Local Area Network
WMAN	-	Wireless Metropolitan Area Network
WPAN	-	Wireless Persona Area Network
WWAN	-	Wide Wireless Area Network



## LIST OF SYMBOLS

$\Delta$	- Minimum value of $d$
$\Delta t_p$	- Time needed by the MT to cover the two RSS sample points
$\varepsilon$	- Zero-mean Gaussian random variable caused by shadow fading
$\Gamma_{BS}$	- Channel coding loss factor
$\mathcal{B}$	- Best network candidate
$\rho$	- RSS sampling time interval
$\Gamma_{BS}$	- Channel coding loss factor
$\lambda_{max}$	- The largest eigenvector
$c$	- Acceleration
$C$	- Cost per bit
$C_{cellular}$	- Average cost per Mb offered by cellular network
$C_k$	- Cost per Mb of network candidate $k$
$C_{HO}$	- Handover cost
$C_{wlan}$	- Average cost per Mb offered by WLAN
$C_{SINR}$	- cost per SINR
$d$	- Traveling distance from WLAN boundary ( $P_{entry}$ ) to predefined RSS threshold ( $P_{In\_RSSth}$ )
$d_0$	- Distance between the AP and a reference point
$d_m$	- Maximum $d$ value
$d_{th}$	- Threshold of $d$ value
$d_{thf}$	- Handover failure threshold of $d$ value
$d_{thu}$	- Unnecessary handover threshold of $d$ value
$D$	- Actual traveling distance within the WLAN coverage
$D_S$	- Sampling distance
$DR_{REQ}$	- data rate requirement

$E$	- Path loss exponent
$E_{SNR_{UMTS}}$	- An equivalent SNR value in WLAN
$G$	- Percentage of throughput gain
$h$	-  AP y coordinate – MT y coordinate
$H$	- hysteresis value
$H_{default}$	- predefined default hysteresis values
$H_{maximum}$	- predefined maximum hysteresis values
$H_{minimum}$	- predefined minimum hysteresis values
$k$	- Network candidate
$l$	- Estimate traveling distance within WLAN coverage
$l_{th}$	- Estimate traveling distance threshold
$l_{thf}$	- $l_{th}$ for handover failure
$l_{thu}$	- $l_{th}$ for unnecessary handover
$L_{th}$	- Distance threshold by Hussain <i>et al.</i> 2013
$L_{thfH}$	- $L_{th}$ of handover failure presented by Hussain <i>et al.</i> 2013
$L_{thfY}$	- $L_{th}$ of handover failure presented by Yan <i>et al.</i> 2008
$L_{thuH}$	- $L_{th}$ of unnecessary handovers presented by Hussain <i>et al.</i> 2013
$L_{thuY}$	- $L_{th}$ of unnecessary handovers presented by Yan <i>et al.</i> 2008
$n$	- Number of available network
$N$	- Number of samples
$N_l$	- Normalized network load value
$N_v$	- Normalized velocity value
$NUHO$	- Number of unnecessary handover
$NHO_{ideal}$	- Number of handovers achieved by ideal solution
$P_{entry}$	- Entry point of the micro cell coverage boundary
$P_f$	- Probability of failure
$P_{In\_RSS_{th}}$	- Entry point of the micro cell coverage at $RSS_{th}$
$P_{Out\_RSS_{th}}$	- Exit point of the micro cell coverage at $RSS_{th}$
$P_s$	- Point MT collects the second RSS sample
$P_{TX}$	- AP transmit power
$P_u$	- Probability of unnecessary handover
$P_{exit}$	- Exit point of the micro cell coverage boundary
$PL_0$	- Power loss at the reference point

$Q_{CCN}$	- Quality of current connected network
$Q_k$	- Quality of network candidate $k$
$r$	- Radius of WLAN from AP to $RSS_{th}$
$R$	- Radius of WLAN from AP to coverage boundary
$R_{cellular}$	- Average data rate of UMTS
$R_{wlan}$	- Average data rate of WLAN
$RSS_{CCN}$	- RSS of current connected network
$RSS_{new}$	- RSS of target network
$RSS_{old}$	- RSS of target network and current connected network
$RSS_{P_{In\_RSSth}}$	- RSS value measured at points $P_{In\_RSSth}$
$RSS_{P_{entry}}$	- RSS value measured at points $P_{entry}$
$RSS_{th}$	- Predefined RSS threshold
$RSS_{thCCN}$	- Predefined RSS threshold of current connect network
$SNR_{CCN}$	- SNR of current connected network
$SNR_{REQ}$	- Dynamic SNR threshold defined based on the sum of the data rate required by the telecardiology services applied by the user
$SNR_{REQ\_WLAN}$	- $SNR_{REQ}$ of WLAN
$SNR_{REQ\_CCN}$	- $SNR_{REQ}$ of current connect network
$SNR_{REQ\_UMTS}$	- $SNR_{REQ}$ of UMTS
$SNR_{UMTS}$	- UMTS SNR value
$t_{cellular}$	- total time connected to UMTS
$t_d$	- time taken by the MT to travel from $P_{entry}$ to $P_{In\_RSSth}$
$t_e$	- Time of the MT passes through $P_{entry}$
$t_{P_{In\_RSSth}}$	- Time MT pass through point $P_{In\_RSSth}$
$t_{P_s}$	- Time MT pass through point $P_s$
$t_R$	- Time of the MT passes through $P_{In\_RSSth}$
$t_{wlan}$	- Total time connected to WLAN
$T_i$	- handover latency from macro cell to micro cell
$T_m$	- RSS monitoring time interval
$T_o$	- handover latency from micro cell to macro cell
$T_{Throughput}$	- Total throughput
$T_{UHO}$	- Time consumed by each unnecessary handover
$T_{WLAN}$	- Traveling time within WLAN

$T_{WLAN\_th}$	- Threshold of traveling time within WLAN
$T_s$	- RSS sampling time
$TP_{THODM}$	- Total throughput achieved by THODM
$TP_x$	- Total throughput of RSS or Cellular based scheme
$U_{QNL}$	- Number network candidate in QNL
$v$	- MT velocity
$v_e$	- MT velocity at $P_{entry}$
$v_R$	- MT velocity at $P_{In\_RSSth}$
$w$	- Weight of the parameter
$W_{BS}$	- Network carrier bandwidth
$W_{CCN}$	- Channel bandwidth (Hz) of current connected network
$W_{UMTS}$	- Channel bandwidth (Hz) of UMTS network
$W_{WLAN}$	- Channel bandwidth (Hz) of WLAN network
$z$	- Distance between two RSS sample point by using Yan <i>et al.</i> and Hussain <i>et al.</i> methods

**LIST OF APPENDICES**

<b>APPENDIX</b>	<b>TITLE</b>	<b>PAGE</b>
A	List of Publications	131

## **CHAPTER 1**

### **INTRODUCTION**

#### **1.1 Research Motivation**

Cardiovascular disease (CVDs) is one of the world leading causes of non-communicable disease resulting in death. According to World Health Organization (WHO), it is estimated that 17.5 million people died from CVDs each year and over 80% of people died from CVDs occurred in low and medium income countries [1]. CVD also the main cause of death in Malaysia [2].

The scarcity of cardiologist is a common issue to both underdeveloped and developing countries including Malaysia. The density of cardiologist in Malaysia is 1 per 120,481 people [3]. This is significantly inadequate, as Royal College of Physicians and British Cardiac Society have recommended one cardiologist per 50,000 people [4].

The issue of lack of specialists has been highlighted by Minister of Health and the President of Malaysia Medical Association [5]. Although large number of doctor are graduating annually, the rural areas in Sabah and Sarawak still suffer from shortage of doctors and specialists [6, 7]. The lack of specialist care for the rural people is a major concern, and has been a major concern for a long time.

In order to overcome the barrier of healthcare services such as distance, time, cost, and effectiveness of health care services, deployment of mobile telecardiology system is necessary. Research has proven that telecardiology is useful in the management of acute coronary syndrome, atrial fibrillation, syncope and even in implementing strategies of cardiovascular primary care [8]. Paus *et al.* [9] proved that telecardiology is feasible for cardiovascular patients living in remote region and led to improve follow up rates. Other benefits of telecardiology are improvement in patient compliance with medication, reduced mortality and morbidity, better utilization of resources such as ambulance, tertiary beds, less travel, reduced treatment delay, improvement in hospitals and clinical administration workflow, and is more cost-effective [10-19]. By employing telecardiology, areas with less developed healthcare facilities, for example, rural areas in East Malaysia, can capitalize on the expertise available in the West Coast Peninsular by centrally managing patients across the nation through telecardiology. This would help to improve the overall healthcare quality across the whole country.

## **1.2 Problem Background**

Telecardiology deals with the transmission of electrocardiogram (ECG) signal, vital signs, still image and video over different communication technologies from home or clinics to cardiology specialty centres [20]. It can be operated in two transmission modes, store-and-forward and real-time. In store-and-forward mode the user sends the pre-recorded health information (e.g. ECG data, vital signs) to hospital information system and the healthcare professionals check the patients' electronic health record some time later. In this application, the sender and the receiver do not have to be present simultaneously. On the other hand, real-time telecardiology requires both parties (patient and healthcare professionals) to be present at the same time. The example of real-time telecardiology application is the transmission of ECG data from ambulance to the hospital emergency department. Store-and-forward has been widely used to monitor or follow-up the patients that have chronic cardiac disease due to its expediency. However, real-time telecardiology service is crucial for

the patient in critical condition. It could help patients to get treatment promptly. It also be a strong educational component for the remote practitioners [21].

Most of the existing telecardiology systems are relying on a single wireless technology. These systems are unable to guarantee that users remain continuously connected to the telecardiology service provider due to the imperfection of network coverage and poor network quality. A disconnection of a link will disrupt the telecardiology services and lead to misdiagnosis from healthcare professional. The study of Bergrath *et al.* [22] presented that only 73% of the still images had good signal quality that was helpful for teleconsultation. The reasons of failed transmission or transmission problems were attributed to connection breakdown and poor network condition [23].

In this research, a wireless telecardiology system integrated with heterogeneous wireless technologies is proposed. The proposed system is able to access to different wireless technologies based on the telecardiology services' requirement to maintain the quality of telecardiology service at the highest level. Moreover, the heterogeneous networks based telecardiology system will have larger service coverage compared to the system that relies on single wireless technology.

In terms of wireless technologies for telecardiology applications, high bandwidth wireless technologies are more desirable to guarantee the quality of telecardiology services. Wireless Local Area Network (WLAN) is the most preferable due to the high transmission capacity and low network access cost. However, WLAN based telecardiology system is typically for indoor application (home, clinic and hospital) due to its small network coverage [24-27]. For overcoming the network coverage issue, Worldwide Interoperability for Microwave Access (WiMAX) has been proposed for telecardiology application due to its large bandwidth and ability to support high mobility [28, 29]. The work in [28] was then further extended by providing telecardiology services over an integration of WLAN and WiMAX networks to minimize the cost of service [30]. Moreover, similar work about integration of WLAN and WiMAX networks for telecardiology application is presented in [31]. Generally, both Niyato *et al.* [30] and Yan and Tsunoda [31] proposed WLAN for short range communication within building such as mall, clinic,



hospital and home. On the other hand, WiMAX is used to support the patient at outdoor environment. However, the detail of handover decision between WLAN and WiMAX is not discussed by authors. In addition, the use of WiMAX in telecardiology is limited as major network service providers are ceasing development of WiMAX [32].

Meanwhile, the rapid growth of mobile cellular network and smartphone technologies has boosted the research on mobile health (m-health). Cellular network based telecardiology system is presented in [33-37]. The advantages of cellular based telecardiology system are that it supports high mobility and offers large service coverage. The main drawback of using cellular network is insufficient network capacity to support real-time high quality diagnosis video transmission which requires at least 4 Mbps data transmission rate [38]. The high bandwidth Fourth Generation Long Term Evolution (4G-LTE) system is still under deployment and the coverage of 4G-LTE is imperfect especially in the rural area of underdeveloped and developing countries. Thus, the use of 4G-LTE in telecardiology application for the rural people in the underdeveloped and developing countries is limited.

The seamless handover management is a challenging issue in heterogeneous wireless networks. To provide seamless handover in heterogeneous networks, IEEE organization had approved a new standard in 2008 named IEEE802.21, Media Independent Handover (MIH) which enables handover in heterogeneous network with no perceivable interruption to an on-going voice or video conversation [39]. MIH provides link configuration, radio measurement reporting, new link discovery, and resource availability check. However, it does not include handover decision making which is the heart of the handover process [39, 40]. Handover decision is crucial for selecting the correct target and triggering handover at the right time so that the quality of the telecardiology services can be maintained at the acceptable level.

The existing vertical handover decision making across heterogeneous networks can be categorized into four groups, received signal strength (RSS), function, multi attributes and intelligent. RSS-Threshold based handover scheme is the oldest method which selects the best network station based on single RSS

criterion. It is typically applied in intra-network because RSS of different wireless technologies cannot be compared directly. In vertical handover, the RSS-based handover scheme uses RSS criterion for predicting the distance between Mobile Terminal (MT) and Base Station (BS) or Access Point (AP). It also can be used to determine the MT moving direction either closer to or farther from BS.

The function based handover decision making algorithm combines few parameters such as bandwidth and cost in a function. In this function, different weights are assigned to each parameter depending on the user preference [39]. The candidate network that has the highest weight sum is considered as the best network. The most common function-based handover methods are Quality-of-Service (QoS) based and cost-based. The QoS-based method always chooses the network with the higher throughput in order to keep the service quality at the highest level. This scheme has high handover rate because it always triggers handover to higher bandwidth network whenever available. The cost-based algorithm selects the handover target based on the cost of the available networks. Typically, the wireless network with the lowest cost is the most preferred. The advantage of cost based scheme is that it improves the user satisfaction level by reducing the users' financial expenses, but the quality of service (QoS) or user mobility will be sacrificed.

Multiple Attributes Decision Making (MADM) scheme is mainly for selecting the best target based on the contribution of various handover criteria such as bandwidth, speed, power, security and cost. MADM strategy is a weighting system where different weights are assigned depending on the priority level of each criterion. The network candidate which has the highest weight sum will be considered as a target network. The main disadvantage of this strategy is that the algorithm complexity will be increased when more parameters are taken into account [39]. Consequently, the handover latency of MADM is higher compared to the RSS based as well as the function-based schemes that require less handover criteria. In addition, the weight assignment on input parameters of the existing MADM schemes is mostly done manually based on certain scenarios. This could lead to a degradation of service quality when different services are applied by the user or with the occurrence of unusual situations.

The fluctuation of network parameters such as RSS will lead to unnecessary handover and ping-pong effect [39, 40]. Ping-pong effect is a scenario of repeated handovers between two networks in a short period of time [41]. It usually occurs when the mobile terminal is within the overlap region of two different networks. Ping-pong effect is undesired because it wastes the network resources, consumes power, and leads to signal transmission degradation [42]. To prevent ping-pong effect, intelligent-based handover decision making schemes such as Fuzzy Logic and Artificial Neural Networks (ANN) have been applied in vertical handover process. Intelligent-based schemes are more reliable compared to other approaches. It has high success rate in connecting to the best network and in minimizing the ping-pong effect. However, this scheme has high handover latency caused by ANN learning/training computation, and/or Fuzzy Logic fuzzification or defuzzification processes [40].

### 1.3 Problem Statement

The main problems of the existing vertical handover decision making algorithms for telecardiology application in heterogeneous wireless networks are:

- i. The existing handover algorithms connect to the preferred or the best candidate network even though this network has insufficient capacity to fulfil the telecardiology service requirements such as real-time video, audio, ECG and vital sign transmission. For example, the speed threshold based handover algorithm [43] selects the macro cell when the mobile terminal traveling speed is above the predefined speed threshold, even though the capacity of the macro cell is insufficient for telecardiology service requirement. An insufficient network capacity may cause loss of audio fidelity, choppy video and packet loss which can lead to misdiagnosis from healthcare professional.
- ii. The existing algorithms do not optimize the utilization of WLAN when MT is traveling at high-speed. Most of the existing handover schemes define a

traveling speed threshold for WLAN. MT triggers handover to WLAN if and only if the traveling speed is below the predefined threshold value to avoid the handover failure and unnecessary handover to WLAN. The application of WLAN is restricted to static or pedestrian navigation environment [44]. For example, researchers in [45-48] predefined the traveling speed threshold for WLAN at 10m/s and below. In addition, Fuzzy MADM based handover algorithm presented by Kaleem *et al.* [49] set the fuzzy if-else rule, “if MT velocity is low, then the probability of rejecting WLAN is low; otherwise, the probability of rejecting WLAN is high”.

- iii. Existing handover algorithm cannot fully satisfy the telecardiology user because most of them are biased either to QoS or cost. None of them are sensitive to patient health conditions. There is a trade-off between QoS and cost. In telecardiology application, the cost-based network selection scheme is suitable for stored-and-forward mode which is widely applied to the user in non-critical health monitoring because stored-and-forward mode does not require high service quality. However, cost based network selection scheme is inappropriate for real-time critical health monitoring application which desires a high service quality. For real-time critical health monitoring, QoS-based network selection scheme is more suitable. Therefore, it is important to have a handover algorithm that is responding to the patient health condition that dynamically selects the best QoS network and the lowest cost network for the case of store-and-forward non-critical health monitoring and real-time critical health monitoring, respectively.

## 1.4 Research Objectives

The aim of this research is to propose a novel handover scheme that is able to respond according to the patient health conditions and also optimize the quality of telecardiology service by minimizing unnecessary handover and handover failure rate while maximizing the connection time to the WLAN in high traveling speed

scenario. The following research objectives are set to achieve the aim of this research work.

- i. To design an Adaptive Service Adjustment (ASA) algorithm for guaranteeing telecardiology service quality when there is insufficient network capacity to fulfil the telecardiology service requirements.
- ii. To design a novel Dwelling Time Prediction (DTP) algorithm for minimizing the probability of handover failures and unnecessary handovers while maximizing the connection time to WLAN at high traveling speed scenario.
- iii. To design a Patient Health Condition based Network Evaluation (PHCNE) algorithm for improving the telecardiology user's satisfaction in terms of throughput and cost.
- iv. To evaluate the performance of the proposed ASA, DTP and PHCNE algorithms and benchmark with existing relevant algorithms.

## **1.5 Research Contributions**

This thesis presents a Telecardiology based Handover Decision Making (THODM) mechanism for telecardiology application in heterogeneous networks. In general, THODM mechanism resolves the problems such as limited coverage, insufficient network capacity and mobility issues faced by the existing telecardiology systems. The detail of research contributions are stated as below:

- i. The Adaptive Service Adjustment (ASA) algorithm reduces the user bandwidth requirement by deactivating the telecardiology service at the lowest priority level while the best network has insufficient capacity to support the telecardiology service requirements. This priority deactivation technique allocates more bandwidth for supporting the telecardiology service at higher priority level. As a result, the quality of the telecardiology services

at higher priority level is still maintained at the acceptable level. This algorithm guarantees the telecardiology service and prevents misdiagnosis from healthcare professional.

- ii. The Dwelling Time Prediction (DTP) algorithm predicts the traveling time within the micro cell (WLAN Hotspot) coverage for the MT which is moving in dynamic speed. It overcomes the limitation of the existing prediction scheme [50] that assumes the MT is moving at a constant speed. The proposed DTP algorithm is more suitable for mobile telecardiology application such as transmitting health signal or data from high-mobility-ambulance to hospital, because the traveling speed of ambulance may not be constant at all time.
- iii. The DTP algorithm minimizes the number of handover failures and unnecessary handovers while optimizing the utilization of micro cell in high speed scenario. With the proposed DTP algorithm, MT can benefit high bandwidth and low network access cost from the micro cell network with minimum interruption in high speed scenario. This can improve the quality of telecardiology service and the cost effectiveness.
- iv. The Patient Health Condition based Network Evaluation (PHCNE) algorithm ensures the quality of the connected or targeted network fulfils the telecardiology service requirements. In critical health condition, the highest bandwidth network will be selected by this algorithm to optimize the quality of the telecardiology services. On the other hand, the lowest cost network that fulfils the telecardiology service requirement will be selected if the patient is in normal health condition. This is to minimize the financial expenses of telecardiology user.

## 1.6 Research Scopes

In this research, only wireless networks are considered because wireless technologies support mobility and have larger coverage compared to wired technologies. In addition, WiMAX technology is excluded from this research because the major network service providers are ceasing development of WiMAX and focusing on cellular network [32]. The scopes of this research are:

- i. This research considers the high mobility scenario. The MT is an ambulance equipped with a telecardiology system.
- ii. The mobile telecardiology is powered by an external power supply from vehicle. Therefore, power issue is not considered in this research.
- iii. The MT is traveling at a speed in the range of 10 m/s to 40 m/s (36 km/h to 144 km/h) on the highway.
- iv. It is assumed that the network services providers reserve certain number of network channels at each base station (BS) and access point (AP) for telecardiology purpose [30]. The AP is the WLAN Hotspot provided by network services providers and they have identical network characteristics. The reserved channels guarantee data transmission rate of 1 Mbps and 6 Mbps for cellular network and WLAN, respectively.
- v. The research is focusing on the rural areas of underdeveloped and developing countries. The telecommunication infrastructure and healthcare facilities in these areas are still imperfect.
- vi. The telecardiology user's satisfaction in this research is measured in terms of throughput and cost. The lower the cost per Megabit (Mb), the higher the user satisfaction.
- vii. It is assumed that the Doppler shift problem caused by high mobility can be mitigated by using the Doppler diversity. In [51], Doppler domain multiplexing communication structure is proposed to achieve the maximum Doppler diversity in time varying fading. Furthermore, the Doppler frequency

offset estimation and compensation algorithms presented in [52, 53] can be used to alleviate the Doppler effects in high-speed scenario.

## **1.7 Thesis Organization**

This thesis comprises of six chapters. The remaining of the chapters are organized as follows:

Chapter 2 provides the extensive review of the existing wireless telecardiology systems, background of handover process and followed by a comprehensive survey of vertical handover decision making algorithms.

Chapter 3 describes research methodology, experiment design and verification of proposed Adaptive Service Adjustment (ASA) algorithm, Dwelling Time Prediction (DTP) algorithm and Patient Health Condition based Network Evaluation (PHCNE) algorithm.

Chapter 4 provides the framework of telecardiology system in heterogeneous wireless network. This chapter also explains the details of the proposed Adaptive Service Adjustment (ASA) algorithm, Dwelling Time Prediction (DTP) algorithm and Patient Health Condition based Network Evaluation (PHCNE) algorithm.

Chapter 5 shows the simulation results. The performance of the proposed algorithms evaluated against the performance of the existing relevant schemes.

Chapter 6 concludes the research work in this thesis and suggests possible future research directions.



## REFERENCES

1. Cardiovascular diseases (CVDs). January 2015 [cited 2017 10 January]; Available from: <http://www.who.int/mediacentre/factsheets/fs317/en/>.
2. Health Facts 2016. [Retrieved 20 October 2016]. Available from: [http://vlib.moh.gov.my/cms/documentstorage/com.tms.cms.document.Document\\_4ad54b25-a0188549-187fdf50-1006a4e5/Health%20Facts%202016.pdf](http://vlib.moh.gov.my/cms/documentstorage/com.tms.cms.document.Document_4ad54b25-a0188549-187fdf50-1006a4e5/Health%20Facts%202016.pdf).
3. National Healthcare Establishment & Workshop Statistics (NHEWS) 2012-2013. [Retrieved Available from: <http://www.crc.gov.my/nhsi/national-healthcare-establishment-workforce-statistics-hospital-2012-2013/>.
4. Fifth report on the provision of services for patients with heart disease. *Heart*. 2002. 88 Suppl 3: iii1-56.
5. Chin, C. (2013, 18 August). Too Many Doctors, Too Little Training. *The Star*, Retrieved 20 August, 2013, Available from: <http://www.thestar.com.my/News/Nation/2013/08/18/Too-many-doctors-too-little-training.aspx/>
6. Reporter. (2013, 30 April). Government provides more physicians for better healthcare of Malaysians. *New Sarawak Tribune*, Retrieved 20 September, 2013, Available from: <http://www.newsarawaktribune.com/news/4175/Government-provides-more-physicians-for-better-healthcare-of-Malaysians/>.
7. Ringgit, D.S. (2015, 22 April). More options needed to address doctor shortage in Sarawak. *The Borneo Post*, Retrieved 23 April, 2015, Available from: <http://www.theborneopost.com/2015/04/22/more-options-needed-to-address-doctor-shortage-in-sarawak/>.

8. Backman, W., Bendel, D., and Rakhit, R. The telecardiology revolution: improving the management of cardiac disease in primary care. *J R Soc Med.* 2010. 103(11): 442-6.
9. PausJenssen, A.M., Spooner, B.A., Wilson, M.P., and Wilson, T.W. Cardiovascular risk reduction via telehealth: a feasibility study. *Can J Cardiol.* 2008. 24(1): 57-60.
10. Brunetti, N.D., De Gennaro, L., Dellegrottaglie, G., Procacci, V., and Di Biase, M. Fast and furious: Telecardiology in acute myocardial infarction triage in the emergency room setting. *European Research in Telemedicine / La Recherche Européenne en Télémédecine.* 2013. 2(2): 75-78.
11. Sorensen, J.T., Clemmensen, P., and Sejersten, M. Update: Innovation in cardiology (II). Telecardiology: past, present and future. *Rev Esp Cardiol (Engl Ed).* 2013. 66(3): 212-8.
12. Chaudhry, S.I., Barton, B., Mattera, J., Spertus, J., and Krumholz, H.M. Randomized trial of Telemonitoring to Improve Heart Failure Outcomes (Tele-HF): study design. *J Card Fail.* 2007. 13(9): 709-14.
13. Brown, J.P., Mahmud, E., Dunford, J.V., and Ben-Yehuda, O. Effect of prehospital 12-lead electrocardiogram on activation of the cardiac catheterization laboratory and door-to-balloon time in ST-segment elevation acute myocardial infarction. *Am J Cardiol.* 2008. 101(2): 158-61.
14. Brunetti, N.D., De Gennaro, L., Amodio, G., Dellegrottaglie, G., Pellegrino, P.L., Di Biase, M., and Antonelli, G. Telecardiology improves quality of diagnosis and reduces delay to treatment in elderly patients with acute myocardial infarction and atypical presentation. *European Journal of Cardiovascular Prevention & Rehabilitation.* 2010. 17(6): 615-620.
15. Maarop, N. and Win, K. Understanding the Need of Health Care Providers for Teleconsultation and Technological Attributes in Relation to The Acceptance of Teleconsultation in Malaysia: A Mixed Methods Study. *Journal of Medical Systems.* 2012. 36(5): 2881-2892.
16. Adeogun, O., Tiwari, A., and Alcock, J.R. Models of information exchange for UK telehealth systems. *Int J Med Inform.* 2011. 80(5): 359-70.
17. Brunetti, N.D., Dellegrottaglie, G., Lopriore, C., Di Giuseppe, G., De Gennaro, L., Lanzone, S., and Di Biase, M. Prehospital telemedicine electrocardiogram triage for a regional public emergency medical service: Is

- it worth it? A preliminary cost analysis. *Clinical cardiology*. 2014. 37(3): 140-145.
18. Leahy, R.A. and Davenport, E.E. Home Monitoring for Cardiovascular Implantable Electronic Devices: Benefits to Patients and to Their Follow-up Clinic. *AACN advanced critical care*. 2015. 26(4): 343-355.
  19. Landolina, M., Perego, G.B., Lunati, M., Curnis, A., Guenzati, G., Vicentini, A., Parati, G., Borghi, G., Zanaboni, P., Valsecchi, S., and Marzegalli, M. Remote Monitoring Reduces Healthcare Use and Improves Quality of Care in Heart Failure Patients With Implantable Defibrillators Clinical Perspective. *The Evolution of Management Strategies of Heart Failure Patients With Implantable Defibrillators (EVOLVO) Study*. 2012. 125(24): 2985-2992.
  20. Newton, M.J. The promise of telemedicine. *Survey of Ophthalmology*. 2014. 59(5): 559-567.
  21. Wootton, R. Realtime telemedicine. *J Telemed Telecare*. 2006. 12(7): 328-36.
  22. Bergrath, S., Rortgen, D., Rossaint, R., Beckers, S.K., Fischermann, H., Brokmann, J., Czaplik, M., Felzen, M., Schneiders, M.T., and Skorning, M. Technical and organisational feasibility of a multifunctional telemedicine system in an emergency medical service - an observational study. *J Telemed Telecare*. 2011. 17(7): 371-7.
  23. Sejersten, M., Sillesen, M., Hansen, P.R., Nielsen, S.L., Nielsen, H., Trautner, S., Hampton, D., Wagner, G.S., and Clemmensen, P. Effect on treatment delay of prehospital teletransmission of 12-lead electrocardiogram to a cardiologist for immediate triage and direct referral of patients with ST-segment elevation acute myocardial infarction to primary percutaneous coronary intervention. *Am J Cardiol*. 2008. 101(7): 941-6.
  24. Tejero-Calado, J.C., Lopez-Casado, C., Bernal-Martin, A., Lopez-Gomez, M.A., Romero-Romero, M.A., Quesada, G., Lorca, J., and Rivas, R. IEEE 802.11 ECG monitoring system. *Engineering in Medicine and Biology Society, 2005. IEEE-EMBS 2005. 27th Annual International Conference of the*. 17-18 Jan. 2006. Shanghai, China: IEEE.2005. 7139-7142.
  25. Moein, A. and Pouladian, M. WIH-Based IEEE 802.11 ECG Monitoring Implementation. *Engineering in Medicine and Biology Society, 2007. EMBS*

2007. *29th Annual International Conference of the IEEE*. 22-26 Aug. 2007. Lyon, France.2007. 3677-3680.
26. Sakthi, P.A. and Sukanesh, R. A reliable and fast routing data transmission protocol for Wi-Fi based real-time patient monitoring system. *Electronics and Communication Systems (ICECS), 2014 International Conference on*. 13-14 Feb. 2014. Tamilnadu, India: IEEE.2014. 1-5.
  27. Cai, K. and Liang, X. Development of WI-FI Based Telecardiology Monitoring System. *Intelligent Systems and Applications (ISA), 2010 2nd International Workshop on*. 22-23 May 2010.2010. 1-4.
  28. Niyato, D., Hossain, E., and Diamond, J. IEEE 802.16/WiMAX-based broadband wireless access and its application for telemedicine/e-health services. *Wireless Communications, IEEE*. 2007. 14(1): 72-83.
  29. Chorbev, I., Mihajlov, M., and Jolevski, I. WiMAX supported telemedicine as part of an integrated system for e-medicine. *Information Technology Interfaces, 2008. ITI 2008. 30th International Conference on*. 23-26 June 2008. Cavtat/Dubrovnik, Croatia: IEEE.2008. 589-594.
  30. Niyato, D., Hossain, E., and Camorlinga, S. Remote patient monitoring service using heterogeneous wireless access networks: architecture and optimization. *Selected Areas in Communications, IEEE Journal on*. 2009. 27(4): 412-423.
  31. Yan, Z., Ansari, N., and Tsunoda, H. Wireless telemedicine services over integrated IEEE 802.11/WLAN and IEEE 802.16/WiMAX networks. *Wireless Communications, IEEE*. 2010. 17(1): 30-36.
  32. Pareit, D., Lannoo, B., Moerman, I., and Demeester, P. The History of WiMAX: A Complete Survey of the Evolution in Certification and Standardization for IEEE 802.16 and WiMAX. *Communications Surveys & Tutorials, IEEE*. 2012. 14(4): 1183-1211.
  33. Mitra, S., Mitra, M., and Chaudhuri, B.B. Rural cardiac healthcare system-A scheme for developing countries. *TENCON 2008 - 2008 IEEE Region 10 Conference*. 19-21 Nov. 2008.2008. 1-5.
  34. Elena, M., Quero, J.M., Tarrida, C.L., and Franquelo, L.G. Design of a mobile telecardiology system using GPRS/GSM technology. *Engineering in Medicine and Biology, 2002. 24th Annual Conference and the Annual Fall Meeting of the Biomedical Engineering Society EMBS/BMES Conference*,

2002. *Proceedings of the Second Joint*. 23-26 Oct. 2002. Houston, Texas, USA: IEEE.2002. 1859-1860.
35. Abo-Zahhad, M., Ahmed, S.M., and Elnahas, O. A wireless emergency telemedicine system for patients monitoring and diagnosis. *Int J Telemed Appl*. 2014. 2014: 380787.
  36. Anpeng, H., Chao, C., Kaigui, B., Xiaohui, D., Min, C., Hongqiao, G., Chao, M., Qian, Z., Yingrui, Z., Bingli, J., and Linzhen, X. WE-CARE: An Intelligent Mobile Telecardiology System to Enable mHealth Applications. *Biomedical and Health Informatics, IEEE Journal of*. 2014. 18(2): 693-702.
  37. Mateev, H., Simova, I., Katova, T., and Dimitrov, N. Clinical Evaluation of a Mobile Heart Rhythm Telemonitoring System. *ISRN Cardiology*. 2012. 2012: 8.
  38. Satria, M.H. *Integration of Multimetrix Path Management Into 802.11S for Telemedicine Quality of Service Provision*. PhD. Universiti Teknologi Malaysia; 2014.
  39. Yan, X., Ahmet Şekercioğlu, Y., and Narayanan, S. A survey of vertical handover decision algorithms in Fourth Generation heterogeneous wireless networks. *Computer Networks*. 2010. 54(11): 1848-1863.
  40. Ahmed, A., Boulahia, L.M., Gai, x, and ti, D. Enabling Vertical Handover Decisions in Heterogeneous Wireless Networks: A State-of-the-Art and A Classification. *Communications Surveys & Tutorials, IEEE*. 2014. 16(2): 776-811.
  41. Pollini, G.P. Trends in handover design. *Communications Magazine, IEEE*. 1996. 34(3): 82-90.
  42. Wong, K.D. and Cox, D.C. Two-state pattern-recognition handoffs for corner-turning situations. *Vehicular Technology, IEEE Transactions on*. 2001. 50(2): 354-363.
  43. Zhang, H., Ma, W., Li, W., Zheng, W., Wen, X., and Jiang, C. Signalling Cost Evaluation of Handover Management Schemes in LTE-Advanced Femtocell. *Vehicular Technology Conference (VTC Spring), 2011 IEEE 73rd*. 15-18 May 2011. Budapest, Hungary: IEEE.2011. 1-5.
  44. Song, W. and Zhuang, W. Introduction on Cellular/WLAN Interworking. In: *Interworking of Wireless LANs and Cellular Networks*. Springer New York. p. 1-10; 2012

45. Malathy, E.M. and Muthuswamy, V. Knapsack - TOPSIS Technique for Vertical Handover in Heterogeneous Wireless Network. *PLoS ONE*. 2015. 10(8): e0134232.
46. Khan, M. and Han, K. An Optimized Network Selection and Handover Triggering Scheme for Heterogeneous Self-Organized Wireless Networks. *Mathematical Problems in Engineering*. 2014. 2014: 11.
47. Janevski, T. and Jakimoski, K. Mobility sensitive algorithm for vertical handovers from WiMAX to WLAN. *Telecommunications Forum (TELFOR), 2012 20th*. 20-22 Nov. 2012. 2012. 91-94.
48. Hong, K., Lee, S., Kim, L., and Song, P. Cost-based vertical handover decision algorithm for WWAN/WLAN integrated networks. *EURASIP J. Wirel. Commun. Netw.* 2009. 2009: 1-11.
49. Kaleem, F., Mehbodniya, A., Yen, K.K., and Adachi, F. A Fuzzy Preprocessing Module for Optimizing the Access Network Selection in Wireless Networks. *Advances in Fuzzy Systems*. 2013. 2013: 9.
50. Hussain, R., Malik, S., Abrar, S., Riaz, R., Ahmed, H., and Khan, S. Vertical Handover Necessity Estimation Based on a New Dwell Time Prediction Model for Minimizing Unnecessary Handovers to a WLAN Cell. *Wireless Personal Communications*. 2013. 71(2): 1217-1230.
51. Wu, J. Exploring Maximum Doppler Diversity by Doppler Domain Multiplexing. *IEEE Globecom 2006*. Nov. 27 2006-Dec. 1 2006. 2006. 1-5.
52. Yang, Y., Fan, P., and Huang, Y. Doppler frequency offsets estimation and diversity reception scheme of high speed railway with multiple antennas on separated carriages. *Wireless Communications & Signal Processing (WCSP), 2012 International Conference on*. 25-27 Oct. 2012. Huangshan, China: IEEE. 2012. 1-6.
53. Feukeu, E.A., Djouani, K., and Kurien, A. Compensating the effect of Doppler shift in a vehicular network. *AFRICON, 2013*. 9-12 Sept. 2013. 2013. 1-7.
54. Burbank, J.L., Andrusenko, J., Everett, J.S., and Kasch, W.T.M. *Wireless Networking: Understanding Internetworking Challenges*: Wiley. 2013
55. MAXIS Annual Report 2015. [Retrieved Available from: <http://maxis.my/ar2015/index.html>.

56. Digi Annual Report 2015. [Retrieved Available from: [http://www.digi.com.my/aboutus/media/annual\\_report.do](http://www.digi.com.my/aboutus/media/annual_report.do).
57. Axiata Annual Report 2015. [Retrieved Available from: <http://www.axiata.com/investor/annual-reports>.
58. MAXIS 4G coverage in Malaysia. [cited 2016 21st September 2016]; MAXIS 4G coverage in Malaysia]. Available from: <https://www.maxis.com.my/en/personal/internet/fastest-and-widest-data-network.html>.
59. Chin-Teng, L., Kuan-Cheng, C., Chun-Ling, L., Chia-Cheng, C., Shao-Wei, L., Shih-Sheng, C., Bor-Shyh, L., Hsin-Yueh, L., Ray-Jade, C., Yuan-Teh, L., and Li-Wei, K. An Intelligent Telecardiology System Using a Wearable and Wireless ECG to Detect Atrial Fibrillation. *Information Technology in Biomedicine, IEEE Transactions on*. 2010. 14(3): 726-733.
60. Chorbev, I. and Mihajlov, M. Wireless telemedicine services as part of an integrated system for e-medicine. *Electrotechnical Conference, 2008. MELECON 2008. The 14th IEEE Mediterranean. 5-7 May 2008*. 2008. 264-269.
61. Mair, F., Fraser, S., Ferguson, J., and Webster, K. Telemedicine via satellite to support offshore oil platforms. *J Telemed Telecare*. 2008. 14(3): 129-31.
62. de Barbara, A.H.A. *The Spanish Ministry of Defence (MOD) Telemedicine System*: INTECH Open Access Publisher. 2011
63. Clarke, M., Jones, R.W., and Lioupis, D. AIDMAN-Telecardiology over a high-speed satellite network. *Computers in Cardiology 2000*. 2000. 2000. 657-660.
64. Becvar, Z., Mach, P., and Bestak, R. Impact of Handover on VoIP Speech Quality in WiMAX Networks. *Networks, 2009. ICN '09. Eighth International Conference on*. 1-6 March 2009. Cancun, Mexico: IEEE. 2009. 281-286.
65. Aguado, M., Astorga, J., Matias, J., and Huarte, M. The MIH (Media Independent Handover) Contribution to Mobility Management in a Heterogeneous Railway Communication Context: A IEEE802.11/802.16 Case Study. In: *Communication Technologies for Vehicles*, T. Strang, et al., Editors. Springer Berlin Heidelberg. p. 69-82; 2011

66. Yujae, S., Peng-Yong, K., and Youngnam, H. Power-Optimized Vertical Handover Scheme for Heterogeneous Wireless Networks. *Communications Letters, IEEE*. 2014. 18(2): 277-280.
67. Nasser, N., Hasswa, A., and Hassanein, H. Handoffs in fourth generation heterogeneous networks. *Communications Magazine, IEEE*. 2006. 44(10): 96-103.
68. Kyoung Seok, L. and Ae-Soon, P. Reduction of handover failure for small cells in heterogeneous networks. *Information and Communication Technology Convergence (ICTC), 2014 International Conference on*. 22-24 Oct. 2014. Busan, Korea: IEEE.2014. 707-708.
69. Jung-Min, M., Jungsoo, J., Sungjin, L., Nigam, A., and Sunheui, R. On the trade-off between handover failure and small cell utilization in heterogeneous networks. *Communication Workshop (ICCW), 2015 IEEE International Conference on*. 8-12 June 2015. London, UK: IEEE.2015. 2282-2287.
70. De La Oliva, A., Banchs, A., Soto, I., Melia, T., and Vidal, A. An overview of IEEE 802.21: media-independent handover services. *Wireless Communications, IEEE*. 2008. 15(4): 96-103.
71. la Oliva, A.d., Soto, I., Banchs, A., Lessmann, J., Niephaus, C., and Melia, T. IEEE 802.21: Media independence beyond handover. *Computer Standards & Interfaces*. 2011. 33(6): 556-564.
72. Ahuja, K., Singh, B., and Khanna, R. Network selection algorithm based on link quality parameters for heterogeneous wireless networks. *Optik - International Journal for Light and Electron Optics*. 2014. 125(14): 3657-3662.
73. Oliva, A.d.l., Melia, T., Vidal, A., Bernardos, C.J., Soto, I., and Banchs, A. IEEE 802.21 enabled mobile terminals for optimized WLAN/3G handovers: a case study. *SIGMOBILE Mob. Comput. Commun. Rev.* 2007. 11(2): 29-40.
74. Kassar, M., Kervella, B., and Pujolle, G. An overview of vertical handover decision strategies in heterogeneous wireless networks. *Computer Communications*. 2008. 31(10): 2607-2620.
75. Mohanty, S. and Akyildiz, I.F. A Cross-Layer (Layer 2 + 3) Handoff Management Protocol for Next-Generation Wireless Systems. *Mobile Computing, IEEE Transactions on*. 2006. 5(10): 1347-1360.



76. Ylianttila, M., Pande, M., Makela, J., and Mahonen, P. Optimization scheme for mobile users performing vertical handoffs between IEEE 802.11 and GPRS/EDGE networks. *Global Telecommunications Conference, 2001. GLOBECOM '01. IEEE*. 2001.2001. 3439-3443 vol.6.
77. Dhar Roy, S. and Vamshidhar Reddy, S.R. Signal Strength Ratio Based Vertical Handoff Decision Algorithms in Integrated Heterogeneous Networks. *Wireless Personal Communications*. 2014. 77(4): 2565-2585.
78. Haider, A., Gondal, I., and Kamruzzaman, J. Dynamic Dwell Timer for Hybrid Vertical Handover in 4G Coupled Networks. *Vehicular Technology Conference (VTC Spring), 2011 IEEE 73rd*. 15-18 May 2011. Budapest, Hungary: IEEE.2011. 1-5.
79. Xiaohuan, Y., Mani, N., and Sekercioglu, Y.A. A Traveling Distance Prediction Based Method to Minimize Unnecessary Handovers from Cellular Networks to WLANs. *Communications Letters, IEEE*. 2008. 12(1): 14-16.
80. Nay, P. and Chi, Z. Vertical Handoff Decision Algorithm for Integrated UMTS and LEO Satellite Networks. *Communications and Mobile Computing, 2009. CMC '09. WRI International Conference on*. 6-8 Jan. 2009. Yunnan, China: IEEE.2009. 180-184.
81. Lee, D.W., Gil, G.T., and Kim, D.H. A cost-based adaptive handover hysteresis scheme to minimize the handover failure rate in 3GPP LTE system. *EURASIP J. Wirel. Commun.* 2010. 2010.
82. Kemeng, Y., Gondal, I., and Bin, Q. Multi-Dimensional Adaptive SINR Based Vertical Handoff for Heterogeneous Wireless Networks. *Communications Letters, IEEE*. 2008. 12(6): 438-440.
83. Park, J., Ryu, H., Lee, S.S., Bae, S.J., Lee, J.Y., Kim, M., and Chung, M.Y. Cost-efficient vertical handover between cellular networks and WLAN based on HMIPv6 and IEEE 802.21 MIH. *International Journal of Network Management*. 2013. 23(3): 155-171.
84. Dong, M. and Maode, M. A QoS Oriented Vertical Handoff Scheme for WiMAX/WLAN Overlay Networks. *Parallel and Distributed Systems, IEEE Transactions on*. 2012. 23(4): 598-606.
85. Kemeng, Y., Bin, Q., and Dooley, L.S. Using SINR as Vertical Handoff Criteria in Multimedia Wireless Networks. *Multimedia and Expo, 2007*

- IEEE International Conference on.* 2-5 July 2007. Beijing, China: IEEE.2007. 967-970.
86. Kemeng, Y., Gondal, I., Bin, Q., and Dooley, L.S. Combined SINR Based Vertical Handoff Algorithm for Next Generation Heterogeneous Wireless Networks. *Global Telecommunications Conference, 2007. GLOBECOM '07. IEEE.* 26-30 Nov. 2007.2007. 4483-4487.
  87. Yadav, A., Barooah, M., Chakraborty, S., and Nandi, S. Vertical Handover Over Intermediate Switching Framework: Assuring Service Quality for Mobile Users. *Wireless Personal Communications.* 2014. 77(1): 507-527.
  88. Ben Elhadj, H., Elias, J., Chaari, L., and Kamoun, L. Multi-Attribute Decision Making Handover Algorithm for Wireless Body Area Networks. *Computer Communications.* 2016. 81: 97-108.
  89. Salih, Y.K., See, O.H., Ibrahim, R.W., Yussof, S., and Iqbal, A. A user-centric game selection model based on user preferences for the selection of the best heterogeneous wireless network. *annals of telecommunications - annales des télécommunications.* 2015. 70(5): 239-248.
  90. Wang, Y.-M. and Luo, Y. On rank reversal in decision analysis. *Mathematical and Computer Modelling.* 2009. 49(5–6): 1221-1229.
  91. Jafari, A.H. and Shahhoseini, H.S. A location aware history-based approach for network selection in heterogeneous wireless networks. *Turkish Journal of Electrical Engineering and Computer Sciences.* 2016. 24(4): 2929-2948.
  92. Ishizaka, A. and Nemery, P. *Multi-criteria Decision Analysis: Methods and Software:* Wiley. 2013
  93. Tran, P.N. and Boukhatem, N. *The distance to the ideal alternative (DiA) algorithm for interface selection in heterogeneous wireless networks,* in *Proceedings of the 6th ACM international symposium on Mobility management and wireless access.* 2008, ACM: Vancouver, British Columbia, Canada. p. 61-68.
  94. Saaty, T.L. A scaling method for priorities in hierarchical structures. *Journal of Mathematical Psychology.* 1977. 15(3): 234-281.
  95. Saaty, T.L. *The Analytic Hierarchy Process: Planning, Priority Setting, Resource Allocation:* McGraw-Hill. 1980

96. Tawil, R., Pujolle, G., and Salazar, O. A Vertical Handoff Decision Scheme in Heterogeneous Wireless Systems. *Vehicular Technology Conference, 2008. VTC Spring 2008. IEEE*. 11-14 May 2008. 2008. 2626-2630.
97. Liu, S.-m., Pan, S., Mi, Z.-k., Meng, Q.-m., and Xu, M.-h. A Simple Additive Weighting Vertical Handoff Algorithm Based on SINR and AHP for Heterogeneous Wireless Networks. *Intelligent Computation Technology and Automation (ICICTA), 2010 International Conference on*. 11-12 May 2010. Changsha, China: IEEE. 2010. 347-350.
98. Khan, M. and Han, K. A Vertical Handover Management Scheme based on Decision Modelling in Heterogeneous Wireless Networks. *IETE Technical Review*. 2015. 32(6): 402-412.
99. Mehbodniya, A., Kaleem, F., Yen, K.K., and Adachi, F. A fuzzy extension of VIKOR for target network selection in heterogeneous wireless environments. *Physical Communication*. 2013. 7: 145-155.
100. Salih, Y.K., See, O.H., Yussof, S., Iqbal, A., and Mohammad Salih, S.Q. A Proactive Fuzzy-Guided Link Labeling Algorithm Based on MIH Framework in Heterogeneous Wireless Networks. *Wireless Personal Communications*. 2014. 75(4): 2495-2511.
101. Chamodrakas, I. and Martakos, D. A utility-based fuzzy TOPSIS method for energy efficient network selection in heterogeneous wireless networks. *Applied Soft Computing*. 2011. 11(4): 3734-3743.
102. Chinnappan, A. and Balasubramanian, R. Complexity-consistency trade-off in multi-attribute decision making for vertical handover in heterogeneous wireless networks. *IET Networks*. 2016. 5(1): 13-21.
103. Singhrova, A. and Prakash, N. Vertical handoff decision algorithm for improved quality of service in heterogeneous wireless networks. *Communications, IET*. 2012. 6(2): 211-223.
104. Nasser, N., Guizani, S., and Al-Masri, E. Middleware Vertical Handoff Manager: A Neural Network-Based Solution. *Communications, 2007. ICC '07. IEEE International Conference on*. 24-28 June 2007. Glasgow, Scotland: IEEE. 2007. 5671-5676.
105. Sadiq, A.S., Bakar, K.A., Ghafoor, K.Z., Lloret, J., and Khokhar, R. An Intelligent Vertical Handover Scheme for Audio and Video Streaming in

- Heterogeneous Vehicular Networks. *Mobile Networks and Applications*. 2013. 18(6): 879-895.
106. Sadiq, A.S., Bakar, K.A., Ghafoor, K.Z., Lloret, J., and Mirjalili, S. A smart handover prediction system based on curve fitting model for Fast Mobile IPv6 in wireless networks. *International Journal of Communication Systems*. 2014. 27(7): 969-990.
  107. Goudarzi, S., Haslina Hassan, W., Abdalla Hashim, A.-H., Soleymani, S.A., Anisi, M.H., and Zakaria, O.M. A Novel RSSI Prediction Using Imperialist Competition Algorithm (ICA), Radial Basis Function (RBF) and Firefly Algorithm (FFA) in Wireless Networks. *PLoS ONE*. 2016. 11(7): e0151355.
  108. Gallego, J.R., Hernandez-Solana, A., Canales, M., Lafuente, J., Valdovinos, A., and Fernandez-Navajas, J. Performance analysis of multiplexed medical data transmission for mobile emergency care over the UMTS channel. *IEEE Trans Inf Technol Biomed*. 2005. 9(1): 13-22.
  109. Mohanty, S. VEPSD: a novel velocity estimation algorithm for next-generation wireless systems. *Wireless Communications, IEEE Transactions on*. 2005. 4(6): 2655-2660.
  110. Santi, P. Modeling Next Generation Wireless Networks. In: *Mobility Models for Next Generation Wireless Networks*. John Wiley & Sons, Ltd. p. 19-32; 2012
  111. Nguyen-Vuong, Q.-T., Agoulmine, N., and Ghamri-Doudane, Y. A user-centric and context-aware solution to interface management and access network selection in heterogeneous wireless environments. *Computer Networks*. 2008. 52(18): 3358-3372.
  112. Singh, B. An improved handover algorithm based on signal strength plus distance for interoperability in mobile cellular networks. *Wireless Personal Communications*. 2007. 43(3): 879-887.
  113. Ravi, A. and Peddoju, S. Handoff Strategy for Improving Energy Efficiency and Cloud Service Availability for Mobile Devices. *Wireless Personal Communications*. 2015. 81(1): 101-132.