



Title	A novel range-verification method using ionoacoustic wave generated from spherical gold markers for particle-beam therapy : a simulation study
Author(s)	Takayanagi, Taisuke; Uesaka, Tomoki; Kitaoka, Masanori; Unlu, Mehmet Burcin; Umegaki, Kikuo; Shirato, Hiroki; Xing, Lei; Matsuura, Taeko
Citation	Scientific reports, 9, 4011 https://doi.org/10.1038/s41598-019-38889-w
Issue Date	2019-03-08
Doc URL	http://hdl.handle.net/2115/73906
Rights(URL)	https://creativecommons.org/licenses/by/4.0/
Type	article
File Information	s41598-019-38889-w.pdf



[Instructions for use](#)

SCIENTIFIC REPORTS



OPEN

A novel range-verification method using ionoacoustic wave generated from spherical gold markers for particle-beam therapy: a simulation study

Taisuke Takayanagi^{1,2}, Tomoki Uesaka¹, Masanori Kitaoka², Mehmet Burcin Unlu^{3,4,5}, Kikuo Umegaki^{3,6,7}, Hiroki Shirato^{3,7}, Lei Xing^{3,5} & Taeko Matsuura^{3,6,7}

This study proposes a novel alternative range-verification method for proton beam with acoustic waves generated from spherical metal markers. When proton beam is incident on metal markers, most of the resulting pressure waves are confined in the markers because of the large difference in acoustic impedance between the metal and tissue. However, acoustic waves with frequency equal to marker's resonant frequency escape this confinement; the marker briefly acts as an acoustic transmitter. Herein, this phenomenon is exploited to measure the range of the proton beam. We test the proposed strategy in 3-D simulations, combining the dose calculations with modelling of acoustic-wave propagation. A spherical gold marker of 2.0 mm diameter was placed in water with a 60 MeV proton beam incident on it. We investigated the dependence of pressure waves on the width of beam pulse and marker position. At short beam pulse, specific high-frequency acoustic waves of 1.62 MHz originating from the marker were observed in wave simulations, whose amplitude correlated with the distance between the marker and Bragg peak. Results indicate that the Bragg peak position can be estimated by measuring the acoustic wave amplitudes from the marker, using a single detector properly designed for the resonance frequency.

Spot-scanning proton therapy (SSPT) is an advanced form of proton therapy, and is being widely employed in newly constructed treatment centers¹. By superposing many individual spot doses from the Bragg peaks of proton beam, SSPT can confine the radiation dose to the tumour while sparing the surrounding normal tissue^{2,3}. However, the precision of the treatment plan is limited by uncertainty in the range of the proton beam and the patient setup. These uncertainties can offset the advantage gained by the spot-scanning technique⁴.

Uncertainty in the range of the proton beam arises from multiple sources, including computed tomography (CT) number to stopping-power conversions and anatomical changes during radiation treatment. These uncertainties have been estimated to about 3.5% of the beam range⁵. Several *in vivo* techniques have been proposed to reduce the uncertainty in the range of proton delivery. Examples include positron-emission tomography (PET)⁶, prompt gamma-ray (PG) detection⁷ and detection of ionoacoustic wave, i.e., acoustic signals originating from the impact of proton beams⁸. PET and PG detection estimate the proton range by measuring the cross sections of nuclear interactions during or immediately following dose delivery.

¹Graduate School of Biomedical Science and Engineering, Hokkaido University, North-13 West-8, Kita-ku, Sapporo, Hokkaido, 060-8628, Japan. ²Hitachi Ltd., 1-1 7-chome, Omika-cho, Hitachi-shi, Ibaraki, 319-1292, Japan. ³Global Station for Quantum Medical Science and Engineering, Global Institution for Collaborative Research and Education (GI-CoRE), Hokkaido University, Sapporo, 060-8648, Japan. ⁴Department of Physics, Bogazici University, Bebek, Istanbul, 34342, Turkey. ⁵Department of Radiation Oncology, Division of Medical Physics, Stanford University School of Medicine, Stanford, CA, USA. ⁶Faculty of Engineering, Hokkaido University, North-13 West-8, Kita-ku, Sapporo, Hokkaido, 060-8628, Japan. ⁷Proton Beam Therapy Center, Hokkaido University Hospital, North-15 West-7, Kita-ku, Sapporo, Hokkaido, 060-8638, Japan. Correspondence and requests for materials should be addressed to T.M. (email: matsuura@med.hokudai.ac.jp)




Figure 1. A schematic figure of the simulation environment. The spherical gold marker is placed in a water phantom at depth Z_m . Detectors, marked with small black circles, are aligned parallel to the proton beam (grey arrow) at 0.2 mm intervals.

In contrast, ionoacoustic detection measures the energy transferred by protons at the Bragg peak^{8–20}. Like PET and PG detection, acoustic-signal detection is non-invasive. Additionally, acoustic-signal detection does not require bulky equipment for implementation. So far, several simulation studies have focused on waveform analysis in a simple water medium^{14,17–19} and in prostate- and liver-cancer patients²⁰. Experimental observations of acoustic emissions have been performed with a linac⁹, a synchrotron^{10–13}, a tandem accelerator¹⁴, a hospital-based isochronous cyclotron¹⁵ and a synchrocyclotron¹⁶, with positive results.

Uncertainty in the patient setup can be reduced by exploiting the image guidance techniques. For this purpose, radiographically visible markers placed within or nearby the tumour are commonly employed to monitor the position of the tumour during radiation therapy^{21–23}. For tumours that move with respiratory cycles, the markers are monitored in real-time to control beam delivery^{24,25}. A wide range of markers with different shapes and materials (gold, steel and others) are used for this purpose, and are selected depending on the tumour site and irradiation strategy²².

This article focuses on a spherical-shaped gold fiducial marker and proposes a novel alternative range verification methodology using the acoustic signal emitted from this marker. Gold's material properties act as a strong pressure source compared to the surrounding biological tissue when irradiated with protons. However, since the large difference in acoustic impedance between gold and water prevents the pressure waves from propagating outward into the tissue, most of these pressure waves are confined within the marker. Only waves with the resonant frequency of the marker gain enough amplitude to propagate outward into the surrounding tissue. We demonstrated this phenomenon in simulations of acoustic-wave transport to explore how this phenomenon can be applied to beam range verification.

Results

The results discussed below are from three-dimensional simulations that combined dose calculations with a model of acoustic-wave propagation. As shown in Fig. 1, we modelled a spherical gold spherical marker (diameter = 2 mm) placed in water, and a proton pencil beam (incident beam energy = 60 MeV and the spot size = 5 mm) is incident on it along the z -axis, which passes through the centre of the marker. The x and y -axes are taken as the transverse coordinates. To measure the time-resolved pressure waves, point-like detectors are aligned 20.6 mm away from the z -axis. The details of the simulation environment are described below in section 4. The pressure waves were measured in both temporal and frequency domains with variations of the following parameters: the beam pulse width σ_p and the marker position Z_m .

General characteristics of transferred energy and waveform. In this study, we assume that the proton beam is injected in a short period of time ($\sigma_p = 30–500$ ns), which is within the output range of fast-extraction synchrotron, fixed-field alternating gradient, dielectric wall and laser-driven accelerators¹⁹. Figure 2(a,b) show the transferred energy distributions on the z - x plane without and with the gold marker, respectively. Protons transfer 9.5 times more energy in gold than they do in water due to gold's relatively high electron density. Accordingly, a hot spot arises in Fig. 2(b) and a cold region develops downstream of the gold marker.

Figure 2(c,d) show the time-resolved pressure waves as measured along the line of detectors in the absence and presence of the gold marker, respectively ($\sigma_p = 100$ ns). As shown in Fig. 2(c), with no gold marker, a typical bipolar waveform appears; the z -dependent arrival time at the post-Bragg peak region ($Z \gtrsim 31$ mm) is a feature of γ waves, while the weaker z -dependency of arrival time is a feature of α waves at pre-Bragg peak positions ($Z \lesssim 31$ mm)⁸.

As shown in Fig. 2(d), the presence of a gold marker introduces a distinctive high-frequency component to the original waveform. It is symmetric around the depth of the marker ($Z_m = 22$ mm), and its frequency does not depend on the detector position. The maximum pressure of -5.5×10^{-5} mPa/proton is achieved at $Z = 16.2$ mm at $t = 21.1$ μ s, and the absolute value of the pressure is about 1.5 times higher than it is with no gold marker present (3.6×10^{-5} mPa/proton at $Z = 17.0$ mm and $t = 14.8$ μ s).




Figure 2. Distribution of transferred energy from proton beam on the z - x plane (a) without a gold marker (b) with a 2.0-mm-diameter gold marker placed at $Z_m = 22$ mm. The horizontal axis is depth in water and the vertical axis is position along the x -axis marked in Fig. 1. Voxel volume = $0.2 \times 0.2 \times 0.2$ mm³. (c,d) are the acoustic wave propagation from the transferred energy distributions displayed in (a,b), respectively. An incident pulse width is 100 ns. The horizontal axis indicates the detector position, the vertical axis indicates the time from the incident pulse peak, and the greyscale colour indicates the sound pressure.

Effects of the beam pulse width σ_p . In this section, we discuss how the frequency and amplitude of the acoustic wave were affected by changing the pulse width. These simulations were run with the detector placed at the same depth as the marker, $Z = 22$ mm.

Figure 3(a) shows the pressure wave simulated with no marker present, and with an incident pulse width of 100 ns. The waveform has a bipolar shape originating from the dose distribution of the pencil beam. The peak-to-peak time of the bipolar waveform (~ 6.5 μ s) can be deduced from the lateral size of the pencil beam (FWHM ~ 12 mm), which corresponds to a pressure-wave propagation time of ~ 8 μ s. Since this value was two or three orders of magnitude greater than the range of the tested pulse width ($\sigma_p = 30$ –500 ns), σ_p has a negligible effect on the waveform.

Figure 3(b) shows the waveform simulated with a gold marker present and with $\sigma_p = 100$ ns (thin solid curve) or 200 ns (thick solid curve). The high-frequency wave appears immediately after the first positive peak emitted from the pencil beam. The amplitude of the wave decreases as σ_p increases. Figure 3(c) plots the difference between the waveforms measured with and without the gold marker present. The maximum pressure measured with the gold marker present was -4.7×10^{-5} mPa/proton and occurred at $t = 13.7$ μ s. This maximum pressure was generated at the gold marker, as its timing is close to the distance between the marker and the detector divided by the speed of sound in water, i.e., 13.1 μ s.

Next, a Fourier transform was applied to the pressure waves in Fig. 3(a,b) to examine their frequency distributions, as shown in Fig. 4. With no marker present, the signal energy arising from the pencil beam is concentrated in the low-frequency regime ($f < 200$ kHz). In this regime, the amplitude is independent of σ_p , as expected from the discussion above. With a gold marker present, a specific resonance peak appears at 1.62 MHz. The resonant frequency does not depend on σ_p , and its magnitude is negatively correlated with σ_p . In the cases of $\sigma_p = 30$, 100 ns, a second peak appears at 3.24 MHz, which is twice the frequency of the first peak. These values agree with




Figure 3. (a) Acoustic waveform in the absence of a marker with $\sigma_p = 100$ ns. (b) Acoustic wave in the presence of a marker with $\sigma_p = 100$ and 200 ns. The marker is placed at $Z_m = 22$ mm (same depth as the detector (dotted line in Fig. 2 (d))). (c) Difference between the pressures with and without the gold marker. The horizontal axis gives the time (μ s) from the incident pulse peak and the vertical axis indicates the pressure (mPa/proton).




Figure 4. Frequency spectrums of the acoustic waves generated from a proton beam pulse with $\sigma_p = 30$, 100 and 200 ns. The horizontal axis is frequency, and the vertical axis is amplitude. Results are shown for a 2-mm-diameter marker at $Z_m = 22$ mm (same depth as the detector).




Figure 5. The peak amplitude at the resonant wave at 1.62 MHz in terms of the residual range R_{res} at marker positions defined by Eq. (1). The vertical axis represents the signal amplitude at 1.62 MHz. Results are shown for $\sigma_p = 100$ ns.

the eigenvalues of the wave equations with zero boundary conditions on the sphere surface, as discussed in section 3 below ($n = 1$ and 2 in Eq. (6), respectively).

Effects of marker position Z_m . The resonance-peak frequency and amplitude were next explored by varying the marker positions from $Z_m = 16$ to 38 mm. Like above, these simulations were run with the detector placed at the same depth as the marker ($Z = Z_m$). As with Fig. 4, two peaks were observed at 1.62 and 3.24 MHz and both were independent of Z_m . For the sake of concision, we concentrate on the first resonance peak at 1.62 MHz below.

In Fig. 5, the amplitude of the resonance peak is plotted in terms of the residual range R_{res} at marker position, which is defined as follows

$$R_{res} = R_0 - Z_m. \quad (1)$$

Here, R_0 is the proton range, which is 31 mm in this case (incident beam energy = 60 MeV).

Based on the characteristics of the plot in Fig. 5, we categorise the marker positions into three cases: case 1: $R_{res} > 9$ mm (marker is placed more than 9 mm upstream of the Bragg peak), case 2: $0 < R_{res} \leq 9$ mm (marker is placed less than 9 mm upstream of the Bragg peak), and case 3: $R_{res} \leq 0$ (marker is placed downstream of the Bragg peak).

In case 2, the amplitude of the acoustic wave increases almost linearly with R_{res} at a rate of approximately 15% per mm. In this case, any protons entering the marker transfer all their residual kinetic energy to the marker and are stopped inside the marker. Therefore, the acoustic wave amplitude will be larger if the proton energy is greater, which is the case as the residual range increases. A range verification method that exploits this phenomenon will be discussed in the following section. In case 1, on the other hand, the proton beam passes the marker and the energy transferred to the marker is roughly proportional to the stopping power at R_{res} . As such, the amplitude of the acoustic wave emanating from the gold marker increases as R_{res} decreases. The wave amplitude is affected less by R_{res} as R_{res} increase, reflecting how the proton energy loss decreases slowly at large R_{res} . Finally, no signal appears at all in case 3, since the beam does not reach the gold marker.

Discussion

Our wave-propagation simulations showed that resonant waves are emitted from the gold marker after the injection of the pulsed proton pencil beam. Here, we analytically derive the properties of this resonant wave that appeared in simulations.

The observed waveform can be decomposed into the waves originating within and outside of the marker. Waves originating outside the marker can be approximated as waves originating from the pencil beam dose distribution with no marker present; although the gold marker develops the cold regions downstream as shown in Fig. 2(b), the area of dose distortion is small enough that it negligibly impacts the acoustic waveform. The rest of the waveform originates within the marker, which acts as a spherically-symmetric heat source. Immediately after proton injection, thermalisation occurs due to the high heat conductivity of gold and the marker reaches a uniform temperature very quickly. During proton injection, a large amount of energy is transferred to the gold (9.5 times the amount energy transferred to water). This energy is rapidly converted to a large pressure wave since the Gruneisen coefficient of gold, i.e., the conversion factor from thermal energy to pressure, is 32 times greater than that of water. However, a large part of the acoustic energy does not reach the surrounding water and is reflected back into the gold as there is a large difference in acoustic impedance at the interface between the gold marker and water. As a result, wave pressure in water reduces to 5% of that in the marker.

	$\rho(\text{g/cm}^3)$	$\beta(\text{K}^{-1})$	$v(\text{m/s})$	$C_p(\text{J/K/kg})$	Γ	$Z(\text{Ns/m}^3)$	$\alpha_{th}(\text{m}^2/\text{s})$
Water (27 °C)	1	210×10^{-6}	1500	4180	0.11	1.5×10^6	0.15×10^{-6}
Gold	19.3	42.6×10^{-6}	3240	128.8	3.47	6.3×10^7	128×10^{-6}

Table 1. Material properties of water and gold. ρ : mass density, β : coefficient of volumetric thermal expansion, v : speed of sound, C_p : heat capacity at constant pressure, Γ : Gruneisen coefficient, Z : acoustic impedance, and α_{th} : thermal diffusivity.




Figure 6. Resonance frequencies of the acoustic waves generated from the gold marker. The horizontal axis is the marker diameter φ_m . The simulation results (marked with \circ) are plotted along with the theoretical curve calculated by Eq. (6) for $n = 1$ (dotted curve). The case of $\sigma_p = 100$ ns and $Z_m = 22$ mm is shown here. These results were simulated with a detector placed at $Z = 22$ mm, which is at the same depth as the gold marker.

Since the acoustic reflectance is nearly 1 at the boundary, let us first assume perfect reflection and set the pressure field $p(\vec{r}, t)$ to be zero at the boundary. Considering the spherical symmetry of the system, the pressure field with the frequency f can be denoted as $p(\vec{r}, t) = p(r)e^{i2\pi ft}$, where r is a radial coordinate with its origin located at the centre of the marker. Then, the acoustic wave equation^{8,14,26,27}

$$\nabla^2 p(\vec{r}, t) - \frac{1}{v^2} \frac{\partial^2 p(\vec{r}, t)}{\partial t^2} = 0 \quad (2)$$

becomes

$$\frac{\partial^2 p(r')}{\partial r'^2} + \frac{2}{r'} \frac{\partial p(r')}{\partial r'} + p(r') = 0, \quad (3)$$

where

$$r' = \frac{2\pi f}{v} r. \quad (4)$$

and v is the speed of sound in gold (see Table 1).

Spherical Bessel functions are the solutions of Eq. (3), with the zero-th order expressed as

$$p(r') \propto \frac{\sin(r')}{r'}. \quad (5)$$

If the diameter of the gold marker is φ_m , the boundary condition of complete reflection can be written as $p(\pi f \varphi_m / v) = 0$. The resonance frequency f_{res} is thereby:

$$f_{res} = \frac{v}{\varphi_m} n \quad (n: \text{integer}). \quad (6)$$

This resonant wave would mostly survive inside of the marker. However, as the reflectance is slightly less than 1 in practice, the resonant wave gradually releases energy to the water, generating the acoustic signal that was observed in the simulations. Pressure waves with frequencies other than f_{res} will cancel each other out when reflecting within the marker.




Figure 7. Signal amplitude at the resonance frequency of 1.62 MHz. The horizontal axis is the pulse width σ_p . The vertical axis is the normalised signal amplitude at $\sigma_p = 100$ ns. Circles show results obtained from the simulation, and the dotted curve was calculated from Eq. (7). The simulation with $\varphi_m = 2.0$ mm and $Z_m = 22$ mm is shown here and the detector is placed at the same depth as the gold marker.

To check the validity of the mechanism described above, the peak frequencies obtained by the simulation for different marker sizes are plotted in Fig. 6. The theoretical prediction (Eq. (6)), plotted in a dotted line, perfectly matches the simulation results.

In addition, the above treatment predicts that the resonance-peak amplitude should behave as

$$A(\sigma_p) \propto \exp\left[-\frac{(2\pi f_{res} \sigma_p)^2}{2}\right] \quad (7)$$

with respect to the Gaussian pulse width σ_p . As is shown in the Fig. 7, the simulation results for different pulse widths match perfectly with Eq. (7), which again supports the validity of the above analysis.

Range verification is an important potential application of the acoustic wave signal from a gold marker. As Fig. 5 shows, the amplitude of the acoustic wave emanating from the gold marker is correlated with R_{res} , i.e., the distance between the marker and the Bragg peak. Especially, when R_{res} is in the range of $0 < R_{res} < 9$ mm (case 2), the amplitude of the acoustic wave (A) is almost linearly related to R_{res} , namely:

$$R_{res} = C_1 A + C_0 \quad (8)$$

where C_1 and C_0 are constants that can be derived for each spot by using wave simulations on patient CT image. If these constants are calculated before treatment, R_{res} can be estimated in real time from measurements of A recorded by acoustic transducers. The marker position is measured by X-ray radiography or fluoroscopy^{23,25} and the beam axis is measured by profile monitors installed in the treatment nozzle. Coupled with these data, the acoustic measurements allow the calculation of the absolute position of the Bragg peak within a patient.

Although the range verification method outlined above is applicable only if the marker is placed less than 9 mm upstream of the Bragg peak, a simpler method may be useful if the proton beam stops just before the marker position ($|R_{res}| \sim 0$; case 3 in Fig. 5). In that case, the appearance of a resonant wave would indicate that the proton range is longer than anticipated. This method will be especially useful if the patient's situation allows the marker to be inserted at the boundary between the tumour and critical organs, since it could be used as a binary warning to prevent damage to critical organs. To accurately monitor a respiratory-moving tumour, several markers are generally placed around the tumour. If these markers have different diameters, the warned spot could be identified based on the various resonance frequency of the markers.

The range-verification method proposed here is not limited to proton beams alone, but is rather also applicable to other beam-scanning treatments with heavy ions, such as a carbon-ion beams. Once a spot is selected that is suitable for range verification, i.e. $0 < R_{res} < 9$ mm (case 2), a high dose should be delivered to those spots to generate a large-amplitude pressure wave, while suppressing the dose to surrounding spots in order to avoid inhomogeneity in the dose. Intensity-modulated proton therapy (IMPT) is therefore the ideal application for this range-verification method.

In contrast to the measurement of the acoustic wave from the Bragg peak, the frequency of the resonant wave from the gold marker depends solely on the marker size; it does not depend on the detector positioning or the properties of the incident beam (i.e., the lateral beam size, the width of the Bragg peak, the incident pulse width etc.). Therefore, the position of the Bragg peak can be estimated with only a single acoustic transducer, if it is properly designed for the resonant frequency of the gold marker.

To experimentally validate the clinical application of the proposed method, we discuss the magnitude of the possible range detection errors due to transducer noise and uncertainties in transducer positioning and beam delivery. For simplicity, we perform estimations with the same conditions as those used to derive the plot shown

in Fig. 5 (i.e., a beam energy of 60 MeV, a spot size of 5 mm (1σ), a pulse width of 100 ns, and a marker to detector distance of 20.6 mm), where the linear coefficient C_1 in Eq. (8) is 2.3×10^6 mm·proton/mPa.

Since we obtain the range from the measured acoustic pressure, range precision is limited by thermal noise in the transducer. According to a recent study¹⁸, the noise level of PZT transducers is typically about 0.03 mPa/ $\sqrt{\text{Hz}}$, which indicates that the noise of 35 mPa will be added to the resonant wave of 1.62 MHz. As a result, to detect the range within 1 mm of accuracy, each pulse requires more than 7.8×10^7 particles. Since a range of 3×10^8 to 8×10^8 protons per pulse is required to deliver a 2 Gy dose at the Bragg peak, clinical proton-beam therapy applications should generate sufficiently strong acoustic waves to verify the range. The range detection error increases for higher energy beams with an identical number of protons because higher energy beams have a wider Bragg peak due to energy straggling that causes an increase in the linear coefficient C_1 . Discussed later, the SNR should increase in such cases.

The high frequency of the resonant wave, on the order of MHz, will lead to strong attenuation in the tissue medium which can reduce the acoustic energy significantly. Therefore, attenuation of the acoustic signal must be considered carefully before treatment with acoustic-wave range verification. Accurate detector positioning is also important for the same reason. For instance, if we assume that the *in vivo* attenuation coefficient is 0.54 dB/cm/MHz²⁸, a 5 mm positioning error produces a range detection error of 0.16–0.88 mm ($R_{\text{res}} = 1\text{--}9$ mm).

Beam spot position error also causes range detection error. Assuming that the maximum spot position error during beam delivery is 0.2 mm, which occurs in current treatment machines with implemented position monitors, the deposited energy in the gold marker has a maximum uncertainty of 3.9%. This uncertainty is equivalent to an error range of 0.07–0.36 mm ($R_{\text{res}} = 1\text{--}9$ mm). In general, the influence of the spot position error is large for small-sized spots. For a clinical proton beam, the spot size near the Bragg peak is greater than the value used in the current study ($\sigma = 5$ mm). On the other hand, this effect requires analysis for heavy-ion beams with smaller spot sizes, such as for carbon beams.

In one perspective, the gold marker considered in this study is a possibly new type of *in vivo* point dosimeter. Previous studies have proposed several *in vivo* dosimeters such as implanting wireless dosimeters²⁹ and PET imaging of radioactive markers³⁰, and based on the literature, their range detection error is approximately 1 mm. Therefore, we expect that the present method should yield an equivalent range detection performance. On the other hand, a common limitation for these implanting dosimeters is that the range is determined for only a limited number of beams that pass through the dosimeters.

As demonstrated above, we estimated that the range detection error for the proposed method is less than 1 mm. However, the analysis was limited to a single condition. To establish this method's utility, future systematic studies are required. Simultaneously, increasing the SNR is important when attempting to expand this method's applicable range. Several approaches may be considered to do so. Large area detectors may effectively suppress any influence from thermal noise as noise amplitude is inversely proportional to the area of the PZT element¹⁸. The installation of multiple detectors can also reduce the influence of thermal noise. Previous studies have also explored the array type detector^{31–33}. Signal averaging can be used for noise reduction if the detectors are arranged on a spherical surface around the gold marker.

A specialized transducer may further reduce the SNR. High Q-value transducers with a thin backing layer have a higher sensitivity at similar resonance frequencies than commercially available broadband transducers³⁴ and are therefore suitable for application in the present study.

A shorter beam pulse and larger markers could also increase the SNR. Larger markers are more invasive, and block more of the proton dose downstream of the marker, so this option is not viable in practice. The development of less-invasive acoustically-detectable markers is a challenge for future work. Higher proton intensities and shorter beam pulses require advances in analytical capabilities, such as developments in laser-based proton acceleration and the Fixed-Field Alternating Gradient accelerator (FFAG), which are currently in progress. FFAG is especially a strong candidate because previous studies have shown that this technique provides the clinically required beam performance^{35,36}, i.e., proton beam energy of 150 MeV, pulse width of <100 ns, and intensity of 10^8 proton/pulse. However, delivering a spot dose with one pulse in the spot-scanning method requires further development with respect to the accurate control of the storage beam charge in the FFAG ring.

A conventional time-of-flight (TOF) approach could be used for the analysis involved in range verification^{11,15}. This approach would allow simultaneous detection of the positions of the marker and Bragg peak, which would increase the reliability of range measurements. To implement TOF calculations in practice, the transducer positions would need to be optimised so that the arrival time of the acoustic waves from the Bragg peak could be distinguished from that generated from the marker.

Finally, comparison with previous studies is discussed. Little research in the literature has focused on a similar arrangement of instrumentation to that in our study, in terms of beam pulse parameters and detector positions. Ahmad *et al.* employed similar settings in simulations, with incident beam energy = 100 MeV, pulse duration = 100 ns and spot size = 10 mm¹⁸. They reported that 100 mPa pressure was generated with 2 cGy irradiation dose. Though Ahmad *et al.* did not give the same level of detail about their simulation settings, we found that 2.9×10^6 particles are required to deliver a 2 cGy dose at the Bragg peak in our simulation, and this pulse generates 104 mPa of pressure. The result is therefore consistent with Ahmad *et al.*

In summary, we proposed a new range verification method using the acoustic wave generated from a spherical gold marker currently used in clinics to reduce set-up errors. Specific high-frequency acoustic waves originated from the marker were observed in numerical simulations. The wave frequency depended solely on the marker diameter, and amplitude was determined by the distance between the marker and the Bragg peak. The frequency of the resonance peak agreed well with the value derived analytically from the wave equation with the boundary condition of zero amplitude on the sphere surface, while the relation between the amplitude of the peak and initial pulse width was reproduced well in Fourier analysis. The results indicate that the measurement of the Bragg peak position may require only a single detector that is designed with consideration of the marker's resonance

frequency. Clinical implementation of the method will be easier with proton beams of higher intensity and shorter pulses, and a less-invasive acoustically-detectable marker, and these remain as challenges for future work.

Methods

Simulation setup. In the simulations used in this study, a spherical gold marker is placed in water, and a proton beam impinges upon the marker. Figure 1 diagrams the simulated system. A spherical gold marker is placed in a water phantom at a depth of Z_m . The diameter of the gold sphere marker was 2.0 mm, which is typically used in current clinical practice. The proton pencil beam is incident on the water phantom along the z -axis, which penetrates the centre of the marker. The x and y -axes are taken as the transverse coordinates. The beam energy is 60 MeV and corresponds to proton range of about 31 mm in water. This energy is close to the minimum energy provided by accelerators for proton therapy (~70 MeV) and was chosen to reduce the amount of memory required for the wave-propagation computations. The dose at the Bragg peak is 0.7×10^{-6} cGy/proton in the case. Both calculations, of dose distribution and pressure wave propagation, require spatial discretisation. The calculation grid size was set to $0.2 \times 0.2 \times 0.2$ mm³ in these simulations to properly represent the shape of a mm-sized spherical marker and to accurately simulate the propagation of high-frequency waves. To measure time-resolved pressure waves, point-like acoustic detectors are spaced at 0.2 mm intervals at 20.6 mm from the beam axis. 254 detectors are included in the simulation. In the wave propagation simulations, a perfectly matched layer is applied to the boundary, which absorbs all outgoing acoustic waves and prevents reflection.

Acoustic wave equation with an external pressure source. The proton beam is injected in a very short period in which stress confinement and thermal confinement are both fulfilled; the proton pulse width σ_p is assumed to be much smaller than the stress and thermal relaxation times, t_s and t_{th} respectively³⁷. Here, t_s and t_{th} are expressed as

$$t_s = \frac{d_c}{v}, \quad t_{th} = \frac{d_c^2}{\alpha_{th}}, \quad (9)$$

where d_c denotes the characteristic dimension of the heated region (lateral beam size of ~5 mm); v and α_{th} are the sound velocity and the thermal diffusivity, respectively. The values of these parameters are listed in Table 1. From v and α_{th} for water, t_s and t_{th} were calculated as 3.3 μ s and 170 μ s, respectively.

Under the conditions of stress and thermal confinement, the initial pressure rise due to the transferred energy can be described as follows:

$$p_0(\vec{r}, t) = \Gamma E(\vec{r}, t). \quad (10)$$

$E(\vec{r}, t)$ and $p_0(\vec{r}, t)$ are the cumulative energy and the amplitude of the external pressure source at time t , respectively. Γ is the Gruneisen coefficient, which is a dimensionless quantity that describes the conversion of thermal energy to volumetric expansion and is calculated using the equation below:

$$\Gamma = \frac{\beta v^2}{C_p}. \quad (11)$$

Here β is the coefficient of volumetric thermal expansion and C_p is the heat capacity at constant pressure. Note that Γ for gold is ~32 times and ρ is ~19 times higher than those values for water. This suggests that gold markers will act as strong pressure sources during proton beam therapy.

Given the initial pressure rise $p_0(\vec{r}, t)$, the temporal and spatial propagation of the pressure wave, $p(\vec{r}, t)$, can be described by the following wave equation:

$$\nabla^2 p(\vec{r}, t) - \frac{1}{v^2} \frac{\partial^2 p(\vec{r}, t)}{\partial t^2} = -\frac{1}{v^2} \frac{\partial}{\partial t} \left(\frac{\partial p_0(\vec{r}, t)}{\partial t} \right). \quad (12)$$

The wave equations (Eq. (12)) were solved using the k-wave toolbox in Matlab³⁸. The time step was set to 30 ns to ensure stability (Courant–Friedrichs–Lewy number <1). The temporal and the spatial parts of $E(\vec{r}, t)$ are assumed to be separable as

$$E(\vec{r}, t) = E(\vec{r})G(t). \quad (13)$$

and the beam's pulse structure is assumed to be Gaussian, namely:

$$\frac{\partial G(t)}{\partial t} = \frac{1}{\sqrt{2\pi\sigma_p^2}} \exp \left[-\frac{t^2}{2\sigma_p^2} \right]. \quad (14)$$

Then the right-hand side of the wave equation (12) becomes

$$-\frac{1}{v^2} \frac{\partial}{\partial t} \left(\frac{\partial p_0(\vec{r}, t)}{\partial t} \right) = -\frac{\Gamma}{v^2} E(\vec{r}) \frac{\partial^2 G(t)}{\partial t^2}. \quad (15)$$

The energy transferred by protons, $E(\vec{r})$, was calculated using the Geant4 Monte Carlo simulation software (version 9.3)³⁹. To achieve sufficient statistics, 10^7 proton histories were tracked, and the transferred energy was calculated in terms of voxels. The lateral beam profile was assumed to be Gaussian with $\sigma = 5$ mm at the water surface. This condition can be met by proton-therapy machines currently in clinical use. We used Geant4's Standard Physics List option 3 configuration for the electromagnet processes, and the G4BinaryCascade and G4HadronElastic models for the inelastic and elastic hadron processes, respectively. The cut-off range for the secondary electrons, positrons and photons was 1 mm.

Simulation parameters. The following parameters were varied to understand their effect on the pressure waves: the width of beam pulse σ_p and the marker position Z_m . The width of beam pulse σ_p was varied from 30 to 500 ns. To obtain a clear signal, the pulse width was chosen in the $< \mu\text{s}$ range^{18,19}, which fulfils the confinement conditions. Moreover, Z_m was varied from 16 to 38 mm, which corresponds to a residual range of 15 mm and -7 mm for 60-MeV protons, respectively; the marker was therefore positioned at both pre- and post-Bragg peak regions in the tests.

References

1. Particle therapy co-operative group. Facilities in operation. <https://www.ptcog.ch/index.php/facilities-in-operation>.
2. Pedroni, E. *et al.* The 200-MeV proton therapy project at the Paul Scherrer Institute: conceptual design and practical realization. *Med. Phys.* **22**, 37–53 (1995).
3. Smith, A. *et al.* The M. D. Anderson proton therapy system. *Med. Phys.* **36**, 4068–4083 (2009).
4. Paganetti, H. Range uncertainties in proton therapy and the role of Monte Carlo simulations. *Phys Med Biol* **57**, R99–R117 (2012).
5. Yang, M. *et al.* Comprehensive analysis of proton range uncertainties related to patient stopping-power-ratio estimation using the stoichiometric calibration. *Phys Med Biol* **57**, 4095 (2012).
6. Knopf, A. C. & Lomax, A. *In vivo* proton range verification: a review. *Phys Med Biol* **58**, R131–R160 (2013).
7. Polf, J. C. *et al.* Prompt gamma-ray emission from biological tissues during proton irradiation: a preliminary study. *Phys Med Biol* **54**, 731 (2009).
8. Jones, K. C. *et al.* Proton beam characterization by proton-induced acoustic emission: simulation studies. *Phys Med Biol* **59**, 6549–6563 (2014).
9. Sulak, L. *et al.* Experimental studies of the acoustic signature of proton beams traversing fluid media. *Nucl. Instrum. Methods* **161**, 203–217 (1979).
10. Hayakawa, Y. *et al.* Acoustic pulse generation in excised muscle by pulsed proton beam irradiation and the possibility of clinical application to radiation therapy. *Journal of the Acoustical Society of Japan (E)* **9**, 255–257 (1988).
11. Hayakawa, Y. *et al.* Acoustic pulse generation in water by pulsed proton beam irradiation and its possible application to radiation therapy. *Jpn. J. Appl. Phys.* **28**, 217–219 (1989).
12. Tada, J. *et al.* Time resolved properties of acoustic pulses generated in water and in soft tissue by pulsed proton beam irradiation—a possibility of doses distribution monitoring in proton radiation therapy. *Med. Phys.* **18**, 1100–1104 (1991).
13. Hayakawa, Y. *et al.* Acoustic pulse generated in a patient during treatment by pulsed proton radiation beam. *Radiation Oncology Investigations* **3**, 42–45 (1995).
14. Assmann, W. *et al.* Ionoacoustic characterization of the proton Bragg peak with submillimeter accuracy. *Med. Phys.* **42**, 567–574 (2015).
15. Jones, K. C. *et al.* Experimental observation of acoustic emissions generated by a pulsed proton beam from a hospital-based clinical cyclotron. *Med. Phys.* **42**, 7090–7097 (2015).
16. Lehrack, S. *et al.* Submillimeter ionoacoustic range determination for protons in water at a clinical synchrocyclotron. *Med. Phys.* **62**, L20–L30 (2017).
17. Terunuma, T. *et al.* Waveform simulation based on 3D dose distribution for acoustic wave generated by proton beam irradiation. *Med. Phys.* **34**, 3642–3648 (2007).
18. Ahmad, M. *et al.* Theoretical detection threshold of the proton-acoustic range verification technique. *Med. Phys.* **42**, 5735–5744 (2015).
19. Kipergil, E. A. *et al.* An analysis of beam parameters on proton-acoustic waves through an analytic approach. *Phys. Med. Biol.* **62**, 4694–4710 (2017).
20. Jones, K. C. *et al.* Acoustic-based proton range verification in heterogeneous tissue: simulation studies. *Phys. Med. Biol.* **63**, 025018 (2018).
21. Giebeler, A. *et al.* Dose perturbations from implanted helical gold markers in proton therapy of prostate cancer. *J. Appl. Clin. Med. Phys.* **10**, 63–70 (2009).
22. Habermehl, D. *et al.* Evaluation of different fiducial markers for image-guided radiotherapy and particle therapy. *J. Radiat. Res.* **54**, i61–i68 (2013).
23. Fujii, Y. *et al.* A simulation study on the dosimetric benefit of real-time motion compensation in spot-scanning proton therapy for prostate. *J. Radiat. Res.* **58**, 591–597 (2017).
24. Bert, C. & Durante, M. Motion in radiotherapy: particle therapy. *Phys. Med. Biol.* **56**, R113 (2011).
25. Kanehira, T. *et al.* Impact of real-time image gating on spot scanning proton therapy for lung tumors: a simulation study. *Int. J. Radiat. Oncol. Biol. Phys.* **97**, 173–181 (2017).
26. Kalinichenki, AI. *et al.* Introduction to Radiation Acoustics, The physics and technology of particle and photon beams, volume 9, harwood academic publishers (2001).
27. Wang, LV. (ed.). Photoacoustic Imaging and Spectroscopy. CRC Press (2009).
28. Culjat, M. O. *et al.* A review of tissue substitutes for ultrasound imaging. *Ultrasound. Med. Biol.* **36**, 861–873 (2013).
29. Cho, J. *et al.* Feasibility of proton-activated implantable markers for proton range verification using PET. *Phys. Med. Biol.* **58**, 7497–7512 (2013).
30. Lu, H. M. *et al.* Investigation of an implantable dosimeter for single-point water equivalent path length verification in proton therapy. *Med. Phys.* **37**, 5858–5866 (2010).
31. Patch, S. K. *et al.* Thermoacoustic range verification using a clinical ultrasound array provides perfectly co-registered overlay of the Bragg peak onto an ultrasound image. *Phys. Med. Biol.* **61**, 5621–5638 (2016).
32. Alsanea, F. *et al.* Feasibility of RACT for 3D dose measurement and range verification in a water phantom. *Med. Phys.* **42**, 937–946 (2015).
33. Kellnberger, S. *et al.* Ionoacoustic tomography of the proton Bragg peak in combination with ultrasound and optoacoustic imaging. *Scientific Reports* **6**, Article number 29305, (2016).
34. Hunt, J. W. & ARDITI, M. Ultrasound Transducers for Pulse-Echo Medical Imaging. *IEEE. trans. biomed. eng.* **30**, 453–481 (1983).
35. Ishi, Y. *et al.* Status report on FFAG accelerator complex at KURRI. *Proc of the 12th Annual Meeting of Particle Accel. Soc. of Japan*, 368–370 (2015).

36. Kuriyama, Y. *et al.* Status and Development of a Proton FFAG Accelerator at KURRI for ADSR Study, *Proc of 2011 Particle Accelerator Conference*, THP027 2172–2174 (2011).
37. Wang, L. V. & Wu, H. I. *Biomedical optics: principles and imaging*. John Wiley & Sons (2012).
38. Treeby, B. E. & Cox, B. T. k-Wave: MATLAB toolbox for the simulation and reconstruction of photoacoustic wave fields. *J. Biomed. Opt.* **15**, doi: Artn 02131410.1117/1.3360308 (2010).
39. Allison, J. *et al.* Recent developments in Geant4. *Nucl. Instrum. Methods A* **835**, 186–225 (2016).

Acknowledgements

We would like to thank prof. Nobuki Kudo for helping us understand the instrumentation in acoustic measurement and valuable discussion. This research was supported by MEXT/JSPS KAKENHI Grant Number JP 18K07621 and by the “Global Institution for Collaborative Research and Education (GI-CoRE), Hokkaido University,” founded by the MEXT, Japan.

Author Contributions

T. Matsuura devised the conceptual design and outline of the study. T. Takayanagi and T. Uesaka implemented the numerical simulations. T. Takayanagi, T. Uesaka, M. Kitaoka, M.B. Unlu and T. Matsuura analyzed the results and discussed the theoretical background. T. Takayanagi, M.B. Unlu and T. Matsuura wrote the manuscript in consultation with K. Umegaki, H. Shirato and L. Xing.

Additional Information

Competing Interests: We disclose conflict of interest as following; Authors Taisuke Takayanagi and Masanori Kitaoka are paid from Hitachi, Ltd., Tokyo, Japan.

Publisher’s note: Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article’s Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article’s Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

© The Author(s) 2019