

&

# 8<sup>th</sup> Kurt Schwabe Symposium

# Book of Abstracts

May 27 – 30, 2019, Split, Croatia

7<sup>th</sup> Regional Symposium on Electrochemistry – South East Europe & 8<sup>th</sup> Kurt Schwabe Symposium

Split, Croatia, May 27-30, 2019

### **Book of Abstracts**

Published by International Association of Physical Chemists E-mail: office@iapchem.org, URL: <u>http://www.iapchem.org</u>

For Publisher

Zoran Mandić

Editors

Višnja Horvat-Radošević, Krešimir Kvastek, Zoran Mandić

Design, Page Making and Computer Layout

Aleksandar Dekanski

ISBN 978-953-56942-7-4

On Line version only

## Surface modification of titanium implants by adherent hydroxyapatite/titanium oxide composite coatings using novel *in-situ* synthesis

Marijana Pantović Pavlović<sup>1</sup>, <u>Sanja Eraković<sup>1</sup></u>, Miroslav Pavlović<sup>1</sup>, Ljiljana Veselinović<sup>2</sup>, Jasmina Stevanović<sup>1,4</sup>, Vladimir Panić<sup>1,3,4</sup>, Nenad Ignjatović<sup>5</sup>

<sup>1</sup>Institute of Chemistry, Technology and Metallurgy, Department of Electrochemistry, University of Belgrade, Belgrade, Serbia <sup>2</sup>Vinča Institute of Nuclear Sciences, University of Belgrade, Belgrade, Serbia

<sup>3</sup>Centre of Excellence in Environmental Chemistry and Engineering-ICTM, University of Belgrade, Serbia <sup>4</sup>State University of Novi Pazar, Department of Chemical-Technological Sciences, Novi Pazar, Serbia <sup>5</sup>Institute of Technical Science of the Serbian Academy of Sciences and Arts, Belgrade, Serbia sanja@ihtm.bg.ac.rs



Figure 1. FE-SEM microphotographs of titanium surface after removing anHAP/TiO<sub>2</sub> coating.

The medical devices based on titanium and its alloys are widely used in the repair and replacement of a degraded or inhibited function of locomotor system [1]. Ti and its alloys exhibit high mechanical strength, good workability, resistant to corrosion and low cost. Although, they are widely used as orthopedic and dental implants their inability to interact with living tissue will inhibit their biological fixation and osseointegration [2]. Therefore, to improve the hard-tissue compatibility of Ti various surface treatments have been developed for the inorganic coating formation [3]. The hydroxyapatite (HAp,  $Ca_{10}(PO_4)_6(OH)_2$ ) with superior osteogenic activity is a competitive approach to make novel coatings for titanium implants applications. HAp is a calcium phosphate very similar to the inorganic part of the human bone and hard tissues both in morphology and composition. Herein, in-situ synthesis of HAp/TiO2 coating on titanium was performed via anaphoretic deposition of HAp and simulta-

neous anodization of Ti to produce highly adherent and strengthened composite coating. It can be seen that morphology of Ti substrate of anHAP/TiO<sub>2</sub> coating is of tubular shape, and tube formation occurs mainly due to competing processes of anodization and electrophoretic deposition of HAp. anHAp/TiO<sub>2</sub> coating does not need sintering process, and simultaneous Ti anodization and HAp deposition occur, where HAp crystals incorporate in the anodized Ti surface. From the presented results it can be concluded that novel suggested process of *in situ* simultaneous anHAp/TiO<sub>2</sub> deposition with Ti surface anodization gives much better results that cathaphoretic deposition regarding adhesion.

#### References

- 1. S.A. Ulasevich, A.I. Kulak, S.K. Poznyak, S.A. Karpushenkov, A.D. Lisenkov, E. V Skorb, RSC Adv. 6 (2016) 62540– 62544. doi:10.1039/C6RA10560B.
- 2. J. Hieda, M. Niinomi, M. Nakai, K. Cho, A. Matsubara, Mater. Trans. 56 (2015) 1703–1710.
- N. Eliaz, S. Shmueli, I. Shur, D. Benayahu, D. Aronov, G. Rosenman, Acta Biomater. 5 (2009) 3178–3191. doi:http://dx.doi.org/10.1016/j.actbio.2009.04.005.