brought to you by & CORE

Technical note

1

6

- 2 An experimental evaluation of fracture movement in two alternative
- 3 tibial fracture fixation models using a vibrating platform
- 4 Mehran Moazen¹; Peter Calder²; Paul Koroma²; Jonathan Wright²; Stephen Taylor²;
- 5 Gordon Blunn^{2,3}
- 7 ¹Department of Mechanical Engineering, University College London, Torrington Place,
- 8 London, WC1E 7JE, UK.
- 9 ²Institute of Orthopaedics and Musculo-Skeletal Science, Division of Surgery &
- 10 Interventional Science, University College London, Royal National Orthopaedic Hospital,
- 11 Stanmore HA7 4LP, UK.
- 12 ³School of Pharmacy and Biomedical Sciences, University of Portsmouth, Portsmouth, P01
- 13 2UP, UK.

14

15

19

16 Corresponding author:

- 17 Mehran Moazen, BSc, PhD, CEng, MIMechE, FHEA
- 18 Lecturer in Biomedical Engineering
- 20 UCL Mechanical Engineering,
- 21 University College London,
- 22 Torrington Place,
- 23 London WC1E 7JE, UK
- 24
- 25 E: M.Moazen@ucl.ac.uk
- 26 T: +44 (0) 207 679 3862

Abstract:

Several studies have investigated the effect of low-magnitude-high-frequency vibration on the outcome of fracture healing in animal models. The aim of this study was to quantify and compare the micromovement at the fracture gap in a tibial fracture fixed with an external fixator in both a surrogate model of a tibial fracture and a cadaver human leg under static loading, both subjected to vibration. The constructs were loaded under static axial loads of 50, 100, 150 and 200 N and then subjected to vibration at each load using a commercial vibration platform, using a DVRT sensor to quantify static and dynamic fracture movement. The overall stiffness of the cadaver leg was significantly higher than the surrogate model under static loading. This resulted in a significantly higher facture movement in the surrogate model. Under vibration the fracture movements induced at the fracture gap in the surrogate model and the cadaver leg were 0.024±0.009 mm and 0.016±0.002 mm respectively, at 200N loading. Soft tissues can alter the overall stiffness and fracture movement recorded in biomechanical studies investigating the effect of various devices or therapies. While the relative comparison between the devices or therapies may remain valid, absolute magnitude of recordings measured externally must be interpreted with caution.

Keywords: biomechanics, fracture fixation, stiffness, fracture movement, external fixator, cadaver, tibiae, non-union

Introduction:

Ilizarov Frame hexapods of various design are typically used in the management of long bone fractures in the field of orthopaedic trauma. Non-union fractures remain a challenge however, and account for around 10% of all fractures treated and about 2% of tibial diaphyseal fractures. There are several contributing factors to non-union including the patient, injury and treatment protocols. The stability of initial fracture fixation and post-operative loading of the fracture are among the key treatment related factors. Both contribute to the mechanobiology of the healing fracture where it is well established that there are certain strain thresholds that promote callus formation. For example, interfragmentary motion (IFM) in the range of 0.2-1mm and 2-10% strain is suggested to improve fracture healing.

There are some studies suggesting that the application of low-intensity pulsed ultrasound (LIPUS) and whole body vibration (WBV) may possibly improve fracture healing and potentially address non-union. The exact mechanisms by which these methods improve fracture healing at the molecular and cellular level are still unknown. However, it is generally accepted that LIPUS generates nano-scale motions while WBV generates micro-scale motion at the fracture site leading to different mechanisms of improved healing.

There has been no prior study to quantify the movement induced at the fracture gap as a result of external vibration in a tibial fracture fixed with an external fixator. Surrogate bone models and cadaveric tissue can be used to compare the fracture movement in an *in vitro* study. An *in vitro* fixation of a surrogate bone in the absence of soft tissues should provide little attenuation to vibration applied at the foot when observed at the fracture gap. Whilst, in a cadaver model, the magnitude of the displacement induced by the vibrating platform at the foot may be attenuated by the presence of the soft tissues. Incremental fracture displacements of 1mm/day are usually induced clinically using an external frame, although the soft tissues and bone remodelling stiffness determine the actual mode of distraction at the fracture gap. In this study we were not able to replicate the bone remodelling response, but just the soft tissues. However, the growing bone formed during distraction osteogenesis would have a low modulus of elasticity compared to mature bone.

The aim of this study was to quantify and compare the micromovement at the fracture gap in a tibial fracture fixed with an external fixator in both a surrogate model and cadaver leg under both static loading and subject to vibration. Therefore, the study quantifies and compares the overall stiffness and fracture movement in both models, and investigates if comparable trends exist between the two models. Cadaver studies are more challenging to perform than the surrogate models, but are more realistic.

Materials and Methods:

Specimens: A fourth generation tibia was purchased from Sawbones Worldwide (SKU:3402-overall length: 405mm; tibia plateau diameter: 84mm; distal tibia diameter: 58mm; mid shaft diameter: 10mm - WA, USA) and a left cadaver leg including all the soft tissue from the knee below was obtained from Anatomy gifts registry (Sex: male; Age: 62; body weight: 56 kg - MD, USA). The host institute had all the required approvals to perform this study. A transverse osteotomy was performed in each model using an oscillating saw (D_EWALT - MD, USA). In the surrogate model the sawbone tibia was cut. In the cadaver leg, the tibia and fibula were divided using a minimally invasive technique that preserved the soft tissues. Both transections were made in the mid-disphyseal region.

The tibiae in both cases were stabilized with an external fixator (Taylor Spatial Frame - Smith & Nephew plc, TN, USA). This is shown schematically in Fig 1. A two-ring Taylor Spatial Frame (TSF) construct was used with two proximal half pins and two distal half pins with a 90-degree divergence between the pins on each ring. The external fixator was then extended to produce a 50 mm fracture gap in the surrogate model, this was to ensure that the bony fragments did not come into contact during the experimental loading. In the case of the cadaver specimen a 13 mm fracture gap was produced, and further extension to match the surrogate model was not possible without overstretching the soft tissue (see Fig 1). This is a clinically typical fixation, although such fractures might be fixed with additional pins/wires pending various patient and injury related factors. Considering that in this biomechanical study the surrogate and cadaver models were fixed in the same configuration, the relative differences in outcome should remain valid.

Loading and measurements: The specimens were then fixed proximally to a material testing machine (Zwick Testing Machines Ltd., Herefordshire, UK) and distally rested on a commercial vibrating platform (Juvent, FL, USA - 0.3g's of acceleration at 32-35 Hz with 0.05mm vertical displacement). It must be noted that (1) the vibrating platform first finds the resonant frequency of the system and then initiates the vibrations see the manufacturer website and previous studies describing and evaluating this system. e.g. 20-22; (2) the natural frequency of a complete leg has been reported to be about 0.85Hz²³ while we are not confident if this has been picked up by the vibrating platform but we are confident that the vibrating frequency applied by the platform is well away from the natural frequency of the leg. A titanium "foot" was used to ensure direct contact between the surrogate tibia and the vibrating platform, while in the case of the cadaver leg the specimen was in contact with the vibrating platform through the foot.

The specimens were loaded five times under static axial loads of 50, 100, 150 and 200 N equivalent to partial weight bearing. And 24,25 Note, normal limb loads are approximately of 3xBW, but the use of far lower loads here is due to the fact that the subjects do not weight bear significantly during distraction osteogenesis, and are in line with measurements of frame loads carried out in author's lab. At the end of each loading scenario (1) the overall stiffness of the constructs were calculated based on the load-displacement data from the material testing machine. (2) The displacement at the fracture gap under the static loads was recorded with a caliper (with the resolution of 0.01 mm) on the lateral side. (3) The vibrating platform was turned on to vibrate the tibial shaft along its long axis. The fracture gap vibration (differential displacement across the medial fracture side) and the platform vibration were recorded using displacement sensors (with the resolution of 0.001 mm - DVRT- LORD MicroStrain, VT, USA) configured to LabVIEW (National Instruments, TX, USA).

Independent (two sample) t-test was used to compare the overall stiffness between the surrogate and cadaver models at 200 N loading. A dependent (paired) t-test was used to compare the difference between the displacement applied via the vibrating platform and the fracture movement both in the surrogate model and the cadaveric specimen. Significance level was set at p<0.05.

Results:

123

124125

126

127

128

129

130

131

132

133

134

135

136

137

138

139

- 141 Static loading: The overall stiffness of the surrogate model was 6.39±0.57 N/mm, and of the
- cadaver leg was 47.46±0.74 N/mm, based on the load-displacement data at 200N (p<0.05 -).
- The fracture movement at the lateral side of the surrogate model and cadaver leg increased
- 144 linearly (R²=0.9) from 2.82±0.13mm and 0.23±0.07 mm under 50 N to 10.99±1.40 mm and
- 145 0.96± 0.08 under 200 N respectively (Fig 3).
- 146 Dynamic loading: In the surrogate model, there was no significant difference between the
- displacement applied via the vibrating platform (platform vibration) and the fracture movement
- induced at the fracture gap (fracture gap vibration) under each loading scenario, figure 4. The
- displacement applied via the vibrating platform was however always higher than the fracture
- 150 gap displacement. Average platform and fracture gap displacement (due to the vibration)
- across all loading scenarios were 0.030 ±0.006 mm and 0.025 ±0.008 mm respectively
- 152 (significant difference p<0.05 Fig 4).
- 153 In the cadaver leg, there was a statistically significant difference between the displacement
- applied via the vibrating platform (platform vibration) and the fracture movement induced at
- the fracture gap (fracture gap vibration) under each loading scenario. Average platform and

fracture gap displacement (due to the vibration) across all loading scenarios were 0.027 ±0.002 mm and 0.013 ±0.003 mm respectively (significant difference - p<0.05 – Fig 4).

There was found to be a significant difference between the amount of displacement of the vibrating platform between the surrogate model (0.030 \pm 0.006 mm) and the cadaver leg specimen (0.027 \pm 0.002 mm) during vibration across all loading scenarios.

161

162

163

164

165

166

167

168169

170

171

172

173

174

175

176

177

178

179

180

181

182

183

184

185

186

187

188

158

159

160

Discussion:

A tibial fracture, fixed with an external fixator, was tested experimentally in a surrogate model and a cadaveric leg. The constructs were statically loaded and then subjected to vibration with a commercial vibration platform, at each load interval, to quantify fracture movement as a result of static loading and then with vibration.

The results highlighted a significant difference (eight times) between the overall stiffness of the surrogate model and the cadaveric leg (Fig 2). This is mainly due to the presence of soft tissues and the fibula in the cadaver model. However, other factors could have been contributing to the difference observed here. The frame constructs may not have been identically positioned resulting in different biomechanical properties.

A linear pattern of increase in fracture movement was observed in both cases due to the linear increase of loading from 50 to 200 N (Fig 3). However, there was about one order of magnitude difference between the fracture movement data obtained from the surrogate model and the cadaver leg. This was not surprising given the lower overall stiffness recorded for the surrogate model. In the case of the cadaver leg at 200 N, corresponding to partial weight bearing, fracture movement of 0.96±0.08 mm was measured. This is within the acceptable 0.2-1 mm fracture movement that is suggested to promote callus formation and enhance the healing process.^{4,5,7,11} In distraction osteogenesis, the TSF is typically extended by 1mm/day clinically. From figure 3B this would correspond to 210N at the bone ends. This seems to agree well with data from an instrumented fixator used in a clinical study 26,27 thus indicating that the stiffness of the cadaver tissues is likely to be similar to normal. Distal vibration of the tibia led to vibration at the fracture gap in both the surrogate model and cadaver leg. In the cadaver leg, a significant difference was observed between the displacement applied via the vibrating platform (0.027±0.003 mm - averaged over all tests) and the fracture movement (0.013±0.003 mm- averaged over all tests – see Fig 4). The difference between the two displacements at the fracture gaps is likely to have been altered by the soft tissues in the cadaver leg, and highlights the contribution made by the soft tissues to both static and dynamic stiffness.

This study has several limitations but perhaps the key limitation is that only one surrogate model and one cadaveric leg were used. While the study would have benefited from a larger sample size, the authors think that the differences captured in this study will remain valid with a larger sample size. Note, considering that only one surrogate and one cadaver leg were used in this study (while several tests were carried out) statistical analysis data must be considered with caution. Further in vivo studies are required to test the hypothesis that whole body vibration can improve the fracture healing process in humans and to investigate the effect of different frequencies, since only one frequency band was used here. Depending on the frequency and magnitude of the load, other vibrational regimes may also be osteogenic. In this paper we have chosen to investigate one level and suggest that this would be osteogenic.

In summary, this study has highlighted the effect of soft tissues in biomechanical studies. Soft tissues can alter the overall stiffness and fracture movement recorded in biomechanical studies investigating the effect of various devices or therapies. While the relative comparison between the devices or therapies may remain valid, absolute magnitude of recordings in such studies must be interpreted with caution.

Acknowledgments

This work was supported by EPSRC-NIHR Healthcare Technology Co-operative (HTC) Partnership Award – Unify: A network focussed on the treatment of fracture non-unions (EP/M000230/1) and UNIFY Plus (EP/N027221/1). Moazen was also supported by a Royal Academy of Engineering Research Fellowship (10216/119). We also thank Mr Arsalan Marghoub and Drs Chaozong Liu, Daniel Fong, Kaddour Bouazza-Marouf, Owen Davis and Alastair Campbell Ritchie.

Conflict of interest

The authors confirm that there is no conflict of interest in this manuscript.

References:

219

249250

251

252

253

220 1. Zimmermann G, Moghaddam A. Trauma: non-union: new trends. In Bentley G, editor. 221 European Instructional Lectures. 2010, p.15-19. 222 223 2. Copuroglu C, Calori GM, Giannoudis PV. Fracture non-union: who is at risk? *Injury* 224 2013; 44:1377-1672. 225 226 3. Court-Brown CM, Caesar B. Epidemiology of adult fractures: a review. *Injury* 2006; 227 37:691-697. 228 229 4. Goodship AE, Kenwright J. The influence of induced micromovement upon the healing 230 of experimental tibial fractures. J Bone Joint Surg Br. 1985; 67:650-5. 231 232 5. Augat P, Merk J, Ignatius A, et al. 1996. Early, full weight bearing with flexible fixation 233 delays fracture healing. Clin Orthp Relat Res 1985; 328:194-202. 234 235 6. Hak DJ. Management of aseptic tibial non-union. J Am Acad Orthop Surg 2011;19:563-236 573. 237 238 7. Moazen M, Jones AC, Leonidou A, Jin Z, Wilcox RK, Tsiridis E. Rigid versus flexible 239 plate fixation for periprosthetic femoral fractured computer modelling of a clinical case. Med Eng Phys 2012; 34:1041-1048. 240 241 242 8. Moazen M, Mak JH, Etchels LW, Jin Z, Wilcox RK, Jones AC, Tsiridis E. The effect 243 of fracture stability on the performance of locking plate fixation in periprosthetic 244 femoral fractures. J Arthroplasty 2013; 28:1589-1595. 245 9. Leonidou A, Moazen M, Skrzypiec DM, Graham SM, Pagkalos J, Tsiridis E, Evaluation 246


Leonidou A, Moazen M, Skrzypiec DM, Graham SM, Pagkalos J, Tsiridis E, Evaluation
of fracture topography and bone quality in periprosthetic femoral fractures: A
preliminary radiographic study of consecutive clinical data. *Injury* 2013;44: 1799-1804.

10. Leonidou A, Moazen M, Lepetsos P, Graham SM, Macheras GA, Tsiridis E. The biomechanical effect of bone quality and fracture topography on locking plate fixation in periprosthetic femoral fractures. *Injury* 2015; 46:213-217. 254 11. Claes L, Wilke HJ, Augat P, Rübenacker S, Margevicius KJ. Effect of dynamization of 255 gap healing of diaphyseal fractures under external fixation. Clin Biomech 1995; 8:227-256 234. 257 258 12. Egol KA, Kubiak EN, Fulkerson E, Kummer FJ, Koval KJ. Biomechanics of locked 259 plates and screws. J Orthop Trauma 2004;18:488-493. 260 261 13. Ogrodnik PJ, Moorcroft CI, Thomas PBM. A fracture movement monitoring system to 262 aid in the assessment of fracture healing in humans. Proc Instn Mech Engrs Part H 263 2001;215:405-414. 264 14. Leung KS, Shi HF, Cheung WH, Qin L, Ng WK, Tam KF, Tang N. Low-magnitude high-265 frequency vibration accelerates callus formation, mineralization, and fracture healing in rats. J Orthop Res. 2009; 27:458-465. 266 267 268 15. Kasturi G, Adler RA. Mechanical means to improve bone strength: ultrasound and 269 vibration. Curr Rheumatol Rep. 2011;13:251-256. 270 271 16. Wehrle E, Liedert A, Heilmann A, et al. The impact of low-magnitude high frequency 272 vibration on fracture healing is profoundly influenced by the oestrogen status in mice. 273 Dis Model Mech. 2015; 8: 93-104. 274 275 17. Harrison A, Lin S, Pounder N, Mikuni-Takagaki Y. Mode & mechanism of low intensity 276 pulsed ultrasound (LIPUS) in fracture repair. Ultrasonics 2016;70:45-52. 277 278 18. Thompson WR, Yen SS, Rubin J. Vibration therapy: clinical applications in bone. Curr 279 Opin Endocrinol Diabetes Obes. 2014; 21:447-453. 280 281 19. Goodship AE, Lawes TJ, Rubin CT. Low-magnitude high-frequency mechanical 282 signals accelerate and augment endochondral bone repair: preliminary evidence of 283 efficacy. J Orthop Res. 2009; 27:922-930. 284 285 20. Fritton JC, Rubin CT, Qin YX, McLeod KJ. Whole-body vibration in the skeleton: 286 development of a resonance-based testing device. Ann Biomed Eng. 287 1997;25(5):831-9. 288


21. Rubin C, Pope M, Fritton JC, Magnusson M, Hansson T, McLeod K. Transmissibility of 15-hertz to 35-hertz vibrations to the human hip and lumbar spine: determining the physiologic feasibility of delivering low-level anabolic mechanical stimuli to skeletal regions at greatest risk of fracture because of osteoporosis. Spine (Phila Pa 1976). 2003; 1;28(23): 2621-7. 22. Alizadeh-Meghrazi M, Zariffa J, Masani K, Popovic MR, Craven BC. Variability of vibrations produced by commercial whole-body vibration platforms. J Rehabil Med. 2014; 46(9):937-40. 23. Wagenaar RC, van Emmerik RE. Resonant frequencies of arms and legs identify different walking patterns. J Biomech. 2000; 33; 853-861. 24. Ryf CR, Arraf J. Postoperative fracture treatment: general considerations. In: Ruedi TP, Buckley RE, Moran CG, editors. AO principal of fracture management. 2nd ed. Davos: AO Publishing; 2007. p. 447. 25. Ganse B, Yang PF, Gardlo J, Gauger P, Kriechbaumer A, Pape HC, Koy T, Müller LP, Rittweger J. Partial weight bearing of the tibia. *Injury*. 2016; 47:1777-1782. 26. Laubscher M, El-Tawil S, Ibrahim I, Mitchell C, Smitham PJ, Chen P, Goodier D, Gorjon J, Richards R, Taylor SJG, Calder P. Measurement of forces across a Taylor Spatial Frame during activity. Orthop Proc. 2015; 97-B, supp_5. 27. Simpson AH, Cunningham JL, Kenwright J. The forces which develop in the tissues during leg lengthening. A clinical study. J Bone Joint Surg [Br] 1996;78-B:979-83.

323	Figure captions:
324 325	Fig 1: Schematic of the experimental set up in a lateral view: (A) the surrogate model, (B) the cadaver leg.
326 327	Fig 2: Overall stiffness of the fracture fixation constructs. Note standard deviation is for 5 number of repeats of the axial compression test. * highlight significant difference.
328 329	Fig 3: Fracture movement induced via static loading in the surrogate model (A) and the cadaver leg (B).
330 331 332	Fig 4: Fracture movement induced via the vibrating platform in the surrogate model (A) and the cadaver leg (B). Note standard deviation is for five number of repeats of the axial compression test. * highlight significant difference.
333	
334	
335	
336	
337	
338	
339	
340	
341	
342	
343	
344	
345	
346	
347	


348 Fig 1




350 Fig 2



354 Fig 3





356 Fig 4

