

April 2019

Thermal Infrared Endoscope

Chenggu Wang

Worcester Polytechnic Institute

James Vahe John Kradjian

Worcester Polytechnic Institute

Matthew Sean Collins

Worcester Polytechnic Institute

Ryan James St. Hilaire

Worcester Polytechnic Institute

Wentao Yuan

Worcester Polytechnic Institute

Follow this and additional works at: <https://digitalcommons.wpi.edu/mqp-all>

Repository Citation

Wang, C., John Kradjian, J. V., Collins, M. S., St. Hilaire, R. J., & Yuan, W. (2019). *Thermal Infrared Endoscope*. Retrieved from <https://digitalcommons.wpi.edu/mqp-all/6819>

This Unrestricted is brought to you for free and open access by the Major Qualifying Projects at Digital WPI. It has been accepted for inclusion in Major Qualifying Projects (All Years) by an authorized administrator of Digital WPI. For more information, please contact digitalwpi@wpi.edu.

Thermal Infrared Endoscope

A Major Qualifying Project

Submitted to the Faculty of Worcester Polytechnic Institute
in partial fulfillment of the requirements for the Degree in Bachelor of Science

in

Robotics Engineering

By

Matthew Collins (RBE&CS)

James Kradjian (RBE)

Ryan St. Hilaire (RBE&ME)

Chenggu Wang (RBE&ME)

Wentao Yuan (RBE)

Date: 4/25/19

Project Advisor: Professor Loris Fichera

Project Co-Advisor: Professor Gregory S. Fischer

Abstract

According to a 2017 report by Intuitive Surgical, the da Vinci Robotic Surgical System was used in approximately 877,000 surgeries that year, and this number is projected to increase in the future. During robotic surgical procedures, energy-based tools are frequently used to perform tasks such as cauterization, i.e. heating up tissue to coagulate blood vessels. However, the use of these instruments poses a risk of spreading thermal energy to surrounding tissues and causing accidental damage to vital anatomy. To help mitigate this hazard, here we propose an infrared-based, minimally invasive thermal camera endoscope that can be attached to the existing instrument arms of da Vinci robot. This device will enable surgeons to monitor, for the first time, the thermal interactions between surgical instruments and tissue, thus helping surgeons reduce the risk of accidental injury, secondary to heat. Our thermal endoscope consists of a set of modular tubes and independent transmission, allowing the camera to have a 180-degree view. The acquisition of the video stream generated by our device is processed by a Robot Operating System (ROS) node running on a Raspberry Pi, which then publishes the stream of thermal images as a ROS topic at a rate of 8 Hz (close to the highest frame rate offered by the infrared camera). We provide a sample application that displays the live thermal video on a computer screen and discuss how in the future this application could be integrated on the console of a da Vinci robot.

Acknowledgments

Thank you to Professor Fichera and Professor Fischer for advising this project.

Thank you to Katherine Crighton for facilitating this project by managing budget and ordering supplies.

Thank you to Aatreya Chakravarti and Keion Bisland for their contribution to the electrical and computer engineering aspects of this project.

Contributions

Matthew Collins	Software Design, C++ Publisher Code, MATLAB Subscriber, Camera Debugging, ROS Configuration, Camera Validation, MATLAB Display, Graphing, and Modalities, Text Fixture Design
James Kradjian	Software Design, C++ Publisher Code, Network Design and Configuration, Camera Debugging, ROS Configuration, Camera Validation, MATLAB Display and Graphing, Test Fixture Electrical Assembly and Testing
Ryan St. Hilaire	Mechanical Design, Prototyping, Machining, Shaft and Transmission Design, Test Fixture Coding
Chenggu Wang	Mechanical Design, Prototyping, Machining, Mechanical Integration and Testing, Kinematics Analysis
Wentao Yuan	ROS Integration, C++ Publisher Node, Matlab Thermal Data Display, Output Validation Analysis

Table of Contents

Abstract	1
Acknowledgments	2
Contributions	3
Table of Contents	4
List of Figures	6
List of Tables	7
Executive Summary	8
1. Introduction	11
2. Background	13
2.1 Literature Review	13
2.2 Da Vinci Surgical Platform	13
2.3 Thermal Imaging Camera	14
3. Materials and Methods	15
3.1 Hardware Design Logistics	15
3.1.1 Design Objectives and Functional Requirements	15
3.1.2 Design Constraints	16
3.1.3 Preliminary Designs	17
Notched-Tube Design	17
Modular Linkage Design	18
Comparison of Preliminary Designs	20
3.1.4 Final Modular Shaft Design	21
3.1.5 Designing Arm Transmission	23
3.2 Software and Electronics Logistics	26
3.2.1 Design Objectives and Functional Requirements	26
3.2.2 System Overview	27
ROS	27
Catkin	28
MATLAB	28
Custom PCB	29
Raspberry Pi	29

Networking Devices	30
Da Vinci Host Computer	30
3.2.3 Program Design	31
C++ ROS Publisher	31
MATLAB ROS Subscriber	31
Output Modalities	32
Gray and Log Gray	33
Hot and Log Hot	34
HSV and Log HSV	34
Jet and Log Jet	35
4. Results	37
4.1 Mechanical Results	37
4.1.1 Analysis of Mechanical Design and Functional Requirements	37
4.1.2 Kinematics	41
4.2 Software Results	42
4.2.1 Analysis of Software Design and Functional Requirements	42
4.2.2 Endoscope Output Validation Testing	44
5. Conclusion and Recommendations	49
Bibliography	51
Appendix A: Notched-Tube Model in MATLAB	54
Appendix B: Test Fixture Arduino Code	56
Appendix C: Flowchart of ROS Publisher Program	59
Appendix D: Flowchart of MATLAB Subscriber Program	60

List of Figures

- Figure 1: Da Vinci Cautery Tool and its Thermal Spread
- Figure 2: Final Mechanical Assembly
- Figure 3: Example Colormap, “jet”, with linear (left) and logarithmic (right) scaling
- Figure 4: Energy-based da Vinci instruments and their thermal spread
- Figure 5: Motor specifications for the da Vinci Research Kit(25)
- Figure 6: Notched-tube design for bi-directional bending
- Figure 7: Preliminary Design of the Linkage-Like Tube Model
- Figure 8: Improved Design of the Modular Link Model
- Figure 9: Solidworks model of the modular link
- Figure 10: Sectional view of the camera mount
- Figure 11: Thermal Infrared Endoscope Transmission
- Figure 12: Isometric Solidworks model of the torsional motor pulley
- Figure 13: Test Fixture for the dVRK
- Figure 14: System Architecture Diagram
- Figure 15: Example Colormap, “gray”, with linear (left) and logarithmic (right) scaling
- Figure 16: Example Colormap, “hot”, with linear (left) and logarithmic (right) scaling
- Figure 17: Example Colormap, “hsv”, with linear (left) and logarithmic (right) scaling
- Figure 18: Example Colormap, “jet”, with linear (left) and logarithmic (right) scaling
- Figure 19: Final Mechanical Build
- Figure 20: Modular Link Bend Radius
- Figure 21: Thermal Infrared Endoscope attached to the dVRK surgical robot
- Figure 22: Modular Link Bend Radius
- Figure 23: Boiling Water Validation
- Figure 24: Thermal Frame Publish Frequency
- Figure 25: Camera Validation Setup
- Figure 26: FLIR camera reading on object temperature at 55°C
- Figure 27: Reference camera reading on object temperature at 55°C

List of Tables

Table 1: Mechanical Design Comparison

Table 2: Camera Validation Test Data in Celsius Degrees

Executive Summary

The purpose of this project was to develop a thermal infrared endoscope that was suitable for minimally invasive robotic surgery. The rate of robot-assisted surgery has increased rapidly over the year with 877,000 procedures performed with da Vinci alone in 2017 [1]. Minimally invasive robotic surgery can be much more precise and much safer compared to open surgeries, however, thermal damage can occur because of the use of cauterization equipment. Nerve tissue is particularly sensitive to heat, and excess thermal spread can result in lifelong complications. As of the writing of this paper, no commercial thermal endoscopes exist on the market.




Figure 1: *Da Vinci Cautery Tool and Lateral Temperature Spread*
(adapted from Hefermehl et al. [8])

Using a combination of surgical robots and thermography, cases of thermal damage can be limited. Thermographic visualizations provide surgeons with live area temperature data that allows them to see both the current temperature of and the thermal spread due to the use of the tool (shown in Figure 1). The da Vinci Surgical Platform is a teleoperated robot used for surgery and can be used as a development platform for instruments. Current instruments include optical endoscopes, forceps, and cauterizers. Developments in the Long-Wave Infrared camera technology have allowed for the creation of a distal tip endoscopic thermal imaging camera of miniature scale to be practical in surgical applications.

In terms of mechanical design, the endoscope has two major features: the rotating shaft and the transmission. The rotating shaft is the part of the mechanism that holds the camera, while

the transmission is the feature that controls the motion and bending of the shaft. The two major design constraints for this mechanism is that the system must use the Lepton 3.5 camera and must be able to be mounted on a da Vinci surgical robot through conventional means. The preliminary designs involved building a notched-tube design for the shaft, but this was changed due to the complexity of the part. The design that was ultimately chosen was a linkage-like tube that used a series of modular links to bend. This solution was more robust and easier to create than the notched tube and had the benefit of a more discrete kinematic solution. For prototyping purposes, the modular links, camera enclosure, and shaft were 3D printed using SLA resin.

The modular links were linked between the main shaft and the camera enclosure and was able to bend the enclosure 60° from the central axis. The bending of the links was controlled by the transmission, an adapted da Vinci base plate that used a series of pulleys to pull tensioned cables, resulting in a bend radius. The transmission contained a mix of 3D printed, machined, and existing da Vinci parts, and this mechanism was successful in controlling the location of the infrared camera. For standalone testing, the team also built a test fixture with three stepper motors controlled by Arduino Uno. The kinematics for the modular link bend radius was generated, relating the bend radius to the steps on the stepper motor of the test fixture. The resulting relationship determined the number of steps per degree of bend radius, up to a maximum of 60° .




Figure 2: *Final Mechanical Assembly*

The goal of the software and electronics designs are to facilitate the formatting and presentation in an interpretable way to the surgeon. Several requirements were determined at the start of the project by the team. These requirements were set to minimize the complexity to the surgeon and other staff. The computer connected to the camera, the Raspberry Pi, is running ROS, or Robot Operating System, and is acting exclusively as a publisher node. The job of the Raspberry Pi is to transmit the images from the FLIR Lepton to the ROS network. From there, any subscribers can subscribe to the topic where the frame is at. In this case, the subscriber is a computer that is running MATLAB to interpret and display the frame. By using MATLAB, it is easy to visualize in different color maps and modalities.




Figure 3: *Example Colormap, “jet”, with linear (left) and logarithmic (right) scaling*

Utilizing the *jet* colormap, a surgeon can easily and intuitively determine the hottest regions in the frame. Depending on the preference, either a linear or logarithmic scale can be chosen. The output of the camera was validated using the FLIR A655sc to determine that the software that was written in a way that correctly interpreted and displayed the data.

The final iteration of the thermal infrared endoscope satisfies the majority of the requirements set by the project, but there are several mechanical and software augmentations that can increase the viability of the system as a medical instrument. The final product shows promise as an initial attempt to reduce the amount of heat-related injuries during robotic surgery and is useful in detecting the effects of energy-based instruments.

1. Introduction

The use of surgical robots has increased dramatically in the past few years. Between 2016 and 2017, the number of da Vinci surgeries has increased by approximately 16%, with nearly 877,000 procedures performed in 2017 [1]. These systems are mainly used for laparoscopic, or minimally invasive surgery, involving the use of thin instruments inserted through small cuts on the patient. Among all robot-assisted surgeries, approximately 75% of robot-assisted surgeries are used for urology or gynecology procedures, with nearly 85% of all prostatectomies performed in the United States using a surgical robot [2]. These surgical robots boast a quicker procedure time and a shorter recovery period, but typically have similar long-term recoveries as compared to open surgery. One study has shown that robot-assisted surgery is easier to learn for people without any previous surgery experience [3]. Surgical procedures utilizing the da Vinci are less expensive than open surgery as the latter causes a longer length of stay and additional costs [4].

While minimally invasive robotic surgery can be more precise and safer than conventional surgical methods, it is not without risks. Thermal damage is especially prevalent when using these systems as they use cauterization to seal blood vessels and maintain a clean work environment [6, 7, 8]. However, cauterizers must produce a significant amount of heat to fully seal the vessels, and the radiative heat can be damaging to the body, as shown in Figure 4. This can be very dangerous with surgeries that operate near a large volume of nerves. Nerves can be permanently damaged at temperatures as low as 45°C, and the cauterizers typically run between 40°C and 60°C [11]. This leaves a very small margin of error, and thus controlling the spread of heat is extremely important. An example of this is with the prostatectomies previously mentioned. If enough heat is radiated to the nerves in or near the prostate, the patient could lose all sensation and functionality in that region [5, 9, 10]. Due to these dangers, cautery tools are not normally used for these sensitive procedures, making the procedure more difficult and increasing the risk of potential complications.




Figure 4: *Energy-based da Vinci instruments and their lateral temperature spread*
(adapted from Hefermehl et al. [8])

The robotic surgical systems also introduce the risk of human operation error and mechanical failure. Several components on the da Vinci robot can malfunction or break during operation, compromising control or leaving a residue within the operation area [6]. This can result in complications and longer recovery times. Currently, there are no commercially available thermal endoscopes that can be used for robotic surgery. The motivation for this project is to develop a viable solution to minimize heat trauma during robotic surgery, thus making the surgeries safer and more effective.

2. Background

2.1 Literature Review

This idea of a thermographic-based surgical observation platform was brought up and tested by Lin et al. [1]. Thermographic visualization can provide surgeons with live temperature maps and monitor heat from cautery tool. Their team has experimented with a miniature IR camera Lepton made by FLIR Systems. Controlled experiments were set up with 2 Lepton cameras side by side and a higher resolution FLIR A655sc camera as a ground reference to obtain thermal images. All three cameras observed the thermal spread of a chicken specimen with statically matching data, thus proving the Lepton camera a viable option for observing thermal energy transmission across the soft tissue [11]. The team plans to improve upon this research both mechanically and digitally integrate the newest version of Lepton into an endoscopic IR camera. In addition, with the resources we have on the da Vinci surgical system, we are planning to attach the IR camera as one of the instrument arms on the da Vinci Research Kit (dVRK) and achieve mechanical control of the endoscope.

2.2 Da Vinci Surgical Platform

The da Vinci Surgical Platform is a teleoperated surgery robot. The system has four arms that instruments can be attached to, and each arm can be controlled independently with input from the surgeon. These arms can hold cauterizers, cameras, forceps, and other surgical instruments. During normal operation, three of the four arms are typically in use, with the fourth reserved as a spare. The da Vinci system activates instruments by activating one of the four motors embedded within each arm. These, in turn, rotate a set of pulleys housed within the transmission of the instrument, allowing for motion and actuation depending on the instrument attached. The system is extremely precise and can remain more steady when performing delicate procedures than a surgeon alone.

Da Vinci instruments are typically very small, with a diameter of 8mm and smaller. This is because the smaller the tool, the smaller the incision needs to be to insert the tool within the patient. The overall length of the tool is usually between 30cm-35cm to prevent the surgical arms from interfering with each other. To ensure that the instruments do not malfunction, the instruments are typically only rated for 10-30 uses before being replaced. To prepare the thermal infrared endoscope for similar surgical applications, the mechanism must be similar in build to the current instrumentation, and it will be attached to one of the da Vinci arms.

2.3 Thermal Imaging Camera

LWIR, or Long-Wave Infrared, cameras are commonly used for near room temperature detection by sensing infrared radiation within the range of 8-14 micrometers[12]. For use in minimally invasive surgical applications, size and reliability are key features to consider when choosing the camera model. Through our research, we found that FLIR manufactures a radiometric-capable camera model, the Lepton 3.5. It has a built-in calibration feature, includes 160 by 120 pixels and is also made to be relatively low cost compare to other traditional IR cameras. The Lepton 3.5 consists of a microbolometer array with a lens to focus. A microbolometer is a ranged temperature sensor in miniature scale, and a microbolometer array is many of these sensors put together, and therefore can form an image with temperature reading on every single pixel [13].

3. Materials and Methods

3.1 Hardware Design Logistics

This section lists the design goals and functional requirements for the mechanical aspects of this project. Starting from the given design constraints, this section covers the preliminary design approaches, comparison of design options, and final design details.

3.1.1 Design Objectives and Functional Requirements

The goal of the mechanical design is to build an endoscopic camera that can be inserted within a patient during minimally invasive surgery. This endoscope must be able to provide full visual coverage of the operation space during the surgery and must be able to navigate while under spatial constraints inside the patient, precisely following the teleoperation commands from the surgeon. With these design goals in mind, the team has composed a list of functional requirements listed below. The realization of each functional requirement will be discussed in detail in the following sections.

Functional Requirements:

1. The size of the custom imaging endoscope needs to be in close proximity with existing da Vinci instruments
2. The material of the endoscope needs to be rigid, non-deformable, and potentially bio-compatible
3. The endoscope needs to have at least two degrees of freedom, and be able to reach all designed orientations with proper actuation
4. The mechanism needs to tolerate wear within the average expected number of operations
5. The endoscope needs to be self-contained, all components of the electronics and mechanical transmission must be bounded within the instrument enclosure
6. The mechanism needs to be compatible with existing da Vinci system, allowing easy integration and testing using the dVRK

3.1.2 Design Constraints

There are a few design constraints proposed by the selected camera module. While being one of the smallest commercially available thermographic camera, the Lepton 3.5 camera [13] has a dimension of 9mm by 9mm, setting the lowest bound of the dimensions of the design. The 57° horizontal field of view of the camera also determines the maximum rotation of the shaft needed to fully observe the operation space. Additionally, the mechanism must be installed on and driven by a da Vinci surgical robot. This means that the system must be actuated by the drive motors already present on the dVRK system using a specific motor mounting configuration while staying below the maximum torque of the robot for Outer Roll (0.044 Nm, See Figure 5). Finally, the system must be self-contained and have all wiring held within the mechanism during operation. This includes the signal wires for the camera as well as the actuation pulleys.

Table 8 summarizes the default and the actual operating conditions of the motors used in the 'da Vinci' system.

TABLE 8: PSM ACTUATOR OPERATING CONDITIONS

#	Axis	Motor Type	Default Max.		Actual Max. Current		Torque Const (Nm/A)	Max. Torque (Nm)	Gear Ratio*	Encoder Counts/Rev
			Voltage (V)	Current (A)	(%)	(Amp)				
1	Outer Yaw	RE025-Twin**	24	1.340	150	2.010	0.043800	0.088	56.50	14400
2	Outer Pitch	RE025-Twin**	24	1.340	150	2.010	0.043800	0.088	56.50	14400
3	In/Out or Insertion	RE025-055-38	24	0.670	150	1.005	0.043800	0.044	336.6	14400
4	Outer Roll	RE025-055-38	24	0.670	150	1.005	0.043800	0.044	11.71	4000
5	Wrist Pitch	RE025-055-38	24	0.670	150	1.005	0.043800	0.044	11.71	4000
6	Wrist Yaw1	RE025-055-38	24	0.670	150	1.005	0.043800	0.044	11.71	4000
7	Wrist Yaw2	RE025-055-38	24	0.670	150	1.005	0.043800	0.044	11.71	4000

* Gear Ratio – the gain from the motor shaft to the actual joint

** RE025-Twin: It represents 2 RE025-055-38 in parallel configuration

da Vinci Research Kit User Guide

Figure 5: Motor specifications for the da Vinci Research Kit[14]

3.1.3 Preliminary Designs

The main design challenge for this project was to design a flexible shaft that could be controlled using tensioned pulleys. There were two major design paths in the early stages of this project. Based on common continuum instruments used in surgical applications, the team came up with a couple of preliminary models of the endoscope as described in this section.

Notched-Tube Design

The first design of our consideration was a notched-tube endoscope made out of nitinol, a superelastic nickel-titanium alloy. The notched-tube joint design has been utilized for various medical devices involving miniaturized wrists, and many of the mechanisms achieve directional compliance by creating notches on superelastic nitinol tube [15,16]. Actuation is achieved by applying tension on the tendon, usually through a stainless steel cable routed on the tip, causing the tube to bend. The kinematic behavior of this mechanism depends on the internal material deformation and the structure of the tube, which can be characterized using a series of transformation matrices provided by Eastwood et al. [17]. A script of the equations used written in MATLAB is included in Appendix A. These equations provide a theoretical model of the maximum bending angle, stress, and strain of the endoscope based on its geometry. In particular, once we input the number of notches, depth and width of notches, and the space in between them, the script will generate the theoretical maximum bending without considering the material strength. We then verify if a given material can sustain the amount of strain, and calculate the max stress based on Young's Modulus of the material. This process is repeated until an optimal geometry is found that works for the selected material and produces desired kinematic output.

The team considered three main criteria for the design: tube diameter, wall thickness, and maximum bending angle. To start with, we decided to keep the outer dimensions of the geometry the same as the original da Vinci instrument tubes. In terms of the bending angle, since our Lepton 3.5 camera has a field of view of 57 degrees horizontally, we decided that a 60 degree bending to both sides would achieve roughly 180-degree view. Therefore, with these parameters

in mind, we formulated the transformation matrices and generated the design shown in Figure 6 below.




Figure 6: *Notched-tube design for bi-directional bending*

Since nitinol is expensive to manufacture and difficult to obtain in small quantities, a few proof-of-concept tests were performed before pursuing the design with nitinol tube. The above model was 3D printed for testing using TPU elastic plastic. However, the result from the initial prototype was not ideal. The printout did not endure much tension before shattering and was not able to recover to the original orientation. This was due to the mechanical weakness that lies within the method of layer by layer material jetting (the most common type of 3D printing). It was also observed that the wall sections in between notches were subject to huge strain and became weak points in the bending motion. Overall, the part was not able to sustain bending stress and failed as a testing prototype.

In addition to the elastic-plastic test print, we also tested the notched-tube model on W1 tool steel. This material was chosen because of its similar resilience to spring steel and nitinol. However, due to its high Rockwell hardness and modulus of elasticity, the W1 tool steel was very difficult to machine into the notched tube and required an extreme amount of force to bend. As a result, it was decided this notched tube design is not suited for our endoscope design since the risk of using nitinol is too high after both concept models have demonstrated material failure.

Modular Linkage Design

An alternative design option the team had was a linkage-like modular tube design. Each link section is able to achieve some degree of bending individually and will be connected together by pin joints. Shown in Figure 7 below, the initial design of the modular linkage had an inner tube section and an outer tube section. The modular linkage can be stacked on top of itself and the pin holes will be held by two metal pins. The inside of this tube is completely hollow, leaving all the space for signal cables in addition to having two ditches for the tension cables.




Figure 7: *Preliminary Design of the Linkage-Like Tube Model*

After sending this model for 3D printing, the team soon realized the problem. Since one of these design goals is to keep the endoscope as small as possible, there was little room left for wall thickness especially around the pin connectors, leaving it extremely easy to fracture. Additionally, it was discovered that there was no significant advantage to have two different diameters for either end of the link. Moreover, the idea of adding a separation along the length of the tube was brought up during the design review process with the aim to provide cable separation, reduce tangling and provide additional structural support.




Figure 8: *Improved Design of the Modular Link Model*

Shown in Figure 8 above, the improved modular link design has a uniform diameter throughout its cylindrical shape. The inside has a separation wall running along the length of the linkage, connecting to the pin joints at both ends. To increase the structural stiffness at the pin connectors, we decided to make them into a hinge-like structure with odd and even ends. The connection pins were also replaced by one bolt going through the diameter of the shaft, allowing the hinges to rotate freely. After printing this improved model and testing it with our

setup, all aspects of the design was shown to be functional. The new hinge-like pin joints were rigid enough to complete designed bending, and the center of the shaft still offered enough space for the signal cables and tension cables.

Comparison of Preliminary Designs

In order to decide the feasibility of the two aforementioned designs, the team came up with a design matrix outlining the pros and cons of each design option, shown in Table 1 below.

Mechanical Design Comparison		
	Notched Tube Design	Modular Link Design
Usage in Surgical Applications	Have been proposed and tested as compliant teleoperation instrument (18)	Multi-Link instruments are commonly used for multiple DOF wrist mechanisms(19)
Material Requirements and accessibility	Requires superelastic nitinol tube, extremely difficult to obtain in small quantity	Can be 3D printed or custom machined, both methods are easily accessible
Prototype Test Result	Prototype testing failed, the mechanical properties of nitinol are not easily replaceable by other materials	Prototype testing successful, all components fulfilled designed purposes
Cost of Production	The high cost of manufacturing, cannot afford multiple iterations	Low cost of production, can build as many iterations as needed
Length of Design Cycle	Takes a long time to design, build, and analyze	Has short design cycles, can quickly improve on iterations

Table 1: *Mechanical Design Comparison*

The criteria for comparison are chosen to reflect the feasibility of each design options given our project timeline, budget, and access to equipment. Overall, the modular link design is a more realistic design option given the high cost and long design cycle of the notched tube design.

3.1.4 Final Modular Shaft Design

In the end, the team chose to pursue a modular link mechanism design with a customized transmission. To meet the required functional requirements and comply with the design limitations, a modularly linked shaft was designed. This mechanism consisted of a series of pinned links that interfaced with each other to bend at discrete angles. The mechanism would be manipulated with two tensioned cables and would remain kinematically stable as long as the cables were in proper tension. To house the camera electronics and tensioned cables, the shaft would be hollow with the cables running from the end effector to the shaft transmission. Additionally, the shaft could be rotated using a third tensioned cable wound around the circumference of the shaft.

The modular link was a significant feature to the modular shaft design and was the main feature that determined the bend angle. Shown in Figure 8, the modular link has two 15° facial cuts on both ends, measured from the center of the hinge. The camera mount (Figure 10) and the shaft (Figure 10) have similar 15° cuts, but only on the side where the modular link would attach to. When the hinges are connected between the camera mount, shaft, or another modular link, the total bend angle between the two features is 30° . The addition of the modular link allowed the mechanism to modify lengths as well as change the bend angle by modifying the facial cuts on the link. The link contains the same central partition as the shaft, shown as the central blue walls in Figure 9, and features specific, notched geometry to provide a crevice for the tensioned cable to ride in.




Figure 9: *Solidworks model of the modular link. Walls are highlighted in blue in the left image, and a top-down view is shown in the right image*

Based on the field of view of the camera, our team determined that the manipulator only needed to bend by a total of 60° from the central axis to achieve a full view of the operation space. To achieve this, each joint was designed with a 15° cut at the top and bottom faces. Thus, when bent, each linkage would result in 30° of bending to maximize the kinematic motion. While the mechanism can have as many links as desired, this particular design only required a singular link in the middle, assuming the camera mount and connector shaft had the same facial cuts, resulting in the required 60° bend.

The modular shaft also contained an internal partition to separate the tensioned cables. This was done to ensure that the cables would not twist together during shaft rotation. To maintain a symmetrical bend radius, the camera wires were split equally between the two partitions. This also helped to stabilize the mechanism and prevent the mechanism from moving erratically during testing with the stepper motors.

The camera mount is designed to carry the Lipton camera and the miniaturized PCB. Shown in Figure 10, the camera mount is designed to be as small as possible, with the diameter of the enclosure being slightly larger than the camera diameter. The final head dimensions for the camera mount are 14mm by 14mm, and the length is 40mm to allow enough space to house the soldered camera wires. The camera mount also has a pair of angled holes within the stem to pass the tensioned cables through. The tensioned cables are then fixed to the shaft with a

clamped sleeve such that the cable can not slip back through the holes. Thus, when the cables are pulled, the cable length effectively becomes shorter, bending the manipulator in the pulled direction.




Figure 10: *Sectional view of the camera mount. Note the two 45° holes in the stem of the mount for the cables to pass through*

Since the modular shaft had so many irregular features making it hard to machine, the shaft, modular link, and camera mount were 3D printed using an SLA resin on the Formlabs Form 2 printer. The print had a very high resolution of features, but sometimes had trouble retaining its shape during UV post cure. Additionally, the resin was not as strong as initially expected, any unexpected external force can cause the parts to break during testing and operation. This was especially prevalent between modular links as the hinges were very thin and could easily experience extreme shear forces if the links were pulled or if the shaft was bent beyond the 60° bend radius. The reason for using the SLA resin was because the Formlabs printer has high resolution and quick turnaround time whereas creating the shaft without 3D printing would have been extremely time consuming and less iterable.

3.1.5 Designing Arm Transmission

To control the chosen shaft design, a transmission system was built to interface with the existing dVRK system, shown in Figure 11. As mentioned previously in the design limitations, this transmission would be driven using the motors present on the da Vinci system. These motors would drive a series of transmission mounted disc under the baseplate, which would in turn wind a pulley mounted on each pole. These pulleys each control a singular degree of actuation by

pulling on tensioned cables. With the cables running through the shaft, the resulting motion would cause the end effector to move in direct response to the motor input. Thus, the transmission is used to convert motor rotation into the end effector movement. This simple system makes it straightforward for the team to develop a moving mechanism and helped us determine how the modular shaft system should be designed.




Figure 11: *Thermal Infrared Endoscope Transmission*

Initially, this transmission was to be built from scratch, but the team decided to instead reuse an existing da Vinci tool as it already contained the majority of the parts necessary to control the mechanism. The baseplate was reused, allowing for simple and repeatable attachment to the dVRK, and the motor shafts and plate bearings were also reused to better control the tensioned cables.

In order to control the bend of the modular shaft design, the original transmission system included a set of pulleys that held the ends of the tensioned cables. These pulleys were mounted through the original motor shafts and would clamp to the shafts using set screws. The existing transmission pulleys were able to pull the tensioned cables for the necessary motions, but the cable clamps were too small of a diameter to fit the intended cables. These existing pulleys could not be reused since the process of increasing the cable clamp diameter would have destroyed the part. Therefore, the team created new torsion pulleys by measuring existing pulleys and

modeling it in SolidWorks (see Figure 12). The team started by manually machining the round stepped shape from a ½ inch aluminum rod, but encountered problems with precisely replicating the parts. As a result, the team created a program on the CNC lathe to ensure all parts are of a uniform size. In addition, the cable clamping holes were enlarged and moved further along the radius of the motor shaft. Due to the small diameter of the cable holes, the part had to be machined using a Haas CNC Super MiniMill to attain the 15,000 RPM necessary to cut the part. The pulleys were later modified using a manual mill to accept set screws, allowing them to clamp on to the existing motor shafts.




Figure 12: *Isometric Solidworks model of the torsional motor pulley*

Since the team did not have the expertise with and limited availability with the da Vinci surgical robot, it was difficult to test the modular system in a controlled test environment. Additionally, the team wanted to demonstrate the kinematic behavior of the mechanism in a standalone testing scenario. This pushed the team to create a test fixture, shown in Figure 13, that would emulate the motion and torque of the standard dVRK system. The fixture was based around a dVRK test plate that served as the integration base. The integration base was actuated using three stepper motors that were controlled by an Arduino Uno and a CNC adapter shield. This allowed the team to attach the modified transmission and move each pulley individually through the Arduino program shown in Appendix B.




Figure 13: *Test Fixture for the dVRK*

3.2 Software and Electronics Logistics

3.2.1 Design Objectives and Functional Requirements

The goal of the software design is to format and display the thermal data output by the camera. The output display modality should allow for easy visual interpretation of heat within the image. The design of the software system should allow for the connection of multiple clients for use with multiple output displays. Similarly to the mechanical requirements, the team generated functional requirements for the software system design. The results of the software design and an analysis of the software design requirements will be discussed in the Results section.

Functional Design Requirements:

1. Cables external to the endoscope package must not interfere with normal robot operations, and therefore the count must be minimized.
2. Additional clients must be easy to configure, to allow multiple screens to read and interpret the data with different modalities in the future, depending on the operation.
3. Thermal data should be displayed with a colored image.

4. The system should have latency lower than 75ms because otherwise, the delay becomes noticeable enough to a person to cause perceived performance decrease, and therefore potentially impacting the ability of the surgeon to precisely control the device.[23]
5. Software will be developed using C++ and/or MATLAB.
6. The surgeon should be able to set a temperature threshold for warning.
7. The surgeon should get a warning once the threshold is reached.

3.2.2 System Overview




Figure 14: *System Architecture Diagram*

ROS

ROS, or Robot Operating System, is a software system that is commonly used in robotics applications, especially in prototype robotics. The dVRK uses ROS to communicate with its network host computer, and the tools are controlled by sending target poses to the da Vinci action server. To allow for compatibility with the dVRK, the team chose to design the software in ROS. ROS contains a networked message handling system, which is handled by the ROS

master core. The message handling system allows for a publisher-subscriber model which handles network transport protocol for the user across both C++ and Python systems. The ROS network runs individual programs in nodes, which allow for distributed computing in robotic systems. The node system allows for easy reconfiguration based on node selection. The node system also increases the ease of prototyping, as new nodes can be added to the network to test new functionality. In the case of a surgical endoscope, ROS allows for easy reconfiguration in the number of end users and of display modalities, which is especially useful for prototyping systems. Different surgeons may prefer different display modalities, or different staff and different operating theatres might allow for different setups for the endoscope display. By using ROS, we are able to keep an expandable system built around utilizing the endoscope to the fullest.

Catkin

Catkin is the custom build system used in recent versions of ROS. It is developed for, but also independently of ROS. Catkin facilitates easy management of packages and dependencies. Catkin is responsible for linking the ROS and the GroupGets Lepton Module thermal library [24]. Catkin is very light in dependencies, and only requires Python and CMake to run. Programs that are built for the Raspberry Pi in this project are built using the Catkin build system.

MATLAB

MATLAB is a software package that is used in engineering and scientific applications for analysis and design. MATLAB contains powerful tools for graphing and data analysis of matrix data and can generate color-mapped images with value labeled color bars, which is why it was picked for use in this project. MATLAB also has a built-in ROS toolbox which allows for data to be captured in real-time from a running ROS network in a MATLAB subscriber node. One major benefit of MATLAB is that it is designed to work with engineering equations like kinematics. In the future, given the exact kinematics of the endoscope mechanism and the thermal data stream, MATLAB can generate custom target pose messages to be published to the dVRK over the ROS network. The built-in mathematics functions of MATLAB are well-suited to manipulating

equations for prototyping and control, and the graphing functions of MATLAB allow for easily customizable data analysis.

Custom PCB

A custom Lepton interface board was designed on-campus by Keion Bisland for use in this project. The PCB breakout allows the camera to connect to a SPI and I²C host and also provides power regulation and a clock signal. Although FLIR produces a proprietary PCB for the camera, the package size of this PCB is significantly greater than is necessary by design. A custom designed, stacked multi-layered board allows for more compact package size. If the PCB is re-designed into a multi-layered circuit board, the physical size of the board is constrained almost entirely by the external width of the camera socket.

The primary objective in the design of the new PCB was to shrink the footprint as much as possible. Along with this goal, a number of considerations had to be made; these include the lack of documentation on the required supporting components for this development board, size constraints of the components, reduction of noise induces on the communication lines and the production process and selection of components to ensure the final product was free of lead and any other hazardous chemicals.

After designing the schematic for the board the footprints were assigned and the board outline made. It was then decided to split the components between two circuit boards that would be split apart after assembly. The upper of these two boards would house the camera module holder, the cameras associated 25MHz crystal oscillator and 1.27mm pin headers to connect to the second board. The lower of these two boards would contain the power regulation and supporting components as well as the communication and power connections to connect to the Raspberry Pi. This two-board system allowed us to reduce the footprint of the overall development board to the dimensions of the camera socket, being 11.5mm x 11.5mm, the smallest it could be made continuing to use the FLIR Lepton Camera.

Raspberry Pi

To reduce the interference that would be caused by an extremely long serial connection, and to increase the range of the thermal camera signal beyond the maximum length of a serial

data line, a Raspberry Pi 3B is used as a host for the Lepton camera to transfer raw data to the ROS network. In the current version of the prototype, the breakout board wires connect directly to the GPIO pins of the Raspberry Pi, which allows the ROS thermal data publisher to connect to the Lepton. The Raspberry Pi then publishes data to the ROS network over the Ethernet connection.

Networking Devices

The router is simply used as a DHCP host. Using a DHCP-equipped router was the easiest way to allow the Raspberry Pi and the host computer to switch between the private network and the internet for package installs and updates, along with connecting to version control software. If the capability was required, the router would also let the endoscope to connect to wireless devices.

The switch chosen for use in this project is a TP-Link 8-Port Fast Ethernet switch. This switch was chosen for its Power over Ethernet feature. Power Over Ethernet facilitates all communications between the host computer and the Pi with less danger of wire tangling by reducing the number of wires, and by increasing the maximum distance between the endoscope and the host computer significantly. Ethernet wire is also cheap and easy to crimp, which is a significant benefit for a research project. Another benefit of a single Ethernet wire is that Ethernet is thickly jacketed, which helps provide some limited protection against external interference or any possible mechanical pinching, especially when compared to the small, easy to tear serial wires used to connect to the breakout board.

Da Vinci Host Computer

The way the entire project was set up allows the use of the host computer of the da Vinci robot to connect to the Pi directly over ROS. The host computer of the dVRK runs the ROS master host, which coordinates ROS messages for the da Vinci. While this iteration of the project did not directly make use of dVRK host computer, it allows for easy expansion into the robot host computer in the future by simply switching startup configuration by running the publisher and subscriber nodes in the native da Vinci network. The da Vinci Host computer would run the only master node in the system, and any nodes set up control the da Vinci can be run on the host

computer, allowing the display host computer to fully focus its resources on rendering the incoming data quickly.

3.2.3 Program Design

C++ ROS Publisher

The C++ ROS Publisher is responsible for publishing thermal data to the ROS network. The camera is connected to the Raspberry Pi using the custom PCB over an SPI and I²C. Upon a successful connection, the camera will send thermal data packets to the Raspberry Pi over SPI. The thermal data packets are then assembled into complete image frames. A single image frame is a fixed-width integer image vector representing the temperature in Kelvin. The temperature data is initially represented in a scaled integer format. To clarify, the output data is scaled by 100 and truncated at the decimal point, therefore, the lowest two integer digits of the temperature represents the first two decimal digits of the temperature in Kelvin. This integer scaling is for image transport only, the thermal data will be converted to floating point and scaled using the greater resources of the display desktop. The image vector is published to the ROS network using a single node, which repeatedly publishes to the /image topic as quickly as images are received from the Lepton. A flowchart depicting the functional flow of the ROS publisher program is included in Appendix C.

MATLAB ROS Subscriber

The end user receives thermal image packets from the network, does any transformations necessary to the picture for the desired output modality, and outputs the transformed data for display. To reduce the amount of work that runs on the relatively lower processor power Raspberry Pi, most of the graphics and matrix computations are left to the implementation and processing of the desktop end user. For the purposes of this project, a MATLAB image vector subscriber was determined to be sufficient for prototyping data coloring and analysis. When the thermal data is received, the scaled image is first converted to a floating point representation which accurately represents the temperature in Kelvin. Elementwise matrix operations are then applied to efficiently convert the entire image from Kelvin to degrees Celsius. Once the Celsius

image data is formatted correctly, the data can be transformed into an appropriate color scale for display on the UI using MATLAB's built-in colormap selection. A flowchart of the MATLAB subscriber program is included in Appendix D.

Output Modalities

The built-in graphing features of MATLAB has a dynamic color map with many preset options. Depending on the colormap selected, the range of temperature values in the thermal image are mapped to a colored gradient. The purpose of color mapping is to allow for easy visual interpretation of sensor data. In our project, the main goal of the display modality is for the surgeon to be able to recognize the temperatures throughout the thermal image without spending time understanding the color scale, while still highlighting variation throughout the data. Along with a colormap, a color bar with labeled ticks is generated as a key to the mapping. The hottest and coldest temperatures in the frame determine the range of the color bar and of the range of the color gradient. As the colorbar and colormap are updated dynamically with each new frame, the colors adjust based on the temperatures in the frame. Creating an application-specific custom static colormap was outside the scope of this project, but developing a custom static colormap which would allow for a high-contrast highlighting of dangerous temperatures. Another consideration is that the temperature range in basic demonstration images or in validation images is very different from the temperature range in a real application of the medical endoscope, and so this should taken into account when selecting a display modality.

Logarithmic scaling is a secondary method of displaying the thermal image data. Logarithmic scaling would have allowed a wider range for colder temperatures while compressing the higher temperatures together. The benefit of displaying this way is that even if a specific area in the frame of the image remained very hot, such as while the cautery tool is being tracked, the rest of the frame would not change significantly in dynamic scaling. As a result, variations in temperature towards the lower range of temperatures retain a higher amount of detail. The drawback is that the area with the highest temperature would not have as wide of a range to display the finer variance in recorded temperature. Whether logarithmic scaling is useful

is very dependent on the range of temperatures in the image, and thus it would be significantly more useful over a static, given range.

Gray and Log Gray




Figure 15: *Example Colormap, “Gray”, with linear (left) and logarithmic (right) scaling*

A monochromatic image such as the one provided by the *Gray* colormap accurately represents the linear dimension of the thermal data, so it is in a manner the most direct mapping of value to color considered. Black and white provide strong visual contrast for understanding the data being displayed, which results in a good amount of detail given the single hue gradient. A similar display modality would be a good candidate for modification in the future. For instance, modality could use a few highly contrasting colors to identify near-dangerous, dangerous, and damaging temperatures, so that information could be interpreted at first glance.

Hot and Log Hot




Figure 16: *Example Colormap, “Hot”, with linear (left) and logarithmic (right) scaling*

The range of colors used in *Hot* is both easy to interpret immediately and provides excellent contrast between the lowest colors in the frame with the hottest (shown in Figure 16). If the camera was being used inside of a patient, the majority of the image would be in the black to red range, and the hottest part of the image, the heated tool, would transition between orange to white. The low number of hues in the gradient results in a loss of detail over wide temperature ranges. With a small temperature range in the image, or by cropping to the minimum color above the majority of the image, this difficulty could be resolved.

HSV and Log HSV




Figure 17: *Example Colormap, “HSV”, with linear (left) and logarithmic (right) scaling*

The *HSV* colormap goes between bright blue at the mid-range to dark blue, then bright again pink at the higher end of the scale range, which is too visually jarring and chaotic for easy interpretation (shown in Figure 17). Due to the bright transitions and high amount of color variation, *HSV* is well suited for showing small differences in very low contrast images. During a robotic operation, the heated tip and the flesh temperature would both be in the frame, which is a large difference in temperatures. *HSV* works well for low-contrast images, or for depicting a small range in temperatures.

A significant downside of HSV is that the scale wraps around at the color red, which creates confusion between the highest and lowest values. Another major issue with *HSV* is that red is normally thought of as a hot color, and is typically grouped with yellow and orange, which are at the bottom of the scale. The hue transitions between blocks of temperature are very bright and are harder to interpret at a first glance, which is why *HSV* was not used for demonstration purposes. HSV is not recommended in the future while using a dynamic color bar, except for the specific applications where the majority of the image would be in only a small fraction of the temperature range, which should never be the case in robotic surgery, or in a low-contrast mode display mode.

Jet and Log Jet




Figure 18: Example Colormap, “Jet”, with linear (left) and logarithmic (right) scaling

Jet is the colormap chosen for our demonstration and poster images because it provides an easy to interpret the image for the range of values used (shown in Figure 18). The rainbow is a natural color spectrum that people are instantly familiar with, and thus is a powerful aid in reading thermal images coded using this colormap. The changing hues contrast well against the background color and show details without oversaturating. The range of colors in the gradient transition smoothly, but also in a way that allows for immediate interpretation, unlike *HSV*, and the wide range of colors in the gradient allows for a more granular visual resolution than simpler gradients like *Gray* or *Hot*, which is why *Jet* is the default modality used to generate thermal images in our project.

4. Results

4.1 Mechanical Results

4.1.1 Analysis of Mechanical Design and Functional Requirements

The final mechanical design, shown in Figure 19, was successfully able to meet the majority of the functional requirements stated in section 3.1.1 and performed well while on the test fixture. These functional requirements were created to set a baseline for the design and operation of the mechanism, and based on the results the mechanism shows potential, but there are still areas where the mechanism can be improved.




Figure 19: *Final Mechanical Build*

The first functional requirement stated that the overall size of the endoscope should be similar to existing da Vinci instruments. While the camera mechanism is limited by the 9mm by 9mm camera, the addition of the stacked PCB and the camera enclosure increased the height and

width of the mechanism to 14mm by 14mm. Typical da Vinci mechanisms have a diameter of 8mm, but by modifying the PCB and changing the camera enclosure material from SLA to stainless steel or aluminum, the enclosure can be smaller and more compact. As such, the team believes that this system can be made smaller in the future to become more viable in the laparoscopic environment.

The material of the mechanism is quite important, as the mechanism is intended to be inserted within a patient. As such, the material used must be rigid, non-deformable, and biocompatible. For this system, the camera enclosure, modular link, and shaft were all made out of SLA resin. This resin was not biocompatible and was mostly rigid, making it not very usable as a final product. This material was initially selected because the team needed to make quick iterations of the mechanism design and the design had small, detailed features that could only be printed using a high-resolution printer. The only printer available and suited for such purpose was the Formlabs Form 2 printer. The company offers very limited biocompatible resin at a high cost, as such the team decided that using regular rigid material will be sufficient for prototyping knowing that the final mechanism would need a more applicable material. The SLA parts for this system had some trouble holding the intended shape during the curing process, resulting in bent and skewed shafts. The shafts were usable for demonstration and prototype purposes, but require a stronger material, such as stainless steel to reinforce. For example, the hinge pins and tensioning cable were made from stainless steel to provide additional structural support.

The final mechanism was successfully able to reach all desired orientations using a two-degree of freedom design shown in Figure 20. The tensioned stainless steel cables can bend the modular shaft by 60° from the central axis and can twist the shaft by 360° in both directions. The combination of the bendable link and a rotating shaft allows the camera to observe the space in a 360° cone from the base of the modular link. With the 60° bend and a field of view of 57° , the camera is able to view up to 88.5° from the central axis, making it more than sufficient to view the proposed surgical area.




Figure 20: *Modular Link Bend Radius*

In terms of mechanical repeatability, the mechanism was able to fully cycle more than 30 times without mechanical failure. However, the team was not able to generate a full fatigue analysis for the proposed link and camera module. Existing da Vinci instruments are typically rated for 10-30 uses, where in each operating the tools are articulated numerous times. The goal of the mechanism is to match this operating condition with no breaks or fractures, however the team was unable to verify this claim due to unexpected fractures. These fractures occurred when extreme external forces were applied to the links while the mechanism is connected, where the spike of shear stress caused the hinge connections to fail. The robustness and repeatability of the mechanism can be improved by switching from the SLA resin to a stronger material as mentioned before. In particular, stainless steel, aluminum, or injection-molded plastic would be ideal, but the SLA was suitable for the initial prototyping.

Due to the nature of minimally invasive surgery, it is important to ensure that the cables and wires are contained within the mechanism at all times during the surgery. The hollow shaft and partitioned modular links allow the camera wires and tensioned cables to pass through the mechanisms to the camera enclosure without outside exposure. To fully enclose this mechanism, however, it would be necessary to include a flexible, biocompatible sheath around the area of bending. The angled cuts of the mechanism could also potentially pinch the patient during the operation, and the interior would be open to contamination. In the past, low-density polyethylene

has been used in medical applications, and the ductility of the plastic could allow this material to be used as a sheath.

By using an existing da Vinci instrument baseplate as the basis for the transmission, the mechanism was able to attach easily to the dVRK system, as shown in Figure 21. Additionally, since the test fixture has the same interfacing geometry as the dVRK, the mechanism can be attached to the test fixture and manipulated using Arduino inputs. This saved the team time in developing a testable solution and ensured that the mechanism would perform as expected when attached to the dVRK.




Figure 21: *Thermal Infrared Endoscope attached to the dVRK surgical robot*

4.1.2 Kinematics




Figure 22: *Modular Link Bend Radius*

To determine the relationship between the stepper motor steps and the bend angle, the change in cable length was correlated with the resulting bend. In order to bend the mechanism, the cable length is shortened by reeling in the tension pulley. This reduces the link separation (denoted as L in Figure 22), and this change is directly correlated with the change in angle. Note that with multiple links, the change in cable length is equal to the total L . To determine how many stepper motor steps are necessary to produce a certain bending angle, the change in cable length can be divided by the pulley circumference and converted to the angle of rotation at the pulley. With the degrees per step known for the particular stepper motor (1.8° per step), the resulting pulley rotation can be divided by the degrees per step to obtain the total number of steps needed to achieve the desired bend angle. This process is shown below:

$$L = 3.16mm, \Delta L = 6.32mm$$

$$r = 2.36mm \text{ (pulley radius)} \rightarrow C \text{ (circumference)} = 2\pi r = 14.83mm$$

Steps for 60° bend : $(L/C)/(1.8^\circ/\text{step}) = 85.24 \approx 85 \text{ steps}$

Steps per degree : $85 \text{ steps} / 60^\circ = 1.42 \text{ steps/deg}$

4. 2 Software Results

4.2.1 Analysis of Software Design and Functional Requirements

The use of cables to one coming out of the endoscope was minimized by making use of Power Over Ethernet. By utilizing Power over Ethernet, a single cable to the unit allows for both power and data to be transmitted. Ethernet is a low-cost wire which can be manually crimped at the desired length, which allows for the system to be easily equipped with a proper length cable. Without Power Over Ethernet, the Raspberry Pi requires a micro-USB power cable separate from the Ethernet cable used to connect to the ROS network. The issue with using a micro-USB power cable to power the Raspberry Pi is that the maximum length of USB cables is strictly limited by the USB standard, which poses a tangle risk with objects or people in the vicinity of the robot, or with other arms of the robot itself. Although the PoE solution was successful, the electronics of the endoscope must be consolidated in order to provide a fully safe solution.

ROS allows for easy configuration of multiple clients by utilizing the centralized ROS core combined with the publisher and subscriber model. The Raspberry Pi publishes the thermal image data over the ROS network to the /image topic. From there, other ROS clients can become subscribers to the /image topic and receive those images without impacting other subscribers by connecting to the ROS network using the router.




Figure 23: *Boiling Water Validation*

MATLAB has built-in tools for the automatic coloring of generic matrix data utilizing colormaps. Colormaps allow for easy interpretation of temperature data. Switching the colormap of a MATLAB image only requires changing a single parameter in the code. The built-in color maps provide different display modalities which are contextually useful. For instance, a colormap such as “jet” or “hot” allows for quick recognition of peak temperature by highlighting the hottest end of the spectrum with a bright color. The built-in colormap can be modified further to juxtapose an extra bright contrast color for specific damaging temperatures by creating a custom colormap. A color map such as “hsv” allows for a clearer interpretation of thermal gradients at all temperatures at the cost of a more chaotic image.

We were able to reach a publishing rate of 8 Hz over the fully integrated network. This does not include the screen update rate which is variable depending on the specific screen used. There were significant difficulties due to limited resources of the Raspberry Pi and with the optimization of MATLAB updates. The Raspberry Pi still seemed to be inefficient for running ROS on top of headless Raspbian. Building using the ROS library initially took multiple hours due to the strict memory and processor limitations of the Raspberry Pi. Switching to a faster

board such as a Beaglebone Black would allow for quicker rates of processing and publishing which would reduce the latency. The system latency could be further reduced on the side of the end user by using a faster graphics processor, by further optimizing the MATLAB code, and by solving the driver issues with hardware rendering. Another potential solution is to replace the middle computer with a proprietary board specialized for the Lepton such as the PureThermal 2. The PureThermal 2 would be able to connect the camera to a host over USB in place of the Raspberry Pi, and the host would act as the publisher server.

All portions of code on the Raspberry Pi are coded in C++ using ROS Kinetic. The code running on the Raspberry Pi is the most performance critical, since the Raspberry Pi is the single source of the image from the thermal camera, yet has little resources to spare. By using C++, we can compile the code to native bytecode. While Python may have been easier for programming on the Pi, the Pi did not have resources that could be spared on the interpretation of the code.

The image processing was written in MATLAB. The MATLAB code exclusively runs on the desktop running Ubuntu. While programming in MATLAB was successful, using Matplotlib on Python may have performed better on the application of live-plotting multiple data views. Using Python would allow for more expandability as well as portability since any Python library could be used.

Highest temperature in frame is tracked and output to console, MATLAB has a command for creating tones, but our demonstration would have been either been unable to produce an object hot enough to trigger the sensor, or the sensor would have been going off constantly because we would have been unable to heat things in our demonstration space, so the functionality was not tested fully. However, implementation given the hottest temperature in the frame can easily be modified with a boolean comparison with the set threshold, which would trigger MATLAB to generate the tone.

4.2.2 Endoscope Output Validation Testing

In order to validate the thermal data outputs from the software system, the team verified the thermal data from the endoscope camera and publishing as a ROS topic are actual temperature readings. The accuracy of the FLIR Lepton camera was expected to be within ± 5 °C

[13]. The thermal outputs displayed in MATLAB is thus expected to have that accuracy. To validate the outputs, the team used a calibrated thermal camera, FLIR A655sc, as a reference camera. The FLIR A655sc has up to 50 mK precision, $\pm 2^{\circ}\text{C}$ or $\pm 2\%$ of accuracy, with a resolution of 640 x 480 [20], which enables us to assess the accuracy and precision of the endoscope software outputs by comparison.

The team positioned the endoscope camera and the A655sc side by side facing the test object in order to test for any significant aberrations in temperature measurement or in calibration. If the camera been misconfigured, calibrated incorrectly, or damaged in some way, then the Lepton output data would likely vary significantly from the measurements of the accurate camera. Similarly, any faults in the software or electrical systems would result in a significant error in the final image output.

In order to clearly compare test images between thermal cameras, a metal test object was chosen. A metal test object equalizes heat more quickly than other materials, resulting in more a uniform surface temperature. Emissivity ratio is important when making quantitative temperature measurements using a thermal imaging camera [21]. Due to thermal emissivity, a thermal imaging camera may give different readings given the same objects at the same temperature, which can cause problems in metallic test objects. To reduce the effect of thermal emissivity on accurate camera measurements, an aluminum cube with black electrical tape covering the surface of the object body was used. The emissivity of electrical tape is approximately 0.95, which should reduce the effect of emissivity on the camera output [22]. The error between the camera and reference camera output can give us a sense of the accuracy of the integrated camera system.




Figure 25: *Camera Validation Setup*

The independent control variable in this validation is the actual temperature of the object at a given time. The team varies the temperature of the object by using a heating unit. Heat is applied until the test object reaches 65 °C, at which point the heating base is disabled. The object then slowly cools to room temperature. While the test object cools, multiple images are taken from both cameras simultaneously for later analysis. The relatively high temperature of the heating unit resulted in a significant difference in temperatures between the top and the bottom of the test object. A strong thermal gradient is less preferable for validation as it is easier to visually and analytically compare between images in regions that vary less. To mitigate this issue, images are only taken after the heating unit has been disabled, and the test object has cooled back into the testing temperature range. The high thermal conductivity of the metal test object results in a relatively even heating across the object by the time it had cooled to the testing range.

Once the test object reached the testing temperature range, the two camera outputs were saved and exported as MATLAB matrix files. The images were then analyzed in MATLAB. In order to ensure that the output images were consistent across cameras, both sets of validation images were generated using the same code used to display thermal images for the camera. The team manually selected the region of interests (ROI) for each camera in MATLAB, which in this test was the surface of the cube which was oriented to face both cameras. The two cameras were positioned to face the nearest surface of the aluminum piece as closely as possible, but as a result of physical constraints, each camera has a slightly different angle of vision. Figure 26 shows the FLIR A655sc camera output displayed in MATLAB. The labeled color bar on the side displays the mapping of color to temperature in degrees Celsius. The indicated area shows the ROI that was selected, and the temperature is the average of that region. The team's camera system measured 55°C for the front face of the test object.




Figure 26: *FLIR camera reading on object temperature at 55°C*

Figure 27 shows the temperature measured by the higher resolution A655sc to be at 50°C. From both Figure 26 & 27, the area with the highest temperature is the heating base, which provides the heat temperature to control the temperature of the tested object.




Figure 27: *Reference camera reading on object temperature at 55°C*

The team measured 4 sets of data with both camera outputs saved simultaneously. Table 2 shows the data set and the error between the two readings captured simultaneously. We assume the actual temperature of the object surface is the same when two readings are taken. The validation measures the object temperature at around 45 °C, 50 °C, 55 °C and 65 °C with an average error of approximately 3.63 °C.

Image Set	FLIR 3.5	A655sc	Error
1	41.4828	44.6817	3.1989
2	47.034	50.3004	3.2664
3	51.4598	54.9415	3.4817
4	60.0038	64.5961	4.5923

Table 2: *Camera Validation Test Data in Celsius Degrees*

The validation indicates that the endoscope camera outputs are within the expected range of error. This result gives us confidence that the readings that would be provided to the surgeon are fairly accurate in the necessary range of 45 °C to 65 °C, and that our integrated camera system is functioning properly.

5. Conclusion and Recommendations

Technology is always improving, and future infrared cameras may allow this endoscope to become as small as regular endoscopes. Future projects, however, can improve on the designs created in this project with the current technology. Integration with the technology on the da Vinci did not happen this project and would be a good place to continue. One way is to augment the regular, optical endoscope on the da Vinci robot is to improve the camera PCB design to fit the endoscope better. Additionally, as mentioned in the functional requirements section, the camera enclosure and shaft features could be made from a stronger material than SLA resin, such as aluminum or nitinol. These additions would serve as initial improvements to significantly increase the viability and integrity of the mechanism.

To improve the usability of the endoscope, there are multiple software improvements that can be done. In particular, it would be helpful for the surgeon to see a side-by-side of the optical camera view and the infrared view during the operation procedure. This would provide significantly more information about the thermal impact of an energy-based instrument, while not obstructing the optical view. This could be tested by performing a benchtop experiment of a cauterizer on a tissue sample. This would allow the user to view the thermal changes within the tissue and determine how much value is added with the thermal endoscope.

Using the index of the maximum temperature spot and a MATLAB ROS node subscribed to the da Vinci encoder position, inverse kinematics can be used to orient the endoscope toward the cautery tool in order to autonomously monitor thermal heat spreading. Given the kinematics of the endoscope and the thermal data, MATLAB can generate target pose messages to be published to the da Vinci action server. MATLAB has built-in tools for control analysis which could be used to analyze the endoscope response and to generate a reliable controller for the endoscope tool. Another compelling potential application of the endoscope in robotic surgery is for autonomous control of cautery tool heat. A MATLAB thermal node could monitor the thermal data stream and could automatically turn off the cautery tool when dangerous conditions are detected, reducing the risk for thermal damage. If a simple temperature threshold is not

sufficient for determining dangerous conditions, machine learning methods could be applied to approximate the risk of thermal damage.

In the future, a ROS Python node using Matplotlib and NumPy could replace the MATLAB subscriber and allow for potentially faster image-plotting of real-time thermal data. Python can process animated subplots of various statistics along with thermal data using the already implemented publisher node. The Python approach would allow for greater customization in the final output and might perform better at multi-plot rendering. Python also supports hardware accelerated vectorization using the NumPy package for use in performing matrix operations, although MATLAB natively handles math.

In conclusion, the design of a thermal infrared endoscope has the potential to reduce the amount of heat-related injuries during robotic surgery. While there is much that can be improved upon, this project has served as a first step to developing a robust and viable solution. With a thermal endoscope alongside the optical camera, robotic surgeries can be performed safer and more effectively.

Bibliography

- [1] Intuitive Surgical - 2017 Financial Report. (2018, February). Retrieved from <https://www.sec.gov/Archives/edgar/data/1035267/000103526718000013/isrg-20171231x10k.htm#sF8F12A5E7D3D558687633AE22CF94405>
- [2] Kirkpatrick, T., & LaGrange, C. (2016, February). Robotic Surgery: Risks vs. Rewards. Retrieved from <https://psnet.ahrq.gov/webmm/case/368/robotic-surgery-risks-vs-rewards->
- [3] R. Sangha, D. I. Eisenstein, A. George, A. Munkarah, and G. Wegienka, “Surgical outcomes for robotic-assisted laparoscopic myomectomy compared to abdominal myomectomy,” *Journal of Robotic Surgery*, vol. 4, no. 4, pp. 229–233, 2010.
- [4] B. S. Peters, P. R. Armijo, C. Krause, S. A. Choudhury, and D. Oleynikov, “Review of emerging surgical robotic technology,” *Surgical Endoscopy*, vol. 32, no. 4, pp. 1636–1655, 2018.
- [5] Penson, D. F., McLerran, D., Feng, Z., Li, L., Albertsen, P. C., Gilliland, F. D., Hamilton, A., Hoffman, R. M., Stephenson, R. A., Potosky, A. L., et al., “5-year urinary and sexual outcomes after radical prostatectomy: results from the prostate cancer outcomes study,” *The Journal of Urology* 173(5), 1701–1705 (2005).
- [6] Niemz, M. H., [Laser-Tissue Interactions: Fundamentals and Applications], Springer Science & Business Media (2007).
- [7] Lantis, J. C., Durville, F. M., Connolly, R., and Schwaitzberg, S. D., “Comparison of coagulation modalities in surgery,” *Journal of Laparoendoscopic & Advanced Surgical Techniques* 8(6), 381–394 (1998).
- [8] Hefermehl, L. J., Largo, R. A., Hermanns, T., Poyet, C., Sulser, T., and Eberli, D., “Lateral temperature spread of monopolar, bipolar and ultrasonic instruments for robot-assisted laparoscopic surgery,” *British Journal of Urology International* 114(2), 245–252 (2014).
- [9] Ahlering, T. E., Skarecky, D., and Borin, J., “Impact of cautery versus cautery-free preservation of neurovascular bundles on early return of potency,” *Journal of Endourology* 20(8), 586–589 (2006).

- [10] Wondergem, J., Haveman, J., Rusman, V., Sminia, P., and Van Dijk, J., “Effects of local hyperthermia on the motor function of the rat sciatic nerve,” *International Journal of Radiation Biology* 53(3), 429–438 (1988).
- [11] Fichera, L., [Cognitive Supervision for Robot-Assisted Minimally Invasive Laser Surgery], Springer Publishing Company, Incorporated (2016).
- [12] P. Francis, K. W. Eastwood, V. Bodani, T. Looi, and J. M. Drake, “Design, Modelling and Teleoperation of a 2 mm Diameter Compliant Instrument for the da Vinci Platform,” *Annals of Biomedical Engineering*, vol. 46, no. 10, pp. 1437–1449, 2018.
- [13] FLIR, “HIGH RESOLUTION MICRO THERMAL CAMERA FLIR LEPTON 3 & 3.5” datasheet
- [14]“da Vinci Research Kit wiki community”,
https://research.intusurg.com/index.php/Main_Page
- [15] P. Francis et al., "Miniaturized Instruments for the da Vinci Research Kit: Design and Implementation of Custom Continuum Tools," in *IEEE Robotics & Automation Magazine*, vol. 24, no. 2, pp. 24-33, June 2017. doi: 10.1109/MRA.2017.2680547
- [16] York, P. A., Swaney, P. J., Gilbert, H. B., & Webster, R. J. (2015). A Wrist for Needle-Sized Surgical Robots. *IEEE International Conference on Robotics and Automation : ICRA : [proceedings]*. *IEEE International Conference on Robotics and Automation*, 2015, 1776-1781.
- [17] K. W. Eastwood, H. Azimian, B. Carrillo, T. Looi, H. E. Naguib and J. M. Drake, "Kinetostatic design of asymmetric notch joints for surgical robots," 2016 IEEE/RSJ International Conference on Intelligent Robots and Systems (IROS), Daejeon, 2016, pp. 2381-2387. doi: 10.1109/IROS.2016.7759371
- [18]P. Francis, K. W. Eastwood, V. Bodani, T. Looi, and J. M. Drake, “Design, Modelling and Teleoperation of a 2 mm Diameter Compliant Instrument for the da Vinci Platform,” *Annals of Biomedical Engineering*, vol. 46, no. 10, pp. 1437–1449, 2018.
- [19] TaoMing Liu, “Design and Prototyping of a Three Degrees of Freedom Robotic Wrist Mechanism for a Robotic Surgery System”, Case Western Reserve University, Jan, 2011

- [20] FLIR. (2014). FLIR A655sc [Brochure]. Retrieved from www.flir.com
- [21] C. Mcdaid and Y. Zhang, “Wall temperature measurements using a thermal imaging camera with temperature-dependent emissivity corrections,” *Measurement Science and Technology*, vol. 22, no. 12, p. 125503, 2011.
- [22] FLUKE, “Emissivity values of common materials”, 3038318 H-EN-N Rev A, 9/2007
- [23] R. Dabrowski, C. Manuel, R. Smieja, “ The Effects of Latency on Player Performance and Experience in a Cloud Gaming System”, May 7, 2014
- [24] GroupGets “LeptonModule”, <https://github.com/groupgets/LeptonModule>

Appendix A: Notched-Tube Model in MATLAB

```
% notches-tendon endscope design

syms a b c g h ri ro ybar phio phii Ao Ai ybari ybaro k s dl theta thetaMAX rhoMin
S

a = 10; %length before joint      ALL LENGTH UNITS IN MM
b = 10; %length after joint
c = 6; %spacing inbetween cut
g = 8.5; %cut depth
h = 2; %cut width
ri = 4.25; %inner radius
ro = 4.75; %outter radius
n = 10; %number of notches
% dl = 1; %distance pulled
E = 205000; %young's modulus in N/mm^2

% ybar calculation: neutral axis location
phio = 2*acos((g-ro)/ro);
phii = 2*acos((g-ro)/ri);
Ao = (ro.^2*(phio-sin(phio)))/2;
Ai = (ri.^2*(phii-sin(phii)))/2;
ybari = (4*ro*sin(phio/2).^3)/(3*(phio-sin(phio)));
ybaro = (4*ri*sin(phii/2).^3)/(3*(phii-sin(phii)));
ybar = (ybaro*Ao-ybari*Ai)/(Ao-Ai);

% k calculation: cuvature of a single notch (turn radius = 1/k)
% k = dl/(h*(ri+ybar)-dl*ybar);
k = 1/100;

% s calculation: arc length
s = h/(1+ybar*k);

theta = n*k*s; % rotation angle of each notch
% Transformation matrix of a single bending notch
T = [1 0 0 0;
     0 cos(theta) -sin(theta) (cos(theta)-1)/k;
     0 sin(theta) cos(theta) sin(theta)/k;
     0 0 0 1];

% % dl calculation (from k&s): tendon displacement
% dl = h-2*(1/k - ri)*sin(k*h/2*(1+ybar*k));

% thetaMAX: max bending angle
```



```
thetaMAX = n*h/(ro+ybar);

% rhoMIN: minimum radius of curvature
rhoMIN = ro+(n-1)*c/thetaMAX;

% S: approximate circular arc length
S = n*(ro*h/(ro+ybar)+c)-c;

% max strain at outer wall
esp = k*(ro-ybar)/(1+ybar*k);

% max stress
sigma = E * esp;
sGPA = sigma/1000; %max stress in GPA
```

Appendix B: Test Fixture Arduino Code

```
//Global Variables
int xStep = 2;
int xDir = 5;
int xLim = 9;
int yStep = 3;
int yDir = 6;
int yLim = 10;
int zStep = 4;
int zDir = 7;
int zLim = 11;

void setup() {
  // put your setup code here, to run once:
  //Serial.begin(9600);
  pinMode(xStep, OUTPUT);
  pinMode(xDir, OUTPUT);
  pinMode(xLim, INPUT);
  pinMode(yStep, OUTPUT);
  pinMode(yDir, OUTPUT);
  pinMode(yLim, INPUT);
  pinMode(zStep, OUTPUT);
  pinMode(zDir, OUTPUT);
  pinMode(zLim, INPUT);
  digitalWrite(xDir, HIGH);
  digitalWrite(yDir, HIGH);
  digitalWrite(zDir, HIGH);
  delay(4000);
}

void loop()
{
  //SIDE TO SIDE
  //Goes UP
  digitalWrite(xDir, HIGH);
  digitalWrite(yDir, HIGH);
  digitalWrite(zDir, HIGH);
  for (int i = 0; i < 40; i++)
  {
    xStepMotor();
    zStepMotor();
  }
  delay(1000);

  //Goes DOWN
  digitalWrite(xDir, LOW);
  digitalWrite(yDir, LOW);
  digitalWrite(zDir, LOW);
}
```

```

for (int i = 0; i < 80; i++)
{
    xStepMotor();
    zStepMotor();
}
delay(1000);

//Rotates Shaft
rotate();

//Goes UP
digitalWrite(xDir, HIGH);
digitalWrite(yDir, HIGH);
digitalWrite(zDir, HIGH);
for (int i = 0; i < 40; i++)
{
    xStepMotor();
    zStepMotor();
}
delay(1000);
}

void rotate()
{
    //ROTATE
    //Goes Clockwise
    digitalWrite(xDir, LOW);
    digitalWrite(yDir, LOW);
    digitalWrite(zDir, LOW);
    for (int i = 0; i < 200; i++)
    {
        yStepMotor();
    }
    delay(1000);


    //Goes Counterclockwise
    digitalWrite(xDir, HIGH);
    digitalWrite(yDir, HIGH);
    digitalWrite(zDir, HIGH);
    for (int i = 0; i < 200; i++)
    {
        yStepMotor();
    }
    delay(1000);
}

void xStepMotor()
{
    digitalWrite(xStep, HIGH);
    delayMicroseconds(500);
    digitalWrite(xStep, LOW);
    delayMicroseconds(500);
}

```

```
}  
  
void yStepMotor()  
{  
    digitalWrite(yStep,HIGH);  
    delayMicroseconds(2000);  
    digitalWrite(yStep,LOW);  
    delayMicroseconds(2000);  
}  
  
void zStepMotor()  
{  
    digitalWrite(zStep,HIGH);  
    delayMicroseconds(500);  
    digitalWrite(zStep,LOW);  
    delayMicroseconds(500);  
}
```

Appendix C: Flowchart of ROS Publisher Program



Appendix D: Flowchart of MATLAB Subscriber Program

