



Topological Data Analysis for Arrhythmia Detection through Modular Neural Networks

Meryll Dindin, Yuhei Umeda, Frédéric Chazal

► To cite this version:

Meryll Dindin, Yuhei Umeda, Frédéric Chazal. Topological Data Analysis for Arrhythmia Detection through Modular Neural Networks. CanadianAI 2020 - 33rd Canadian Conference on Artificial Intelligence, May 2020, Ottawa, Canada. hal-02155849

HAL Id: hal-02155849

<https://hal.inria.fr/hal-02155849>

Submitted on 14 Jun 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Topological Data Analysis for Arrhythmia Detection through Modular Neural Networks

Meryll Dindin^[1], Yuhei Umeda^[2], Frederic Chazal^[3]

^[1]meryll.dindin@student.ecp.fr ^[2]umeda.yuhei@jp.fujitsu.com ^[3]frederic.chazal@inria.fr

Abstract

This paper presents an innovative and generic deep learning approach to monitor heart conditions from ECG signals. We focus our attention on both the detection and classification of abnormal heartbeats, known as arrhythmia. We strongly insist on generalization throughout the construction of a deep-learning model that turns out to be effective for new unseen patient. The novelty of our approach relies on the use of topological data analysis as basis of our multichannel architecture, to diminish the bias due to individual differences. We show that our structure reaches the performances of the state-of-the-art methods regarding arrhythmia detection and classification.

1 Introduction

Countless artificial intelligence breakthroughs are observed in healthcare on a daily basis. They currently target improved monitoring of vital signs, better diagnostics and more reliable clinical decisions. Among the many on-going developments, heart monitoring is of particular importance as heart attack and strokes are among the five first causes of death in the US. Developing wearable medical devices would help to deal with a larger proportion of the population, and reduce the time used by cardiologists to make their diagnosis. This paper focuses on both the detection and classification of arrhythmia, which is an umbrella term for group of conditions describing irregular heartbeats. Detection deals with spotting any abnormal heartbeat, while classification deals with giving the right label to the spotted abnormal heartbeats.

Among the several existing studies, some developed descriptive temporal features to feed SVM [Houssein *et al.*, 2017] or neural networks [Shirin and Behbood, 2016], sometimes mixed with optimization methods [Houssein *et al.*, 2017; S.M. and E.S., 2013]. Others [Yochum *et al.*, 2016] dealt with wavelet transforms and Daubechies wavelets. The general approach of those papers enables arrhythmia classification through machine learning. However, most papers [Padmavathi and K.Sri, 2016; Jun *et al.*, 2017; Jianning, 2018; Hassanien *et al.*, 2018] reduce the classification to specific arrhythmia, or limit it to a few classes only. On the other hand, [Rajpurkar *et al.*, 2017] sought to improve multi-class

classification. But the more the classes, the faster the performances did vanish. To overcome this issue, [Jianning, 2018; Clifford *et al.*, 2017] have introduced deep learning methods based on convolutional networks. Other teams focused on unsupervised learning, such as auto-encoders [Lu *et al.*, 2018], with promising results. Nonetheless, the methods presented so far have low performance for unknown patient, mainly due to individual differences. Generalization, which means independence on individual differences, is a serious issue for any application in the healthcare sector.

The proposed approach consists in the analysis of ECG through a modular multi-channel neural network whose originality is to include a new channel relying on the theory of topological data analysis, that is able to capture robust topological patterns of the ECG signals. That information describes best the geometry of each heartbeat, independently of the values of the signal or the individual heart rhythms. By combining topological data analysis, handcrafted features and deep-learning, we aimed for better generalization.

Our paper is organized as follows. After presenting Topological Data Analysis, we condensed our approach in the presentation of the datasets, our preprocessing and the general deep-learning architecture. We then develop our testing methodology, which is used to quantify generalization. The last sections provide comparisons with benchmarks and state-of-the-art results, and conclude with our experimental results. We introduce a new benchmark for arrhythmia classification, underlining the strengths of topological data analysis and auto-encoders to tackle the issue of individual differences. Finally, remarks and thoughts are provided as conclusion at the end of the paper.

2 Topological Data Analysis

Among the main challenges faced for arrhythmia classification generalization, we find individual differences, and specifically bradycardia and tachycardia. We dealt with it by introducing Topological Data Analysis, and by merging theory with our deep-learning approach. Topological Data Analysis (TDA) is a recent and fast growing field that provides mathematically well-founded methods [Chazal and Michel, 2017] to efficiently exhibit topological patterns in data and to encode them into quantitative and qualitative features. In our setting, TDA, and more precisely *persistent homology theory* [Edelsbrunner and Harer, 2010], powerfully characterizes the

shape of the ECG signals in a compact way, avoiding complex geometric feature engineering. Thanks to fundamental stability properties of persistent homology [Chazal *et al.*, 2016], the TDA features appear to be very robust to the deformations of the patterns of interest in the ECG signal, especially expansion and contraction in the time axis direction. This makes them particularly useful to overcome individual differences and potential issues raised by bradycardia and tachycardia.

Persistence Homology. To characterize the heartbeats, we consider the persistent homology of the so called sub-level (resp. upper-level) sets filtration of the considered time series. Seeing the signal as a function f defined on an interval I and given a threshold value α , we consider the connected components of $F_\alpha = \{t \in I : f(t) \leq \alpha\}$ (resp. $F^\alpha = \{t \in I : f(t) \geq \alpha\}$). As α increases (resp. decreases) some components appear and some others get merged together. Persistent homology keeps track of the evolution of these components and encodes it in a *persistence barcode*, i.e. a set of intervals - see Figure 2 for an example of barcode computation on a simple example. The starting point of each interval corresponds to a value α where a new component is created while the end point corresponds to the value α' where the created component gets merged into another one. In our practical setting, the function f is the piecewise linear interpolation of the ECG time series and persistence barcodes can be efficiently computed in $O(n \log n)$ time, using, e.g., the GUDHI library [Maria *et al.*, 2014], where n is the number of nodes of the time series.




Figure 2: The persistence barcode of a 1D signal

To clarify the construction of a persistence barcode, one may observe Figure 2 with the following notations: $y = f(t)$: for $\alpha < \alpha_1$, F_α is empty. A first component appears in F_α as α reaches α_1 , resulting in the beginning of an interval. Similarly when α reaches α_2 and then α_3 , new components

appear in F_α giving birth to the starting point of new intervals. When α reaches α_4 , the two components born at α_1 and α_3 get merged, resulting in the “death” of the most recently born component (persistence rule), i.e. the one that appeared at α_3 and creation of the interval $[\alpha_3, \alpha_4]$ in the persistence barcode. Similarly when α reaches α_5 the interval $[\alpha_2, \alpha_5]$ is added to the barcode. The component appeared at α_1 remains until the end of the sweeping-up process, resulting in the interval $[\alpha_1, \alpha_6]$.

Betti Curves. As an unstructured set of intervals, the persistence barcodes are not suitable for direct integration in machine-learning models. To tackle this issue, we use a specific representation of the barcode diagrams, the so-called Betti curves [Umeda, 2016]: for each α , the Betti curve value at α is defined as the number of intervals containing the value α . The Betti curves are computed and discretized on the interval delimited by the minimum and maximum of the birth and death values of each persistent diagram, both for the time-series and its opposite (in order to study the sub-levels and upper-levels of the signal). One may observe that a fundamental property of Betti curves of 1D signal that follows from the definition of barcodes is their stability with respect to time re-parametrization and signal value rescaling, as stated in the following theorem. This allows us to build a uniform input for classical 1D convolutional deep-learning models, thus tackling the main issue of individual differences.

Theorem: *Time Independence of Betti Curves*

Given a function $f : I \rightarrow \mathbb{R}$ and a real number $a > 0$ the Betti curves of $t \rightarrow f(t)$ and $t \rightarrow f(at)$ are the same.

Moreover, if $g(t) = bf(t)$ for some $b > 0$, then the Betti curves of f and g are related by $BC_g(\alpha) = BC_f(\frac{\alpha}{b})$.

This theorem is a particular case of a more general statement resulting from classical properties of general persistence theory [Chazal *et al.*, 2016]. Intuitively, the invariance to time rescaling follows from the observation that persistence intervals measure the relative height of the peaks of the signal and not their width. The value-rescaling of the signal by a factor b results in a stretching of the persistence intervals by the same factor resulting in the above relation between the Betti curves of the signal and its rescaled version.




Figure 1: From three Consecutive Heartbeats to their corresponding persistence barcode and Betti curve

3 Deep-Learning Approach

3.1 Datasets

To facilitate comparison to other existing methods, our approach is experimented on a family of open-source data sets that have already been studied among the literature. Those are provided by the Physionet platform, and named after the diseases they describe: *MIT-BIH Normal Sinus Rhythm Database* [Goldberger *et al.*, 2000], *MIT-BIH Arrhythmia Database* [Goldberger *et al.*, 2000; Moody and Mark, 2001], *MIT-BIH Supraventricular Arrhythmia Database* [Goldberger *et al.*, 2000; Greenwald, 1990], *MIT-BIH Malignant Ventricular Arrhythmia Database* [Goldberger *et al.*, 2000; Greenwald, 1986] and *MIT-BIH Long Term Database* [Goldberger *et al.*, 2000]. Those databases present single-channel ECGs, each sampled at 360 Hz with 11-bit resolution over a 10 mV range. Two or more cardiologists independently annotated each record, whose disagreements were resolved to obtain the reference annotations for each beat in the databases. Each heartbeat is annotated independently, making peak detection thus unnecessary.

Database	# Patients	# Labels	Duration (Hours)
Arrhythmia	48	109494	24
LongTerm	7	668735	147
NormalSinus	18	1729629	437
SupraVentricular	78	182165	38
TWave	90	790565	180

Table 1: Physionet Datasets Description

3.2 Preprocessing

Every machine-learning comes with its data preprocessing. We first focused on the standardization of all the available ECG. Different methods have been applied in order to enhance the signal, and reduce noise and bias. After resampling at 200 Hz, we removed the baselines [Blanco-Velasco *et al.*, 2008] and applied filters, based on both a FIR filter and a Kalman filter. The signal is then rescaled between 0 and 1 before being translated to get a mean of the signal close to 0, for deep-learning stability.

Baseline Wander. The method dealing with the baseline drift [Blanco-Velasco *et al.*, 2008] is based on the Daubechies wavelets theory. It consists in consecutive processes of de-composition and reconstruction of the signal thanks to convolution windows. By removing the outlying components, we are able to identify and suppress the influence of the baseline in the signal, which generally corresponds to muscular and respiratory artifacts.

Filtering. The first applied filter to each ECG is a FIR (Finite Impulse Response) filter. It performs particularly well on ECG, and wavelets-based signals in general. It behaves basically as a band-filter. We chose 0.05 Hz and 50 Hz as cut frequencies to minimize the resulting distortion, according to our tests and the literature [Buendía-Fuentes *et al.*, 2012; Urganlawar and Chowhan, 2014; Goras and Fira, 2010].

Heartbeats Slicing. Once preprocessed, each ECG is segmented into partially overlapping elementary sequences made of a fix number of consecutive heartbeats. Each sequence is extracted according to the previous and next heartbeat. This extraction being patient-dependent, it reduces the influence of diverging heartbeat rhythms, *e.g* bradycardia and tachycardia. This extraction can be done for as many consecutive heartbeats as wanted. The labels are attributed by the central peak (whose index is the integer value of half the number of peaks). Once the windows are defined, we use interpolation to standardize the vectors, making them suitable for deep-learning purposes.

Feature Engineering. Once those heartbeats are extracted, we build relative features. Literature [Awais *et al.*, 2017; Pyakillya *et al.*, 2017; Blanco-Velasco *et al.*, 2008; Luz *et al.*, 2016] screening brought us to the discrete Fourier transform of each window, the linear relationships between each temporal components (P, Q, R, S, T), and the statistical values given by the extrema, mean, standard deviation, kurtosis, skewness, entropy, crossing-overs and PCA reduction to 10 components.

3.3 Auto-Encoder

Table 1 gives a good overview of other issues we faced, such as an uneven distribution of labels and extreme minority classes. Furthermore, a challenging imbalance between normal and abnormal samples is noticeable, as it would be for any anomaly problem.




Figure 3: Reconstruction Error on Heartbeat Signals

We decided to take advantage of the large amount of normal samples compared to abnormal samples, through unsupervised learning with auto-encoders [Baldi, 2012]. The structure is made of six fully-connected hidden layers, developed in a symmetric fashion, with an input dimension of size 400 and a latent space of size 20. The loss is defined by the *mean squared error*. The model is then trained on all the normal beats available, minimizing the reconstruction error between the input and the output. Once frozen, this model is integrated into our larger architecture. This reconstruction error, as presented in Figure 3, is already a good indicator for anomaly detection, but not satisfactory for classification.

Such a constructor may be used in two different ways to deal with binary classification: by either using the encoded inputs as new features, or using the reconstruction error through a subtraction layer between the input signal and the reconstructed signal. Those solutions are respectively referred to by *encoder* and *auto-encoder* in our architectures. Another way of using this structure is to directly integrate it into the deep-learning model. The concurrent optimization of two models is thus necessary, building a relational encoding space relative to the task. This is the strategy that has been applied for multi-class classification.

3.4 Architecture

Once the preprocessing done, we undertook the construction of our deep-learning approach to deal with the multi-modality of inputs. Our first objective was to determine whether the heartbeats are normal or abnormal, before determining their classification. The aim of such strategy is to avoid the issue of great imbalance between normal and abnormal samples, while focusing on an easier task before multi-class classification. An overview of our architecture is given in Figure 4.

Channels. For the auto-encoder, we use a convolution channel to deal with the subtracting layer, and a fully-connected layer to deal with the feature map given by the latent space. The input signals and the Betti curves are fed in convolution channels, aiming at extracting the right patterns [Kachuee *et al.*, 2018; Xia *et al.*, 2018; Isin and Ozdalili, 2017; M.M. Al *et al.*, 2016; Clifford *et al.*, 2017; Rajpurkar *et al.*, 2017]. The other inputs (both features and discrete Fast Fourier Transform coefficients) are injected into fully connected networks.

Annealed DropOut. As we launched a first battery of tests, we were confronted to the unexpected strong influence of the DropOut parameter. Its value could dramatically change the results. Since DropOut is of great help for generalization, we sought a way to deal with that issue. A solution came from the annealing dropout technique [Rennie *et al.*, 2014; Jimmy Ba Brendan Frey, 2013], which consists in scheduling the decrease of the dropout rate. It helped us stabilizing the results.

4 Experimental Results

From the problem presentation, we highlighted two issues, relative to any healthcare machine-learning problem: imbalanced datasets and individual differences. The fewer the patients and the bigger the imbalance, the greater the influence of individual differences. To deal with the issue of imbalance, we introduced our auto-encoder architecture, while dealing with the individual differences by introducing Topological Data Analysis. Once we established our solution to both imbalance and individual differences, we aimed at developing our own approach and validation. As we mentioned earlier, two ways have been explored, both for performance enhancement and reduction of the influence of imbalance. The first one has been to detect whether a heartbeat is normal or abnormal, in order to get a first classification. The second one has been multi-class classification (13 classes) on the arrhythmic heartbeats only. Our objective is to introduce a new benchmark to attest that TDA (and auto-encoders) do improve generalization for arrhythmia classification.

Training Parameters. Different methods have been used for the model training and optimization. Firstly, all the channels described previously are concatenated into one fully-connected network, dealing with all the obtained feature maps concurrently. Secondly, all the activation layers used are *PReLU*, initialized with *he_normal* [He *et al.*, 2015; Srivastava *et al.*, 2015]. Thirdly, the dropout has been parametrized according to the strategy of the *annealing dropout*, from a rate of 0.5 to a rate of 0.0 after 100 epochs. Concerning the losses, we used the *categorical_crossentropy* or *binary_crossentropy* for the classification model, and *mean_squared_error* for the auto-encoder structure. *Adadelta* was used for optimization with an initial learning rate of 1.0.




Figure 4: General Overview of Deep-Learning Architecture

Testing Methodology. Dealing with a health issue, the testing methodology has to be rigorously defined to accurately analyze the performances. A great importance was given to generalization abilities of the developed models. For that purpose, our strategy aimed at performing **patient-based cross-validation**, which means that for each model, train and validation sets were build on a fraction of the available patients, while the remaining patients constituted the test set. This way, the validation score would demonstrate the ability of the model to dissociate arrhythmias on known patients, while the test score would demonstrate the ability of the model to detect arrhythmias on new patients. By using permutations of all the available patients, we were able to train, validate and test each model on all patients. The results presented in the following parts stem from a cross-validation keeping 5 unique patients for testing at each cross-validation permutation.

4.1 Channel Comparison

We first quantified the importance of each introduced channel, by turning them off and on. Such strategy allowed us to specifically quantify the influence of the introduced TDA channel, as improving the general ability of our architecture to both detect and classify arrhythmias. We tested our architecture through a patient-based cross-validation. Our format consisted in 10 experiments, made with the underlying will of generalization: **for each experiment**, 225 patients are used for both training (70%) and validation (30%), while 15 patients are kept for testing. For the purpose of validation, each subset of 15 patients is not overlapping between experiments.

ID	Arrhythmia Detection		Arrhythmia Classification	
	With TDA	Without TDA	With TDA	Without TDA
0	0.99	0.98	0.73	0.68
1	0.96	0.90	0.75	0.69
2	0.85	0.86	0.68	0.65
3	0.94	0.95	0.95	0.96
4	0.85	0.80	0.97	0.97
5	0.87	0.77	0.96	0.93
6	0.78	0.80	0.94	0.93
7	0.81	0.63	0.90	0.80
8	0.79	0.65	0.85	0.78
9	0.84	0.86	0.68	0.47

Table 2: Weighted Test Accuracy for Channel Comparisons. TDA improves both detection and classification, with a major improvement in arrhythmia classification, but also greatly enhances performances reliability.

Finally, a closer look at Table 2 supports the importance of TDA. Its role is emphasized for multi-class classification, with a general greater improvement of performances. With this right combination of channels, we aimed for testing through patient-based cross-validation for both binary and multi-class classification. **For the purpose of the demonstration, the scores are weighted in order to compensate for the general imbalance.** Moreover, multi-class classification is not biased by normal samples since they have been extracted beforehand. This finally supports the generalization

role of TDA, that is expected to bring improvements combined with other deep-learning architectures as well.

4.2 Arrhythmia Detection

Our first benchmark dealt with arrhythmia detection (binary classification). It consisted in using our architecture, enhanced with the (auto-) encoder trained in an unsupervised manner on normal beats. The model determined by channel comparison has thus been used for cross-validation. Each instance of cross-validation has been made by randomly undersampling the majority class to obtain balanced datasets. It takes approximately 10 hours to train on a GPU (GeForce GTX). We used the data structure previously presented to test over the 240 patients we have in our datasets. Moreover, to tackle the issue of anomaly detection, and accelerate the process of validation, each cross-validation round is respectively built out of a set of 5 unknown patients. The mean accuracy score is **98%** for validation and **90%** for test. This approach shows great generalization abilities. Unfortunately, no other paper do use those test settings for comparison. With a closer look on the results, the low performances appear on patients for which it was hard to recognize their normal beats. It also means that more patients may improve the generalization abilities of the model. However, its performances on the validation results prove its abilities to learn about specific patients (suitable for personalized devices issues).

4.3 Arrhythmia Classification

The same strategy has been applied for multi-class classification. The greatest channel influence is provided by TDA and the encoder. As a consequence, we reduced the original model to the one composed of four channels in the same fashion than we did for anomaly detection. Moreover, the influence of those channels is greater than observed for binary classification. Since the previous approach was not enough, we went further with 13-class classification. The models proved their ability to learn about heartbeat condition through cross-validation, with a mean validation score of **97.3%**, while being able to generalize this acquired knowledge on patients it never saw, with a mean testing accuracy of **80.5%**. Once again, literature does not provide comparable settings. The use of cross-validation focused on the generalization ability of the model. By also removing the normal beats, we focus on differences between the different arrhythmias, and remove the influence of imbalance that is generally found in the scores presented in the literature.

5 Benchmarks Comparison

5.1 Premature Ventricular Heartbeats Detection

Since those open-source datasets have been exploited by others, we sought to compare our own architecture to existing benchmarks. Our claim is an enhanced generalization thanks to TDA and the auto-encoder architecture we developed. To support it, our first confrontation has been made with [Jianing, 2018], which focuses on the detection of premature ventricular contractions (PVC). The detection of a specific arrhythmia is a particular case of anomaly detection (one-vs-all), for which our architecture is suitable. The results we

obtained are presented in tables 3 and 4, and support the generalization ability of our model. For the comparison, we applied the settings used in the paper. Out of the 48 initial patients in the MIT-BIH Arrhythmia Database, 4 patients are discarded. The remaining patients are split in two groups: 22 are used for training and validation, 22 are used for testing. The objective of such approach is to use the premature ventricular contractions, which are the majority class among the available arrhythmias. The results we obtained with such configuration are given in Table 3 for the MIT-BIH Arrhythmia, where *PPV* stands for *Positive Predictive Value*.

Paper	Accuracy	PPV	Sensitivity
Proposed	99.2%	95.1%	95.5%
Li et al. (2018)	99.4%	93.9%	98.2%

Table 3: Detection of PVC - MIT-BIH Database

However, they went further by considering the five databases. Our model being suited to that larger amount of data, we also compared our performances to their new settings. This time they split the group of 240 patients in two, both 120 for training and validation, and 120 for testing. Once again, our performances are given in Table 4. This experiment shows a great enhancement due to samples augmentation in this case of one-vs-all classification.

Paper	Accuracy	PPV	Sensitivity
Proposed	96.1%	97.0%	96.0%
Li et al. (2018)	95.6%	94.1%	92.7%

Table 4: Detection of PVC - Augmented Databases. With more patients, those settings imply a greater need for generalization, which emphasizes our model ability to learn the concept of PVC and apply it to unknown patients.

5.2 8-Classes Classification

The previous experiment is a specific use-case for our architecture. In this second comparison, we focus on 8-classes classification [Jun *et al.*, 2017]. Once again, our claim is better generalization, which is done thanks to patient-based cross-validation. Nonetheless, their settings imply a limitation to the MIT-BIH Arrhythmia Database, from which they select 8 classes, comprising normal beats. Unexpectedly, this selection does not correspond to the majority classes. Our performances are compared in Table 5, extending the results they present in their paper. We pinpoint, thanks to those results, the generalization ability of our model, which has better positive predictive value (here precision) and sensitivity. It finally underlines a more efficient classification.

6 Conclusion

We developed a new approach to deal with the issue of generalization in arrhythmia detection and classification. Our innovative architecture uses common source of information, Topological Data Analysis and auto-encoders. We supported

Paper	Classes	Acc	PPV	Sen
Proposed	8	99.0%	99.0%	98.5%
Jun et al. (2017)	8	99.0%	98.5%	97.8%
Kiranyaz et al. (2017)	5	96.4%	79.2%	68.8%
Gler et al. (2018)	4	96.9%	-	96.3%
Melgani et al. (2008)	6	91.7%	-	93.8%

Table 5: 8-Classes Classification - MIT-BIH Database

our claim of improved generalization with scores reaching the performances of state of the art methods, and above. Our experiments pinpoint the strengths of TDA and auto-encoders to improve generalization results. Moreover, the modularity of such model allows us to build and add new channels, such as a possible channel based on the Wavelet transform [Xia *et al.*, 2018], which also gives a good description of the ECG time-series. Finally, we give a new benchmark on five open-source datasets, and as it is often the case in deep-learning, we still envision greater performances with larger datasets such as [Clifford *et al.*, 2017].

References

- [Awais *et al.*, 2017] Muhammad Awais, Nasreen Badruddin, and Micheal Driberg. A hybrid approach to detect driver drowsiness utilizing physiological signals to improve system performance and Wearability. *Sensors (Switzerland)*, 2017.
- [Baldi, 2012] P Baldi. Autoencoders, unsupervised learning, and deep architectures. 27:37–50, 2012.
- [Blanco-Velasco *et al.*, 2008] Manuel Blanco-Velasco, Binwei Weng, and Kenneth E. Barner. ECG signal denoising and baseline wander correction based on the empirical mode decomposition. *Computers in Biology and Medicine*, 2008.
- [Buendía-Fuentes *et al.*, 2012] F. Buendía-Fuentes, M. A. Arnau-Vives, A. Arnau-Vives, Y. Jiménez-Jiménez, J. Rueda-Soriano, E. Zorio-Grima, A. Osa-Sáez, L. V. Martínez-Dolz, L. Almenar-Bonet, and M. A. Palencia-Pérez. High-Bandpass Filters in Electrocardiography: Source of Error in the Interpretation of the ST Segment. *ISRN Cardiology*, 2012.
- [Chazal and Michel, 2017] Frederic Chazal and Bertrand Michel. An introduction to topological data analysis: fundamental and practical aspects for data scientists. *Submitted to the Journal de la Societe Francaise de Statistiques*, 2017.
- [Chazal *et al.*, 2016] Frédéric Chazal, Vin de Silva, Marc Glisse, and Steve Oudot. *The structure and stability of persistence modules*. SpringerBriefs in Mathematics. Springer, 2016.
- [Clifford *et al.*, 2017] Gari D Clifford, Chengyu Liu, Benjamin Moody, Li-wei H Lehman, Ikaro Silva, Qiao Li, A E Johnson, and Roger G Mark. AF Classification from a short single lead ECG recording: the PhysioNet/Computing in Cardiology Challenge 2017. 2017.

- [Edelsbrunner and Harer, 2010] Herbert Edelsbrunner and John Harer. *Computational Topology: An Introduction*. AMS, 2010.
- [Goldberger *et al.*, 2000] AL Goldberger, LAN Amaral, L Glass, JM Hausdorff, PCh Ivanov, RG Mark, JE Mietus, GB Moody, C-K Peng, and HE Stanley. PhysioBank, PhysioToolkit, and PhysioNet: Components of a New Research Resource for Complex Physiologic Signals. 2000.
- [Goras and Fira, 2010] Liviu Goras and Monica Fira. Pre-processing Method For Improving Ecg Signal Classification And Compression Validation. 2010.
- [Greenwald, 1986] SD Greenwald. Development and analysis of a ventricular fibrillation detector. *MIT Dept. of Electrical Engineering and Computer Science*, 1986.
- [Greenwald, 1990] SD Greenwald. Improved detection and classification of arrhythmias in noise-corrupted electrocardiograms using contextual information. *Harvard-MIT Division of Health Sciences and Technology*, 1990.
- [Hassanien *et al.*, 2018] Aboul Ella Hassanien, Moataz Kilany, and Essam H. Houssein. Combining support vector machine and elephant herding optimization for cardiac arrhythmias. *CoRR*, abs/1806.08242, 2018.
- [He *et al.*, 2015] Kaiming He, Xiangyu Zhang, Shaoqing Ren, and Jian Sun. Delving Deep into Rectifiers: Surpassing Human-Level Performance on ImageNet Classification. 2015.
- [Houssein *et al.*, 2017] Essam Houssein, Moataz Kilany, and Aboul Ella Hassanien. Ecg signals classification: a review. 5:376, 01 2017.
- [Isin and Ozdalili, 2017] Ali Isin and Selen Ozdalili. Cardiac arrhythmia detection using deep learning. In *Procedia Computer Science*, 2017.
- [Jianning, 2018] Li Jianning. Detection of Premature Ventricular Contractions Using Densely Connected Deep Convolutional Neural Network with Spatial Pyramid Pooling Layer. 2018.
- [Jimmy Ba Brendan Frey, 2013] Lei Jimmy Ba Brendan Frey. Adaptive dropout for training deep neural networks. 2013.
- [Jun *et al.*, 2017] Tae Joon Jun, Hoang Minh Nguyen, Daeyoun Kang, Dohyeun Kim, Young-Hak Kim, and Daeyoung Kim. Ecg arrhythmia classification using a 2-d convolutional neural network (submitted). 04 2017.
- [Kachuee *et al.*, 2018] Mohammad Kachuee, Shayan Fazeli, and Majid Sarrafzadeh. ECG Heartbeat Classification: A Deep Transferable Representation. 2018.
- [Lu *et al.*, 2018] Weijia Lu, Jie Shuai, Shuyan Gu, and Joel Xue. Method to annotate arrhythmias by deep network. 06 2018.
- [Luz *et al.*, 2016] Eduardo José da S. Luz, William Robson Schwartz, Guillermo Cámara-Chávez, and David Menotti. ECG-based heartbeat classification for arrhythmia detection: A survey. *Computer Methods and Programs in Biomedicine*, 2016.
- [Maria *et al.*, 2014] C. Maria, J.-D. Boissonnat, M. Glisse, and M. Yvinec. The gudhi library: Simplicial complexes and persistent homology. In *International Congress on Mathematical Software*, pages 167–174. Springer, 2014.
- [M.M. Al *et al.*, 2016] Rahhal M.M. Al, Bazi Yakoub, Al-Hichri Haikel, Alajlan Naif, Melgani Farid, and Yager R.R. Deep learning approach for active classification of electrocardiogram signals. *Information Sciences*, 345:340 – 354, 2016.
- [Moody and Mark, 2001] GB Moody and RG Mark. The impact of the MIT-BIH Arrhythmia Database. *IEEE Eng in Med and Biol*, 2001.
- [Padmavathi and K.Sri, 2016] Kora Padmavathi and Rama Krishna K.Sri. Hybrid firefly and particle swarm optimization algorithm for the detection of bundle branch block. *International Journal of the Cardiovascular Academy*, 2(1):44 – 48, 2016.
- [Pyakillya *et al.*, 2017] B. Pyakillya, N. Kazachenko, and N. Mikhailovsky. Deep Learning for ECG Classification. In *Journal of Physics: Conference Series*, 2017.
- [Rajpurkar *et al.*, 2017] Pranav Rajpurkar, Awni Y Hannun, Masoumeh Haghpanahi, Codie Bourn, and Andrew Y Ng. Cardiologist-level arrhythmia detection with convolutional neural networks. 2017.
- [Rennie *et al.*, 2014] Steven J Rennie, Vaibhava Goel, and Samuel Thomas. Annealed Dropout Training Of Deep Networks. 2014.
- [Shirin and Behbood, 2016] Shadmam Shirin and Mashoufi Behbood. A new personalized ecg signal classification algorithm using block-based neural network and particle swarm optimization. *Biomedical Signal Processing and Control*, 25:12 – 23, 2016.
- [S.M. and E.S., 2013] Abd-Elazim S.M. and Ali E.S. A hybrid particle swarm optimization and bacterial foraging for optimal power system stabilizers design. *International Journal of Electrical Power and Energy Systems*, 46:334 – 341, 2013.
- [Srivastava *et al.*, 2015] Rupesh Kumar Srivastava, Klaus Greff, and Urgan Schmidhuber. Training Very Deep Networks. 2015.
- [Umeda, 2016] Yuhei Umeda. Time series classification via topological data analysis. *Transactions of the Japanese Society for Artificial Intelligence*, Vol. 32, 2016.
- [Uppanlawar and Chowhan, 2014] Isha V Uppanlawar and Harshal Chowhan. Pre-processing of ECG Signals Using Filters. *International Journal of Computer Trends and Technology*, 11(4), 2014.
- [Xia *et al.*, 2018] Yong Xia, Naren Wulan, Kuanquan Wang, and Henggui Zhang. Detecting atrial fibrillation by deep convolutional neural networks. *Computers in Biology and Medicine*, 2018.
- [Yochum *et al.*, 2016] Maxime Yochum, Charlotte Renaud, and Sabir Jacquir. Automatic detection of p, qrs and t patterns in 12 leads ecg signal based on cwt. *Biomedical Signal Processing and Control*, 25:46 – 52, 2016.