

CONF - 820321--5

CONF-820321--5

DE82 009636

NEUTRON-EXPOSURE PARAMETERS FOR THE FOURTH HSST SERIES
OF METALLURGICAL IRRADIATION CAPSULES*

F. B. K. Kam, F. W. Stallmann, C. A. Baldwin, and A. Fabry[†]

Oak Ridge National Laboratory
Oak Ridge, Tennessee 37830

[†]CEN/SCK, Mol, Belgium

MASTER

Paper to be presented at the
Fourth ASTM-EURATOM Symposium on Reactor Dosimetry
March 22-26, 1982 at NBS, Washington, DC

By acceptance of this article, the publisher or recipient acknowledges the U.S. Government's right to retain a nonexclusive, royalty-free license in and to any copyright covering the article.

- DISCLAIMER

the first time in history that the people of the United States have been asked to make a choice between two political parties, each of which has a different program for the future of the country.

NOTICE

PORTIONS OF THIS REPORT ARE ILLEGIBLE. It has been reproduced from the best available copy to permit the broadest possible availability.

*Research sponsored by the Division of Reactor Safety Research, U.S. Nuclear Regulatory Commission under Interagency Agreement DOE 40-551-75 with the U.S. Department of Energy under contract W-7405-eng-26 with the Union Carbide Corporation.

DISTRIBUTION OF THIS DOCUMENT IS UNLIMITED

NEUTRON EXPOSURE PARAMETERS FOR THE FOURTH HSST SERIES
OF METALLURGICAL IRRADIATION CAPSULES*

F. B. K. Kam, F. W. Stallmann, C. A. Baldwin, and A. Fabry[†]

Oak Ridge National Laboratory
Oak Ridge, Tennessee 37830

[†]CEN/SCK, Mol, Belgium

ABSTRACT

The neutron exposure parameters for the Heavy Section Steel Technology (HSST) Experiments performed at the Oak Ridge National Laboratory (ORNL) can be determined conservatively to $\pm 10\%$ (1σ) variance.

The neutron exposure parameters used for this study were fluence greater than 1 MeV, fluence greater than 0.1 MeV, and displacements per atom (dpa). Measured reaction rates, calculated neutron transport fluxes, and cross sections values were combined in the logarithmic least square adjustment code LSL.⁽¹⁾

1. Introduction

The U.S. Nuclear Regulatory Commission (NRC) is conducting an extensive research program^(2,3) to study fracture toughness of irradiated pressure vessel materials in the upper transition region and to investigate the applicability of small specimen test results to thick-section materials. This study has been extended to the study of upper-shelf behavior (plastic behavior). The first three irradiation experiments (nine capsules) contained fracture toughness specimens of four-inch (4T-CS) and a number of smaller specimens of low-shelf weldments.

*Research sponsored by the Division of Reactor Safety Research, U.S. Nuclear Regulatory Commission under Interagency Agreement DOE 40-551-75 with the U.S. Department of Energy under contract W-7405-eng-26 with the Union Carbide Corporation.

The fourth HSST irradiation series (four capsules) is primarily designed to obtain statistical data on the fracture toughness of "current practice" weldments. Each of the capsules contains one-inch-thick fracture toughness (1T-CS), Charpy V-notch, and tension test specimens (Figs. 1 and 2).

In support of the material irradiation experiments, a neutron characterization program was initiated to provide accurate exposure parameters for correlation with the property change rate data. The dosimetry results of the second and third HSST series have been reported in Refs. 4 and 5. The experience gained in these two experiments have led to modifications in the composition and distribution of the dosimeters which monitor the flux spectrum in the irradiated steel specimens. In addition, multiple foil sets were irradiated in simulated HSST irradiation capsules to obtain detailed neutron spectrum information. This dosimetry experiment was a joint effort between CEN/SCK, Mol, Belgium and ORNL. The methods and techniques of measurement, calculation, and analysis are the same as applied to the neutron spectral characterization of the PCA experiments and Blind Test.⁽⁶⁾

2. Results

The results reported in this paper are for capsules A and B of the fourth HSST irradiation series (Fig. 3). Tables 1 and 2 show the exposure parameters (fluence > 1 MeV, fluence > 0.1 MeV and dpa) for each specimen in capsule A. For the 1T-CT specimens, the exposure values represent values at the crack tip. For the charpy specimens, the values are given at the apex of the v-notch. Similarly Tables 3 and 4 represent the exposure values for capsule B. A 3-dimensional map for the exposure parameters has been determined. This map has the form

$$\phi(x,y,z) = \phi_0 \cos B_x (x-x_0) \cos B_y (y-y_0) e^{-\lambda z} . \quad (1)$$

Preliminary estimates yield a $\pm 10\%$ (1σ) variance for the exposure parameters obtained from equation (1).

A technical report will be issued at a later date with a detailed uncertainty analysis and the methodology that was used to arrive at the

ORNL-DWG 82-10483




Fig. 1. Fourth HSST Series of Irradiation Capsules

1	21	41	42	141
		43	44	
2	22	45	46	142
		47	48	
3	23	49	50	143
		51	52	
4	24	53	54	144
		55	56	
5	25	57	58	145
		59	60	
6	26	61	62	146
		63	64	
7	27	65	66	147
		67	68	
8	28	69	70	148
		71	72	
9	29	73	74	149
		75	76	
10	30	77	78	149
		79	80	
11	31	81	82	150
		83	84	
12	32	85	86	151
		87	88	
13	33	89	90	152
		91	92	
14	34	93	94	153
		95	96	
15	35	97	98	154
		99	100	
16	36	101	102	155
		103	104	
17	37	105	106	156
		107	108	
18	38	109	110	157
		111	112	
19	39	113	114	158
		115	116	
20	40	117	118	159
		119	120	
		121	122	160
		123	124	
		125	126	
		127	128	
		128	130	
		131	132	
		133	134	
		135	136	
		137	138	
		139	140	

Fig. 2. Specimen Position Numbers for Fourth HSST Series Irradiation Capsules.




Fig. 3. Fourth HSST Experimental Configuration.

Table 1. Exposure Parameters for Capsule A 1T-CT Specimens

Specimen Position No.	Fluence >1 MeV	Fluence >.1 MeV	DPA	Specimen Position No.	Fluence >1 MeV	Fluence >.1 MeV	DPA	Specimen Position No.	Fluence >1 MeV	Fluence >.1 MeV	DPA
1	7.9485E+18	2.6420E+19	1.2545E-02	21	1.0480E+19	3.6404E+19	1.6890E-02	141	8.8131E+18	3.0132E+19	1.4112E-02
2	1.0668E+19	3.5955E+19	1.6933E-02	22	1.3109E+19	4.6066E+19	2.1232E-02	142	1.0917E+19	3.7778E+19	1.7570E-02
3	1.3254E+19	4.5022E+19	2.1107E-02	23	1.5589E+19	5.5174E+19	2.1232E-02	143	1.2899E+19	4.4977E+19	2.0826E-02
4	1.5676E+19	5.3500E+19	2.5012E-02	24	1.7891E+19	6.3618E+19	2.9125E-02	144	1.4737E+19	5.1645E+19	2.3842E-02
5	1.8467E+19	6.3250E+19	2.9508E-02	25	2.0517E+19	7.3227E+19	3.3452E-02	145	1.6830E+19	5.9224E+19	2.7275E-02
6	2.0406E+19	7.0000E+19	3.2627E-02	26	2.2321E+19	7.9801E+19	3.6419E-02	146	1.8266E+19	6.4404E+19	2.9626E-02
7	2.2092E+19	7.5836E+19	3.5331E-02	27	2.3871E+19	8.5414E+19	3.8960E-02	147	1.9497E+19	6.8822E+19	3.1636E-02
8	2.3502E+19	8.0684E+19	3.7587E-02	28	2.5148E+19	9.0000E+19	4.1045E-02	148	2.0510E+19	7.2428E+19	3.3284E-02
9	2.4875E+19	8.5332E+19	3.9767E-02	29	2.6358E+19	9.4272E+19	4.3006E-02	149	2.1467E+19	7.5780E+19	3.4829E-02
10	2.5600E+19	8.7715E+19	4.0904E-02	30	2.5966E+19	9.6334E+19	4.3972E-02	150	2.1945E+19	7.7392E+19	3.5586E-02
11	2.6007E+19	8.8953E+19	4.1521E-02	31	2.7266E+19	9.7237E+19	4.4424E-02	151	2.2177E+19	7.8090E+19	3.5935E-02
12	2.6090E+19	8.9031E+19	4.1611E-02	32	2.7255E+19	9.6969E+19	4.4357E-02	152	2.2160E+19	7.7864E+19	3.5872E-02
13	2.5728E+19	8.7457E+19	4.0963E-02	33	2.6793E+19	9.4949E+19	4.3524E-02	153	2.1781E+19	7.7864E+19	3.5195E-02
14	2.5079E+19	8.4921E+19	3.9862E-02	34	2.6982E+19	9.2060E+19	4.2289E-02	154	2.1206E+19	7.3960E+19	3.4209E-02
15	2.4118E+19	8.1277E+19	3.8254E-02	35	2.5073E+19	8.8062E+19	4.0560E-02	155	2.0393E+19	7.0792E+19	3.2826E-02
16	2.2857E+19	7.6573E+19	3.6160E-02	36	2.3777E+19	8.3005E+19	3.8356E-02	156	1.9351E+19	6.6788E+19	3.1067E-02
17	2.0848E+19	6.9168E+19	3.2842E-02	37	2.1743E+19	7.5151E+19	3.4914E-02	157	1.7717E+19	6.0572E+19	2.8322E-02
18	1.8969E+19	6.2305E+19	2.9752E-02	38	1.9858E+19	6.7933E+19	3.1737E-02	158	1.6204E+19	5.4859E+19	2.5789E-02
19	1.6854E+19	5.4629E+19	2.6285E-02	39	1.7746E+19	5.9897E+19	2.8189E-02	159	1.4510E+19	4.8498E+19	2.2960E-02
20	1.4530E+19	4.6241E+19	2.2483E-02	40	1.5432E+19	5.1141E+19	2.4311E-02	160	1.2652E+19	4.1563E+19	1.9867E-02

Table 2. Exposure Parameters for Capsule A Charpy Specimens

Specimen Position No.	Fluence >1 MeV	Fluence <.1 MeV	DPA	Specimen Position No.	Fluence >1 MeV	Fluence >>1 MeV	DPA
41	7.6558E+18	2.7335E+19	1.2598E-02	42	7.5128E+18	2.6789E+19	1.2372E-02
43	8.5618E+18	3.0795E+19	1.4133E-02	44	8.4019E+18	3.0179E+19	1.3880E-02
45	9.4528E+18	3.4196E+19	1.5643E-02	46	9.2762E+18	3.3513E+19	1.5362E-02
47	1.0327E+19	3.7534E+19	1.7124E-02	48	1.0134E+19	3.6784E+19	1.6816E-02
49	1.1183E+19	4.0802E+19	1.8573E-02	50	1.0974E+19	3.9987E+19	1.8240E-02
51	1.2019E+19	4.3993E+19	1.9990E-02	52	1.1795E+19	4.3114E+19	1.9631E-02
53	1.2834E+19	4.7102E+19	2.1370E-02	54	1.2594E+19	4.6161E+19	2.0986E-02
55	1.3626E+19	5.0123E+19	2.2711E-02	56	1.3372E+19	4.9122E+19	2.2303E-02
57	1.4394E+19	5.3050E+19	2.4011E-02	58	1.4125E+19	5.1990E+19	2.3580E-02
59	1.5137E+19	5.5878E+19	2.5268E-02	60	1.4854E+19	5.4762E+19	2.4874E-02
61	1.6437E+19	6.0820E+19	2.7465E-02	62	1.6129E+19	5.9605E+19	2.6971E-02
63	1.7100E+19	6.3337E+19	2.8585E-02	64	1.6780E+19	6.2072E+19	2.8071E-02
65	1.7733E+19	6.5736E+19	2.9653E-02	66	1.7401E+19	6.4423E+19	2.9121E-02
67	1.8334E+19	6.8012E+19	3.0667E-02	68	1.7992E+19	6.6654E+19	3.0117E-02
69	1.8903E+19	7.0161E+19	3.1626E-02	70	1.8550E+19	6.8759E+19	3.1058E-02
71	1.9439E+19	7.2178E+19	3.2527E-02	72	1.9076E+19	7.0736E+19	3.1943E-02
73	1.9940E+19	7.4061E+19	3.3370E-02	74	1.9568E+19	7.2581E+19	3.2770E-02
75	2.0406E+19	7.5805E+19	3.4151E-02	76	2.0025E+19	7.4290E+19	3.3538E-02
77	2.0836E+19	7.7407E+19	3.4871E-02	78	2.0447E+19	7.5861E+19	3.4245E-02
79	2.1229E+19	7.8864E+19	3.5527E-02	80	2.0833E+19	7.7289E+19	3.4889E-02
81	2.1860E+19	8.1181E+19	3.6576E-02	82	2.1452E+19	7.9559E+19	3.5919E-02
83	2.2144E+19	8.2210E+19	3.7045E-02	84	2.1730E+19	8.0568E+19	3.6380E-02
85	2.2389E+19	8.3086E+19	3.7447E-02	86	2.1971E+19	8.1426E+19	3.6774E-02
87	2.2594E+19	8.3806E+19	3.7781E-02	88	2.2172E+19	8.2132E+19	3.7332E-02
89	2.2760E+19	8.4569E+19	3.8046E-02	90	2.2334E+19	8.2684E+19	3.7363E-02
91	2.2885E+19	8.4775E+19	3.8242E-02	92	2.2457E+19	8.3081E+19	3.7556E-02
93	2.2969E+19	8.5022E+19	3.8369E-02	94	2.2540E+19	8.3324E+19	3.7680E-02
95	2.3013E+19	8.5110E+19	3.8426E-02	96	2.2583E+19	8.3410E+19	3.7736E-02
97	2.3016E+19	8.5039E+19	3.8414E-02	98	2.2586E+19	8.3340E+19	3.7724E-02
99	2.2979E+19	8.4808E+19	3.8332E-02	100	2.2550E+19	8.3114E+19	3.7643E-02
101	2.3030E+19	8.3967E+19	3.7997E-02	102	2.2377E+19	8.2290E+19	3.7315E-02
103	2.2650E+19	8.3288E+19	3.7718E-02	104	2.2227E+19	8.1624E+19	3.7040E-02
105	2.2458E+19	8.2453E+19	3.7370E-02	106	2.2038E+19	8.0806E+19	3.6699E-02
107	2.2225E+19	8.1463E+19	3.6954E-02	108	2.1810E+19	7.9836E+19	3.6290E-02
109	2.1953E+19	8.0321E+19	3.6471E-02	110	2.1543E+19	7.8717E+19	3.5816E-02
111	2.1643E+19	7.9029E+19	3.5922E-02	112	2.1238E+19	7.7451E+19	3.5277E-02
113	2.1294E+19	7.7589E+19	3.5307E-02	114	2.0896E+19	7.6040E+19	3.4673E-02
115	2.0907E+19	7.6004E+19	3.4629E-02	116	2.0517E+19	7.4486E+19	3.4007E-02
117	2.0484E+19	7.4277E+19	3.3888E-02	118	2.0101E+19	7.2794E+19	3.3279E-02
119	2.0024E+19	7.2411E+19	3.3085E-02	120	1.9650E+19	7.0965E+19	3.2491E-02
121	1.9083E+19	6.8614E+19	3.1446E-02	122	1.8726E+19	6.7243E+19	3.0882E-02
123	1.8524E+19	6.6372E+19	3.0477E-02	124	1.8178E+19	6.5046E+19	2.9930E-02
125	1.7913E+19	6.4007E+19	2.9452E-02	126	1.7598E+19	6.2728E+19	2.8923E-02
127	1.7311E+19	6.1521E+19	2.8374E-02	128	1.6987E+19	6.0292E+19	2.7664E-02
129	1.6657E+19	5.8921E+19	2.7244E-02	130	1.6346E+19	5.7744E+19	2.6755E-02
131	1.5974E+19	5.6210E+19	2.6064E-02	132	1.5676E+19	5.5088E+19	2.5597E-02
133	1.5974E+19	5.3395E+19	2.4838E-02	134	1.4978E+19	5.2328E+19	2.4392E-02
135	1.4525E+19	5.0479E+19	2.3566E-02	136	1.4254E+19	4.9471E+19	2.3143E-02
137	1.3762E+19	4.7469E+19	2.2252E-02	138	1.3505E+19	4.6521E+19	2.1852E-02
139	1.2973E+19	4.4370E+19	2.0897E-02	140	1.2731E+19	4.3484E+19	2.0522E-02

Table 3. Exposure Parameters for Capsule B 1T-CT Specimens

Specimen Position No.	Fluence 1 MeV	Fluence .1 MeV	DPA	Specimen Position No.	Fluence 1 MeV	Fluence .1 MeV	DPA	Specimen Position No.	Fluence 1 MeV	Fluence .1 MeV	DPA
1	5.5006E+18	1.8511E+19	8.7261E-03	21	6.9316E+18	2.4142E+19	1.1131E-02	141	5.6745E+18	1.9713E+19	9.0076E-03
2	7.1566E+18	2.4418E+19	1.1410E-02	22	8.8069E+18	3.0995E+19	1.4212E-02	142	7.1586E+18	2.4395E+19	1.1409E-02
3	8.7272E+18	3.0017E+19	1.3955E-02	23	1.0579E+19	3.7464E+19	1.7121E-02	143	8.5609E+18	2.9382E+19	1.3677E-02
4	1.0194E+19	3.5238E+19	1.6330E-02	24	1.2226E+19	4.3471E+19	1.9824E-02	144	9.8654E+18	3.4015E+19	1.5785E-02
5	1.1879E+19	4.1222E+19	1.9055E-02	25	1.4109E+19	5.0315E+19	2.2909E-02	145	1.1358E+19	3.9299E+19	1.8193E-02
6	1.3047E+19	4.5350E+19	2.0939E-02	26	1.5405E+19	5.5005E+19	2.5028E-02	146	1.2387E+19	4.2925E+19	1.8193E-02
7	1.4060E+19	4.8907E+19	2.2567E-02	27	1.6521E+19	5.9014E+19	2.6844E-02	147	1.3275E+19	4.6033E+19	2.1273E-02
8	1.4905E+19	5.1849E+19	2.3919E-02	28	1.7442E+19	6.2294E+19	2.8338E-02	148	1.4011E+19	4.8584E+19	2.2448E-02
9	1.5724E+19	5.4649E+19	2.5219E-02	29	1.8318E+19	6.5355E+19	2.9745E-02	149	1.4717E+19	5.0984E+19	2.3565E-02
10	1.6153E+19	5.6064E+19	2.5888E-02	30	1.8760E+19	6.6839E+19	3.0442E-02	150	1.5079E+19	5.2166E+19	2.4127E-02
11	1.6390E+19	5.6773E+19	2.6242E-02	31	1.8982E+19	6.7495E+19	3.0772E-02	151	1.5270E+19	5.2717E+19	2.4407E-02
12	1.6432E+19	5.6768E+19	2.6275E-02	32	1.8982E+19	6.7318E+19	3.0731E-02	152	1.5285E+19	5.2632E+19	2.4401E-02
13	1.6203E+19	5.5731E+19	2.5854E-02	33	1.8659E+19	6.5894E+19	3.0146E-02	153	1.5054E+19	5.1608E+19	2.3983E-02
14	1.5803E+19	5.4117E+19	2.5164E-02	34	1.8156E+19	6.3846E+19	2.9272E-02	154	1.4676E+19	5.0092E+19	2.3334E-02
15	1.5215E+19	5.1822E+19	2.4167E-02	35	1.7440E+19	6.1009E+19	2.8045E-02	155	1.4131E+19	4.7971E+19	2.2413E-02
16	1.4445E+19	4.8874E+19	2.2875E-02	36	1.6518E+19	5.7418E+19	2.6481E-02	156	1.3424E+19	4.5270E+19	2.1229E-02
17	1.3221E+19	4.4249E+19	2.0833E-02	37	1.5070E+19	5.1839E+19	2.4036E-02	157	1.2307E+19	4.1057E+19	1.9369E-02
18	1.2077E+19	3.9969E+19	1.8934E-02	38	1.3727E+19	4.6714E+19	2.1780E-02	158	1.1266E+19	3.7172E+19	1.7646E-02
19	1.0789E+19	3.5187E+19	1.6805E-02	39	1.2223E+19	4.1011E+19	1.9262E-02	159	1.0097E+19	3.2838E+19	1.5716E-02
20	9.3728E+18	2.9962E+19	1.4470E-02	40	1.0575E+19	3.4801E+19	1.512E-02	160	8.8122E+18	2.8107E+19	1.3603E-02

Table 4. Exposure Parameters for Capsule B Charpy Specimens

SPECIMEN POSITION NO.	FLUENCE >1 MeV	FLUENCE >.1 MeV	DPA	SPECIMEN POSITION NO.	FLUENCE >1 MeV	FLUENCE >.1 MeV	OPA
41	5.5840E+18	1.9859E+19	9.1671E-03	42	5.4686E+18	1.9381E+19	8.9778E-03
43	6.2505E+18	2.2376E+19	1.0289E-02	44	6.1213E+18	2.1838E+19	1.0076E-02
45	6.9059E+18	2.4851E+19	1.1392E-02	46	6.7631E+18	2.4253E+19	1.1157E-02
47	7.5491E+18	2.7280E+19	1.2474E-02	48	7.3931E+18	2.6623E+19	1.2217E-02
49	8.1790E+18	2.9657E+19	1.3534E-02	50	8.0100E+18	2.8944E+19	1.3254E-02
51	8.7794E+18	3.1980E+19	1.4569E-02	52	8.6127E+18	3.1210E+19	1.4258E-02
53	9.3945E+18	3.4242E+19	1.5578E-02	54	9.2002E+18	3.3419E+19	1.5256E-02
55	9.9778E+18	3.6441E+19	1.6559E-02	56	9.7715E+18	3.5564E+19	1.6217E-02
57	1.05044E+19	3.8571E+19	1.7509E-02	58	1.0226E+19	3.7644E+19	1.7148E-02
59	1.1091E+19	4.0630E+19	1.8428E-02	60	1.0861E+19	3.9652E+19	1.8048E-02
61	1.2049E+19	4.4228E+19	2.0036E-02	62	1.1799E+19	4.3164E+19	1.9622E-02
63	1.2538E+19	4.6060E+19	2.0855E-02	64	1.2278E+19	4.4952E+19	2.0425E-02
65	1.3004E+19	4.7807E+19	2.1637E-02	66	1.2736E+19	4.6657E+19	2.1191E-02
67	1.3448E+19	4.9464E+19	2.2380E-02	68	1.3170E+19	4.8274E+19	2.1918E-02
69	1.3868E+19	5.1029E+19	2.3082E-02	70	1.3582E+19	4.9802E+19	2.2606E-02
71	1.4264E+19	5.2499E+19	2.3743E-02	72	1.3969E+19	5.1236E+19	2.3252E-02
73	1.4635E+19	5.3870E+19	2.4360E-02	74	1.4332E+19	5.2574E+19	2.3857E-02
75	1.4979E+19	5.5141E+19	2.4933E-02	76	1.4670E+19	5.3815E+19	2.4418E-02
77	1.5297E+19	5.6309E+19	2.5461E-02	78	1.4981E+19	5.4954E+19	2.4935E-02
79	1.5589E+19	5.7371E+19	2.5943E-02	80	1.5266E+19	5.5991E+19	2.5407E-02
81	1.6057E+19	5.9061E+19	2.6714E-02	82	1.5725E+19	5.7641E+19	2.6162E-02
83	1.6268E+19	5.9813E+19	2.7059E-02	84	1.5932E+19	5.8374E+19	2.6500E-02
85	1.6451E+19	6.0452E+19	2.7355E-02	86	1.6110E+19	5.8998E+19	2.6791E-02
87	1.6604E+19	6.0979E+19	2.7602E-02	88	1.6261E+19	5.9512E+19	2.7032E-02
89	1.6728E+19	6.1392E+19	2.7799E-02	90	1.6382E+19	5.9915E+19	2.7225E-02
91	1.6823E+19	6.1690E+19	2.7945E-02	92	1.6475E+19	6.0206E+19	2.7368E-02
93	1.6888E+19	6.1873E+19	2.8041E-02	94	1.6539E+19	6.0384E+19	2.7462E-02
95	1.6923E+19	6.1940E+19	2.8086E-02	96	1.6573E+19	6.0450E+19	2.7506E-02
97	1.6928E+19	6.1891E+19	2.8080E-02	98	1.6578E+19	6.0402E+19	2.7501E-02
99	1.6904E+19	6.1727E+19	2.8024E-02	100	1.6554E+19	6.0242E+19	2.7445E-02
101	1.6780E+19	6.1121E+19	2.7785E-02	102	1.6433E+19	5.9851E+19	2.7212E-02
103	1.6671E+19	6.0630E+19	2.7585E-02	104	1.6326E+19	5.9171E+19	2.7015E-02
105	1.6532E+19	6.0026E+19	2.7334E-02	106	1.6190E+19	5.8582E+19	2.6770E-02
107	1.6364E+19	5.9309E+19	2.7034E-02	108	1.6026E+19	5.7883E+19	2.6476E-02
109	1.6168E+19	5.8482E+19	2.6685E-02	110	1.5833E+19	5.7075E+19	2.6.34E-02
111	1.5942E+19	5.7545E+19	2.6287E-02	112	1.5613E+19	5.6161E+19	2.5744E-02
113	1.5689E+19	5.6501E+19	2.5842E-02	114	1.5365E+19	5.5142E+19	2.5308E-02
115	1.5408E+19	5.5352E+19	2.5350E-02	116	1.5089E+19	5.4020E+19	2.4826E-02
117	1.5099E+19	5.4099E+19	2.4812E-02	118	1.4787E+19	5.2797E+19	2.4300E-02
119	1.4764E+19	5.2745E+19	2.4229E-02	120	1.4459E+19	5.1476E+19	2.3729E-02
121	1.4078E+19	4.9988E+19	2.3039E-02	122	1.3787E+19	4.7875E+19	2.2563E-02
123	1.3671E+19	4.8361E+19	2.2334E-02	124	1.3388E+19	4.7197E+19	2.1873E-02
125	1.3239E+19	4.6643E+19	2.1588E-02	126	1.2965E+19	4.5521E+19	2.1143E-02
127	1.2784E+19	4.4838E+19	2.0804E-02	128	1.2520E+19	4.3759E+19	2.0374E-02
129	1.2307E+19	4.2949E+19	1.9982E-02	130	1.2052E+19	4.1916E+19	1.9569E-02
131	1.1807E+19	4.0980E+19	1.9124E-02	132	1.1563E+19	3.9995E+19	1.8729E-02
133	1.1282E+19	3.8935E+19	1.8231E-02	134	1.1054E+19	3.7998E+19	1.7854E-02
135	1.0747E+19	3.6817E+19	1.7305E-02	136	1.0525E+19	3.5931E+19	1.6947E-02
137	1.0188E+19	3.4630E+19	1.6347E-02	138	9.9775E+18	3.3797E+19	1.6010E-02
139	9.6111E+18	3.2378E+19	1.5360E-02	140	9.4124E+18	3.1599E+19	1.5043E-02

values. Preliminary results of the uncertainty analysis can be found in Ref. 7.

3. Conclusions and Recommendations

The results achieved in the analysis of the fourth HSST experiments indicate that for a reasonable amount of time and funds, a considerable improvement in accuracy is attainable. The following steps are suggested for the analysis of exposure parameters from test reactor experiments.

1. ^{237}Np (n,f) and ^{238}U (n,f) sensors with gadolinium or CdO covers be included in the dosimetry multiple foil sets at select locations.
2. Single wire sensors should be placed at as many locations as feasible to obtain a complete and accurate spatial distribution of fast fluxes.
3. Good calculations for the source term in the core and the transport of neutrons from the core to the experiment.
4. A cross section data base for the reaction rate cross section and their variances and covariances.
5. A least square adjustment method to combine the reaction rate data, the calculated fluxes, and cross section values and arrive at an overall uncertainty for the exposure parameter of interest.

REFERENCES

1. F. W. Stallmann, "LSL, A Logarithmic Least Squares Adjustment Method," Proceedings of the Fourth ASTM-EURATOM Symposium on Reactor Dosimetry (to be published).
2. F. J. Witt, "The USAEC Heavy Section Steel Technology Program: Objectives and Status," Nucl. Eng. Des. 20(1), pp. 169-180 (June 1972).
3. G. D. Whitman and R. H. Bryan, Heavy-Section Steel Technology Program Quarterly Progress Report for Jan.-Mar. 1981, NUREG/CR-2141/VI, ORNL/TM-7822 (June 1981).
4. F. W. Stallmann and F. B. K. Kam, Neutron Spectral Characterization of the Second Nuclear Regulatory Commission Heavy Section Steel Technology 4T-CT Irradiation Experiment, ORNL/NUREG/TM-285 (Dec. 1978).
5. F. W. Stallmann and F. B. K. Kam, "Neutron Spectral Characterization of the NRC-HSST Experiments," Proceedings of the Third ASTM-EURATOM Symposium on Reactor Dosimetry, EUR 5667, Vol. 1, pp. 198-207 (1980).

6. W. N. McElroy (Editor), LWR Pressure Vessel Surveillance Dosimetry Improvement Program: PCA Experiments and Blind Test, NUREG/CR-1861 (July 1981).
7. F. W. Stallmann, "Uncertainties in the Estimation of Radiation Damage Parameters," Proceedings of the Fourth ASTM-EURATOM Symposium on Reactor Dosimetry (to be published).