

9-5-2013

Design Optimization of an Above-Knee Prosthesis with Energy Regeneration


Taylor Barto
Cleveland State University

Holly Warner
Cleveland State University

Rick Rarick
Cleveland State University

Dan Simon
Cleveland State University, d.j.simon@csuohio.edu

Follow this and additional works at: https://engagedscholarship.csuohio.edu/u_poster_2013

 Part of the [Biomechanics and Biotransport Commons](#)
How does access to this work benefit you? Let us know!

Recommended Citation

Barto, Taylor; Warner, Holly; Rarick, Rick; and Simon, Dan, "Design Optimization of an Above-Knee Prosthesis with Energy Regeneration" (2013). *Undergraduate Research Posters 2013*. 10.
https://engagedscholarship.csuohio.edu/u_poster_2013/10

This Article is brought to you for free and open access by the Undergraduate Research Posters at EngagedScholarship@CSU. It has been accepted for inclusion in Undergraduate Research Posters 2013 by an authorized administrator of EngagedScholarship@CSU. For more information, please contact library.es@csuohio.edu.



This digital edition was prepared by MSL Academic Endeavors, the imprint of the Michael Schwartz Library at Cleveland State University.

Design Optimization of an Above-Knee Prosthesis with Energy Regeneration

Fenn College of Engineering

Department of Electrical and Computer Engineering

Student Researchers: Taylor Barto; Holly Warner; Rick Rarick

Faculty Advisor: Dan Simon

Abstract

Above-knee amputees who use a prosthetic leg typically have to compensate for its shortcomings with unnatural hip motions. This compensation eventually leads to adverse health issues such as arthritis. We propose an active prosthesis to improve performance. The motor in our prosthetic knee allows the patient to move his hip normally, thus reducing the possibility of ancillary health issues. To improve the efficiency of the prosthesis, we use the braking phase of the prosthesis to regenerate energy. By storing energy in a supercapacitor during braking, the prosthesis lasts longer between each charge than it would without regenerative braking. We are considering two knee motor designs—a gear drive and a ball screw drive. Both designs appear to have the potential for regeneration. Several parameters characterize the prosthesis design. We use biogeography-based optimization (BBO) to determine these parameters. We are currently optimizing the prosthesis design to achieve accurate tracking of the knee angle. Future optimization criteria will include efficient energy use and generation.