

9-6-2012

Microcontroller Based ECG Amplifier

Berney Montavon

Cleveland State University, b.montavon@csuohio.edu

Follow this and additional works at: https://engagedscholarship.csuohio.edu/u_poster_2012

 Part of the [Biomedical Commons](#)

How does access to this work benefit you? Let us know!

Recommended Citation

Montavon, Berney, "Microcontroller Based ECG Amplifier" (2012). *Undergraduate Research Posters 2012*. 29.
https://engagedscholarship.csuohio.edu/u_poster_2012/29

This Book is brought to you for free and open access by the Undergraduate Research Posters at EngagedScholarship@CSU. It has been accepted for inclusion in Undergraduate Research Posters 2012 by an authorized administrator of EngagedScholarship@CSU. For more information, please contact library.es@csuohio.edu.



This digital edition was prepared by MSL Academic Endeavors, the imprint of the Michael Schwartz Library at Cleveland State University.

Microcontroller Based ECG Amplifier

Fenn College of Engineering

Department of Electrical and Computer Engineering

Student Researcher: Berney Montavon

Faculty Advisor: Dan Simon

Abstract

A microprocessor that reads ECG data and detects heart arrhythmias can help keep people safe while working in high-risk environments. Mining professionals, firefighters, and athletes have strenuous jobs that would benefit from non-invasive ECG monitoring. We use a Microchip Peripheral Interface Controller (PIC) for serial peripheral interface (SPI™) communication with a Texas Instruments ADS1298 development board. The ADS1298 is an 8-channel analog-to-digital converter (ADC) that is configured to record ECG data at a rate of 250 samples per second. The PIC is a master to the ADS1298 slave and is responsible for writing registers, receiving ECG data, and transmitting the data to a PC through an RS-232 serial connection. Our ADS1298 is configured to receive four channels of ECG data and up to four channels of accelerometer data, which are used to measure the noise due to acceleration that is embedded within the ECG signal.