

# CO<sub>2</sub>, CH<sub>4</sub>, and CO with CRDS technique at the Izaña Global GAW station: instrumental tests, developments and first measurement results

**Angel J. Gómez-Peláez**, Ramón Ramos,  
Vanessa Gómez-Trueba, Emilio Cuevas,  
Enrique Reyes  
Izaña Atmospheric Research Center,  
Meteorological State Agency of Spain (AEMET)

GGMT-2017, 27-31 August 2017, Dübendorf, Switzerland


## Acquisition of a CRDS for CO<sub>2</sub>, CH<sub>4</sub> and CO



- At the end of 2015, a **CO<sub>2</sub>/CH<sub>4</sub>/CO CRDS** was installed at the **Izaña Global GAW station (Tenerife, Spain)** to improve the Izaña GHG GAW measurement programme.
- The acquisition of the instrument was largely financed by European FEDER funds through a Spanish R+D infrastructure project we got: “Equipamiento para la Monitorización e Investigación en la estación Global VAG (Vigilancia Atmosférica Global) de Izaña (Tenerife) de componentes atmosféricos que provocan y modulan el cambio climático” AEDM15-BE-3319.

European Union

European Regional  
Development Fund  
“A way to build Europe”



## Initial/Acceptance tests performed to the CRDS

November 2015



- **Precision test** (ICOS): measuring a gas cylinder (filled with dry natural air) over **25 hours**; first hour rejected (stabilization time). **Good results:**

	Raw data average length	CO <sub>2</sub>	CH <sub>4</sub>	CO
Std. Dev. (1-sigma)	1 minute	0.013 ppm	0.19 ppb	0.87 ppb
	60 minutes	0.009 ppm	0.14 ppb	0.16 ppb

- **Repeatability test** (ICOS): Measuring alternately a gas cylinder (filled with dry natural air) during 30 minutes and ambient air (not dried) during 270 minutes over **72 hours**. Statistics based on the last 10 minute average data of each cylinder injection.

**We use 2 cylinders** (each one measured every 5 hours). **Good results:**

	CO <sub>2</sub>	CH <sub>4</sub>	CO
Std. Dev. (1-sigma) 10' average raw data	0.016 ppm	0.23 ppb	0.23 ppb
	0.016 ppm	0.23 ppb	0.35 ppb

- **Ambient pressure sensitivity of the measurements during the 72-h test:**  
**CO:** 0.4 ppb/10mb; **CO<sub>2</sub>:** 0.038 ppm/10mb; **CH<sub>4</sub>:** 0.47 ppb/10mb

## Initial/Acceptance tests performed to the CRDS

November 2015



GOBIERNO  
DE ESPAÑA

MINISTERIO  
DE AGRICULTURA Y PESCA,  
ALIMENTACIÓN Y MEDIO AMBIENTE



- **Inlet pressure sensitivity** when measuring a gas cylinder filled with dry natural air:

**CO:** 0.1 ppb/6 psi;      **CO<sub>2</sub>:** 0.04 ppm/6 psi;      **CH<sub>4</sub>:** 0.16 ppb/6 psi

The sensitivity for CO<sub>2</sub> is quite large.

- **Fitting response curves when calibrating with 4 WMO tertiary standards:**

**Linear fitting for CH<sub>4</sub>.** RMS residual = 0.143 ppb. Excellent.

**Linear fitting for CO.** RMS residual = 0.067 ppb. Excellent.

**Linear fitting for CO<sub>2</sub>.** RMS residual = 0.0395 ppb. Worse than our NDIR-based system with quadratic fitting.

**Quadratic fitting for CO<sub>2</sub>.** RMS residual = 0.0284 ppm. Still slightly worse than our NDIR-based system.

**Quadratic fitting for CO<sub>2</sub> correcting approximately from outletvalve aperture** (related with the inlet pressure). **RMS residual = 0.0219 ppm.**

Good. There were small differences in the inlet pressure for air of the different cylinders (we have improved our skills getting smaller differences between them since then).



## Calibrations and Response Functions



- **Calibration scheme:** 4 standards + 2 target gases. 30 minutes per tank every cycle.
  - From Dec. 2015 till Aug. 2016: **5 cycles** per calib. A calib. **every 3 weeks**.
  - From Sept. 2016 till present : **2 cycles** per calib. A calib. **every month**. Enough.

**The regulators of the standards remain closed between calibrations.**

For CO<sub>2</sub> and CH<sub>4</sub>, the **last 10 mins** of each injections are used. For CO, the **last 20 mins**. Stabilization times more than enough (numeric details not presented here).

I have developed a **Fortran 90 code** for processing the calibrations. We **use raw values** (not dry values).

- **Long-term drift of the CRDS.** Raw response drifts for a **virtual tank** with 400 ppm of CO<sub>2</sub>, 1850 ppb of CH<sub>4</sub> and 100 ppb of CO:

CO <sub>2</sub>	CH <sub>4</sub>	CO
0.104 ppm/year	2.22 ppb/year	0.544 ppb/year

- Using the empirical sensitivities (partial derivatives) provided in Sect. 3.3.6 of Yver Kwok et al. (2015), we have determined that **our CRDS has a long-term drift of 0.152°C/Torr and 0.446 Torr/year in the cavity sensors.**

## Calibrations and Response Functions



- **Raw signal used:** raw CO<sub>2</sub> (not dry), raw CH<sub>4</sub> (not dry), and since the raw CO value is not provided by the CRDS (but only the dry one), we compute, using peak84\_raw, peak\_14 and b\_h2o\_pct, **an uncalibrated raw CO value not dry** (indeed, **peak84\_spec\_wet**) corrected only from CO<sub>2</sub> and H<sub>2</sub>O baseline interference (Rella, 2016, Private communication).

- **Response functions used:**


**For CH<sub>4</sub> and CO, a linear function:**

$C_{\text{raw}} = a + b * C$ , where C is the real dry mole fraction (the gas standards are dry), and C\_raw the raw signal.

**For CO<sub>2</sub>, a quadratic function** with raw signal slightly corrected in outletvalve aperture (OV):

$$CO2_{\text{raw}} + (OV-26468.15)*0.04/7700 = a + b * CO2 + c * CO2^2$$



# Calibration and Response Functions





## Water Vapour Correction: water droplet method


**Crushed** (to increase the Surface/volumen ratio) **Silica Gel balls soaked with deionized water**






## Water Vapour Correction: water droplet method

Our correction function is similar to that of Chen et al. (2010). Indeed, for **CO<sub>2</sub>** we use exactly their function.



We use means of 100 instantaneous values for all the variables


**CO signal is very noisy**, but the long duration of the experiment allows us using **4000-instantaneous-value running means** without compromising the accuracy of the data:



We use a cubic correction function for CO

## Ambient Measurements

- Ambient air/gas standard **plumbing configuration** operative since 28 Nov 2016.
- Before that date, there were no “Dedicated inlet”, **no drying** (no cooled flasks), no solenoid nor needle valves, and ambient air entered through the MPV.



- **Operative ambient air measurements started on 27 Nov 2015.**
- **Target gas measurements started** on 18 Dec 2015 with a **7-hour cycle** (to monitor better the behaviour of the CRDS), which became a **21-hour cycle after 24 June 2016.**
- With the new plumbing configuration, **ambient air is alternatively sampled from the two inlet lines** within the 21-hour cycle (5 hours from general inlet and 15 hours from dedicated inlet).
- **Cooler bath temperature: -30 °C**, therefore, **no complete drying.**

## Ambient Air Measurement Processing



- The **processing software** has been **developed in Fortran 90** (there are some features and refinements that still need to be implemented).
- The **first step is computing raw-data 30-second means** using the DataLog\_User files (no the synchronized ones), taking into account the “species” field (i.e., which fields have been updated in each file line; there are 1.7 lines/second).
- **For not discarding a 30-second mean:**
  - **85%** of the expected **data** needs to be **present**.
  - All the instantaneous data need to have the **same MPV position and solenoid valve configuration**.
  - The **mean values** of the following variables **need to be within** the indicated ranges: cavity pressure, **140 +- 0.035 Torr**, cavity temperature, **45 +- 0.02 °C**, and **outletvalve 20000-40000**.
  - It needs to exit a calibration before and after the ambient mean considered, separated in time less than 180 days between them (as in ICOS).

## Ambient Air Measurement Processing




- **Processing scheme:**
- Computation of the 30-second mean for the derived variable: **peak84\_spec\_wet**, and multiplication of this variable and CH4\_raw by 1000.
- Application of the **water vapour correction** (dilution and pressure broadening effects).
- Application of the **calibration curves** interpolated linearly in time.
- This scheme is also used to **assign mole fractions to the target and calibration gas injections**, and it is **checked that the water vapour correction for them is smaller than: 0.01 ppm for CO2 and 0.1 ppb for CH4 and CO.**
- **Discarding due to stabilization time:** 10 minutes for ambient measurements, and 20 minutes for target and calibration gas injections.



## Target Gas Injections


- 30-second-mean assignments to target gases.
- Good results.
- Note that 30-second is a too short time for CO, and it is necessary to consider longer averages for decreasing noise.



Tank/months	CO2 (ppm)	Std.Dev.	CH4 (ppb)	Std.Dev.	CO (ppb)	Std.Dev.
CA07080/ 7m	381.96	<b>0.020</b>	1825.43	<b>0.32</b>	148.60	<b>0.97</b>
CA05038/ 7m	368.85	<b>0.020</b>	1777.04	<b>0.33</b>	93.56	<b>0.99</b>
CA06812/ 13m	372.48	<b>0.020</b>	1784.80	<b>0.27</b>	142.04	<b>1.01</b>
CA05034/ 13m	363.71	<b>0.020</b>	1775.89	<b>0.27</b>	139.11	<b>0.98</b>

## Comparison with other Izaña continuous measurements

- Comparison of **daily nighttime means** with other Izaña continuous measurements.
- **The data sets are still not final** and the origin of the discrepancies in some periods needs to be investigated.
- It seems **our set of CRDS laboratory standards** (WMO tertiaries) **might be drifting up significantly for CO** (increase of fit RMS, drowndrift in targets, increase of the difference with RGA3). **Maybe, there is also a small contribution from Primaries of the X2014A scale drifting less than initially computed by CCL.**



## Some additional novelties in the Izaña GHG instrumentation



- **Improvements introduced in the dedicated inlet lines:** 1) **back-pressure regulators** for the vents located downstream the pumps and rotameters for those vents; 2) **needle valves** in low flow vents installed downstream the cryotrap; 3) glass flask cryotrap with **Ultra-Torr connections**; 4) hermetic plugs for unused ports of the rotary Valco valves.
- **Preparation of two CO<sub>2</sub> laboratory standards of 418.7 ppm for the Izaña NDIRs** Li7000 and Li6252 and calibration of them against our CRDS laboratory standards.
- **Reprocessing of the Izaña time series of CH<sub>4</sub> and CO in the scales X2004A and X2014A**, respectively, **taking into account also the drift of the 5 laboratory standards used in the Izaña RGA-3.**



**Thanks for your attention!**



**Teide Peak from Izaña Observatory  
(Picture: Ángel J. Gómez-Peláez, Christmas 2016)**