

Article

Alpha Radiation-Induced Luminescence by Am-241 in Aqueous Nitric Acid Solution

Thomas Kerst ^{1,2,*}, Rikard Malmbeck ³, Nidhu lal Banik ³ and Juha Toivonen ¹

¹ Photonics Laboratory, Physics Unit, Tampere University, P.O. Box 692, 33101 Tampere, Finland; juha.toivonen@tuni.fi

² Helsinki Institute of Physics, Helsinki University, P.O. Box 64, 00014 Helsinki, Finland

³ European Commission, Joint Research Centre (JRC), Directorate G–Nuclear Safety and Security, Advanced Nuclear Knowledge, 76125 Karlsruhe, Germany; rikard.malmbeck@ec.europa.eu (R.M.); nidhu.banik@ec.europa.eu (N.B.)

* Correspondence: thomas.kerst@tuni.fi; Tel.: +358-50-300-5968

Received: 13 February 2019; Accepted: 28 March 2019; Published: 2 April 2019

Abstract: When exposed to air, alpha particles cause the production of light by exciting the molecules surrounding them. This light, the radioluminescence, is indicative of the presence of alpha radiation, thus allowing for the optical sensing of alpha radiation from distances larger than the few centimeters an alpha particle can travel in air. While the mechanics of radioluminescence in air and other gas compositions is relatively well understood, the same cannot be said about the radioluminescence properties of liquids. Better understanding of the radioluminescence properties of liquids is essential to design methods for the detection of radioactively contaminated liquids by optical means. In this article, we provide radioluminescence images of Am-241 dissolved in aqueous nitric acid (HNO₃) solution and present the recorded radioluminescence spectrum with a maximum between 350 nm and 400 nm, and a steep decrease at the short wavelength side of the maximum. The shape of the spectrum resembles a luminescence process rather than Cerenkov light, bremsstrahlung, or other mechanisms with broadband emission. We show that the amount of light produced is about 150 times smaller compared to that of the same amount of Am-241 in air. The light production in the liquid is evenly distributed throughout the sample volume with a slight increase on the surface of the liquid. The radioluminescence intensity is shown to scale linearly with the Am-241 concentration and not be affected by the HNO₃ concentration.

Keywords: alpha radiation; radioluminescence; liquid phase luminescence; americium

1. Introduction

Radioluminescence describes the spontaneous emission of light as a consequence of interaction of luminescent material with ionizing radiation. The ionizing particle typically originates from, but is not limited to, a form of radioactive decay. In the process of creating radioluminescence, the ionizing particle or an induced secondary electron collides with a luminescent material, resulting in the elevation of an orbital electron [1]. The excited electron then has a chance to radiatively decay, thereby emitting a photon of light. Photons created this way are said to be radioluminescence, since the presence of ionizing radiation ultimately lead to its production. Thus, the presence of radioluminescence is to be taken as an indicator for the presence of ionizing radiation itself.

Radioluminescence induced by alpha particles allows for their remote detection by optical means. In air, alpha particles come to a halt after having travelled for about 4 cm, losing almost all their kinetic energy in the process [2]. In contrast, beta and gamma radiation travel a few meters and tens of meters in air, respectively [3], making alpha radiation a comparably short ranged type of nuclear radiation. This makes it rather easy to avoid exposure to a known alpha source by keeping distance, while at

the same time making it difficult to detect unknown alpha sources with methods that rely on direct interaction with the particles. Alpha induced radioluminescence is not limited to a travel distance of a few cm, thus making it possible to remotely detect alpha radiation by collecting radioluminescent light. [4].

How little alpha radiation can reliably be detected by collecting radioluminescence depends on the conditions surrounding the source. To understand the reasons for this, it is necessary to recognize that even high-activity alpha sources create amounts of photons that can be considered almost negligible in comparison to the number of photons a typical ambient light source, such as the sunlight on earth or a fluorescent door sign, emits. A single alpha particle being stopped in air leads to the production of about 100 photons [5,6], each of which is isotropically radiated away from the location of the alpha source. Thus, photon counting devices must be employed to make radioluminescence detectable. In a recent study, Sand et al. [7] thoroughly investigated the question of how small an alpha contamination can one detect by optical means alone, given that one can use state-of-the art equipment to detect single UV photons and the site of contamination is exposed to normal air. Under ideal conditions, e.g., the only present light source being radioluminescence, they were able to detect sources with an alpha activity of 4 kBq or surface contamination with an activity of 300 Bq cm^{-2} , respectively, by using an UV sensitive low-noise PMT at a distance of about 1 m with a measurement time of about 10 s.

Once background lighting must be taken into account, a straightforward approach of collecting light is no longer feasible. The spectral patterns of both the air radioluminescence, being mostly made up of N_2 emissions [1], and of the background lighting must be considered. Furthermore, the optical design must be optimized to mitigate the effects of the typically high ambient light levels [7–11]. Another way of overcoming the problem of background lighting is by only collecting photons with wavelengths that are in the UVC wavelength range at wavelengths shorter than 280 nm. Daylight, one of the most common source background lighting containing UV parts, does not extend into the UVC [12]. Thus, choosing equipment that is both sensitive to those wavelengths and able to detect single photons can be used for remote detection of alpha radiation. Ivanov et al. [13,14] are one of the first reporting on such an approach using cameras, whereas Crompton et al. [15] pioneered the field with using a detector that uses the photoelectric effect and gas multiplication to detect individual UVC photons. However, restricting the detection to the UVC limits the sensitivity of the optical detection as UVC range contains only a small fraction of the radioluminescence [16]. One way to overcome this impediment is by relying, when possible, on an artificial atmosphere that allows for enhanced yield of radioluminescence [17–19].

The ability of liquids to exhibit luminescence upon irradiation with ionizing particles has not been as thoroughly studied as it has been done for gases. Even though there have been a few studies reaching back into the 60s [20], renewed interest into the topic has been partly sparked by the observations of Yamamoto and colleagues [21]. They showed that faint radioluminescence is produced by placing an Am-241 source in water. In a later study, it was estimated using long-pass filters that the spectrum is a very broad one with increasing intensity towards shorter wavelengths [22]. However, a fully resolved spectrum was not presented.

In this paper, we present radioluminescence images originating from a liquid with Am-241 dissolved in an aqueous solution of nitric acid (HNO_3). The imaging reveals equal distribution of light production throughout the sample volume with a slight increase at the surface of the liquid interfacing with air. We present a fully resolved radioluminescence spectrum in the spectral range between 280 nm and 550 nm. Lastly, we investigate the intensity of the radioluminescence as a function of the amount of Am-241 and the HNO_3 concentration, and discuss about the possible mechanisms producing the observed radioluminescence.

In section 2 we describe our experimental arrangements and give detailed information on how data was obtained. We also given detailed information about the image processing involved and how the reader can reproduce the algorithms and methods used in the digital post-processing. The results

are presented and discussed in section 3. Section 4 concludes this work and summarizes the main findings.

2. Materials and Methods

All experiments have been carried out in the laboratories and premises of the Joint Research Centre (JRC) in Karlsruhe, Germany, during two measurement campaigns. For safety reasons, all radioactive material used was kept and handled in a plexiglass glovebox while most of the other equipment remained outside. The only modifications to the glovebox were a replacement of one glove with a quartz window (Sico Technology GmbH, SQ1, Austria) and the insertion of another quartz glass window into the plexiglass ceiling of the box. The chosen quartz glass has a flat transmission spectrum down to 180 nm wavelength and is thus well suited for this experiment [23].

The box has been shielded from ambient light with black rubberized fabric (Thorlabs, BK5, Newton, NJ, USA). A very thorough shielding with multiple layers was required to block all ambient light, especially given that the laboratory could not be made totally dark. We made sure that radioluminescence was the only significant source of light during the data acquisition by recording a control image without the alpha source prior to actual data recording with different radioactive samples.

The radioactive samples used during the measurements included a planchet with a thin layer of Pu-239 evaporated onto it (from now on referred to as: the solid sample) and several liquids containing Am-241 (from now on referred to as: the liquid samples). The surface activity of the solid sample was verified by alpha spectroscopy to be 5.6 MBq. We prepared the liquid samples by evaporating a pure Am-241 solution to dryness and then adding nitric acid (HNO₃) solution to it. Depending on the sample, we used 1 M, 3 M or 7 M, with the solvent having been deionized water. M is the molar concentration, i.e., mols per liter. We prepared each liquid sample to measure a volume of 2 mL. We denote the concentration of Am-241 dissolved in each sample by dividing the amount of μg of Am-241 with the number of grams of the liquid sample. Thus, the concentration is defined by mass, and throughout the paper we express it with the unit ppm. We prepared a total of 7 liquid samples divided on 2 distinct sets. The first set comprised a total of 3 different solutions each containing 150 ppm of Am-241, but with varying HNO₃ concentrations of 1 M, 3 M and 7 M. The second set comprised a total of 4 different solutions each with a HNO₃ concentration of 3 M, but with varying Am-241 concentrations being 50 ppm, 75 ppm, 150 ppm and 750 ppm. Using the specific activity of Am-241 [24], the liquid samples have activities of roughly 13 MBq, 19 MBq, 38 MBq and 190 MBq, respectively. When using a liquid sample in an experiment, we filled it into a standard quartz glass fluorescence cuvette that transmits light at the wavelength range of 200 nm - 2500 nm (12.5 mm length and width, 45 mm height, 10 mm light path length, Hellma analytics GmbH & Co. KG, type: 101-QS, material: QS).

2.1. Radioluminescence Imaging

We imaged the radioluminescence by placing a sample on a holder close to the quartz window and an EMCCD camera (Andor iXon3 897) on the other side of the window. Sample holder and camera objective were kept at a constant distance of about 30 cm throughout all experiments. The camera sensor was internally cooled down to -80°C to reduce thermal noise. We placed the sample and the holder inside a light-tight PVC tube and wrapped the camera in black fabric to further improve the ambient light shielding. The imaging lens was a UV-objective (Universe Kogaku, UV1228CM) with a focal length of 12 mm and a light collection efficiency of $f/2.8$. A schematic illustration of the arrangement is shown in Figure 1.




Figure 1. Schematic of the arrangement designed to image radioluminescence: A liquid or solid alpha active sample (green) resides on a holder placed inside black and light-tight PVC tube. The tube is placed around a glove port of box which is holding a quartz glass window instead of a glove. Light exiting the glove box through the window form an image on the sensor of an EMCCD camera. Multiple layers of black fabric wrapped around the camera and the glove port prevent stray light form entering the setup.

For each sample we took one background image in normal laboratory lighting conditions and then darkened the setup to obtain images of the radioluminescence. For each liquid sample we took 30 images with an exposure time of 60 s each, resulting in a total integration of 30 min. For the solid sample, a single exposure of 100 s length proofed sufficient to acquire image material with good enough signal-to-noise ratio. The individual images accumulated numerous hot pixels, e.g., pixels with maximum value, which likely occurs due to exposure of the sensor to gamma radiation predominantly from the 59.6 keV line of Am-241.

All image material has been post-processed by taking the pixel-wise median of the 30 consecutive images. The so-constructed median image is largely free of hot pixels all the while leaving the features and resolution of the image unaffected. The procedure is illustrated in Figure 2 with data from the experiments serving as example. For image processing we used python 3.6.3 as programming language and made use of the libraries OpenCV 3.3.1, matplotlib 2.1.0 and NumPy 1.13.3.




Figure 2. Demonstration of the post-processing using 30 images with exposure times of 60 s each. (a) A single capture, raw data. The cuvette and the liquid are clearly visible, hot pixels appear as red dots. (b) The pixel-wise median over a total of 30 images. The z-scale shows raw pixel values.

2.2. Spectral Measurement

The radioluminescence spectrum was measured by transporting light with a system of liquid light guides outside the glove box, where it was analyzed. A schematic illustration of the arrangement is presented in Figure 3. The sample was placed on a height-adjustable pedestal inside the glovebox. The liquid light guide (Lumatec GmbH, Series 300, 8mm core diameter) was installed next to the pedestal to collect the light emanating from the sample. The fiber was put as close as possible to the sample to eliminate the need for collimating optics. When the liquid sample was spectrally analyzed, the fiber was installed as close as possible to the sample holder to make sure that only radioluminescence from the bulk of the liquid reaches the detector. From there the light was guided to a 10 mm thick quartz window in the ceiling of the glove box. Another liquid light guide at the other side of the window then picked up the light and further transported it to a motorized monochromator (Horiba Scientific Inc., iHR 550), where it was spectrally dispersed. Light of only one selected wavelength leaves the monochromator at a time. The monochromator was operated with a UV sensitive grating (model 51050, 300 gr mm^{-1} , 250 nm blaze) and an entrance slit width of 2 mm, resulting in a spectral resolution of about 6 nm. The light passing the monochromator was measured with a low-noise photon counting PMT (Perkin Elmer GmbH, MP-1982, $< 1 \text{ cps DC}$) and read out via computer.




Figure 3. Schematic of the setup designed to measure the radioluminescence spectrum. A liquid or solid alpha active sample (green) resides on a holder inside holder in the glove box. A system of liquid light guides transports radioluminescent light to a monochromator, where it is spectrally separated and detected by a PMT.

A spectrum was obtained by counting the photons with the PMT at each individual wavelength and then arranging the so-measured counts according to their associated wavelength. The monochromator scanned from 280 nm to 550 nm in 3 nm steps, integrating for a total of 10 s at each step. The scan was repeated multiple times both to improve the signal-to-noise ratio by having more data points available and to compensate for drifts of the PMT. A drift of the PMT could not have been counteracted against if the measurement consisted of one long single scan.

During post-processing, the spectral data we accounted for effects that alter the measured spectrum. The cuvette, the light guide, and the PMT have a flat response. They can change the intensity of the light they respond to, but they cannot alter the shape of the spectrum of this light. The reflectance of the grating has been calibrated to give a flat spectral response. A 10 mm thick water column has a negligible absorption in the measurement wavelength range [25], thus it has no effect on the measured spectral shape.

3. Results and Discussion

3.1. Imaging

Radioluminescence of 750 ppm Am-241 which has been dissolved in 3 M HNO₃ was imaged with the EMCCD camera in darkness. In addition, we took a photograph image under normal lighting conditions for a reference. We repeated the procedure for the Pu-239 solid sample planchet, where we expected radioluminescence of air to occur above it. That served as an intensity calibration regarding to previously published results [26]. The recorded images are shown in Figure 4.




Figure 4. Images of radioluminescence and the locations where they occur. The gray scale images each show a single 0.1 s capture of the samples under normal lighting conditions. The colorful images next to them show the same scenes in darkness using much longer exposure times. Dotted lines throughout the images (c–e) frame the area of the images where the surface of the liquid with air is displayed. The z-scale used in the images is the same as the one used in Figure 2. (a) Image of the Pu-239 coated planchet under normal lighting conditions. (b) Radioluminescence image of the Pu-239 coated planchet. (c) Image of the sample holder and the cuvette which resides in it taken under normal lighting conditions. The camera is slightly tilted and reveals part of the sample holder behind the cuvette. The cuvette holds the transparent liquid, the camera tilt makes it possible to see the boundary of the liquid with the air. (d) Radioluminescence image of the cuvette containing 750 ppm of Am-241 in 3 M HNO₃ solution. (e) Overlay of the cuvette radioluminescence image (d) on the image of the cuvette (c).

The image of the solid planchet radioluminescence in Figure 4b shows radioluminescence emanating from the surface of the planchet, in a fashion similar to what has been found and described in greater detail by Sand et al. in an experiment under similar conditions [26]. The imaging of radioluminescence emanating from the liquid was performed with a total integration time of 1800 s,

whereas the imaging the solid sample was performed in 100 s. This is due to the much lower light yield of the liquid, even though the overall alpha activity is higher. We estimated the light yield per alpha particle in the liquid by integrating the pixel intensities of the background-corrected planchet image and dividing with activity and exposure time. A similar procedure has been used for to the central parts of liquid radioluminescence image, and the resulting light yield was further scaled to take into account the total liquid area in the cuvette. The energies of alpha particles emitted by Pu-239 are 5.157 MeV and by Am-241 are 5.486 MeV, respectively[27], thus the two samples are quite similar from an alpha radiation point of view. From this analysis, we find that the light yield of an alpha particle in the liquid is roughly 150 times lower than in air. A similar effect has been noticed by Yamamoto et al. in their research on liquid radioluminescence [21], where they observe a decreased light yield with a factor of about 100.

The pixel intensities at the liquid-air boundary in Figure 4d are higher compared to the pixel intensities further down the water column. The shape of this area is coincidental with the shape of the liquid-air boundary in Figure 4c, which can be even better seen in the overlay image in Figure 4e. The camera is slightly tilted towards the cuvette, which makes it possible to look on top of the interface of the liquid-air boundary. The increase in the intensity might be due to an optical effect or an increased light production at the surface. The optical refraction at the liquid-air boundary can redirect more radioluminescent light towards the camera. Also, it is known that alpha particles are emitted out from the liquid[28], and thus can create radioluminescence in air [5]. The radioluminescence yield in air was previously shown to be approximately 150 times more efficient, which can result in the increased intensity at the surface in Figure 4d.

It can also be seen there is an area of slightly increased pixel intensities at the lower end of the cuvette. The camera looks with a slightly tilted angle at the cuvette, making a part of the cuvette which is surrounded by the holder visible to the camera. It is those parts of the image that show increased pixel intensities. In the overlay image in Figure 4e this becomes even clearer. Thus, it is very likely that the increased intensity in the radioluminescence image in this particular area is not created by an increased radioluminescence production but rather by reflection of radioluminescence photons from the sample holder towards the camera.

Vertical and horizontal cross sections of the radioluminescence image of the cuvette are shown in Figure 5 for detailed analysis of the intensity distribution. In the horizontal cross section in Figure 5a it can be seen that the radioluminescence is evenly distributed throughout the liquid. We notice that the glass walls have no influence on the light production. In the vertical cross section in Figure 5b the differences in radioluminescence intensity as a function of height become clearly visible. The rightmost peak corresponds to the slightly increased intensity seen at the point where the cuvette is submerged in the holder. The maximum in the vertical cross section shows the increased light production at the surface of the liquid. Dotted lines indicate the position where the interface of the liquid-air boundary faces the camera. They correspond to the same dotted lines shown in Figure 4. Between these two peaks, the radioluminescence intensity is notably evenly distributed.

The local maxima to the left in Figure 5b correspond to the increased radioluminescence observed at the top end of the cuvette. This is very likely radioluminescence from dried liquid. The same cuvette was used to hold all liquid samples. Changing the sample made it necessary to empty and refill the cuvette. During refilling contact of the liquid with the walls was difficult to avoid, thus contaminating it. Even careful cleaning rarely removed all contamination, and it were the edges of the cuvette that proved most difficult to clean. It is this contamination that is seen in the radioluminescence image. Though the radioluminescence from contamination of the cuvette walls is somewhat of a nuisance for this particular work, it shows once again that radioluminescence imaging is a very efficient method to find alpha contamination in difficult to access areas.




Figure 5. Cross sections of parts of the radioluminescence image of 750 ppm Am-241 in 3 M HNO₃. (a) Horizontal cross section across the cuvette. The uncertainty displayed in gray is identical to the standard deviation of the pixel values in vertical direction. (b) Vertical cross section across the cuvette. The uncertainty displayed is identical to the standard deviation of the pixel values in the horizontal direction. Labels point to notable features and name the position where they occur.

3.2. Influence of Am Concentration and Acidity

A dependency of the radioluminescence intensity on alpha active Am-241 and nitric acid concentrations was studied, since it is not clear what exactly causes the light emission. The role of nitric acid was tested with samples, where the activity was held constant by having 150 ppm of Am-241 while the nitric acid concentration was varied. The effect of Am-241 was studied with samples, where the HNO₃ concentration was held constant at 3 M while the americium concentration was varied. In all the experiments, the post-processing of the radioluminescence images was done as outlined in section 2. The radioluminescence signal level was calculated as an average of pixels corresponding the liquid volume of the cuvette. Then this average was corrected for the background level of the image. The resulting data is presented in Figure 6. The radioluminescence shows little or no dependency on the nitric acid concentration. However, the data shows that the radioluminescence is linearly dependent on the Am-241 concentration. Thus, we can exclude nitric acid from potential luminescent species in the solutions. We can further reduce the number of possible sources of radioluminescence by comparing our results with experiments by using an ion beam as an ionizing source [22]. In that work, a water phantom was irradiated with a carbon-ion beam producing radioluminescence with properties similar to what we found. This leads us to conclude that Am-241 is unlikely linked to the luminescent properties in other ways than as a source of the ionizing radiation.




Figure 6. Dependence of the measured radioluminescence intensity on the nitric acid concentration and the americium concentration. The pixel values displayed as radioluminescence have been corrected for the image background. (a) Change of the radioluminescence with a varying nitric acid concentration. A constant has been fitted to the data. (b) Change of the radioluminescence with varying Am-241 concentration. A linear function has been fitted to the data.

3.3. Radioluminescence Spectrum

Radioluminescence properties of the both solid and liquid samples were further studied by measuring the spectra of the luminescence using a monochromator and a photomultiplier tube. The measurements of the solid and liquid samples were otherwise performed in very similar way, except the liquid sample was with an extra quartz glass wall of the cuvette in the optical path and different spectral scanning parameters were used. For the solid sample, only a single scan from 280 nm to 500 nm in 0.5 nm steps was used and the integration time at each step was set to 30 s. To record spectrally narrow nitrogen radioluminescence lines in air over the solid sample, we used the grating model 53020 with a groove density of 1800 gr mm^{-1} blazed at 250 nm yielding to 1 nm spectral resolution. These modifications allowed us to record the well-known radioluminescence spectrum of air, shown in Figure 7a, in less than four hours. For the liquid sample, the procedure followed the steps described in section 2. The spectral resolution for a liquid sample was 6 nm and the scanning was performed from 280 nm to 550 nm in 3 nm steps, integrating for 10 s at each step. The spectral scan was repeated overnight for 88 times to achieve reasonable signal-to-noise ratio for the weak liquid radioluminescence resulting total recording time of 22 hours. Figure 7b shows the resulted spectrum of the aqueous solution under ionizing radiation. The correct spectral shape of the well-known radioluminescence of air in Figure 7a verifies that the calibration of spectral response has been successful and that the spectral shape of the liquid radioluminescence in Figure 7b can be trusted. The only additional elements in the optical path are the quartz cuvette and the water column, which both have negligible absorption at the measured wavelength range from 280 nm to 550 nm.




Figure 7. The normalized radioluminescence spectra of (a) light created by the solid sample (black) contrasted with this of air (red, data from [19]) (b) 750 ppm Am-241 in 3 M HNO₃ solution.

The normalized spectrum of the light produced by the solid sample in Figure 7a shows the N₂ lines [16], which is commonly associated with radioluminescence in air. This is congruent with other works investigating air radioluminescence [1,7,29]. To demonstrate the validity of the measurement, the spectrum has been superimposed with the normalized spectrum of radioluminescence in air which has been measured at a higher spectral resolution [19]. Contrary to the radioluminescence of air, the spectrum of aqueous solution under ionizing radiation has not been fully resolved earlier apart from a rough estimation using several long-pass filters [22]. Thus, the spectrum of the radioluminescence of aqueous solution in Figure 7b is, to the best of our knowledge, reported here for the first time. The radioluminescence is weak and spectrally broad, which makes it difficult to record. The spectrum shows a broad maximum between 350 nm and 400 nm with a gradual decrease until 550 nm at the longer wavelength side. On the other hand, the spectrum exhibits a pronounced decrease at shorter wavelength side reaching down to the background level already at 300 nm. This steep decreasing slope at shorter wavelength side contrasts with the earlier study done with long-pass filters [22], where they estimated a continuous increase towards the shorter wavelengths resembling the spectral shape of the Cerenkov light. Cerenkov light has a spectrum that is proportional to λ^{-2} , where λ is the wavelength of the emitted light [30]. The data in Figure 7 (b) clearly shows this is not the case with the radioluminescence of aqueous solutions, as there is the steep decrease at short wavelength side of the spectrum. Furthermore, alpha particles are not capable of producing Cerenkov light directly or through secondary mechanisms [31] and therefore we conclude that the measured radioluminescence must have an origin other than Cerenkov radiation. The spectral response of the measurement system is compensated in the results and the air radioluminescence spectrum with the same setup demonstrates the correct spectral shape down to 300 nm, thus validating the spectral response of the measurements. The only optical difference between the measurement arrangements is the quartz cuvette, which has a flat spectral transmittance over the whole measurement range and therefore no effect on the results. The steep decrease at the short wavelength side of the spectrum makes it uncharacteristic for broadband emissions such as Cerenkov light and bremsstrahlung emission. We rather speculate that some of the radiolysis products [32] of water, such as H₂O₂, OH or H₂, might be responsible for the recorded luminescence of water under the ionizing radiation.

4. Conclusions

We studied the radioluminescence produced by alpha active Am-241 dissolved into an aqueous solution of nitric acid. We showed that the amount of light created is about 150 times smaller than what is produced with same alpha activity by the radioluminescence of air. The low amounts of light made it necessary to use photon counting devices and intensified cameras to record the data. We analyzed the spatial distribution of the radioluminescence showing a very evenly distributed light emission. Furthermore, we showed for the first time the aqueous radioluminescence spectrum in the wavelength range between 280 nm and 550 nm, where we found a broad spectrum between 330 nm and 500 nm. The demonstrated radioluminescence is potentially useful in applications where visible background light can be avoided by light shielding, such as in radioactive liquid monitoring in closed chambers. We concluded that Cerenkov light, bremsstrahlung or other mechanisms producing broadband emission cannot be the source of the observed radioluminescence due to steep decrease at the short wavelength side of the recorded spectrum, which leaves intrinsic processes of water and its radiolysis products as strong candidates for the luminescence.

Acknowledgments: The authors wish to thank Business Finland, Finland Distinguished Professors (FiDiPro), Project Novel Instrumentation for Nuclear Safety, Security and Safeguard (NINS3), for the funding that made this project possible. The authors also wish to acknowledge the Actinide User Laboratory (AUL) in Karlsruhe, for providing access to their laboratories, sources, and infrastructure. The authors wish to especially thank Adrian Nicholl for his constant support before, during, and after the experiments.

Author Contributions: T.K. participated in designing the experiments, was part of execution of the experiments and handled the optics referred to in this study. R.M. and N.B. participated in designing the experiment, were

part in execution of the experiments and handled all the chemistry and radioactive sources referred to in this study. R.M. and J.T. planned, guided, and oversaw the collaboration leading to the conception of the experiments.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Arqueros, F.; Blanco, F.; Rosado, J. Analysis of the fluorescence emission from atmospheric nitrogen by electron excitation, and its application to fluorescence telescopes. *New J. Phys.* **2009**, *11*, 065011. [[CrossRef](#)]
2. Brown, A.; Suit, H. The centenary of the discovery of the Bragg peak. *Radiother. Oncol.* **2004**, *73*, 265–268. [[CrossRef](#)]
3. Williams, W.S.C. Nuclear Instability. In *Nuclear and Particle Physics*; Oxford University Press: Oxford, UK, 1991; pp. 66–80.
4. Baschenko, S.M. Remote optical detection of alpha particle sources. *J. Radiol. Prot.* **2004**, *24*, 27–92. [[CrossRef](#)]
5. Sand, J.; Ihantola, S.; Peräjärvi, K.; Toivonen, H.; Toivonen, J. Radioluminescence yield of alpha particles in air. *New J. Phys.* **2014**, *16*, 053022. [[CrossRef](#)]
6. Colin, P.; Chukanov, A.; Grebenyuk, V.; Naumo, D.; Nédélec, P.; Nefedov, Y.; Onofre, A.; Porokhovoi, S.; Sabirov, B.; Tkatchev, L; et al. Measurement of air and nitrogen fluorescence light yields induced by electron beam for UHECR experiment. *Astropart. Phys.* **2007**, *27*, 317–325. [[CrossRef](#)]
7. Sand, J.; Nicholl, A.; Hrncek, E.; Toivonen, H.; Toivonen, J.; Peräjärvi, K. Stand-Off Radioluminescence Mapping of Alpha Emitters under Bright Lighting. *IEEE Trans. Nucl. Sci.* **2016**, *63*, 1777–1783. [[CrossRef](#)]
8. Pineau, J.F.; Imbard, G. Remote α source location device and method, US Patent 6281502B1, 2001.
9. Lamadie, F.; Delmas, F.; Mahe, C.; Gironès, P.; Goaller, C.L.; Coestes, J.R. Remote Alpha Imaging in Nuclear Installations: New Results and Prospects. *IEEE Trans. Nucl. Sci.* **2005**, *52*, 3035–3039. [[CrossRef](#)]
10. Chichester, D.L.; Watson, S.M. Multispectral UV-Visual Imaging as a Tool for Locating and Assessing Ionizing Radiation in Air. *IEEE Trans. Nucl. Sci.* **2011**, *58*, 2512–2518. [[CrossRef](#)]
11. Feener, J.S.; Charlton, W.S. Preliminary results of nuclear fluorescence imaging of alpha and beta emitting sources. In Proceedings of the 2013 3rd International Conference on Advancements in Nuclear Instrumentation, Measurement Methods and their Applications (ANIMMA), Marseille, France, 23–27 June 2013.
12. National Renewable Energy Laboratory: Direct and Global 37 Deg Tilt: ASTM G-173. Available online <https://www.astm.org/Standards/G173.html>. (accessed on 27 February 2019).
13. Ivanov, O.; Danilovich, A.; Stepanov, V.; Smirnov, S.; Volkovich, A. Visualization of Radioactive Sources without Gamma-Radiation with UV Imaging Systems. In Proceedings of the 12th International Conference on Environmental Remediation and Radioactive Waste Management, Liverpool, UK, 11–15 October 2009; pp. 321–325.
14. Ivanov, O.P.; Stepanov, V.E.; Smirnov, S.V.; Volkovich, A.G. Development of method for detection of alpha contamination with using UV-camera “DayCor” by OFIL. In Proceedings of the 2011 IEEE Nuclear Science Symposium Conference Record. Valencia, Spain, October 23–29, 2011, pp. 2192–2194.
15. Crompton, A.; Gamage, K.; Bell, S.; Wilson, A.; Jenkins, A.; Trivedi, D. First Results of Using a UVTron Flame Sensor to Detect Alpha-Induced Air Fluorescence in the UVC Wavelength Range. *Sensors* **2017**, *17*, 2756. [[CrossRef](#)]
16. Lofthus, A.; Krupenie, P.H. The spectrum of molecular nitrogen. *J. Phys. Chem* **1977**, *6*, 113–307. [[CrossRef](#)]
17. Crompton, A.; Gamage, K.; Bell, S.; Wilson, A.; Jenkins, A.; Trivedi, D. Gas Flow to Enhance the Detection of Alpha-Induced Air Radioluminescence Based on a UVTron Flame Sensor. *Sensors* **2018**, *18*, 1842. [[CrossRef](#)]
18. Dondes, S.; Harteck, P.; Kunz, C. A Spectroscopic Study of Alpha-Ray-Induced Luminescence in Gases: Part I. *Radiat. Res.* **1966**, *27*, 174–210. [[CrossRef](#)]
19. Kerst, T.; Toivonen, J. Intense radioluminescence of NO/N₂-mixture in solar blind spectral region. *Opt. Express* **2018**, *26*, 33764–33771. [[CrossRef](#)] [[PubMed](#)]
20. Duquesne, M.; Kaplan, I. Mesure de la luminescence induite par le rayonnement α du ²¹⁰Po dans l’air et dans l’eau. *J. Phys. Radium* **1960**, *21*, 708–716. [[CrossRef](#)]
21. Yamamoto, S.; Komori, M.; Koyama, S.; Toshito, T. Luminescence imaging of water during alpha particle irradiation. *Nucl. Instrum. Meth. A* **2016**, *819*, 6–13. [[CrossRef](#)]

22. Yamamoto, S.; Akagi, T.; Yamashita, T.; Toivonen, J.; Yamaguchi, M.; Komori, M.; Kawachi, N. Source of luminescence of water lower energy than the Cerenkov-light threshold during irradiation of carbon-ion. *J. Phys. Commun.* **2018**, *2*, 065010. [CrossRef]
23. Sico GmbH: Transmission of Synthetic Quartz Glass (SQ) and Silu. Available online <http://www.sico.at/assets/files/transmission.pdf>. (accessed on 6 March 2019).
24. Live Chart of Nuclides. Available online <https://www-nds.iaea.org/relnsd/vcharthtml/VChartHTML.html>. (accessed on 22 January 2019).
25. Segelstein, D.J. The Complex Refractive Index of Water; Ph.D. Thesis, University of Missouri, Kansas City, MO, USA, 1981.
26. Sand, J.; Ihantola, S.; Peräjärvi, K.; Nicholl, A.; Hrncek, E.; Toivonen, H.; Toivonen, J. Imaging of alpha emitters in a field environment. *Nucl. Instrum. Meth. A* **2015**, *782*, 13–19. [CrossRef]
27. Radioactive elements: Table of nucleids. Available online: http://www.kayelaby.npl.co.uk/atomic_and_nuclear_physics/4_6/4_6_1_part09_090_099.html. (accessed on 18 March 2019).
28. Egorov, O.B.; Addleman, R.S.; O'Hara, M.J.; Marks, T.; Grate, J.W. Direct measurement of alpha emitters in liquids using passivated ion implanted planar silicon (PIPS) diode detectors. *Nucl. Instrum. Methods Phys. Res. Sect. A* **2005**, *537*, 600–609. [CrossRef]
29. Waldenmaier, T. Spectral Resolved Measurement of the Nitrogen Fluorescence Yield in air Induced by Electrons. *Astropart. Phys.* **2008**, *29*, 205–222 [CrossRef]
30. Jelley, J. Cerenkov radiation and its applications. *Br. J. Appl. Phys.* **1955**, *6*, 227. [CrossRef]
31. Ackerman, N.; Graves, E. The potential for Cerenkov luminescence imaging of alpha-emitting radionuclides. *Phys. Med. Biol.* **2012**, *57*, 771. [CrossRef]
32. Christensen, H.; Sunder, S. Current State of Knowledge of Water Radiolysis Effects on Spent Nuclear Fuel Corrosion. *Nucl. Technol.* **2000**, *131*, 102–123. [CrossRef]



© 2019 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).