

Measurement of ultra-trace level of intact oxytocin in plasma using SALLE combined with nano-LC-MS

Dan Liu^{a#}, Xiaoli Han^{b#}, Xinxin Liu^{a#}, Mengchun Cheng^a, Meixi He^b, Gregor Rainer^c, Huiyuan Gao^{b*}, Xiaozhe Zhang^{a*}

^a CAS Key Laboratory of Separation Sciences for Analytical Chemistry, Dalian Institute of Chemical Physics, Chinese Academy of Sciences, 457 Zhongshan Road, Dalian, 116023, China

^b Key Laboratory of Structure-Based Drug Design & Discovery of Ministry of Education, Shenyang Pharmaceutical University, Shenyang 110016, China.

^c Department of Medicine, University of Fribourg, Chemin de Musee 5, Fribourg, CH-1700, Switzerland.

UPLC-QTOF

The detection of model compounds were performed using Agilent 1290 Series UPLC system with Agilent Poroshell 120 SB-C4 column (150 × 3.0 mm, 1.8 μm). A mobile phase consisting of solvent A 0.5% formic acid and solvent B acetonitrile was applied at a flow rate of 0.3 mL/min. Gradient elution was 0-5 min, 5-65% B; 5-8 min, 65-100% B; 8-13 min, 100% B; 13-15 min, 5% B. Each sample volume injected was 5 μL. Mass spectrometry was carried out by using an Agilent 6520 Q-TOF MS (Agilent Corporation, CA, USA) equipped with an electrospray ionization (ESI) interface. The optimum operating parameters was as follows: drying gas (N₂) flow rate of 9.0 L/min; drying gas temperature of 350°C; voltage of capillary of 4000 V; skimmer of 65 V; fragmentor voltage of 180V. Full-scan data acquisition in positive ion mode was operated from 450~550 with accurate mass of 5 ppm and acquisition rate of 3 spectra per second. The divalent ions of OT (Calc. m/z: 504.2271) and stable isotope-labeled OT (as IS, Calc. m/z: 507.2325) were detected by extracted ion chromatograms (EIC). Data were handled on the MassHunter WorkStation Data Acquisition software (Version B.02.01) and Quantitative Analysis software (Version B.04.00).

E-mail address: sypugaohy@163.com (H.Y. Gao) and zhangxz@dicp.ac.cn (X.Z. Zhang)

Protein precipitation (PPT) method

Protein precipitation was performed by adding ice-cold AcN and whirl mixing 30s, then centrifuged 15min at 15,000g. The supernatant was evaporated to dryness.

SDS-PAGE method

Each sample lyophilized was dissolved in 1x SDS-PAGE sample buffer (2ml 0.5mol/L Tris-HCl (pH 6.8), 1.6ml glycerol, 3.2ml 10% SDS, 0.8ml 2-mercaptoethanol, 0.4ml 1.0% bromophenol blue) and incubated at 90°C for 7 min. 5ul each sample or 3ul standard polypeptide marker was applied to a 15% commercial polyacrylamide gels (Willget biotech Co., Ltd in shanghai, China.). The gel was run at a constant voltage of 120V for 1h in a Mini Protein Tetra System (BioRad) and then stained by 0.1% silver nitrate solution.