

2018

# Towards smart self-clearing glaucoma drainage device.

Hyunsu Park

Amir Hossein Rafiee

Simon W M John

*The Jackson Laboratory*, [simon.john@jax.org](mailto:simon.john@jax.org)

Arezoo M. Ardekani

Hyowon Lee

Follow this and additional works at: <https://mouseion.jax.org/stfb2018>

---

## Recommended Citation

Park, Hyunsu; Rafiee, Amir Hossein; John, Simon W M; Ardekani, Arezoo M.; and Lee, Hyowon, "Towards smart self-clearing glaucoma drainage device." (2018). *Faculty Research 2018*. 251.

<https://mouseion.jax.org/stfb2018/251>

This Article is brought to you for free and open access by the Faculty Research at The Mouseion at the JAXlibrary. It has been accepted for inclusion in Faculty Research 2018 by an authorized administrator of The Mouseion at the JAXlibrary. For more information, please contact [ann.jordan@jax.org](mailto:ann.jordan@jax.org).

ARTICLE

Open Access

# Towards smart self-clearing glaucoma drainage device

Hyunsu Park<sup>1</sup>, Amir Hossein Raffiee<sup>2</sup>, Simon W. M. John<sup>3</sup>, Arezoo M. Ardekani<sup>2</sup> and Hyowon Lee<sup>1</sup>

## Abstract

For patients who are unresponsive to pharmacological treatments of glaucoma, an implantable glaucoma drainage devices (GDD) are often used to manage the intraocular pressure. However, the microscale channel that removes excess aqueous humor from the anterior chamber often gets obstructed due to biofouling, which necessitates additional surgical intervention. Here we demonstrate the proof-of-concept for smart self-clearing GDD by integrating magnetic microactuators inside the drainage tube of GDD. The magnetic microactuators can be controlled using externally applied magnetic fields to mechanically clear biofouling-based obstruction, thereby eliminating the need for surgical intervention. In this work, our prototype magnetic microactuators were fabricated using low-cost maskless photolithography to expedite design iteration. The fabricated devices were evaluated for their static and dynamic mechanical responses. Using transient numerical analysis, the fluid–structure interaction of our microactuator inside a microtube was characterized to better understand the amount of shear force generated by the device motion. Finally, the anti-biofouling performance of our device was evaluated using fluorescein isothiocyanate labeled bovine serum albumin. The microactuators were effective in removing proteinaceous film deposited on device surface as well as on the inner surface of the microchannel, which supports our hypothesis that a smart self-clearing GDD may be possible by integrating microfabricated magnetic actuators in chronically implanted microtubes.

## Introduction

Glaucoma is a group of eye diseases that causes progressive damage to optic nerve. It is commonly known as “the silent thief of sight” due to the lack of symptoms during the early stages<sup>1</sup>. Because of this difficulty in early diagnosis, glaucoma remains as one of the leading causes of blindness and visual impairments in the world<sup>2</sup>. It currently affects around 64.3 million people in the world and this number is expected to double by 2040<sup>3,4</sup>. In the United States, there are more than 3 million patients with glaucoma and it disproportionally affects African Americans and Hispanics<sup>5–9</sup>. Glaucoma is a major healthcare

issue with the annual cost for treatment in the US that exceeds \$2.9 billion<sup>10</sup>.

Typically, glaucoma patients experience poor drainage of aqueous humor (AH) through the natural outflow pathways (i.e., trabecular meshwork and Schlemm’s canal)<sup>11</sup>. The imbalance between the rate of production and the outflow of AH from the eye causes an increased intraocular pressure (IOP), which is a major risk factor that leads to subsequent damage to optic nerve and the loss of eyesight<sup>12,13</sup>. Unfortunately, there is no cure for glaucoma. However, the progression of disease can significantly be delayed using pharmaceutical and surgical interventions that maintain the IOP in a safe range to minimize optic nerve damage<sup>14</sup>. Glaucoma drugs are typically designed to decrease the production of AH or to increasing its outflow through trabecular meshwork or uveoscleral pathway<sup>15–18</sup>. As with most pharmaceutical interventions, however, these drugs have several undesirable side effects including bitter taste, headache,

Correspondence: Hyowon Lee ([hwlee@purdue.edu](mailto:hwlee@purdue.edu))

<sup>1</sup>Weldon School of Biomedical Engineering, Birck Nanotechnology Center, Center for Implantable Devices, Purdue University, West Lafayette, IN 47907, USA

<sup>2</sup>School of Mechanical Engineering, Purdue University, West Lafayette, IN 47907, USA

Full list of author information is available at the end of the article.

© The Author(s) 2018



**Open Access** This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article’s Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article’s Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

conjunctivitis, visual blurring, eyelid inflammation, and eye pain<sup>19–21</sup>. Surgical treatments such as trabeculectomy and laser trabeculoplasty can also be used to increase AH outflow but these invasive procedures often lead to serious post-operative complications such as hypotony, cataract, and bleb-related infections<sup>22–25</sup>. Moreover, the surgical ablation of trabecular meshwork often results in coagulative necrotic tissue, which can cause difficulty in chronic management of optimal AH outflow<sup>26</sup>.

For patients with refractory or inflammatory glaucoma who are unresponsive to conventional pharmacological or surgical procedures, glaucoma drainage device (GDD) are often implanted. These devices offer several advantages over conventional trabeculectomy including better IOP control, ease of surgery, and minimum post-surgical complications<sup>27–29</sup>. Traditional GDDs consists of a short polymeric microscale tube that connects the anterior chamber to a thin silicone plate for drainage of excess AH<sup>30–32</sup>. While GDDs have been used to manage IOP for glaucoma patients for the past 40 years, 15.1% of implanted devices fail within 3 years and more than 29.8% fail within 5 years post-implantation<sup>33,34</sup>. Clinical studies have shown that up to 10% of glaucoma patients require additional medications and surgical intervention because of the tube blockage<sup>35</sup>. The hydrophobic polymer materials from which GDDs are constructed (e.g., polypropylene, polymethylmethacrylate, and polydimethylsiloxane) typically have high affinity for interstitial proteins such as fibrinogen, immunoglobulin, and albumin that adsorbs onto the device surface within minutes after the implantation<sup>36–38</sup>. Once it forms, the proteinaceous layer triggers the inflammatory response that can lead to premature implantable device failure<sup>39</sup>. Since GDDs generally have a drainage tube with an inner diameter that ranges from 50 to 600  $\mu\text{m}$ , the microscale channel can easily be occluded by various biofouling materials including vitreous, fibrin, or blood clot<sup>40–43</sup>.

One promising approach that can remove adsorbed biofouling material is to ablate occlusion using neodymium-doped yttrium aluminum garnet (Nd:YAG) laser. However, there are several potential risks associated with laser treatments including focal cataracts, prolonged elevation of the intraocular pressure, posterior capsule rupture, retinal injury, and laser injury<sup>44–50</sup>. Tissue plasminogen activator (tPA), which is a serine protease involved in the breakdown of fibrin or blood clots, has also been used to clear occluded glaucoma shunts<sup>51,52</sup>. However, tPA may cause additional undesirable complications such as hyphema, active bleeding, and vitreous hemorrhage<sup>53–55</sup>.

Establishing a method to non-invasively remove biofouling without causing side effects can significantly improve the reliability and functionality of many chronically implanted devices. Here we report on the

design, fabrication, and testing of anti-biofouling microtube integrated with an array of magnetic microactuators as a part of a self-clearing GDD that can actively combat against proteinaceous biofouling in situ without the need for additional surgical or pharmaceutical interventions. We believe our strategy to remove bioaccumulation on-demand using externally applied magnetic field is a way to significantly improve the functional lifetime of implantable devices that suffer from biofouling-related performance degradation. By integrating thin-film magnetic microactuators fabricated out of liquid crystal polymer (LCP) using maskless lithography, here we demonstrate a low-cost prototype of self-clearing GDD drainage tube (Fig. 1). Using fluorescent-tagged bovine serum albumin, we show the protein-clearing capabilities of these prototype GDD microtubes using time-varying magnetic fields, which may eliminate the need for additional surgical or pharmaceutical interventions for glaucoma patients.

## Results


### Device fabrication and mechanical characterization

Figure 2a shows our microfabricated LCP-based device. The needle-shape was chosen to accommodate the relatively small tube diameter. We can control the deflection direction and amplitude of the microactuator by adjusting the strength and the direction of the externally applied magnetic field (Fig. 2b). We assembled the microdevices into a prototype GDD drainage tube using an anchor to demonstrate protein-removal performance inside the tube (Fig. 2c). Once manually placed into the microtube, we heated the tube and applied tensile stress at both ends to decrease the diameter of the microtube and fix the microactuators in position, which prevented any shifting of devices during actuation in a continuous fluid flow.

To characterize the actuation capabilities of our magnetic microactuators, we evaluated the static and dynamic mechanical responses. A magnetic moment of the soft ferromagnetic element is generated when the magnetic microactuator is placed in a static magnetic field. The microactuators can deflect out of plane when the direction of the applied magnetic field is normal to the magnetization direction of the ferromagnetic element (Fig. 2b). The deflection angle of magnetic microactuator can be described in the ref.<sup>56</sup>

$$\phi = \frac{V_m (\vec{M} \times \vec{H})}{k_{beam}} \quad (1)$$

with the angular deflection  $\phi$ , magnet volume  $V_m$ , magnetization  $\vec{M}$ , applied magnetic field  $\vec{H}$ , and the flexure stiffness  $k_{beam}$ . The beam geometry and the material property affect the mechanical stiffness of the flexure with



following

$$k_{beam} = \frac{E_c w t^3}{12l} \tag{2}$$

with the elastic modulus  $E_c$ , beam width  $w$ , beam thickness  $t$ , and beam length  $l$ <sup>57,58</sup>.


As shown in Fig. 3a, the measured deflection angle corresponded closely with the theoretical values. We varied the frequency and amplitude of the externally-applied, time-varying magnetic fields using a custom electromagnet to obtain the frequency response (10–200 Hz) of our microactuators (Fig. 3b). As expected, the amplitude of deflection increased as a function of applied magnetic field strength. Furthermore, we determined that the actuation frequency of 20 Hz to be the primary resonance, which can be used to generate the highest dynamic deflection amplitude. The increase in dynamic deflection amplitude may be attributable to the increase in mean fluid velocity around the microactuator, which leads to an increase in the wall shear stress on the microactuator and the tube<sup>59,60</sup>. Therefore, we used a fixed

actuation frequency (20 Hz) with the highest actuation amplitude (64°) for all experiments and simulation.

To verify these Ti-coated LCP based microactuator is robust enough to withstand a large number of actuation cycles in physiological condition, we examined the changes in the dynamic responses of these microdevices in 37 °C phosphate buffered saline (PBS, ThermoFisher Scientific, Waltham, MA, USA). After 10.9 million actuation cycles, we saw no visible damage to the LCP-based microactuators and no change in the resonant frequency of tested devices ( $n = 4$ , Supplementary Fig. 1). If we assume a 5-min weekly actuation, this equates to up to 35 years of lifetime, which suggests adequate robustness for our LCP-based microactuators against fatigue related failure.

**Fluid–structure interaction**

We evaluated the shear stress distribution generated by the microactuation motion using finite element modeling. The simulation results showed that the maximum shear stress is generated near the perimeter of the actuator (Fig. 4a). When the device is integrated into the microtube, the actuation leads to a larger shear stress as the



microactuator approaches the tube wall (Fig. 4c). During the actuation, the maximum shear stress of  $\sim 8$  and  $\sim 10$   $\text{dyn/cm}^2$  are generated periodically on the surface of the actuator and the tube, respectively (Fig. 4b, d).

#### BSA-FITC adsorption and desorption

To maximize the fluorescence intensity, we incubated Ti-coated LCP samples in various concentrations of BSA-FITC (1–8 mg/ml) for 2 h. The fluorescence intensity of absorbed BSA-FITC plateaued around 5 mg/ml (Fig. 5a), therefore, all subsequent BSA-FITC evaluations used this concentration. The jet impingement technique is widely used to analyze the shear stress required to remove cells by corresponding the size of a lesion created by a perpendicular jet of fluid to a well-characterized shear stress profile<sup>61,62</sup>. To quantify the adhesion strength of BSA-FITC on Ti-coated LCP surface, we used the theoretical description of the wall shear stress under the impinging jet proposed by Phares et al.<sup>63</sup>. For this analysis, we assumed that the AH is incompressible Newtonian fluid in a steady and laminar flow. In the theoretical description of the wall shear stress in normally impinging jet with jet height  $H$ , the wall shear stress  $\tau$  at a radial distance  $r$  can be described by

$$\frac{\tau}{\tau_m} = 0.18 \left( \frac{1 - e^{-114\lambda^2}}{\lambda} \right) - 0.943\lambda e^{-114\lambda^2} \quad (3)$$


with the maximum shear stress  $\tau_m$  and non-dimensionalized jet height ( $\lambda = r/H$ ). The maximum shear stress  $\tau_m$  is given by

$$\tau_m = 0.16 \frac{\rho u_o^2}{(H/D)^2} \quad (4)$$

with the fluid density  $\rho$ , the average flow velocity at the nozzle exit  $u_o$ , and diameter for the nozzle  $D$ . The critical shear stress ( $\tau_c$ ) required to remove the adsorbed BSA-FITC can then be calculated by measuring the radius of lesion (Fig. 5b).

In the jet impingement test, the fluid jet was delivered at a flow rate of 1.18 ml/min for 5 s, which corresponds to Reynolds number of 100 in laminar flow range to be used for Eqs. (3) and (4). The fluid jet created  $\tau_m \sim 30$   $\text{dyn/cm}^2$  which is in line with published shear stress value required to rupture protein-ligand interaction<sup>64</sup>. Figure 5b shows an image of BSA-FITC lesion created by jet impingement and a plot of non-dimensional wall shear stress as a function of non-dimensional lesion size for jet radius  $R_{\text{jet}} = 125$   $\mu\text{m}$ . With an average lesion radius of 284  $\mu\text{m}$  ( $n = 4$ ), the estimated shear stress required to remove BSA-FITC ( $\tau_c$ ) was 10.2  $\text{dyn/cm}^2$ . The numerical analysis results (Fig. 4) showed that our magnetic microactuators can generate up to 10  $\text{dyn/cm}^2$ . Taken together with the results from our jet impingement study, we expected to





show a robust protein removal using our prototype GDD drainage tube.


### Protein biofouling removal in GDD

The main function of our magnetic microactuators is to remove the protein adsorbed on the device surface and the inner wall of GDD microtube to prevent the initiation of inflammatory cascade. As such, we quantified the decrease in fluorescence intensity due to device actuation on device surface and the inner wall of the microtube as simulated in Fig. 4. To study anti-biofouling capability of the actuator itself, we actuated BSA-FITC coated devices with different actuation durations at 20 Hz. The maximum actuation duration was set to 5 min based on prior literature<sup>65,66</sup> and for practical consideration assuming that a shorter actuation protocol would be less burdensome on clinicians and patients. The minimum actuation duration was set to be 30 s, which is 10% of the maximum actuation duration.

Figure 6 demonstrates BSA-FITC removal due to actuation of magnetic microdevices. We compared the

decreased fluorescence intensity values using one-way analysis of variance (ANOVA) with Tukey’s HSD post-hoc pairwise analysis. The results showed that BSA-FITC coated on microactuators ( $n = 3$ ) was significantly reduced compared to non-actuated control regardless of deflection amplitude or actuation duration ( $p < 0.01$ ). Without actuation, the fluorescence intensity decreased by approximately 10–20% depending on treatment duration. However, the difference in fluorescent intensity between the small ( $8^\circ$ ) and large ( $64^\circ$ ) deflection magnitudes was not statistically significant. The impact of actuation duration was also statistically significant. When actuated for 30 s, BSA-FITC amount reduced by 42%, whereas up to 85% protein clearance can be seen on device when actuated for 5 min. Thus, in subsequent evaluations to determine the impact of actuation on removing protein adsorbed on the microtube inner wall, we actuated all samples for 5 min to maximize protein clearance.

To demonstrate the in situ anti-biofouling performance of our smart GDD, we coated the inner lumen of




300- $\mu\text{m}$ -diameter microtube integrated with our magnetic microactuator using BSA-FITC (Fig. 2c). Figure 7a highlights the difference in fluorescence intensity between actuated versus non-actuated GDD prototype. Without actuation, we saw virtually no difference in fluorescence intensity. Following actuation, however, we saw a significant decrease in fluorescence intensity in areas surrounding the microdevice. The pattern of cleared area closely resembles the shear stress distribution predicted by our numerical analysis (Fig. 4). We then quantified the amount of fluorescence intensity decrease from the end of the beams to the actuator tip and compared the results using a two-sample *t*-test. The results show that the microactuation can remove significant amount of adsorbed BSA from the tube wall compared to the non-actuated control ( $p < 0.01$ , Fig. 7b). However, the decrease in fluorescence intensity (<40%) in microtube was much smaller than the 85% decrease we saw from the device surface following a 5 min actuation (Fig. 6b).

## Discussion

Here we demonstrated that the LCP-based microactuators can easily be fabricated at low-cost using our maskless photolithography. LCP is widely used polymer in


biomedical applications due to their near hermetic properties, biocompatibility and superior chemical resistance<sup>67–70</sup>. By using commercially available low-cost LCP sheets as the substrate in combination of highly scalable microfabrication processes such as maskless photolithography, oxygen etching, electroplating, and polymer coating, it is possible to manufacture these LCP-based devices at extremely low costs. Moreover, the simple integration process that we employed to immobilize microdevices inside a small drainage tube may be used to create other smart MEMS-enabled catheter-based devices.

Both the static and the dynamic responses of these microfabricated LCP-based actuators corresponded well with the theoretical values, which suggests a good control of our fabrication process. The *in vitro* evaluation using BSA-FITC showed that, as expected, the actuation from our device can effectively reduce proteinaceous biofouling on the actuator surface and the inner wall of the microtube. The results from the *in vitro* experiments demonstrated a good agreement with our results from the numerical analysis that predicted the pattern of biofilm clearance by quantifying the shear stress distribution and the jet impingement study that quantified the adhesion



strength of the BSA-FITC. This approach to quantify the adhesion properties of specific biofilm and to model the shear stress profile of a device actuation may be used in future iterations to design novel microactuator arrays that are tailored for bespoke implantable application against specific biofouling materials.


To confirm that the biofouling removal process does occur via mechanical shear generated by the microactuation and not by the heat generated from the microactuation, we also measured the amount of heat generated during actuation (Supplementary Fig. 2). When we actuated our microdevices ( $n = 4$ ) for 5 min at 20 Hz using 40 mT in room temperature PBS, no temperature increase was seen in thermal camera images (FLIR



A325sc, FLIR, Wilsonville, OR, USA). It is important to ensure that no thermal effect occurs due to microactuation since excessive heat may lead to unintended damage to the surrounding tissue.

The potential implication of utilizing active mechanism for combating biofouling is enormous since many chronically implantable devices including biosensors, neural interface electrodes, and drug delivery and drainage devices suffer from significant performance





**Fig. 7 Protein cleaning in the microtube.** **a** Fluorescence images of BSA-FITC removal using the microactuator located within the lumen of microtube with an actuation for 5 min at 20 Hz (a. u., scale bar = 200  $\mu\text{m}$ ). **b** The area used for measuring fluorescence intensity (red dotted box (scale bar = 200  $\mu\text{m}$ )) and decreased fluorescence intensity near the actuator

degradation due to biofouling<sup>39,71</sup>. Although there is a number of proposed mechanisms for actively addressing biofouling using electrical and mechanical transducers<sup>65</sup>, the magnetically-powered actuators have several key advantages. First, the magnetic device can be activated in situ wirelessly via externally applied magnetic field with low power requirements (Supplementary Fig. 3) without the need for any invasive procedure. Second, the magnetic microactuators can be tailored to deliver large disruptive forces to remove multi-scale biofouling materials including protein, bacteria, and cells. Third, the lack of integrated circuit and internal power source can facilitate the integration and packaging of these type of devices into existing medical devices, which can accelerate clinical adoption. Finally, as mentioned, the simple design is compatible with many scalable microfabrication technologies that can significantly reduce the cost of manufacturing.

Despite these key benefits, there are several remaining questions to be answered. First, the amount of protein

removed from the microtube wall was much lower than that from the actuator surface despite our numerical analysis demonstrating a higher maximum shear stress on the wall. This may be due to the fact that each microscope image was focused on microactuator surface, which is located at the center of the microtube. As can be seen from Fig. 4 and the Supplementary Video, the magnitude of shear stress distribution around the mid-plane of microtube is much lower than the top and bottom of the microtube. It may be interesting to characterize protein distribution using a confocal microscope in the future to verify this hypothesis. If not, it is possible to leverage our predictive modeling to redesign microactuators that can provide a greater average shear stress to ensure a more efficient protein-removal. Secondly, additional experiments are needed to determine optimum actuation duty cycle that will ensure a protein-free GDD microtube. Although the microactuators were able to demonstrate good protein reduction in just 5 min, it may be possible to reduce this actuation duration further by performing a

systematic evaluation. Thirdly, although it is possible to integrate many microactuators into a long microtube (Supplementary Fig. 4), the manual assembly process can be further streamlined if 1D arrays of microactuators are fabricated to better control the device spacing. Finally, additional *in vitro* and *in vivo* work is necessary to ascertain whether periodically removing biofilm using our self-clearing implants will actually prolong the device lifetime and improve patient outcome. A critical question to address is to determine what happens with the displaced biofouling material. A detailed histopathological evaluations using animal models must be performed to ensure that the displaced biomaterial will not cause undesirable downstream effects.

## Materials and methods

### Device fabrication

We fabricated the microactuators from copper-(Cu)-cladded LCP using a custom maskless photolithography setup previously described<sup>72</sup>. We used a computer connected to a conventional home theater projector with a digital micromirror device (HD142X, Optoma, Fremont, CA, USA) to project and expose a desired pattern<sup>72–75</sup>. The projector was vertically fixed on a stereo-microscope (SM-4B, Amscope, Irvine, CA, USA) using a custom machined bracket. To improve the resolution and reduce the size of the image, we optimized the alignment between the lens of the microscope, the sample stage, and the projector. We used Microsoft PowerPoint to design and project various mask patterns. We adjusted the exposure intensity by modifying pattern color in the software. Figure 1c shows the overall process flow for the device fabrication.

The commercially available LCP sheet (Ultralam 3850, Rogers corporation, Chandler, AZ, USA) has a thickness of 25  $\mu\text{m}$ . To improve compliance of the cantilevers, we reduced the LCP thickness to 8  $\mu\text{m}$  using a reactive ion etcher (RIE, PlasmaPro80, Oxford Instruments plc, Abingdon, Oxfordshire, United Kingdom) after removing Cu from one side using a wet Cu etchant (CE-100, Transene, Danvers, MA, USA). We then mounted the 8- $\mu\text{m}$ -thick single clad LCP sheet onto a carrier wafer using a positive photoresist (PR) (AZ9260, Microchem, Westborough, MA, USA) with the Cu on top. We spin coated AZ9260 onto the Cu layer and exposed the cantilevers designs using our custom maskless photolithography setup. After etching the Cu layer using a wet Cu etchant (CE-100, Transene, Danvers, MA, USA), we removed PR using acetone. Next, we defined the Ni magnet electroplating mold on spin-coated AZ9260 using the same maskless photolithography procedure. We electroplated Ni to achieve a final thickness of 20  $\mu\text{m}$ . After removing the PR, we etched the cantilever pattern on bare LCP layer using an RIE and removed the remaining Cu

layer using a chemical etchant (BTP, Transene, Danvers, MA, USA). Finally, we coated the device with 100 nm thick titanium (Ti) using a sputterer (Magnetron sputtering systems, PVD Products, Inc., Wilmington, MA, USA) on both sides to improve biocompatibility.

### Mechanical characterization

We measured the magnitudes of angular deflections for a range of applied magnetic flux density up to 40.9 mT. We used a bespoke iron-core electromagnet to generate the magnetic field. We quantified the strength of magnetic field using a commercial gaussmeter (Model 8010, Pacific Scientific OECO, Milwaukie, OR, USA). We then placed the device on top of the electromagnet and applied the magnetic field of varying amplitude and frequency. We imaged the deflected actuators using a digital microscope KH8700, Hirox, Hackensack, NJ, USA) and calculated the deflection angles from the images using imageJ software (version 1.50i). To better characterize the motion of the devices in liquid, we also characterized the dynamic responses of the magnetic microactuators in deionized water using a custom laser deflecting setup. Using a mirror, we placed a laser beam onto the metallic surface of the device, which then reflected the laser beam onto a position sensitive diode (PSD) sensor. We recorded the two-dimensional position data from the PSD using a custom data acquisition system (LabView 2014, Austin, TX, USA).

### Fatigue evaluation

As a baseline, we photographed and measured the dynamic responses of each test sample ( $n = 4$ ). We immobilized each microactuator on glass slide using polyimide tape for the fatigue evaluation. We then placed the glass slide fixture in a beaker filled with PBS at 37 °C. We actuated the device for 6 days to achieve 10.9 million cycles at 12 mT and 20 Hz sinusoidal signal. Following the continuous actuation, the microactuators were removed from the beaker, photographed, and analyzed for post-actuation dynamic response.

### Fluid–structure interaction

To determine the shear stress generated by the actuation, we used a finite volume method to simulate shear stress on the surface of microactuator and the lumen of the microtube by numerically solving Navier–Stokes equations. We discretized the computational domain using a uniform, staggered, cartesian grid. We used Euler explicit method for time discretization and spatial derivatives in convective and computed diffusive terms using the quadratic upstream interpolation for convective kinematics and central difference schemes, respectively<sup>76</sup>. Furthermore, we coupled the pressure and velocity using a projection method<sup>77</sup>. We implemented a distributed

Lagrange multiplier method to simulate the motion of microactuator in a viscous fluid, which allowed us to accurately capture the hydrodynamic interaction between the microactuator and a surrounding fluid and evaluate the shear stress acting on the surface<sup>78,79</sup>.

### Protein-based biofouling adhesion

We tested the anti-biofouling performance of our magnetic microactuators using fluorescent-tagged bovine serum proteins (BSA-FITC, ThermoFisher Scientific, Waltham, MA, USA), which is readily coated onto the implant surface via non-specific binding and subsequently initiates the inflammatory response *in vivo*<sup>80</sup>. We incubated the devices and samples for jet impingement test in the BSA-FITC solution of various concentrations (1–8 mg/ml) in PBS ( $n = 5$ , each) for 2 h and rinsed with deionized water. We captured the images of protein coated samples using a fluorescence microscope (Axio Observer Z1, Carl Zeiss Microscopy, LLC) and a filter set 17 (excitation, BP 485/20, and emission BP 515–565, Carl Zeiss Microscopy, LLC), and quantified the fluorescence intensity using imageJ. We normalized each image using the imageJ against bare non-coated sample.

We quantified the magnitude of shear stress required to remove the absorbed protein on Ti-coated surface using a jet impingement experiment. We vertically placed the tip of a 15-ml syringe with the needle having an inner diameter of 250  $\mu\text{m}$  (7018333, Nordson EFD, East Providence, RI, USA) 1 mm over BSA-FITC coated substrate and delivered the jet flow using a syringe pump (NE-300, New Era Pump Systems, Inc., Farmingdale, NY, USA). We created a total of four lesions. Using the same fluorescence microscope describe above, we imaged of each lesion and measured the diameter of each using imageJ. We subtracted the background fluorescence from bare substrate to normalize fluorescence intensity,

### Protein biofouling removal in GDD

We investigated the impact of deflection amplitude (8° vs. 64°) and actuation duration (30 s vs. 5 min) of our magnetic microactuators by quantifying the amount of BSA-FITC ( $n = 3$ , each). We placed each sample in deionized water in a custom testing chamber to block the ambient light during actuation. We captured the fluorescence images of protein-coated devices before and after the actuation and quantified the difference in fluorescence intensity using imageJ. We normalized the background fluorescence by subtracting fluorescence from a bare actuator surface. To demonstrate the anti-biofouling capability of magnetic microactuators inside a polytetrafluoroethylene microtube, we coated the lumen of assembled GDD drainage tube by flowing BSA-FITC at 2.7  $\mu\text{l}/\text{min}$ , which is the average flow rate of AH in human eyes. We then actuated the micro-devices at 20 Hz for 5 min to remove the adsorbed protein

layer ( $n = 3$ ). Finally, we quantified the decrease in fluorescence level due to actuation and compared with the non-actuated controls.

### Acknowledgements

This work was supported in part with The Jackson Laboratory - Purdue University Collaborative Seed Grant, NIH NINDS (R21NS095287), and NIH/NCRR Indiana CTSI (UL1TR001108).

### Authors' contributions

H.P., S.J., and H.L. conceived the experiments, H.P. fabricated the samples and conducted experiments, A.R. and A. A. performed the numerical analysis, and all authors analyzed the results and reviewed the manuscript.

### Author details

<sup>1</sup>Weldon School of Biomedical Engineering, Birk Nanotechnology Center, Center for Implantable Devices, Purdue University, West Lafayette, IN 47907, USA. <sup>2</sup>School of Mechanical Engineering, Purdue University, West Lafayette, IN 47907, USA. <sup>3</sup>Howard Hughes Medical Institute, The Jackson Laboratory, Bar Harbor, ME 04609, USA

### Conflict of interest

The authors declare that they have no conflict of interest.

**Supplementary Information** accompanies this paper at <https://doi.org/10.1038/s41378-018-0032-3>.

Received: 7 June 2018 Revised: 16 August 2018 Accepted: 18 August 2018.  
Published online: 5 November 2018

### References

- Lee, D. A. & Higginbotham, E. J. Glaucoma and its treatment. *Am. J. Health Syst. Pharm.* **62**, 691–699 (2005).
- Mariotti, S. P. Global data on visual impairments 2020. Technical Report (World Health Organization, 2010).
- Quigley, H. A. & Broman, A. T. The number of people with glaucoma worldwide in 2010 and 2020. *Br. J. Ophthalmol.* **90**, 262–267 (2006).
- Tham, Y. C. et al. Global prevalence of glaucoma and projections of glaucoma burden through 2040: a systematic review and meta-analysis. *Ophthalmology* **121**, 2081–2090 (2014).
- Broman, A. T. et al. Estimating the rate of progressive visual field damage in those with open-angle glaucoma, from cross-sectional data. *Invest. Ophthalmol. Vis. Sci.* **49**, 66–76 (2008).
- Ko, F. et al. Diabetes, triglyceride levels, and other risk factors for glaucoma in the National Health and Nutrition Examination Survey 2005–2008. *Invest. Ophthalmol. Vis. Sci.* **57**, 2152 (2016).
- Drance, S., Anderson, D. R. & Schulzer, M., Collaborative Normal-Tension Glaucoma Study Group. Risk factors for progression of visual field abnormalities in normal-tension glaucoma. *Am. J. Ophthalmol.* **131**, 699–708 (2001).
- Muñoz, B. et al. Causes of blindness and visual impairment in a population of older Americans: The Salisbury Eye Evaluation Study. *Arch. Ophthalmol.* **118**, 819–825 (2000).
- Gupta, P. et al. Prevalence of Glaucoma in the United States: The 2005–2008 National Health and Nutrition Examination Survey. *Invest. Ophthalmol. Vis. Sci.* **57**, 2905–2913 (2016).
- Rein, D. B. et al. The economic burden of major adult visual disorders in the United States. *Arch. Ophthalmol.* **124**, 1754 (2006).
- Seiler, T. & Wollensak, J. The resistance of the trabecular meshwork to aqueous humor outflow. *Clin. Exp. Ophthalmol.* **229**, 265–270 (1982).
- Gabelt, B. T. & Kaufman, P. L. Changes in aqueous humor dynamics with age and glaucoma. *Prog. Retin. Eye Res.* **24**, 612–637 (2005).
- Mark, H. H. Aqueous humor dynamics in historical perspective. *Surv. Ophthalmol.* **55**, 89–100 (2010).
- Kenneth Schwartz, D. B. Current management of glaucoma. *Curr. Opin. Ophthalmol.* **8**, 339–342 (2003).
- Bitó, L. Z. Prostaglandins: a new approach to glaucoma management with a new, intriguing side effect. *Surv. Ophthalmol.* **41 Suppl 2**, S1–S14 (1997).

16. Zimmerman, T. J. W. P. B. I. The beta-adrenergic blocking agents and the treatment of glaucoma. *Surv. Ophthalmol.* **23**, 347–362 (1979).
17. Thom, J. & Herbert, E. Timolol A beta-adrenergic blocking agent for the treatment of glaucoma. *Arch. Ophthalmol.* **95**, 601–604 (1977).
18. Coakes, R. L. & Brubaker, R. F. The mechanism of timolol in lowering intraocular pressure. *Arch. Ophthalmol.* **96**, 2045 (1978).
19. Strahlman, E., Tipping, R., Vogel, R. & Group, tl. D. S. A double-masked, randomized 1-Year study comparing dorzolamide (Trusopt), timolol, and betaxolol. *J. Chem. Inf. Model.* **53**, 1689–1699 (2013).
20. Bloch, S., Rosenthal, A. R., Friedman, L. & Caldarolla, P. Patient compliance in glaucoma. *Br. J. Ophthalmol.* **61**, 531–534 (1977).
21. Bansal, R. K. & Tsai, J. C. Compliance/adherence to glaucoma medications—a challenge. *J. Curr. Glaucoma Pract.* **1**, 22–25 (2007).
22. Jones, E., Clarke, J. & Khaw, P. T. Recent advances in trabeculectomy technique recent advances in trabeculectomy technique. *Curr. Opin. Ophthalmol.* **16**, 107–113 (2005).
23. Mchugh, D., Marshall, J., Timothy, J., Hamilton, P. A. M. & Raven, A. Diode laser trabeculectomy (DLT) for primary open-angle glaucoma and ocular hypertension. *Br. J. Ophthalmol.* **74**, 743–747 (1990).
24. Juzych, M. S. et al. Comparison of long-term outcomes of selective laser trabeculectomy versus argon laser trabeculectomy in open-angle glaucoma. *Ophthalmology* **111**, 1853–1859 (2004).
25. Spurny, R. C. & Lederer, C. M. Krypton laser trabeculectomy a clinical. *Arch. Ophthalmol.* **102**, 1626–1628 (1984).
26. Melamed, S., Pei, J. & Epstein, D. L. Delayed response to argon laser trabeculectomy in monkeys morphological and morphometric analysis. *Arch. Ophthalmol.* **104**, 1078–1083 (1986).
27. Gedde, S. J. et al. Treatment outcomes in the tube versus trabeculectomy (TVT) study after five years of follow-up. *Am. J. Ophthalmol.* **153**, 789–803 (2012).
28. Ceballos, E. M., Parrish, R. K. & Schiffman, J. C. Outcome of Baerveldt glaucoma drainage implants for the treatment of uveitic glaucoma. *Ophthalmology* **109**, 2256–2260 (2002).
29. Ayyala, R. S., Duarte, J. L. & Sahiner, N. Glaucoma drainage devices: state of the art. *Expert. Rev. Med. Devices* **3**, 509–521 (2006).
30. Melamed, S. & Fiore, P. M. Molten implant surgery in refractory glaucoma. *Surv. Ophthalmol.* **34**, 441–448 (1990).
31. Rittenbach, T. L. Proptosis from a baerveldt tube shunt implant. *Optom. Vis. Sci.* **91**, e145–8 (2014).
32. Sarkisian, S. R. Tube shunt complications and their prevention. *Curr. Opin. Ophthalmol.* **20**, 126–130 (2009).
33. Souza, C. et al. Long-term outcomes of Ahmed glaucoma valve implantation in refractory glaucomas. *Am. J. Ophthalmol.* **144**, 893–900 (2007).
34. Gedde, S. J. et al. Surgical complications in the tube versus trabeculectomy study during the first year of follow-up. *Am. J. Ophthalmol.* **143**, 804–814 (2007).
35. Schlote, T., Ziemssen, F. & Bartz-Schmidt, K. U. Pars plana-modified Ahmed glaucoma valve for treatment of refractory glaucoma: a pilot study. *Graefes Arch. Clin. Exp. Ophthalmol.* **244**, 336–341 (2006).
36. Hill, R. A., Pirouzian, A. & Liaw, L. H. Pathophysiology of and prophylaxis against late Ahmed glaucoma valve occlusion. *Am. J. Ophthalmol.* **129**, 608–612 (2000).
37. Choritz, L. et al. Surface topographies of glaucoma drainage devices and their influence on human tenon fibroblast adhesion. *Invest. Ophthalmol. Vis. Sci.* **51**, 4047–4053 (2010).
38. Ihlenfeld, J. V. et al. Transient in vivo thrombus deposition onto polymeric biomaterials: role of plasma fibronectin. *Trans. Am. Soc. Artif. Intern. Organs XXIV*, 727–735 (1978).
39. Kenneth Ward, W. A review of the foreign-body response to subcutaneously-implanted devices: the role of macrophages and cytokines in biofouling and fibrosis. *J. Diabetes Sci. Technol.* **2**, 768–777 (2008).
40. Coleman, A. L. et al. Initial clinical experience with the Ahmed glaucoma valve implant in pediatric patients. *Arch. Ophthalmol.* **115**, 186 (1997).
41. Dubey, S., Pegu, J., Agarwal, M. & Agrawal, A. Vitreous occlusion of tube implant in a phakic patient with traumatic glaucoma. *Oman J. Ophthalmol.* **7**, 2014–2016 (2014).
42. McClintock, M. & MacCumber, M. W. Lowered intraocular pressure in a glaucoma patient after intravitreal injection of ocriplasmin. *Clin. Ophthalmol.* **9**, 1995–1998 (2015).
43. Christakis, P. G. et al. The ahmed versus baerveldt study: three-year treatment outcomes. *Ophthalmology* **120**, 2232–2240 (2013).
44. Tsai, W. F., Chen, Y. C. & Su, C. Y. Treatment of vitreous floaters with neodymium YAG Laser. *Br. J. Ophthalmol.* **77**, 485–488 (1993).
45. Kanner, E. M., Netland, P. A., Sarkisian, S. R. & Du, H. Ex-PRESS miniature glaucoma device implanted under a scleral flap alone or combined with phacoemulsification cataract surgery. *J. Glaucoma* **18**, 488–491 (2009).
46. Tanito, M., Sano, I. & Ohira, A. A case report of progressive obstruction of Ex-PRESS miniature glaucoma shunt after transient flat anterior chamber and treatment using Nd:YAG laser. *BMC Ophthalmol.* **15**, 4–6 (2015).
47. Song, J. Complications of selective laser trabeculectomy: a review. *Clin. Ophthalmol.* **10**, 137–143 (2016).
48. Perez, C. I., Chansangpetch, S., Hsia, Y. C. & Lin, S. C. Use of Nd:YAG laser to recanalize occluded Cypass Micro-Stent in the early post-operative period. *Am. J. Ophthalmol. Case Rep.* **10**, 114–116 (2018).
49. Koo, E. H., Haddock, L. J., Bhardwaj, N. & Fortun, J. A. Cataracts induced by neodymium-yttrium-aluminium-garnet laser lysis of vitreous floaters. *Br. J. Ophthalmol.* **101**, 709–711 (2017).
50. Hahn, P., Schneider, E. W., Tabandeh, H., Wong, R. W. & Emerson, G. G. Reported complications following laser vitreolysis. *JAMA Ophthalmol.* **135**, 973–976 (2017).
51. Smith, M. F. & Doyle, J. W. Use of tissue plasminogen activator to revive blebs following intraocular surgery. *Arch. Ophthalmol.* **119**, 809–812 (2001).
52. Raczynska, D., Lipowski, P., Zorena, K., Skorek, A. & Glasner, P. Enzymatic vitreolysis with recombinant tissue plasminogen activator for vitreomacular traction. *Drug Des. Dev. Ther.* **9**, 6259–6268 (2015).
53. Zalta, A. H., Sweeney, C. P., Zalta, A. K. & Kaufman, A. H. Intracameral tissue plasminogen activator use in a large series of eyes with valved glaucoma drainage implants. *Arch. Ophthalmol.* **120**, 1487 (2002).
54. Sidoti, P. A. et al. Tissue plasminogen activator and glaucoma drainage. *J. Glaucoma* **4**, 258–262 (1995).
55. Lundy, D. C., Sidoti, P., Winarko, T., Minckler, D. & Heuer, D. K. Intracameral tissue plasminogen activator after glaucoma surgery: indications, effectiveness, and complications. *Ophthalmology* **103**, 274–282 (1996).
56. Judy, J. W. & Muller, R. S. Magnetic microactuation of torsional polysilicon structures. In *The 8th International Conference on Solid-State Sensors and Actuators*, vol. 1770, 332–335 (1995).
57. Young, W. & Budynas, R. *Roark's Formulas for Stress and Strain*. 7th edn, (McGraw-Hill, New York, 2002).
58. Judy, J. W., Muller, R. S., Fellow, L. & Actuation, A. S. Magnetically actuated, addressable microstructures. *J. Micro. Syst.* **6**, 249–256 (1997).
59. Große, W., Schröder. Mean wall-shear stress measurements using the micropillar shear-stress sensor MPS<sup>2</sup>. *Meas. Sci. Technol.* **19**, 015403 (2008).
60. Khaled, A. R. A., Vafai, K., Yang, M., Zhang, X., & Ozkan, C. S. Analysis, control and augmentation of microcantilever deflections in bio-sensing systems. *Sensors Actuators, B Chem.* **94**, 103–115 (2003).
61. Visser, C. W. et al. Quantifying cell adhesion through impingement of a controlled microjet. *Biophys. J.* **108**, 23–31 (2015).
62. Bayouh, S., Ponnsonnet, L., Ouada, H. B., Bakhruf, A. & Othmane, A. Bacterial detachment from hydrophilic and hydrophobic surfaces using a microjet impingement. *Colloids Surf. A: Physicochem. Eng. Asp.* **266**, 160–167 (2005).
63. Phares, D. J., Smedley, G. T. & Flagan, R. C. The wall shear stress produced by the normal impingement of a jet on a flat surface. *J. Fluid. Mech.* **418**, 351–375 (2000).
64. Weisel, J. W., Shuman, H. & Litvinov, R. I. Protein – protein unbinding induced by force: single-molecule studies. *Curr. Opin. Struct. Biol.* **13**, 227–235 (2003).
65. Yeh, P. Y., Jayachandran, J. K., N. Madden, J. D. & Chiao, M. Electric field and vibration-assisted nanomolecule desorption and anti-biofouling for biosensor applications. *Colloids Surf. B.59*, 67–73 (2007).
66. Meyer, G. D., Moran-Mirabal, J. M., Branch, D. W. & Craighead, H. G. Nonspecific binding removal from protein microarrays using thickness shear mode resonators. *IEEE Sens. J.* **6**, 254–261(2006).
67. Wang, X., Engel, J. & Liu, C. Liquid crystal polymer for MEMS: processes and applications. *J. Micromech. Microeng.* **13**, 628–633 (2003).
68. Wang, K., Liu, C.-c., Member, S. & Durand, D. M. Oxide sputtered on liquid crystal polymer. *IEEE Trans. Biomed. Eng.* **56**, 6–14 (2009).
69. Chow, E. Y., Chlebowski, A. L. & Irazoqui, P. P. A miniature-implantable RF-wireless active glaucoma intraocular pressure monitor. *IEEE Trans. Biomed. Circuits Syst.* **4**, 340–349 (2010).
70. Min, K. S. et al. A liquid crystal polymer-based neuromodulation system: an application on animal model of neuropathic pain. *Neuromodulation* **17**, 160–169 (2014).

71. Yu, L., Kim, B. & Meng, E. Chronically implanted pressure sensors: challenges and state of the field. *Sensors* **14**, 20620–20644 (2014).
72. Park, H., John, S. & Lee, H. Low-cost rapid prototyping of liquid crystal polymer based magnetic microactuators for glaucoma drainage devices. *Proc. Annu. Int. Conf. IEEE Eng. Med. Biol. Soc.* **2016**, 4212–4215 (2016).
73. Dunham, J. S., Musgraves, J. D., Close, B. T. & Tanenbaum, D. M. A maskless photolithographic prototyping system using a low-cost consumer projector and a microscope. *Am. J. Phys.* **73**, 980–990 (2005).
74. Horiuchi, T., Koyama, S. & Kobayashi, H. Microelectronic engineering simple maskless lithography tool with a desk-top size using a liquid-crystal-display projector. *Microelectron. Eng.* **141**, 37–43 (2015).
75. Li, Y. et al. Rapid fabrication of microfluidic chips based on the simplest LED lithography. *J. Micromech. Microeng.* **055020**, 1–7 (2015).
76. Leonard, B. P. Based on quadratic upstream interpolation. *Comput. Methods Appl. Mech. Eng.* **19**, 59–98 (1979).
77. Alexandre, B. & Chorin, J. Numerical solution of the Navier-Stokes equations. *Math. Comput.* **22**, 745–762 (1968).
78. Ardekani, A. M., Dabiri, S. & Rangel, R. H. Collision of multi-particle and general shape objects in a viscous fluid. *J. Comput. Phys.* **227**, 10094–10107 (2008).
79. Zhang, Y. & Li, A. G. Reduced viscosity for flagella moving in a solution of long polymer chains. *Phys. Rev. Fluids* **3**, 023101 (2018).
80. Kim, Y. K., Chen, E. Y. & Liu, W. F. Biomolecular strategies to modulate the macrophage response to implanted materials. *J. Mater. Chem. B* **4**, 1600–1609 (2015).