

Design, Fabrication and Testing of a Piezoresistive-Based Tactile Sensor for Minimally Invasive Surgery

Ahmad Atieh

Department of Mechanical and Industrial Engineering
Concordia University
Montreal, Canada
a_atieh@encs.concordia.ca

Javad Dargahi

Department of Mechanical and Industrial Engineering
Concordia University
Montreal, Canada
dargahi@encs.concordia.ca

Abstract—Minimally invasive robotic surgery (MIRS) has become a preferred method for patients for the last two decades, thanks to its crucial advantages over classical open surgeries. Although MIRS has some advantages, it has a few drawbacks. Since MIRS technology includes performing surgery through small incisions using long slender tools, one of the main drawbacks of MIRS becomes the loss of direct contact with the patient's body in the site of operation. Therefore, the surgeon loses the sense of touch during the operation which is one of the important tools to investigate the health condition of the treated tissue. This paper presents a novel piezoresistive-based multifunctional tactile sensor that is able to measure the contact force, relative hardness of the contact object, and the position of a hidden lump at the same time. A prototype of the designed sensor has been fabricated, simulated, analyzed, and tested numerically and experimentally. The experiments have been performed on hyperelastic materials, which are silicone rubber samples with different hardness values that resemble different biological tissues. The ability of the sensor to measure the contact force and relative hardness of the contact objects is tested with several experiments. A finite element (FE) model has been built in COMSOL Multiphysics (v3.4) environment to simulate both the mechanical behavior of the silicone rubber samples, and the interaction between the sensor and the silicone rubbers. Both numerical and experimental analysis proved the capability of the sensor to distinguish between different silicone rubber samples.

Index Terms—Minimally invasive surgery/devices, Tactile sensor, Tissue modeling/contact, Hyperelastic material modeling, Piezoresistive films.