



# University of Groningen

# RAS testing in metastatic colorectal cancer

Boleij, Annemarie; Tops, Bastiaan B. J.; Rombout, Paul D. M.; Dequeker, Elizabeth M.; Ligtenberg, Marjolijn J. L.; van Krieken, J. Han; Dutch RAS EQA Initiative; Elst, ter, Arja; Schuuring, Ed; Diepstra, Arjan

Published in: Oncotarget

10.18632/oncotarget.3804

IMPORTANT NOTE: You are advised to consult the publisher's version (publisher's PDF) if you wish to cite from it. Please check the document version below.

Document Version Publisher's PDF, also known as Version of record

Publication date: 2015

Link to publication in University of Groningen/UMCG research database

Citation for published version (APA):
Boleij, A., Tops, B. B. J., Rombout, P. D. M., Dequeker, E. M., Ligtenberg, M. J. L., van Krieken, J. H., & Dutch RAS EQA Initiative (2015). RAS testing in metastatic colorectal cancer: excellent reproducibility amongst 17 Dutch pathology centers. Oncotarget, 6(17), 15681-15689. https://doi.org/10.18632/oncotarget.3804

Other than for strictly personal use, it is not permitted to download or to forward/distribute the text or part of it without the consent of the author(s) and/or copyright holder(s), unless the work is under an open content license (like Creative Commons).

Take-down policy

If you believe that this document breaches copyright please contact us providing details, and we will remove access to the work immediately and investigate your claim.

Downloaded from the University of Groningen/UMCG research database (Pure): http://www.rug.nl/research/portal. For technical reasons the number of authors shown on this cover page is limited to 10 maximum.

Download date: 12-11-2019

# RAS testing in metastatic colorectal cancer: excellent reproducibility amongst 17 Dutch pathology centers

Annemarie Boleij<sup>1</sup>, Bastiaan B.J. Tops<sup>2</sup>, Paul D.M. Rombout<sup>1</sup>, Elizabeth M. Dequeker<sup>3</sup>, Marjolijn J.L. Ligtenberg<sup>1,2</sup>, J. Han van Krieken<sup>1</sup> and Dutch RAS EQA Initiative<sup>4,5,6,7,8,9,10,11,12,13,14,15,16,17,18,19,20</sup>

Correspondence to: J.Han van Krieken, email: Han.vanKrieken@radboudumc.nl

Keywords: RAS, colorectal cancer, metastasis, quality control, next generation sequencing

Received: February 09, 2015 Accepted: March 18, 2015 Published: April 12, 2015

This is an open-access article distributed under the terms of the Creative Commons Attribution License, which permits unrestricted use, distribution, and reproduction in any medium, provided the original author and source are credited.

# **ABSTRACT**

In 2013 the European Medicine Agency (EMA) restricted the indication for anti-EGFR targeted therapy to metastatic colorectal cancer (mCRC) with a wild-type RAS gene, increasing the need for reliable RAS mutation testing. We evaluated the completeness and reproducibility of RAS-testing in the Netherlands.

From 17 laboratories, tumor DNA of the first 10 CRC cases tested in 2014 in routine clinical practice was re-tested by a reference laboratory using a custom next generation sequencing panel. In total, 171 CRC cases were re-evaluated for hotspot mutations in *KRAS*, *NRAS* and *BRAF*.

Most laboratories had introduced complete RAS-testing (65%) and BRAF-testing

Department of Pathology, Radboud University Medical Center, Nijmegen, The Netherlands

<sup>&</sup>lt;sup>2</sup> Department of Human Genetics, Radboud University Medical Center, Nijmegen, The Netherlands

<sup>&</sup>lt;sup>3</sup> Department of Public Health and Primary Care, Biomedical Quality Assurance Research Unit, KU Leuven - University of Leuven, Leuven, Belgium

<sup>&</sup>lt;sup>4</sup>C.J.M. van Noesel, Academic Medical Center (AMC), Department of Pathology, Amsterdam

<sup>&</sup>lt;sup>5</sup> C.C. Scheidel-Jacobse (Technical specialist), J.A. Kummer (Pathologist/KMBP), P. Roepman (KMBP in training), St. Antonius Ziekenhuis, Department of Pathology, Nieuwegein

<sup>&</sup>lt;sup>6</sup> C.F.M. Prinsen (KMBP), S.H.M. van den Berg-van Erp (Pathologist), Canisius Wilhelmina Ziekenhuis (CWZ), Department of Pathology, Nijmegen

<sup>&</sup>lt;sup>7</sup> J.M.H.H. van Gorp, Diakonessenhuis, Laboratory for Pathology, Utrecht

<sup>&</sup>lt;sup>8</sup> P.M. Nederlof, Dutch Cancer Institute (NKI), Amsterdam.

<sup>&</sup>lt;sup>9</sup> E. Caspers, St. Elisabeth Ziekenhuis, Department of Molecular Pathology, Tilburg

<sup>&</sup>lt;sup>10</sup> W.N.M. Dinjens, E.C.W. Beerens, Erasmus MC, Department of Pathology, Molecular Diagnostics, Rotterdam

<sup>&</sup>lt;sup>11</sup> N.A. 't Hart, Isala, Department of Pathology, Zwolle

<sup>12</sup> A.J.C. van den Brule, Jeroen Bosch Ziekenhuis, Molecular Diagnostics, 's-Hertogenbosch

<sup>13</sup> R. van der Geize (KMBP), S.A. Riemersma (Pathologist), Laboratory for Pathology Oost-Nederland (LABPON), Hengelo

<sup>&</sup>lt;sup>14</sup>T. van Wezel (KMBP), H. Morreau (Pathologist), R. van Eijk (Technical specialist), Leiden University Medical Center (LUMC), Department of Pathology, Leiden

<sup>&</sup>lt;sup>15</sup> J.W.M. Jeuken, Laboratory for Pathology and Medical Microbiology (PAMM), Eindhoven

<sup>&</sup>lt;sup>16</sup> A. Dirkx, Pathan B.V., Molecular Diagnostics, Rotterdam

<sup>&</sup>lt;sup>17</sup> M. Klomp, Rijnstate Ziekenhuis, Department of Pathology, Arnhem

<sup>18</sup> W.T.M van Blokland, University Medical Center (UMC) Utrecht, Molecular Pathology, Utrecht

<sup>&</sup>lt;sup>19</sup> A. ter Elst (Technical specialist/KMBP in training), E. Schuuring (KMBP), A. Diepstra (Pathologist), University Medical Center Groningen (UMCG), Department of Pathology, Groningen

<sup>&</sup>lt;sup>20</sup> D.A.M. Heideman (KMBP), N.C.T. van Grieken (Pathologist), D. Sie (Technical specialist), VU-University Medical Center (VUMC), Department of Pathology, Amsterdam

(71%) by January 2014. The most employed method for all hotspot regions was Sanger sequencing (range 35.7 – 49.2%). The reference laboratory detected all mutations that had been found in the participating laboratories (n=92), plus 10 additional mutations. This concerned three *RAS* and seven *BRAF* mutations that were missed due to incomplete testing of the participating laboratory. Overall, the concordance of tests performed by both the reference and participating laboratory was 100% (163/163;  $\kappa$ -static 1.0) for *RAS* and 100% (144/144;  $\kappa$ -static 1.0) for *BRAF*.

Our study shows that RAS and BRAF mutations can be reproducibly assessed using a variety of testing methods.

# **INTRODUCTION**

Anti-Epithelial Growth Factor Receptor (EGFR) targeted therapy, such as panitumumab and cetuximab, is effectively reducing the risk of tumor progression and improving overall survival (OS), progression free survival (PFS) and quality of life in metastatic colorectal cancer (mCRC) patients whose tumor is RAS wild-type [1, 2]. Conversely, mCRC patients with mutated RAS tumors (KRAS and NRAS mutations) who received panitumumab in combination with oxaliplatin-based chemotherapy had a significantly worse outcome in OS and PFS [3]. This notion has led to a new incentive by the European Medicine Agency (EMA) for Vectibix (panitumumab) in June 2013 [4] and for Erbitux (cetuximab) in November 2013 [5], to indicate administration of EGFR targeted therapy only to patients with wild-type RAS mCRC. This has increased the need for reliable RAS mutation testing methods to assure the quality of *RAS* status determination.

Most molecular testing methods that are used nowadays accurately assess mutational status of RAS genes in samples with >30-50% tumor cells, or alternatively with 15-25% of mutated alleles in the test sample [6, 7]. However, with lower number of mutated alleles in the sample, the limit of accurate detection of a method declines depending on the test method used [8]. Even when using the same method, differences in protocols between laboratories can result in different outcomes. It has been suggested that the reproducibility amongst different testing methods is not as high as anticipated for based on previous EQA schemes for KRAS exon 2 testing[9]. In a recent study, in 29 out of 182 KRAS exon 2 wild-type tumors (15.9%), as assessed with Sanger sequencing, a KRAS exon 2 mutation was found with next generation sequencing (NGS) [10]. This suggests a higher variability in reproducibility between test methods and laboratories than initially measured [11].

Given the clinical impact of *RAS*-testing, it is of utmost importance to control for reliable performance of routine *RAS*-testing methods used in clinical practice. In this study, we evaluated the inter-laboratory agreement of *RAS*-testing amongst 171 mCRC patients of 17 Dutch laboratories.

# **RESULTS**

## **Integration of full RAS-testing in the Netherlands**

In January 2014, 11 of 17 participating Dutch laboratories had introduced full *RAS* testing (65%). All 17 participating centers performed *KRAS* exon 2 (codon 12 and 13) and *KRAS* exon 3 (codon 61) testing. The method most frequently used for *KRAS* exon 2 was Sanger sequencing of PCR products, either directly (5 laboratories; 51 samples (29.8%)) or to specify the mutation detected with high resolution melting (HRM) analysis or to confirm a real-time PCR result (7 laboratories; 70 samples (40.9%)) [Table 1].

Of the 6 laboratories that had not introduced full *RAS*-testing, one laboratory did not test for *NRAS* exon 2, 3, 4 and *KRAS* exon 4, three had not introduced *NRAS* exon 4 (codon 117 and 146), one laboratory did not have *KRAS* exon 4 codon 117 in their test panel and one laboratory did not test for KRAS exon 3 codon 59 [see Table 1]. One of these 6 laboratories introduced full *RAS* testing early 2014 and had only tested 3 out of 10 (30%) CRC cases with full *RAS*-testing.

For full *RAS*-testing direct Sanger sequencing of PCR products was used most often either with or without a prescreen (range 35.7% - 49.2%) [Table 1]. The use of Mass Spectrometry (Sequenom) or CE-IVD kits (Therascreen) was reported by 2 laboratories (11.6%). In 3 laboratories (17.5%) next generation sequencing (NGS) with MiSeq (Illumina) or Ion Torrent (Life technologies) was used for *RAS* testing. In conclusion, full *RAS* testing had been introduced in the majority of laboratories participating in this study and relied mostly on Sanger sequencing methods.


# Mutation frequencies of KRAS, NRAS and BRAF

The amount and quality of 167 out of 171 received DNA samples was sufficient for successful evaluation by the reference laboratory using NGS of at least one of the target sites (165 samples for KRAS, 163 for NRAS and 163 for BRAF; see materials and methods). Overall, in 102 samples (61.1% (95% CI 53.5 – 68.2)) a mutation

Table 1: Methods used to test for KRAS, NRAS and BRAF mutations.

·	codon	codon	KRAS codon	codon	codon	codon	codon	NRAS codon	codon	codon	BRAF codon
	12/13	59	61	117	146	12/13	59	61	117	146	600
Method <sup>†</sup>	N (%)	N (%)	N (%)	N (%)	N (%)	N (%)	N (%)	N (%)	N (%)	N (%)	N (%)
Next generation sequencing											
	10	10	10	10	10	10	10	10	0	0	10
Illumina MiSeq Oncopanel (Illumina)[25]	(5.8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(0,0)	(0,0)	(5,8)
	20	20	20	20	20	20	20	20	10	20	20
IonTorrent (Life Sciences)	(11,7)	(11,7)	(11,7)	(11,7)	(11,7)	(11,7)	(11,7)	(11,7)	(5,8)	(11,7)	(11,7)
Sequencing											
	51	61	61	73	84	74	84	74	83	73	51
Sanger Sequencing	(29,8)	(35,7)	(35,7)	(42,7)	(49,2)	(43,3)	(49,2)	(43,3)	(48,6)	(42,7)	(29,8)
	60	60	60	30	30	30	30	30	0	0	40
HRM + Sanger Sequencing	(35,1)	(35,1)	(35,1)	(17,5)	(17,5)	(17,5)	(17,5)	(17,5)	(0,0)	(0,0)	(23,4)
	10	0	0	0	0	10	0	10	0	0	10
Real-time PCR + Sanger sequencing	(5,8)	(0,0)	(0,0)	(0,0)	(0,0)	(5,8)	(0,0)	(5,8)	(0,0)	(0,0)	(5,8)
	0	0	0	0	0	0	0	0	10	10	0
Pyrosequencing	(0,0)	(0,0)	(0,0)	(0,0)	(0,0)	(0,0)	(0,0)	(0,0)	(5,8)	(5,8)	(0,0)
Other assays											
Therascreen KRAS/NRAS/BRAF Pyrokit	10	0	10	0	0	10	0	10	0	0	10
(Qiagen)	(5,8)	(0,0)	(5,8)	(0,0)	(0,0)	(5,8)	(0,0)	(5,8)	(0,0)	(0,0)	(5,8)
Therascreen RAS extension Pyrokit	0	10	0	10	10	0	10	0	10	10	
(Qiagen)	(0,0)	(5,8)	(0,0)	(5,8)	(5,8)	(0,0)	(5,8)	(0,0)	(5,8)	(5,8)	
	10	0	10	10	10	10	10	10	10	10	10
Sequenom MassArray (Sequenom)	(5,8)	(0,0)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)	(5,8)
	171	161	171	153	164	164	164	164	123	123	151
Total samples tested	(100)	(94,2)	(100)	(89,5)	(95,9)	(95,9)	(95,9)	(95,9)	(71,9)	(71,9)	(88,3)*

Total 171 samples received from 17 labs, 10 samples per lab, 1 lab sent 11 DNA samples; \* 4 laboratories only test *BRAF* on request by the physician. Note: for each codon, the most frequently used method is highlighted in bold



**Figure 1: Mutation prevalence of** *KRAS*, *NRAS* and *BRAF*. A) Pie chart showing all evaluated CRC cases (n = 167; 167 of 171 samples could be evaluated for at least one of the target sites). Of these CRC cases, 61.1% had a mutation. *KRAS* mutations alone were most frequently observed (43.0%). *BRAF* mutations were found in 23 CRC cases (All in exon 15; c.1799T>A (p.Val600Glu)). The 95% CI was calculated with Jeffry's method. **B**) A total of 71 *KRAS* mutations was found; 59 in *KRAS* exon 2 (83.2%), 6 in *KRAS* exon 3 (8.4%), and 6 in *KRAS* exon 4 (8.4%). Gly12Val (29.6%) and Gly12Asp (23.9%) were the most common mutations found followed by Gly13Asp (12.7%). C) A total of 8 NRAS mutations was detected of which 2 were detected in exon 2 (Gly12Asp) and 6 in exon 3. The most common mutation in *NRAS* was Gln61Arg (37.5%).

Table 2: Additional mutations found by the reference laboratory.

Gene	Exon	N samples	Amino acid change	Reason not found by participating laboratory
KRAS	4	1	p.Ala146Thr	not in test repertoire
NRAS	2	1	p.Gly12Asp	not in test repertoire
NRAS	3	1	p.Gln61Lys	not in test repertoire
BRAF	15	4	p.Val600Glu	BRAF tested only on request by physician
BRAF	15	3	p.Val600Glu	not in test repertoire

in *KRAS*, *NRAS* or *BRAF* was found at the reference laboratory [Figure 1A].

A KRAS exon 2 mutation was detected in 59 mCRC cases (35.8%), other RAS mutations were found in KRAS exon 3 (n = 3 (3.6%)), KRAS exon 4 (n = 3 (3.6%)) NRAS exon 2 (n = 1 (1.2%)) and NRAS exon 3 (n = 3 (3.6%)). No mutations in NRAS exon 4 were detected. This resulted in a total 79 RAS mutations (47.6% (95% CI 40.1-55.2)) of which only 8 were NRAS mutations. The mean percentage of mutated alleles was 42.1% (SD 15.7%) for KRAS and 46.1% (SD 23.25%) for NRAS [Supplementary Table 1]. The majority of KRAS mutations affected codon 12 (70.5%), especially p.Gly12Asp (23.9%) and p.Gly12Val were common (29.6%) [Figure 1B]; 5 out of 8 NRAS mutations were found in codon 61 (62.5%) [Figure 1C]. None of the samples harbored both or more than one KRAS and/or NRAS mutation.

*BRAF* mutations occurred in 23 of 163 CRC cases (14.1% (95%CI 9.4-20.1)) [Figure 1A] and were all c.1799T>A (p.Val600Glu) mutations in exon 15. The mean percentage of mutated alleles for *BRAF* was 34.5% (SD 14.3%) [Supplementary Table 1]. Both *KRAS* and *NRAS* mutations in our sample were mutually exclusive with *BRAF* mutations (OR =2.15 (95%CI 1.80-2.57); Chisquare p<0.01).

# High concordance of RAS and BRAF-testing

For 167 samples that could be evaluated by the reference laboratory, all mutations that had been found in the participating laboratories were verified. In addition, the reference lab detected three RAS and seven BRAF mutations that had not been detected in the participating laboratories. This concerned mutations in KRAS exon 4 (c.436G>A; p.Ala146Thr), NRAS exon 2 (c.35G>A; p.Gly12Asp), NRAS exon 3 (c.181C>A; p.Gln61Lys) and seven mutations in BRAF (c.1799T>A; p.Val600GLu) that the participating laboratory had not tested for [Table 2]. Altogether, the inter-laboratory agreement for tests performed by both the reference and participating laboratory was 100% (163/163) (k statistic 1.0) for full RAS and 100% (144/144) ( $\kappa$  statistic 1.0) for BRAF-testing [Table 3]. All three additionally found RAS-mutations were reported back to the respective participating laboratory, and were confirmed with their newly installed Sanger sequencing or NGS approaches for the respective target sites.

# Association between percentage of neoplastic cells and mutation detection

Testing sensitivity of RAS-mutations is dependent on the percentage of neoplastic cells represented in the test sample; low neoplastic cell percentages may result in missing of RAS-mutations depending on the technique that is used[8]. For 158 samples the percentage of neoplastic cells as estimated by the pathologist of the participating laboratory was known. The median estimated neoplastic cell percentage represented in the DNA samples was 50% (inter quartile range (IQR)=30). Of the 158 samples, 47 (29.7%) had neoplastic cell percentages below 40%. When comparing the distribution of estimated neoplastic cell percentages amongst RAS-mutated (N = 73) and RAS-wild-type (N = 85) samples, the median neoplastic cell percentage was significantly lower in RAS-mutated samples (50.0% (IQR=28) compared to RAS-wild type (60.0% (IQR=20)) samples (Mann-Whitney U, p < 0.01) [Figure 2]. In RAS-mutated samples, the mean percentage of mutated alleles correlated with the percentage of neoplastic cells in the test sample (Pearson r = 0.433 (p < 0.01) for KRAS and Pearson r = 0.792 (p = 0.034) for NRAS; Supplementary figure 1).

The frequency of RAS mutations amongst samples with neoplastic cell percentages  $\leq$ 40% was significantly higher than in samples >40% neoplastic cells (OR 2.45 (95%CI 1.22 – 4.94), chi-square p=0.011). A similar result was obtained when lowering the cut-off to 30 or 20% neoplastic cells, suggesting that the sensitivity of the RAS mutation analyses were adequate, even in samples with low neoplastic cells. However, care should be taken while reporting wild-type in samples with low neoplastic cells depending on the lower limit of detection of the technique applied to prevent reporting false-negatives.

#### **DISCUSSION**


Our results show an excellent concordance (100%, kappa 1.0) of *RAS*-test results of the reference lab oratory and 17 other Dutch laboratories in routine clinical practice, despite differences in testing methods used. Overall, full *RAS*-testing of *KRAS* exon 2, 3 and 4 and *NRAS* exon 2, 3 and 4 has been introduced in the majority of the participating laboratories (65%). Three *RAS* mutations and seven *BRAF* mutation were missed due to incomplete testing.

Table 3: Inter-laboratory agreement between participating and reference laboratory.

			Wildt	ype	Muta	ited	Inter-laboratory agreement <sup>†</sup>		
	Samples	N (%) <sup>‡</sup>	participating laboratory	reference laboratory	participating laboratory	reference laboratory	concordance	k-statistic	
RAS	163*	(95,3)	87	87	76	76	100%	1,0	
KRAS	164	(95,9)	94	94	70	70	100%	1,0	
exon 2	165	(96,5)	106	106	59	59	100%	1,0	
exon 3	165	(96,5)	159	159	6	6	100%	1,0	
exon 4	159	(93,0)	154	154	5	5	100%	1,0	
NRAS	157	(91,8)	151	151	6	6	100%	1,0	
exon 2	157	(91,8)	156	156	1	1	100%	1,0	
exon 3	157	(91,8)	152	152	5	5	100%	1,0	
exon 4	117	(68,4)	117	117	0	0	100%	1,0	
BRAF	144**	(84,2)	128	128	16	16	100%	1,0	

wild-type and mutated as found by the reference laboratory

<sup>\*\* 7</sup> of the 151 samples tested for *BRAF* by the participating laboratory could not be evaluated in the reference laboratory due to low read coverage as result of low or poor DNA quality.



**Figure 2: Neoplastic cell percentage in** *RAS***-mutated and** *RAS***-wild-type tumors.** The distribution of the estimated percentage of tumor cells in the test sample was compared between *RAS*-mutated and *RAS*-wild type CRC cases. Frequency of CRC cases is shown on the Y-axis; the X-axis represents the percentage of neoplastic cells. Mann-Whitney U test of the comparison indicates that the median neoplastic cell percentage of *RAS*-mutated mCRC cases is significantly lower (Median 50 (IQR 28), Mean Rank = 66,65) than RAS-wild-type mCRC cases (Median 60 (IQR 20), Mean Rank = 90,54; p = 0.001) and thus that the populations have distinct neoplastic cell percentage distributions.

<sup>\*</sup>Samples that were not included could not be evaluated due to low DNA concentration or low read coverage, or were not tested by both the reference laboratory and the participating laboratory

<sup>\*</sup> in 163 of the 171 samples at least 1 of the target exons could be evaluated in the reference laboratory that was also tested by the participating laboratory. The following samples that were tested by the participating laboratory could not be evaluated in the reference laboratory due to low read coverage as result of low or poor DNA quality: *KRAS* exon 2: 6 of the 171 samples tested; *KRAS* exon 3: 6 of the 171 samples tested; *KRAS* exon 4; 5 of the 164 samples tested; *NRAS* exon 2; 7 of the 164 samples tested; *NRAS* exon 3; 7 of the 164 samples tested; *NRAS* exon 4; 6 of the 123 samples tested.

The frequency of RAS mutations reported in our study matches those previously reported in the literature. Approximately 35-45% of all CRCs contain RAS-mutations in KRAS exon 2 [3, 12–16], whereas approximately 10% of RAS-mutations occurs in KRAS exons 3 and 4 or NRAS exons 2, 3 and 4 [3, 17–19]. In addition, KRAS, NRAS and BRAF mutations have a strong tendency towards mutual exclusivity [20]. BRAF mutations occurred in 14% of our samples, which seems higher than the reported 9% in the Cancer Genome Atlas [20]. While BRAF mutation analysis is incorporated in the test repertoire of most laboratories (71%), some laboratories only test for BRAF on special request by the physician when tumors were previously tested RAS wildtype. BRAF mutation frequency amongst KRAS exon 2 wild-type tumors is reported to be around 8-15% [10, 18].

To assure accurate determination of RAS wildtype and mutant status, the quality of mutation detection for each KRAS and NRAS exon needs to be determined. Reproducibility is one of the measures that signify the quality of diagnostic tests. Poor reproducibility can have several causes: 1) the testing methods used have a difference in sensitivity (ability to identify tumors with the mutation) or, 2) there is variation between different persons/laboratories performing the test (inter-observer variability). In a recently published Italian study, it was found that amongst KRAS exon 2 wild-type tumors, as assessed with Sanger sequencing, a KRAS exon 2 mutation was found with NGS in 28 out of 182 mCRC cases (15.9%). The difference in sensitivity between the testing methods could have partly accounted for the discrepancy found in this study [10]. In our study the discrepancy between the test results generated in 17 different laboratories using a variety of testing methods and the reference laboratory using a NGS approach was naught, even in samples with a neoplastic cell percentage between 20-40% and in laboratories using Sanger sequencing. In the 2012 KRAS external quality assessment program of the European Society of Pathology, less than 5% of all samples were wrongly genotyped for KRAS amongst 100 laboratories in 26 countries[11].

Testing sensitivity is reflected by the limit of detection of the method, but is also limited by the percentage of neoplastic cells represented in the test sample. In fact, low neoplastic cell percentage (~10%) led to 16 of the 29 false-negatives in the 2012 KRAS EQA scheme[11]. For PCR and Sanger sequencing a minimum amount of 20-30% of tumor cells is required [21]. Our study has high reproducibility even in samples with a low neoplastic cell percentage and the percentage of mutated alleles correlates with the neoplastic cell content in the sample [Supplementary figure 1]. In fact, the percentage of RAS mutant samples was significantly higher in samples with neoplastic cell percentages below 40%, which demonstrates that mutations could also be detected in samples with a low neoplastic cell percentage. However,

this unexpected observation could be merely coincidental due to small subgroups and difference in tumor sampling in combination with variation in estimating the amount of neoplastic cells, which is known to be high among pathologists [22, 23].

Potential sources of bias in our study could have affected the reported reproducibility. First, centers were asked to send DNA of the first 10 CRC cases tested in 2014. Because KRAS exon 2 mutations are the most frequent, there were some laboratories where only KRAS exon 2 mutations or wild-type RAS-status was reported. This could have led to an overestimation of the reproducibility because other RAS-mutations were not present. Secondly, the likelihood of finding falsenegatives is higher when only wild-type samples are re-tested. In this series only 54 samples were wild-type for all the tested hotspot mutations and thus could have led to an underestimation of false-negative wild-type samples. The small number of NRAS-mutations found has likely overestimated the reproducibility of this group of mutations. Future studies need to verify whether newly installed methods reproducibly detect NRAS mutations, especially in laboratories that did not report any NRAS mutations in the tested CRC panel.

In conclusion, our study clearly shows that *RAS*-status can be reproducibly assessed between laboratories in routine clinical practice using similar or different testing methods. With constant improvement of testing methodologies and quality controls, this offers good expectations for the future of molecular testing in mCRC. Nevertheless, more in depth analyses with regard to the effect of testing sensitivity and the percentage of tumor cells in the test sample on *RAS* mutation detection are warranted.

#### PATIENTS AND METHODS

# Patient selection and data collection

All 22 Dutch institutes that participated in the European Quality Assurance (EQA) scheme of the European society of pathology (ESP) in 2013 were invited to participate in the study. Seventeen laboratories responded and were subsequently requested to send 10  $\mu$ l DNA ( $\geq$ 30 ng) of the first 10 mCRCs tested in routine diagnostics for anti-EGFR targeted therapy from 1st January 2014. In addition, each laboratory was asked to indicate the testing method used, the mutations tested for and the percentage of neoplastic cells in the tissue the DNA was extracted from. Moreover, the *KRAS*, *NRAS* and *BRAF* mutations found by the participating laboratories were reported and stored by a third party until study end. Retesting of the samples at the reference laboratory was performed blinded. Approval by a medical ethics

committee was not required. All data were reported in deidentified form and are in agreement with the Dutch Data Protection Act.

# **Next generation sequencing**

A total of 171 DNA samples received from the participating laboratories (one lab contributed 11 samples) were quantified on the Qubit platform (Life technologies). Samples with a DNA concentration below 0.5 ng/µl were excluded from further analysis. Bar-coded libraries were prepared from 10-100 ng DNA using a custom AmpliSeq panel targeting frequently mutated regions in the KRAS, NRAS, EGFR, PIK3CA, ERBB2, AKT1, BRAF genes and the AMELX/Y gene as a control gene to determine the sex of the patient. Libraries were equimolar pooled and clonal amplification was performed by emulsion PCR using the One Touch 2 system (Life Technologies) and subsequently run on the Ion Torrent Personal Genome Machine (Life Technologies). Torrent Suite Software v.3.4.2. was used to pre-process the raw data. Subsequent mapping and variant calling was performed using SeqNext software v.4.1.2. (JSI medical systems GmbH). The following genes and exons were included in the data analysis: KRAS exon 2, 3 and 4 (NM 004985.3), NRAS exon 2,3 and 4 (NM 002524.3) and BRAF exon 15 (NM 004333.4). The minimum read coverage allowed to call mutant alleles was set at 100 reads. In general, read coverage was more than 500 resulting in a sensitivity of 5% mutant alleles. For a read coverage between 100-500 reads, a limit of 10% mutant alleles was employed. When coverage was low or mutation calling could not be established due to other technical reasons, the run was repeated. When there was still insufficient coverage of the target sites after repeating the run, the samples were excluded from analysis. In total 6 KRAS, 8 NRAS and 8 BRAF target sites had insufficient coverage due to low amount or poor DNA quality. Therefore, we could successfully evaluate a total of 165 KRAS, 163 NRAS and 163 BRAF target sites; for 167 samples at least one of the target sites could be successfully evaluated.

# Data analysis

Variants were filtered for known SNPs and systematic sequencing artifacts. Next, the somatic mutations found in the DNA-samples by the reference laboratory were compared with the mutations originally found by the participating laboratories to assess the percentage of agreement of the results. The mutations were compared at the genotype level. Each participating pathology center was informed about their individual results and performance.

# Statistical analysis

All statistical analyses were performed in IBM SPSS statistics version 20. Mutation frequencies were calculated for KRAS, NRAS and BRAF. Binominal confidence intervals were calculated using the Jeffrey's interval. The detected mutation(s) in the DNA-samples at the genotype level were compared between the reference laboratory and the participating laboratories and expressed as percentage agreement. Data were evaluated for hotspot mutations in KRAS exon 2, 3 and 4, NRAS exon 2, 3 and 4 and BRAF exon 15. Concordance (inter-laboratory agreement, kappa-statistic[24]) was only calculated when the mutation was targeted in the test panel of both the reference and participating laboratory. Odds ratios with 95% confidence intervals and chi-square statistics were calculated to evaluate associations between RAS-mutations and BRAF. The distribution of tumor cell percentage in the DNA-samples amongst RAS-mutated and RAS-wild-type samples was assessed with the Mann-Whitney U test.

# **ACKNOWLEDGMENTS**

We thank all the laboratories that participated in this study for their cooperation and fruitful discussions on the topic. We thank Véronique Tack from the University of Leuven for communication with and recruitment of the laboratories.

#### **FUNDING**

This study was funded by MERCK Group KGaA. The source of funding did not have any influence on the design and the analysis of the results.

#### **DISCLOSURE**

E. Schuuring received honoraria for expertise from commercial partners, is part of scientific advisory boards from Amgen, Abbott, Roche and Pfizer, and received travel reimbursements (on invitation) from Abbott and Roche. T. van Wezel is part of the scientific advisory board of Amgen. N 't Hart is part of the scientific advisory boards of Amgen, Pfizer, Astra and Roche. Han van Krieken is part of the scientific advisory boards of Amgen and Merck KGaA. All remaining authors have declared no conflict of interest.

#### REFERENCES

 Stintzing S, Kapaun C, Laubender RP, Jung A, Neumann J, Modest DP, Giessen C, Moosmann N, Wollenberg A, Kirchner T, Heinemann V. Prognostic value of cetuximabrelated skin toxicity in metastatic colorectal cancer patients and its correlation with parameters of the epidermal growth

- factor receptor signal transduction pathway: results from a randomized trial of the GERMAN AIO CRC Stu. Int. J. Cancer 2013; 132: 236–245.
- Schwartzberg LS, Rivera F, Karthaus M, Fasola G, Canon J-L, Hecht JR, Yu H, Oliner KS, Go WY. PEAK: a randomized, multicenter phase II study of panitumumab plus modified fluorouracil, leucovorin, and oxaliplatin (mFOLFOX6) or bevacizumab plus mFOLFOX6 in patients with previously untreated, unresectable, wild-type KRAS exon 2 metastatic colorectal . J. Clin. Oncol. 2014; 32: 2240–2247.
- Douillard J-Y, Oliner KS, Siena S, Tabernero J, Burkes R, Barugel M, Humblet Y, Bodoky G, Cunningham D, Jassem J, Rivera F, Kocákova I, Ruff P, Błasińska-Morawiec M, Šmakal M, Canon JL, Rother M, Williams R, Rong A, Wiezorek J, Sidhu R, Patterson SD. Panitumumab-FOLFOX4 treatment and RAS mutations in colorectal cancer. N. Engl. J. Med. 2013; 369: 1023–1034.
- European Medicines Agency. Vectibix Panitumumab. 2013.Available from: http://www.ema.europa.eu/ema/ index.jsp?curl=pages/medicines/human/medicines/000741/ human med 001128.jsp&mid=WC0b01ac058001d124.
- European Medicines Agency. Erbitux Cetuximab. 2013. Available from: http://www.ema.europa.eu/ema/index. jsp?curl=pages/medicines/human/medicines/000558/ human med 000769.jsp&mid=WC0b01ac058001d124.
- Querings S, Altmüller J, Ansén S, Zander T, Seidel D, Gabler F, Peifer M, Markert E, Stemshorn K, Timmermann B, Saal B, Klose S, Ernestus K, Scheffler M, Engel-Riedel W, Stoelben E, Brambilla E, Wolf J, Nürnberg P, Thomas RK. Benchmarking of mutation diagnostics in clinical lung cancer specimens. PLoS One 2011; 6: e19601.
- Tsiatis AC, Norris-Kirby A, Rich RG, Hafez MJ, Gocke CD, Eshleman JR, Murphy KM. Comparison of Sanger sequencing, pyrosequencing, and melting curve analysis for the detection of KRAS mutations: diagnostic and clinical implications. J. Mol. Diagn. 2010; 12: 425–432.
- Dijkstra JR, Heideman DAM, Meijer GA, Boers JE, 't Hart NA, Diebold J, Hirschmann A, Hoefler G, Winter G, Miltenberger-Miltenyi G, Pereira S V, Richman SD, Quirke P, Rouleau EL, Guinebretiere JM, Tejpar S, Biesmans B, van Krieken JHJM. KRAS mutation analysis on low percentage of colon cancer cells: the importance of quality assurance. Virchows Arch. 2013; 462: 39–46.
- Van Krieken JH, Siebers AG, Normanno N. European consensus conference for external quality assessment in molecular pathology. Ann. Oncol. 2013; 24: 1958–1963.
- 10. Ciardiello F, Normanno N, Maiello E, Martinelli E, Troiani T, Pisconti S, Giuliani F, Barone C, Carteni G, Rachiglio a M, Montesarchio V, Tonini G, Rizzi D, Cinieri S, Bordonaro R, Febbraro A, De Vita F, Orditura M, Fenizia F, Lambiase M, Rinaldi A, Tatangelo F, Botti G, Colucci G. Clinical activity of FOLFIRI plus cetuximab according to extended gene mutation status by next generation sequencing: findings from the CAPRI-GOIM trial. Ann.

- Oncol. 2014; 25: 1756-1761.
- Tembuyser L, Ligtenberg MJL, Normanno N, Delen S, van Krieken JH, Dequeker EMC. Higher quality of molecular testing, an unfulfilled priority: Results from external quality assessment for KRAS mutation testing in colorectal cancer. J. Mol. Diagn. 2014; 16: 371–377.
- Amado RG, Wolf M, Peeters M, Van Cutsem E, Siena S, Freeman DJ, Juan T, Sikorski R, Suggs S, Radinsky R, Patterson SD, Chang DD. Wild-type KRAS is required for panitumumab efficacy in patients with metastatic colorectal cancer. J. Clin. Oncol. 2008; 26: 1626–1634.
- 13. Ciardiello F, Tejpar S, Normanno N, Mercadante D, Teague T, Wohlschlegel B, Van Cutsem E. Uptake of KRAS mutation testing in patients with metastatic colorectal cancer in Europe, Latin America and Asia. Target. Oncol. 2011; 6: 133–145.
- 14. De Roock W, Claes B, Bernasconi D, De Schutter J, Biesmans B, Fountzilas G, Kalogeras KT, Kotoula V, Papamichael D, Laurent-Puig P, Penault-Llorca F, Rougier P, Vincenzi B, Santini D, Tonini G, Cappuzzo F, Frattini M, Molinari F, Saletti P, De Dosso S, Martini M, Bardelli A, Siena S, Sartore-Bianchi A, Tabernero J, Macarulla T, Di Fiore F, Gangloff AO, Ciardiello F, Pfeiffer P, et al. Effects of KRAS, BRAF, NRAS, and PIK3CA mutations on the efficacy of cetuximab plus chemotherapy in chemotherapy-refractory metastatic colorectal cancer: a retrospective consortium analysis. Lancet Oncol. 2010; 11: 753–762.
- Peeters M, Douillard J-Y, Van Cutsem E, Siena S, Zhang K, Williams R, Wiezorek J. Mutant KRAS codon 12 and 13 alleles in patients with metastatic colorectal cancer: assessment as prognostic and predictive biomarkers of response to panitumumab. J. Clin. Oncol. 2013; 31: 759– 765.
- 16. Wong A, Ma BBY. Personalizing therapy for colorectal cancer. Clin. Gastroenterol. Hepatol. 2014; 12: 139–144.
- 17. Negru S, Papadopoulou E, Apessos A, Stanculeanu DL, Ciuleanu E, Volovat C, Croitoru A, Kakolyris S, Aravantinos G, Ziras N, Athanasiadis E, Touroutoglou N, Pavlidis N, Kalofonos HP, Nasioulas G. KRAS, NRAS and BRAF mutations in Greek and Romanian patients with colorectal cancer: a cohort study. BMJ Open 2014; 4: e004652.
- Vaughn CP, Zobell SD, Furtado L V, Baker CL, Samowitz WS. Frequency of KRAS, BRAF, and NRAS mutations in colorectal cancer. Genes. Chromosomes Cancer 2011; 50: 307–312.
- Peeters M, Oliner KS, Parker A, Siena S, Van Cutsem E, Huang J, Humblet Y, Van Laethem J-L, André T, Wiezorek J, Reese D, Patterson SD. Massively parallel tumor multigene sequencing to evaluate response to panitumumab in a randomized phase III study of metastatic colorectal cancer. Clin. Cancer Res. 2013; 19: 1902–1912.
- Cancer Genome Atlas Network. Comprehensive molecular characterization of human colon and rectal cancer. Nature 2012; 487: 330–337.

- 21. Thunnissen E, Bovée JVMG, Bruinsma H, van den Brule AJC, Dinjens W, Heideman DAM, Meulemans E, Nederlof P, van Noesel C, Prinsen CFM, Scheidel K, van de Ven PM, de Weger R, Schuuring E, Ligtenberg M. EGFR and KRAS quality assurance schemes in pathology: generating normative data for molecular predictive marker analysis in targeted therapy. J. Clin. Pathol. 2011; 64: 884–892.
- Viray H, Li K, Long TA, Vasalos P, Bridge JA, Jennings LJ, Halling KC, Hameed M, Rimm DL. A prospective, multi-institutional diagnostic trial to determine pathologist accuracy in estimation of percentage of malignant cells. Arch. Pathol. Lab. Med. 2013; 137: 1545–1549.
- 23. Smits AJJ, Kummer JA, de Bruin PC, Bol M, van den Tweel JG, Seldenrijk KA, Willems SM, Offerhaus GJA, de Weger RA, van Diest PJ, Vink A. The estimation of tumor cell percentage for molecular testing by pathologists is not accurate. Mod. Pathol. 2014; 27: 168–174.
- 24. Viera AJ, Garrett JM. Understanding Interobserver Agreement: The Kappa Statistic. 2005; 37: 360–363.
- 25. Sie D, Snijders PJF, Meijer GA, Doeleman MW, van Moorsel MIH, van Essen HF, Eijk PP, Grünberg K, van Grieken NCT, Thunnissen E, Verheul HM, Smit EF, Ylstra B, Heideman DAM. Performance of amplicon-based next generation DNA sequencing for diagnostic gene mutation profiling in oncopathology. Cell. Oncol. (Dordr). 2014; 37: 353–361.