

Periodontal ligament influence on the stress distribution in a removable partial denture supported by implant: a finite element analysis

Carlos Marcelo ARCHANGELO¹, Eduardo Passos ROCHA², João Antônio PEREIRA³, Manoel MARTIN JUNIOR¹, Rodolfo Bruniera ANCHIETA⁴, Amilcar Chagas FREITAS JÚNIOR⁴

1-DDS, PhD, Professor, Federal Institute of Paraná – IFPR, Londrina, PR, Brazil.

2-DDS, PhD, Associate Professor, Department of Dental Materials and Prosthodontics, Faculty of Dentistry of Araçatuba, UNESP - Univ. Estadual Paulista, Araçatuba, SP, Brazil.

3-MEng, PhD, Associate Professor, Department of Mechanical Engineering, Faculty of Engineering of Ilha Solteira, UNESP - Univ. Estadual Paulista, Ilha Solteira, SP, Brazil.

4-DDS, MS, PhD student, Department of Dental Materials and Prosthodontics, Faculty of Dentistry of Araçatuba, UNESP - Univ. Estadual Paulista, Araçatuba, SP, Brazil.

Corresponding address: Prof. Adj. Dr. Carlos Marcelo Archangelo - Rua José Bonifácio, 1193, Vila Mendonça - Araçatuba - SP - Brasil - 16010-050 - Phone: 55-18-36363290 - e-mail: carlosarchangelo@uol.com.br

Received: October 18, 2010 - **Modification:** August 7, 2011 - **Accepted:** September 1, 2011

ABSTRACT

Objective: The non-homogenous aspect of periodontal ligament (PDL) has been examined using finite element analysis (FEA) to better simulate PDL behavior. The aim of this study was to assess, by 2-D FEA, the influence of non-homogenous PDL on the stress distribution when the free-end saddle removable partial denture (RPD) is partially supported by an osseointegrated implant. Material and Methods: Six finite element (FE) models of a partially edentulous mandible were created to represent two types of PDL (non-homogenous and homogenous) and two types of RPD (conventional RPD, supported by tooth and fibromucosa; and modified RPD, supported by tooth and implant [10.00x3.75 mm]). Two additional FE models without RPD were used as control models. The non-homogenous PDL was modeled using beam elements to simulate the crest, horizontal, oblique and apical fibers. The load (50 N) was applied in each cusp simultaneously. Regarding boundary conditions the border of alveolar ridge was fixed along the x axis. The FE software (Ansys 10.0) was used to compute the stress fields, and the von Mises stress criterion (σ_vM) was applied to analyze the results. Results: The peak of σ_vM in non-homogenous PDL was higher than that for the homogenous condition. The benefits of implants were enhanced for the non-homogenous PDL condition, with drastic σ_vM reduction on the posterior half of the alveolar ridge. The implant did not reduce the stress on the support tooth for both PDL conditions. Conclusion: The PDL modeled in the non-homogeneous form increased the benefits of the osseointegrated implant in comparison with the homogeneous condition. Using the non-homogenous PDL, the presence of osseointegrated implant did not reduce the stress on the supporting tooth.

Key words: Removable partial denture. Dental implant. Finite element analysis. Periodontal ligament.

INTRODUCTION

Conventional removable partial dentures (RPD) show a complex biomechanical behavior, mainly when the RPD is supported by teeth and fibromucosa, such as in Kennedy Class I²⁵. This behavior has been better controlled when an

osseointegrated implant is placed and acts as an additional support, retainer, or both, improving the stability and retention of the RPD, which reduces the demand on the support structures¹⁴.

In this context, the use of RPD and implants has been presented in a series of clinical possibilities that reflect individual experiences with some

singularities according to the number and position of the remaining teeth in the arch, as well as the size of the prosthetic space and degree of bone loss^{2-4,9-11,13,14,16,17,20,26,28}. The usual clinical responses from patients are represented by better RPD stability, retention and comfort when this association is done¹⁴.

Mitrani, et al.¹⁸(1993) investigated the free-end saddle RPD combined with an implant on the distal extension wearers for a mean period of 2.5 years. Those authors showed that the implant placed on the distal extension of RPD increases patient satisfaction, reduces the marginal bone loss, and establishes the peri-implant tissues¹⁸. Nonetheless, data about the abutment tooth behavior over time are still not reported and remain inconclusive.

Finite element (FE) studies can provide valuable information about the stress reduction on the support structures promoted by implants. However this information is controversial because the benefits of the osseointegrated implant are more evident on the alveolar ridge and hardly conclusive for the RPD abutment tooth²³.

There are some concerns about the tooth behavior in a RPD supported by implant and tooth through finite element analysis (FEA), mainly because the periodontal ligament (PDL) has not been properly modeled as a non-homogenous structure. The homogenous aspect of the PDL in FEA might negatively change the results. Atmaram and Mohammed¹ (1981) reported that the homogeneous condition of the PDL is closer to more uniform stress distribution and unable to direct the forces applied in accordance with the set of fibers *in vivo*. In addition, the magnitude of stresses is shown to be less pronounced than under the non-homogeneous condition. This is particularly important in face of forces with non-axial incidence, producing terminal torqueing forces against the abutment teeth and the soft tissue¹⁹.

Other authors have pointed out that the non-homogeneous PDL was necessary in order to understand how occlusal loads are absorbed by the PDL²². Therefore, this condition is decisive to analyze the results of RPD associated with an osseointegrated implant by FEA. Nevertheless, the benefits of the combination between RPD and osseointegrated implant may be analyzed from this perspective. The analysis of the stress distribution in the studies by Atmaram and Mohammed¹ (1981) and Rees and Jacobsen²² (1997) did not show the behavior of the abutment tooth.

The objective of this study was to evaluate the influence of PDL condition (homogenous and non-homogenous) on the stress distribution of a free-end saddle RPD associated with an osseointegrated implant through 2-D FEA. The hypothesis tested was that non-homogeneous PDL increases the benefits

of implant, reducing the stress on the supporting tooth of the RPD supported by tooth and implant.

MATERIAL AND METHODS

Six mandibular models (A, B, C, D, E, and F) were modeled using graphic software (AutoCAD, Autodesk Inc., San Rafael, CA, USA). In a sagittal view, all models simulated partially edentulous hemi-arches without posterior dental support (Kennedy Class I). In addition, all models had the remaining tooth 33 and a distal extension ridge.

The characteristics of the abutment - the length of the distal extension ridge, the dimensions and characteristics of the support and protective periodontium, the mandibular bone height, the thickness of the CoCr metal structure, and the number of artificial teeth - were kept constant. The difference was the PDL in models D, E, and F, which was reproduced in accordance with the *in vivo* characteristic, with representation of the following 4 groups of fibers: crest, horizontal, oblique and apical (Figure 1).

After the models had been created, the files were exported to ANSYS 8.0 (Swanson Analysis Systems, Houston, PA, USA) to discriminate the regions and generate the FE mesh. Models A, B, and C, (Figure 1) were considered homogeneous, isotropic, and linearly elastic. The mechanical properties adopted for all materials (elastic modulus and Poisson's ratio) were established according to the literature (Table 1)^{1,8,15,23,24,29}.

Models D, E, and MF, similar to models A, B, and C, respectively, were considered homogeneous, isotropic, and linearly elastic, except for the PDL, which was considered a non-homogeneous structure. Thus, models D, E, and F differed from models A, B and C only by their structural characteristic of the PDL, being similar in all other factors.

The element used to generate the mesh was the plane 2 element, a triangular element defined by 6 nodes, with 2 degrees of freedom *per* node, and quadratic displacement behavior. The FE mesh showed up to 80,000 nodes and up to 40,000 elements. This configuration allowed the appropriate refinement of the mesh in thin structures, such as the cortical bone and the PDL, reaching the convergence norm.

As far as the creation of the non-homogeneous PDL is concern, using the methodology established by Atmaram and Mohammed¹ (1981) with some modifications, the composition of the PDL in the FE model was established under the *in vivo* condition⁵. For this, beam elements were used to represent 4 groups of PDL fibers in a sagittal cut (crest, horizontal, oblique and apical fibers) and their directions. To support the bending movements, as

well as tensile and compression loads, the cross-sectional area and the moment of inertia for the beam elements were used to describe the PDL. Thus, the following equations were used:

Cross sectional area: $A=\pi \times (r)^2$

Moment of inertia: $I=\pi \times (d)^4 / 64$

where:

A=cross sectional area; I=moment of inertia;
 $\pi=3.14$; $d=4 \mu\text{m}$ or $4 \times 10^{-6} \text{ m}$; $r=2 \times 10^{-6} \text{ m}$.

Grant, et al.¹² (1988) determined a value of 4 μm for the PDL fibers of an RPD abutment tooth. Thus, the values found for the cross-sectional area and the moments of inertia of each element were $1,256 \times 10^{-5} \text{ mm}^2$ and $0.785 \times 10^{-12} \text{ mm}^4$, respectively.

In accordance with that established by Atmaram and Mohammed¹ (1981), the number of periodontal fibers adopted under the non-homogeneous condition was half the number of elements present




Figure 1- Description of finite element models, finite element mesh and boundary condition. A: Models A, B, and C. B - Model B with the removable partial denture (RPD). C: Model C with RPD and implant. Areas (1 to 6) selected to perform the analysis in all models. D:- The beam elements used to build the non-homogeneous periodontal ligament (PDL). E and F: Mesh refinements for tooth apex and implant, respectively. L shows the distributed loading scheme. The entire right and left sides were fixed on the x axis. The bone at the bottom of the model was fixed on the x and y axes

Table 1- Mechanical properties of the materials. E – Elastic modulus. v – Poisson’s ratio. PDL – Periodontal ligament. Ti - Titanium

Structures	E (GPa)	References	v	References
Enamel	41	17	0.3	17
Dentin	18.6	18	0.31	18
Homogenous PDL	0.175	15	0.45	15
Non-homogenous PDL	0.35	15	0.45	15
Fibromucosa	0.68	17	0.45	17
Cortical bone	13.7	18	0.3	18
Medullar bone	1.37	18	0.3	18
Implant (Ti)	103.4	20	0.35	20
Healing abutment (Ti)	103.4	20	0.35	20
CoCr cast alloy	185	21	0.35	21
Acrylic resin	8.3	14	0.28	14
Artificial teeth	8.3	14	0.28	14

in the PDL under the homogeneous condition. Thus, all models with a homogeneous PDL (A, B, and C) contained 824 elements in the PDL region, and the models with a non-homogeneous PDL (D, E, and F) contained 412 fibers for composition of the PDL.

As the boundary condition, the left and right sides of the models were fixed only in the x direction, in order to prevent lateral movement of the structures. This allowed the vertical movement of the RPD base over the fibromucosa and, consequently, the deformation of the cortical and medullar bone beneath it, as well as simulation of the bilateral mandibular condition. Only the cortical bone at the base of the models was fixed in the x and y directions (Figure 1).

The models were loaded with vertical forces (50

N) distributed at each cusp, fractionated into 5 point loads of 10 N each to prevent the occlusal contact characterized by only one point load (Figure 1).

The von Mises equivalent stress criterion (σ_{VM}) was adopted to analyze the stress distribution. Each structure was individually analyzed following specific areas (Figure 1), as well as to allow the data to be refined into regions of interest, as follows:

Area 1: Root apex; Area 2: Mesial side of the abutment tooth and adjacent structures; Area 3: Distal side of the abutment tooth and adjacent structures; Area 4: Distal bone crest of the abutment tooth; Area 5: Anterior half of the alveolar ridge; Area 6: Posterior half of the alveolar ridge; Area 7: Osseointegrated implant.

For more details and to allow the comparison

Table 2- von Mises stress (σ_{VM}), in MPa, for the cortical bone according to specific areas in models A to F. Note the subdivision of regions 2 and 3 into 10 segments (1 to 10) around the abutment tooth

Models	Areas	Subdivision of the regions 2 and 3 around the abutment tooth									
		1	2	3	4	5	6	7	8	9	10
A	1	10.5									
	2	10.5	10.5	12.2	12.2	14	15.7	17.5	19.2	21	19.2
	3	8.7	5.2	5.2	7	7	8.7	8.7	10.5	12.2	19.2
	4	8.7									
	5	3.5									
	6	1.7									
B	1	33.4									
	2	14.3	14.3	20.7	20.7	20.7	27	27	27	27	20.7
	3	20.7	20.7	27	27	27	33.4	33.4	39.7	77.8	77.8
	4	27									
	5	20.7									
	6	14.3									
C	1	25.1									
	2	25.1	12.9	19	19	25.1	25.1	31.2	31.2	31.2	19
	3	25.1	19	25.1	25.1	25.1	31.2	37.4	43.5	74	74
	4	25.1									
	5	19									
	6	12.9									
D	1	8.6									
	2	103	42.9	42.9	42.9	42.9	42.9	42.9	42.9	51.2	51.2
	3	77.2	17.2	8.6	8.6	8.6	8.6	8.6	8.6	17.2	25.7
	4	77.2									
	5	17.2									
	6	8.6									
E	1	34.8									
	2	86.3	43.4	34.8	34.8	34.8	34.8	34.8	34.8	34.8	34.8
	3	86.3	52	43.4	52	52	52	52	60.6	103.5	103.5
	4	86.3									
	5	26.3									
	6	17.7									
F	1	33.5									
	2	91.6	41.8	41.8	41.8	41.8	41.8	41.8	33.5	33.5	25.2
	3	83.3	50.1	50.1	50.1	50.1	50.1	50.1	58.4	100	100
	4	83.3									
	5	25.2									
	6	8.6									

Table 3- von Mises stress (σ_{VM}), in MPa, for the medullar bone according to specific areas (1 to 6) in the models (A to F)

Models	Areas					
	1	2	3	4	5	6
A	8.3	2.8	2.8	1.4	0.7	0.7
B	17.8	5.1	5.1	3.7	9.3	6.5
C	18.1	4.4	6.4	4.4	8.3	4.4
D	11.4	5.7	1.9	3.8	1.9	1.9
E	17.4	4.4	5.8	11.6	10.2	7.3
F	17.6	4.9	6.1	11.8	11.9	4.3

Table 4- von Mises stress (σ_{VM}), in MPa, for the fibromucosa according to specific areas (4 to 6) in the models (A to F)

Models	Areas		
	4	5	6
A	5.7	0.9	0.9
B	7	8.6	7.8
C	7.1	8	3.7
D	95	8.3	8.3
E	59.7	14.1	14.1
F	61.3	13.7	6.9

with the study of Atmaram and Mohamed¹ (1981), the previously determined areas 2 and 3 were subdivided into 10 parts.

RESULTS

Cortical bone

For the homogeneous PDL, the stress concentration on the cortical bone around tooth 33 gradually increased in the bone crest (tooth apex direction) in areas 2 and 3 of models A, B, and C. The presence of the conventional RPD (model B) drastically increased the stress on areas 5 and 6 in comparison with MA, and the presence of the osseointegrated implant (model C) reduced the σ_{VM} on the posterior half of the ridge, mainly in area 6 (Table 2).

For the non-homogeneous PDL (models D, E, and F), the σ_{VM} in areas 2 and 3 of model D diminished in the bone crest (root apex direction). The presence of the conventional RPD (model E) increased the σ_{VM} in areas 1, 4, 5, and 6 compared with model D, and the presence of the osseointegrated implant (model F) drastically reduced the stresses in area 6 of the alveolar ridge (Table 2).

Medullar bone

The incorporation of the RPD (model B) drastically increased the σ_{VM} on the trabecular bone for areas 5 and 6 by approximately 1235% and 834%, respectively, when compared with the

respective areas of model A (Table 3). The presence of the osseointegrated implant (model C) reduced the σ_{VM} in those areas, similar to what occurred in the cortical bone.

For the non-homogeneous PDL, model E increased the σ_{VM} in areas 5 and 6 (Table 3) compared with areas 5 and 6 of model B for the homogeneous PDL. The osseointegrated implant (model F) also reduced the stress levels in area 6 of the ridge (Table 3).

Fibromucosa

The high σ_{VM} in areas 5 and 6 of the fibromucosa in model B, corroborate those found for the same cortical and trabecular bone areas in relation to the homogeneous PDL. The osseointegrated implant (model C) also provided the fibromucosa with stress relief in areas 5 and 6 of the ridge, when compared with the σ_{VM} of model B (Table 4).

For the PDL under the non-homogeneous condition in areas 5 and 6, the RPD (model E) increased the σ_{VM} , however, in lower ratio than those observed under the homogeneous condition, approximately 69% in both areas. The osseointegrated implant also provided model F with a reduction in σ_{VM} in areas 5 and 6 of the fibromucosa (Table 4).

Implant

The osseointegrated implant showed a similar behavior for the homogeneous and non-homogeneous PDL conditions. The pitch of the internal threads of the implants was responsible for the high σ_{VM} in models C and F.

DISCUSSION

The conventional RPD applies moment of force or binary forces on the abutment tooth and alveolar ridge^{6,7}. The associations of the RPD with an osseointegrated implant aims to increase the retention and stability, as well as provide a reduction of the stress on the support tooth, fibromucosa and alveolar ridge¹⁴.

In a previous FEA study with the RPD and

implant²³, this reduction of stress on the support tooth was not confirmed²³. It was considered that the homogeneous PDL condition might reduce the stress in the alveolar ridge, but was not efficient in reducing the stress on the abutment tooth when an osseointegrated implant was distally supporting the acrylic resin base of the RPD.

In the present study, when the homogeneous PDL condition was simulated, the stress concentration on the cortical and medullar bone remained high at the abutment tooth apex. The σ_{VM} increased gradually in the bone crest – tooth apex direction in areas 2 and 3 (models A, B, and C). The greatest benefit of the osseointegrated implant was the reduction of the stresses on the posterior half of the alveolar ridge. The peak stress in the implant was observed on the neck closer to the cortical bone.

In agreement with another study³, the higher σ_{VM} found in the medullar bone for the homogeneous PDL condition occurred at the apex of the osseointegrated implant. It is pointed out that the stress found on the cortical bone around the osseointegrated implant remained below the stress levels found at the abutment tooth apex, showing that RPD associated with an osseointegrated implant to be a feasible and safe alternative, according Mitrani, et al.¹⁸ (2003).

Another benefit of the implant (model C) was the stress reduction in the fibromucosa in the posterior half compared with the model B. The implant provided anchorage for the acrylic resin base of the RPD, limiting its vertical intrusion movement and reducing the σ_{VM} when it was compared with the model B.

Nevertheless, the use of the non-homogeneous PDL (models D, E, and F) changed the stress distribution pattern in practically all regions analyzed. The peak of stress was higher for almost all structures with the non-homogeneous condition than with the homogeneous condition (models A, B, and C). However, the stress variation around the abutment tooth was lower compared with models D, E, and F, and its distribution was similar to that established by others authors^{1,21,27}.

The modeling type for PDL has a significant effect on the nature and magnitude of the alveolar stress^{1,21,27}. It occurs because the modeling of homogeneous PDL is like an assumption of a hypothetical soft interactive medium in which the stress is shared uniformly; hence, it uniformly distributes the stresses to bone. In contrast, in the heterogeneous PDL assumption with the main fibers of the PDL incorporated, there is no direct contact between the individual fibers, and the stresses among the fibers are not shared equally, resulting in the wide variation of the stresses in the adjoining structure, such as the cortical bone, reproducing more accurately the events of the PDL in function.

Similarly to Atmaram and Mohammed¹ (1981), in the present study there was an inversion in the direction of stress, since the stresses around tooth 33 gradually diminished in the direction of the bone crest to the tooth apex when the non-homogeneous PDL was simulated. Furthermore, the increase of σ_{VM} on the anterior and posterior halves of the alveolar ridge for areas 5 and 6 (Table 4) were accentuated in model B. A similar behavior was not verified in models D and E, as the σ_{VM} increased 69.1% and 69.1% for those areas, respectively.

The results found in models A, B, and C, with the homogeneous PDL condition, might be not represent the potentially artificial, since the peak of σ_{VM} , as well the stress variation was not repeated in the equivalent models, D, E, and F, respectively. In addition, the stress around the abutment tooth, and in the posterior area of the alveolar ridge in models C and F, should be closer to that observed in models A and D, with no RPD, respectively. It was observed that the result closest to this hypothesis was showed by the posterior half of the fibromucosa. Even with the use of non-homogeneous PDL, the stress variation around the abutment tooth in models E and F was small, rejecting the hypothesis of the present study. This aspect partially supports the results pointed by Keltjens, et al.¹⁴ (1993) that the osseointegrated implant does not reduce the stress on the abutment tooth.

The results of the present study keep valid the data from Atmaram and Mohammed¹ (1981) because the direction of demand on the abutment tooth and the behavior of the bone showed correlation with the experiment¹. This aspect reinforces the condition that the PDL used in its non-homogeneous form appropriately reproduces the PDL in function^{21,27}.

The main benefit of the osseointegrated implant placed in the position as performed in the present study is reducing the stress on the fibromucosa and alveolar ridge. These benefits are more evident when the PDL is modeled in its non-homogeneous form.

CONCLUSION

The PDL modeled in the non-homogeneous form increases the benefits of the osseointegrated implant in comparison with the homogeneous condition. Using the non-homogeneous PDL, the presence of osseointegrated implant did not reduce the stress on the supporting tooth.

REFERENCES

- 1- Atmaram GH, Mohammed H. Estimation of physiologic stresses with a natural tooth considering fibrous PDL structure. *J Dent Rest.* 1981;60:873-7.
- 2- Attard NJ, Zarb GA. Long-term treatment outcomes in edentulous patients with implant-fixed prostheses: the Toronto study. *Int J Prosthodont.* 2004;17:417-24.
- 3- Barbier L, Sloten LV, Krzesinski G, Schepers E, Van Der Perre G. Finite element analysis of non-axial versus axial loading of oral implants in the mandible of the dog. *J Oral Rehabil.* 1998;25:847-58.
- 4- Battistuzzi PGFCM, Van Slooten H, Käyser AF. Management of an anterior defect with a removable partial denture supported by implants and residual teeth: a case report. *Int J Oral Maxillofac Implants.* 1992;7:112-5.
- 5- Berkovitz BKB, Moxham BJ, Newman HN. The periodontal ligament in health and disease. 2nd ed. London: Ed. Mosby- Wolfe; 1995. 446 p.
- 6- Chou TM, Caputo AA, Moore JD, Xiao B. Photoelastic analysis and comparison of force-transmission characteristics of intracoronal attachments with clasp distal-extension removable partial dentures. *J Prosthet Dent.* 1989;62:313-9.
- 7- Chou TM, Eick JD, Moore DJ, Tira DE. Stereophotogrammetric analysis of abutment tooth movement in distal-extension removable partial dentures with intracoronal attachments and clasps. *J Prosthet Dent.* 1991;66:343-9.
- 8- Farah JW, Craig RG, Meroueh KA. Finite element analysis of a mandibular model. *J Oral Rehabil.* 1988;15:615-24.
- 9- Ganz SD. Combination natural tooth and implant-borne removable partial denture: a clinical report. *J Prosthet Dent.* 1991;66:1-5.
- 10- George MA. Removable partial denture design assisted by osseointegrated implants. *J Calif Dent Assoc.* 1992;20:64-6.
- 11- Giffin KM. Solving the distal extension removable partial denture base movement dilemma: a clinical report. *J Prosthet Dent.* 1996;76:347-9.
- 12- Grant DA, Stern IB, Listgarten MA. *Periodontics.* 6th ed. St. Louis: Mosby; 1988. 1154 p.
- 13- Jang Y, Emtiaz S, Tarnow DP. Single implant-supported crown used as an abutment for a removable cast partial denture: a case report. *Implant Dent.* 1998;7:199-204.
- 14- Keltjens HMAM, Käyser AF, Hertel R, Battistuzzi PGF. Distal extension removable partial dentures supported by implants and residual teeth: considerations and case reports. *Int J Oral Maxillofac Implants.* 1993;8:208-13.
- 15- Ko CC, Chu CS, Chung KH, Lee MC. Effects of posts on dentin stress distribution in pulpless teeth. *J Prosthet Dent.* 1992;68:421-7.
- 16- McAndrew R. Prosthodontic rehabilitation with a swing-lock removable partial denture and a single osseointegrated implant: a clinical report. *J Prosthet Dent.* 2002;88:128-31.
- 17- Mijiritsky E and Karas S. Removable partial denture design involving teeth and implants as an alternative to unsuccessful fixed implant therapy: a case report. *Implant Dent.* 2004;13:218-22.
- 18- Mitrani R, Brudvik JS, Phillips KM. Posterior implants for distal extension removable prostheses: a retrospective study. *Int J Periodontics Restorative Dent.* 2003;23:353-9.
- 19- Ohkubo C, Kobayashi M, Suzuki Y, Hosoi T. Effect of implant support on distal-extension removable partial dentures: *in vivo* assessment. *Int J Oral Maxillofac Implants.* 2008;23:1095-101.
- 20- Pellecchia M, Pellecchia R, Emtiaz, S. Distal extension mandibular removable partial denture connected to an anterior fixed implant-supported prosthesis: a clinical report. *J Prosthet Dent.* 2000;83:607-12.
- 21- Qian H, Chen J, Katona TR. The influence of the PDL principal fibers in a 3-dimensional analysis of orthodontic tooth movement. *Am J Orthod Dentofacial Orthop.* 2001;120:272-9.
- 22- Rees JS, Jacobsen PH. Elastic modulus of the periodontal ligament. *Biomaterials.* 1997;18:995-9.
- 23- Rocha EP, Luersen MA, Pellizzer EP, Del Bel Cury AA. Distal-extension removable partial denture associated with an osseointegrated implant. Study by finite element method [Abstract. No. 1948]. *J Dent Res.* 2003;82:254.
- 24- Sertgoz A, Gunever S. Finite element analysis of the effect of cantilever and implant length on stress distribution in an implant-supported fixed prosthesis. *J Prosthet Dent.* 1996;76:165-9.
- 25- Shahmiri RA, Atieh MA. Mandibular Kennedy Class I implant-tooth-borne removable partial denture: a systematic review. *J Oral Rehabil.* 2010;37:225-34.
- 26- Starr NL. The distal extension case: an alternative restorative design for implant prosthetics. *Int J Periodontics Restorative Dent.* 2001;21:61-7.
- 27- Toms SR, Eberhardt AW. A nonlinear finite element analysis of the periodontal ligament under orthodontic tooth loading. *Am J Orthod Dentofacial Orthop.* 2003;123:657-65.
- 28- Uludag B, Celik G. Fabrication of a maxillary implant-supported removable partial denture: a clinical report. *J Prosthet Dent.* 2006;95:19-21.
- 29- Williams DF. *Biocompatibility of clinical implant materials.* Boca Raton: CRC Press; 1981. p. 99-127.