

BIOMECHANICAL ANALYSIS OF DISTAL STENT GRAFT INDUCED NEW ENTRY (SINE) FORMATION IN AORTIC DISSECTION PATIENT

Wei Ting Tan¹, Mohd Jamil Mohamed Mokhtarudin², Yih Miin Liew¹
, Nasrul Hadi Johari², and Einly Lim¹

1 Department of Biomedical Engineering, Faculty of Engineering, University Malaya, 50603 Kuala Lumpur, Malaysia, weitingtan14@gmail.com, liewym, einly_lim@um.edu.my

2 Faculty of Mechanical Engineering, Universiti Malaysia Pahang, 26600 Pekan, Pahang, Malaysia, {mohdjamil, [nhadi](mailto:nhadi@ump.edu.my)}@ump.edu.my

Abstract:

This is the first study to investigate the effect of material models and properties on stress and strain distributions in Type IIIb aortic dissection patient with distal stent graft induced new entry. Both elastic and hyperelastic models consistently showed a reduction in all stress and strain parameters (i.e. 5.4% - 59.1%) at neighboring region distal to the stent-wall composite when the disease progressed from acute to chronic. We found that the maximum shear stress (SS) presents a potential inducer for SINE formation, while the maximum shear elastic strain (SE) could be an indicator for the location of such formation.

Keywords: Aortic Dissection; Distal Stent Graft Induced Entry (SINE); Thoracic Endovascular Aneurysm Repair (TEVAR); Disease Stage Based

ACKNOWLEDGEMENT

The authors wish to acknowledge the help provided by Prof Xiao Yun Xu of Imperial College London and her research team, especially Claudia Menichini and Selene Pirola in sharing the reconstructed patient specific geometry used in this study. This work is supported by Royal Society Newton Advanced Fellowship (NA170094) and UMP RDU1803120.