



Article

Oxygen Saturation Measurements from Green and Orange Illuminations of Multi-Wavelength Optoelectronic Patch Sensors

Samah Alharbi ¹, Sijung Hu ^{1,*} , David Mulvaney ¹, Laura Barrett ², Liangwen Yan ³, Panagiotis Blanos ¹, Yasmin Elsahar ¹  and Samuel Adema ¹ 

¹ Wolfson School of Mechanical, Electrical and Manufacturing Engineering, Loughborough University, Ashby Road, Loughborough, Leicestershire LE11 3TU, UK; S.M.Alharbi@lboro.ac.uk (S.A.); D.J.Mulvaney@lboro.ac.uk (D.M.); P.Blanos@lboro.ac.uk (P.B.); Y.Elsahar@lboro.ac.uk (Y.E.); S.T.Adema@lboro.ac.uk (S.A.)

² School of Sport, Exercise and Health Sciences, Loughborough University, Ashby Road, Loughborough, Leicestershire LE11 3TU, UK; L.A.Barrett@lboro.ac.uk

³ School of Mechatronic Engineering and Automation, Shanghai University, Shanghai 200072, China; lw_yan@staff.shu.edu.cn

* Correspondence: S.Hu@lboro.ac.uk; Tel.: +44-1509-227059

Received: 21 October 2018; Accepted: 20 December 2018; Published: 31 December 2018



Abstract: Photoplethysmography (PPG) based pulse oximetry devices normally use red and infrared illuminations to obtain oxygen saturation (SpO₂) readings. In addition, the presence of motion artefacts severely restricts the utility of pulse oximetry physiological measurements. In the current study, a combination of green and orange illuminations from a multi-wavelength optoelectronic patch sensor (mOEPS) was investigated in order to improve robustness to subjects' movements in the extraction of SpO₂ measurement. The experimental protocol with 31 healthy subjects was divided into two sub-protocols, and was designed to determine SpO₂ measurement. The datasets for the first sub-protocol were collected from 15 subjects at rest, with the subjects free to move their hands. The datasets for the second sub-protocol with 16 subjects were collected during cycling and walking exercises. The results showed good agreement with SpO₂ measurements ($r = 0.98$) in both sub-protocols. The outcomes promise a robust and cost-effective approach of physiological monitoring with the prospect of providing health monitoring that does not restrict user physical movements.

Keywords: oxygen saturation (SpO₂); green and orange illuminations; optoelectronic patch sensor (mOEPS); pulse oximetry; physical movement

1. Introduction

Many researchers have contributed to improving the performance of physiological monitoring systems, including the simplification of the setting-up of equipment, generating more accurate readings or reducing the cost of obtaining measurements. One of the monitoring approaches commonly used is photoplethysmography (PPG), a non-invasive optical technique able to measure volume changes in blood vessels and to estimate vital health signs, such as heart rate (HR), heart rate variability, respiration rate, blood pressure and oxygen saturation (SpO₂) [1]. PPG sensors are normally designed to operate either in transmission mode or reflectance mode. In transmission mode, a photo detector (PD) is placed to detect light after it has been transmitted through a subject's body part (such as a finger or earlobe). In a reflectance mode, the PD is usually located on the same side of the body part as the illumination source and captures back-scattered light from skin and tissue. Transmission mode is widely used in

hospital and clinical settings, such as in pulse oximetry, whereas the reflectance approach is mainly used in wearable sensors, such as smart watches. To date, a number of wearable sensors used to detect physiological parameters have been developing commercially, such as HRM-3200 (H3, Daejeon, Korea) [2], Honour S1 (Huawei, Shenzhen China) [3], MZ-3Belt (Myzone, Nottingham, UK) [4], NoninWristOx 3100 [5], AMON (Advanced Medical Monitor) physiological monitor [6] LifeShirt [7], Smart 32 Shirt [8], SenseWear [9] and Smart Vest [10]. The transmission mode is restrictive in terms of the number of options available for measurement sites, e.g., finger, earlobe and toe, in addition to the limitation of the patients' freedom of movement [11]. For example, fingertip sensors, such as those used for pulse oximetry, interfere with daily activities and it can be difficult to measure vital signs reliably during hand movements. By using a reflectance multi-wavelength optoelectronic sensor (mOEPS) in this work, the number of suitable sites available on the body increases dramatically. Furthermore, it has been reported that the reflectance mode PPG sensor probe MaxFast (Nellcor™, Minneapolis, MN, USA) has been used clinically to provide non-invasive continuous monitoring, albeit giving occasional erroneous readings [11]. An mOEPS was applied to carry out a specific study using green and orange illuminations. The raw reflected PPG signal is often considered to be composed of two separate components, namely a direct current (DC) and an alternating current (AC) [11].

The DC component arises from non-pulsatile tissue such as bones and veins that maintain a constant absorption characteristic during measurement. The AC component is the variation in the intensity of reflected light that results from variations in absorption characteristics as the blood pulses through tissue. The amplitude of the AC component is typically only 1% to 2% of that of the DC component [12], making its extraction more susceptible to the presence of measurement and electrical noise. In this study, we proposed that the most encouraging research area would consider reflectance mode operation, as this approach is more likely to provide comfort and convenience for end users.

Current PPG based pulse oximetry sensors mainly use two separate wavelength illuminations, namely red and infrared (IR). However, these two wavelength illuminations have been reported to produce signals of poor reliability during the movements [13]. It would be appropriate to consider green as an alternative illumination colour as it has been reported to provide signals with improved resilience to motion artefacts [11]. In addition, the measurement of certain physiological parameters, such as SpO₂, requires that at least two separate signals be obtained from incident light of different wavelengths. Many less expensive health monitoring products, e.g., smartwatches, tend to use only one wavelength and are consequently restricted to HR measurement only.

In this study, an mOEPS with at least four wavelength illumination sources was used to multiply illumination on a designated tissue and to extract SpO₂ from same device. Conventionally, pulse oximeters with red and IR illuminations are used to measure SpO₂ non-invasively as haemoglobin is the oxygen-transport metalloprotein in red blood cells [14] and the light absorption of oxy-haemoglobin (HbO₂) relative to that of deoxyhaemoglobin (Hb) is dependent on the illumination wavelength. Under red illumination, the absorption of Hb is greater than that of HbO₂, whereas the relative absorption is reversed under IR light. By measuring the absorption by the blood of both red and IR light, it is possible to estimate the value of SpO₂. As tissues in the body generally contain water, they generally absorb light more strongly in the longer wavelength IR region [12], whereas melanin is known to exhibit greater absorption of light at shorter wavelengths [11]. Red blood cells are reported to have greater light absorption at red and near IR wavelengths [15] and this could be used to facilitate the measurement of blood volume changes. Consequently, red and IR illuminations are commonly used as illuminated light sources in pulse oximetry [16]. There are several studies that have used red light (650–750 nm) [17] and IR light (850–1000 nm) [18] to measure SpO₂ using the PPG technique [18]. However, PPG signals are easily affected by the sources of noise from human tissue deformation, ambient light interference and electromagnetic field, thus making good readings of physiological parameters difficult, as anticipated in practice [19].

In this study, 525 nm (green) and 595 nm (orange) wavelength illuminations were investigated as alternative illuminated light sources to be used in the extraction of SpO₂ reading. Green and orange

illuminations are absorbed closer to the surface of the skin in comparison with the depth of absorption of incident red and IR light. Consequently, although green and orange illumination penetrate the skin to a lesser degree, their signal-to-noise ratio is improved as they suffer from less attenuation due to their comparatively shorter path lengths [20]. Figure 1 shows the skin penetration depths of light signals at wavelengths in the range of 400 nm to 1000 nm.




Figure 1. Optical penetration depth δ shown for a range of light sources of different wavelengths [21].

Figure 2 shows that there is a sufficiently significant difference in the absorption of Hb and HbO₂ at each of the green and orange wavelengths to make the method suitable for extracting SpO₂ measurements.




Figure 2. Absorption of Hb and HbO₂ at green and orange illuminations [22].

The aim of the study is to research how to effectively work out SpO₂ by means of mOEPS with its unique multi-wavelength illumination functionality. As the mOEPS has four wavelengths, e.g., green (525 nm), orange (595 nm), red (650 nm) and near Infrared (IR) (870 nm), the green and orange illuminations were used as primary results to extract SpO₂ [23]. The study was specified to be able to operate under a number of different scenarios including physical movements [24]. To achieve relevant

quality signals, the mOEPS was able to be located at various measurement sites of subjects [25] who had a range of skin pigments [26].

The specific objectives of the paper can be stated as:


- i. To study how effectively to obtain physiological parameters SpO₂ from mOEPS using green and orange illuminations.
- ii. To evaluate the performance of mOEPS against a pulse oximetry through implementation of a setup protocol involving healthy subjects.
- iii. To compare the performance of different illuminations (green and orange, red and infrared) from mOEPS to work out the SpO₂ measurement.

2. Methodology


This section describes the approach taken in the current work in the measurement of SpO₂ and provides a description of the sensing and processing systems.

2.1. Experimental Protocol

Thirty-one (31) healthy subjects (gender: male, age: 25 ± 5 years, height: $179 \text{ cm} \pm 4 \text{ cm}$, and Body Mass Index: $22 \pm 1 \text{ kg/m}^2$) volunteered to take part in the study. The experimental protocol (R18-P006) was approved by the Loughborough University Ethical Advisory Committee on 19 February 2018. The protocol was designed into two different groups of 15 and 16 subjects as two sub-protocols, respectively. Both sub-protocols were used to investigate the performance of mOEPS in the extraction of SpO₂ in sitting mode and during physical activities. For the first sub-protocol, the subjects were sitting and gently moving their hands and the green and orange illumination were compared and validated against the commercial pulse oximetry for extracting SpO₂. For the second sub-protocol, the subjects completed the cycling and walking activities and the mOEPS was solely used to compare the SpO₂ of different wavelengths during exercises. All the subjects in both sub-protocols were required to abstain from drinking alcohol, coffee or other substances that may have affected their performance in the 24 h leading up to the taking of the measurements. The subjects' data was recorded using the patch sensor placed in the palm of the left hand. The subjects were free to move their hands as usual. In order to collect benchmark data, TempIRTM pulse oximetry (Shenzhen Jumper Medical Equipment Co., Ltd., Shenzhen, China), which uses optical PPG techniques, was utilized to match with our mOEPS in the first sub-protocol. The commercial pulse oximeter was used to obtain a reference SpO₂ parameter for comparison against the mOEPS obtained parameters. The data recordings from the commercial sensor and the mOEPS were processed and analysed using a bespoke MATLAB programs (R2016b) (MathWorks, Cambridge, UK). The subjects were first asked to remain in a seated, upright position for one minute while the sensors were attached, before the raw data was captured over a period of three minutes during free hand movement in the first sub-protocol. However, in the second sub-protocol, the subjects were asked to perform different activities, such as cycling (20 rpm) and walking (5 km/h), for a period of four minutes per activity. A time window of 10 s duration was used in the calculation SpO₂. Figure 3 shows typical time duration for the captured data.



(a) Raw data



(b) AC components

Figure 3. Data obtained from the mOEPS system (a) the raw green and orange signals; (b) the signals from the four wavelength illuminations after the band pass filtering (0.8–5 Hz). The peaks and troughs in the signals can be used to obtain SpO_2 .

2.2. Extraction of Oxygen Saturation (SpO_2)

The oxygen saturation of the blood is the degree to which oxygen is chemically combined with haemoglobin. SpO_2 can be worked out as the fractional ratio between the concentration of oxygenated haemoglobin and the total haemoglobin present in the blood, as shown in the following Equation (1):

$$SpO_2(\%) = \frac{[HbO_2]}{([HbO_2] + [Hb])} \times 100\%, \quad (1)$$

where [Hb] is the concentration of the deoxyhaemoglobin form and [HbO₂] is the concentration of the oxy-haemoglobin form.

According to Lambert Beer's law [12], the transmission of illumination decays exponentially with the absorption coefficient and the optical path length. The tissues absorb light differently during the diastolic and systolic phases. In this study, green and orange wavelengths were used to extract the SpO₂ and their absorption is expressed in the following Equations (2) and (3):

$$I_{h,G} = I_{in,G} e^{-\mu(\lambda_G) d_{Gmin}}, \quad (2)$$

$$I_{l,G} = I_{in,G} e^{-\mu(\lambda_G) d_{Gmax}}, \quad (3)$$

where $I_{h,G}$ and $I_{l,G}$ are the intensity of the green light transmission during the diastolic phase and systolic phases, respectively, $I_{in,G}$ is the intensity of incident green light, μ is the absorption coefficient, λ is the wavelength of the green light (525 nm) and d_{Gmin} and d_{Gmax} are the minimum and maximum optical path lengths, respectively.

By taking the logarithm of the ratio of transmissions in the diastolic and systolic phases, and, assuming the absorption coefficient is known, the difference between the optical path lengths for the green illumination can be determined by Equation (4);

$$\ln\left(\frac{I_{h,G}}{I_{l,G}}\right) = -\mu(\lambda_G)(d_{Gmax} - d_{Gmin}) = -\mu(\lambda_G)\Delta d_G. \quad (4)$$

Similarly, for the orange illumination, the difference between the optical path lengths is given by Equation (5);

$$\ln\left(\frac{I_{h,O}}{I_{l,O}}\right) = -\mu(\lambda_O)(d_{Omax} - d_{Omin}) = -\mu(\lambda_O)\Delta d_O. \quad (5)$$


The absorption ratio of green and orange wavelengths (R) can be calculated as follows of Equation (6);

$$R = \frac{\ln\left(\frac{I_{h,G}}{I_{l,G}}\right)}{\ln\left(\frac{I_{h,O}}{I_{l,O}}\right)} = \frac{\mu(\lambda_G)\Delta d_G}{\mu(\lambda_O)\Delta d_O}. \quad (6)$$


From the previous equation, $\ln(I_{h,G}/I_{l,G})$ is the arterial transmittance variation under incident green light $\mu(\lambda_G)$ and the arterial transmittance variation under incident orange light $\mu(\lambda_O)$. If it is assumed that both sources of illumination (green and orange) are placed at the same position on the palm, then the optical path lengths (light travel to the skin) are approximately the same ($\Delta d_G \approx \Delta d_O$). The lengths of the green and orange return optical paths from the arterial blood flow to the detector will also be similar. As a result, the absorption ratio depends principally on the relative absorption coefficients obtained at the two wavelengths and is largely independent of the optical path lengths:

$$R = \frac{\mu(\lambda_G)}{\mu(\lambda_O)}. \quad (7)$$

In general, SpO₂ can also be extracted from signals captured by PPG sensors with two different wavelength illuminations and one photo sensor. In the current work, the green and orange illumination sources were used for both signals of two different wavelength measurements obtained from the DC and AC components, which represent the absorption light. The DC components can be extracted from the raw data as shown in Figure 4a, and the AC amplitude can be determined from the positions of peaks and troughs as shown in Figure 4b.



(a) Determination of the DC component



(b) Determination of AC amplitude

Figure 4. Measurement of PPG signals to extract the DC and AC components. (a) the raw data contains both information of non-pulsatile (DC) and a relatively small variation resulting from pulsatile (AC) absorption; (b) filtering of the raw data can be used to obtain the dynamic AC change.

The absorption ratio can be estimated from a ‘ratio of ratios’ measurement as follows:

$$R = \frac{\mu(\lambda_G)}{\mu(\lambda_O)} = \frac{\left(\frac{AC_{green}}{DC_{green}}\right)}{\left(\frac{AC_{orange}}{DC_{orange}}\right)}. \quad (8)$$

The oxygen saturation can then be calculated from Equation (8);

$$SpO_2 = A - BR, \quad (9)$$

where A and B are coefficients that are obtained from a best-fit linear regression equation relating to the standard pulse oximeter series reading against the ratio R of mOEPS. Table 1 shows the ratio values (R) obtained from the mOEPS and the SpO₂ values obtained from the pulse oximeter.

Table 1. The values of Ratio (R) and SpO_2 (%) used in the calculation of the coefficients A and B.

Ratio of Green/Orange in mOEPS	SpO_2 (Pulse Oximetry)
0.153	98
0.156	98
0.150	98
0.176	97
0.168	97
0.170	97
0.190	96
0.150	99
0.147	96
0.151	99
0.154	98
0.143	99
0.148	99
0.153	99

2.3. Experimental Setup

The mOEPS was developed to capture PPG signals at green, orange, red and IR illuminations. A microcontroller was used to control the four wavelength illuminations provided by the PPG sensor and capture the reflected scattered light via a photodetector as shown in Figure 5. The multi-wavelength illumination was synchronized with sixteen LEDs, arranged as four channels of LEDs with four LEDs in each channel to provide the same wavelength illumination. The LEDs in each channel were driven in a cyclical sequence as the four LEDs in each channel were illuminated at the same time, when the LEDs in the other three channels were turned off. The signals captured from the individual wavelength illuminations of mOEPS were pre-amplified, then demultiplexed to determine each captured signal corresponding to the specified channel of wavelength illuminations, i.e., green, orange, red and IR. The captured signals were thus passed individually to be processed further as requested to extract relevant physiological parameters, i.e., SpO_2 .




Figure 5. Microcontroller control of the multi-wavelength illumination and de-multiplexing operations. In the mOEPS, the detail of four illumination are: the channel 1: green, the channel 2: orange, the channel 3: red, and the channel 4: near infrared(IR).

The mOEPS system consists of (1) signal conditioning analogue front-end microelectronics, (2) digital signal processing unit (DSP), and (3) a communication port for either data storage or wirelessly transmitted to a PC for offline signal processing. All data sets obtained from the mOEPS were collected by the means of a 4-channel PPG board (DISCO4, Dialog Devices Ltd., Reading, Berkshire, UK). The analogue-to-digital conversion used to capture PPG signals was implemented on a 14-bit data acquisition board (DAQ, USB-6009, National Instruments Co., Novato, CA, USA) and the control software of the PPG board was implemented in a LabVIEW GUI (National Instruments Co., USA). A band pass filter range of 0.8 Hz to 5 Hz was implemented to reduce the influence of extraneous noise and a sampling frequency of 256 Hz was sufficient to reconstruct the PPG signals considering the Nyquist criteria.

3. Results

The comparative results of SpO₂ (%) are presented to express the measurements using the commercial pulse oximetry, the green and orange illuminations and red and infrared illuminations of the mOEPS. The normality of the distributions of SpO₂ (%) data was tested. A *t*-test was carried out on the SpO₂ (%) data to find the correlations between the SpO₂ (%) readings in each sub-protocol.

3.1. Oxygen Saturation: SpO₂(%)

3.1.1. First Sub-Protocol: Free Hand Movement

The green and orange illuminations from the mOEPS were used to extract SpO₂, and their results were compared with those of a commercial pulse oximetry device (TempIR™, Shenzhen Jumper Medical Equipment Co., Ltd., Shenzhen, China). Figure 6a (subject I) shows well-matched readings of SpO₂ between mOEPS and pulse oximetry. However, for subject II in Figure 6b, there was a significant difference of three units with a single SpO₂ reading of 98% from TempIR™, and 95% from the mOEPS, respectively. The overall SpO₂ readings, as shown in Figure 6c, for all 15 subjects show a high agreement with a correlation of $r = 0.98$.




Figure 6. Cont.




Figure 6. SpO₂ data obtained from the mOEPS and a standard pulse oximetry device. The graphs of (a–c) display a line correlation for subject I, subject II and all 15 subjects respectively.

The means from the Bland–Altman method [27] were used to evaluate the agreement of SpO₂ readings between the mOEPS and the pulse oximetry. Firstly, the analysis was carried out with each subject to make sure that the data was matched and then the entire data of all 15 subjects were combined together to produce an overall agreement. Figure 7 demonstrates the Bland–Altman plots for all 15 subjects.




Figure 7. Bland–Altman plot showing differences in SpO₂ between the green and orange illuminations of mOEPS and the pulse oximetry device. The graphs display a scatter diagram of the differences against the averages of the pulse oximetry and the green and orange illuminations of mOEPS for all 15 subjects.

A *t*-test was carried out on the SpO₂ data and a strong relationship was found between the SpO₂ measured by using the mOEPS and the pulse oximetry, and vice versa ($r = 0.98$). Figure 8 shows the box-and-whisker plots for subject I, II and 15 subjects, respectively. As per Figure 8, a significant difference can be seen with the reading of subject II. However, when all the data sets of SpO₂ for all subjects were aggregated, the overall box-and-whisker plot is comparable and shows no significant difference with value ($p = 0.88$), as shown in Figure 8.




Figure 8. SpO₂ readings with box-and-whisker plots and median values of the pulse oximetry device and the green and orange illuminations of mOEPS. Subject I (Pulse Oximetry: 97% and mOEPS: 96%), subject II (Pulse Oximetry: 98% and mOEPS: 96%) and all 15 subjects readings (Pulse Oximetry: 98% and mOEPS: 98%). All SpO₂ data sets were processed with box-and-whisker procedures to obtain the lower, median and higher values.

The comparison study between the standard pulse oximetry (TempIRTM) and the green and orange illuminations of mOEPS was performed to obtain SpO₂ readings. The results of SpO₂ indicate that both measurements readings have a very similar range of variation ($\sigma^2 = \pm 1$). The mean difference for the both techniques is 0.03 as anticipated in a higher correlation. The star sign (*) identifies the readings of SpO₂ that are of a very low value due to a subject's sweaty skin during the experiment implementation.

3.1.2. Second Sub-Protocol: Physical Movement

The results of this protocol were extracted by using mOEPS wavelength illuminations. The SpO₂ reading obtained from green and orange illuminations were compared with red and infrared illuminations during physical exercises. Figure 9 shows the overall SpO₂ readings for all 16 subjects in cycling and walking. The result shows well-matched readings of SpO₂ between green/orange and red/infrared in both exercises with a correlation of ($r = 0.98$).




Figure 9. SpO₂ data obtained from the mOEPS (green/orange) and the mOEPS (red/infrared) for different exercises.

The agreements of SpO₂ readings obtained from green/orange and red/infrared were validated by using the Bland–Altman plot. Figure 10a,b demonstrates the Bland–Altman plots for cycling and walking for all 16 subjects.




Figure 10. Bland–Altman plot showing differences in SpO₂ between the green and orange illuminations of mOEPS and the red and infrared illuminations of mOEPS for (a) cycling and (b) walking. The graphs display a scatter diagram of the differences against the averages of the green/orange and red/infrared illuminations of mOEPS for 16 subjects.

The comparison results of mOEPS was performed to obtain SpO₂ readings from different illuminations. The results of SpO₂ indicate that both measurements readings have a very similar range of variation ($\sigma^2 = \pm 1$) with the probability value consistently being 0.88. The star sign (*) identifies the readings of SpO₂ that are of a very low value due to a subject's sweaty skin during the experiment implementation.

4. Discussion

The mOEPS sensor provides multi-wavelength illumination sources to enable optoelectronic sensor based physiological monitoring more optional on signal processing to extract vital signs, i.e., SpO₂. The experimental performance has demonstrated that the measurements with the mOEPS were less affected by small movements compared to those with the commercial pulse oximetry. For example, if the incident light from one LED, or even one channel of LEDs, becomes misaligned with the capillary arteries, at least one of the other source lights aligns with the same lay of the sensor; thus, the mOEPS is still likely to have sufficient alignment with the area of interest to be able to capture a useful reflected signal.

Referring to Figure 2, there are differences in the absorption of Hb and HbO₂ at the illuminated wavelengths identified as green and orange, making these better suited for the extraction of SpO₂ values. At these wavelengths, the penetration into the skin is significantly shallower compared with red and IR illuminations, implying that the use of green and orange illuminations for SpO₂ can work during free hand movement. The subjects as implemented in the first sub-protocol, felt free to do anything, such as typing, pointing, holding and making unconscious hand movements during speaking, unlike some restrictive procedures and setups for measuring SpO₂ from present pulse oximetry devices.

The *t*-tests were applied in this study to determine whether there was a significant difference between means of the mOEPS and SpO₂ worked out. The correlation and the Bland–Altman plot were also utilized to evaluate the agreement between the measurement techniques in the both sub-protocols. The SpO₂ readings from mOEPS were validated against the values of SpO₂ obtained from the pulse oximeter in the first sub-protocol, whereas mOEPS was only used in the second sub-protocol to verify the SpO₂ worked out from two set of illuminations. During the resting mode, the results of SpO₂ readings were found to be in a higher agreement, as shown in Figure 6, with $r = 0.98$. The correlation of cycling and walking showed a good agreement of $r = 0.98$ as shown in Figure 9. As depicted in the Bland–Altman plot in Figures 7 and 10, the comparison of the SpO₂ from both designated sub-protocols are presented, where the points represent individual readings of SpO₂.

All data readings of SpO₂ are gathered together and displayed in the overall box-and-whisker plots. The results are comparable and there is no significant difference, with the value of $p = 0.88$, as shown in Figure 8. However, there was a three unit difference between pulse oximeter and mOEPS reading of SpO₂ for subject II due to sweaty skin during the experiment.

The SpO₂ readings are in the acceptable range of $B \pm 1.96$ SD of the differences between the Pulse oximeter and the mOEPS signals as well as differences between both sets (green/orange and red/infrared) of mOEPS signals.

The average of the bias difference is close to zero and most of the points fell within the upper and lower limit while some points were presented out of the range due to different readings between the commercial device and mOEPS. These points are out of the line regressions and acceptable range, which indicates that there are some variations of these readings between the Pulse oximeter and the mOEPS due to fact that both sensors might have misaligned with area of interest. In addition, the SNR of Pulse oximeter is decreased with movements.

Tables 2 and 3 represent the SpO₂ readings that imply that both measurements of both sub-protocols have a very similar range of variation $\sigma^2 = \pm 1$. The outcome from the present study suggests that both green and orange illuminations of mOEPS are practicable for extracting SpO₂ readings during free hand movement and physical activity.

Table 2. Typical SpO₂ (percentage points) comparison between the standard pulse oximetry and the green and orange illuminations of mOEPS. The star sign (*) identifies the readings of SpO₂ that are of a very low value due to a subject's sweaty skin during the experiment implementation.

Pulse Oximetry	mOEPS (Green/Orange)	Variation (Percentage Points)	Pulse Oximetry	mOEPS (Green/Orange)	Variation (Percentage Points)
94	95	-1	96	93 *	3
93 *	93 *	0	95	96	-1
96	97	1	96	96	0
97	97	0	97	98	-1
96	96	0	97	97	0
96	95	1	97	98	-1
96	96	0	97	98	-1
96	92 *	4	97	97	0
98	97	1	96	95	1
98	97	1	95	94	1
98	98	0	98	98	0
97	97	0	97	96	1
97	96	1	97	94	3
97	95	2	97	96	1
94	94	0	97	97	0
95	90	5	97	97	0
95	95	0	96	96	0
97	97	0	96	96	0
97	98	-1	97	96	1
97	97	0	98	97	1
96	97	-1	97	96	1
97	98	-1	98	98	0
97	93 *	4	97	97	0
97	98	-1	97	98	1
97	97	0	99	99	0

Table 3. Typical SpO₂ (percentage points) comparison between the green/orange and red/infrared illuminations of mOEPS. The star sign (*) identifies the readings of SpO₂ that are of a very low value due to a subject's sweaty skin during the experiment implementation.

Cycling			Walking		
mOEPS (Green/Orange)	mOEPS (Red/Infrared)	Variation (Percentage Points)	mOEPS (Green/Orange)	mOEPS (Red/Infrared)	Variation (Percentage Points)
99	99	0	97	97	0
98	98	0	98	98	0
98	98	0	98	98	0
98	98	0	98	98	0
99	99	0	98	98	0
98	98	0	95	95	1
98	98	0	96	96	1
98	98	0	94	93 *	1
99	99	0	97	97	0
99	99	0	96	96	0
98	97	1	99	99	0
98	98	0	97	97	0
99	98	1	95	95	0
98	97	1	99	99	0
99	99	0	99	99	0
98	98	0	99	99	0
95	96	-1	97	97	0
98	98	0	97	97	0
99	99	0	99	99	0
98	98	0	99	99	0
98	98	0	99	99	0
97	98	-1	96	96	0
99	99	0	98	98	0
98	97	1	98	98	0
97	96	1	98	98	0
99	99	0	98	99	-1

5. Conclusions

The green and orange illuminations of mOEPS demonstrated that the readings range of SpO₂ (98–99%) can be extracted in a range of physiological monitoring application scenarios. The results of SpO₂ extracted from the mOEPS have demonstrated, (1) an excellent correlation $r = 0.98$ between the responses of green and orange illuminations, and the pulse oximetry with a statistical test $p = 0.88$; and (2) a same quality correlation $r = 0.98$ between the responses of green and orange illuminations, and red and IR illuminations in the same mOEPS with a statistical test $p = 0.88$.

Due to the nature of multi-wavelength illuminations in the mOEPS, illumination adaption can be performed in the mOEPS to obtain SpO₂ readings with a certain degree of tolerance to different skin types and skin thicknesses [25]. The forthcoming study to validate SpO₂ readings from the green and orange illuminations from mOEPS will be carried out in compliance with the ISO 80601-2-61, 2017 [28] and the FDA Pulse Oximeters—Premarket Notification Submissions [510(k)s], 2017 [29]. The study will be working closely with clinical professionals along with these standards in a gas controllable chamber/hypoxic condition against a clinical golden standard of blood gas analyzer for SpO₂ reading accuracy.

Thus, the mOEPS could be consolidated with currently available wearable and smart devices for forthcoming real-time physiological monitoring to satisfy the requirements of present clinic monitoring and assessment. In addition, the mOEPS would appear to be of value in applications of sport physiological monitoring and assessment during physical movements, and is also likely to have its applications in personal healthcare, where it can provide essential vital signs with an indeed cost-effective solution.

Furthermore, one observation from the current study is that the mOEPS could provide an alternative opto-physiological monitoring solution that would be well suited to incorporation into wearable devices.

As part of the study of a wearable solution, the simplification of the controlling electronics and the embedding data processing algorithms will be considered. Several challenges remain to be overcome for the mOEPS through improvements in packaging, automated adaption to different subjects and the provision of users with recommendations for suitable operational procedures to satisfy requirements of cost-effective physiological monitoring in individuals or clinical settings.

Author Contributions: S.A. (Samah Alharbi) designed the experimental protocol, carried out the signal processing and prepared the manuscript. D.M. advised on the electronic works. L.Y. implemented the mOEPS system setting, P.B. executed two sub-protocols. Y.E. assisted signal processing and manuscript proofreading. S.A. (Samuel Adema) executed the second sub-protocol. L.B. designed and supervised the second sub-protocol. S.H. supervised the project and guided the project implementation, and organized the manuscript.

Funding: EPSRC CDT-Enabling Technologies.

Acknowledgments: The authors would like to acknowledge Loughborough University, the Ministry of Higher Education in the Kingdom of Saudi Arabia and EPSRC CDT-Enabling Technologies for supporting this research. We also would like to convey our special thanks to Mortara Instrument (Stirling, UK) Ltd. for lending us a trial ECG device.

Conflicts of Interest: The authors declare no conflict of interest.

References

1. Lopez, S. *Pulse Oximeter Fundamentals and Design*; Application Note Document No. AN4327, Rev 1; Free-Scale Semiconductor Inc.: Austin, TX, USA, 2004; p. 4327. Available online: <https://www.nxp.com/docs/en/application-note/AN4327.pdf> (accessed 27 December 2017).
2. H3 SYSTEM HRM-3200 Heart Rate Monitor. Available online: <https://usermanual.wiki/H3-SYSTEM/HRM-3200> (accessed on 20 May 2017).
3. Huawei Honor Watch S1. Available online: <http://www.smartwatchspecifications.com/Device/huawei-honor-watch-s1-glory-watch-s1> (accessed on 10 February 2017).
4. MZ3 Belt | MYZONE. Available online: <http://myzone.org/meet-the-mz3/> (accessed on 28 March 2017).

5. Nonin Medical Inc. 13700 1st Avenue North, P. Nonin Medical Inc. The WristOx 3100 Pulse Oximeter. Available online: <http://www.nonin.com/WristOx3100> (accessed on 28 June 2016).
6. Anliker, U.; Ward, J.; Lukowicz, P.; Troster, G.; Dolveck, F.; Baer, M.; Keita, F.; Schenker, E.; Catarsi, F.; Coluccini, L.; et al. AMON: A Wearable Multiparameter Medical Monitoring and Alert System. *IEEE Trans. Inf. Technol. Biomed.* **2004**, *8*, 415–427. [[CrossRef](#)] [[PubMed](#)]
7. Grossman, P. The LifeShirt: A multi-function ambulatory system monitoring health, disease, and medical intervention in the real world. *Stud. Health Technol. Inform.* **2004**, *108*, 133–141. [[PubMed](#)]
8. Park, S.; Jayaraman, S. E-Health and quality of life: The role of the Wearable Motherboard. *Stud. Health Technol. Inform.* **2004**, *108*, 239–252. [[PubMed](#)]
9. Fruin, M.; Rankin, J. Validity of a Multi-Sensor Armband in Estimating Rest and Exercise Energy Expenditure. *Med. Sci. Sport Exerc.* **2004**, *36*, 1063–1069. [[CrossRef](#)]
10. Pandian, P.; Mohanavelu, K.; Safeer, K.; Kotresh, T.; Shakunthala, D.; Gopal, P.; Padaki, V. Smart Vest: Wearable multi-parameter remote physiological monitoring system. *Med. Eng. Phys.* **2008**, *30*, 466–477. [[CrossRef](#)] [[PubMed](#)]
11. Tamura, T.; Maeda, Y.; Sekine, M.; Yoshida, M. Wearable Photoplethysmographic Sensors—Past and Present. *Electronics* **2014**, *3*, 282–302. [[CrossRef](#)]
12. Jayasree, V.K. Selected Cardiovascular Studies Based on Photoplethysmography Technique. Ph.D. Thesis, International School of Photonics Cochin University of Science and Technology, Cochin, India, 2009.
13. Lee, J.; Matsumura, K.; Yamakoshi, K.; Rolfe, P.; Yamakoshi, T. Comparison between Red, Green and Blue Light Reflection Photoplethysmography for Heart Rate Monitoring During Motion. In Proceedings of the 35th Annual International Conference of the IEEE Engineering in Medicine and Biology Society, Osaka, Japan, 3–7 July 2013; pp. 1724–1727.
14. Strojnik, M.; Paez, G. Spectral dependence of absorption sensitivity on concentration of oxygenated hemoglobin: Pulse oximetry implications. *J. Biomed. Opt.* **2013**, *18*, 108001. [[CrossRef](#)] [[PubMed](#)]
15. Uyuklu, M.; Canpolat, M.; Meiselman, H.; Baskurt, O. Wavelength selection in measuring red blood cell aggregation based on light transmittance. *J. Biomed. Opt.* **2011**, *16*, 117006. [[CrossRef](#)] [[PubMed](#)]
16. Nitzan, M.; de Boer, H.; Turivnenko, S.; Babchenko, A.; Sapoznikov, D. Power spectrum analysis of spontaneous fluctuations in the photoplethysmographic signal. *J. Basic Clin. Physiol. Pharmacol.* **1994**, *5*, 269–276. [[CrossRef](#)] [[PubMed](#)]
17. Bagha, S.; Shaw, L. A real time analysis of PPG signal for measurement of SpO₂ and pulse rate. *Int. J. Comput. Appl.* **2011**, *36*, 45–50.
18. Sonoda, K.; Kishida, Y.; Tanaka, T.; Kanda, K.; Fujita, T.; Maenaka, K.; Higuchi, K. Wearable Photoplethysmographic Sensor System with PSoC Microcontroller. In Proceedings of the 5th International Conference on Emerging Trends in Engineering and Technology, Himeji, Japan, 5–7 November 2012; pp. 61–65.
19. Hu, S.; Azorin-Peris, V.; Zheng, J. Opto-Physiological Modeling Applied to Photoplethysmographic Cardiovascular Assessment. *J. Healthc. Eng.* **2013**, *4*, 505–528. [[CrossRef](#)] [[PubMed](#)]
20. Vizbara, V.; Sološenko, A.; Stankevičius, D.; Marozas, V. *Comparison of Green, Blue and Infrared Light in Wrist*; Biomedical Engineering Institute, Kaunas University of Technology: Kaunas, Lithuania, 2013; pp. 78–81.
21. Abay, T.; Kyriacou, P. Reflectance Photoplethysmography as Noninvasive Monitoring of Tissue Blood Perfusion. *IEEE Trans. Biomed. Eng.* **2015**, *62*, 2187–2195. [[CrossRef](#)] [[PubMed](#)]
22. Optical Absorption of Hemoglobin. Available online: <https://omlc.org/spectra/hemoglobin/> (accessed on 24 May 2017).
23. Alharbi, S.; Hu, S.; Mulvaney, D.; Blanos, P. An applicable approach for extracting human heart rate and oxygen saturation during physical movements using a multi-wavelength illumination optoelectronic sensor system. *Proc. SPIE* **2018**, *10486*. [[CrossRef](#)]
24. Alzahrani, A. A Reflectance Optoelectronic Patch Sensor for Physiological Monitoring with Motion Artefacts Reduction. Ph.D. Thesis, Loughborough University, Loughborough, UK, 2016.
25. Alzahrani, A.; Hu, S.; Azorin-Peris, V.A. Comparative Study of Physiological Monitoring with a Wearable Opto-Electronic Patch Sensor (OEPS) for Motion Reduction. *Biosensors* **2015**, *5*, 288–307. [[CrossRef](#)] [[PubMed](#)]
26. Yan, L.; Hu, S.; Alzahrani, A.; Alharbi, S.; Blanos, P. A Multi-Wavelength Opto-Electronic Patch Sensor to Effectively Detect Physiological Changes against Human Skin Types. *Biosensors* **2017**, *7*, 22. [[CrossRef](#)] [[PubMed](#)]

27. Bland, J.; Altman, D. Statistical methods for assessing agreement between two methods of clinical measurement. *Int. J. Nurs. Stud.* **2010**, *47*, 931–936. [[CrossRef](#)]
28. ISO 80601-2-61:2017. *Medical Electrical Equipment Part 2-61: Particular Requirements for Basic Safety and Essential Performance of Pulse Oximeter Equipment*. Available online: <https://www.iso.org/standard/67963.html> (accessed on 31 December 2018).
29. US Food and Drug Administration (FDA). *Pulse Oximeters—Premarket Notification Submissions [510(k)s]: Guidance for Industry and Food and Drug Administration Staff*; FDA: Silver Spring, MD, USA, 2013. Available online: <https://www.fda.gov/RegulatoryInformation/Guidances/ucm341718.htm> (accessed on 31 December 2018).



© 2018 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).