

This item was submitted to Loughborough's Institutional Repository (<https://dspace.lboro.ac.uk/>) by the author and is made available under the following Creative Commons Licence conditions.



For the full text of this licence, please go to:  
<http://creativecommons.org/licenses/by-nc-nd/2.5/>

# **The CURE (Community Urgent Response Environment): Pods and Packs for Pre-Hospital Care**

Sue Hignett<sup>1</sup>, Mike Fray<sup>1</sup>, Jonathan Bengler<sup>2</sup>, Andrew Jones<sup>3</sup>, David Coates<sup>4</sup>, Neil Mansfield<sup>1</sup>

<sup>1</sup>*Healthcare Ergonomics and Patient Safety Unit, Loughborough University, UK*

<sup>2</sup>*Academic Dept of Emergency Care, University of the West of England, Bristol, UK*

<sup>3</sup>*Openhouse Products Ltd, Birkenhead, UK*

<sup>4</sup>*Great Western Ambulance Service NHS Trust, Chippenham, UK*

## *Context of Research*

About 40% of the 10.3 million visits to NHS emergency departments in England in 2009/10 ended with the patient just needing advice and no actual treatment<sup>1</sup>. It has previously been suggested that these needs could be met in the community through the delivery of urgent (or pre-hospital) care<sup>2</sup> by Emergency Care Practitioners (ECP)<sup>3</sup>. A first research project (Smart Pods) was carried out with 6 NHS Trusts (acute, primary care and ambulance) with 125 staff and 88 patients to explore the technology requirements needed to support this new professional role<sup>4</sup>.

## *Research question*

This paper describes how the outputs from Smart Pods have been taken forward to develop a market-ready pod/pack system and a design specification for the ambulance interface in a knowledge transfer collaboration between 2 universities, 2 manufacturers and the NHS to develop a Community Urgent Response Environment (CURE).

## *Method*

An iterative prototyping approach was taken (figure one) starting with the 8 design concepts and equipment/consumable requirements from Smart Pods. These manifested as pod/pack prototype 1.0 and were reviewed by clinical, design and ergonomic experts to produce prototype 1.1. This was tested in 4 simulated clinical scenarios with ECPs (n=10) and patient actors. The face validity of the simulations was maximized by using, for example, pig skin for suturing. After Pod/Pack User Trials (1) the vehicle interface design commenced (February 2011). This followed a similar iterative prototyping approach (in conjunction with the pod/pack prototype ergonomic and design review). User trials (2) will use clinical scenarios to test the functionality and usability of pod/pack and vehicle interface designs. The final output will be a market-ready pod/pack system and a design specification for the ambulance interface.

### *Results and conclusion*

The results from Pod/Pack User trial (1) will be reported at the conference.

### *References*

1. HESonline. A&E Data (experimental statistics) 2009-10  
<http://www.hesonline.nhs.uk/Ease/servlet/ContentServer?siteID=1937&categoryID=1272>  
(accessed 21 January 2011)
2. DH (2005). *Taking Healthcare to the Patient. Transforming NHS Ambulance Services.*  
<http://www.dh.gov.uk/assetRoot/04/11/42/70/04114270.pdf> (Accessed 2 Sept. 2005)
3. Department of Health (2004) *The Emergency Care Practitioner report – Right skill, Right time, Right place,*  
[http://www.dh.gov.uk/en/Publicationsandstatistics/Publications/PublicationsPolicyAndGuidance/DH\\_4093086](http://www.dh.gov.uk/en/Publicationsandstatistics/Publications/PublicationsPolicyAndGuidance/DH_4093086) (Accessed 21<sup>st</sup> October 2009).
4. Hignett, S., Jones, A., Bengner, J. (2010) Designing a modular work space for urgent care. *Journal of Paramedic Practice*, 2, 3, 116-122




Figure 1. Iterative prototyping process used in CURE