



A comparison of the force applied on oral structures during intubation attempts between the Pentax-AWS airwayscope and the Macintosh laryngoscope: a high-fidelity simulator-based study

The Harvard community has made this article openly available. Please share how this access benefits you. Your story matters.

Citation	Goto, Tadahiro, Yasuaki Koyama, Takashiro Kondo, Yusuke Tsugawa, and Kohei Hasegawa. 2014. "A comparison of the force applied on oral structures during intubation attempts between the Pentax-AWS airwayscope and the Macintosh laryngoscope: a high-fidelity simulator-based study." BMJ Open 4 (10): e006416. doi:10.1136/bmjopen-2014-006416. http://dx.doi.org/10.1136/bmjopen-2014-006416.	
Published Version	doi:10.1136/bmjopen-2014-006416	
Accessed	February 17, 2015 2:40:40 AM EST	
Citable Link	http://nrs.harvard.edu/urn-3:HUL.InstRepos:13347425	
Terms of Use	This article was downloaded from Harvard University's DASH repository, and is made available under the terms and conditions applicable to Other Posted Material, as set forth at <a href="http://nrs.harvard.edu/urn-3:HUL.InstRepos:dash.current.terms-of-use#LAA">http://nrs.harvard.edu/urn-3:HUL.InstRepos:dash.current.terms-of-use#LAA</a>	

(Article begins on next page)

**Open Access** Research

# BMJ Open A comparison of the force applied on oral structures during intubation attempts between the Pentax-AWS airwayscope and the Macintosh laryngoscope: a high-fidelity simulator-based study

Tadahiro Goto, <sup>1</sup> Yasuaki Koyama, <sup>2</sup> Takashiro Kondo, <sup>3</sup> Yusuke Tsuqawa. <sup>4</sup> Kohei Hasegawa<sup>5</sup>

To cite: Goto T, Koyama Y, Kondo T, et al. A comparison of the force applied on oral structures during intubation attempts between the Pentax-AWS airwayscope and the Macintosh laryngoscope: a high-fidelity simulator-based study. BMJ Open 2014;4: e006416. doi:10.1136/ bmjopen-2014-006416

Prepublication history for this paper is available online. To view these files please visit the journal online (http://dx.doi.org/10.1136/ bmjopen-2014-006416).

Received 19 August 2014 Revised 15 September 2014 Accepted 23 September 2014

# **ABSTRACT**

Objective: We sought to determine whether the use of Pentax-AWS Airwayscope (AWS) applied less force on oral structures during intubation attempts than a conventional direct laryngoscope (DL).

**Design:** Prospective cross-over study.

Participants: A total of 37 physicians (9 transitionalyear residents, 20 emergency medicine residents and 8 emergency physicians) were enrolled.

**Interventions:** We used four simulation scenarios according to the difficulty of intubation and devices and used a high-fidelity simulator to quantify the forces applied on the oral structures.

**Outcome measures:** Primary outcomes were the maximum force applied on the maxillary incisors and tongue. Other outcomes of interest were time to intubation and glottic view during intubation attempts. **Results:** The maximum force applied on the maxillary incisors in the normal airway scenario was higher with the use of AWS than that with DL (107 newton (N) vs 77 N, p=0.02). By contrast, the force in the difficult airway scenario was significantly lower with the use of AWS than that of the DL (89 N vs 183 N, p<0.01). Likewise, the force applied on the tongue was significantly lower with the use of AWS than the use of DL in both airway scenarios (11 N vs 27 N, p<0.001 in the normal airway scenario; 12 N vs 40 N, p<0.01 in the difficult airway scenario).

**Conclusions:** The use of AWS during intubation attempts was associated with decreased forces applied to oral structures in the simulated difficult airway scenario.

# Strengths and limitations of this study

- This cross-over study is the first to assess the difference in the actual force on oral structures during intubation attempts with the Pentax-AWS Airwayscope and the direct laryngoscope, by using high-fidelity simulators.
- Approximately 50% of the participants had experienced 100 or fewer intubations in their medical career.
- As our study used simulators, caution should be exercised when extrapolating these findings to a clinical setting.

requiring additional forces. Several studies have reported that the use of AWS is associated with faster intubation in cases of difficult airways1 4 and lower incidence of dental injury than DL. Potential mechanism by which the use of AWS reduces airway trauma is a reduction in applied forces on the oral structures. However, to our knowledge, differences in applied forces during intubation attempts between the AWS and the DL remain unclear.

To address the knowledge gap, using highfidelity simulators, we sought to determine whether the use of an AWS decreases applied forces on the maxillary incisors and tongue during intubation attempts compared to those with the use of a DL.



For numbered affiliations see end of article.

Correspondence to

Dr Tadahiro Goto; gto@u-fukui.ac.jp, qq\_gto@yahoo.co.jp

# INTRODUCTION

While direct laryngoscopes (DLs) have been widely used for tracheal intubations, video laryngoscopes have been used in the operating room, intensive care unit and emergency department. 1-3 The Pentax-AWS Airwayscope (AWS) is a rigid video laryngoscope that uses an anatomically-shaped blade and video system to obtain a better glottic view without

# **METHODS**

# Study design and settings

We conducted a prospective cross-over study to examine the forces applied on the oral structures during intubation attempts according to intubation devices. This design enabled each study participant to serve as his or her own control, thereby removing both measured and unmeasured time-invariant confounding.<sup>6</sup>

We recruited 37 physicians across Japan in August 2013, including transitional-year residents (postgraduate year (PGY) 1 or 2), emergency medicine residents (PGY 3, 4 and 5) and emergency physicians (PGY ≥6). The study was approved by the institutional ethics committee of St. Marianna University School of Medicine Hospital, and informed consent was obtained from the participants.

# **Data collection**

### Baseline characteristics

We documented the characteristics of each participant (PGY and level of training), and self-reported the number of airway management encounters prior to this study (the number of intubations, difficult airway encounters and intubations using an AWS).

# Simulation scenarios

We used a high-fidelity airway management simulator (Waseda Kyoto Airway No.5, Kyoto-Kagaku, Kyoto, Japan) to quantify the forces applied on the oral structures,—the maxillary incisors and tongue—during intubation attempts. The implanted sensors in the simulator automatically quantify the forces. The intubation difficulty can be changed by limiting mouth opening.

We used four simulation scenarios according to the difficulty of intubation and devices (a size-3 Macintosh DL and AWS) as follows: (1) intubation of a normal airway using a DL, (2) intubation of a difficult airway using a DL, (3) intubation of a normal airway using an AWS and (4) intubation of a difficult airway using an AWS. We defined a 'difficult airway scenario' as a scenario with a limited mouth opening. Prior to simulations, all participants received a 10 min lecture and 5 min practice session to ensure that they were familiar with proper DL and AWS techniques. The participants were then randomly assigned to one of the four simulation scenarios, and sequentially underwent the other simulation scenarios. All participants were blinded to the difficulty of intubation in all scenarios.

### **Outcome measures**

Primary outcomes were the maximum applied forces on the maxillary incisors and tongue during intubation attempts. Other outcomes of interest were time to intubation and glottic view during intubation attempts. Time to intubation was defined as time to successful placement of an endotracheal tube (ie, an appropriate positioning of the endotracheal tube tip and confirmation of ventilation), regardless of the number of attempts. The glottic view at each intubation attempt was scored by the participant using the Cormack-Lehane grades. 9

# Statistical analysis

To account for the natural pairing of the observations within each participant, and non-normal distribution of the outcome variables, we compared the outcomes between two devices (AWS vs DL) using Wilcoxon signed

rank test. In addition, to compare the outcomes with consideration for the intubator's experience, we performed stratified analysis by categorising the intubators into two groups based on the previous study: experienced intubators (n=18) and inexperienced intubators (n=19). Experienced intubators were defined as those who had intubated 100 or more cases, while inexperienced intubators were defined as those who had intubated less than 100 cases. p Values <0.05 were considered statistically significant. All data analyses were performed with Stata software V.13.1 (StataCorp, College Station, Texas, USA).

### **RESULTS**

The characteristics of participants are shown in table 1. Of the 37 participants, 20 were emergency medicine residents. The overall median number of intubations was 80 (IQR, 35–150), with a median of 4 (IQR, 1–10) among those using an AWS.

The maximum force applied on the maxillary incisors in the normal airway scenario was higher with the use of an AWS (107 newton (N) vs 77 N, p=0.02; table 2) than that of a DL. By contrast, the force in the difficult airway scenario was significantly lower with the use of an AWS than that of a DL (89 N vs 183 N, p<0.01). Likewise, the force applied on the tongue was significantly lower with the use of an AWS than that of the use of a DL in both airway scenarios (11 N vs 27 N (p<0.01) in the normal airway scenario; 12 N vs 40 N (p<0.01) in the difficult airway scenario). There were no significant differences in time to intubation by airway device in both scenarios. By contrast, the Cormack-Lehane grade score was significantly lower with the use of AWS than that of a DL in the difficult airway scenario (median, 1.0 vs 2.0, p<0.01).

The sensitivity analysis stratified by the experience of intubators showed the robustness of our findings. For example, in the experienced intubator group, the forces applied on oral structures were significantly lower with

Table 1 Characteristics of participants	
Characteristics	All participants n=37
PGY median (IQR)	4 (2.5–6)
Level of training (%)  Transitional year residents*	9 (24)
Emergency medicine residents	20 (54)
Emergency physicians†	8 (22)
Airway management	
Total number of intubations, median (IQR)	80 (35–150)
Number of difficult airway	5 (2–10)
encounters, median (IQR)	4 (4 40)
Number of intubation using AWS, median (IQR)	4 (1–10)
*Defined as PGY 1 or 2. †Defined as PGY ≥6. AWS, Airwayscope; PGY, post-graduate year	S.

Table 2 Comparisons of Air	rwayscope with the direct laryngoscop	e for intubation	
Outcomes	AWS n=37	DL n=37	p Value
Forces on maxillary incisors	(N), median (IQR)		
Normal airway	107 (54–127)	77 (44–96)	0.02
Difficult airway	89 (6–141)	183 (170–186)	<0.01
Applied forces on tongue (N)	, median (IQR)		
Normal airway	11 (8–14)	27 (21–39)	<0.01
Difficult airway	12 (9–16)	40 (37–42)	<0.01
Time to intubation (seconds),	, median (IQR)		
Normal airway	30 (24–45)	26 (23–34)	0.99
Difficult airway	28 (20–48)	36 (25–50)	0.95
Cormack-Lehane grades, me	edian (IQR)		
Normal airway	1 (1–2)	1 (1–2)	0.50
Difficult airway	1 (1–2)	2 (1–2)	<0.01
AWS, Airwayscope; DL, direct la	ryngoscope; PGY, post graduate years.		

the use of an AWS than that with a DL in the difficult airway scenario (table 3). Similarly, in the inexperienced intubator group, the forces applied on oral structures were lower with the use of an AWS than that with a DL in the difficulty airway scenarios (table 4).

In comparison of the outcomes with the use of a DL according to the airway scenarios, the forces applied on oral structures were significantly lower in the normal airway scenario (the maximum force applied on the maxillary incisors, 77 N vs 183 N, p<0.01; the force applied on the tongue, 27 N vs 40 N, p<0.01, table 5) compared to those in the difficult airway scenario. Likewise, the Cormack-Lehane grade score was significantly lower in the normal airway scenario (median, 1.0 vs 2.0. p<0.01). In contrast, with the use of an AWS, there were no significant differences in any of the outcomes between the two scenarios.

# DISCUSSION

In this prospective cross-over study, we found that the use of AWS, compared to DL, was associated with a

lower maximum force on the oral structures during intubation attempts in the difficult airway scenario. Additionally, we also found that the use of AWS was associated with an improved glottic view in the difficult airway scenario.

Applying excessive forces on the oral structures during intubation attempt is associated with direct and indirect adverse events. 11 Dental injury—one of the common direct adverse events—occurs when the force applied to the maxillary incisors exceeds 150 N.12 Our study demonstrated that the forces applied on the maxillary incisors with the use of AWS are significantly lower than those applied with DL in the difficult airway scenario. Additionally, in the same scenario, the highest quartile of applied forces on the maxillary incisors with the use of AWS did not exceed 150 N. In agreement with these data, a previous simulation study reported that the use of AWS decreases the number of dental clicks compared with the use of DL.5 These findings collectively suggest that the use of AWS might decrease incidences of dental injuries during intubation attempts.

Table 3	Comparison of the outcomes between Airwayscope and the direct laryngoscope for intubation by experienced
intubato	rs (n=18)

	AWS	DL	
Outcomes	n=18	n=18	p Value
Forces on maxillary incisors	(N), median (IQR)		
Normal airway	110 (62–126)	78 (16–94)	0.09
Difficult airway	105 (7–153)	182 (172–187)	<0.01
Applied forces on tongue (N)	), median (IQR)		
Normal airway	10 (8–16)	32 (22–42)	<0.01
Difficult airway	12 (10–15)	39 (36–40)	< 0.01
Time to intubation (seconds)	, median (IQR)		
Normal airway	32 (20–45)	25 (21–30)	0.09
Difficult airway	24 (18–38)	27 (23–39)	0.58
Cormack-Lehane grades, me	edian (IQR)		
Normal airway	2 (1–2)	1 (1–2)	0.40
Difficult airway	1 (1–1)	2 (2–2)	<0.01
AWS, Airwayscope; DL, direct la	ryngoscope; PGY, post graduate years.		

**Table 4** Comparison of the outcomes between Airwayscope and the direct laryngoscope for intubation by inexperienced intubators (n=19)

	AWS	DL	
Outcomes	n=19	n=19	p Value
Forces on maxillary incisors (	(N), median (IQR)		
Normal airway	107 (44–142)	76 (47–99)	0.17
Difficult airway	77 (4–120)	183 (160–186)	<0.01
Applied forces on tongue (N)	, median (IQR)		
Normal airway	12 (8–13)	22 (19–32)	<0.01
Difficult airway	13 (5–17)	41 (39–43)	<0.01
Time to intubation (seconds),	, median (IQR)		
Normal airway	28 (26–47)	38 (35–71)	0.89
Difficult airway	40 (22–61)	28 (25–63)	0.07
Cormack-Lehane grades, me	dian (IQR)		
Normal airway	1 (1–2)	1 (1–1)	0.72
Difficult airway	2 (1–2)	2 (1–3)	0.02
AWS, Airwayscope; DL, direct la	ryngoscope; PGY, post graduate years.		

To avoid indirect adverse events, such as vasovagal responses and haemodynamic alterations, emergency airway management requires intubation attempts with minimal forces to the oropharyngeal structures. <sup>11</sup> <sup>13</sup> <sup>14</sup> A previous report suggested that the use of AWS was associated with a decreased haemodynamic alteration during intubation attempts <sup>13</sup>; however, the authors did not evaluate the actual forces applied on oral structures. In contrast, in the present study, we found that the forces applied on the tongue were significantly lower with the use of AWS compared to that of DL. Our prospective cross-over study corroborated the previous knowledge and extended prior research by demonstrating the mechanism by which the use of AWS reduces intubation-related adverse events.

The reasons for the observed difference in the applied forces between the airway devices in the difficult airway are likely multifactorial. A plausible explanation is that the use of AWS enabled intubations without an excessive effort to achieve a direct line of sight. This is supported by the observed improvement in the glottic view with the

use of an AWS. When a DL is used during intubation attempt in a difficult airway, maxillary incisors are reportedly used as the pivot point to lever the soft tissues upward, thereby leading to an excessive force to the maxillary incisors and tongue. 15 However, the reason for the finding that the maximum force applied on the maxillary incisors in the normal airway scenario was higher with the use of AWS than that with DL was unclear and also likely multifactorial. In our study, most intubators were familiar with the use of a DL, while they were less familiar with the use of AWS. Indeed, approximately 80% of the participants had performed less than 10 intubations using an AWS. Therefore, one may surmise that the unfamiliarity with AWS led to this finding. However, the sensitivity analysis demonstrated no significant difference in the applied force on the maxillary incisor between the devices in the experienced intubator group. Alternatively, the size of an AWS-bigger than that of a DL-might have contributed to the finding.

To the best of our knowledge, this is the first study to report the association between the use of AWS and

Table 5 Comparisons of the outcomes according to the airway scenarios in the use of direct laryngoscope and Airway scope

	Normal airway	Difficult airway	
Device and outcomes	n=37	n=37	p Value
Direct laryngoscope			
Forces on maxillary incisors (N), median (IQR)	77 (44–96)	183 (170–186)	< 0.01
Applied forces on tongue (N), median (IQR)	27 (21–39)	40 (37–42)	< 0.01
Time to intubation, (seconds), median (IQR)	26 (23–34)	36 (25–50)	0.06
Cormack-Lehane grades, median (IQR)	1 (1–2)	2 (1–2)	< 0.01
Airway scope			
Forces on maxillary incisors (N), median (IQR)	107 (54–127)	89 (6–141)	0.07
Applied forces on tongue (N), median (IQR)	11 (8–14)	12 (9–16)	0.63
Time to intubation (seconds), median (IQR)	30 (24–45)	28 (20–48)	0.74
Cormack-Lehane grades, median (IQR)	1 (1–2)	1 (1–2)	0.31

decreased forces on the oral structures during intubation attempts. Multiple studies have reported the advantages of AWS for intubation, not only in the routine general anaesthesia setting <sup>1</sup> <sup>4</sup> <sup>16</sup> but also in several clinical situations—for example, cardiopulmonary resuscitation, pregnant patients with lateral tilt-position and out-of-hospital setting. <sup>17–19</sup> Several simulation studies also reported that novice intubators intubate more successfully and safely with the use of AWS compared with that of DL. <sup>20</sup> <sup>21</sup> Furthermore, our findings support the systematic use of an AWS in the critical settings, such as in patients with hypotension and those who cannot tolerate haemodynamic alterations (eg, intracranial haemorrhage).

This study has potential limitations. Approximately half of the participants had experienced 100 or fewer intubations in their medical career, and the median number of AWS uses was 4. However, even the less experienced intubator was able to intubate with less applied forces. In addition, as our study used simulators, an extrapolation of the findings to the clinical setting requires caution.

### CONCLUSIONS

In this prospective cross-over study, we found that the use of AWS, compared to DL, was associated with lower maximum forces on the oral structures in the difficult airway scenario. We also found that the use of AWS was associated with an improved glottic view in the difficult airway scenario.

### **Author affiliations**

- <sup>1</sup>Department of Emergency Medicine, University of Fukui Hospital, Fukui, Japan
- <sup>2</sup>Department of Emergency and Critical Care Medicine, University of Tsukuba Hospital, Ibaraki, Japan
- <sup>3</sup>Department of Healthcare Epidemiology, Graduate School of Medicine and Public Health, Kyoto University, Kyoto, Japan
- <sup>4</sup>Harvard Interfaculty Initiative in Health Policy, Harvard University, Cambridge, Massachusetts, USA
- <sup>5</sup>Department of Emergency Medicine, Massachusetts General Hospital, Harvard Medical School, Boston, Massachusetts, USA

Contributors TG was involved in study concept and design, analysis and interpretation of the data and drafting of the manuscript. YK took part in study concept and design, analysis and interpretation of the data and statistical expertise. TK was involved in acquisition of the data, analysis and interpretation of the data. YT took part in analysis and interpretation of the data and critical revision of the manuscript for important intellectual content. KH was involved in study concept and design, analysis and interpretation of the data and critical revision of the manuscript for important intellectual content.

**Funding** This research received no specific grant from any funding agency in the public, commercial or not-for-profit sectors.

Competing interests None.

Ethics approval St. Marianna University School of Medicine Hospital.

Provenance and peer review Not commissioned; externally peer reviewed.

Data sharing statement No additional data are available.

Open Access This is an Open Access article distributed in accordance with the Creative Commons Attribution Non Commercial (CC BY-NC 4.0) license, which permits others to distribute, remix, adapt, build upon this work non-commercially, and license their derivative works on different terms, provided the original work is properly cited and the use is non-commercial. See: http://creativecommons.org/licenses/by-nc/4.0/

#### REFERENCES

- Asai T, Liu EH, Matsumoto S, et al. Use of the Pentax-AWS in 293 patients with difficult airways. Anesthesiology 2009;110:898–904.
- Sakles JC, Mosier J, Chiu S, et al. A comparison of the C-MAC video laryngoscope to the Macintosh direct laryngoscope for intubation in the emergency department. Ann Emerg Med 2012;60:739–48
- Mosier JM, Whitmore SP, Bloom JW, et al. Video laryngoscopy improves intubation success and reduces esophageal intubations compared to direct laryngoscopy in the medical intensive care unit. Crit Care 2013:17:R237.
- Hirabayashi Y, Seo N. Tracheal intubation by non-anesthesia residents using the Pentax-AWS airway scope and Macintosh laryngoscope. J Clin Anesth 2009;21:268–71.
- Liu L, Tanigawa K, Kusunoki S, et al. Tracheal intubation of a difficult airway using Airway Scope, Airtraq, and Macintosh laryngoscope: a comparative manikin study of inexperienced personnel. Anesth Analg 2010;110:1049–55.
- Maclure M, Mittleman MA. Should we use a case-crossover design? *Annu Rev Public Health* 2000:21:193–221.
- Suissa S. The case-time-control design: further assumptions and conditions. Epidemiology 1998;9:441–5.
- Yohan N, Wang C, Tokumoto M, et al. Development of Airway Management Training System WKA-4: Provide Useful Feedback of Trainee Performance to Trainee during Airway Management. International Conference on Complex Medical Engineering. 2012.
- Cormack RS, Lehane J. Difficult tracheal intubation in obstetrics. Anaesthesia 1984;39:1105–11.
- Bishop MJ, Harrington RM, Tencer AF. Force applied during tracheal intubation. Anesth Analg 1992;74:411–14.
- Mort TC. Complications of emergency tracheal intubation: hemodynamic alterations—part I. J Intensive Care Med 2007:22:157–65.
- Fontijn-Tekamp FA, Slagter AP, Van Der Bilt A, et al. Biting and chewing in over dentures, full dentures, and natural dentitions. J Dent Res 2000;79:1519–24.
- Koyama Y, Nishihama M, Inagawa G, et al. Comparison of haemodynamic responses to tracheal intubation using the Airway Scope((R)) and Macintosh laryngoscope in normotensive and hypertensive patients. Anaesthesia 2011;66:895–900.
- Kovac AL. Controlling the hemodynamic response to laryngoscopy and endotracheal intubation. J Clin Anesth 1996;8:63–79.
- Lee RA, van Zundert AA, Maassen RL, et al. Forces applied to the maxillary incisors during video-assisted intubation. Anesth Analg 2009;108:187–91.
- Suzuki A, Onodera Y, Mitamura SM, et al. Comparison of the Pentax-AWS airway scope with the Macintosh laryngoscope for nasotracheal intubation: a randomized, prospective study. J Clin Anesth 2012;24:561–5.
- Han SK, Shin DH, Choi PC. Utility of the Pentax-AWS without interruption of chest compression: comparison of the Macintosh laryngoscope with the Pentax-AWS in manikin model. *Resuscitation* 2010;81:69–73.
- Komasawa N, Ueki R, Itani M, et al. Validation of the Pentax-AWS Airwayscope utility as an intubation device during cardiopulmonary resuscitation on the ground. J Anesth 2010;24:582–6.
- Kohama H, Komasawa N, Ueki R, et al. Utility of the Pentax-AWS Airwayscope and Macintosh laryngoscope for airway management during chest compressions in 27 degrees left-lateral tilt: a manikin simulation study of maternal cardiopulmonary resuscitation. J Anesth 2013;27:671–5.
- Komasawa N, Ueki R, Itani M, et al. Evaluation of tracheal intubation in several positions by the Pentax-AWS Airway Scope: a manikin study. J Anesth 2010;24:908–12.
- Malik MA, Hassett P, Carney J, et al. A comparison of the Glidescope, Pentax AWS, and Macintosh laryngoscopes when used by novice personnel: a manikin study. Can J Anaesth 2009;56:802–11.