

# A simple dilute-and-shoot approach for the determination of ultra-trace levels of arsenic in biological fluids *via* ICP-MS using CH<sub>3</sub>F/He as a reaction gas

M.R. Flórez,<sup>a,b</sup> E. García-Ruiz,<sup>a</sup> E. Bolea-Fernández,<sup>b</sup> F. Vanhaecke<sup>b</sup> and M. Resano<sup>a\*</sup>

<sup>a</sup> University of Zaragoza, Department of Analytical Chemistry, Aragón Institute of Engineering Research (I3A), Pedro Cerbuna 12, 50009 Zaragoza, Spain

<sup>b</sup> Ghent University, Department of Analytical Chemistry, Krijgslaan 281-S12, 9000 Ghent, Belgium

## Abstract

The performance of a mixture of CH<sub>3</sub>F/He (1/9) as reaction gas for the determination of As in biological fluids using a quadrupole ICP-MS instrument has been explored. A simple (dilute-and-shoot), interference-free method has been developed to quantify As concentrations at trace and ultra-trace levels in matrices with high Cl content. As<sup>+</sup> reacts with CH<sub>3</sub>F (through CH<sub>3</sub>F addition, followed by HF elimination) with high efficiency forming AsCH<sub>2</sub><sup>+</sup> as the primary reaction product, which can be monitored at a mass-to-charge ratio of 89, free from the Cl-based interferences (e.g., <sup>40</sup>Ar<sup>35</sup>Cl<sup>+</sup> and <sup>40</sup>Ca<sup>35</sup>Cl<sup>+</sup>) that hamper the monitoring of <sup>75</sup>As<sup>+</sup>. Matrix effects are overcome by the use of Te as an internal standard and the addition of 3% v/v ethanol to all samples and calibration standard solutions. The method presented was validated by analysing a set of reference materials (blood, serum and urine) and by assessing As recovery from a set of real blood samples. With this method, the limit of detection was calculated to be 0.8 ng L<sup>-1</sup> As, comparing favourably with the vast majority of the values reported in the

literature, even with those obtained using more sophisticated sector-field instrumentation.

## 1. Introduction

Arsenicosis or As poisoning is most commonly due to long-term exposure to As-rich drinking water.<sup>1,2</sup> The World Health Organization (WHO) sets a guideline value of  $10 \mu\text{g L}^{-1}$  of As in drinking waters, but this value is mainly governed by the traditional limits of detection (LODs).<sup>3</sup> There is still a demand for simple, direct and interference-free methodologies that allow better LODs to be obtained and, therefore, provide more reliable results for lower As concentrations, for instance when analysing biological fluids as a proxy for tracking potentially elevated As exposure and the related risks.

Several spectroscopic techniques have been employed for As determination, although none of them is free of hindrances.<sup>4</sup> Due to its high detection power, inductively coupled plasma-mass spectrometry (ICP-MS) has been widely used to monitor low As contents in a variety of samples. However, accurate determination of As at trace and ultra-trace levels by ICP-MS is strongly jeopardized by the occurrence of spectral overlap of the signals from  $\text{ArCl}^+$  ( $^{40}\text{Ar}^{35}\text{Cl}^+$ ,  $^{38}\text{Ar}^{37}\text{Cl}^+$ ) and  $^{40}\text{Ca}^{35}\text{Cl}^+$  ions with that of the mono-isotopic target element at the mass-to-charge ratio ( $m/z$ ) of 75.<sup>5</sup> This interference represents a great challenge when aiming to determine As in matrices such as blood, urine or serum, due to the high amount of dissolved chloride salts present and the very low As concentrations expected.

Several analytical methodologies have been developed to overcome the aforementioned spectral interferences, but many of them require a separation of the analyte from the matrix prior to the measurement. An obvious choice for a direct and complete resolution of the spectral interferences would be using a sector-field ICP-MS (SF-ICP-MS) instrument, capable of working at higher mass resolution.<sup>6</sup> However, at the resolution required, there is a significant drop in

sensitivity, which is not desirable when aiming to determine very low concentrations of the analyte.<sup>7</sup>

The use of quadrupole-based ICP-MS (Q-ICP-MS) instrumentation equipped with a collision/reaction cell is a good alternative, especially as such instrumentation is more affordable to routine labs than SF-ICP-MS and, thus, it is more widespread.<sup>8-</sup>

<sup>10</sup> Several gases/mixtures of gases have been explored in the context of As determination in Cl-containing matrices, in an attempt to alleviate the spectral interferences. Examples of reactive gases widely used and documented in the literature are H<sub>2</sub>, O<sub>2</sub> or N<sub>2</sub>O. H<sub>2</sub> reacts with ArCl<sup>+</sup> helping in reducing the interference at m/z = 75 to some extent.<sup>11-15</sup> On the other hand, O<sub>2</sub> reacts with As<sup>+</sup>, forming the oxide ion <sup>75</sup>As<sup>16</sup>O<sup>+</sup>, which can be monitored at m/z = 91, free from Cl-based polyatomic interferences.<sup>16,17</sup> However, it has to be kept in mind that Co can form <sup>59</sup>Co<sup>32</sup>O<sub>2</sub><sup>+</sup>, which also shows a m/z = 91.<sup>18</sup> It is likely that this fact led Funk *et al.* to state that “quantification of As in the DRC mode resulted in interfering ions at m/z 91” when aiming at blood analysis.<sup>19</sup> N<sub>2</sub>O reacts with As<sup>+</sup> in the same way as O<sub>2</sub> does, as O-atom donor.<sup>20,21</sup>

NH<sub>3</sub> could be another possibility. NH<sub>3</sub> reacts with <sup>40</sup>Ar<sup>35</sup>Cl<sup>+</sup> successfully removing the interference, but it also reacts with As<sup>+</sup> according to what has been described as an “unusually complex condensation reaction” by Baramov and Tanner,<sup>22</sup> leading to the formation of various species, of which As(NH<sub>3</sub>)NH<sub>2</sub><sup>+</sup> seems to be the most abundant ion. To the best of the authors’ knowledge, the formation of such product has not been used yet for analytical purposes.

He is also typically used as collision gas in combination with kinetic energy discrimination to reduce the contribution from polyatomic interferences, although such approach tends to significantly reduce the sensitivity, providing a signal-to-

noise ratio that is not competitive with those obtained with the reactions described above.<sup>23</sup>

Despite these possibilities, As determination at the low levels required in some complex clinical samples remains a challenge. A clear example of these difficulties can be found in a very recent publication, that reports on As values obtained using various approaches, based on collision/reaction cell ICP-MS and SF-ICP-MS, for a new blood CRM material. The results obtained showed so much scatter (one order of magnitude range) that the As value could not be certified finally, and only an informative  $<5 \mu\text{g L}^{-1}$  value was provided instead.<sup>4</sup> It is worthwhile mentioning that the typical As level in human blood is precisely a few  $\mu\text{g L}^{-1}$ , and thus clinical labs should be able to routinely determine such levels reliably.

High efficiency has been reported for the reaction of  $\text{As}^+$  with  $\text{CH}_3\text{F}$ , with predominant formation of  $\text{AsCH}_2^+$  ( $m/z=89$ ), upon  $\text{CH}_3\text{F}$  addition and subsequent elimination of  $\text{HF}$ .<sup>24</sup> However, this reactive gas has been tested very seldom in reaction cells.<sup>25</sup> Nevertheless, very good results (both in terms of sensitivity and LODs) have been obtained recently for As determination in some biological CRMs (tissues and plants) using a tandem ICP-mass spectrometer (ICP-MS/MS), a new type of ICP-MS instrumentation with a collision/reaction cell situated in-between two quadrupole units.<sup>26</sup> Such instrumentation is more expensive than a traditional quadrupole ICP-MS unit, but enables a better control of the reactions taking place within the reaction cell (an octopole) by removing all ions with a  $m/z$  different from that defined by the first quadrupole, and subsequently selecting the desired  $m/z$  ratio to be monitored using the second quadrupole.

It is the purpose of this work to investigate the potential of  $\text{CH}_3\text{F}$  for As determination in a more simple and standard (and thus, more widespread)

reaction cell ICP-MS instrument. Plasma, urine and, especially, blood have been selected as target samples owing to their intrinsic interest and the problems mentioned above. The goal of the work was to develop a simple, but accurate dilute-and-shoot approach.

## **2. Experimental**

### **2.1. Instrumentation**

All the measurements were carried out using a quadrupole-based ICP-MS instrument (NexION 300X), which is commercially available from Perkin Elmer (Waltham, USA). This instrument is equipped with a cell (a quadrupole) that can be used both as a collision cell in combination with kinetic energy discrimination (KED) and as a dynamic reaction cell (DRC). The instrument is equipped with a triple cone interface, with an additional hyper skimmer cone, providing a more gradual pressure reduction within the interface, which results in less dispersion of the ion beam. A quadrupole ion deflector reflects the ion beam over a 90 degree angle, focusing it into the cell.

The sample introduction system comprises a 0.4 mL min<sup>-1</sup> concentric quartz nebulizer and a quartz cyclonic spray chamber, with 0.38 mm inner diameter PVC flared tubing for the peristaltic pump.

### **2.2. Samples and standards**

#### **2.2.1. Standards and reagents**

Purified water was obtained from a Milli-Q system (Millipore, Billerica, USA). Moreover, high purity water TraceSELECT Ultra was purchased from Fluka Analytical (Sigma-Aldrich, St. Louis, USA).

HNO<sub>3</sub> solutions were prepared from 14 mol L<sup>-1</sup> HNO<sub>3</sub> SupraPur, obtained from Merck Millipore (Darmstadt, Germany), and used for dilutions.

As and Te solutions were prepared from commercially available 1 g L<sup>-1</sup> single-element standards (Merck) by appropriate dilution with 0.14 mol L<sup>-1</sup> HNO<sub>3</sub> Suprapur. Anhydrous, denatured ethanol of spectrophotometric purity grade (90% alcoholic purity) was purchased from Alfa Aesar (Karlsruhe, Germany). KCl solid salt (pro analysis purity grade) was obtained from Merck.

### 2.2.2. Samples

Reference materials with different As contents were analysed in order to validate the method developed. Seronorm trace elements in whole blood level I, II and III, and Seronorm trace elements in serum level I, all of them available as lyophilized materials, were purchased from Sero AS (Billingstad, Norway). ClinChek trace elements in urine control level I and II, also available as lyophilized materials, were purchased from Recipe (Munich, Germany).

In addition, whole blood samples from healthy volunteers were obtained from the University Hospital Miguel Servet (Zaragoza, Spain).

### 2.3. Analytical method for sample analysis

All the samples and reference materials were analysed using the instrumental settings and data acquisition settings listed in **Table 1**.

The samples were not treated prior to analysis, except for dilution. Different dilution factors were evaluated (25-, 50-, 100- and 200-fold), but the final solutions always contained 3% ethanol and 1% HNO<sub>3</sub>.

A new set of 5 calibration standard solutions of suitable concentrations, adapted to the expected As sample levels and also containing 3% ethanol and 1% HNO<sub>3</sub>, was prepared for every measurement session.

SupraPur grade HNO<sub>3</sub> and TraceSelect purity grade water were used for dilutions, in order to assure low blank signals.

Te was added as internal standard to all sample and standard solutions for a final concentration of  $20 \mu\text{g L}^{-1}$  Te.

### 3. Results and discussion

#### 3.1. Study of the reaction between As and $\text{CH}_3\text{F}/\text{He}$ within the reaction cell

In order to avoid the major interferences found at  $m/z=75$ , the possibility to promote a reaction between As and  $\text{CH}_3\text{F}$  was investigated.

Zhao *et al.*<sup>24</sup> listed  $\text{AsCH}_2^+$  (in a 97% distribution) as the primary product of the reaction between As and  $\text{CH}_3\text{F}/\text{He}$ , with a minor (3%) formation of the addition product  $\text{As}(\text{CH}_3\text{F})^+$ , using ICP/selected-ion flow tube-MS (ICP/SIFT-MS). Bolea-Fernandez *et al.*<sup>26</sup> confirmed this behaviour using ICP-MS/MS. This type of reaction with  $\text{CH}_3\text{F}$  ( $\text{CH}_3\text{F}$  addition and ulterior elimination of HF) is not very common. Fluorination or  $\text{CH}_3\text{F}$  addition are much more usual. In fact, As is the only element listed in ref. 18 to undergo this process as the main reaction with  $\text{CH}_3\text{F}$ . Recently, it has been shown that the same is true for Se, but in a much less effective way.<sup>26</sup> In any case, this reaction between As and  $\text{CH}_3\text{F}$  is quite characteristic and has been demonstrated to be very efficient in ICP-MS/MS.<sup>26</sup>

In this work,  $\text{CH}_3\text{F}/\text{He}$  is studied as reaction gas for the determination of As, making use of a more conventional reaction cell-ICP-MS instrumentation. To evaluate the efficiency of the reaction and the products generated under the reaction cell conditions, a  $5 \mu\text{g L}^{-1}$  As aqueous standard solution was nebulized and measured at different  $m/z$  ratios. The reaction cell was pressurized with different  $\text{CH}_3\text{F}/\text{He}$  flow rates to find the best settings in terms of sensitivity. The results, represented in **Figure 1**, show an evident preference of the reaction to evolve towards the formation of  $\text{AsCH}_2^+$ , also in the present study.



As shown in **Figure 1**, the maximum  $\text{AsCH}_2^+$  signal that can be attained is lower than the initial  $^{75}\text{As}^+$  signal. The  $\text{AsCH}_2^+$  signal reaches approx. 40% of the initial  $^{75}\text{As}^+$  signal at a flow of  $1.6 \text{ mL min}^{-1} \text{ CH}_3\text{F/He}$ , and even decreases at higher flows. This fact is not due to poor reaction efficiency, as the remaining  $^{75}\text{As}^+$  signal and the signals corresponding to other As species are rather low at such  $\text{CH}_3\text{F/He}$  flows. Instead, this is most likely due to the presence of He, which, at these high flows, is expected to result in scattering losses. Nevertheless, the addition of He is highly recommended for this type of reaction in order to slow down the ions sufficiently to obtain a good reaction efficiency.<sup>24</sup> In any case, this loss of raw sensitivity (roughly a factor of 3) seems like a reasonable price to pay for an interference-free detection of As, as will be demonstrated later on.

### 3.2. Optimization of the reaction cell parameters

A further optimization of the reaction gas flow rate was carried out by monitoring the signal intensity of  $\text{AsCH}_2^+$  in a 50-fold diluted blood sample at different  $\text{CH}_3\text{F/He}$  flow rates. A blank solution containing  $4 \text{ g L}^{-1} \text{ Cl}$  (the concentration typically found in blood), added as KCl, was also measured in the same way, in order to find the conditions providing the best S/N ratio.

**Figure 2** shows how the signal intensity at  $m/z = 89$  increases (left y-axis) with the reaction gas flow rate, reaching a maximum sensitivity at  $1.6 \text{ mL min}^{-1} \text{ CH}_3\text{F/He}$ . The use of higher flow rates leads to signal losses, probably due to scattering. On the other hand, the blank signal (right y-axis) remains rather low (below 40 cps) and stable along all the conditions evaluated. Therefore, an optimum signal-to-noise ratio is obtained at  $1.6 \text{ mL min}^{-1}$ .

The rejection parameter  $q$  of the DRC (RP $q$ ), associated with the low-mass cut-off, was also adjusted. An RP $q$  value of 0.55-0.65 was found to bring the best results in

terms of LOD. A summary of the optimized instrument settings is presented in **Table 1**.

In order to further assure that no spectral interference is affecting the measurements under these optimized conditions, two sets of 0.5, 1, 2.5 and 5  $\mu\text{g L}^{-1}$  As aqueous standard solutions were prepared, a first one diluted with 1%  $\text{HNO}_3$  and a second one diluted with 1%  $\text{HCl}$ . A calibration curve was plotted based on the results obtained for each set of standard solutions so to compare the slopes. These results are presented in **Figure 3**, where the lack of any significant difference can be appreciated. From this direct comparison, it can be concluded that the presence of a high Cl content in the matrix does not influence the results when monitoring the  $\text{AsCH}_2^+$  species at  $m/z = 89$ , thus enabling the determination of As free from spectral interference.

### **3.3. Overcoming non-spectral interferences**

As has a high ionization potential (9.79 eV),<sup>27</sup> and displays a very pronounced matrix-induced signal enhancement in some matrices, especially in the presence of organic C (C-effect).<sup>5,28-35</sup> Due to this fact, some strategies to correct for matrix effects need to be explored.

On the one hand, a suitable internal standard is required, in order to correct for any instrumental instability or signal drift and improve the precision of the measurements. An adequate internal standard has to be absent from the sample matrix, should not lead to or suffer from spectral overlap and must have a mass number close to that of the analyte and a similar ionization potential.<sup>36</sup> Additionally, in this case, particular attention has to be paid not to introduce an element that will react with  $\text{CH}_3\text{F}/\text{He}$ , thus creating a new spectral interference. Ge, for instance, would react with  $\text{CH}_3\text{F}/\text{He}$  via F atom transfer,<sup>24</sup> leading to the

formation of  $^{70}\text{Ge}^{19}\text{F}^+$  at  $m/z = 89$ , making the selection of this element, often used for this task when As determination is aimed at, prohibited for our method. Overall, Te (IP= 9.01 eV)<sup>27</sup> seems to be the best choice for this work and was used as internal standard in all following experiments.

On the other hand, differences in the matrix composition may preclude a direct quantification of As in the samples based on external calibration *versus* aqueous standard solutions. When aiming to determine As in very low concentrations, sample treatment needs to be minimized in order to avoid analyte losses or significant external contamination. Moreover, depending on the concentration, it might not be possible to dilute the sample until no matrix effect is observed. The high organic load of blood and, to a lower extent, serum samples, significantly affects the plasma conditions. Even when applying internal standard correction, such effect may not be completely compensated for.

Thus, the approach finally developed was based on using a sufficiently high sample dilution factor (at least 1:24) to minimize matrix effects and, also, on adding an amount of ethanol to both the samples and the standard solutions, such that basically its presence would control the plasma conditions, helping in matching the potentially different behaviour expected between standard solutions and samples. Moreover, it is well-known that the addition of an appropriate amount of an organic compound, such as ethanol, will lead to a more complete ionization of As though the C-effect, thus increasing the sensitivity for this element.<sup>5,28-35</sup>

An optimization of the amount of ethanol added was carried out, as shown in **Figure 4**. As can be seen (**Figure 4A**), the presence of ethanol initially increases the sensitivity, until a value of 3-5%. However, the LOD does not follow the same trend. After initially dropping, the LOD tends to increase with higher values,

because the blanks are rising even more than the sensitivity. A value between 1 and 3% ethanol seems optimum from this point of view. Even more important, **Figure 4B** shows that overestimated results are obtained if blood is analysed in the absence of ethanol, because the organic matrix tends to increase the As signal. This phenomenon is corrected for when ethanol is added, and good agreement with the reference value is attained at 3% ethanol. The same trend was observed for all the blood samples tested: results biased 15-40% high (depending on the dilution) in the absence of ethanol, and values in agreement with the expected ones when samples and standards contained 3% ethanol. Thus, this value was used in further work.

Under these conditions, the LOD was calculated to be  $0.8 \text{ ng L}^{-1}$ . A comparison with the values found in the literature for different interference-free As determination strategies is presented in **Table 2**. As can be seen, the LOD presented in this work compares favourably with the vast majority of them, even with those obtained using SF-ICP-MS. It seems that only when using  $\text{CH}_3\text{F}/\text{He}$  as reaction gas and the ICP-MS/MS mode a better LOD can be achieved. If  $\text{CH}_3\text{F}/\text{He}$  is used in a triple quadrupole instrument but the mass window of the first quadrupole is maintained open (no mass filtering), a very similar LOD as the one reported in the current work is obtained.<sup>26</sup>

It has to be mentioned that the LODs presented in the table are instrumental LODs. Therefore, any dilution required as a consequence of sample pretreatment needs to be factored in as well. Thus, the method proposed in this work ultimately provides a LOD of  $20 \text{ ng L}^{-1}$  and a LOQ of  $70 \text{ ng L}^{-1}$  (for a dilution factor of 25). These values are sufficiently low to determine As in real blood samples, as will be demonstrated in the next section.

### 3.4. Method validation and sample results

In order to validate the method developed, all the reference samples listed in section 2.2.2. were analysed. All of them were diluted in a 3% ethanol solution to four different dilution levels: 25-, 50-, 100- and 200-fold.

As determination was carried out according to the optimized instrumental settings listed in **Table 1**. The results obtained are presented in **Table 3**. As can be seen, for every sample, the results obtained were found to be in good agreement with the corresponding reference value. The confidence intervals obtained for the different dilutions factors tested always overlapped among them and, also, with the certified range. However, it was also observed that the values obtained with higher dilution factors tended to be a bit lower (see 1:199 values), even if this difference could not be considered as statistically significant. In any case, using such an extreme dilution factor is not really necessary, and it seems preferable to use intermediate dilution factors (1:49 or 1:99) to guarantee the best accuracy, while minimizing residues in the spray chamber.

It is worth mentioning here that, although all the samples and standards were diluted and measured under the same conditions, urine samples do not really require the addition of ethanol, as no significant matrix effect was observed for such samples. It was experimentally confirmed that very similar results were attained for urine when no ethanol was added.

Finally, five real blood samples were also analysed. Since no reference value was available for As, recovery assays were carried out by doping the samples (diluted 1:99) with a  $0.1 \mu\text{g L}^{-1}$  As standard solution. Samples were measured with and without As spike and the results thus obtained are presented in **Table 4**. In all cases, recoveries were higher than 90%, with an average recovery of  $96.8 \pm 3.9\%$ ,

further proving the validity of the approach proposed for analysis of real blood samples.

#### **4. Conclusions**

This work presents a simple and straightforward analytical method for the interference-free determination of As at ultra-trace levels in biological fluids with high Cl content and a organic matrix.

A 1/9 gas mixture of CH<sub>3</sub>F/He was evaluated as reaction gas in a dynamic reaction cell of a quadrupole ICP-MS instrument. As reacts very efficiently and with a high selectivity with CH<sub>3</sub>F/He, leading to the formation of AsCH<sub>2</sub><sup>+</sup> as the primary product.

Despite the strong matrix effects due to the high organic load of the samples, it was possible to develop a simple “dilute and shoot” approach by diluting the samples using 3% of ethanol.

This simple method provides very competitive limits of detection (0.8 ng L<sup>-1</sup>), as well as high accuracy and precision (typically better than 7% RSD).

#### **Acknowledgements**

This work has been funded by the Spanish Ministry of Economy and Competitiveness (Project CTQ2012-33494) and the Aragón Government (Fondo Social Europeo). The authors acknowledge Dr. Luis Rello from the Miguel Servet University Hospital (Zaragoza, Spain) for providing the real blood samples.

## References

1. A. Gomez-Caminero, P. Howe, M. Hughes, E. Kenyon, D. R. Lewis, M. Moore, J. Ng, A. Aitio and G. Becking, *Arsenic and Arsenic compounds*, World Health Organization, 2 edn., Geneva, 2001.
2. J. C. Ng, J. Wang and A. Shraim, *Chemosphere*, 2003, **52**, 1353–1359.
3. WHO, *Guidelines for drinking-water quality*, World Health Organization, 4 edn., Geneva, 2011.
4. R. L. Paul, W. C. Davis, L. Yu, K. E. Murphy, W. F. Guthrie, D. D. Leber, C. E. Bryan, T. W. Vetter, G. Shakirova, G. Mitchell, D. J. Kyle, J. M. Jarrett, K. L. Caldwell, R. L. Jones, S. Eckdahl, M. Wermers, M. Maras, C. D. Palmer, M. F. Verostek, C. M. Geraghty, A. J. Steuerwald and P. J. Parsons, *J. Radioanal. Nucl. Chem.*, 2014, **299**, 1555–1563.
5. J. Goossens, F. Vanhaecke, L. Moens and R. Dams, *Anal. Chim. Acta*, 1993, **280**, 137–143.
6. F. Vanhaecke and L. Moens, *Anal. Bioanal. Chem.*, 2004, **378**, 232–240.
7. C. Turetta, G. Cozzi, C. Barbante, G. Capodaglio and P. Cescon, *Anal. Bioanal. Chem.*, 2004, **380**, 258–268.
8. S. D. Tanner, V. I. Baranov and D. R. Bandura, *Spectrochim. Acta B*, 2002, **57**, 1361–1452.
9. D. R. Bandura, V. I. Baranov, A. E. Litherland and S. D. Tanner, *Int. J. Mass Spectrom.*, 2006, **255-256**, 312–327.
10. S. D'Illo, N. Violante, C. Majorani and F. Petrucci, *Anal. Chim. Acta*, 2011, **698**, 6–13.
11. K. Neubauer and U. Völlkopf, *Atom. Spectrosc.*, 1999, **20**, 64–68.
12. M. Resano, E. García Ruiz, V. G. Mihucz, Á. M. Móricz, Gy. Záráy and F.

- Vanhaecke, *J. Anal. At. Spectrom.*, 2007, **22**, 1158–1162.
13. J. Darrouzès, M. Bueno, G. Lespès, M. Holeman and M. Potin-Gautier, *Talanta*, 2007, **71**, 2080–2084.
  14. C. D. Pereira, E. E. Garcia, F. V. Silva, A. R. A. Nogueira and J. A. Nóbrega, *J. Anal. At. Spectrom.*, 2010, **25**, 1763–1768.
  15. M. Colon, M. Hidalgo and M. Iglesias, *Talanta*, 2011, **85**, 1941–1947.
  16. K. Kawabata, Y. Kishi and R. Thomas, *Anal. Chem.*, 2003, **75**, 422 A–428 A.
  17. S. D'Ilio, N. Violante, M. Di Gregorio, O. Senofonte and F. Petrucci, *Anal. Chim. Acta*, 2006, **579**, 202–208.
  18. [http://www.chem.yorku.ca/profs/bohme/research/selection\\_table.html](http://www.chem.yorku.ca/profs/bohme/research/selection_table.html), last accessed, July 2015.
  19. W. E. Funk, J. K. McGee, A. F. Olshan and A. J. Ghio, *Biomarkers*, 2013, **18**, 174–177.
  20. V. Blagojevic, E. Flaim, M. J. Y. Jarvis, G. K. Koyanagi and D. K. Bohme, *J. Phys. Chem. A*, 2005, **109**, 11224–11235.
  21. M. Grotti and R. Frache, *J. Anal. At. Spectrom.*, 2007, **22**, 1481–1487.
  22. V. I. Baranov and S. D. Tanner, *J. Anal. At. Spectrom.*, 1999, **14**, 1133–1142.
  23. M. Niemelä, P. Perämäki, H. Kola and J. Piispanen, *Anal. Chim. Acta*, 2003, **493**, 3–12.
  24. X. Zhao, G. K. Koyanagi and D. K. Bohme, *J. Phys. Chem. A*, 2006, **110**, 10607–10618.
  25. E. Bolea-Fernández, L. Balcaen, M. Resano and F. Vanhaecke, *Anal. Chem.*, 2014, **86**, 7969–7977.
  26. E. Bolea-Fernández, L. Balcaen, M. Resano and F. Vanhaecke, *Anal. Bioanal. Chem.*, 2015, **407**, 919–929.



27. D. R. Lide, Ed., *CRC Handbook of Chemistry and Physics*, CRC Press (Taylor & Francis group), Boca Ratón, 89 edn. 2008.
28. P. Allain, L. Jaunault, Y. Mauras, J. M. Mermet and T. Delaporte, *Anal. Chem.*, 1991, **63**, 1497–1498.
29. E. H. Larsen and S. Stürup, *J. Anal. At. Spectrom.*, 1994, **9**, 1099–1105.
30. F. Vanhaecke, J. Riondato, L. Moens and R. Dams, *Fresenius J. Anal. Chem.*, 1996, **355**, 397–400.
31. V. L. Dressler, D. Pozebon and A. J. Curtius, *Anal. Chim. Acta*, 1999, **379**, 175–183.
32. Z. Hu, S. Hu, S. Gao, Y. Liu and S. Lin, *Spectrochim. Acta B*, 2004, **59**, 1463–1470.
33. Z. Hu, S. Gao, S. Hu, H. Yuan, X. Liu and Y. Liu, *J. Anal. At. Spectrom.*, 2005, **20**, 1263–1269.
34. W. Guo, S. Hu, X. Li, J. Zhao, S. Jin, W. Liu and H. Zhang, *Talanta*, 2011, **84**, 887–894.
35. G. Grindlay, J. Mora, M. de Loos-Vollebregt and F. Vanhaecke, *Spectrochim. Acta B*, 2013, **86**, 42–49.
36. F. Vanhaecke, H. Vanhoe, R. Dams and C. Vandecasteele, *Talanta*, 1992, **39**, 737–742.
37. J. Frank, M. Krachler and W. Shotyk, *Anal. Chim. Acta*, 2005, **530**, 307–316.
38. B. P. Jackson, A. Liba and J. Nelson, *J. Anal. At. Spectrom.*, 2015, **30**, 1179–1183.
39. C. D. B. Amaral, R. S. Amais, L. L. Fialho, D. Schiavo, T. Amorim, A. R. A. Nogueira, F. R. P. Rocha and J. A. Nóbrega, *Anal. Methods*, 2015, **7**, 1215–1220.

**Table 1.** Instrumental settings and data acquisition parameters for the NexION 300X Q-ICP-MS instrument

NexION 300X	As determination
Mode	DRC
Reaction cell gas	CH <sub>3</sub> F/He (1/9)
Reaction cell gas flow rate	1.6 mL min <sup>-1</sup>
Nebulizer gas flow rate	1.02 L min <sup>-1</sup>
Auxiliary gas flow rate	1.20 L min <sup>-1</sup>
Plasma gas flow rate	18.00 L min <sup>-1</sup>
RF power	1600 W
RPa	0.00
RPq	0.65
Nuclides monitored	<sup>89</sup> AsCH <sub>2</sub> <sup>+</sup> , <sup>126</sup> Te <sup>+</sup>
Sweeps/Reading	100
Readings/Replicate	1
Replicates	10
Dwell time	20 ms

**Table 2.** Comparison of the LODs reported in the literature for As determination by means of ICP–MS.<sup>1</sup>

Species monitored	System* (gas)	LODs/ng L <sup>-1</sup>	Sample	Publication
As <sup>+</sup>	CC (H <sub>2</sub> /He)	147	Biological samples	Ref. 23
As <sup>+</sup>	RC (H <sub>2</sub> )	25	Rain water	Ref. 13
As <sup>+</sup>	CRI (H <sub>2</sub> )	19-95	Biological samples	Ref. 14
AsO <sup>+</sup>	DRC (O <sub>2</sub> , N <sub>2</sub> O)	40	Sea water	Ref. 21
AsO <sup>+</sup>	DRC (O <sub>2</sub> )	2.0	Digested blood	Ref. 17
As <sup>+</sup>	SF–ICP–MS	10	Solutions	Ref. 26
As <sup>+</sup>	SF–ICP–MS	3.0	Digested plants	Ref. 37
AsO <sup>+</sup>	MS/MS (O <sub>2</sub> )	1.0	Digested plants	Ref. 38
AsO <sup>+</sup>	MS/MS (O <sub>2</sub> )	1.6	Drinking water	Ref. 39
AsO <sup>+</sup>	MS/MS (O <sub>2</sub> )	7.0	Solutions	Ref. 26
AsCH <sub>2</sub> <sup>+</sup>	MS/MS (CH <sub>3</sub> F)	0.2	Biological samples	Ref. 26
AsCH <sub>2</sub> <sup>+</sup>	MS/MS (CH <sub>3</sub> F) SQ mode	1.0	Biological samples	Ref. 26
AsCH <sub>2</sub> <sup>+</sup>	DRC (CH <sub>3</sub> F)	0.8	Blood, urine, serum	This work

<sup>1</sup>LODs were calculated as 3 times the standard deviation of the blank signal intensities divided by the slope of the calibration curve.

\*CC= collision cell; RC: reaction cell; CRI: collision reaction interface; DRC=dynamic reaction cell; SF: sector field; SQ: single quadrupole.

**Table 3.** Results obtained for As determination of the biological certified reference materials mentioned in section 2.2.2. Uncertainty is expressed as 95% confidence intervals (n=4).

Reference material	Reference value	Dil. factor 1:24	Dil. factor 1:49	Dil. factor 1:99	Dil. factor 1:199
Whole blood, level I / $\mu\text{g L}^{-1}$	$2.4 \pm 0.5$	$3.1 \pm 0.2$	$2.9 \pm 0.2$	$2.9 \pm 0.2$	$2.9 \pm 0.2$
Whole blood, level II / $\mu\text{g L}^{-1}$	$14.3 \pm 2.9$	$15.3 \pm 1.1$	$14.6 \pm 1.2$	$13.4 \pm 1.3$	$12.6 \pm 1.7$
Whole blood, level III / $\mu\text{g L}^{-1}$	$30.4 \pm 7.3$	$30.0 \pm 3.1$	$27.1 \pm 3.1$	$26.5 \pm 3.7$	$26.1 \pm 3.4$
Serum, level I / $\mu\text{g L}^{-1}$	0.40*	$0.45 \pm 0.07$	$0.52 \pm 0.10$	$0.42 \pm 0.08$	---
Urine, level I / $\mu\text{g L}^{-1}$	$43.0 \pm 8.6$	$42.9 \pm 3.4$	$43.4 \pm 3.9$	$44.1 \pm 3.5$	$38.8 \pm 3.6$
Urine, level II / $\mu\text{g L}^{-1}$	$83.3 \pm 16.6$	$83.2 \pm 4.9$	$81.2 \pm 5.8$	$78.5 \pm 5.1$	$75.9 \pm 5.4$

\* Indicative value

**Table 4.** Results obtained for As determination in real blood samples. Uncertainty is expressed as 95% confidence intervals (n=4).

Sample code	As content / $\mu\text{g L}^{-1}$	Recovery assay / %
MSHBlood01	$5.10 \pm 0.48$	94.5
MSHBlood02	$5.18 \pm 0.46$	91.9
MSHBlood03	$3.13 \pm 0.25$	98.5
MSHBlood04	$2.84 \pm 0.24$	102.1
MSHBlood05	$14.6 \pm 1.3$	97.0

**Figure captions.**

**Figure 1.** Overview of the As-based reaction products generated within the dynamic reaction cell at different CH<sub>3</sub>F/He flow rates

**Figure 2.** A) Optimization of the CH<sub>3</sub>F/He reaction gas flow rate for the generation of AsCH<sub>2</sub><sup>+</sup> (monitored at m/z=89) from a 1:49 diluted blood sample (left y-axis) and from an equivalent Cl-containing blank (right y-axis).

**Figure 3.** Direct comparison of a calibration curve prepared by dilution with 1% HNO<sub>3</sub> and a calibration curve prepared by dilution with 1% HCl, both sets of standard solutions measured under the conditions listed in **Table 1**.

**Figure 4.** Optimization of the concentration of ethanol in samples and standards. A) Variation of the sensitivity and of the limit of detection with every ethanol concentration tested. B) Results obtained for the determination of As in a 1:49 diluted reference blood sample (the certified As concentration,  $30.4 \pm 7.3 \mu\text{g L}^{-1}$ , is represented by a red line), for all different ethanol concentrations studied. Uncertainty is indicated as the standard deviation of 5 replicates.

Figure 1.




Figure 2.






Figure 3.




Figure 4.

