



UNIVERSITY OF LEEDS

This is a repository copy of *Flow cell design for effective biosensing*.

White Rose Research Online URL for this paper:
<http://eprints.whiterose.ac.uk/80645/>

Article:

Pike, DJ, Kapur, N, Millner, PA et al. (1 more author) (2012) Flow cell design for effective biosensing. *Sensors*, 13 (1). 58 - 70. ISSN 1424-8220

<https://doi.org/10.3390/s130100058>

Reuse

Unless indicated otherwise, fulltext items are protected by copyright with all rights reserved. The copyright exception in section 29 of the Copyright, Designs and Patents Act 1988 allows the making of a single copy solely for the purpose of non-commercial research or private study within the limits of fair dealing. The publisher or other rights-holder may allow further reproduction and re-use of this version - refer to the White Rose Research Online record for this item. Where records identify the publisher as the copyright holder, users can verify any specific terms of use on the publisher's website.

Takedown

If you consider content in White Rose Research Online to be in breach of UK law, please notify us by emailing eprints@whiterose.ac.uk including the URL of the record and the reason for the withdrawal request.



eprints@whiterose.ac.uk
<https://eprints.whiterose.ac.uk/>

Supplementary Information

Flow Cell Design for Effective Biosensing

Douglas J. Pike ^{1,*}, Nikil Kapur ², Paul A. Millner ³ and Douglas I. Stewart ¹

¹ Pathogen Control Engineering (PaCE) Institute, School of Civil Engineering, University of Leeds, Leeds, West Yorkshire, LS2 9JT, UK; E-Mail: d.i.stewart@leeds.ac.uk

² Institute of Engineering Thermofluids, Surfaces & Interfaces (iETSI), School of Mechanical Engineering, University of Leeds, Leeds, West Yorkshire, LS2 9JT, UK; E-Mail: n.kapur@leeds.ac.uk

³ Institute of Membrane and Systems Biology, Faculty of Biological Sciences, University of Leeds, Leeds, West Yorkshire, LS2 9JT, UK; E-Mail: p.a.millner@leeds.ac.uk

* Author to whom correspondence should be addressed; E-Mail: ee09djp@leeds.ac.uk; Tel.: +44-11-3343-2399.

Modelled Flow Animations

Four animation videos are presented (alongside this document) that show flow simulations for the circular flow cell design at flow rates of 0.1, 1.75, 5.0 & 7.5 mL/min (Videos SV1–SV4), taken as a plan-view through the mid-plane of the flow cell. Flow cell dimensions on each video are shown in mm. These videos directly supplement Figure 5(A) of the main paper which presented snapshot images after five flow cell volume changes for the same flow conditions. The response of the circular flow cell at the four selected flow rates is presented in Figure 1(A).

Video Details


Video SV1 Flow simulation animation for circular flow cell design at a flow rate of 0.1 mL/min. Animation is based upon numerical solutions at 1s intervals and an output rate of 15 frames per second. Video length is 33.3 s.

Video SV2 Flow simulation animation for circular flow cell design at a flow rate of 1.75 mL/min. Animation is based upon numerical solutions at 1s intervals and an output rate of 15 frames per second. Video length is 13.3 s.

Video SV3 Flow simulation animation for circular flow cell design at a flow rate of 5 mL/min. Animation is based upon numerical solutions at 1s intervals and an output rate of 15 frames per second. Video length is 6.7 s.

Video SV4 Flow simulation animation for circular flow cell design at a flow rate of 7.5 mL/min. Animation is based upon numerical solutions at 1s intervals and an output rate of 15 frames per second. Video length is 6.7 s. Flow cell dimensions are shown in mm.

Figure S1. Computational results showing response of flow cell to injection of analyte (scaled between 0—no analyte and 1—analyte concentration within cell equal to that of injected fluid) for circular cell at four selected flow rates.



© 2011 by the authors; licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution license (<http://creativecommons.org/licenses/by/3.0/>).