

Cite this: DOI: 10.1039/c3lc51277k

www.rsc.org/xxxxxx

PAPER

Dynamically reconfigurable Fibre Optical Spanner

Thorsten Kolb, Sahradha Albert, Michael Haug and Graeme Whyte*

Received (in XXX, XXX) Xth XXXXXXXXXX 20XX, Accepted Xth XXXXXXXXXX 20XX

DOI: 10.1039/c3lc51277k

In this paper we describe a pneumatically actuated fibre-optic spanner integrated into a microfluidic Lab-on-a-Chip device for the controlled trapping and rotation of living cells. The dynamic nature of the system allows interactive control over the rotation speed with the same optical power. The use of a multi-layer device makes it possible to rotate a cell both in the imaging plane and also in a perpendicular plane allowing tomographic imaging of the trapped living cell. The integrated device allows easy operation and by combining it with high-resolution confocal microscopy we show for the first time that the pattern of rotation can give information regarding the sub-cellular composition of a rotated cell.

Optical trapping is a powerful tool for manipulating microscopic objects¹ including living cells. In addition to holding the trapped object in place, it is also possible to rotate it using the intrinsic spin^{2,3}, orbital angular momentum^{4,5} or anisotropic beam shape⁶ however all of these systems require complex optical equipment.

Dual-beam optical traps⁷ are created when two, diverging counter propagating beams overlap and form a stable trap in the centre of the two beams. The large area and diverging nature provides a less photo-damaging technique for holding and manipulating living cells^{8,9,6}. Recent developments have increased the range of possible manipulations which can be carried out with dual beam traps to include mechanical phenotyping of living cells^{9,10}, optical rotation⁶ and optofluidic rotation. The development of dual-beam fibre traps¹¹ simplifies the creation and integration of such traps into microfluidic systems¹²⁻¹⁴ allowing for high throughput measurements and more complex microfluidic processing.

The fibre optic spanner¹⁵ is a form of dual-beam trap where the optical fibres are intentionally misaligned to provide a torque on the trapped object from the asymmetry of the trapping light beams, and it can be used to rotate cells around an axis perpendicular to the misalignment. In theory this allows cells to be rotated around any angle by correctly arranging the fibre positions, however accurate control of the fibre position and angle limits the usability of such a system.

Here we present a dynamically reconfigurable system for positioning the optical fibres within a dual-beam trap. The positioning is achieved using the elastic nature of PDMS devices and by pneumatically applying air pressure to specifically designed membranes^{16,17}. Multilayer PDMS microfabrication allows such membranes to be placed both in the plane of the fibre as well as above it, to allow adjustment of the fibre in orthogonal directions, opening the door to rotate cells around an axis perpendicular to the imaging plane of the microscope for single cell tomography.

Experimental

The device is made up of a PDMS chip housing two optical fibres aligned to each other across a microfluidic flow channel, and pneumatic actuators next to the fibres to dynamically adjust the positions of the fibres. The actuators move the fibres by applying




Figure 1 a) Schematic representation of the device, showing the aligned dual-beam trap in the absence of external pressure and the repositioning of the optical fibres when additional pressure is applied to the fibre pads, causing controlled rotation of the trapped cell. b) Image of a fabricated device next to a 1 Euro coin, showing the flow channel (blue), fibre channels (yellow) and fibre pad features (green). c) Micrograph of the trap region with the fibres inserted d) Measure of the offset of one arm of the trap (measured at the middle of the flow channel) after repeated cycling of the fibre pads. e) Detail of the time response of the device to a pressurisation of the fibre pad (blue) and the time-response fit (red) to the data giving a time constant of 10 μ s. f-i) Micrographs of the movement of an optical fibre by applying an external pressure. Scale bars are 200 μ m.

a pressure to one side of a thin PDMS membrane which in turn pushes the fibre away. By moving the fibres from their aligned position, it is possible to create a reconfigurable fibre optic spanner capable of controlling the rotation speed of single cells held in the optical trap.

The additional freedom of controlling the fibre misalignment allows more detailed characterization of the rotating cell.




Figure 2a) Sequential micrographs of a trapped U2OS GFP-Lamin A expressing cell rotating in the reconfigurable spanner trap. GFP shown in green and the DNA is visualised using DRAQ5 stain shown in red. Each image is taken after 1/8 of the rotation period. Scale bar is 10 μm . b) Rotation rate as a function of total optical power for a single cell showing the highly linear relationship. c) Rotation rate as a function of applied pressure showing the optimum fibre arrangement. d) Decreasing the ratio of PDMS curing agent to elastomer of a device results in a larger displacement of the optical fibre within the device for the same pressure.

Device Fabrication

Devices were fabricated using standard PDMS moulding techniques from a photolithography defined SU8 master. Briefly SU8-2025 (MicrochemInc) was spin coated onto 3" Si wafers (Silicon Materials) to form 125 μm thick layer to allow the easy insertion of the optical fibres¹⁸, and then features were created using UV photolithography from an acetate mask. Single-layer devices were produced from 10:1 Elastomer to Curing agent (Syl-Guard185), which was mixed and poured onto the SU8 master. After baking for at least 3hrs at 80 $^{\circ}\text{C}$, the device was peeled from the SU8 and plasma bonded to a microscope coverslip using air plasma generated by a plasma oxidiser (Diener) and further baked for 1 hr at 80 $^{\circ}\text{C}$.

Multi-layer devices were produced in a similar manner to Hulme¹⁹. Briefly, the top layer containing the z-adjustment pads were produced as above, however instead of bonding to a microscope coverslip, they were plasma bonded to a thin layer of PDMS which was spin coated onto a silanised (Sigma-Aldrich) glass slide. Once the top layer had sealed to the thin layer of PDMS, both layers were peeled away from the supporting glass slide. The bottom layer was prepared by pouring 10:1 mixed PDMS: curing agent onto the SU8 master and aligning the top layer to the fibre channels under a microscope. The multi-layer device was then baked for at least 3hrs at 80 $^{\circ}\text{C}$ to cure the liquid PDMS and seal it to the top layer. Finally the multi-layer structure was cut out of the master and sealed to a 150 μm thick layer of PDMS using air plasma treatment, to form the bottom wall of the fluidic channel.

Elastomer was used to fill the fibre channels to provide refractive-index matching between the device walls and the optical fibre^{20,21}. In order to minimise the possibility of trapping air bubbles in front of the fibres, which would cause the light to be scattered, the elastomer was filled into the channels under vacuum and the optical fibres were dipped in elastomer before insertion to ensure the surface was fully wetted. Optical fibres were then manually inserted into the fibre channels and externally secured in place with the fibre aligned with the end of the fibre

channel, leaving a small volume of liquid PDMS to allow the fibre to easily move. The system was then visually checked to ensure that there was no debris or air between the end of the fibre and the channel wall, which could scatter the light and disrupt the trapping.

Cell Culture

HEK293 and U2OS cells (ATCC: HBT-96 and ATCC: CRL-1573 respectively) were cultivated at 37 $^{\circ}\text{C}$, 5% CO_2 in DMEM medium containing 10% fetal calf serum and 4 mM L-glutamine. Cells were regularly passaged every three days and kept in the exponential growth phase. Nuclear staining was performed utilizing DRAQ5 (abcam) according to the manufacturer's manual.

Device Operation

Microfluidic flow was regulated using custom, automated air-over-liquid pumps. The microfluidic channels of the devices were first flushed with 3% BSA (Sigma-Aldrich) to reduce cell

adhesion to the device walls²², then with 1x PBS (Sigma-Aldrich) and finally with the cell samples.

Control of the fibre position was achieved by inflating the triangular fibre pad features using pneumatic regulators (Marsh Bellofram) as can be seen in Figure 1. Imaging of the trapped and rotated cell was performed using either a CCD camera (ProSilica 680, AVT Technologies) or a laser-scanning confocal (SP5, Leica microsystems).

Data Analysis

Videos of the cells were processed using bespoke analysis written in MatLab. Images were aligned using the imregister function, thresholded and the position of contrast features were tracked over time. The average rotation rate was calculated by linear fitting of the angle vs time curve. Instantaneous rotation rate was calculated for each frame by taking the numerical differentiation of the curve. The resulting data was then low-passed.

Device Characterisation

Applying pressure to one side of the membrane can provide enough force to displace the fibre (Figure 1f-i) even at moderate pressures (~700 mbar), well below the breaking pressure of the device at ~4 bar). The pneumatic displacement of the fibre is reversible and repeatable (Figure 1d) and can be actuated many times with little change in the resulting position. The movement of the fibre is fast (Figure 1e) with the fibre moving to its new position in less than 100 ms. The elastic properties of the PDMS determine the size of the displacement for a given pressure. By decreasing the curing agent to elastomer ratio of the PDMS, softer devices can be produced²³ which are capable of larger displacements (Figure 2 d), however the breaking pressure of the device is lower (see Supporting Figure S1) and the fabrication is made more difficult with the softer material. The curing temperature has also been shown to affect the material properties of the resulting devices, however we saw very little difference in the fibre displacements between devices baked at different temperatures between 45 $^{\circ}\text{C}$ and 100 $^{\circ}\text{C}$ (see Supporting Figure S2). Also changes in the refractive index of PDMS have been reported when the preparation method was varied. This effect is especially important, when PDMS devices are used as optical




Figure 3 Rotation rate dependence on cell orientation. a) Graph of rotation velocity vs cell-angle. Periodic structure can be seen in the rotation rate and analysis of the images show that cell rotation speed depends on the alignment between the optical trap and the cell contents, particularly the nucleus (DRAQ5 nuclear stain shown in red). b) Repeatability of the nuclear orientation and cell rotation speed. Cells rotate faster when the long axis of an elliptical nucleus is collinear with the trap axis. c) Angular plot of the orientation of the nucleus at the fastest (green and yellow) and slowest (red and blue) points during the rotation. Solid lines are the mean angle and the coloured areas correspond to the standard deviation.

tools, for example as optofluidic lenses¹⁷. In our case, a different refractive index will also result in changes to the rotation for a given offset. Analysis of the trap position by ray optics shows a negligible shift in the trap displacement over the range of PDMS refractive indices used.

In-plane Rotation

Such displacements are enough to be able to rotate a cell, as can be seen in Figure 2. The fibres were first aligned with no additional pressure to form a stable non-rotating trap where the long axis of the cell aligns along the optical trap, before pressure was applied. Image analysis of the rotating cell provides details of the angle and rotation speed as shown in Figure 2b.

Control of the rotation rate can be achieved by modulating either the optical power or the offset between the fibres. Figure 2 demonstrates the control over the rotation rate obtainable using the device. Figure 2b shows the linear dependence of the rotation rate on the optical power in the trap. As expected, increasing the power linearly increases the rotation rate. Figure 2c shows the rotation rate against fibre offset, showing the increase in rotation rate until a maximum is reached, beyond which the offset is too large and the rotation rate decreases again.

Sub-cellular dependencies

Detailed analysis of the rotation reveals that the rotation in such a system is not constant, but has periodic features. Figure 3a is a typical rotation rate curve showing the periodic features present in optical spanner rotation. Further investigation shows that it is the orientation of the sub-cellular organelles within the optical trap which determines this substructure in the rotation rate. The optical forces acting on the trapped cell originate from differences in refractive index and so can be separated into those acting on the surface of the spheroidal cell and those acting on the organelles with a refractive index contrast to the cytoplasm. Because the distribution of refractive index inside the cell is not isotropic, as the cell rotates, the forces generated by interaction with the light beam will change, giving rise to this periodic structure. The nucleus, the largest organelle, with the large, dense packing of DNA provides most of this anisotropic force as can be seen in Figure 3a and c where the orientation of the nuclear long axis lies parallel to the optical axis when the cell is rotating at its fastest and perpendicularly to the trap axis when the cell is rotating at its slowest. Analysis of these periodic structures in the rotation rate could provide novel insight into the size, shape and refractive index of nuclei, in a non-contact, label-free technique and may lead to rapid diagnostic tools for diseases which affect nuclear morphology such as Emery-Dreifuss dystrophy²⁴, Hutchinson–Gilford progeria syndrome²⁵ as well as a wide range of other laminopathies^{26,27}. As well as diseases related to the nuclear proteins, it may also be possible to diagnose and characterise other serious diseases. Changes in nuclear morphology are one of the primary markers for the diagnosis of cancer^{28,29} and the ability to rapidly measure, at the single-cell level, properties of the nuclear morphology may provide a powerful diagnostic tool.

Tomographic Rotation

By applying pressure from above the fibre in a multi-layer device, it is possible to push the fibre downwards into the PDMS layer below the device and thus create a vertical offset between the fibres. In order to test this possibility we created a multi-layer version of the device which included additional pressure pads above one of the optical fibres, separated by a flexible, thin PDMS membrane. Such a displacement induces a rotation around the fluidic flow axis, as shown in Figure 4e and f. In contrast to the in-plane rotation, such tomographic rotation visualises the trapped cell from different directions. Although rotation properties in this direction is harder to measure, it offers the possibility to perform tomography on the trapped cell by imaging it from a series of different angles. It can be assumed that the tomographic rotation would experience the same periodic features as seen in the in-plane rotation, resulting in an uncertainty regarding the orientation of the cell during image acquisition. However this can be compensated for in post-processing using algorithms to register the unknown angles during tomographic reconstruction^{30,31}.




Figure 4 Rotation of optically trapped cells around an axis perpendicular to the trapping axis. a) Schematic of the xz cross-section through the device showing the multilayer device. b) Vertical adjustment is provided by the top layer pads (red) which when pressurized push the fibre downwards into the soft PDMS base of the device. c) false-coloured micrograph of the multi-layer device showing the features of the device - flow channel in blue, optical fibre channels in yellow, same layer pads in green and top layer pads in red. d) yz cross section through the device. Scale bar is 100 μm . e) Transmission images of a HEK293 cell rotating through 360 degrees. f) Confocal fluorescence images of the nuclear envelope of a the same HEK293 cell expressing GFP-Lamin A as the cell is rotated around the vertical axis allowing for a more detailed view of the nuclear envelope structure. Scale bar is 10 μm .

Conclusions

The integration of pneumatic valves into a dual-beam optical trap allows for the reconfigurable rotation of living cells around both the optical axis and a perpendicular axis. Rotation around the optical axis allows for accurate observation of the rotation rate, which is dependent on the size and composition of the particular cell. The system allows further details of the rotation rate properties of living cells held in the trap and may offer new insights and diagnostic possibilities for rapid cell characterisation. The reconfigurable nature of the device also allows the rotator to be optimised for the particular cell type, or even the particular cell, to give the optimum rotation rate for the available optical power and account for differences in cell size between samples. The rotation rate observed in the system is sensitive to the distribution of refractive index within the cell and it may be possible to gain further insight into the current cell state by observing the details of the rotation curve. In particular, changes to the morphology and localization of the nucleus may give rise to observable changes in the substructure of the rotation pattern, giving rise to a contactless, non-destructive, marker-free analysis of the sub-cellular conditions. The extension of the system into multi-layer devices allows the optical fibres to be manipulated in two orthogonal directions and

creates the possibility to rotate a living cell around an axis orthogonal to the optical axis and allows the cell to be visualised from different directions and opens the door to using the system for single-cell tomography.

Notes and references

1. A. Ashkin, J. M. Dziedzic, J. E. Bjorkholm, and S. Chu, *Opt. Lett.*, 1986, **11**, 288.
2. M. Friese and T. Nieminen, *Nature*, 1998, **350**, 348–350.
3. S. L. Neale, M. P. MacDonald, K. Dholakia, and T. F. Krauss, *Nat. Mater.*, 2005, **4**, 530–3.
4. N. B. Simpson, K. Dholakia, L. Allen, and M. J. Padgett, *Opt. Lett.*, 1997, **22**, 52.
5. A. O’Neil, I. MacVicar, L. Allen, and M. Padgett, *Phys. Rev. Lett.*, 2002, **88**, 053601.
6. M. K. M. Kreysing, A. Fritsch, C. Dietrich, T. Kiessling, J. R. Guck, and J. A. Kaes, *Opt. Express*, 2008, **16**, 16984–16992.
7. A. Ashkin, *Phys. Rev. Lett.*, 1970, **24**, 156–159.
8. J. Guck, R. Ananthakrishnan, T. J. Moon, C. C. Cunningham, and J. Käs, *Phys. Rev. Lett.*, 2000, **84**, 5451–4.
9. A. E. Ekpenyong, G. Whyte, K. Chalut, S. Pagliara, F. Lautenschläger, C. Fiddler, S. Paschke, U. F. Keyser, E. R. Chilvers, and J. Guck, *PLoS One*, 2012, **7**, e45237.
10. J. Guck, S. Schinkinger, B. Lincoln, F. Wottawah, S. Ebert, M. Romeyke, D. Lenz, H. M. Erickson, R. Ananthakrishnan, D. Mitchell, J. Käs, S. Ulvick, and C. Bilby, *Biophys. J.*, 2005, **88**, 3689–98.
11. A. Constable, J. Kim, and J. Mervis, *Opt. Lett.*, 1993, **18**, 1867–1869.
12. B. Lincoln, S. Schinkinger, K. Travis, F. Wottawah, S. Ebert, F. Sauer, and J. Guck, *Biomed. Microdevices*, 2007, **9**, 703–10.
13. N. Bellini, F. Bragheri, I. Cristiani, J. Guck, R. Osellame, and G. Whyte, *Biomed. Opt. Express*, 2012, **3**, 2658–68.
14. J. T. Blakely, R. Gordon, and D. Sinton, *Lab Chip*, 2008, **8**, 1350–6.
15. B. J. Black and S. K. Mohanty, *Opt. Lett.*, 2012, **37**, 5030–2.
16. a. R. Abate and D. a. Weitz, *Appl. Phys. Lett.*, 2008, **92**, 243509.
17. W. Song, A. E. Vasdekis, and D. Psaltis, *Lab Chip*, 2012, **12**, 3590–7.
18. M. J. Kennedy, S. J. Stelick, L. G. Sayam, A. Yen, D. Erickson, and C. A. Batt, *Lab Chip*, 2011, **11**, 1138–43.
19. S. Hulme, S. Shevkoplyas, and G. Whitesides, *Lab Chip*, 2009, **9**, 79–86.
20. L. M. Fidalgo, G. Whyte, B. T. Ruotolo, J. L. P. Benesch, F. Stengel, C. Abell, C. V Robinson, and W. T. S. Huck, *Angew. Chem. Int. Ed. Engl.*, 2009, **48**, 3665–8.
21. C. L. Bliss, J. N. McMullin, and C. J. Backhouse, *Lab Chip*, 2007, **7**, 1280–7.
22. Z. Long, E. Nugent, A. Javer, P. Cicuta, B. Sclavi, M. Cosentino Lagomarsino, and K. D. Dorfman, *Lab Chip*, 2013, **13**, 947–54.
23. X. Q. Brown, K. Ookawa, and J. Y. Wong, *Biomaterials*, 2005, **26**, 3123–9.
24. A. Fidziańska and I. Hausmanowa-Petrusewicz, *J. Neurol. Sci.*, 2003, **210**, 47–51.

-
25. K. N. Dahl, P. Scaffidi, M. F. Islam, A. G. Yodh, K. L. Wilson, and T. Misteli, *Proc. Natl. Acad. Sci. U. S. A.*, 2006, **103**, 10271–6.
26. M. Zwerger, C. Y. Ho, and J. Lammerding, *Nuclear mechanics in disease.*, 2011, vol. 13.
- 5 27. H. J. Worman, L. G. Fong, A. Muchir, and S. G. Young, *J. Clin. Invest.*, 2009, **119**, 1825–36.
28. D. Zink, A. H. Fischer, and J. a Nickerson, *Nat. Rev. Cancer*, 2004, **4**, 677–87.
- 10 29. P. Dey, *Diagn. Cytopathol.*, 2010, **38**, 382–90.
30. O. Renaud, J. Viña, Y. Yu, C. Machu, A. Trouvé, H. Van der Voort, B. Chalmond, and S. L. Shorte, *Biotechnol. J.*, 2008, **3**, 53–62.
31. B. Le Saux, B. Chalmond, Y. Yu, a Trouvé, O. Renaud,
15 and S. L. Shorte, *J. Microsc.*, 2009, **233**, 404–16.

*Biophysics Group, Department of Physics, Friedrich-Alexander
Universität Erlangen-Nürnberg, Henkestrasse 9, 91052 Erlangen
Germany E-mail: gwhyte@biomed.uni-erlangen.de*

20 † Electronic Supplementary Information (ESI) available: [details of any
supplementary information available should be included here]. See
DOI: 10.1039/b000000x/