

Fundamental Review to Ozone Gas Sensing Using Optical Fibre Sensors

Michael David^{1*}, Mohd Haniff Ibrahim², Sevia Mahdaliza Idrus³, Tay Ching En Marcus⁴

^{1,2,3,4}Lightwave Communication Research Group, Innovative Engineering Research Alliance,
Department of Telecommunication Engineering, Faculty of Electrical Engineering,
Universiti Teknologi Malaysia, 81310 Skudai, Johor, Malaysia

¹Department of Telecommunication Engineering, School of Engineering and Engineering Technology,
Federal University of Technology, Minna, Nigeria

*Corresponding author, email: hanif@fke.utm.my¹; mdavid2@live.utm.my²

Abstract

The manuscript is a review of basic essentials to ozone gas sensing with optical methods. Optical methods are employed to monitor optical absorption, emission, reflectance and scattering of gas samples at specific wavelengths of light spectrum. In the light of their importance in numerous disciplines in analytical sciences, necessary integral information that serves both as a basis and reference material for intending researchers and others in the field is inevitable. This review provides insight into necessary essentials to gas sensing with optical fibre sensors. Ozone gas is chosen as a reference gas. Simulation results for ozone gas absorption cross section in the ultraviolet (UV) region of the spectrum using spectralcalc.com simulation have also been included.

Keywords: Absorption spectroscopy, Beer- Lambert law, optical fibre, optical method, ozone gas, sensors

Copyright © 2015 Universitas Ahmad Dahlan. All rights reserved.

1. Introduction

In the field of analytical sciences, optical methods have become very relevant to numerous disciplines [1]. Optical methods are employed to monitor optical absorption, emission, reflectance and scattering of gas samples at specific wavelengths of light spectrum [2]. Newton's discovery of the solar spectrum in 1666 is considered to be the beginning of spectroscopy [3]. The entire spectrometric methods solely rely on emission or absorption of electromagnetic radiation [4]. Optical method relevance to science and other disciplines has made it necessary to put together in one piece essential fundamentals which could be a ready guide for all users. The necessity of a review manuscript which is intended to be a reference material is inevitable. This review provides insight into vital fundamentals to gas sensing with optical fibre sensors. It is comprised of optical sensor mechanism [5], advantages of optical sensors [6, 7], optical sensor classification [8], optical gas cells classification [9], Beer-Lambert law [10] and ozone gas and its research challenges [11-13]. Ozone is a trace gas in the atmosphere [14] and is discovered in 1839 [15]. Ozone is a useful gas, but it is a threat to human life [16-19]. Ozone gas relevance has been previously emphasised by the authors [13]. Significant volume of research activities which are not just limited to detection and monitoring are devoted to ozone gas [20-28]. These activities are summarised in Figure 1. Relevant simulation software (spectralcalc.com) was used to obtain simulation results for ozone gas absorption cross section.

2. Mechanism of Optical Sensors

"An Optical Sensor (OS) is a photonic system in which an input signal (U_i), modulates certain characteristics (absorption, dispersion, reflection, transmission, etc) of light in an optical system, such that after detection at the receiver, it is also processed and conditioned, the system will deliver an output electrical signal (U_o), which will be an exact reproduction of the object variable. If any of the processes or parts of it use fibre optic technology, a subgroup of the optical sensor known as Optical Fibre Sensors (OFS) or Fibre-Optic Sensors (FOS), is created" [29].




Figure 1. Research Activities on Ozone gas

The interaction of light with matter can be in any one of the following ways: absorption, diffraction, dispersion, reflectance, and interference [1]. Electromagnetic radiations are absorbed by chemical compounds containing covalent bonds. This is as a result of different mechanisms whose effects are seen throughout the electromagnetic spectrum [5]. Absorption of light by a molecule at a given frequency is caused by electron resonance at that given frequency [5, 30]. Light absorption by ozone gas in ultra violet (UV) region (200 to 400 nm, 6.2 to 3.0 electron volt ((eV)) as well as in the visible region (400 to 780 nm, 3.1 to 1.6eV) of the light spectrum [5, 31] is caused by excitation of valence electrons in the atoms of molecules. Light absorption in microwave region (0.3 to 300 cm.) is due to a change in rotation of the bonds in a molecule. Absorption of light in the infrared region (3 to 50 μm , 0.4 to 0.025 eV) and near infrared (0.78 to 3 μm , 1.6 to 0.4 eV) occurs due to the vibration of the bonds of a molecule [32].

While discussion of this paper is focus on absorption spectroscopy, there are other classes of optical sensors such as:

- Reflection spectroscopy [33, 34]
- Luminescence intensity spectroscopy [35]
- Fluorescence lifetime spectroscopy [36]
- Refractive index spectroscopy [37]
- Surface Plasmon resonance or ellipsometric spectroscopy [38]

The classifications are meant to give a clear picture and are not discussed further. Ozone, the gas of interest in this article, absorbs light intensity and hence absorption spectroscopy is dwelled upon in other sections of this article.

Measurement of radiation absorbed by atoms is described as atomic absorption spectroscopy (3). The history of optical sensors can be traced back to when pH indicator strips were developed by immobilizing pH-sensitive indicators on cellulose. The absorption spectrum of each species is unique and can be used to identify and quantify presence of that specie.

3. Merits of Fibre Sensors

The authors have previously [13] highlighted quite a number of different methods for detecting ozone gas such as: cavity enhanced absorption spectroscopy (CEAS) [39, 40], cavity ring down spectroscopy [41], chemiluminescence [42, 43], electrochemical concentration cells [44], photo-acoustic sensors [45, 46], photo reductive [47], solid state sensors [48] and UV absorption [49]. Authors of reference [50] have shown the compatibility of fibre sensors with optical communication systems and their application in electrical noisy systems and explosion

prone scenarios. They offer good resistance to corrosion prone environments and high-voltage and high-temperature environments. In Table 1, we compare the performance of optical spectroscopy with other sensing methods.

Table 1. Gas sensors comparison

Sensor Type	Merits	Demerits
Photo acoustic Spectroscopy	<ul style="list-style-type: none"> ▪ High sensitivity ▪ Response time is fast ▪ Measurement is free from background noise ▪ Requires no reference as a result of noise(51) 	<ul style="list-style-type: none"> ▪ Selectivity is poor for photo acoustic system that utilises infrared light sources (52)
Photo reductive gas sensor	<ul style="list-style-type: none"> ▪ Good sensitivity ▪ Short response time ▪ Inexpensive 	<ul style="list-style-type: none"> ▪ Temperature requirement is high ▪ Energy dissipation is high (53, 54)
Electro-chemical Sensors	<ul style="list-style-type: none"> ▪ They are portable ▪ Exhibits high sensitivity. ▪ They are inexpensive (55) 	<ul style="list-style-type: none"> ▪ There is the depletion of electrolyte when used for sensing high ozone concentrations. ▪ It requires frequent maintenance (56, 57)
Metal oxide ozone sensors	<ul style="list-style-type: none"> ▪ Broad range of application (58) 	<ul style="list-style-type: none"> ▪ High temperature requirements of detectors which translate into: ▪ High energy consumption. ▪ High cost ▪ Fabrication and size limitations (55, 56, 59) ▪ Characteristic activity is high ▪ Film sensor thickness requirement is large when applied for ozone sensing (56, 60, 61)
Solid State	<ul style="list-style-type: none"> ▪ Consumes less energy ▪ Good sensitivity ▪ Fast response time ▪ Inexpensive ▪ Light weight 	<ul style="list-style-type: none"> ▪ Requires to be calibrated within every one hour (every 1 to 60 minutes) (43). It is not absolute.
Chemiluminescence.	<ul style="list-style-type: none"> ▪ Fast response time (43) 	<ul style="list-style-type: none"> ▪ Gas sample must be able in a distinct manner to either absorb, emit, or scatter transmitted light rays at specific region of the light spectrum (7, 57, 62);
Optical spectroscopy	<ul style="list-style-type: none"> • It is a rapid and direct means of sensing gases with good cross sensitivity (57) • Require no consumables either for calibration or operation • Anti-electric magnetic interference, • Excellent electrical insulativity, and • Suitability of long-distance online measurement 	<ul style="list-style-type: none"> ▪ Expensive ▪ Large in size (6)

4. Sensor Classification

Fibre optic sensors can be classified based on method of fibre application in sensor system and modulation mechanism [8]. The classification is illustrated in Figure 2.




Figure 2. Optical sensor classification

4.1. Classification Based on Fibre Application

Fibre optic sensors are categorised into intrinsic and extrinsic types.

4.1.1. Intrinsic Optical Sensors

In an intrinsic fibre optic sensor, light is restricted within the optical fibre and modulation of the light signal is within the fibre [8, 63]; it is illustrated in Figure 3.




Figure 3. An Intrinsic Fibre Optic Sensor

4.1.2. Extrinsic Optical Sensors

In an extrinsic sensor, interaction between light signal (i.e. light signal modulation) and the sample to be measured takes place outside the optical fibre cable in a gas cell generally referred to a cuvette [8, 64]. It is illustrated in Figure 4.




Figure 4. An Extrinsic Fibre Optic Sensor

4.2. Classification Based on Modulation Mechanism

In the application of light for sensing in fibre optic sensors, different characteristics of light are modulated to achieve sensing. These characteristics include: intensity, phase, polarization and wavelength [8]. Ozone gas measurement with optical absorption spectroscopy is detected by light intensity modulation.

5. Basic Experimental Setup for Ozone Detection via Optical Absorption Spectroscopy

A typical absorption spectroscopic experimental setup is made up of the following components: source of light radiation, a monochromator (except when light source is a laser). Light sources can either be broadband or chromatic [65]. Light emanating from a broadband light source must be propagated through a collimating lens to eliminate scattering effects. Light coupler, waveguide (fibre, fibre bundle, planar wave guide), variable attenuator, lenses (optical), cuvette (absorption cell or gas cell), light detection unit (spectrometer, photo detector), amplifier, secondary filter, transducer, data acquisition unit, data processing unit, and display unit [1, 66]. Figure 5 is a typical experimental setup for ozone measurements using optical absorption spectroscopy. It is the typical extrinsic setup.




Figure 5. A basic layout of an optical absorption spectroscopy for ozone measurements

5.2. Classification of Gas Cells

The design of an optical gas cell in absorption spectroscopy is a major factor that affects the overall system performance in the form of sensitivity and speed of response. Authors of reference [9] have classified gas cells based on the principles of light transmission. The classification includes transmission type, reflective type, slow light and refractive index periodic change. More information on this can be obtained from reference [9].

6. Beer-Lambert Law

Absorption spectroscopy is the quantification of the energy that molecules absorb and is translated to the bending and stretching of the bonds between the atoms in the molecules [67]. The working principle of gas cells in optical gas sensor is based on the Beer-Lambert law. According to Beer and Lambert, the concentration of a sample can be determined by detecting the intensity of the output light. Beer-Lambert law describes the relation of the input light and the output light that are affected by the measuring gas.

Beer's law: it states that the fraction of the incident light absorbed is proportional to the number of the absorbing molecules in the light-path and will increase with increasing concentration or sample thickness [10].

Lambert's law: it states that the fraction of monochromatic light absorbed by a homogeneous medium (sample) is independent of the intensity of the incident light and each successive unit layer absorbs an equal fraction of the light incident on it [10].

The combination of the two laws together yields the Beer-Lambert law. If radiation of intensity I_0 (zero sample concentration) is directed at a sample in a path length L , radiation of intensity I_t leaves the sample [68]. Beer-Lambert law shows the mathematical expression of the relation between the absorbing samples concentration (c) and absorbance (A). It written as follow:

$$A = \varepsilon \times c \times L \quad (1)$$

Where:

ε = molar absorption coefficient ($\text{m}^2 \text{mol}^{-1}$)

c = sample concentration (mol m^{-3}) and

L = optical path length in (m)

In an experimental scenario, measurements are obtained in the form of transmittance T defined as:

$$T = e^{-\varepsilon c l} = \frac{I_t}{I_0} \quad (2)$$

The ratio $\frac{I_t}{I_0}$ is defined as the transmittance T :

From equation 2, absorbance A can also be defined as:

$$A = \ln \frac{I_0}{I_t} = \varepsilon c L = \text{Optical density } (D), \text{optical depth (69) or optical thickness (70)} \quad (3)$$

7. Absorption of Light by Ozone

Ozone gas detection via optical absorption spectroscopy is generally accepted [71]. This method has an inherent advantage to measure ozone absolutely without the requirement for consumables to operate or calibrate [7]. Whereas, ozone measurement with the method of chemiluminescence is not absolute, it has to be frequently calibrated. Chemiluminescence technique requires to be calibrated every 1 to 60 minutes [43]. Ozone absorbs light in the Hartley band (200–310 nm) [72], the Huggins band (310–375 nm), the Chappius band (375–603 nm), and the Wulf band (beyond 700 nm). It has peak absorption at 253.65nm ($\sigma_{253.65} = 1.147 \times 10^{-17} \text{cm}^2/\text{molecule}$) (73) and 603nm ($\sigma_{603} = 5.18 \times 10^{-21} \text{cm}^2/\text{molecule}$) (64).

7.1. The Absorption Cross Section of Ozone

Error free measurement of ozone gas is dependent upon ozone gas absorption cross section [74]. Hence, lots of research efforts are devoted to investigate the accurate value of ozone absorption cross section [64, 75-78]. Spectralcalc.com simulator has been used in this review to show the effect of temperature on absorption cross section in the Hartley band. Figure 6 shows absorption cross section of ozone gas obtained by simulation with spectralcalc.com at temperatures of 200 K and 300 K respectively. Ozone gas absorption cross section at 253.65 (actual spectral line is 253.6526 nm) is $1.16 \times 10^{-17} \text{cm}^2/\text{molecule}$ and $1.14 \times 10^{-17} \text{cm}^2/\text{molecule}$ at temperatures of 200 K and 300 K respectively. Absorption cross section decreases with increase in temperature from 200 K to 300 K. The percentage decrease is 0.95 % at a measurement wavelength of 253.6 nm. Malicet *et al* reported a decrease of 1 % in absorption cross section for a temperature rise from 218 K to 295 K [79]. Similarly, Serdyuchenko *et al* reported a slight decrease in absorption cross section with temperature increase in the Hartley band [80]. The result thus obtained is in good agreement with previous works.

8. Materials Compatibility with Ozone

Not all materials are compatible with ozone gas. Ozone gas compatibility with common materials used for ozone sensing in literature is compared in Table 2.

The rating in the table depicts chemical effect of ozone on the listed materials. A material rated "A" (excellent) implies ozone has no effect; "B" (good) ozone has minor effect. Other categories not included in the table are "C" (fair), which implies ozone effect is moderate and "D" means ozone has a severe effect on the material. The rating as defined by Ozone solutions is for ozone gas concentrations greater than 1000 ppm [81].




Figure 6. Spectralcalc.com simulation of ozone absorption cross section at 200K and 300K

Table 2. Materials compatibility with ozone (81)

Material	Rating	Example of applications
Aluminium	B - Good	(64)
Brass	B - Good	(82)
Glass	A - Excellent	(53, 83)
PTFE (Teflon®)	A - Excellent	(30)
Silicone	A - Excellent	(84, 85)
Stainless steel - 304	B - Good/Excellent	(86)
Stainless steel - 316	A - Excellent	(86)
Viton®	A - Excellent	(87)

9. Research Challenges

Recent research activities on ozone gas sensing with optical absorption spectroscopy include sensitivity enhancement through optical path length and ozone gas absorption cross section optimization [88] and effect of noise reduction on absorption cross section of ozone gas in the visible spectrum [89]. Redefinition of the value of ozone gas absorption cross section in the UV for accurate measurements of ozone gas [90] and preservation of linearity of Beer-Lamberts law by measuring ozone gas concentration at an alternate sampling wavelength of 279.95 nm [85, 91]. Ozone gas measurement in the visible spectrum using LED as a light source at 605 nm [92] and sensitivity enhancement through light propagation at incident angle [93]. Temperature and pressure dependence of ozone gas absorption cross section in the UV and visible spectrums [80, 94]. Performance indicators/metrics of ozone sensors and sensors in general include selectivity, sensitivity, accuracy, resolution, response time, fabrication cost, dynamic range, precision and linearity [58, 95-99]. Sensor requirements either in performance, physical, or cost, are application dependent [100]. Research activities on sensors in general and ozone sensors in particular, are aimed towards meeting recent sensing requirements, strengthening and upgrading some or all of the aforementioned parameters [11, 12, 49].

10. Conclusions

The review paper summarises necessary information. It is a ready reference material for new researchers in the field of absorption spectroscopy for ozone sensor application. Issues discussed include basic operating principles of optical sensors and its mechanism. Optical sensors as well as optical gas cells were classified. Specific properties of ozone gas were also highlighted. Recent research activities have been enumerated. Spetralcalc.com simulation software was used to demonstrate possibility of obtaining preliminary results before experiments are conducted.

Acknowledgements

The authors would like to thank Universiti Teknologi Malaysia (UTM) for sponsoring this publication under Research University Grant (RUG) Scheme, grant no: 05J60 and 04H35. The Ministry of Higher Education (MOE) Malaysia is acknowledged for provision of Fundamental Research Grant Scheme (FRGS) grant no: 4F317 and 4F565. The Nigerian Education Trust Fund (ETF) is also acknowledged for the financial support giving inform of Tertiary Education Trust Fund (TET-Fund).

References

- [1] Otto SW. *Fibre Optic Chemical Sensors and Biosensors Volume I*. CRC Press Boca Raton Ann Boston London. 1991: 2 & 26.
- [2] Department of Chemistry. Beer-Lambert Law. The University of Adelaide Australia <http://www.chemistry.adelaide.edu.au/external/soc-rel/content/beerslaw.htm> Accessed online on 12th February, 2013.
- [3] Ebdon L, Evans EH. *An introduction to analytical atomic spectrometry*. John Wiley & Sons. 1998.
- [4] Fifield F, Kealey D. *Analytical Chemistry: principles and practice*. Blackwell Science, Ltd.: Malden, MA; Fifth Edition. 2000: 270.
- [5] Mulrooney J, Clifford J, Fitzpatrick C, Chambers P, Lewis E, editors. Detection of carbon dioxide emissions from a land transport vehicle using a mid-infrared optical fiber based sensor. *Optics East 2006; 2006: International Society for Optics and Photonics*.
- [6] Gaddari A, Berger F, Amjoud M, Sanchez J, Lahcini M, Rhouta B, et al. A novel way for the synthesis of tin dioxide sol-gel derived thin films: Application to O₃ detection at ambient temperature. *Sensors and Actuators B: Chemical*. 2013; 176: 811-7.
- [7] Gomez A, Rosen E. Fast response cavity enhanced ozone monitor. *Atmospheric Measurement Techniques*. 2013; 6(2).
- [8] Yu M. Fiber Optic Sensor Technology. Sensors and Actuators Laboratory, Department of Mechanical Engineering, University of Maryland, USA <https://www.google.com/#q=Fiber+Optic+Sensor+Technology+Miao+Yu>. 2008.
- [9] Zhao Y, Bai L, Zhang YN, Hou W, Wang Q. Review On Structures And Principles Of Gas Cells In The Absorption Spectrum-Based Optical Fiber Gas Sensor Systems. *Instrumentation Science & Technology*. 2012; 40(5): 385-401.
- [10] Singh D, Deshwal B, Kumar VS. *The absorption laws and measurement of absorption intensity*. *Comprehensive Engineering Chemistry*. IK International Publishing House, New Delhi. 2008: 259.
- [11] Fang J, Park SC, Schlag L, Stauden T, Pezoldt J, Jacobs HO. Localized Collection of Airborne Analytes: A Transport Driven Approach to Improve the Response Time of Existing Gas Sensor Designs. *Advanced Functional Materials*. 2014.
- [12] Colindres SC, Aguir K, Cervantes Sodi F, Vargas LV, Salazar JAM, Febles VG. Ozone Sensing Based on Palladium Decorated Carbon Nanotubes. *Sensors*. 2014; 14(4): 6806-18.
- [13] David M, Ibrahim MH, Idrus SM, Azmi AI, Ngajikin NH, Marcus TCE, et al. Progress in Ozone Sensors Performance: A Review. *Jurnal Teknologi*. 2015; 73(6).
- [14] Janssen C, Simone D, Guinet M. Preparation and accurate measurement of pure ozone. *Review of Scientific Instruments*. 2011; 82(3): 034102.
- [15] Von Sonntag C, Von Gunten U. *Chemistry of Ozone in Water and Wastewater Treatment*. IWA Publishing. 2012.
- [16] Rubin MB. The history of ozone. The Schönbein period, 1839–1868. *Bull Hist Chem*. 2001; 26(1): 40-56.
- [17] Brimblecombe P. Interest in air pollution among early Fellows of the Royal Society. Notes and records of the Royal Society of London. 1978; 32(2): 123-9.
- [18] Roscoe HK, Clemittshaw KC. Measurement techniques in gas-phase tropospheric chemistry: A selective view of the past, present, and future. *Science*. 1997; 276(5315): 1065-72.
- [19] Udhayakumar G, Rashmi M, Patel K, Ramesh G, Suresh A. Supply Power Factor Improvement in Ozone Generator System Using Active Power Factor Correction Converter. *International Journal of Power Electronics and Drive Systems (IJPEDS)*. 2015; 6(2).
- [20] David M, Marcus TCE, Yaacob M, Salim MR, Hussin N, Ibrahim MH, et al. A New Ozone Concentration Regulator. *TELKOMNIKA Indonesian Journal of Electrical Engineering*. 2015; 13(2).
- [21] Facta M, Salam Z, Buntat Z. A New Type of Planar Chamber for High Frequency Ozone Generator System. *Advanced Materials Research*. 2014; 896: 726-9.
- [22] Facta M, Sutikno T, Salam Z. The Application of FPGA in PWM Controlled Resonant Converter for an Ozone Generator. *International Journal of Power Electronics and Drive Systems (IJPEDS)*. 2013; 3(3): 336-43.
- [23] Cole J, Su S, Blakeley R, Koonath P, Hecht A. Radiolytic yield of ozone in air for low dose neutron and x-ray/gamma-ray radiation. *Radiation Physics and Chemistry*. 2015; 106: 95-8.

- [24] Kumharn W, Sudhibrabha S. Study of ozone and sulfur dioxide using Thailand based Brewer Spectrophotometers. *Advances in Space Research*. 2014; 53(5): 802-9.
- [25] Li YL. Experimental Investigation on Ozone Mass Transfer Coefficient Enhanced by Electric Field in Liquid Phase. *Advanced Materials Research*. 2014; 864: 2139-44.
- [26] Costagliola MA, Murena F, Prati MV. Exhaust emissions of volatile organic compounds of powered two-wheelers: Effect of cold start and vehicle speed. Contribution to greenhouse effect and tropospheric ozone formation. *Science of the Total Environment*. 2014; 468: 1043-9.
- [27] Lefohn AS, Emery C, Shadwick D, Wernli H, Jung J, Oltmans SJ. Estimates of background surface ozone concentrations in the United States based on model-derived source apportionment. *Atmospheric Environment*. 2014; 84: 275-88.
- [28] Guo WQ, Yin RL, Zhou XJ, Du JS, Cao HO, Yang SS, et.al. Sulfamethoxazole degradation by ultrasound/ozone oxidation process in water: Kinetics, mechanisms, and pathways. *Ultrasonics sonochemistry*. 2015; 22: 182-7.
- [29] López-Higuera JM. Optical sensors: Ed. Universidad de Cantabria; 1998.
- [30] O'Keeffe S, Dooly G, Fitzpatrick C, Lewis E, editors. Optical fibre sensor for the measurement of ozone. *Journal of Physics: Conference Series*. 2005: IOP Publishing.
- [31] Arshak K, Hickey G, Forde E, Harris J, editors. Development of novel room temperature ozone sensors for health and safety applications. Electronics Technology, 30th International Spring Seminar on; 2007: IEEE.
- [32] Degner M, Damaschke N, Ewald H, Lewis E, editors. High resolution LED-spectroscopy for sensor application in harsh environment: A sensor system based on LED-light sources and standard photodiode receiver is shown as an example of this sensor concept for in-situ gas measurements down to the ppb range. 2010 IEEE International Instrumentation and Measurement Technology Conference, I2MTC 2010, May 3, 2010 - May 6, 2010; 2010; Austin, TX, United states: IEEE Computer Society.
- [33] Reynolds CS. Measuring black hole spin using x-ray reflection spectroscopy. *The Physics of Accretion onto Black Holes*: Springer; 2015. p. 277-94.
- [34] Garcia JA, Dauser T, Steiner JF, McClintock JE, Keck ML, Wilms J. On Estimating the High-Energy Cutoff in the X-ray Spectra of Black Holes via Reflection Spectroscopy. arXiv preprint arXiv:150503616. 2015.
- [35] Cavalli E, Angiuli F, Belletti A, Boutinaud P. Luminescence spectroscopy of YVO₄: Ln³⁺, Bi³⁺ (Ln³⁺= Eu³⁺, Sm³⁺, Dy³⁺) phosphors. *Optical Materials*. 2014; 36(10): 1642-8.
- [36] Coda S, Thompson AJ, Kennedy GT, Roche KL, Ayaru L, Bansi DS, et al. Fluorescence lifetime spectroscopy of tissue autofluorescence in normal and diseased colon measured ex vivo using a fiber-optic probe. *Biomedical optics express*. 2014; 5(2): 515-38.
- [37] Zarzana KJ, Cappa CD, Tolbert MA. Sensitivity of Aerosol Refractive Index Retrievals Using Optical Spectroscopy. *Aerosol Science and Technology*. 2014; 48(11): 1133-44.
- [38] Del Rosso T, Sánchez J, Carvalho R, Pandoli O, Cremona M. Accurate and simultaneous measurement of thickness and refractive index of thermally evaporated thin organic films by surface plasmon resonance spectroscopy. *Optics express*. 2014; 22(16): 18914-23.
- [39] Darby SB, Smith PD, Venables DS. Cavity-enhanced absorption using an atomic line source: application to deep-UV measurements. *Analyst*. 2012; 137(10): 2318-21.
- [40] Gomez A, Rosen E. Fast response cavity enhanced ozone monitor. *Atmospheric Measurement Techniques Discussions*. 2012; 5(5): 7223-41.
- [41] Washenfelder R, Wagner N, Dube W, Brown S. Measurement of atmospheric ozone by cavity ring-down spectroscopy. *Environmental science & technology*. 2011; 45(7): 2938-44.
- [42] Ermel M, Oswald R, Mayer JC, Moravek A, Song G, Beck M, et al. Preparation methods to optimize the performance of sensor discs for fast chemiluminescence ozone analyzers. *Environmental Science and Technology*. 2013; 47(4): 1930-6.
- [43] Zahn A, Weppner J, Widmann H, Schlote-Holubek K, Burger B, Kühner T, et al. A fast and precise chemiluminescence ozone detector for eddy flux and airborne application. *Atmospheric Measurement Techniques*. 2012; 5(2): 363-75.
- [44] Eastman JA, Stedman DH. A fast response sensor for ozone eddy-correlation flux measurements. *Atmospheric Environment (1967)*. 1977; 11(12): 1209-11.
- [45] Bottger S, Kohring M, Willer U, Schade W. Off-beam quartz-enhanced photoacoustic spectroscopy with LEDs. *Applied Physics B: Lasers and Optics*. 2013; 113(2): 227-32.
- [46] Gondal MA, Dastageer A, Yamani ZH. Laser-induced photoacoustic detection of ozone at 266 nm using resonant cells of different configuration. *Journal of Environmental Science and Health - Part A Toxic/Hazardous Substances and Environmental Engineering*. 2009; 44(13): 1457-64.
- [47] Chien FSS, Wang CR, Chan YL, Lin HL, Chen MH, Wu RJ. Fast-response ozone sensor with ZnO nanorods grown by chemical vapor deposition. *Sensors and Actuators, B: Chemical*. 2010; 144(1): 120-5.

- [48] Mastelaro VR, Zílio SC, da Silva LF, Pelissari PI, Bernardi MI, Guerin J, et al. Ozone gas sensor based on nanocrystalline SrTi_{1-x}Fe_xO₃ thin films. *Sensors and Actuators B: Chemical*. 2013.
- [49] Gao R, Ballard J, Watts L, Thornberry T, Ciciora S, McLaughlin R, et al. A compact, fast UV photometer for measurement of ozone from research aircraft. *Atmospheric Measurement Techniques*. 2012; 5(9): 2201-10.
- [50] Diemeer M, Trommel E. Fiber-optic microbend sensors: sensitivity as a function of distortion wavelength. *Optics letters*. 1984; 9(6): 260-2.
- [51] Kohring M, Willer U, Böttger S, Pohlkötter A, Schade W. Fiber-coupled ozone sensor based on tuning fork-enhanced interferometric photoacoustic spectroscopy. *IEEE Journal on Selected Topics in Quantum Electronics*. 2012; 18(5): 1566-72.
- [52] Mao X, Zhou X, Zhai L, Yu Q. Dissolved Gas-in-Oil Analysis in Transformers Based on Near-Infrared Photoacoustic Spectroscopy. *International Journal of Thermophysics*. 2014: 1-7.
- [53] Wu RJ, Chiu YC, Wu CH, Su YJ. Application of Au/TiO₂-WO₃ material in visible light photoreductive ozone sensors. *Thin Solid Films*. 2015; 574: 156-61.
- [54] Chen MH, Lu CS, Wu RJ. Novel Pt/TiO₂-WO₃ materials irradiated by visible light used in a photoreductive ozone sensor. *Journal of the Taiwan Institute of Chemical Engineers*. 2014; 45(3): 1043-8.
- [55] Wang Z, Qiu X, Tang R, Oiler J, Zhu J, Huang H, et al., editors. Ozone sensor using ZnO based film bulk acoustic resonator. 2011 16th International Solid-State Sensors, Actuators and Microsystems Conference, TRANSDUCERS'11, June 5, 2011 - June 9, 2011; 2011; Beijing, China: IEEE Computer Society.
- [56] Wang CY, Becker R, Passow T, Pletschen W, Köhler K, Cimalla V, et al. Photon stimulated sensor based on indium oxide nanoparticles I: wide-concentration-range ozone monitoring in air. *Sensors and Actuators B: Chemical*. 2011; 152(2): 235-40.
- [57] Dakin JP, Chambers P. Review of methods of optical gas detection by direct optical spectroscopy, with emphasis on correlation spectroscopy. *Optical Chemical Sensors: Springer*. 2006: 457-77.
- [58] Da Silva LF, Catto AC, Avansi Jr W, Cavalcante LS, Andres J, Aguir K, et al. A novel ozone gas sensor based on one dimensional (1D) α-Ag₂WO₄ nanostructures. *Nanoscale*. 2014.
- [59] Yu JH, Yang HJ, Mo HS, Kim TS, Jeong TS, Youn CJ, et al. Sensing mechanism and behavior of sputtered ZnCdO ozone sensors enhanced by photons for room- temperature operation. *Journal of Electronic Materials*. 2013; 42(4): 720-5.
- [60] Thirumalairajan S, Mastelaro VR, Escanhoela Jr CA. In-Depth Understanding of the Relation between CuAlO₂ Particle Size and Morphology for Ozone Gas Sensor Detection at a Nanoscale Level. *ACS applied materials & interfaces*. 2014; 6(23): 21739-49.
- [61] Korotcenkov G, Cho B. Ozone measuring: What can limit application of SnO₂-based conductometric gas sensors? *Sensors and Actuators B: Chemical*. 2012; 161(1): 28-44.
- [62] Chambers P. A study of a correlation spectroscopy gas detection method: University of Southampton; 2005.
- [63] Tracey PM. Intrinsic fiber-optic sensors. *IEEE Transactions on Industry Applications*. 1991; 27(1): 96-8.
- [64] O'Keeffe S, Fitzpatrick C, Lewis E. An optical fibre based ultra violet and visible absorption spectroscopy system for ozone concentration monitoring. *Sensors and Actuators B: Chemical*. 2007; 125(2): 372-8.
- [65] Fein H, Liu SY. Chemical sensing techniques employing liquid-core optical fibers. *Google Patents*. 2000.
- [66] O'Keeffe S, Fitzpatrick C, Lewis E. Ozone measurement in visible region: an optical fibre sensor system. *Electronics Letters*. 2005; 41(24): 1317-9.
- [67] Hardiman R, Mckee D, Kimmerle K. Qualitative Comparison of Cavity Ring-Down vs. Direct Measurement Absorption Spectroscopy of Determining ppb Moisture Levels in UHP Gases. *Gases and Technology*. 2004; 3(3).
- [68] Denney RC. *Dictionary of spectroscopy*. Willey 2nd Edition. 1982: 119-20.
- [69] En Marcus TC, David M, Yaacob M, Salim MR, Ibrahim MH, Ngajikin NH, et al. Absorption Cross Section Simulation: a Preliminary Study of Ultraviolet Absorption Spectroscopy for Ozone Gas Measurement. *Jurnal Teknologi*. 2013; 64(3).
- [70] Larsen ML, Clark AS. On the link between particle size and deviations from the Beer-Lambert-Bouguer law for direct transmission. *Journal of Quantitative Spectroscopy and Radiative Transfer*. 2014; 133: 646-51.
- [71] Köhring M, Willer U, Böttger S, Pohlkötter A, Schade W. Fiber-Coupled Ozone Sensor Based on Tuning Fork-Enhanced Interferometric Photoacoustic Spectroscopy. *IEEE Journal of Selected Topics in Quantum Electronics*. 2012; 18(5): 1566-72.
- [72] Murata I, Sato K, Okano S, Tomikawa Y. Measurements of stratospheric ozone with a balloon-borne optical ozone sensor. *International Journal of Remote Sensing*. 2009; 30(15-16): 3961-6.

- [73] Hearn A. *The absorption of ozone in the ultra-violet and visible regions of the spectrum*. Proceedings of the Physical Society. 1961; 78(5): 932.
- [74] Bogumil K, Orphal J, Homann T, Voigt S, Spietz P, Fleischmann O, et al. Measurements of molecular absorption spectra with the SCIAMACHY pre-flight model: instrument characterization and reference data for atmospheric remote-sensing in the 230–2380 nm region. *Journal of Photochemistry and Photobiology A: Chemistry*. 2003; 157(2): 167-84.
- [75] Griggs M. Absorption coefficients of ozone in the ultraviolet and visible regions. *The Journal of Chemical Physics*. 1968; 49: 857-9.
- [76] Inn EC, Tanaka Y. Absorption coefficient of ozone in the ultraviolet and visible regions. *JOSA*. 1953; 43(10): 870-2.
- [77] Vigroux E. Contribution à l'étude expérimentale de l'absorption de l'ozone, par Ernest Vigroux: Masson; 1953.
- [78] Brion J, Chakir A, Daumont D, Malicet J, Parisse C. High-resolution laboratory absorption cross section of O_3 . Temperature effect. *Chemical physics letters*. 1993; 213(5): 610-2.
- [79] Malicet J, Daumont D, Charbonnier J, Parisse C, Chakir A, Brion J. Ozone UV spectroscopy. II. Absorption cross-sections and temperature dependence. *Journal of atmospheric chemistry*. 1995; 21(3): 263-73.
- [80] Serdyuchenko A, Gorshchev V, Weber M, Chehade W, Burrows J. High spectral resolution ozone absorption cross-sections—Part 2: Temperature dependence. *Atmospheric Measurement Techniques*. 2014; 7(2): 625-36.
- [81] Solutions O. Material Compatibility with Ozone. http://www.ozoneapplications.com/info/ozone_compatible_materials.htm. 2015.
- [82] Marcus TCE, David M, Yaacob M, Salim MR, Ibrahim MH, Ngajikin NH, et al. Interchangeable Range of Ozone Concentration Simulation for Low Cost Reconfigurable Brass Gas Cell. *Jurnal Teknologi*. 2014; 69(8).
- [83] Yu J, Yang H, Mo H, Kim T, Jeong T, Youn C, et al. Sensing Mechanism and Behavior of Sputtered ZnCdO Ozone Sensors Enhanced by Photons for Room-Temperature Operation. *Journal of electronic materials*. 2013; 42(4): 720-5.
- [84] Acuautla Mn, Bernardini S, Bendahan M. Ozone Sensor on Flexible Substrate by ZnO Nanoparticles. *Key Engineering Materials*. 2014; 605: 163-6.
- [85] Marcus TC, David M, Yaacob M, Salim MR, Hussin N, Ibrahim MH, et al. Alternative wavelength for linearity preservation of Beer–Lambert Law in ozone concentration measurement. *Microwave and Optical Technology Letters*. 2015; 57(4): 1013-6.
- [86] Teranishi K, Shimada Y, Shimomura N, Itoh H. Investigation of Ozone Concentration Measurement by Visible Photo Absorption Method. *Ozone, Science & Engineering*. 2013; 35(3): 229-39.
- [87] Kondo T, Sakai K, Watanabe T, Einaga Y, Yuasa M. Electrochemical detection of lipophilic antioxidants with high sensitivity at boron-doped diamond electrode. *Electrochimica Acta*. 2013; 95: 205-11.
- [88] Marcus TCE, Ibrahim MH, Ngajikin NH, Azmi AI. Optical path length and absorption cross section optimization for high sensitivity ozone concentration measurement. *Sensors and Actuators B: Chemical*. 2015; 221: 570-5.
- [89] David M, Ibrahim MH, Idrus SM. Sampling frequency effect on the absorption cross section of ozone in the Visible Spectrum. *Journal of optoelectronics and advanced materials*. 2015; 17(3-4): 403 -8.
- [90] Viallon J, Lee S, Moussay P, Tworek K, Petersen M, Wielgosz R. Accurate measurements of ozone absorption cross-sections in the Hartley band. *Atmospheric Measurement Techniques*. 2015; 8(3): 1245-57.
- [91] Marcus TC, David M, Yaacob M, Salim MR, Hussin N, Ibrahim MH, et al. Erratum for: Alternative wavelength for linearity preservation of Beer–Lambert law in ozone concentration measurement. *Microwave and Optical Technology Letters*. 2015; 57(7): 1768-.
- [92] Jodpimai S, Boonduang S, Limsuwan P. Inline ozone concentration measurement by a visible absorption method at wavelength 605nm. *Sensors and Actuators B: Chemical*. 2015.
- [93] David M, Marcus TCE, Yaacob M, Salim MR, Hussin N, Ibrahim MH, et al., editors. Incident Angle Approach to Sensitivity Enhancement for Ozone Sensor. *Applied Mechanics and Materials*; 2015: Trans Tech Publ.
- [94] Marcus TCE, David M, Yaacob M, Salim MR, Hussin N, Ibrahim MH, et al., editors. Pressure Dependence of Ozone Absorption Cross Section. *Applied Mechanics and Materials*; 2015: Trans Tech Publ.
- [95] Proffitt MH, McLaughlin RJ. Fast-response dual-beam UV-absorption ozone photometer suitable for use on stratospheric balloons. *Review of scientific instruments*. 1983; 54(12): 1719-28.
- [96] Jadsadapattarakul D, Thanachayanont C, Nukeaw J, Sooknoi T. Improved selectivity, response time and recovery time by [010] highly preferred-orientation silicalite-1 layer coated on SnO₂ thin film sensor for selective ethylene gas detection. *Sensors and Actuators B: Chemical*. 2010; 144(1): 73-80.

-
- [97] Sarfraz J, Ihalainen P, Määttä A, Gulin T, Koskela J, Wilén C-E, et al. A printed H₂S sensor with electro-optical response. *Sensors and Actuators B: Chemical*. 2014; 191: 821-7.
- [98] Tsitron J, Kreller CR, Sekhar PK, Mukundan R, Garzon FH, Brosha EL, et al. Bayesian decoding of the ammonia response of a zirconia-based mixed-potential sensor in the presence of hydrocarbon interference. *Sensors and Actuators B: Chemical*. 2014; 192: 283-93.
- [99] Liu H, Li M, Voznyy O, Hu L, Fu Q, Zhou D, et al. Physically Flexible, Rapid-Response Gas Sensor Based on Colloidal Quantum Dot Solids. *Advanced Materials*. 2014.
- [100] Burgess LW. Absorption-based sensors. *Sensors and Actuators B: Chemical*. 1995; 29(1): 10-5.