



Thomas Jefferson University
Jefferson Digital Commons

Phase 1

Class of 2022

1-2020

3D Printing of Bone Spurs Before Surgical Removal

Matthew Chadwick

John Paul Prodoehl

Katelyn Koons

Chris Li

Lauren Schlegel

See next page for additional authors

Follow this and additional works at: https://jdc.jefferson.edu/si_des_2022_phase1

 Part of the [Art and Design Commons](#), [Orthopedics Commons](#), and the [Surgery Commons](#)

[Let us know how access to this document benefits you](#)

This Article is brought to you for free and open access by the Jefferson Digital Commons. The Jefferson Digital Commons is a service of Thomas Jefferson University's [Center for Teaching and Learning \(CTL\)](#). The Commons is a showcase for Jefferson books and journals, peer-reviewed scholarly publications, unique historical collections from the University archives, and teaching tools. The Jefferson Digital Commons allows researchers and interested readers anywhere in the world to learn about and keep up to date with Jefferson scholarship. This article has been accepted for inclusion in Phase 1 by an authorized administrator of the Jefferson Digital Commons. For more information, please contact: JeffersonDigitalCommons@jefferson.edu.

Authors

Matthew Chadwick; John Paul Prodoehl; Katelyn Koons; Chris Li; Lauren Schlegel; and Jeffrey Henstenburg, MD

Project Title: 3D Printing of Bone Spurs Before Surgical Removal

Student Name: Matthew Chadwick

Co-author(s): John Paul Prodoehl**, Katelyn Koons**, Chris Li**, Lauren Schlegel**, Jeffrey Henstenburg, MD*

Project Background: In the US alone, total knee arthroplasty is the most common performed orthopedic surgery, with over 700,000 cases per year.¹ Overall, 21-25% underwent revision due to instability.² The stability of a TKA depends largely on soft tissue balance for proper alignment in flexion and extension. Soft tissue balancing in TKA depends on posterior femoral condylar osteophytes, which prevent full extension of the knee and increase tension posteriorly. Pre-operative soft tissue visualization is difficult and this leaves an increased chance for implant failure and revision surgery. Having a 3D model of the bone spur before and during the surgery may decrease risk of complication and enhance soft tissue modification for proper knee balancing post TKA.

Proposed Methods: We will 3D print a knee via MRI or CT that has been de-identified and provided by our mentor. The orthopedic surgeon will use this print out before or during surgery to address any concerns they have during the surgery in terms of proper balancing of the soft tissue of the knee. After approximately 10 uses within the OR, we will interview the surgeons and patients for their feedback. Patients undergoing total knee arthroplasty using the 3D printed knee will be matched with historic records of those who had undergone TKA. Rates of revision, patient satisfaction, and OR time will be compared across groups using chi-square and t-tests where appropriate.

Results: We anticipate reduced OR time, increased patient satisfaction and decreased rates of revision.

Conclusions: Our study demonstrates an initial use for 3D models as an aid or guide for total knee arthroplasty. Using a 3D model for TKA helps the surgeon visualize osteophytes and reduces the need for revision surgery. A larger study will need to be conducted in order to test the feasibility and practicality of 3D printing for surgery.

WORD COUNT: 300

-
1. ¹Gromov, K., Korchi, M., Thomsen, M. G., Husted, H., & Troelsen, A. (2014). What is the optimal alignment of the tibial and femoral components in knee arthroplasty?. *Acta orthopaedica*, 85(5), 480–487. doi:10.3109/17453674.2014.940573
 2. Kebbach, M., Grawe, R., Geier, A., Winter, E., Bergschmidt, P., Kluess, D., ... Bader, R. (2019). Effect of surgical parameters on the biomechanical behaviour of bicondylar total knee endoprostheses - A robot-assisted test method based on a musculoskeletal model. *Scientific reports*, 9(1), 14504. doi:10.1038/s41598-019-50399-3

DESIGN PROJECT Abstract – MAX 300 WORDS – DUE 12/18/19 ON EXAMSOFT