

Accepted Manuscript

Multicenter evaluation of the RAPIDEC[®] CARBA NP test for rapid screening of carbapenemase-producing *Enterobacteriaceae* and Gram-negative nonfermenters from clinical specimens

Marco Coppi, Alberto Antonelli, Tommaso Giani, Teresa Spanu, Flora Marzia Liotti, Carla Fontana, Walter Mirandola, Raffaele Gargiulo, Agostino Barozzi, Carola Mauri, Luigi Principe, Gian Maria Rossolini

PII: S0732-8893(17)30138-4
DOI: doi: [10.1016/j.diagmicrobio.2017.04.009](https://doi.org/10.1016/j.diagmicrobio.2017.04.009)
Reference: DMB 14338

To appear in: *Diagnostic Microbiology and Infectious Disease*

Received date: 17 January 2017
Revised date: 18 April 2017
Accepted date: 20 April 2017

Please cite this article as: Coppi Marco, Antonelli Alberto, Giani Tommaso, Spanu Teresa, Liotti Flora Marzia, Fontana Carla, Mirandola Walter, Gargiulo Raffaele, Barozzi Agostino, Mauri Carola, Principe Luigi, Rossolini Gian Maria, Multicenter evaluation of the RAPIDEC[®] CARBA NP test for rapid screening of carbapenemase-producing *Enterobacteriaceae* and Gram-negative nonfermenters from clinical specimens, *Diagnostic Microbiology and Infectious Disease* (2017), doi: [10.1016/j.diagmicrobio.2017.04.009](https://doi.org/10.1016/j.diagmicrobio.2017.04.009)

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.



Multicenter evaluation of the RAPIDEC® CARBA NP test for rapid screening of carbapenemase-producing *Enterobacteriaceae* and Gram-negative nonfermenters from clinical specimens

Marco Coppi¹, Alberto Antonelli¹, Tommaso Giani², Teresa Spanu³, Flora Marzia Liotti³, Carla Fontana^{4,5}, Walter Mirandola⁵, Raffaele Gargiulo⁶, Agostino Barozzi⁶, Carola Mauri⁷, Luigi Principe⁷ and Gian Maria Rossolini^{1,8#}

¹Department of Experimental and Clinical Medicine, University of Florence, Italy;

²Department of Medical Biotechnologies, University of Siena, Siena, Italy;

³Institute of Microbiology, Catholic University of the Sacred Heart, Agostino Gemelli Hospital, Rome, Italy;

⁴Department of Experimental Medicine and Surgery, Tor Vergata University of Rome, Rome, Italy;

⁵Clinical Microbiology Laboratories, Polyclinic of Tor Vergata Foundation;

⁶Provincial Laboratory of Clinical Microbiology, S. Agostino-Estense Hospital, Modena, Italy;

⁷Microbiology and Virology Unit, A. Manzoni Hospital, Lecco, Italy;

⁸Microbiology and Virology Unit, Florence Careggi University Hospital, Florence, Italy.

Keywords: CPE, blood culture, fast-track workflow, rapid screening.

Running title: RAPIDEC® CARBA NP on isolates from blood cultures

#Corresponding author:

Gian Maria Rossolini, MD

Department of Experimental and Clinical Medicine, University of Florence, Florence, Italy; SOD Microbiologia e Virologia, Dipartimento dei Servizi, Azienda Ospedaliera Universitaria Careggi, 50134 Firenze, Italy. Phone +39-0557949239; e-mail: gianmaria.rossolini@unifi.it

Abstract

The rapid diagnosis of carbapenemase-producing (CP) bacteria is essential for the management of therapy and infection control. In this study, RAPIDEC[®] CARBA NP (RCNP) was evaluated for the rapid screening of CP *Enterobacteriaceae*, *Acinetobacter baumannii* complex and *Pseudomonas aeruginosa* from clinical specimens collected at five Italian hospitals. Firstly, each site tested 20 well-characterized strains in a blinded fashion. Secondly, each center prospectively tested 25 isolates from blood cultures processed with a rapid workflow (6 hours after subculture) and 25 isolates from other specimens processed after an overnight culture. The presence of carbapenemases was confirmed by multiplex Real-Time-PCRs targeting carbapenemase genes. RCNP presented an overall sensitivity, specificity, positive predictive value, and negative predictive value of 70%, 94%, 82%, and 89%, respectively, with a higher performance in detection of CP *Enterobacteriaceae* and a poorer performance in detection of CP *A. baumannii* complex. With isolates from blood cultures, RCNP could significantly reduce the time required for identification of CP *Enterobacteriaceae* (less than 9 hours since the positivization of blood cultures).

1. Introduction

Antibiotic resistance is an issue of growing importance for public health, and involves a large variety of pathogenic bacteria responsible for healthcare-associated and community-acquired infections (Tang et al., 2014). Carbapenems are considered among the last resort antibiotics for treatment of resistant Gram-negatives (Papp-Wallace et al., 2011), but carbapenem-resistant strains of *Enterobacteriaceae* and Gram-negative nonfermenters are now spreading worldwide (Ruppé et al., 2015). The main mechanisms of resistance to carbapenems in Gram-negative pathogens are represented by the production of carbapenemases, reduction of outer membrane permeability mediated by the loss of porin function, and upregulation of efflux systems (Papp-Wallace et al., 2011). The spread of carbapenemase-producing (CP) strains of Gram-negative bacteria (GNB), including *Enterobacteriaceae*, *Pseudomonas aeruginosa* and *Acinetobacter* spp., is of notable

concern since these strains often carry additional resistance determinants and exhibit complex multidrug-resistant (MDR) phenotypes. Moreover, carbapenemase genes are usually associated with mobile genetic elements and their expression can be associated with higher-level carbapenem resistance (Kaye and Pogue, 2015; Rossolini et al., 2014; Ruppé et al., 2015).

Therefore, rapid identification of CP-GNB is important to implement infection control strategies that limit their spread in hospitals, and to the selection of appropriate antimicrobial therapy (Miriagou et al., 2010). Several approaches can be used for rapid identification of CP-GNB, including phenotypic and genotypic methods (Osei et al., 2015). Among the phenotypic methods, the RAPIDEC[®] CARBA NP test (bioMérieux, Marcy l'Etoile, France) is a commercial test for rapid screening of CP-GNB developed basing on the original CARBA NP colorimetric method (Nordmann et al., 2012). The RAPIDEC[®] CARBA NP test is easy to use and provides results in two hours, while being cheaper than molecular assays and able to detect also unknown carbapenemase genes. This test, based on the colorimetric detection of hydrolysis of imipenem using phenol red as indicator, has been previously validated or compared with other tests in several studies in which it was retrospectively applied on a collection of isolates previously characterized for the presence of carbapenemases (Poirel and Nordmann, 2015; Dortet et al., 2015; Hombach et al., 2015; Garg et al., 2015; Lifshitz et al., 2016; Kabir et al., 2016; Österblad et al., 2016; Aktaş et al., 2016), or prospectively applied on *Enterobacteriaceae* isolates (Noël et al., 2016).

In this work, we carried out a multicenter evaluation of the RAPIDEC[®] CARBA NP test, including a proficiency test with well-characterized strains, followed by further testing for the detection of CP-GNB among bacterial isolates prospectively collected from various clinical specimens. Moreover, a fast-track workflow for the detection of CP-GNB using the RAPIDEC[®] CARBA NP from blood cultures was implemented.

2. Materials and methods

2.1 Participating centers

Five laboratories associated with hospitals located in northern (Lecco and Modena) and central (Florence and Rome) Italy, representative of different Italian Regions, were involved in the study carried out from April to September 2015.

2.2 Proficiency test

A collection of 20 well-characterized strains, previously confirmed as CP (n=14) or carbapenem-resistant but carbapenemase-non-producers (CNP, n=4) or carbapenem-susceptible (n=2) (Table 1), was provided to each participating center in a blinded fashion. Each strain was cultured for 18-24 hours on blood agar and then tested with the RAPIDEC[®] CARBA NP test according to the Manufacturer's instructions.

2.3 Test on clinical isolates

A total of 250 (50 per participating center) consecutive, non-replicate clinical isolates of *Enterobacteriaceae* and Gram-negative nonfermenters (*P. aeruginosa* and *A. baumannii* complex) were tested with RAPIDEC[®] CARBA NP. Of them, 125 isolates (25 per participating center) were from blood cultures processed with a fast-track workflow, and 125 isolates (25 per participating center) were from other clinical specimens (surveillance specimens were not included). Positive blood cultures from BACTEC[™] (Becton Dickinson, Franklin Lakes, NJ, USA) or BacT/ALERT[®] (bioMérieux) systems were evaluated with Gram staining and plated onto blood agar plates (bioMérieux). The fast-track workflow foresaw that after six hours of incubation (35±2°C, 5% CO₂), bacterial isolates were identified by MALDI-TOF with the VITEK[®] MS system (bioMérieux) and, if they belonged to the target species, they were included in the study and tested with the RAPIDEC[®] CARBA NP test. Since the laboratories did not process positive blood cultures on a 24/7 schedule, only the blood cultures that became positive during the night or in the morning (until 12 a.m.) were processed with the fast-track workflow, by the staff in charge of the afternoon shift. Blood cultures yielding Gram-positive bacteria or mixed Gram-positive/Gram-negative bacteria and/or yeasts at Gram staining were excluded (Figure 1). Urine samples were cultured on chromID[®] CPS[®] Elite medium (bioMérieux) for 18-24 hours, while other materials were cultured on blood

agar (bioMérieux) for 18-24 hours. Bacterial isolates were identified by MALDI-TOF with the VITEK[®] MS system (bioMérieux) and, if they belonged to the target species, they were included in the study and tested with the RAPIDEC[®] CARBA NP (Figure 1).

2.4 RAPIDEC[®] CARBA NP test

The RAPIDEC[®] CARBA NP test was performed according to the Manufacturer's instructions, as follows. In case of isolated colonies from 18-24 hour-old cultures, several colonies were deposited in the dedicated well. For the 6-hour bacterial growth from blood cultures, the bacterial growth was transferred directly to the well of RAPIDEC[®] CARBA NP, until the indicated turbidity was reached. Samples presenting an insufficient bacterial growth were excluded. Strips were incubated at 35±2 °C for up to 120 minutes, and inspected at 30, 60 and 120 minutes. Results were interpreted by comparing the test well and the control well colors. A test was considered positive when a change of color of the well (from red to red-orange, orange or yellow) was observed.

2.5 Antimicrobial susceptibility testing

Antimicrobial susceptibility testing (AST) was performed using reference broth microdilution according to CLSI guidelines (CLSI, 2015) and results were interpreted according to EUCAST criteria v 6.0 (http://www.eucast.org/clinical_breakpoints/).

2.6 Molecular detection of carbapenemase genes

After performing the RAPIDEC[®] CARBA NP, each isolate was processed with three homebrew multiplex Real-Time-PCR mixes for the detection of the main carbapenemase genes, including *bla*_{KPC}, *bla*_{VIM}, *bla*_{NDM}, *bla*_{OXA-48-like} genes in *Enterobacteriaceae*, *bla*_{IMP}, *bla*_{VIM}, *bla*_{FIM-1}, *bla*_{GES} genes in *P. aeruginosa*, and *bla*_{OXA-23-like}, *bla*_{OXA-24-like}, *bla*_{OXA-58-like}, *ISAbal*+*bla*_{OXA-51-like} genes for *A. baumannii* complex. Primers and probes used in every reaction, and reaction conditions are described in Table 2. An internal control, consisting of phocine herpesvirus DNA (PhHV), and primers and probe targeting PhHV, was added in each reaction mix as a positive amplification control (Van Doornum et al., 2003).

2.7 Spectrophotometric assay

All meropenem non-susceptible isolates which tested negative with molecular assays for the detection of carbapenemase genes were further investigated by a spectrophotometric assay with crude extracts, using imipenem as substrate, as previously described (Lauretti et al., 1999), for detection of carbapenemase activity to exclude the presence of carbapenemase types not included in the molecular assay.

3. Results

3.1 Proficiency test with RAPIDEC[®] CARBA NP

Considering that RAPIDEC[®] CARBA NP is based on a colorimetric method and results are assigned by visual inspection, a proficiency test was initially performed with a collection of 20 well-characterized strains provided to each participating center in a blinded fashion, to evaluate the reproducibility of interpretation of results obtained at different centers.

All the CNP (n=4) and the susceptible (n=2) strains were correctly identified as carbapenemase-negative with the RAPIDEC[®] CARBA NP test by all laboratories, while four of the 14 CP strains were not reported as carbapenemase-positive by one or more laboratories. In particular, the OXA-58-positive *A. baumannii* complex strain was not detected as CP by two laboratories, while the FIM-1-positive and the IMP-13-positive *P. aeruginosa* strains, and the OXA-372-positive *Citrobacter freundii* strain, were not detected as CP by one laboratory each (Table 1).

The overall good results of the proficiency test (specificity for detection of CP, 100%; sensitivity for detection of CP, 93%) prompted us to proceed with the analysis of isolates from clinical samples in each center.

3.2 Performance of the RAPIDEC[®] CARBA NP test on isolates from clinical specimens other than blood cultures

Among 125 isolates from various materials (mainly urine and respiratory specimens), 31 were confirmed as CP by molecular methods. These isolates included 10 KPC-producing *K. pneumoniae*,

one OXA-48-producing *K. pneumoniae*, one KPC- and VIM-coproducing *K. pneumoniae*, and 19 class D carbapenemase-producing *A. baumannii* complex (n=16 OXA-23-like, n=2 OXA-24-like, n=1 OXA-51-like overexpressed by an *ISAbal* inserted upstream). Of these, the RAPIDEC[®] CARBA NP test correctly detected the 12 CP *Enterobacteriaceae*, but failed to identify 11 of the 19 CP *A. baumannii* complex (including 9 positive for a *bla*_{OXA-23-like} gene, one positive for a *bla*_{OXA-24-like} gene, and one carrying a *bla*_{OXA-51-like} gene preceded by an *ISAbal* insertion sequence). Moreover, the RAPIDEC[®] CARBA NP test was positive with one *K. pneumoniae* and five *P. aeruginosa* for which the molecular tests and the spectrophotometric assay had not identified carbapenemases (Table 3). Consequently, overall sensitivity, specificity, positive predictive values (PPV), and negative predictive values (NPP) were found to be 65%, 94%, 77%, and 89%, respectively. If *A. baumannii* complex strains were not included in the study, the sensitivity, specificity, PPV and NPP of RAPIDEC[®] CARBA NP would be 100%, 93%, 67%, 100%, respectively. Considering only the 61 *Enterobacteriaceae* isolates, the test yielded an even better performance (100%, 98%, 92%, 100%, respectively) (Table 4).

3.3 Performance of the RAPIDEC[®] CARBA NP test with positive blood cultures in a fast-track workflow

Of the 125 isolates from blood cultures tested with RAPIDEC[®] CARBA NP, 122 gave interpretable results, while three isolates (one *P. aeruginosa*, one *A. baumannii* complex and one *Enterobacter cloacae*) yielded insufficient growth to perform the test at 6 hours after subculture. Among the 122 evaluable isolates, 36 were confirmed as CP by molecular methods, including 18 KPC-producing *K. pneumoniae*, two OXA-48-producing *K. pneumoniae*, one VIM-producing *K. pneumoniae*, three VIM-producing *P. aeruginosa* and 12 class D carbapenemase-producing *A. baumannii* complex (n=10, OXA-23-like; n=1, OXA-24-like; n=1, OXA-23-like and OXA-24-like co-producer). Of them, RAPIDEC[®] CARBA NP assay correctly identified 27 isolates as CP, but missed nine *A. baumannii* complex carrying *bla*_{OXA-23-like} and/or *bla*_{OXA-24-like} genes (Table 3). RAPIDEC[®] CARBA NP correctly categorized as CNP 82 out of 86 isolates from blood cultures, while false-positive

results were obtained with one *K. pneumoniae* and three *P. aeruginosa* isolates for which no carbapenemase genes nor carbapenemase activity were detected using Real-Time-PCRs and spectrophotometric activity, respectively. According to these results, the RAPIDEC[®] CARBA NP showed a sensitivity of 75%, a specificity of 95%, a PPV of 87%, and a NPV of 90%. However, excluding the *A. baumannii* complex isolates, sensitivity, specificity, PPV and NPV were 100%, 95%, 86%, 100%, respectively, and these percentages further increased when considering only the 81 *Enterobacteriaceae* isolates (100%, 98%, 95%, 100%, respectively) (Table 4).

3.4 Overall performance of the RAPIDEC[®] CARBA NP test with clinical isolates

Overall, considering all types of clinical specimens, the RAPIDEC[®] CARBA NP test yielded sensitivity, specificity, PPV and NPP of 70%, 94%, 82%, and 89%, respectively. The best results were observed with *Enterobacteriaceae*, as described in Table 4.

No significant differences were observed among the performance of the RAPIDEC[®] CARBA NP test carried out at the five centers.

4. Discussion

The global spread of CP-GNB represents a major public health challenge. Clinical Microbiology laboratories are increasingly asked for rapid detection of CP strains for infection control and antimicrobial stewardship purposes.

The RAPIDEC[®] CARBA NP test is a simple phenotypic test that does not require any specific technical skills or expensive equipment (Poirel and Nordmann, 2015; Garg et al., 2015), which allows rapid detection of CP strains in a timeframe comparable to that of molecular tests (i. e. 1-2 hrs) but at a substantially lower cost.

In this multicenter work, involving five different Italian laboratories, we evaluated the performance of RAPIDEC[®] CARBA NP in a proficiency test carried out by each laboratory with a collection of 20 well-characterized carbapenem-resistant strains representative of different Gram-negative species and resistance mechanisms, and then in a field test with 50 clinical isolates from each

laboratory.

Results revealed an overall high specificity of RAPIDEC[®] CARBA NP for detection of CP strains, similar to that previously reported by Poirel and Nordmann (2015) (96%), Garg et al. (2015) (96.2%) and Kabir et al. (2016) (98.5%). The false-positive results, observed with *K. pneumoniae* and *P. aeruginosa* isolates, and reported also by other Authors (Poirel and Nordmann, 2015; Lifshitz et al., 2016; Österblad et al., 2016), could be attributable to a reduced stability of the imipenem substrate used in the test toward strains producing enzymes that are not true carbapenemases but have some weak carbapenemase activity (e. g. AmpCs, CMY, CTX-M-type producers), and possibly also to the use of the inoculum recommended in the Manufacturer's instructions. Indeed, Dortet et al. (2015) identified a critical impact of the bacterial inoculum in the performance of the test, and recommended the use of a much higher inoculum to avoid false positive results.

On the other hand, the test frequently failed to detect *A. baumannii* complex strains producing class D β -lactamases, a problem reported also by other authors (Poirel and Nordmann, 2015; Kabir et al., 2016) and probably due to the overall weak carbapenemase activity of class D carbapenemases (Queenan and Bush, 2007). A lower overall sensitivity (70%) was detected in this work compared to previously reported studies by Garg et al. (2015) (92.6%), Kabir et al. (2016) (97.8%), and Poirel and Nordmann (2015) (96%). This difference could be partially ascribed to the higher percentage of class D producing *A. baumannii* complex (12.5%) tested in our study (Garg et al. [2015] 0%, Kabir et al. [2016] 4.7%, and Poirel and Nordmann [2015] 8.5%). Also Noël et al. (2016) showed that RAPIDEC[®] CARBA NP performed poorly for the detection of class D carbapenemase-producing *Acinetobacter* spp. isolates (sensitivity 36.4%, specificity 75%), leading us to conclude that this test should not be used with *A. baumannii* complex isolates in its present format. A much higher bacterial inoculum, compared to the Manufacturer's instructions, could increase the sensitivity of the tests (Lifshitz et al., 2016). Indeed, Dortet et al. (2015) recommended to perform the test using a standardized inoculum (a full 10 μ l loop), which is critical for test reliability.

A limitation of the prospective evaluation of RAPIDEC[®] CARBA NP carried out in this work is

represented by the relatively low number of isolates producing some types of carbapenemases (e. g. OXA-48 or VIM), and of carbapenemase-producing *P. aeruginosa* isolates, which reflected the local epidemiology of infections.

Considering *Enterobacteriaceae* isolates only, the values of sensitivity and specificity were high and comparable with other works (Dortet et al., 2015; Hombach et al., 2015; Lifshitz et al., 2016), showing that RAPIDEC[®] CARBA NP test could be useful also for detecting CP *Enterobacteriaceae* from positive blood cultures processed with a fast-workflow approach. With this approach, RAPIDEC[®] CARBA NP can be used to reduce the time required for identification of CP *Enterobacteriaceae* to less than 9 hours since positivization of blood cultures (from at least 24 to 48 hours of routine methods) (Morgenthaler and Kostrzewa, 2015). The rapid identification of CP Gram-negatives from blood cultures can be of remarkable importance to antimicrobial stewardship (Barlem et al., 2016). It should be noted, however, that the possibility of using a fast workflow is dependent on the laboratory schedule. For instance, in the laboratories participating in this work, which do not process positive blood cultures on a 24/7 schedule, the fast workflow could only be performed with blood cultures that were found to be positive or became positive in the morning. When the RAPIDEC[®] CARBA NP is used with positive blood cultures in the rapid workflow, another limitation could be represented by an insufficient bacterial growth for inoculum at 6 hours. However, in our experience this occurred only with a small number of cases (3 of 122, 2.4%).

A limitation of the RAPIDEC[®] CARBA NP test is that it cannot discriminate the type of carbapenemase produced. This information is important for antimicrobial stewardship since the new antibiotics active against CP-GNB that are entering clinical practice (e. g ceftazidime-avibactam) may not cover all types of CP strains. In this perspective, the RAPIDEC[®] CARBA NP could have a role as a screening test for excluding carbapenemase production and selecting the most suitable candidates for characterization of the carbapenemase type by molecular platforms or immunoenzymatic assays (Banerjee et al., 2015; Raich and Powell, 2015; Meunier et al., 2016).

Acknowledgements

Results of this work were partially presented at the 44th National Congress of the Italian Society of Microbiology (SIM). September 2016, Pisa. Poster P32.

This work was partially funded by a bioMérieux grant.

References

- Aktaş E, Malkoçoğlu G, Otlu B, Çopur Çiçek A, Külah C, Cömert F et al. Evaluation of the Carbapenem Inactivation Method for Detection of Carbapenemase-Producing Gram-Negative Bacteria in Comparison with the RAPIDEC CARBA NP. *Microb Drug Resist*. August 2016, ahead of print. doi:10.1089/mdr.2016.0092.
- Antonelli A, Di Palo DM, Galano A, Becciani S, Montagnani C, Pecile P et al. Intestinal carriage of *Shewanella xiamenensis* simulating carriage of OXA-48-producing *Enterobacteriaceae*. *Diagn Microbiol Infect Dis* 2015a; 82:1–3. doi: 10.1016/j.diagmicrobio.2015.02.008.
- Antonelli A, D'Andrea MM, Vaggelli G, Doçquier JD, Rossolini GM. OXA-372, a novel carbapenem-hydrolysing class D β -lactamase from a *Citrobacter freundii* isolated from a hospital wastewater plant. *J Antimicrob Chemother* 2015b; 70 (10):2749–56. doi: 10.1093/jac/dkv181.
- Antonelli A, D'Andrea MM, Di Pilato V, Viaggi B, Torricelli F, Rossolini GM. Characterization of a Novel Putative Xer-Dependent Integrative Mobile Element Carrying the *bla*_(NMC-A) Carbapenemase Gene, Inserted into the Chromosome of Members of the *Enterobacter cloacae* complex. *Antimicrob Agents Chemother* 2015c; 59(10):6620–4. doi: 10.1128/AAC.01452-15.
- Antonelli A, Arena F, Giani T, Colavecchio OL, Valeva SV, Paule S et al. Performance of the BD MAX™ instrument with Check-Direct CPE real-time PCR for the detection of carbapenemase genes from rectal swabs, in a setting with endemic dissemination of carbapenemase-producing *Enterobacteriaceae*. *Diagn Microbiol Infect Dis* 2016; 86(1):30–4. doi: 10.1016/j.diagmicrobio.2016.06.002.
- Arena F, Giani T, Becucci E, Conte V, Zanelli G, D'Andrea MM et al. Large oligoclonal outbreak due to *Klebsiella pneumoniae* ST14 and ST26 producing the FOX-7 AmpC β -lactamase in a neonatal intensive care unit. *J Clin Microbiol* 2013; 51(12):4067–72. doi: 10.1128/JCM.01982-13.
- Banerjee R, Teng CB, Cunningham SA, Ihde SM, Steckelberg JM, Moriarty JP et al. Randomized trial of rapid multiplex polymerase chain reaction-based blood culture identification and susceptibility testing. *Clin Infect Dis* 2015; 61(7): 1071–1080. doi: 10.1093/cid/civ447.
- Barlam TF, Cosgrove SE, Abbo LM, MacDougall C, Schuetz AN, Septimus EJ et al. Implementing an Antibiotic Stewardship Program: Guidelines by the Infectious Diseases Society of America and the Society for Healthcare Epidemiology of America. *Clin Infect Dis* 2016; 62(10):e51–77. doi: 10.1093/cid/ciw118.

- CLSI. Methods for Dilution Antimicrobial Susceptibility Tests for Bacteria That Grow Aerobically; Approved Standard—Tenth Edition. CLSI document M07-A10. 2015. Wayne, PA: Clinical and Laboratory Standards Institute.
- Corvec S, Poirel L, Naas T, Drugeon H, Nordmann P. Genetics and expression of the carbapenem-hydrolyzing oxacillinase gene *bla*_{OXA-23} in *Acinetobacter baumannii*. *Antimicrob Agents Chemother* 2007; 51(4):1530–3.
- D'Andrea MM, Giani T, D'Arezzo S, Capone A, Petrosillo N, Visca P et al. Characterization of pABVA01, a plasmid encoding the OXA-24 carbapenemase from Italian isolates of *Acinetobacter baumannii*. *Antimicrob Agents Chemother* 2009; 53(8):3528–33. doi: 10.1128/AAC.00178-09
- D'Andrea MM, Venturelli C, Giani T, Arena F, Conte V, Bresciani P et al. Persistent Carriage and Infection by Multidrug-Resistant *Escherichia coli* ST405 Producing NDM-1 Carbapenemase: Report on the First Italian Cases. *J Clin Microbiol* 2011; 49(7):2755–8. doi: 10.1128/JCM.00016-11
- Dortet L, Agathine A, Naas T, Cuzon G, Poirel L, Nordmann P. Evaluation of the RAPIDEC® CARBA NP, the Rapid CARB Screen® and the Carba NP test for biochemical detection of carbapenemase-producing *Enterobacteriaceae*. *J Antimicrob Chemother* 2015; 70(11):3014–22. doi: 10.1093/jac/dkv213.
- Garg A, Garg J, Upadhyay GC, Agarwal A, Bhattacharjee A. Evaluation of the Rapidec Carba NP Test Kit for Detection of Carbapenemase-Producing Gram-Negative Bacteria. *Antimicrob Agents Chemother* 2015; 59(12):7870–2. doi: 10.1128/AAC.01783-15.
- Giani T, D'Andrea MM, Pecile P, Borgianni L, Nicoletti P, Tonelli F et al. Emergence in Italy of *Klebsiella pneumoniae* sequence type 258 producing KPC-3 Carbapenemase. *J Clin Microbiol* 2009; 47(11):3793–4. doi: 10.1128/JCM.01773-09.
- Giani T, Conte V, Di Pilato V, Aschbacher R, Weber C, Larcher C et al. *Escherichia coli* from Italy producing OXA-48 carbapenemase encoded by a novel Tn1999 transposon derivative. *Antimicrob Agents Chemother* 2012; 56(4):2211–3. doi: 10.1128/AAC.00035-12.
- Hindiyeh M, Smollen G, Grossman Z, Ram D, Davidson Y, Mileguir F et al. Rapid detection of *bla*_{KPC} carbapenemase genes by real-time PCR. *J Clin Microbiol* 2008; 46:2879–2883. doi: 10.1128/JCM.00661-08.
- Hombach M, von Gunten B, Castelberg C, Bloemberg GV. Evaluation of the Rapidec Carba NP Test for Detection of Carbapenemases in *Enterobacteriaceae*. *J Clin Microbiol* 2015; 53(12):3828–33. doi: 10.1128/JCM.02327-15
- Kabir MH, Meunier D, Hopkins KL, Giske CG, Woodford N. A two-centre evaluation of RAPIDEC® CARBA NP for carbapenemase detection in *Enterobacteriaceae*, *Pseudomonas aeruginosa* and *Acinetobacter* spp. *J Antimicrob Chemother* 2016; 71(5):1213–6. doi: 10.1093/jac/dkv468.
- Kaye KS, Pogue JM. Infections Caused by Resistant Gram-Negative Bacteria: Epidemiology and Management. *Pharmacotherapy* 2015; 35(10):949–62. doi: 10.1002/phar.1636
- Lauretti L, Riccio ML, Mazzariol A, Cornaglia G, Amicosante G, Fontana R et al. Cloning and

- Characterization of *bla*_{VIM}, a New Integron-Borne Metallo- β -Lactamase Gene from a *Pseudomonas aeruginosa* Clinical Isolate. *Antimicrob Agents and Chemother* 1999; 43(7):1584–90
- Lifshitz Z, Adler A, Carmeli Y. Comparative Study of a Novel Biochemical Assay, the Rapidec Carba NP Test, for Detecting Carbapenemase-Producing *Enterobacteriaceae*. *J Clin Microbiol* 2016; 54(2):453–6. doi: 10.1128/JCM.02626-15.
- Luzzaro F, Docquier JD, Colinson C, Endimiani A, Lombardi G, Amicosante G et al. Emergence in *Klebsiella pneumoniae* and *Enterobacter cloacae* clinical isolates of the VIM-4 metallo-beta-lactamase encoded by a conjugative plasmid. *Antimicrob Agents Chemother* 2004; 48(2):648–50.
- Meunier D, Vickers A, Pike R, Hill RL, Woodford N, Hopkins KL. Evaluation of the K-SeT R.E.S.I.S.T. immunochromatographic assay for the rapid detection of KPC and OXA-48-like carbapenemases. *J Antimicrob Chemother* 2016; 71(8):2357–9. doi: 10.1093/jac/dkw113.
- Miriagou V, Cornaglia G, Edelstein M, Galani I, Giske CG, Gniadkowski M et al. Acquired carbapenemases in Gram-negative bacterial pathogens: detection and surveillance issues. *Clin Microbiol Infect* 2010; 16(2):112–22. doi: 10.1111/j.1469-0691.2009.03116.x.
- Morgenthaler NG, Kostrzewa M. Rapid identification of pathogens in positive blood culture of patients with sepsis: review and meta-analysis of the performance of the sepsi typer kit. *Int J Microbiol* 2015; 2015:827416. doi: 10.1155/2015/827416.
- Noël A, Huang TD, Berhin C, Hoebek M, Bouchahrouf W, Yunus S et al. Comparative evaluation of four phenotypic tests for the detection of carbapenemase-producing Gram-negative bacteria. *J Clin Microbiol* 2017; 55(2):510–518. doi: 10.1128/JCM.01853-16.
- Nordmann P, Poirel L, Dortet L. Rapid detection of carbapenemase-producing *Enterobacteriaceae*. *Emerg Infect Dis* 2012; 18:1503–1507. doi: 10.3201/eid1809.120355.
- Ong DC, Koh TH, Syahidah N, Krishnan P, Tan TY. Rapid detection of the *bla*_{NDM-1} gene by real-time PCR. *J Antimicrob Chemother* 2011; 66:1647–1649. doi: 10.1093/jac/dkr184.
- Osei Sekyere J, Govinden U, Essack SY. Review of established and innovative detection methods for carbapenemase-producing Gram-negative bacteria. *J Appl Microbiol* 2015; 119(5):1219–33. doi: 10.1111/jam.12918.
- Österblad M, Lindholm L, Jalava J. Evaluation of two commercial carbapenemase gene assays, the Rapidec Carba NP test and the in-house Rapid Carba NP test, on bacterial cultures. *J Antimicrob Chemother* 2016; 71(7):2057–9. doi: 10.1093/jac/dkw077.
- Papp-Wallace KM, Endimiani A, Taracila MA, Bonomo RA. Carbapenems: past, present, and future. *Antimicrob Agents Chemother* 2011; 55(11):4943–60. doi: 10.1128/AAC.00296-11.
- Poirel L, Weldhagen GF, Naas T, De Champs C, Dove MG, Nordmann P. GES-2, a class A beta-lactamase from *Pseudomonas aeruginosa* with increased hydrolysis of imipenem. *Antimicrob Agents Chemother* 2001; 45(9):2598–603.
- Poirel L, Nordmann P. Rapidec Carba NP Test for Rapid Detection of Carbapenemase Producers. *J Clin Microbiol* 2015; 53(9):3003–8. doi: 10.1128/JCM.00977-15.

- Pollini S, Maradei S, Pecile P, Olivo G, Luzzaro F, Docquier JD et al. FIM-1, a new acquired metallo- β -lactamase from a *Pseudomonas aeruginosa* clinical isolate from Italy. *Antimicrob Agents Chemother* 2013; 57(1):410–6. doi: 10.1128/AAC.01953-12
- Queenan AM, Bush K. Carbapenemases: the Versatile β -Lactamases. *Clin Microbiol Rev* 2007; 20(3): 440–458. doi: 10.1128/CMR.00001-07
- Raich T, Powell S. Identification of bacterial and fungal pathogens from positive food culture bottles: a microarray-based approach. *Methods Mol Biol* 2015; 1237:73–90. doi: 10.1007/978-1-4939-1776-1_8.
- Rossolini GM, Arena F, Pecile P, Pollini S. Update on the antibiotic resistance crisis. *Curr Opin Pharmacol* 2014; 18:56–60. doi: 10.1016/j.coph.2014.09.006
- Ruppé É, Woerther PL, Barbier F. Mechanisms of antimicrobial resistance in Gram-negative bacilli. *Ann Intensive Care* 2015; 5(1):61. doi: 10.1186/s13613-015-0061-0.
- Tang SS, Apisarnthanarak A, Hsu LY. Mechanisms of β -lactam antimicrobial resistance and epidemiology of major community- and healthcare-associated multidrug-resistant bacteria. *Adv Drug Deliv Rev* 2014; 78:3–13. doi: 10.1016/j.addr.2014.08.003
- Van Doornum, GJ, Guldemeester J, Osterhaus AD, Niesters HG. Diagnosing herpesvirus infections by real-time amplification and rapid culture. *J Clin Microbiol* 2003; 41:576–580.

Figure 1. Workflow for the detection of carbapenemase producers in different clinical isolates. Significant pathogen indicates the presence of a suspect pathogen in a significant count (if applicable for the specimen type), present as a pure culture or as a mixed population with commensals derived from sampling the nonsterile site, considered to be clinically significant and subjected to further identification, AST and reporting.




Table 1

Gram-negative strains selected for the evaluation of RAPIDEC® CARBA NP proficiency with the test results obtained in each site.

Strain	Species	Principal Acquired β -lactamase	Reference	MIC Meropenem ($\mu\text{g/mL}$)	RAPIDEC® results (center)					
					Expected	1	2	3	4	5
6-419	<i>Escherichia coli</i>	No-one	-	0.5	-	-	-	-	-	-
23-1786	<i>Enterobacter ludwigii</i>	NMC-A	Antonelli et al., 2015c	32	+	+	+	+	+	+
7-556	<i>K. pneumoniae</i>	NDM-1	-	32	+	+	+	+	+	+
22-1706	<i>E. coli</i>	NDM-5	-	>32	+	+	+	+	+	+
7728	<i>P. aeruginosa</i>	IMP-13	-	4	+	+	-	+*	+	+
ATCC 25922	<i>E. coli</i>	no-one	-	0.5	-	-	-	-	-	-
47-3752	<i>E. cloacae</i> complex	IMI-2	-	>32	+	+	+	+	+*	+
CVB-1	<i>E. coli</i>	NDM-1	D'Andrea et al., 2011	32	+	+	+	+	+	+
ECBZ-1	<i>E. coli</i>	OXA-48	Giani et al., 2012	1	+	+	+	+	+*	+
FIPP-1	<i>K. pneumoniae</i>	KPC-3	Giani et al., 2009	>32	+	+	+	+	+	+
VA-417/02	<i>E. cloacae</i> complex	VIM-4	Luzzaro et al., 2004	32	+	+	+	+	+	+
FI-14/157	<i>P. aeruginosa</i>	FIM-1	Pollini et al., 2013	>32	+	+	-	+	+	+
Cfr-FI-07	<i>C. freundii</i>	OXA-372	Antonelli et al., 2015b	16	+	+	+	+	-	+
45A02	<i>K. pneumoniae</i>	FOX-7 + porin deficiency	Arena et al., 2013	4	-	-	-	-	-	-
NV132	<i>A. baumannii</i> complex	OXA-58	-	8	+	+*	-	+	+*	-
8-27	<i>K. pneumoniae</i>	CTX-M-1-like + OMP deficient	-	2	-	-	-	-	-	-
10-52	<i>K. pneumoniae</i>	CTX-M-1-like + OMP deficient	-	2	-	-	-	-	-	-
VA-416/02	<i>K. pneumoniae</i>	VIM-4	Luzzaro et al., 2004	32	+	+	+	+	+	+
GW1	<i>P. aeruginosa</i>	GES-2	Poirel et al., 2001	16	-	-	-	-	-	-
PIEcl	<i>E. cloacae</i> complex	VIM-1	-	>32	+	+	+	+	+	+

* Borderline positive: there was correctly a change of color in the test well, but it was not clear as described in manufacturer's instructions.

Table 2
Primers and probes used in this study for three different multiplex Real Time PCR.

Investigated bacteria	Target	Primer name	Sequence (5'-3')	Reference	Positive control	Concentration of use in reaction mix (nM)
<i>Enterobacteriaceae</i> ^a	<i>bla</i> _{OXA-48-like} genes	OXA-48-like-rt-F	GTAGCAAAGGAATGGCAAGAAA	Antonelli et al., 2015a	<i>E. coli</i> ECBZ-1 (OXA-48) (Giani et al., 2012)	500
		OXA-48-like-rt-R	GATGCGGGTAAAAATGCTTG	Antonelli et al., 2015a		500
		OXA-48-like-rt-P	HEX-CTCTGGAATGAGAATAAGCAGCAAGG-BHQ-1	Antonelli et al., 2015a		125
	<i>bla</i> _{KPC} genes	kpc-hind-fwd	GATACCACGTTCCGCTCGG	Hindiyeh et al., 2008	<i>K. pneumoniae</i> FIPP-1 (KPC-3) (Giani et al., 2009)	500
		kpc-hind-rev	GCAGGTTCCGTTTGTCTC	Hindiyeh et al., 2008		500
		kpc-hind-tq	FAM-AGCGGCAGCAGTTTGTGATTG-BHQ-1	Hindiyeh et al., 2008		125
	<i>bla</i> _{VIM} genes ^b	VIM-rt-fwd	TGGTCTCATTTGCCGTGATG	Antonelli et al., 2016	<i>K. pneumoniae</i> VA-416/02 (VIM-4), (Luzzaro et al., 2004)	500
		VIM-rt-rev	CATGAAAGTGCCTGGAGA	Antonelli et al., 2016		500
		VIM-rt-tq	ROX-AAGCAAATGGACTTCCCGTAACGC-BHQ-2	Antonelli et al., 2016		125
	<i>bla</i> _{NDM} genes	<i>bla</i> _{NDM1_F}	CGCAACACAGCCTGACTTT	Ong et al., 2008	<i>E. coli</i> CVB-1 (NDM-1) (D'Andrea et al., 2011)	500
		<i>bla</i> _{NDM1_R}	TCGATCCCAACGGTGATATT	Ong et al., 2008		500
		<i>bla</i> _{NDM1_P}	CY5-CAACTTGGCCCGCTCAAGGTATT-BHQ-3	Ong et al., 2008		125
<i>P. aeruginosa</i> ^c	<i>bla</i> _{FIM-1} gene	FIM-rt-F	CGCCTTAACACCCGTCGTGA	This study	<i>P. aeruginosa</i> FI-14/157 (FIM-1) (Pollini et al., 2013)	500
		FIM-rt-R	GTCICCTTTTCAACGATTAGCC	This study		500
		FIM-rt-P	HEX- CTGGCGTACAAGCGGCTCAACCCAA- BHQ-1	This study		125
	<i>bla</i> _{GES} genes	GES-rt-F	AGAATTGACTCAGGCACCGAG	This study	<i>P. aeruginosa</i> GW1 (GES-2) (Poirel et al., 2001)	500
		GES-rt-R	GTTAGTAGCCCATTTGTCCG	This study		500
		GES-rt-P	CY5- GAACCGTCATGTGTCCCGATGCTAG-BHQ-3	This study		125
	<i>bla</i> _{IMP} genes	IMP-rt-F	GANGCYTAYHTRATWGAYACTCCA	This study	<i>P. aeruginosa</i> FI-5/7728 (IMP-13), unpublished	2000
		IMP-rt-R	GRRATDGAYYGAGARTTAAGCCA	This study		2000
		IMP-rt-P	FAM- ATTCNCSCYGHRCRTRCRCYATGRAAATG- BHQ-1	This study		250
<i>A. baumannii</i> complex ^c	<i>bla</i> _{OXA-23-like} genes	oxa-23-like-rt-F	GATTGTCAAGGACATAATCAGGTG	This study	<i>A. baumannii</i> Ab13 (OXA-23) (Coverc et al., 2007)	500
		oxa-23-like-rt-R	GGTTCCTCCAAATCCGATCAGGG	This study		500
		oxa-23-like-rt-P	FAM- AGGCTGGCACATATTCGTATTGCGG- BHQ-3	This study		125
	<i>bla</i> _{OXA-24-like} genes	oxa-24-like-rt-F	CTTCCTATHYTCAGCATTTCTATTCTAG	This study	<i>A. baumannii</i> VA-566/00 (OXA-40/24), (D'Andrea et al., 2009)	1000
		oxa-24-like-rt-R	ATCTTAAATGTTGAYGCAGGGAC	This study		1000
		oxa-24-like-rt-P	HEX- GCTATTTTGTGAAGCTCAACACARGGT- BHQ-1	This study		250
	<i>bla</i> _{OXA-58-like} genes	oxa-58-like-rt-F	AAAGCATGGGACAAAGATTTTAC	This study	<i>A. baumannii</i> NV132 (OXA-58), unpublished	500
		oxa-58-like-rt-R	CAAACCTTACTTCTGTATAGGTGT	This study		500
		oxa-58-like-rt-P	ROX- CAGTGCCTGATATCAAGAATTGGCAC- BHQ-2	This study		125
	<i>ISAbal</i> + <i>bla</i> _{OXA-51-like} genes	ISABA1-oxa-51-rt-F	ATAATCACAAGCATGATGAGCG	This study	<i>A. baumannii</i> 696/03 (OXA-58), unpublished	500
		ISABA1-oxa-51-rt-R	GTGARCAGGCTGAAATARRAATAG	This study		500
		ISABA1-oxa-51-rt-P	CY5- ATGAACATTAAGCACTCTTACTTATAACAAG- BHQ-3	This study		125
Everyone	PhHV (internal control)	PhHV-267s	GGG CGA ATC ACA GAT TGA ATC	Van Doornum et al., 2003	PhHV DNA cloned in pGEM-T-easy <i>E. coli</i> DH5α	500
		PhHV-337as	GCG GTT CCA AAC GTA CCA A	Van Doornum et al., 2003		500
		PhHV-305tq	Cy5.5 -TTTTATGTGTCCGCCACCATCTGGATC-BHQ-3	Van Doornum et al., 2003		125

^aThe amplification program consisted of 35 two-step cycles of 15 s at 95°C and 60 s at 60°C.

^b*bla*_{VIM} genes were also target of the multiplex Real Time PCR used for *P. aeruginosa* isolates.

^cThe amplification program consisted of 35 three-step cycles of 15 s at 95°C, 30 s at 50°C and 30 s at 60°C.

Table 3

Comparison of the results obtained from RAPIDEC® CARBA NP test with the molecular detection of carbapenemase genes. Discrepant RAPIDEC® CARBA NP results are underlined.

Origin of Isolates	Species	No. of isolates	MIC MEM ($\mu\text{g/ml}$)	Carbapenemase genes	RAPIDEC® Result
various materials	<i>Enterobacteriaceae</i>	10	16 to >128	<i>bla</i> _{KPC}	+
		1	32	<i>bla</i> _{OXA-48-like}	+
		1	16	<i>bla</i> _{KPC} + <i>bla</i> _{VIM}	+
		48	≤0.03 to 4	No one	–
		1	0.25	<u>No one</u>	+
	<i>P. aeruginosa</i>	4	16 to >128	<u>No one</u>	+
		35	≤0.03 to 64	No one	–
		1	2	<u>No one</u>	+
		7	64 to >128	<i>bla</i> _{OXA-23-like}	+
		9	8 to 128	<i>bla</i> _{OXA-23-like}	–
	<i>A. baumannii</i> complex	1	64	<i>bla</i> _{OXA-24-like}	+
		1	2	<i>bla</i> _{OXA-24-like}	–
		1	2	<u><i>ISAbal</i>+<i>bla</i>_{OXA-51-like}</u>	–
		5	0.25 to 8	No one	–
		blood cultures	<i>Enterobacteriaceae</i>	18	8 to >128
2	1 to 4			<i>bla</i> _{OXA-48-like}	+
1	4			<i>bla</i> _{VIM}	+
59	≤0.03 to 2			No one	–
1	≤0.03			<u>No one</u>	+
<i>P. aeruginosa</i>	1		≤0.03	No one	^a
	3		8 to >128	<i>bla</i> _{VIM}	+
	1		>128	<i>bla</i> _{VIM}	^a
	1		>128	<u>No one</u>	+
	20		0.12 to >128	No one	–
<i>A. baumannii</i> complex	2		0.12; 0.5	<u>No one</u>	+
	1		128	<i>bla</i> _{OXA-23-like} and <i>bla</i> _{OXA-24-like}	–
	7		16 to >128	<i>bla</i> _{OXA-23-like}	–
	3		16 to 64	<i>bla</i> _{OXA-23-like}	+
	1		>128	<i>bla</i> _{OXA-24-like}	–
	1	>128	<i>bla</i> _{OXA-23-like}	^a	
	3	0.25	No one	–	

^a insufficient biomass 6 hours after subculture.

Table 4
Performance of RAPIDEC® CARBA NP test with clinical isolates.

Sources	Species/Family	Sensitivity %	Specificity %	PPV %	NPV %
All	All (N=247)	70	94	82	89
	Excluding <i>A. baumannii</i> complex (N=208)	100	94	78	100
	<i>Enterobacteriaceae</i> (N=142)	100	98	94	100
Blood cultures	All (N=122)	75	95	87	90
	Excluding <i>A. baumannii</i> complex (N=107)	100	95	86	100
	<i>Enterobacteriaceae</i> (N=81)	100	98	95	100
Other materials	All (N=125)	65	94	77	89
	Excluding <i>A. baumannii</i> complex (N=101)	100	93	67	100
	<i>Enterobacteriaceae</i> (N=61)	100	98	92	100

Highlights

- RAPIDEC[®] CARBA NP could detect carbapenemase producers from positive blood cultures, after 6 hours of subculture
- The test showed low sensitivity with carbapenemase-producing *A. baumannii*
- Excluding *A. baumannii* the test showed high sensitivity (100%) and specificity (94%)

ACCEPTED MANUSCRIPT