

# Minerva Anestesiologica

## EDIZIONI MINERVA MEDICA

### ARTICLE ONLINE FIRST

This provisional PDF corresponds to the article as it appeared upon acceptance.

A copyedited and fully formatted version will be made available soon.

The final version may contain major or minor changes.

## Use of the CompuFlo® system to identify the epidural space in obstetric-gynecological area. A single-center retrospective study.

Alessandro DI FILIPPO, Ubaldo BITOSSO, Viviana MARCELLINO, Vittorio LIMATOLA, Martina SICURANI, Tommaso BORRACCI, Massimo MICAGLIO

*Minerva Anestesiologica* 2019 Sep 03

DOI: 10.23736/S0375-9393.19.13857-6

Article type: Letter to the Editor (Correspondence)

© 2019 EDIZIONI MINERVA MEDICA

Article first published online: September 03, 2019

Manuscript accepted: August 2, 2019

Manuscript revised: June 27, 2019

Manuscript received: May 6, 2019

Subscription: Information about subscribing to Minerva Medica journals is online at:

<http://www.minervamedica.it/en/how-to-order-journals.php>

Reprints and permissions: For information about reprints and permissions send an email to:

[journals.dept@minervamedica.it](mailto:journals.dept@minervamedica.it) - [journals2.dept@minervamedica.it](mailto:journals2.dept@minervamedica.it) - [journals6.dept@minervamedica.it](mailto:journals6.dept@minervamedica.it)

**Use of the CompuFlo® system to identify the epidural space in obstetric-gynecological area.  
A single-center retrospective study.**

**A. Di Filippo, U. Bitossi\*, V. Marcellino\*, V. Limatola\*, M. Sicurani\*, T. Borracci\*, M.  
Micaglio\***

University of Florence. Department of Health Sciences.

\*Azienda Ospedaliero Universitaria Careggi, Florence, Italy. Department of Anesthesia and Intensive Care. Unit of Obstetric and Gynecologic Anesthesia.

Key Words:

Regional Anesthesia; Epidural Space; Drug Delivery System

Dear Editor,

although epidural anesthesia (EA) is frequently used, the success rate is different and complications, as accidental dural puncture (ADP), may occur due to incorrect needle placement [1]. This is mainly observed with the commonly used technique of “Loss-Of-Resistance” at the pressure exerted by the anesthesiologist's thumb on the plunger of a syringe containing saline, or air, which is injected through an epidural “Tuohy” needle. Pressure control during injection of local anesthetics has recently been studied both on simulator and in large a review of the literature concerning peripheral blocks [2-3]. To limit complications, a new device is being proposed (CompuFlo<sup>®</sup> Epidural System, CompuFlo<sup>®</sup>, Milestone Scientific, Livingston, NJ) [4]. It is a computer-controlled drug-delivery system, able to measure the pressure at the tip of the needle during its progression through the tissues. The needle is connected via an arterial line extension tubing to a saline-filled syringe, whose plunger is controlled with real time feedback by CompuFlo<sup>®</sup>. Thanks to the fluid column through the epidural needle, CompuFlo<sup>®</sup> measures pressures obtained at the injection site four times a second, displaying a pressure/time curve on a dedicated digital monitor. During insertion, a continuous acoustic signal produced by CompuFlo<sup>®</sup> guides the operator too: when the needle runs into higher density tissues (i.e. in ligamenta) and the pressure at the injection site rises, the device produces a higher tone; when the pressure suddenly decreases (i.e. in the epidural space), a sudden fall in the tone of the audio output is clearly audible [5].

After the device has obtained the qualification at European level (CE), our Unit has inserted it into the clinical practice for difficult cases. After obtaining the approval of the local ethics committee, and registering in the National Clinical Trials registry, a retrospective study was conducted on the case series collected at the tertiary referral obstetrics gynecological center of Careggi University Hospital.

From January to December 2017, 141 consecutive women were recruited. In all patients a lumbar epidural catheter was placed using the CompuFlo<sup>®</sup> according to the following method: after identification of the interspinous space and local anesthesia of the skin, a Tuohy needle was inserted subcutaneously. The needle was connected to CompuFlo<sup>®</sup> via a dedicated tubing set (Epidural Disposable Kit, Milestone Scientific, Livingston, NJ), the procedure was started and the epidural needle was advanced slowly to obtain a continuous reading of the pressure in the crossed tissues, while a pressure/time curve was displayed on the screen. A sudden drop in pressure once the ligamentum flavum has been crossed, maintained for more than 5 seconds and the fall in the tone of

the acoustic signal emitted by the device, indicated the identification of the epidural space (Figure 1).

In 85% of cases, the epidural catheter was correctly positioned within the first two attempts. The average time of the procedure was 76.07 sec (range 48.9-98.15). A third attempt was necessary in 15% of cases. Of the 141 patients who underwent the catheter positioning using CompuFlo<sup>®</sup>, none developed complications. All the epidural blocks worked properly.

CompuFlo<sup>®</sup> epidural system seems to be effective and safe in identifying the epidural space, also in difficult cases, minimizing the incidence of adverse events. Our retrospective study needs to be validated by larger RCTs.

## References

1. Gaiser RR. Postdural puncture headache: a headache for the patient and a headache for the anesthesiologist. *Curr Opin Anaesthesiol* 2013;26(3):296–303
2. A New Pressure Guided Management Tool for Epidural Space Detection: Feasibility Assessment on a Simulator. Carassiti M, Mattei A, Quarta R, Massaroni C, Saccomandi P, Tesei M, Setola R, Schena E. *ArtifOrgans*. 2017;41:E320-E325.
3. Injection pressures measuring for a safe peripheral nerve block. Carassiti M, De Filippis A, Palermo P, Valenti C, Costa F, Massaroni C, Schena E. *Minerva Anesthesiol*. 2019 May 21. doi: 10.23736/S0375-9393.19.13518-3. [Epub ahead of print]
4. Ghelber O, Gebhard RE, Vora S, Hagberg CA, Szmuk P. Identification of the Epidural Space Using Pressure Measurement With the CompuFlo Injection Pump - A Pilot Study. *Reg Anesth Pain Med* 2008;33(4):346–52
5. Capogna G, Camorcia M, Coccoluto A, Micaglio M, Velardo M. Experimental validation of the CompuFlo<sup>®</sup> epidural controlled system to identify the epidural space and its clinical use in difficult obstetric cases. *Int J Obstet Anesth* 2018;36:28-33

Figure 1: Pressure/time curve at the tip of Tuohy needle and saline infused volume during epidural space identification registered using the CompuFlo<sup>®</sup> system.

