

1 **Anticoagulants used in plasma collection affect adipokine multiplexed**  
2 **measurements**

3

4 Alessandra Allione<sup>a,b,\*</sup>, Cornelia Di Gaetano<sup>a,b</sup>, Nadia Dani<sup>c</sup>, Davide Barberio<sup>c</sup>, Sabina  
5 Sieri<sup>d</sup>, Vittorio Krogh<sup>d</sup>, Giuseppe Matullo<sup>a,b</sup>

6

7 <sup>a</sup> Human Genetics Foundation- HuGeF, Via Nizza 52, 10126 Turin, Italy

8 <sup>b</sup> Department of Medical Sciences, University of Turin, Via Santena 19, 10100  
9 Turin, Italy

10 <sup>c</sup> Bioclarma Srl, Via Nizza 52, 10126 Turin, Italy

11 <sup>d</sup> IRCCS Foundation, Istituto Nazionale dei Tumori, Via Venezian, 1, 20133 Milan,  
12 Italy

13

14 \* Corresponding author:

15 Human Genetics Foundation – HuGeF, Via Nizza 52, 10126 Turin, Italy

16 E-mail: [alessandra.allione@hugef-torino.org](mailto:alessandra.allione@hugef-torino.org)

17

18

19 **Abstract**

20 Obesity is an important health problem worldwide. Adipose tissue acts as an  
21 endocrine organ that secretes various bioactive substances, called adipokines,  
22 including pro-inflammatory biomarkers such as TNF- $\alpha$ , IL-6, leptin and C-reactive  
23 protein (CRP) and anti-inflammatory molecules such as adiponectin. The  
24 deregulated production of adipokines in obesity is linked to the pathogenesis of  
25 various disease processes and monitoring their variation is critical to understand  
26 metabolic diseases.

27 The aim of this study was to determine the plasma concentration of adipokines in  
28 healthy subjects by multiplexed measurements and the effect of anticoagulants on  
29 their levels.

30 Plasma samples from 10 healthy donors were collected in two different  
31 anticoagulants (sodium citrate or heparin).

32 All markers, excluding TNF- $\alpha$ , showed significantly higher concentrations in  
33 heparinized compared to citrate plasma. However, levels of adipokines in different  
34 plasma samples highly correlated for most of these markers.

35 We reported that different anticoagulants used in the preparation of the plasma  
36 samples affected the measurements of some adipokines. The importance of the  
37 present results in epidemiology is relevant when comparing different studies in  
38 which blood samples were collected with different anticoagulants.

39

40 **Keywords:** adipokines, anticoagulants, cytokines, microbead assay, plasma

## 41 **1. Introduction**

42 Adipokines are defined as cell signaling mediators secreted by the adipose tissue.  
43 They have both pro-inflammatory and anti-inflammatory activities, and the  
44 imbalance between the different factors secreted by adipose tissue contributes to  
45 metabolic dysfunction. [1]. When adipocyte dysfunction is developed as a result of  
46 adipose tissue expansion, the deregulation of adipokine levels can produce several  
47 effects on inflammatory responses, thereby contributing to the initiation and  
48 progression of obesity-induced metabolic and cardiovascular complications [1].  
49 Therefore, further elucidation of the functions and mechanisms of key adipokines  
50 will lead to a better understanding of the pathogenesis of obesity-linked disorders.  
51 Adipokines are involved in the regulation of metabolism and insulin sensitivity.  
52 Moreover, inflammation and blood concentrations of various adipokines are  
53 associated with obesity, metabolic and cardiovascular diseases. It is, therefore,  
54 emerging that serum adipokine levels may serve as biomarkers of obesity-related  
55 illnesses [2].

56 Recent papers demonstrated the effect of anticoagulants on multiplexed  
57 measurement of cytokines/chemokines in healthy subjects [3-7], highlighting the  
58 importance of sample preparation in biomarkers measurements in plasma. The  
59 Luminex multiplex platform system is a highly efficient fluorescent (or magnetic)  
60 bead-based capture/detection sandwich immunoassay that allows for  
61 measurements of multiple analytes simultaneously in a single reaction with small  
62 sample volumes [8, 9]. This technology can measure up to 100 different analytes  
63 using as little as 50  $\mu$ l sample volume, making it an assay very useful in clinical  
64 trials or epidemiology studies, especially when volumes are limited.

65 The aim of this study was to evaluate the effect of two different anticoagulants  
66 (lithium heparin (LiEP) and sodium citrate (NaCitr)) on the levels of adipokines in  
67 plasma collected from 10 healthy subjects. Multiplex microbead immunoassay was  
68 performed for measuring IL-6, TNF- $\alpha$  and leptin while adiponectin and C-reactive  
69 protein (CRP) were measured as a single measurement.

70

## 71 **2. Materials and Methods**

### 72 **2.1 Samples**

73 Samples were obtained concurrently from 10 healthy donors. Twenty milliliters of  
74 peripheral venous blood were drawn into vacutainer tubes containing two  
75 different anticoagulants (LiEp or NaCitr). Samples were centrifuged immediately  
76 after the blood withdrawal, and plasma was analysed within 1 hour. Cells were  
77 removed from plasma by centrifugation for 15 minutes at 2,000 x g at room  
78 temperature. Written informed consent was achieved from all subjects. The study  
79 protocol was approved by the ethics committee of the Fondazione IRCCS (Istituto  
80 per la Ricerca e la Cura del Cancro, National Institute for Research and Treatment  
81 of Cancer) Istituto Nazionale dei Tumori (Milan, Italy).

82 Aliquots of the samples were stored at -80<sup>0</sup>C and analysed ten days later.

83

### 84 **2.2 Experimental measurements**

85 The adipokine analysis was performed with the Luminex technology, which  
86 combines the principle of a sandwich immunoassay with fluorescent bead-based  
87 technology. In this way, individual and multiplex analysis of different analytes in a  
88 single microwell plate are allowed [10]. Capture antibodies directed against the

89 biomarker of interest are covalently coupled to fluorescently dyed magnetic  
90 microspheres, each with a different color code or spectral address to allow  
91 discrimination of individual tests within a multiplex suspension. Coupled beads  
92 react with the sample containing the analyte of interest. After a series of washes to  
93 remove unbound protein, a biotinylated detection antibody is added to create a  
94 sandwich complex. The final detection complex is formed by the addition of  
95 streptavidin-phycoerythrin (SA-PE) conjugate. Phycoerythrin serves as a  
96 fluorescent indicator.

97 The assay for human IL-6, TNF- $\alpha$ , leptin, adiponectin (Bio-Rad Laboratories,  
98 Hercules California, USA) and CRP (Merk Millipore, Darmstadt, Germany) was  
99 carried out on fresh plasma samples using 96-well microplates accordingly to the  
100 recommendations of manufacturers. The contents of each well were moved into  
101 the Bio-Plex 100 System array reader (Bio-Rad Laboratories Bio-Rad Laboratories,  
102 Hercules California, USA), which identifies and quantifies each specific reaction  
103 based on bead color and fluorescent signal intensity. The data were finally  
104 processed using Bio-Plex Manager software (version 6.1) using five-parametric  
105 curve fitting and converted in pg/ml.

106

### 107 **2.3 Statistical analyses**

108 All results are showed as mean  $\pm$  standard deviation (SD). Data were compared by  
109 nonparametric analyses with Wilcoxon's matched pairs test, in which the median  
110 was used to calculate significant differences. Spearman correlation coefficients  
111 were calculated to investigate the correlation between measurements obtained  
112 with the two anticoagulants. All statistical analyses were performed using the

113 statistical software GraphPad Prism 5, version 5.04 (GraphPad Software, San  
114 Diego, CA).

115 When measurements were below the lower limit of detection (LLD, <5% of  
116 measurements), we assigned a value equal to the midpoint between the LLD and  
117 zero.

118

### 119 **3. Results and Discussion**

120 Significant differences in the measurements of all adipokines, except TNF- $\alpha$ , were  
121 observed in plasma samples collected in different anticoagulants (Fig. 1). IL-6,  
122 leptin, adiponectin and CRP showed a significantly higher concentration in heparin  
123 plasma compared to citrate.

124 However, despite these differences in absolute levels, the measurements of all  
125 adipokines were highly correlated each other, except for TNF- $\alpha$  (Fig. 2). We can  
126 speculate that, even though the measured levels differ, the same variation between  
127 individuals can be detected, as demonstrated by the significant correlations.

128 TNF- $\alpha$  measurements differed from the other mediators both with regard to the  
129 effect of anticoagulant and the lack of correlation between the levels in the paired  
130 samples. These differences could be due to the very low levels, quite undetectable,  
131 of this cytokine in our samples. To explain these results we performed also an  
132 ELISA assay for TNF- $\alpha$  on the same samples: by this way we confirmed the absence  
133 of significant differences in the levels of TNF- $\alpha$  in plasma samples obtained using  
134 different anticoagulants. On the contrary, correlation analysis of ELISA results  
135 obtained with the two anticoagulants was statistically significant ( $r = 0.811$ ;  $p =$

136 0.004), confirming the results obtained with the other adipokines (data not  
137 shown).

138 We carried out this study with explorative purposes only on ten subjects, with the  
139 aim to understand if the cross-comparison of levels or profiles of adipokines  
140 performed in different anticoagulants (e.g. results collected from various studies)  
141 could be performed.

142 Adipokines are involved in the regulation of metabolism, insulin sensitivity, and  
143 inflammation and serum concentrations of various adipokines are associated with  
144 obesity, metabolic and cardiovascular diseases [1, 2]. It has been therefore  
145 hypothesized that serum adipokine levels may serve as predictors of obesity-  
146 related diseases or the individual disease outcomes [11].

147 Our study focused the attention on the measurement of five adipokines.  
148 Anticoagulants effect on IL-6 and TNF- $\alpha$  measurements has been already described  
149 in recent papers [3-6], while only a few of them described the effect on leptin [3, 5]  
150 or CRP [5], and none, to our knowledge, reported anticoagulants effect on  
151 adiponectin measurement. However, the importance of such a study is underlined  
152 by the proposal of using adiponectin as a clinical biomarker for several diseases  
153 [12]. Lifestyle modification with visceral fat reduction combined with targeted  
154 therapeutic interventions designed to improve adiponectinemia seem to be  
155 potential clinically useful strategies to prevent obesity-related diseases, including  
156 type 2 diabetes, cardiovascular diseases and malignancies [12].

157 Significant differences observed between different plasma samples in IL-6, leptin  
158 and CRP levels were not in agreement with previous data reported by Biancotto et  
159 al. [5], however, we measured analyte levels in fresh sample while they performed

160 the assay on frozen specimens. When we did the same test on frozen samples,  
161 statistically significant differences between plasma samples (LiEp vs. NaCitr) were  
162 no more observed for IL-6 measurements but were still significant for leptin  
163 measurements ( $p=0.002$ , Supplementary Materials Fig. S1). However, we observed  
164 a significant difference between fresh and frozen plasma samples with Li Ep for IL-  
165 6 measurements ( $p<0.001$ , Fig. S1) and between fresh and frozen plasma samples  
166 both with LiEp and NaCitr for leptin measurements ( $p=0.002$ , Fig. S1). Moreover,  
167 these statistically significant results always showed a significant correlation (IL-6  
168 LiEp fresh vs. frozen  $r=0.988$ ,  $p<0.0001$ ; Leptin LiEp fresh vs. frozen  $r=1$ ,  
169  $p<0.0001$ ; Leptin NaCitr fresh vs. frozen  $r=0.939$ ,  $p<0.002$ ; Supplementary  
170 Materials Fig. S2). Previous papers on anticoagulant effect on cytokine  
171 measurements often showed undetectable levels of IL-6, TNF- $\alpha$  or leptin, probably  
172 because samples were stored at  $-80^{\circ}\text{C}$  for a different time and affected by multiple  
173 freeze-thawing cycles [3, 5, 6].

174 We dedicated our attention only at plasma samples, but differences between  
175 plasma and serum cytokine measurements have been also described. However,  
176 results are often discordant: for many analytes, high correlation were observed  
177 between serum and plasma levels even when absolute levels differed. In fact, the  
178 removal of fibrinogen, platelets and other circulating proteins from the plasma  
179 during serum preparation could influence the presence or detection of an analyte.  
180 Also during the cascade of coagulation, activation of cellular elements can release  
181 inflammatory mediators, that may affect cytokine levels. This kind of information  
182 are available for IL-6 and TNF- $\alpha$  [3-6], leptin [3, 5] or CRP [5], but are not available  
183 for adiponectin. Possible differences between serum and plasma samples should



184 be taken into consideration also when measuring adipokine levels and further  
185 evaluations are needed.

186 We used Luminex multiplex bead-based technology, with the antibody kits  
187 purchased from Bio-Rad or Merck (see Section 2). Previous studies compared the  
188 performance of multiplex microbead assays between various vendors, and it was  
189 noted that the consistency between the vendors was high only for some analytes  
190 [13, 14]. Though our conclusions are broadly applicable, specific care must be  
191 taken when other kits or suppliers are used.

192

#### 193 4. **Conclusions**

194 Our results emphasize, as already demonstrated for cytokines/chemokines [3-7],  
195 the need to consider sample collection methods carefully when a study requires  
196 measurement of adipokines in the peripheral blood. Cross comparison of levels or  
197 profiles of adipokines performed in different anticoagulants should be avoided.  
198 However, we observed a correlation between levels in plasma samples obtained  
199 with different anticoagulants, and a possible solution might be the normalization  
200 of values from one type of samples to an alternative type. Despite this possibility,  
201 Jung and Wu demonstrated that this kind of conversion is complicated, and  
202 calculations based on correlation analyses alone are not sufficient [15].

203

#### 204 **Acknowledgements**

205 This work was supported by grants from the Associazione Italiana per la Ricerca  
206 sul Cancro (project number IG13273, V.K.). Partial funding for this project has also  
207 been received from the Compagnia di San Paolo ([www.compagniadisanpaolo.it](http://www.compagniadisanpaolo.it))

208 (G.M.) and by the Human Genetics Foundation-Torino ([www.hugef-torino.org](http://www.hugef-torino.org))

209 (G.M.).

210

## 211 **References**

212 [1] N. Ouchi, J.L. Parker, J.J. Lugus, K. Walsh, Adipokines in inflammation and  
213 metabolic disease, *Nat Rev Immunol*, 11 (2011) 85-97.

214 [2] M. Bluher, C.S. Mantzoros, From leptin to other adipokines in health and  
215 disease: facts and expectations at the beginning of the 21st century, *Metabolism*,  
216 64 (2015) 131-145.

217 [3] T.H. Tvedt, K.P. Rye, H. Reikvam, A.K. Brenner, O. Bruserud, The importance of  
218 sample collection when using single cytokine levels and systemic cytokine profiles  
219 as biomarkers--a comparative study of serum versus plasma samples, *J Immunol*  
220 *Methods*, 418 (2015) 19-28.

221 [4] V.V. Krishnan, R. Ravindran, T. Wun, P.A. Luciw, I.H. Khan, K. Janatpour,  
222 Multiplexed measurements of immunomodulator levels in peripheral blood of  
223 healthy subjects: Effects of analytical variables based on anticoagulants, age, and  
224 gender, *Cytometry B Clin Cytom*, 86 (2014) 426-435.

225 [5] A. Biancotto, X. Feng, M. Langweiler, N.S. Young, J.P. McCoy, Effect of  
226 anticoagulants on multiplexed measurement of cytokine/chemokines in healthy  
227 subjects, *Cytokine*, 60 (2012) 438-446.

228 [6] W. de Jager, K. Bourcier, G.T. Rijkers, B.J. Prakken, V. Seyfert-Margolis,  
229 Prerequisites for cytokine measurements in clinical trials with multiplex  
230 immunoassays, *BMC Immunol*, 10 (2009) 52.

231 [7] H.L. Wong, R.M. Pfeiffer, T.R. Fears, R. Vermeulen, S. Ji, C.S. Rabkin,  
232 Reproducibility and correlations of multiplex cytokine levels in asymptomatic  
233 persons, *Cancer Epidemiol Biomarkers Prev*, 17 (2008) 3450-3456.

234 [8] C. Belabani, S. Rajasekharan, V. Poupon, T. Johnson, A. Bar-Or, A condensed  
235 performance-validation strategy for multiplex detection kits used in studies of  
236 human clinical samples, *J Immunol Methods*, 387 (2013) 1-10.

237 [9] K. Skogstrand, Multiplex assays of inflammatory markers, a description of  
238 methods and discussion of precautions - Our experience through the last ten years,  
239 *Methods*, 56 (2012) 204-212.

240 [10] D.A. Vignali, Multiplexed particle-based flow cytometric assays, *J Immunol*  
241 *Methods*, 243 (2000) 243-255.

242 [11] L.F. Van Gaal, I.L. Mertens, C.E. De Block, Mechanisms linking obesity with  
243 cardiovascular disease, *Nature*, 444 (2006) 875-880.

244 [12] K. Kishida, T. Funahashi, I. Shimomura, Adiponectin as a routine clinical  
245 biomarker, *Best Pract Res Clin Endocrinol Metab*, 28 (2014) 119-130.

246 [13] S.S. Khan, M.S. Smith, D. Reda, A.F. Suffredini, J.P. McCoy, Jr., Multiplex bead  
247 array assays for detection of soluble cytokines: comparisons of sensitivity and  
248 quantitative values among kits from multiple manufacturers, *Cytometry B Clin*  
249 *Cytom*, 61 (2004) 35-39.

250 [14] A.M. Dupuy, N. Kuster, G. Lizard, K. Ragot, S. Lehmann, B. Gallix, J.P. Cristol,  
251 Performance evaluation of human cytokines profiles obtained by various

252 multiplexed-based technologies underlines a need for standardization, Clin Chem  
253 Lab Med, 51 (2013) 1385-1393.  
254 [15] K. Jung, C.Y. Wu, Methodological weakness in using correlation coefficients for  
255 assessing the interchangeability of analyte data between samples collected under  
256 different sampling conditions--the example of matrix metalloproteinase 9  
257 determined in serum and plasma samples, Clin Chem Lab Med, 48 (2010) 733-736.  
258