



Aalborg Universitet

AALBORG UNIVERSITY
DENMARK

Is Benign Paroxysmal Positional Vertigo treated successfully with the TRV reposition chair?

Pedersen, Morten Falshøj; Abrahamsen, Emil Riis; Eriksen, Helle Højmark; Hougaard, Dan Dupont

Publication date:
2018

Document Version
Publisher's PDF, also known as Version of record

[Link to publication from Aalborg University](#)

Citation for published version (APA):
Pedersen, M. F., Abrahamsen, E. R., Eriksen, H. H., & Hougaard, D. D. (2018). *Is Benign Paroxysmal Positional Vertigo treated successfully with the TRV reposition chair?*. Abstract from 6th Dubai Otology Conference, Dubai, United Arab Emirates.

General rights

Copyright and moral rights for the publications made accessible in the public portal are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

- ? Users may download and print one copy of any publication from the public portal for the purpose of private study or research.
- ? You may not further distribute the material or use it for any profit-making activity or commercial gain
- ? You may freely distribute the URL identifying the publication in the public portal ?

Take down policy

If you believe that this document breaches copyright please contact us at vbn@aub.aau.dk providing details, and we will remove access to the work immediately and investigate your claim.

Submission group: Oral presentation

Submission subgroup: Vestibular/Balance

Submission name: Dan Dupont Hougaard

Presentation category: BPPV

Presentation Title: Is Benign Paroxysmal Positional Vertigo treated successfully with the TRV reposition chair?

Authors:

Pedersen MF^{1,2}, Abrahamsen ER^{1,2}, Eriksen HH³, Hougaard DD^{1,2}

¹Department of Otolaryngology, Head & Neck Surgery and Audiology, Aalborg University Hospital, Aalborg, Denmark.

²Department of Clinical Medicine, Aalborg University, Aalborg, Denmark

³Unit of Clinical Biostatistics and Bioinformatics, Aalborg University Hospital, Aalborg, Denmark

Abstract:

Primary objective was to evaluate how successful reposition of BPPV was with the TRV reposition chair. Designed as a prospective clinical trial. Seventy-five patients were diagnosed and treated in the TRV reposition chair. Overall, patients were successfully treated after an average of 2.23 (\pm 1.66 SD) treatments in the TRV-chair. There was a significant difference between the numbers of treatments needed for a single semicircular canal (SCC) affection (average 1.8) compared to multi SCC affection (average 3.13). We found the TRV reposition chair to be very successful for treatment of patients with retractable as well as atypical BPPV.

Introduction:

Benign Paroxysmal Positional Vertigo (BPPV) is a very common inner ear disease. Despite the fact that the majority of these patients are successfully treated by manual repositioning maneuvers, still a large number of patients experience treatment failures. Therefore, new therapeutic modalities have been proposed, among these the biaxial chairs like the TRV reposition chair. The primary objective of this study was to evaluate how successful reposition of BPPV was when treating patients with the TRV reposition chair.

Materials and Methods:

Designed as a prospective clinical trial. Seventy-five patients, referred to the Tertiary ENT Dizziness Clinic at Aalborg University Hospital, were included following a diagnosis of BPPV. All patients were diagnosed and treated in the TRV reposition chair. Video Frenzel goggles were used to enable both precise diagnostics and optimal treatment.

Results:

Overall, patients were successfully treated after an average of 2.23 (\pm 1.66 SD) treatments in the TRV-chair. There was a significant difference between the numbers of treatments needed for a single semicircular canal (SCC) affection (average 1.8) compared to multi SCC affection (average 3.13).

Conclusions:

We found the TRV reposition chair to be very successful for treatment of patients with retractable as well as atypical BPPV.